

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Vermont's Original, LLC

SDG No.: P4432

Contract: DAIR01

Lab Code: CHEM

Case No.: P4432

SAS No.: P4432

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164220BS Lead	mg/Kg	50.0	46.1		92	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Vermont's Original, LLC

SDG No.: P4432

Contract: DAIR01

Lab Code: CHEM

Case No.: P4432

SAS No.: P4432

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164258BS Mercury	mg/Kg	0.26	0.26		99	80 - 120	CV