

Surrogate Summary

SW-846

SDG No.: P4432

Client: Vermont's Original, LLC

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P4432-01	LOT-4078A	2-Methylnaphthalene-d10	0.4	0.16	40		17	161
		Fluoranthene-d10	0.4	0.16	40		23	138
		Nitrobenzene-d5	0.4	0.32	80		33	121
		2-Fluorobiphenyl	0.4	0.20	50		32	121
		Terphenyl-d14	0.4	0.20	50		21	130
PB164230BL	PB164230BL	2-Methylnaphthalene-d10	0.4	0.33	83		17	161
		Fluoranthene-d10	0.4	0.33	83		23	138
		Nitrobenzene-d5	0.4	0.34	85		33	121
		2-Fluorobiphenyl	0.4	0.37	93		32	121
		Terphenyl-d14	0.4	0.38	96		21	130
PB164230BS	PB164230BS	2-Methylnaphthalene-d10	0.4	0.32	81		17	161
		Fluoranthene-d10	0.4	0.31	77		23	138
		Nitrobenzene-d5	0.4	0.34	84		33	121
		2-Fluorobiphenyl	0.4	0.36	90		32	121
		Terphenyl-d14	0.4	0.40	100		21	130