

FORM 10B-IN
ICP-AES INTERELEMENT CORRECTION FACTORS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51682 MA No.: SDG No.: MJHJ00
Instrument ID: P5 Date : 03/15/2024

Analyte	Wave- Length (nm)	Interelement Correction Factors For				
		Cr	Cu	Mn	Mo	Ni
Arsenic	189.042	0	0	0	0	0
Lead	220.353	0	0.0014	0.00034	-0.00148	0.00063

FORM 10B-IN
ICP-AES INTERELEMENT CORRECTION FACTORS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51682 MA No.: SDG No.: MJHJ00
Instrument ID: P5 Date : 03/15/2024

Analyte	Wave- Length (nm)	Interelement Correction Factors For				
		Pb	Ti	V	Zn	
Arsenic	189.042	0	0	0	0	
Lead	220.353	0	0	0	0	

FORM 10B-IN
ICP-AES INTERELEMENT CORRECTION FACTORS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51682 MA No.: SDG No.: MJHJ00
Instrument ID: P5 Date : 03/15/2024

Analyte	Wave- Length (nm)	Interelement Correction Factors For				
Arsenic	189.042	0	0	0	0	
Lead	220.353	0	0	0	0	