

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51682 MA No.: _____ SDG No.: MJHJ00
Instrument ID: P5 Start Date: 10/24/2024
Analytical Method: ICP-AES Run Batch: LB133118
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	0	0.0	0	10	7.8	-22	1250	1200	-4
Lead	0	-0	0	10	13	30	6250	6400	2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51682 MA No.: SDG No.: MJHJ00
Instrument ID: P5 Start Date: 10/24/2024
Analytical Method: ICP-AES Run Batch: LB133118
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	2500	2500	0	5000	5000	0	10000	10000	0
Lead	12500	13000	4	25000	25000	0	50000	49000	-2