

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51682 MA No. : SDG No.: MJHJ00Initial Calibration Verification Source : MP82450Continuing Calibration Verification Source : MP82453Run Batch: LB133118 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV010				ID: CCV072				ID: CCV073		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic	1000	1000	104	0	5000	5000	100	0	5000	100	0
Lead	1000	990	99	0	25000	24000	95	0	24000	94	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51682 MA No. : SDG No.: MJHJ00
Initial Calibration Verification Source : MP82450
Continuing Calibration Verification Source : MP82453
Run Batch: LB133118 Analytical Method: ICP-AES
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV074				ID: CCV075		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic					5000	5100	101	1	5200	105	0
Lead					25000	24000	95	0	24000	95	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51682 MA No. : SDG No.: MJHJ00
 Initial Calibration Verification Source : MP82450
 Continuing Calibration Verification Source : MP82453
 Run Batch: LB133118 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV076				ID: CCV077		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic					5000	5300	106	0	5300	106	0
Lead					25000	24000	96	0	24000	95	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51682 MA No. : SDG No.: MJHJ00
Initial Calibration Verification Source : MP82450
Continuing Calibration Verification Source : MP82453
Run Batch: LB133118 Analytical Method: ICP-AES
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV078				ID: CCV079		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic					5000	5300	107	0	5100	102	0
Lead					25000	24000	96	0	24000	94	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51682 MA No. : SDG No.: MJHJ00
 Initial Calibration Verification Source : MP82450
 Continuing Calibration Verification Source : MP82453
 Run Batch: LB133118 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV080				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic					5000	4600	91	0			
Lead					25000	23000	94	0			