

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4438 SAS No.: P4438 SDG No.: P4438

Instrument ID: BNA_N Calibration Date/Time: 10/25/2024 20:41

Lab File ID: BN034732.D Init. Calib. Date(s): 10/24/2024 10/24/2024

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 09:11 12:48

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.550	0.499		-9.3	50.0
Fluoranthene-d10	0.901	0.778		-13.7	50.0
2-Fluorophenol	1.189	1.284		8.0	50.0
Phenol-d6	1.568	1.673		6.7	50.0
Nitrobenzene-d5	0.336	0.311		-7.4	50.0
2-Fluorobiphenyl	1.590	1.375		-13.5	50.0
2,4,6-Tribromophenol	0.124	0.129		4.0	50.0
Terphenyl-d14	0.817	0.831		1.7	50.0
1,4-Dioxane	0.532	0.481		-9.6	50.0

All other compounds must meet a minimum RRF of 0.010.