

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P4438 SAS No.: P4438 SDG NO.: P4438
Lab File ID: VN084428.D Date Analyzed: 10/21/2024
Instrument ID: MSVOA_N Time Analyzed: 11:38
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	127638	13.788				
UPPER LIMIT	255276	14.288				
LOWER LIMIT	63819	13.288				
EPA SAMPLE NO.						
BP-VPB-190-TB-20241015	110339	13.79				
BP-VPB-190-EB-20241016	102477	13.79				
VPB190-HYD-20241016	104212	13.79				
BP-VPB-190-GW-58-60	106943	13.79				
BP-VPB-190-GW-98-100	102361	13.79				
BP-VPB-190-GW-158-160	98336	13.79				
BP-VPB-190-GW-198-200	95771	13.79				
VN1021WBL01	99954	13.79				
VN1021WBS01	138325	13.79				
VN1021WBSD01	124118	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.