

DATA PACKAGE
SEMI-VOLATILE ORGANICS

PROJECT NAME : CTO WE13

TETRA TECH NUS, INC.
661 Andersen Drive
Suite 200
Pittsburgh, PA - 15220-2745
Phone No: 412-921-7090

ORDER ID : P4440
ATTENTION : Ernie Wu



Laboratory Certification ID # 20012



1) SEMI-VOLATILE DATA	2
2) Signature Page	4
3) Case Narrative	5
4) Qualifier Page	7
5) Conformance/Non Conformance	8
6) QA Checklist	10
7) Chronicle	11
8) Hit Summary	12
9) QC Data Summary For SVOC-SIMGroup1	13
9.1) Deuterated Monitoring Compound Summary	14
9.2) LCS/LCSD Summary	15
9.3) Method Blank Summary	17
9.4) GS/MS Tune Summary	18
9.5) Internal Standard Area and RT Summary	21
10) Sample Data	25
10.0.1) BPOW6-11-HYD-20241016	26
10.0.2) BPOW6-11-HYD-20241016RE	37
11) Calibration Data Summary	48
11.1) Initial Calibration Data	3
11.1.1) BN102424	49
11.2) Continued Calibration Data	236
11.2.1) BN034701.D	236
11.2.2) BN034719.D	262
11.2.3) BN034721.D	288
11.2.4) BN034732.D	314
12) QC Sample Data	340
12.1) Tune Raw Data	341
12.2) Method Blank Data	353
12.3) LCS Data	363
12.4) LCSD Data	387
13) Manual Integration	411
14) Analytical Runlogs	413
15) Extraction Logs	420
15.1) PB164282.pdf	420
15.2) PB164282IC.pdf	422

Table Of Contents for P4440

16) Standard Prep Logs	423
17) Shipping Document	490
17.1) Chain Of Custody	491
17.2) Lab Certificate	492
17.3) Internal COC	493
18) Not Reviewed Data	494

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Cover Page

Order ID : P4440

Project ID : CTO WE13

Client : Tetra Tech NUS, Inc.

Lab Sample Number

P4440-01
P4440-02

Client Sample Number

BPOW6-11-HYD-20241016
BPOW6-11-TB-20241016

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : N. N. Pandya

NYDOH CERTIFICATION NO - 11376



NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: CTO WE13

Project Manager : Ernie Wu

Chemtech Project # P4440

Test Name: SVOC-SIMGroup1

A. Number of Samples and Date of Receipt:

2 Water samples were received on 10/18/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1 and VOCMS Group1. This data package contains results for SVOC-SIMGroup1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_N using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOC-SIMGroup1 was based on method 8270-Modified and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for BPOW6-11-HYD-20241016 [2-Methylnaphthalene-d10 - 23%,], BPOW6-11-HYD-20241016RE [2-Methylnaphthalene-d10 - 24%], Failure sample for surrogate was reanalyzed to confirm the failure and both run were reported in Hard Copy.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration File ID BN034701.D met the requirements except for 2,4,6-Tribromophenol, The failure compound not associated with the client parameters list, therefore no corrective action was taken.

The Tuning criteria met requirements.

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).”

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

N. N. Pandya

APPROVED

Nimisha Pandya, QA/QC Supervisor , 10/28/2024, 12:44:42 PM

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: P4440

MATRIX: Water

METHOD: 8270-Modified/3510

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements . The Continuous Calibration File ID BN034701.D met the requirements except for 2,4,6-Tribromophenol, The failure compound not associated with the client parameters list, therefore no corrective action was taken.			✓
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges. The Surrogate recoveries met the acceptable criteria except for BPOW6-11-HYD- 20241016 [2-Methylnaphthalene-d10 - 23%,], BPOW6-11-HYD-20241016RE [2- Methylnaphthalene-d10 - 24%], Failure sample for surrogate was reanalyzed to confirm the failure and both run were reported in Hard Copy.		✓	
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The Blank Spike met requirements for all samples . The Blank Spike Duplicate met requirements for all samples .			✓

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

(CONTINUED)

		NA	NO	YES
9.	Internal Standard Area/Retention Time Shift Meet Criteria			✓
	Comments:			
10.	Extraction Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			
11.	Analysis Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)."

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

____ S. M. Jodhani _____
QA REVIEW

REVIEWED

By Sohil Jodhani, QA/QC Director at 10:36 am, Oct 28, 2024

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: P4440

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 10/28/2024

LAB CHRONICLE

OrderID:	P4440	OrderDate:	10/18/2024 10:22:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	CTO WE13
Contact:	Ernie Wu	Location:	K51,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P4440-01	BPOW6-11-HYD-2024 1016	Water			10/16/24			10/18/24
			SVOC-SIMGroup1	8270-Modified		10/21/24	10/25/24	
P4440-01RE	BPOW6-11-HYD-2024 1016RE	Water			10/16/24			10/18/24
			SVOC-SIMGroup1	8270-Modified		10/21/24	10/25/24	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet SW-846

SDG No.: P4440
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID :				0.000					
Total Svoc :					0.00				
Total Concentration:					0.00				



QC SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Surrogate Summary

SW-846

SDG No.: P4440

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P4440-01	BPOW6-11-HYD-20241016	2-Methylnaphthalene-d10	0.4	0.091	23	*	30	150
		Fluoranthene-d10	0.4	0.35	86		30	150
		Nitrobenzene-d5	0.4	0.31	78		55	111
		2-Fluorobiphenyl	0.4	0.34	84		53	106
		Terphenyl-d14	0.4	0.38	96		58	132
P4440-01RE	BPOW6-11-HYD-20241016RE	2-Methylnaphthalene-d10	0.4	0.095	24	*	30	150
		Fluoranthene-d10	0.4	0.37	93		30	150
		Nitrobenzene-d5	0.4	0.32	81		55	111
		2-Fluorobiphenyl	0.4	0.33	83		53	106
		Terphenyl-d14	0.4	0.39	98		58	132
PB164282BL	PB164282BL	2-Methylnaphthalene-d10	0.4	0.33	83		30	150
		Fluoranthene-d10	0.4	0.34	86		30	150
		Nitrobenzene-d5	0.4	0.34	85		55	111
		2-Fluorobiphenyl	0.4	0.37	91		53	106
		Terphenyl-d14	0.4	0.39	97		58	132
PB164282BS	PB164282BS	2-Methylnaphthalene-d10	0.4	0.50	125		30	150
		Fluoranthene-d10	0.4	0.34	84		30	150
		Nitrobenzene-d5	0.4	0.34	84		55	111
		2-Fluorobiphenyl	0.4	0.36	90		53	106
		Terphenyl-d14	0.4	0.36	91		58	132
PB164282BSD	PB164282BSD	2-Methylnaphthalene-d10	0.4	0.50	125		30	150
		Fluoranthene-d10	0.4	0.33	83		30	150
		Nitrobenzene-d5	0.4	0.34	84		55	111
		2-Fluorobiphenyl	0.4	0.36	89		53	106
		Terphenyl-d14	0.4	0.37	93		58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4440

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified DataFile: BN034706.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	
								Qual		High	RPD
PB164282BS	1,4-Dioxane	0.4	0.30	ug/L	75				70	130	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4440

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified DataFile: BN034711.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB164282BSD	1,4-Dioxane	0.4	0.31	ug/L	78	3			70	130	20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB164282BL

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: P4440

SAS No.: P4440 SDG NO.: P4440

Lab File ID: BN034705.D

Lab Sample ID: PB164282BL

Instrument ID: BNA_N

Date Extracted: 10/21/2024

Matrix: (soil/water) Water

Date Analyzed: 10/24/2024

Level: (low/med) LOW

Time Analyzed: 23:26

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB164282BS	PB164282BS	BN034706.D	10/25/2024
PB164282BSD	PB164282BSD	BN034711.D	10/25/2024
BPOW6-11-HYD-20241016	P4440-01	BN034717.D	10/25/2024

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: TETR06Lab Code: CHEMSAS No.: P4440 SDG NO.: P4440Lab File ID: BN034683.DDFTPP Injection Date: 10/24/2024Instrument ID: BNA_NDFTPP Injection Time: 07:59

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	66.1
68	Less than 2.0% of mass 69	0.9 (1.7) 1
69	Mass 69 relative abundance	54.4
70	Less than 2.0% of mass 69	0.3 (0.5) 1
127	10.0 - 80.0% of mass 198	60.3
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.5
275	10.0 - 60.0% of mass 198	20.6
365	Greater than 1% of mass 198	2.5
441	Present, but less than mass 443	7.8
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	9.1 (18.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN034684.D	10/24/2024	09:11
SSTDICC0.2	SSTDICC0.2	BN034685.D	10/24/2024	09:47
SSTDICCC0.4	SSTDICCC0.4	BN034686.D	10/24/2024	10:23
SSTDICC0.8	SSTDICC0.8	BN034687.D	10/24/2024	10:59
SSTDICC1.6	SSTDICC1.6	BN034688.D	10/24/2024	11:35
SSTDICC3.2	SSTDICC3.2	BN034689.D	10/24/2024	12:11
SSTDICC5.0	SSTDICC5.0	BN034690.D	10/24/2024	12:48
SSTDCCC0.4EC	SSTDCCC0.4	BN034699.D	10/24/2024	19:06



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: TETR06Lab Code: CHEMSAS No.: P4440SDG NO.: P4440Lab File ID: BN034700.DDFTPP Injection Date: 10/24/2024Instrument ID: BNA_NDFTPP Injection Time: 20:22

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	73.1
68	Less than 2.0% of mass 69	0.9 (1.6) 1
69	Mass 69 relative abundance	59.5
70	Less than 2.0% of mass 69	0.3 (0.5) 1
127	10.0 - 80.0% of mass 198	65.2
197	Less than 2.0% of mass 198	0.5
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.5
275	10.0 - 60.0% of mass 198	20.7
365	Greater than 1% of mass 198	2.6
441	Present, but less than mass 443	7.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	9.2 (21.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN034701.D	10/24/2024	21:01
PB164282BL	PB164282BL	BN034705.D	10/24/2024	23:26
PB164282BS	PB164282BS	BN034706.D	10/25/2024	00:02
PB164282BSD	PB164282BSD	BN034711.D	10/25/2024	03:02
BPOW6-11-HYD-20241016	P4440-01	BN034717.D	10/25/2024	06:39
SSTDCCC0.4EC	SSTDCCC0.4	BN034719.D	10/25/2024	07:51



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: TETR06Lab Code: CHEMSAS No.: P4440 SDG NO.: P4440Lab File ID: BN034720.DDFTPP Injection Date: 10/25/2024Instrument ID: BNA_NDFTPP Injection Time: 09:07

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	66.6
68	Less than 2.0% of mass 69	0.9 (1.6) 1
69	Mass 69 relative abundance	56.4
70	Less than 2.0% of mass 69	0.2 (0.4) 1
127	10.0 - 80.0% of mass 198	61.5
197	Less than 2.0% of mass 198	0.3
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	21
365	Greater than 1% of mass 198	2.8
441	Present, but less than mass 443	7.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	9.5 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN034721.D	10/25/2024	09:46
BPOW6-11-HYD-20241016RE	P4440-01RE	BN034724.D	10/25/2024	11:34
SSTDCCC0.4EC	SSTDCCC0.4	BN034732.D	10/25/2024	20:41

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG NO.: P4440

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 10/24/2024

Lab File ID: BN034701.D Time Analyzed: 21:01

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	5885	7.712	17350	10.48	8078	14.33
UPPER LIMIT	11770	8.212	34700	10.979	16156	14.825
LOWER LIMIT	2942.5	7.212	8675	9.979	4039	13.825
EPA SAMPLE NO.						
01 BPOW6-11-HYD-20241016	5535	7.71	15645	10.48	6994	14.32
02 PB164282BL	6454	7.71	17746	10.48	7655	14.33
03 PB164282BS	6277	7.71	17435	10.48	7557	14.32
04 PB164282BSD	5974	7.71	16564	10.48	7143	14.33

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG NO.: P4440

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 10/24/2024

Lab File ID: BN034701.D Time Analyzed: 21:01

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	14577	17.064	8841	21.248	8336	23.464
UPPER LIMIT	29154	17.564	17682	21.748	16672	23.964
LOWER LIMIT	7288.5	16.564	4420.5	20.748	4168	22.964
EPA SAMPLE NO.						
01 BPOW6-11-HYD-20241016	12994	17.06	7954	21.24	7258	23.46
02 PB164282BL	13670	17.07	7475	21.24	6736	23.46
03 PB164282BS	13505	17.06	7661	21.24	6455	23.46
04 PB164282BSD	13002	17.06	7027	21.24	5873	23.46

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG NO.: P4440

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 10/25/2024

Lab File ID: BN034721.D Time Analyzed: 09:46

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	6565	7.705	19566	10.48	9604	14.33
UPPER LIMIT	13130	8.205	39132	10.979	19208	14.825
LOWER LIMIT	3282.5	7.205	9783	9.979	4802	13.825
EPA SAMPLE NO.						
01 BPOW6-11-HYD-20241016RE	6120	7.71	17531	10.48	8276	14.32

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG NO.: P4440

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 10/25/2024

Lab File ID: BN034721.D Time Analyzed: 09:46

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	17577	17.064	13068	21.248	14018	23.459
UPPER LIMIT	35154	17.564	26136	21.748	28036	23.959
LOWER LIMIT	8788.5	16.564	6534	20.748	7009	22.959
EPA SAMPLE NO.						
01 BPOW6-11-HYD-20241016RE	15513	17.06	11157	21.24	11491	23.46

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



SAMPLE DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/16/24	
Project:	CTO WE13		Date Received:	10/18/24	
Client Sample ID:	BPOW6-11-HYD-20241016		SDG No.:	P4440	
Lab Sample ID:	P4440-01		Matrix:	Water	
Analytical Method:	SW8270SIM		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN034717.D	1	10/21/24 08:45	10/25/24 06:39	PB164282

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.091	*	30 - 150		23%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.35		30 - 150		86%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.31		55 - 111		78%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.34		53 - 106		84%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.38		58 - 132		96%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	5540	7.712				
1146-65-2	Naphthalene-d8	15600	10.479				
15067-26-2	Acenaphthene-d10	6990	14.318				
1517-22-2	Phenanthrene-d10	13000	17.057				
1719-03-5	Chrysene-d12	7950	21.241				
1520-96-3	Perylene-d12	7260	23.458				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034717.D
Acq On : 25 Oct 2024 06:39
Operator : RC/JU
Sample : P4440-01
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016

Quant Time: Oct 25 07:09:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.712	152	5535	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	15645	0.400	ng	0.00
13) Acenaphthene-d10	14.318	164	6994	0.400	ng	-0.01
19) Phenanthrene-d10	17.057	188	12994	0.400	ng	#-0.01
29) Chrysene-d12	21.241	240	7954	0.400	ng	# 0.00
35) Perylene-d12	23.458	264	7258	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.322	112	400	0.024	ng	0.00
5) Phenol-d6	6.882	99	94	0.004	ng	0.00
8) Nitrobenzene-d5	8.845	82	4108	0.313	ng	0.00
11) 2-Methylnaphthalene-d10	12.067	152	1956	0.091	ng	0.00
14) 2,4,6-Tribromophenol	15.815	330	370	0.171	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	9314	0.335	ng	0.00
27) Fluoranthene-d10	19.094	212	10124	0.346	ng	0.00
31) Terphenyl-d14	19.698	244	6233	0.383	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.187	149	1931	0.112	ng	# 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034717.D
Acq On : 25 Oct 2024 06:39
Operator : RC/JU
Sample : P4440-01
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016


Quant Time: Oct 25 07:09:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.712 min Scan# 669
Delta R.T. 0.000 min
Lab File: BN034717.D
Acq: 25 Oct 2024 06:39

Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016







Tgt Ion:152 Resp: 5535
Ion Ratio Lower Upper
152 100
150 149.7 122.7 184.1
115 61.8 51.0 76.6



#4
2-Fluorophenol
Concen: 0.024 ng
RT: 5.322 min Scan# 338
Delta R.T. 0.000 min
Lab File: BN034717.D
Acq: 25 Oct 2024 06:39

Tgt Ion:112 Resp: 400
Ion Ratio Lower Upper
112 100
64 73.3 53.0 79.6
63 50.2 27.6 41.4#









#5
Phenol-d6
Concen: 0.004 ng
RT: 6.882 min Scan# 554
Delta R.T. 0.000 min
Lab File: BN034717.D
Acq: 25 Oct 2024 06:39




Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016


Tgt Ion: 99 Resp: 94
Ion Ratio Lower Upper
99 100
42 55.3 21.8 32.8#
71 111.7 26.6 40.0#



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 955
Delta R.T. 0.000 min
Lab File: BN034717.D
Acq: 25 Oct 2024 06:39

Tgt Ion: 136 Resp: 15645
Ion Ratio Lower Upper
136 100
137 11.3 9.1 13.7
54 11.4 9.8 14.8
68 7.4 6.4 9.6









#8
Nitrobenzene-d5
Concen: 0.313 ng
RT: 8.845 min Scan# 80
Delta R.T. 0.000 min
Lab File: BN034717.D
Acq: 25 Oct 2024 06:39




Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016


Tgt Ion:	82	Resp:	4108
Ion	Ratio	Lower	Upper
82	100		
128	38.7	29.7	44.5
54	65.2	53.0	79.6



#11
2-Methylnaphthalene-d10
Concen: 0.091 ng
RT: 12.067 min Scan# 1200
Delta R.T. -0.004 min
Lab File: BN034717.D
Acq: 25 Oct 2024 06:39

Tgt Ion:	152	Resp:	1956
Ion	Ratio	Lower	Upper
152	100		
151	21.0	16.7	25.1









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.318 min Scan# 1503
Delta R.T. -0.011 min
Lab File: BN034717.D
Acq: 25 Oct 2024 06:39




Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016


Tgt Ion:164 Resp: 6994
Ion Ratio Lower Upper
164 100
162 106.5 81.4 122.0
160 54.9 40.6 61.0



#14
2,4,6-Tribromophenol
Concen: 0.171 ng
RT: 15.815 min Scan# 1639
Delta R.T. 0.000 min
Lab File: BN034717.D
Acq: 25 Oct 2024 06:39

Tgt Ion:330 Resp: 370
Ion Ratio Lower Upper
330 100
332 117.6 76.6 115.0#
141 59.2 54.6 81.8









#15
2-Fluorobiphenyl
Concen: 0.335 ng
RT: 12.950 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN034717.D
Acq: 25 Oct 2024 06:39




Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016


Tgt Ion:	172	Resp:	9314
Ion Ratio	Lower	Upper	
172	100		
171	34.7	28.3	42.5
170	23.8	19.8	29.6



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.057 min Scan# 1739
Delta R.T. -0.012 min
Lab File: BN034717.D
Acq: 25 Oct 2024 06:39

Tgt Ion:	188	Resp:	12994
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	20.6	8.9	13.3#





#27

Fluoranthene-d10

Concen: 0.346 ng

RT: 19.094 min Scan# 1982

Delta R.T. 0.000 min

Lab File: BN034717.D

Acq: 25 Oct 2024 06:39





Instrument :

BNA_N

ClientSampleId :

BPOW6-11-HYD-20241016

Tgt Ion:212	Resp:	10124
Ion Ratio	Lower	Upper
212	100	
106	23.5	17.8 26.8
104	13.3	10.2 15.4



#29

Chrysene-d12

Concen: 0.400 ng




RT: 21.241 min Scan# 2420


Delta R.T. -0.009 min

Lab File: BN034717.D

Acq: 25 Oct 2024 06:39

Tgt Ion:240	Resp:	7954
Ion Ratio	Lower	Upper
240	100	
120	30.7	14.2 21.2#
236	30.4	23.1 34.7









#31
Terphenyl-d14
Concen: 0.383 ng
RT: 19.698 min Scan# 2112
Delta R.T. -0.005 min
Lab File: BN034717.D
Acq: 25 Oct 2024 06:39




Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016


Tgt Ion	Ratio	Lower	Upper
244	100		
212	11.7	8.6	13.0
122	30.7	22.3	33.5



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.112 ng
RT: 21.187 min Scan# 2414
Delta R.T. -0.009 min
Lab File: BN034717.D
Acq: 25 Oct 2024 06:39

Tgt Ion	Ratio	Lower	Upper
149	100		
167	24.9	19.7	29.5
279	3.3	1.8	2.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.458 min Scan# 2910

Delta R.T. -0.009 min

Lab File: BN034717.D


Acq: 25 Oct 2024 06:39

Instrument :

BNA_N

ClientSampleId :

BPOW6-11-HYD-20241016





Tgt Ion:264 Resp: 7258

Ion Ratio Lower Upper

264 100

260 27.1 21.9 32.9

265 88.7 60.6 91.0



Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/16/24	
Project:	CTO WE13		Date Received:	10/18/24	
Client Sample ID:	BPOW6-11-HYD-20241016RE		SDG No.:	P4440	
Lab Sample ID:	P4440-01RE		Matrix:	Water	
Analytical Method:	SW8270SIM		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN034724.D	1	10/21/24 08:45	10/25/24 11:34	PB164282

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.095	*	30 - 150		24%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.37		30 - 150		93%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		55 - 111		81%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.33		53 - 106		83%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.39		58 - 132		98%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	6120	7.712				
1146-65-2	Naphthalene-d8	17500	10.479				
15067-26-2	Acenaphthene-d10	8280	14.318				
1517-22-2	Phenanthrene-d10	15500	17.057				
1719-03-5	Chrysene-d12	11200	21.241				
1520-96-3	Perylene-d12	11500	23.461				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102524\
Data File : BN034724.D
Acq On : 25 Oct 2024 11:34
Operator : RC/JU
Sample : P4440-01RE
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016RE

Quant Time: Oct 25 12:33:55 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.712	152	6120	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	17531	0.400	ng	0.00
13) Acenaphthene-d10	14.318	164	8276	0.400	ng	-0.01
19) Phenanthrene-d10	17.057	188	15513	0.400	ng	#-0.01
29) Chrysene-d12	21.241	240	11157	0.400	ng	# 0.00
35) Perylene-d12	23.461	264	11491	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.322	112	471	0.026	ng	0.00
5) Phenol-d6	6.882	99	99	0.004	ng	0.00
8) Nitrobenzene-d5	8.845	82	4759	0.324	ng	0.00
11) 2-Methylnaphthalene-d10	12.067	152	2286	0.095	ng	0.00
14) 2,4,6-Tribromophenol	15.815	330	550	0.215	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	10852	0.330	ng	0.00
27) Fluoranthene-d10	19.089	212	12945	0.371	ng	0.00
31) Terphenyl-d14	19.698	244	8939	0.392	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.187	149	3662	0.151	ng	# 98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102524\
Data File : BN034724.D
Acq On : 25 Oct 2024 11:34
Operator : RC/JU
Sample : P4440-01RE
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016RE

Quant Time: Oct 25 12:33:55 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.712 min Scan# 669
Delta R.T. 0.000 min
Lab File: BN034724.D
Acq: 25 Oct 2024 11:34




Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016RE


Tgt Ion:152 Resp: 6120
Ion Ratio Lower Upper
152 100
150 150.2 122.7 184.1
115 63.2 51.0 76.6



#4
2-Fluorophenol
Concen: 0.026 ng
RT: 5.322 min Scan# 338
Delta R.T. -0.000 min
Lab File: BN034724.D
Acq: 25 Oct 2024 11:34

Tgt Ion:112 Resp: 471
Ion Ratio Lower Upper
112 100
64 74.7 53.0 79.6
63 57.3 27.6 41.4#









#5
Phenol-d6
Concen: 0.004 ng
RT: 6.882 min Scan# 554
Delta R.T. -0.000 min
Lab File: BN034724.D
Acq: 25 Oct 2024 11:34




Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016RE


Tgt Ion: 99 Resp: 99
Ion Ratio Lower Upper
99 100
42 82.8 21.8 32.8#
71 140.4 26.6 40.0#



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 955
Delta R.T. -0.000 min
Lab File: BN034724.D
Acq: 25 Oct 2024 11:34

Tgt Ion: 136 Resp: 17531
Ion Ratio Lower Upper
136 100
137 11.6 9.1 13.7
54 11.3 9.8 14.8
68 7.7 6.4 9.6









#8
Nitrobenzene-d5
Concen: 0.324 ng
RT: 8.845 min Scan# 80
Delta R.T. -0.000 min
Lab File: BN034724.D
Acq: 25 Oct 2024 11:34




Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016RE


Tgt Ion: 82 Resp: 4759
Ion Ratio Lower Upper
82 100
128 41.4 29.7 44.5
54 64.3 53.0 79.6



#11
2-Methylnaphthalene-d10
Concen: 0.095 ng
RT: 12.067 min Scan# 1200
Delta R.T. -0.004 min
Lab File: BN034724.D
Acq: 25 Oct 2024 11:34

Tgt Ion: 152 Resp: 2286
Ion Ratio Lower Upper
152 100
151 21.9 16.7 25.1









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.318 min Scan# 1503
Delta R.T. -0.011 min
Lab File: BN034724.D
Acq: 25 Oct 2024 11:34




Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016RE


Tgt Ion:164 Resp: 8276
Ion Ratio Lower Upper
164 100
162 106.4 81.4 122.0
160 53.1 40.6 61.0



#14
2,4,6-Tribromophenol
Concen: 0.215 ng
RT: 15.815 min Scan# 1639
Delta R.T. -0.000 min
Lab File: BN034724.D
Acq: 25 Oct 2024 11:34

Tgt Ion:330 Resp: 550
Ion Ratio Lower Upper
330 100
332 119.6 76.6 115.0#
141 58.0 54.6 81.8









#15
2-Fluorobiphenyl
Concen: 0.330 ng
RT: 12.950 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN034724.D
Acq: 25 Oct 2024 11:34




Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016RE


Tgt Ion:	172	Resp:	10852
Ion Ratio	Lower	Upper	
172	100		
171	34.8	28.3	42.5
170	23.8	19.8	29.6



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.057 min Scan# 1739
Delta R.T. -0.012 min
Lab File: BN034724.D
Acq: 25 Oct 2024 11:34

Tgt Ion:	188	Resp:	15513
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	21.4	8.9	13.3









#27
Fluoranthene-d10
Concen: 0.371 ng
RT: 19.089 min Scan# 1981
Delta R.T. -0.005 min
Lab File: BN034724.D
Acq: 25 Oct 2024 11:34




Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016RE


Tgt Ion:	212	Resp:	12945
Ion Ratio	Lower	Upper	
212	100		
106	22.7	17.8	26.8
104	12.7	10.2	15.4



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.241 min Scan# 2420
Delta R.T. -0.009 min
Lab File: BN034724.D
Acq: 25 Oct 2024 11:34

Tgt Ion:	240	Resp:	11157
Ion Ratio	Lower	Upper	
240	100		
120	30.3	14.2	21.2#
236	29.9	23.1	34.7









#31
Terphenyl-d14
Concen: 0.392 ng
RT: 19.698 min Scan# 2112
Delta R.T. -0.005 min
Lab File: BN034724.D
Acq: 25 Oct 2024 11:34




Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016RE


Tgt Ion	Ratio	Lower	Upper
244	100		
212	10.5	8.6	13.0
122	27.6	22.3	33.5



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.151 ng
RT: 21.187 min Scan# 2414
Delta R.T. -0.009 min
Lab File: BN034724.D
Acq: 25 Oct 2024 11:34


Tgt Ion	Ratio	Lower	Upper
149	100		
167	25.6	19.7	29.5
279	2.8	1.8	2.6







#35
Perylene-d12
Concen: 0.400 ng
RT: 23.461 min Scan# 2911
Delta R.T. -0.006 min
Lab File: BN034724.D
Acq: 25 Oct 2024 11:34

Instrument :
BNA_N
ClientSampleId :
BPOW6-11-HYD-20241016RE



Tgt Ion:264 Resp: 11491
Ion Ratio Lower Upper
264 100
260 27.3 21.9 32.9
265 72.1 60.6 91.0





CALIBRATION SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
Method File : 8270-SIM-BN102424.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Oct 24 13:17:02 2024
Response Via : Initial Calibration

Calibration Files

0.1 =BN034684.D 0.2 =BN034685.D 0.4 =BN034686.D 0.8 =BN034687.D 1.6 =BN034688.D 3.2 =BN034689.D 5.0 =BN034690.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5.0	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.621	0.564	0.485	0.545	0.530	0.503	0.479	0.532	9.40
3)	n-Nitrosodimet...	0.775	0.766	0.704	0.842	0.827	0.761	0.749	0.775	6.06
4) S	2-Fluorophenol	1.188	1.202	1.070	1.268	1.253	1.168	1.177	1.189	5.46
5) S	Phenol-d6	1.556	1.554	1.386	1.658	1.657	1.569	1.597	1.568	5.85
6)	bis(2-Chloroet...	1.306	1.281	1.140	1.380	1.346	1.243	1.220	1.274	6.37
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.345	0.323	0.294	0.354	0.349	0.337	0.348	0.336	6.27
9)	Naphthalene	1.126	1.100	0.987	1.182	1.157	1.094	1.101	1.107	5.61
10)	Hexachlorobuta...	0.175	0.169	0.149	0.178	0.170	0.160	0.159	0.166	6.23
11) SURR	2-Methylnaphth...	0.542	0.534	0.481	0.583	0.586	0.557	0.566	0.550	6.53
12)	2-Methylnaphth...	0.675	0.670	0.600	0.729	0.730	0.697	0.705	0.687	6.50
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.110	0.118	0.096	0.127	0.133	0.136	0.146	0.124	13.66
15) S	2-Fluorobiphenyl	1.661	1.577	1.379	1.703	1.666	1.540	1.601	1.590	6.84
16)	Acenaphthylene	1.900	1.860	1.581	2.040	2.038	1.964	2.064	1.921	8.76
17)	Acenaphthene	1.318	1.286	1.122	1.435	1.407	1.329	1.370	1.324	7.79
18)	Fluorene	1.598	1.609	1.388	1.764	1.750	1.663	1.663	1.634	7.68
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.048	0.043	0.055	0.059	0.063	0.067	0.056		16.66
21)	4-Bromophenyl-...	0.217	0.211	0.190	0.224	0.221	0.212	0.219	0.213	5.22
22)	Hexachlorobenzene	0.238	0.234	0.211	0.243	0.238	0.226	0.229	0.231	4.55
23)	Atrazine	0.151	0.158	0.140	0.183	0.181	0.181	0.179	0.168	10.55
24)	Pentachlorophenol	0.065	0.064	0.083	0.087	0.092	0.100	0.082		17.73
25)	Phenanthrene	1.214	1.208	1.065	1.273	1.244	1.191	1.215	1.202	5.49
26)	Anthracene	1.006	1.043	0.931	1.124	1.127	1.102	1.131	1.067	7.17
27) SURR	Fluoranthene-d10	0.870	0.885	0.784	0.994	0.951	0.913	0.910	0.901	7.35
28)	Fluoranthene	1.188	1.222	1.092	1.380	1.333	1.290	1.269	1.253	7.66
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	2.066	2.035	1.880	2.152	2.164	2.039	2.048	2.055	4.57
31) S	Terphenyl-d14	0.813	0.797	0.740	0.867	0.865	0.820	0.820	0.817	5.27
32)	Benzo(a)anthra...	1.502	1.445	1.373	1.640	1.613	1.553	1.570	1.528	6.19
33)	Chrysene	1.642	1.568	1.442	1.709	1.643	1.533	1.531	1.581	5.68
34)	Bis(2-ethylhex...	0.901	0.779	0.722	0.843	0.878	0.915	1.030	0.867	11.48
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN102424.M

36)	Indeno(1,2,3-c...	1.562	1.447	1.418	1.596	1.620	1.517	1.624	1.541	5.39
37)	Benzo(b)fluora...	1.543	1.516	1.481	1.714	1.699	1.614	1.615	1.597	5.57
38)	Benzo(k)fluora...	1.511	1.491	1.386	1.723	1.712	1.596	1.613	1.576	7.75
39) C	Benzo(a)pyrene	1.203	1.188	1.157	1.369	1.383	1.323	1.350	1.282	7.45
40)	Dibenzo(a,h)an...	1.236	1.139	1.113	1.246	1.283	1.201	1.288	1.215	5.58
41)	Benzo(g,h,i)pe...	1.397	1.287	1.262	1.377	1.387	1.290	1.378	1.340	4.28

(#) = Out of Range

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034684.D
Acq On : 24 Oct 2024 09:11
Operator : RC/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Oct 24 12:42:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 12:41:40 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.712	152	5558	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	16023	0.400	ng	0.00
13) Acenaphthene-d10	14.329	164	7436	0.400	ng	0.00
19) Phenanthrene-d10	17.069	188	15281	0.400	ng	0.00
29) Chrysene-d12	21.250	240	9023	0.400	ng	0.00
35) Perylene-d12	23.467	264	9104	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.322	112	1651	0.097	ng	0.00
5) Phenol-d6	6.882	99	2162	0.099	ng	0.00
8) Nitrobenzene-d5	8.846	82	1381	0.104	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	2173	0.096	ng	0.00
14) 2,4,6-Tribromophenol	15.815	330	205	0.065	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	3087	0.105	ng	0.00
27) Fluoranthene-d10	19.094	212	3322	0.089	ng	0.00
31) Terphenyl-d14	19.703	244	1833	0.105	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.285	88	863	0.128	ng	89
3) n-Nitrosodimethylamine	3.588	42	1077	0.132	ng	# 95
6) bis(2-Chloroethyl)ether	7.142	93	1815	0.109	ng	99
9) Naphthalene	10.533	128	4509	0.102	ng	96
10) Hexachlorobutadiene	10.831	225	703	0.099	ng	# 99
12) 2-Methylnaphthalene	12.147	142	2704	0.097	ng	98
16) Acenaphthylene	14.041	152	3532	0.099	ng	100
17) Acenaphthene	14.393	154	2450	0.101	ng	97
18) Fluorene	15.377	166	2971	0.098	ng	100
21) 4-Bromophenyl-phenylether	16.275	248	830	0.099	ng	96
22) Hexachlorobenzene	16.374	284	909	0.098	ng	97
23) Atrazine	16.548	200	577	0.079	ng	# 87
25) Phenanthrene	17.106	178	4637	0.102	ng	100
26) Anthracene	17.193	178	3843	0.095	ng	100
28) Fluoranthene	19.127	202	4537	0.089	ng	99
30) Pyrene	19.485	202	4660	0.116	ng	99
32) Benzo(a)anthracene	21.232	228	3389	0.099	ng	97
33) Chrysene	21.286	228	3705	0.109	ng	98
34) Bis(2-ethylhexyl)phtha...	21.196	149	2033	0.087	ng	# 97
36) Indeno(1,2,3-cd)pyrene	25.695	276	3556	0.101	ng	96
37) Benzo(b)fluoranthene	22.806	252	3512	0.107	ng	# 65
38) Benzo(k)fluoranthene	22.847	252	3438	0.107	ng	# 62
39) Benzo(a)pyrene	23.370	252	2737	0.098	ng	# 55
40) Dibenzo(a,h)anthracene	25.718	278	2812	0.102	ng	# 78
41) Benzo(g,h,i)perylene	26.370	276	3179	0.105	ng	# 89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034684.D
Acq On : 24 Oct 2024 09:11
Operator : RC/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Oct 24 12:42:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 12:41:40 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.712 min Scan# 60
Delta R.T. 0.000 min
Lab File: BN034684.D
Acq: 24 Oct 2024 09:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 5558
Ion Ratio Lower Upper
152 100
150 152.0 122.7 184.1
115 62.4 51.0 76.6



#2
1,4-Dioxane
Concen: 0.128 ng
RT: 3.285 min Scan# 56
Delta R.T. 0.000 min
Lab File: BN034684.D
Acq: 24 Oct 2024 09:11

Tgt Ion: 88 Resp: 863
Ion Ratio Lower Upper
88 100
43 40.9 32.1 48.1
58 77.6 73.7 110.5









#3
n-Nitrosodimethylamine
Concen: 0.132 ng
RT: 3.588 min Scan# 91
Delta R.T. 0.007 min
Lab File: BN034684.D
Acq: 24 Oct 2024 09:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 42 Resp: 1077
Ion Ratio Lower Upper
42 100
74 93.9 77.2 115.8
44 8.6 13.4 20.2#



#4
2-Fluorophenol
Concen: 0.097 ng
RT: 5.322 min Scan# 338
Delta R.T. 0.000 min
Lab File: BN034684.D
Acq: 24 Oct 2024 09:11

Tgt Ion: 112 Resp: 1651
Ion Ratio Lower Upper
112 100
64 67.8 53.0 79.6
63 35.0 27.6 41.4









#5
Phenol-d6
Concen: 0.099 ng
RT: 6.882 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN034684.D
Acq: 24 Oct 2024 09:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 99 Resp: 2162
Ion Ratio Lower Upper
99 100
42 31.1 21.8 32.8
71 33.9 26.6 40.0



#6
bis(2-Chloroethyl)ether
Concen: 0.109 ng
RT: 7.142 min Scan# 590
Delta R.T. 0.000 min
Lab File: BN034684.D
Acq: 24 Oct 2024 09:11

Tgt Ion: 93 Resp: 1815
Ion Ratio Lower Upper
93 100
63 87.8 71.5 107.3
95 31.8 25.8 38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 955
Delta R.T. 0.000 min
Lab File: BN034684.D
Acq: 24 Oct 2024 09:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	136	Resp:	16023
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.1	13.7
54	11.3	9.8	14.8
68	7.6	6.4	9.6



#8
Nitrobenzene-d5
Concen: 0.104 ng
RT: 8.846 min Scan# 802
Delta R.T. 0.000 min
Lab File: BN034684.D
Acq: 24 Oct 2024 09:11

Tgt Ion:	82	Resp:	1381
Ion	Ratio	Lower	Upper
82	100		
128	41.7	29.7	44.5
54	67.2	53.0	79.6





#9

Naphthalene

Concen: 0.102 ng

RT: 10.533 min Scan# 9

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:128 Resp: 4509

Ion Ratio Lower Upper

128 100

129 13.0 9.0 13.6

127 14.5 10.5 15.7



#10

Hexachlorobutadiene

Concen: 0.099 ng

RT: 10.831 min Scan# 988

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11




Tgt Ion:225 Resp: 703


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 63.7 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 0.096 ng
RT: 12.071 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN034684.D
Acq: 24 Oct 2024 09:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 2173
Ion Ratio Lower Upper
152 100
151 21.4 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.097 ng
RT: 12.147 min Scan# 1221
Delta R.T. 0.000 min
Lab File: BN034684.D
Acq: 24 Oct 2024 09:11

Tgt Ion:142 Resp: 2704
Ion Ratio Lower Upper
142 100
141 87.3 70.4 105.6
115 39.2 29.8 44.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.329 min Scan# 1504
Delta R.T. 0.000 min
Lab File: BN034684.D
Acq: 24 Oct 2024 09:11




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:164 Resp: 7436
Ion Ratio Lower Upper
164 100
162 102.7 81.4 122.0
160 52.0 40.6 61.0



#14
2,4,6-Tribromophenol
Concen: 0.065 ng
RT: 15.815 min Scan# 1639
Delta R.T. 0.000 min
Lab File: BN034684.D
Acq: 24 Oct 2024 09:11

Tgt Ion:330 Resp: 205
Ion Ratio Lower Upper
330 100
332 95.1 76.6 115.0
141 66.3 54.6 81.8





#15

2-Fluorobiphenyl

Concen: 0.105 ng

RT: 12.950 min Scan# 1375

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:172 Resp: 3087

Ion Ratio Lower Upper

172 100

171 36.0 28.3 42.5

170 25.7 19.8 29.6



#16

Acenaphthylene

Concen: 0.099 ng

RT: 14.041 min Scan# 1477

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11




Tgt Ion:152 Resp: 3532


Ion Ratio Lower Upper

152 100

151 19.4 15.6 23.4

153 13.5 10.6 15.8





#17

Acenaphthene

Concen: 0.101 ng

RT: 14.393 min Scan# 1510

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:154 Resp: 2450

Ion Ratio Lower Upper

154 100

153 113.7 88.2 132.2

152 59.7 46.9 70.3



#18

Fluorene

Concen: 0.098 ng

RT: 15.377 min Scan# 1602

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11




Tgt Ion:166 Resp: 2971


Ion Ratio Lower Upper

166 100

165 98.5 78.8 118.2

167 13.1 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.069 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:188 Resp: 15281

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.1 8.9 13.3



#21

4-Bromophenyl-phenylether

Concen: 0.099 ng

RT: 16.275 min Scan# 1676

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11




Tgt Ion:248 Resp: 830


Ion Ratio Lower Upper

248 100

250 104.0 80.9 121.3

141 55.5 40.5 60.7





#22

Hexachlorobenzene

Concen: 0.098 ng

RT: 16.374 min Scan# 1684

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:284 Resp: 909

Ion Ratio Lower Upper

284 100

142 55.8 43.6 65.4

249 36.0 26.6 40.0



#23

Atrazine

Concen: 0.079 ng

RT: 16.548 min Scan# 1698

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11




Tgt Ion:200 Resp: 577


Ion Ratio Lower Upper

200 100

173 31.6 19.7 29.5#

215 52.3 35.7 53.5









#25
Phenanthrene
Concen: 0.102 ng
RT: 17.106 min Scan# 1743
Delta R.T. 0.000 min
Lab File: BN034684.D
Acq: 24 Oct 2024 09:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:178 Resp: 4637
Ion Ratio Lower Upper
178 100
176 19.5 15.4 23.2
179 15.2 12.2 18.4



#26
Anthracene
Concen: 0.095 ng
RT: 17.193 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BN034684.D
Acq: 24 Oct 2024 09:11

Tgt Ion:178 Resp: 3843
Ion Ratio Lower Upper
178 100
176 18.7 14.9 22.3
179 15.3 12.2 18.2





#27

Fluoranthene-d10

Concen: 0.089 ng

RT: 19.094 min Scan# 1982

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:212 Resp: 3322

Ion Ratio Lower Upper

212 100

106 23.1 17.8 26.8

104 13.2 10.2 15.4



#28

Fluoranthene

Concen: 0.089 ng

RT: 19.127 min Scan# 1989

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11




Tgt Ion:202 Resp: 4537


Ion Ratio Lower Upper

202 100

101 15.5 12.8 19.2

203 17.2 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.250 min Scan# 2421

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:240 Resp: 9023

Ion Ratio Lower Upper

240 100

120 20.9 14.2 21.2

236 29.7 23.1 34.7



#30

Pyrene

Concen: 0.116 ng

RT: 19.485 min Scan# 2066

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11




Tgt Ion:202 Resp: 4660


Ion Ratio Lower Upper

202 100

200 21.1 16.6 25.0

203 18.3 14.2 21.4









#31
Terphenyl-d14
Concen: 0.105 ng
RT: 19.703 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BN034684.D
Acq: 24 Oct 2024 09:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:244 Resp: 1833
Ion Ratio Lower Upper
244 100
212 13.7 8.6 13.0#
122 29.6 22.3 33.5



#32
Benzo(a)anthracene
Concen: 0.099 ng
RT: 21.232 min Scan# 2419
Delta R.T. 0.000 min
Lab File: BN034684.D
Acq: 24 Oct 2024 09:11

Tgt Ion:228 Resp: 3389
Ion Ratio Lower Upper
228 100
226 28.7 22.0 33.0
229 21.2 15.6 23.4





#33

Chrysene

Concen: 0.109 ng

RT: 21.286 min Scan# 2425

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:228 Resp: 3705

Ion Ratio Lower Upper

228 100

226 30.7 24.1 36.1

229 21.1 15.8 23.8



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.087 ng

RT: 21.196 min Scan# 2415

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11




Tgt Ion:149 Resp: 2033


Ion Ratio Lower Upper

149 100

167 23.0 19.7 29.5

279 1.7 1.8 2.6#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.467 min Scan# 2913

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:264 Resp: 9104

Ion Ratio Lower Upper

264 100

260 27.6 21.9 32.9

265 80.7 60.6 91.0



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.101 ng

RT: 25.695 min Scan# 3675

Delta R.T. 0.000 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11




Tgt Ion:276 Resp: 3556


Ion Ratio Lower Upper

276 100

138 47.7 35.1 52.7

277 25.1 19.7 29.5





#37

Benzo(b)fluoranthene

Concen: 0.107 ng

RT: 22.806 min Scan# 2

Delta R.T. 0.003 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:252 Resp: 3512

Ion Ratio Lower Upper

252 100

253 35.4 20.0 30.0#

125 50.6 20.9 31.3#



#38

Benzo(k)fluoranthene

Concen: 0.107 ng

RT: 22.847 min Scan# 2701

Delta R.T. -0.003 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11




Tgt Ion:252 Resp: 3438


Ion Ratio Lower Upper

252 100

253 36.4 20.6 31.0#

125 54.9 21.6 32.4#





#39

Benzo(a)pyrene

Concen: 0.098 ng

RT: 23.370 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:252 Resp: 2737

Ion Ratio Lower Upper

252 100

253 42.1 22.2 33.4#

125 66.8 26.0 39.0#



#40

Dibenzo(a,h)anthracene

Concen: 0.102 ng

RT: 25.718 min Scan# 3683

Delta R.T. 0.003 min

Lab File: BN034684.D

Acq: 24 Oct 2024 09:11




Tgt Ion:278 Resp: 2812


Ion Ratio Lower Upper

278 100

139 42.6 25.4 38.2#

279 41.6 22.4 33.6#





#41

Benzo(g,h,i)perylene

Concen: 0.105 ng

RT: 26.370 min Scan# 3906

Delta R.T. -0.003 min

Lab File: BN034684.D


Acq: 24 Oct 2024 09:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:276 Resp: 3179

Ion Ratio Lower Upper

276 100

277 34.0 21.3 31.9#

138 45.5 31.9 47.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034685.D
Acq On : 24 Oct 2024 09:47
Operator : RC/JU
Sample : SSTDICC0.2
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Quant Time: Oct 24 12:42:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 12:41:40 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.712	152	5856	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	17008	0.400	ng	0.00
13) Acenaphthene-d10	14.329	164	8253	0.400	ng	0.00
19) Phenanthrene-d10	17.069	188	17136	0.400	ng	0.00
29) Chrysene-d12	21.250	240	10548	0.400	ng	0.00
35) Perylene-d12	23.467	264	9753	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.322	112	3519	0.196	ng	0.00
5) Phenol-d6	6.882	99	4551	0.199	ng	0.00
8) Nitrobenzene-d5	8.845	82	2747	0.195	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	4541	0.190	ng	0.00
14) 2,4,6-Tribromophenol	15.815	330	486	0.139	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	6506	0.200	ng	0.00
27) Fluoranthene-d10	19.094	212	7583	0.181	ng	0.00
31) Terphenyl-d14	19.703	244	4203	0.206	ng	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.285	88	1652	0.232	ng	93
3) n-Nitrosodimethylamine	3.588	42	2242	0.260	ng	# 97
6) bis(2-Chloroethyl)ether	7.142	93	3752	0.214	ng	100
9) Naphthalene	10.532	128	9353	0.199	ng	98
10) Hexachlorobutadiene	10.831	225	1441	0.192	ng	# 97
12) 2-Methylnaphthalene	12.147	142	5699	0.192	ng	99
16) Acenaphthylene	14.040	152	7676	0.194	ng	99
17) Acenaphthene	14.393	154	5306	0.196	ng	98
18) Fluorene	15.377	166	6638	0.197	ng	99
20) 4,6-Dinitro-2-methylph...	15.452	198	410	0.175	ng	# 58
21) 4-Bromophenyl-phenylether	16.275	248	1806	0.193	ng	99
22) Hexachlorobenzene	16.374	284	2002	0.192	ng	99
23) Atrazine	16.548	200	1350	0.165	ng	94
24) Pentachlorophenol	16.721	266	558	0.137	ng	98
25) Phenanthrene	17.106	178	10352	0.203	ng	100
26) Anthracene	17.193	178	8939	0.196	ng	99
28) Fluoranthene	19.127	202	10471	0.183	ng	100
30) Pyrene	19.484	202	10731	0.228	ng	100
32) Benzo(a)anthracene	21.232	228	7621	0.190	ng	99
33) Chrysene	21.286	228	8267	0.208	ng	99
34) Bis(2-ethylhexyl)phtha...	21.196	149	4110	0.150	ng	# 99
36) Indeno(1,2,3-cd)pyrene	25.700	276	7054	0.188	ng	98
37) Benzo(b)fluoranthene	22.803	252	7395	0.210	ng	# 88
38) Benzo(k)fluoranthene	22.847	252	7269	0.211	ng	# 89
39) Benzo(a)pyrene	23.370	252	5791	0.193	ng	# 84
40) Dibenzo(a,h)anthracene	25.715	278	5556	0.188	ng	91
41) Benzo(g,h,i)perylene	26.367	276	6274	0.193	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034685.D
Acq On : 24 Oct 2024 09:47
Operator : RC/JU
Sample : SSTDICC0.2
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Quant Time: Oct 24 12:42:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 12:41:40 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.712 min Scan# 669
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 5856
Ion Ratio Lower Upper
152 100
150 155.7 122.7 184.1
115 63.5 51.0 76.6



#2
1,4-Dioxane
Concen: 0.232 ng
RT: 3.285 min Scan# 56
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47

Tgt Ion: 88 Resp: 1652
Ion Ratio Lower Upper
88 100
43 45.2 32.1 48.1
58 85.9 73.7 110.5









#3
n-Nitrosodimethylamine
Concen: 0.260 ng
RT: 3.588 min Scan# 91
Delta R.T. 0.007 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion: 42 Resp: 2242
Ion Ratio Lower Upper
42 100
74 95.2 77.2 115.8
44 21.1 13.4 20.2#



#4
2-Fluorophenol
Concen: 0.196 ng
RT: 5.322 min Scan# 338
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47

Tgt Ion: 112 Resp: 3519
Ion Ratio Lower Upper
112 100
64 65.2 53.0 79.6
63 34.5 27.6 41.4









#5
Phenol-d6
Concen: 0.199 ng
RT: 6.882 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion: 99 Resp: 4551
Ion Ratio Lower Upper
99 100
42 26.6 21.8 32.8
71 33.2 26.6 40.0



#6
bis(2-Chloroethyl)ether
Concen: 0.214 ng
RT: 7.142 min Scan# 590
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47

Tgt Ion: 93 Resp: 3752
Ion Ratio Lower Upper
93 100
63 89.0 71.5 107.3
95 31.9 25.8 38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	136	Resp:	17008
Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.1	13.7
54	12.0	9.8	14.8
68	8.0	6.4	9.6



#8
Nitrobenzene-d5
Concen: 0.195 ng
RT: 8.845 min Scan# 802
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47

Tgt Ion:	82	Resp:	2747
Ion	Ratio	Lower	Upper
82	100		
128	38.9	29.7	44.5
54	69.0	53.0	79.6





#9

Naphthalene

Concen: 0.199 ng

RT: 10.532 min Scan# 9

Delta R.T. 0.000 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:128 Resp: 9353

Ion Ratio Lower Upper

128 100

129 12.1 9.0 13.6

127 13.6 10.5 15.7



#10

Hexachlorobutadiene

Concen: 0.192 ng

RT: 10.831 min Scan# 988

Delta R.T. 0.000 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47




Tgt Ion:225 Resp: 1441


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 61.8 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 0.190 ng
RT: 12.071 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 4541
Ion Ratio Lower Upper
152 100
151 21.4 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.192 ng
RT: 12.147 min Scan# 1221
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47

Tgt Ion:142 Resp: 5699
Ion Ratio Lower Upper
142 100
141 88.2 70.4 105.6
115 39.1 29.8 44.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.329 min Scan# 1504
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:164 Resp: 8253
Ion Ratio Lower Upper
164 100
162 102.7 81.4 122.0
160 51.8 40.6 61.0



#14
2,4,6-Tribromophenol
Concen: 0.139 ng
RT: 15.815 min Scan# 1639
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47

Tgt Ion:330 Resp: 486
Ion Ratio Lower Upper
330 100
332 88.7 76.6 115.0
141 65.0 54.6 81.8





#15

2-Fluorobiphenyl

Concen: 0.200 ng

RT: 12.950 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.2





Tgt Ion:172 Resp: 6506

Ion Ratio Lower Upper

172 100

171 36.7 28.3 42.5

170 25.3 19.8 29.6



#16

Acenaphthylene

Concen: 0.194 ng

RT: 14.040 min Scan# 1477

Delta R.T. 0.000 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47




Tgt Ion:152 Resp: 7676


Ion Ratio Lower Upper

152 100

151 19.2 15.6 23.4

153 12.7 10.6 15.8









#17
Acenaphthene
Concen: 0.196 ng
RT: 14.393 min Scan# 1510
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:154 Resp: 5306
Ion Ratio Lower Upper
154 100
153 112.7 88.2 132.2
152 60.1 46.9 70.3



#18
Fluorene
Concen: 0.197 ng
RT: 15.377 min Scan# 1602
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47

Tgt Ion:166 Resp: 6638
Ion Ratio Lower Upper
166 100
165 99.3 78.8 118.2
167 13.5 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.069 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:188 Resp: 17136

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.4 8.9 13.3



#20

4,6-Dinitro-2-methylphenol

Concen: 0.175 ng

RT: 15.452 min Scan# 1609

Delta R.T. 0.000 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47




Tgt Ion:198 Resp: 410


Ion Ratio Lower Upper

198 100

51 174.3 91.8 137.8#

105 67.3 44.3 66.5#









#21
4-Bromophenyl-phenylether
Concen: 0.193 ng
RT: 16.275 min Scan# 1676
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion	Ratio	Lower	Upper
248	100		
250	102.2	80.9	121.3
141	51.7	40.5	60.7



#22
Hexachlorobenzene
Concen: 0.192 ng
RT: 16.374 min Scan# 1684
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47

Tgt Ion	Ratio	Lower	Upper
284	100		
142	55.4	43.6	65.4
249	33.6	26.6	40.0









#23
Atrazine
Concen: 0.165 ng
RT: 16.548 min Scan# 1698
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:200 Resp: 1350
Ion Ratio Lower Upper
200 100
173 28.0 19.7 29.5
215 47.8 35.7 53.5



#24
Pentachlorophenol
Concen: 0.137 ng
RT: 16.721 min Scan# 1712
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47

Tgt Ion:266 Resp: 558
Ion Ratio Lower Upper
266 100
264 64.3 49.4 74.2
268 62.4 50.5 75.7









#25
Phenanthrene
Concen: 0.203 ng
RT: 17.106 min Scan# 1743
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:178 Resp: 10352
Ion Ratio Lower Upper
178 100
176 19.2 15.4 23.2
179 15.4 12.2 18.4



#26
Anthracene
Concen: 0.196 ng
RT: 17.193 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47

Tgt Ion:178 Resp: 8939
Ion Ratio Lower Upper
178 100
176 18.2 14.9 22.3
179 15.1 12.2 18.2





#27

Fluoranthene-d10

Concen: 0.181 ng

RT: 19.094 min Scan# 1982

Delta R.T. 0.000 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:212 Resp: 7583

Ion Ratio Lower Upper

212 100

106 22.5 17.8 26.8

104 13.0 10.2 15.4



#28

Fluoranthene

Concen: 0.183 ng

RT: 19.127 min Scan# 1989

Delta R.T. 0.000 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47




Tgt Ion:202 Resp: 10471


Ion Ratio Lower Upper

202 100

101 16.0 12.8 19.2

203 16.9 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.250 min Scan# 2421

Delta R.T. 0.000 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.2





Tgt Ion:240 Resp: 10548

Ion Ratio Lower Upper

240 100

120 20.0 14.2 21.2

236 29.2 23.1 34.7



#30

Pyrene

Concen: 0.228 ng

RT: 19.484 min Scan# 2066

Delta R.T. 0.000 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47




Tgt Ion:202 Resp: 10731


Ion Ratio Lower Upper

202 100

200 20.9 16.6 25.0

203 17.6 14.2 21.4









#31
Terphenyl-d14
Concen: 0.206 ng
RT: 19.703 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:244 Resp: 4203
Ion Ratio Lower Upper
244 100
212 12.2 8.6 13.0
122 29.9 22.3 33.5



#32
Benzo(a)anthracene
Concen: 0.190 ng
RT: 21.232 min Scan# 2419
Delta R.T. 0.000 min
Lab File: BN034685.D
Acq: 24 Oct 2024 09:47

Tgt Ion:228 Resp: 7621
Ion Ratio Lower Upper
228 100
226 28.6 22.0 33.0
229 19.7 15.6 23.4





#33

Chrysene

Concen: 0.208 ng

RT: 21.286 min Scan# 2425

Delta R.T. 0.000 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.2





Tgt Ion:228 Resp: 8267

Ion Ratio Lower Upper

228 100

226 30.8 24.1 36.1

229 20.4 15.8 23.8



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.150 ng

RT: 21.196 min Scan# 2415

Delta R.T. 0.000 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47




Tgt Ion:149 Resp: 4110


Ion Ratio Lower Upper

149 100

167 25.2 19.7 29.5

279 2.7 1.8 2.6#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.467 min Scan# 2913

Delta R.T. 0.000 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.2





Tgt Ion:264 Resp: 9753

Ion Ratio Lower Upper

264 100

260 27.2 21.9 32.9

265 75.9 60.6 91.0



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.188 ng

RT: 25.700 min Scan# 3677

Delta R.T. 0.006 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47




Tgt Ion:276 Resp: 7054


Ion Ratio Lower Upper

276 100

138 45.8 35.1 52.7

277 24.7 19.7 29.5





#37

Benzo(b)fluoranthene

Concen: 0.210 ng

RT: 22.803 min Scan# 2

Delta R.T. 0.000 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2




Tgt Ion:252 Resp: 7395

Ion Ratio Lower Upper

252 100

253 28.7 20.0 30.0

125 34.8 20.9 31.3#



#38

Benzo(k)fluoranthene

Concen: 0.211 ng

RT: 22.847 min Scan# 2701

Delta R.T. -0.003 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47




Tgt Ion:252 Resp: 7269


Ion Ratio Lower Upper

252 100

253 28.7 20.6 31.0

125 35.3 21.6 32.4#





#39

Benzo(a)pyrene

Concen: 0.193 ng

RT: 23.370 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.2





Tgt Ion:252 Resp: 5791

Ion Ratio Lower Upper

252 100

253 33.5 22.2 33.4#

125 43.6 26.0 39.0#



#40

Dibenzo(a,h)anthracene

Concen: 0.188 ng

RT: 25.715 min Scan# 3682

Delta R.T. 0.000 min

Lab File: BN034685.D

Acq: 24 Oct 2024 09:47




Tgt Ion:278 Resp: 5556


Ion Ratio Lower Upper

278 100

139 36.0 25.4 38.2

279 33.5 22.4 33.6





#41

Benzo(g,h,i)perylene

Concen: 0.193 ng

RT: 26.367 min Scan# 3905

Delta R.T. -0.006 min

Lab File: BN034685.D


Acq: 24 Oct 2024 09:47

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:276 Resp: 6274

Ion Ratio Lower Upper

276 100

277 28.7 21.3 31.9

138 41.2 31.9 47.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
 Data File : BN034686.D
 Acq On : 24 Oct 2024 10:23
 Operator : RC/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Oct 24 12:42:59 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 24 12:41:40 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.712	152	6324	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	18497	0.400	ng	0.00
13) Acenaphthene-d10	14.329	164	9311	0.400	ng	0.00
19) Phenanthrene-d10	17.069	188	18185	0.400	ng	0.00
29) Chrysene-d12	21.250	240	10628	0.400	ng	0.00
35) Perylene-d12	23.467	264	9595	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.322	112	6764	0.350	ng	0.00
5) Phenol-d6	6.882	99	8764	0.354	ng	0.00
8) Nitrobenzene-d5	8.845	82	5433	0.355	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	8901	0.342	ng	0.00
14) 2,4,6-Tribromophenol	15.815	330	897	0.227	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	12842	0.350	ng	0.00
27) Fluoranthene-d10	19.094	212	14254	0.320	ng	0.00
31) Terphenyl-d14	19.703	244	7865	0.382	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.285	88	3065	0.398	ng	100
3) n-Nitrosodimethylamine	3.581	42	4451	0.478	ng	100
6) bis(2-Chloroethyl)ether	7.142	93	7208	0.380	ng	100
9) Naphthalene	10.532	128	18256	0.358	ng	100
10) Hexachlorobutadiene	10.831	225	2759	0.338	ng	# 100
12) 2-Methylnaphthalene	12.147	142	11104	0.344	ng	100
16) Acenaphthylene	14.040	152	14720	0.329	ng	100
17) Acenaphthene	14.393	154	10443	0.342	ng	100
18) Fluorene	15.377	166	12923	0.340	ng	100
20) 4,6-Dinitro-2-methylph...	15.452	198	774	0.311	ng	100
21) 4-Bromophenyl-phenylether	16.275	248	3462	0.348	ng	100
22) Hexachlorobenzene	16.374	284	3844	0.348	ng	100
23) Atrazine	16.548	200	2545	0.292	ng	100
24) Pentachlorophenol	16.721	266	1164	0.269	ng	100
25) Phenanthrene	17.106	178	19369	0.358	ng	100
26) Anthracene	17.193	178	16930	0.350	ng	100
28) Fluoranthene	19.127	202	19859	0.327	ng	100
30) Pyrene	19.484	202	19977	0.420	ng	100
32) Benzo(a)anthracene	21.232	228	14596	0.360	ng	100
33) Chrysene	21.286	228	15321	0.382	ng	100
34) Bis(2-ethylhexyl)phtha...	21.196	149	7675	0.278	ng	100
36) Indeno(1,2,3-cd)pyrene	25.694	276	13606	0.368	ng	100
37) Benzo(b)fluoranthene	22.803	252	14206	0.410	ng	100
38) Benzo(k)fluoranthene	22.850	252	13300	0.393	ng	100
39) Benzo(a)pyrene	23.367	252	11106	0.377	ng	100
40) Dibenzo(a,h)anthracene	25.715	278	10684	0.368	ng	100
41) Benzo(g,h,i)perylene	26.373	276	12106	0.378	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034686.D
Acq On : 24 Oct 2024 10:23
Operator : RC/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Quant Time: Oct 24 12:42:59 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 12:41:40 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.712 min Scan# 669
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 6324
Ion Ratio Lower Upper
152 100
150 153.4 122.7 184.1
115 63.8 51.0 76.6



#2
1,4-Dioxane
Concen: 0.398 ng
RT: 3.285 min Scan# 56
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23

Tgt Ion: 88 Resp: 3065
Ion Ratio Lower Upper
88 100
43 40.1 32.1 48.1
58 92.1 73.7 110.5









#3
n-Nitrosodimethylamine
Concen: 0.478 ng
RT: 3.581 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 42 Resp: 4451
Ion Ratio Lower Upper
42 100
74 96.5 77.2 115.8
44 16.8 13.4 20.2



#4
2-Fluorophenol
Concen: 0.350 ng
RT: 5.322 min Scan# 338
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23

Tgt Ion: 112 Resp: 6764
Ion Ratio Lower Upper
112 100
64 66.3 53.0 79.6
63 34.5 27.6 41.4









#5
Phenol-d6
Concen: 0.354 ng
RT: 6.882 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 99 Resp: 8764
Ion Ratio Lower Upper
99 100
42 27.3 21.8 32.8
71 33.3 26.6 40.0



#6
bis(2-Chloroethyl)ether
Concen: 0.380 ng
RT: 7.142 min Scan# 590
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23

Tgt Ion: 93 Resp: 7208
Ion Ratio Lower Upper
93 100
63 89.4 71.5 107.3
95 32.2 25.8 38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 955
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	136	Resp:	18497
Ion	Ratio	Lower	Upper
136	100		
137	11.4	9.1	13.7
54	12.3	9.8	14.8
68	8.0	6.4	9.6



#8
Nitrobenzene-d5
Concen: 0.355 ng
RT: 8.845 min Scan# 802
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23

Tgt Ion:	82	Resp:	5433
Ion	Ratio	Lower	Upper
82	100		
128	37.1	29.7	44.5
54	66.3	53.0	79.6





#9

Naphthalene

Concen: 0.358 ng

RT: 10.532 min Scan# 960

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:128 Resp: 18256

Ion Ratio Lower Upper

128 100

129 11.3 9.0 13.6

127 13.1 10.5 15.7



#10

Hexachlorobutadiene

Concen: 0.338 ng

RT: 10.831 min Scan# 988

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23




Tgt Ion:225 Resp: 2759


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 64.2 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 0.342 ng
RT: 12.071 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 8901
Ion Ratio Lower Upper
152 100
151 20.9 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.344 ng
RT: 12.147 min Scan# 1221
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23

Tgt Ion:142 Resp: 11104
Ion Ratio Lower Upper
142 100
141 88.0 70.4 105.6
115 37.2 29.8 44.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.329 min Scan# 1504
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:164 Resp: 9311
Ion Ratio Lower Upper
164 100
162 101.7 81.4 122.0
160 50.8 40.6 61.0



#14
2,4,6-Tribromophenol
Concen: 0.227 ng
RT: 15.815 min Scan# 1639
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23

Tgt Ion:330 Resp: 897
Ion Ratio Lower Upper
330 100
332 95.8 76.6 115.0
141 68.2 54.6 81.8









#15
2-Fluorobiphenyl
Concen: 0.350 ng
RT: 12.950 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	172	Resp:	12842
Ion Ratio	Lower	Upper	
172	100		
171	35.4	28.3	42.5
170	24.7	19.8	29.6



#16
Acenaphthylene
Concen: 0.329 ng
RT: 14.040 min Scan# 1477
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23

Tgt Ion:	152	Resp:	14720
Ion Ratio	Lower	Upper	
152	100		
151	19.5	15.6	23.4
153	13.2	10.6	15.8





#17

Acenaphthene

Concen: 0.342 ng

RT: 14.393 min Scan# 1510

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:154 Resp: 10443

Ion Ratio Lower Upper

154 100

153 110.2 88.2 132.2

152 58.6 46.9 70.3



#18

Fluorene

Concen: 0.340 ng

RT: 15.377 min Scan# 1602

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23




Tgt Ion:166 Resp: 12923


Ion Ratio Lower Upper

166 100

165 98.5 78.8 118.2

167 13.3 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.069 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:188 Resp: 18185

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.1 8.9 13.3



#20

4,6-Dinitro-2-methylphenol

Concen: 0.311 ng

RT: 15.452 min Scan# 1609

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23




Tgt Ion:198 Resp: 774


Ion Ratio Lower Upper

198 100

51 114.8 91.8 137.8

105 55.4 44.3 66.5









#21
4-Bromophenyl-phenylether
Concen: 0.348 ng
RT: 16.275 min Scan# 1676
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:248 Resp: 3462
Ion Ratio Lower Upper
248 100
250 101.1 80.9 121.3
141 50.6 40.5 60.7



#22
Hexachlorobenzene
Concen: 0.348 ng
RT: 16.374 min Scan# 1684
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23

Tgt Ion:284 Resp: 3844
Ion Ratio Lower Upper
284 100
142 54.5 43.6 65.4
249 33.3 26.6 40.0









#23
Atrazine
Concen: 0.292 ng
RT: 16.548 min Scan# 1698
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:200 Resp: 2545
Ion Ratio Lower Upper
200 100
173 24.6 19.7 29.5
215 44.6 35.7 53.5



#24
Pentachlorophenol
Concen: 0.269 ng
RT: 16.721 min Scan# 1712
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23

Tgt Ion:266 Resp: 1164
Ion Ratio Lower Upper
266 100
264 61.8 49.4 74.2
268 63.1 50.5 75.7









#25
Phenanthrene
Concen: 0.358 ng
RT: 17.106 min Scan# 1743
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:178 Resp: 19369
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.2
179 15.3 12.2 18.4



#26
Anthracene
Concen: 0.350 ng
RT: 17.193 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23

Tgt Ion:178 Resp: 16930
Ion Ratio Lower Upper
178 100
176 18.6 14.9 22.3
179 15.2 12.2 18.2





#27

Fluoranthene-d10

Concen: 0.320 ng

RT: 19.094 min Scan# 1982

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:212 Resp: 14254

Ion Ratio Lower Upper

212 100

106 22.3 17.8 26.8

104 12.8 10.2 15.4



#28

Fluoranthene

Concen: 0.327 ng

RT: 19.127 min Scan# 1989

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23




Tgt Ion:202 Resp: 19859


Ion Ratio Lower Upper

202 100

101 16.0 12.8 19.2

203 16.9 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.250 min Scan# 2421

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:240 Resp: 10628

Ion Ratio Lower Upper

240 100

120 17.7 14.2 21.2

236 28.9 23.1 34.7



#30

Pyrene

Concen: 0.420 ng

RT: 19.484 min Scan# 2066

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23




Tgt Ion:202 Resp: 19977


Ion Ratio Lower Upper

202 100

200 20.8 16.6 25.0

203 17.8 14.2 21.4









#31
Terphenyl-d14
Concen: 0.382 ng
RT: 19.703 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	Ratio	Lower	Upper
244	100		
212	10.8	8.6	13.0
122	27.9	22.3	33.5



#32
Benzo(a)anthracene
Concen: 0.360 ng
RT: 21.232 min Scan# 2419
Delta R.T. 0.000 min
Lab File: BN034686.D
Acq: 24 Oct 2024 10:23

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.5	22.0	33.0
229	19.5	15.6	23.4





#33

Chrysene

Concen: 0.382 ng

RT: 21.286 min Scan# 2425

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:228 Resp: 15321

Ion Ratio Lower Upper

228 100

226 30.1 24.1 36.1

229 19.8 15.8 23.8



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.278 ng

RT: 21.196 min Scan# 2415

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23




Tgt Ion:149 Resp: 7675


Ion Ratio Lower Upper

149 100

167 24.6 19.7 29.5

279 2.2 1.8 2.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.467 min Scan# 2913

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:264 Resp: 9595

Ion Ratio Lower Upper

264 100

260 27.4 21.9 32.9

265 75.8 60.6 91.0



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.368 ng

RT: 25.694 min Scan# 3675

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23




Tgt Ion:276 Resp: 13606


Ion Ratio Lower Upper

276 100

138 43.9 35.1 52.7

277 24.6 19.7 29.5





#37

Benzo(b)fluoranthene

Concen: 0.410 ng

RT: 22.803 min Scan# 2686

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:252 Resp: 14206

Ion Ratio Lower Upper

252 100

253 25.0 20.0 30.0

125 26.1 20.9 31.3



#38

Benzo(k)fluoranthene

Concen: 0.393 ng

RT: 22.850 min Scan# 2702

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23




Tgt Ion:252 Resp: 13300


Ion Ratio Lower Upper

252 100

253 25.8 20.6 31.0

125 27.0 21.6 32.4





#39

Benzo(a)pyrene

Concen: 0.377 ng

RT: 23.367 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:252 Resp: 11106

Ion Ratio Lower Upper

252 100

253 27.8 22.2 33.4

125 32.5 26.0 39.0



#40

Dibenzo(a,h)anthracene

Concen: 0.368 ng

RT: 25.715 min Scan# 3682

Delta R.T. 0.000 min

Lab File: BN034686.D

Acq: 24 Oct 2024 10:23




Tgt Ion:278 Resp: 10684


Ion Ratio Lower Upper

278 100

139 31.8 25.4 38.2

279 28.0 22.4 33.6





#41

Benzo(g,h,i)perylene

Concen: 0.378 ng

RT: 26.373 min Scan# 3907

Delta R.T. 0.000 min

Lab File: BN034686.D


Acq: 24 Oct 2024 10:23

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:276 Resp: 12106

Ion Ratio Lower Upper

276 100

277 26.6 21.3 31.9

138 39.9 31.9 47.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
 Data File : BN034687.D
 Acq On : 24 Oct 2024 10:59
 Operator : RC/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024

Quant Time: Oct 24 12:43:27 2024

Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Oct 24 12:41:40 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.712	152	6279	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	18298	0.400	ng	0.00
13) Acenaphthene-d10	14.325	164	8788	0.400	ng	0.00
19) Phenanthrene-d10	17.063	188	18528	0.400	ng	# 0.00
29) Chrysene-d12	21.247	240	12066	0.400	ng	# 0.00
35) Perylene-d12	23.467	264	10772	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.321	112	15922	0.829	ng	0.00
5) Phenol-d6	6.881	99	20820	0.847	ng	0.00
8) Nitrobenzene-d5	8.845	82	12949	0.854	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	21347	0.829	ng	0.00
14) 2,4,6-Tribromophenol	15.810	330	2227	0.598	ng	0.00
15) 2-Fluorobiphenyl	12.957	172	29939	0.864	ng	0.00
27) Fluoranthene-d10	19.096	212	36834	0.812	ng	0.00
31) Terphenyl-d14	19.705	244	20915	0.896	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.285	88	6845	0.895	ng	98
3) n-Nitrosodimethylamine	3.581	42	10574	1.144	ng	# 97
6) bis(2-Chloroethyl)ether	7.141	93	17336	0.921	ng	99
9) Naphthalene	10.532	128	43252	0.857	ng	99
10) Hexachlorobutadiene	10.831	225	6528	0.807	ng	# 99
12) 2-Methylnaphthalene	12.146	142	26668	0.836	ng	100
16) Acenaphthylene	14.047	152	35850	0.849	ng	99
17) Acenaphthene	14.389	154	25224	0.876	ng	99
18) Fluorene	15.373	166	31002	0.864	ng	100
20) 4,6-Dinitro-2-methylph...	15.448	198	2025	0.799	ng	# 77
21) 4-Bromophenyl-phenylether	16.269	248	8290	0.818	ng	# 80
22) Hexachlorobenzene	16.381	284	9007	0.801	ng	98
23) Atrazine	16.542	200	6777	0.764	ng	# 93
24) Pentachlorophenol	16.716	266	3059	0.693	ng	99
25) Phenanthrene	17.101	178	47181	0.856	ng	100
26) Anthracene	17.200	178	41660	0.845	ng	100
28) Fluoranthene	19.129	202	51141	0.827	ng	100
30) Pyrene	19.486	202	51944	0.963	ng	100
32) Benzo(a)anthracene	21.238	228	39570	0.860	ng	99
33) Chrysene	21.283	228	41243	0.906	ng	99
34) Bis(2-ethylhexyl)phtha...	21.203	149	20342	0.650	ng	# 99
36) Indeno(1,2,3-cd)pyrene	25.698	276	34379	0.828	ng	100
37) Benzo(b)fluoranthene	22.803	252	36922m	0.948	ng	
38) Benzo(k)fluoranthene	22.850	252	37118	0.977	ng	# 91
39) Benzo(a)pyrene	23.371	252	29490	0.892	ng	# 88
40) Dibenzo(a,h)anthracene	25.718	278	26850	0.823	ng	96
41) Benzo(g,h,i)perylene	26.373	276	29665	0.824	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034687.D
Acq On : 24 Oct 2024 10:59
Operator : RC/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Quant Time: Oct 24 12:43:27 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 12:41:40 2024
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.712 min Scan# 60
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:152 Resp: 6279
Ion Ratio Lower Upper
152 100
150 154.4 122.7 184.1
115 63.2 51.0 76.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#2
1,4-Dioxane
Concen: 0.895 ng
RT: 3.285 min Scan# 56
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion: 88 Resp: 6845
Ion Ratio Lower Upper
88 100
43 41.3 32.1 48.1
58 93.9 73.7 110.5









#3
n-Nitrosodimethylamine
Concen: 1.144 ng
RT: 3.581 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion: 42 Resp: 10574
Ion Ratio Lower Upper
42 100
74 95.9 77.2 115.8
44 9.3 13.4 20.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#4
2-Fluorophenol
Concen: 0.829 ng
RT: 5.321 min Scan# 338
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion:112 Resp: 15922
Ion Ratio Lower Upper
112 100
64 66.1 53.0 79.6
63 34.5 27.6 41.4









#5
Phenol-d6
Concen: 0.847 ng
RT: 6.881 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion: 99 Resp: 20820
Ion Ratio Lower Upper
99 100
42 27.0 21.8 32.8
71 33.1 26.6 40.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#6
bis(2-Chloroethyl)ether
Concen: 0.921 ng
RT: 7.141 min Scan# 590
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion: 93 Resp: 17336
Ion Ratio Lower Upper
93 100
63 88.1 71.5 107.3
95 32.1 25.8 38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 955
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion:	136	Resp:	18298
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.1	13.7
54	12.4	9.8	14.8
68	8.2	6.4	9.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#8
Nitrobenzene-d5
Concen: 0.854 ng
RT: 8.845 min Scan# 802
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion:	82	Resp:	12949
Ion	Ratio	Lower	Upper
82	100		
128	36.2	29.7	44.5
54	65.0	53.0	79.6









#9
Naphthalene
Concen: 0.857 ng
RT: 10.532 min Scan# 960
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion:128 Resp: 4325
Ion Ratio Lower Upper
128 100
129 11.1 9.0 13.6
127 12.8 10.5 15.7




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#10
Hexachlorobutadiene
Concen: 0.807 ng
RT: 10.831 min Scan# 988
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion:225 Resp: 6528
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.3 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 0.829 ng
RT: 12.071 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Tgt Ion:152 Resp: 2134
Ion Ratio Lower Upper
152 100
151 20.9 16.7 25.1




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#12
2-Methylnaphthalene
Concen: 0.836 ng
RT: 12.146 min Scan# 1221
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59


Tgt Ion:142 Resp: 26668
Ion Ratio Lower Upper
142 100
141 88.0 70.4 105.6
115 37.3 29.8 44.6





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.325 min Scan# 1504
Delta R.T. -0.004 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8




Tgt Ion:164 Resp: 878
Ion Ratio Lower Upper
164 100
162 106.2 81.4 122.0
160 53.4 40.6 61.0

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024




#14
2,4,6-Tribromophenol
Concen: 0.598 ng
RT: 15.810 min Scan# 1638
Delta R.T. -0.006 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59



Tgt Ion:330 Resp: 2227
Ion Ratio Lower Upper
330 100
332 96.3 76.6 115.0
141 67.4 54.6 81.8





#15

2-Fluorobiphenyl

Concen: 0.864 ng

RT: 12.957 min Scan# 11

Delta R.T. 0.007 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59

Instrument :

BNA_N

ClientSampleId :





SSTDIC0.8

Tgt Ion:	172	Resp:	29939
Ion Ratio	Lower	Upper	
172	100		
171	34.3	28.3	42.5
170	23.0	19.8	29.6

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#16

Acenaphthylene

Concen: 0.849 ng




RT: 14.047 min Scan# 1477


Delta R.T. 0.007 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59

Tgt Ion:	152	Resp:	35850
Ion Ratio	Lower	Upper	
152	100		
151	19.1	15.6	23.4
153	12.8	10.6	15.8









#17
Acenaphthene
Concen: 0.876 ng
RT: 14.389 min Scan# 1509
Delta R.T. -0.004 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion:154 Resp: 25224
Ion Ratio Lower Upper
154 100
153 109.5 88.2 132.2
152 57.2 46.9 70.3




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#18
Fluorene
Concen: 0.864 ng
RT: 15.373 min Scan# 1601
Delta R.T. -0.004 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion:166 Resp: 31002
Ion Ratio Lower Upper
166 100
165 98.5 78.8 118.2
167 13.3 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.063 min Scan# 11

Delta R.T. -0.006 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	15.9	8.9	13.3

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#20

4,6-Dinitro-2-methylphenol

Concen: 0.799 ng




RT: 15.448 min Scan# 1608


Delta R.T. -0.004 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59


Tgt Ion	Ratio	Lower	Upper
198	100		
51	79.0	91.8	137.8
105	52.7	44.3	66.5





#21
4-Bromophenyl-phenylether
Concen: 0.818 ng
RT: 16.269 min Scan# 1675
Delta R.T. -0.006 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:248 Resp: 8290
Ion Ratio Lower Upper
248 100
250 93.8 80.9 121.3
141 81.7 40.5 60.7




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#22
Hexachlorobenzene
Concen: 0.801 ng
RT: 16.381 min Scan# 1684
Delta R.T. 0.007 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59


Tgt Ion:284 Resp: 9007
Ion Ratio Lower Upper
284 100
142 56.0 43.6 65.4
249 34.1 26.6 40.0





#23
Atrazine
Concen: 0.764 ng
RT: 16.542 min Scan# 1710
Delta R.T. -0.006 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:200 Resp: 677
Ion Ratio Lower Upper
200 100
173 32.2 19.7 29.5
215 42.9 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#24
Pentachlorophenol
Concen: 0.693 ng
RT: 16.716 min Scan# 1711
Delta R.T. -0.006 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59


Tgt Ion:266 Resp: 3059
Ion Ratio Lower Upper
266 100
264 61.0 49.4 74.2
268 62.3 50.5 75.7





#25
Phenanthrene
Concen: 0.856 ng
RT: 17.101 min Scan# 1743
Delta R.T. -0.006 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:178 Resp: 4718
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.2
179 15.1 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#26
Anthracene
Concen: 0.845 ng
RT: 17.200 min Scan# 1750
Delta R.T. 0.007 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion:178 Resp: 41660
Ion Ratio Lower Upper
178 100
176 18.8 14.9 22.3
179 15.2 12.2 18.2





#27

Fluoranthene-d10

Concen: 0.812 ng

RT: 19.096 min Scan# 1981

Delta R.T. 0.002 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59

Instrument :

BNA_N

ClientSampleId :





SSTDICC0.8

Tgt Ion:	212	Resp:	36834
Ion Ratio	Lower	Upper	
212	100		
106	22.1	17.8	26.8
104	12.5	10.2	15.4

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#28

Fluoranthene

Concen: 0.827 ng




RT: 19.129 min Scan# 1988


Delta R.T. 0.002 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59


Tgt Ion:	202	Resp:	51141
Ion Ratio	Lower	Upper	
202	100		
101	16.0	12.8	19.2
203	17.2	13.5	20.3





#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.247 min Scan# 2419
Delta R.T. -0.003 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:240 Resp: 12060
Ion Ratio Lower Upper
240 100
120 24.6 14.2 21.2
236 30.0 23.1 34.7




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#30
Pyrene
Concen: 0.963 ng
RT: 19.486 min Scan# 2065
Delta R.T. 0.002 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion:202 Resp: 51944
Ion Ratio Lower Upper
202 100
200 21.0 16.6 25.0
203 17.5 14.2 21.4





#31
Terphenyl-d14
Concen: 0.896 ng
RT: 19.705 min Scan# 2113
Delta R.T. 0.002 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Instrument :

BNA_N





ClientSampleId :

SSTDICC0.8

Tgt Ion	Ratio	Lower	Upper
244	100		
212	10.6	8.6	13.0
122	26.3	22.3	33.5




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#32
Benzo(a)anthracene
Concen: 0.860 ng
RT: 21.238 min Scan# 2418
Delta R.T. 0.006 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.6	22.0	33.0
229	20.0	15.6	23.4





#33

Chrysene

Concen: 0.906 ng

RT: 21.283 min Scan# 2423

Delta R.T. -0.002 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8

Tgt Ion:228 Resp: 4124

Ion Ratio Lower Upper

228 100

226 30.8 24.1 36.1





229 19.1 15.8 23.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.650 ng

RT: 21.203 min Scan# 2414

Delta R.T. 0.006 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59




Tgt Ion:149 Resp: 20342


Ion Ratio Lower Upper

149 100

167 25.0 19.7 29.5


279 1.7 1.8 2.6#





#35
Perylene-d12
Concen: 0.400 ng
RT: 23.467 min Scan# 2913
Delta R.T. 0.001 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8



Tgt Ion:264 Resp: 10772
Ion Ratio Lower Upper
264 100
260 27.0 21.9 32.9
265 70.3 60.6 91.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.828 ng
RT: 25.698 min Scan# 3673
Delta R.T. 0.004 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion:276 Resp: 34379
Ion Ratio Lower Upper
276 100
138 43.6 35.1 52.7
277 24.4 19.7 29.5





#37

Benzo(b)fluoranthene

Concen: 0.948 ng m

RT: 22.803 min Scan# 2690

Delta R.T. 0.001 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8

Tgt Ion:252 Resp: 3692





Ion	Ratio	Lower	Upper
252	100		
253	22.5	20.0	30.0
125	20.2	20.9	31.3

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#38

Benzo(k)fluoranthene

Concen: 0.977 ng

RT: 22.850 min Scan# 2699




Delta R.T. 0.001 min


Lab File: BN034687.D

Acq: 24 Oct 2024 10:59

Tgt Ion:252 Resp: 37118

Ion	Ratio	Lower	Upper
252	100		
253	22.8	20.6	31.0
125	20.9	21.6	32.4





#39

Benzo(a)pyrene

Concen: 0.892 ng

RT: 23.371 min Scan# 21

Delta R.T. 0.004 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion:	252	Resp:	29490
Ion Ratio	Lower	Upper	
252	100		
253	23.4	22.2	33.4
125	23.9	26.0	39.0

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#40

Dibenzo(a,h)anthracene

Concen: 0.823 ng




RT: 25.718 min Scan# 3680


Delta R.T. 0.004 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59

Tgt Ion:	278	Resp:	26850
Ion Ratio	Lower	Upper	
278	100		
139	30.3	25.4	38.2
279	25.0	22.4	33.6





#41

Benzo(g,h,i)perylene

Concen: 0.824 ng

RT: 26.373 min Scan# 3907

Delta R.T. 0.001 min

Lab File: BN034687.D


Acq: 24 Oct 2024 10:59

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8



Tgt Ion:276 Resp: 2966

Ion Ratio Lower Upper

276 100

277 24.2 21.3 31.9



138 38.6 31.9 47.9

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
 Data File : BN034688.D
 Acq On : 24 Oct 2024 11:35
 Operator : RC/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Oct 24 12:43:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 24 12:41:40 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.712	152	6253	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	18478	0.400	ng	0.00
13) Acenaphthene-d10	14.329	164	9335	0.400	ng	0.00
19) Phenanthrene-d10	17.069	188	19334	0.400	ng	0.00
29) Chrysene-d12	21.250	240	11924	0.400	ng	0.00
35) Perylene-d12	23.467	264	10103	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.322	112	31333	1.638	ng	0.00
5) Phenol-d6	6.882	99	41454	1.694	ng	0.00
8) Nitrobenzene-d5	8.845	82	25808	1.686	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	43281	1.664	ng	0.00
14) 2,4,6-Tribromophenol	15.815	330	4968	1.255	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	62224	1.691	ng	0.00
27) Fluoranthene-d10	19.094	212	73515	1.553	ng	0.00
31) Terphenyl-d14	19.703	244	41261	1.788	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.285	88	13254	1.741	ng	99
3) n-Nitrosodimethylamine	3.581	42	20688	2.247	ng	# 98
6) bis(2-Chloroethyl)ether	7.142	93	33672	1.797	ng	99
9) Naphthalene	10.532	128	85545	1.679	ng	99
10) Hexachlorobutadiene	10.831	225	12572	1.540	ng	# 100
12) 2-Methylnaphthalene	12.147	142	53951	1.675	ng	99
16) Acenaphthylene	14.040	152	76104	1.696	ng	99
17) Acenaphthene	14.393	154	52549	1.718	ng	98
18) Fluorene	15.377	166	65352	1.714	ng	100
20) 4,6-Dinitro-2-methylph...	15.452	198	4583	1.734	ng	# 50
21) 4-Bromophenyl-phenylether	16.275	248	17079	1.615	ng	98
22) Hexachlorobenzene	16.374	284	18410	1.569	ng	99
23) Atrazine	16.548	200	14024	1.516	ng	100
24) Pentachlorophenol	16.721	266	6756	1.467	ng	99
25) Phenanthrene	17.106	178	96240	1.673	ng	100
26) Anthracene	17.193	178	87194	1.695	ng	100
28) Fluoranthene	19.127	202	103093	1.598	ng	100
30) Pyrene	19.489	202	103194	1.936	ng	100
32) Benzo(a)anthracene	21.232	228	76954	1.693	ng	100
33) Chrysene	21.286	228	78358	1.741	ng	99
34) Bis(2-ethylhexyl)phtha...	21.196	149	41892	1.354	ng	# 100
36) Indeno(1,2,3-cd)pyrene	25.697	276	65482	1.682	ng	99
37) Benzo(b)fluoranthene	22.803	252	68644	1.880	ng	# 90
38) Benzo(k)fluoranthene	22.850	252	69183	1.942	ng	# 88
39) Benzo(a)pyrene	23.370	252	55896	1.802	ng	# 85
40) Dibenzo(a,h)anthracene	25.715	278	51867	1.696	ng	95
41) Benzo(g,h,i)perylene	26.373	276	56042	1.661	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034688.D
Acq On : 24 Oct 2024 11:35
Operator : RC/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Quant Time: Oct 24 12:43:58 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 12:41:40 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.712 min Scan# 669
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 6253
Ion Ratio Lower Upper
152 100
150 154.5 122.7 184.1
115 63.8 51.0 76.6



#2
1,4-Dioxane
Concen: 1.741 ng
RT: 3.285 min Scan# 56
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35

Tgt Ion: 88 Resp: 13254
Ion Ratio Lower Upper
88 100
43 40.1 32.1 48.1
58 93.0 73.7 110.5









#3
n-Nitrosodimethylamine
Concen: 2.247 ng
RT: 3.581 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 20688
Ion Ratio Lower Upper
42 100
74 96.2 77.2 115.8
44 11.1 13.4 20.2#



#4
2-Fluorophenol
Concen: 1.638 ng
RT: 5.322 min Scan# 338
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35

Tgt Ion: 112 Resp: 31333
Ion Ratio Lower Upper
112 100
64 65.7 53.0 79.6
63 34.8 27.6 41.4









#5
Phenol-d6
Concen: 1.694 ng
RT: 6.882 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 99 Resp: 41454
Ion Ratio Lower Upper
99 100
42 27.2 21.8 32.8
71 33.2 26.6 40.0



#6
bis(2-Chloroethyl)ether
Concen: 1.797 ng
RT: 7.142 min Scan# 590
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35

Tgt Ion: 93 Resp: 33672
Ion Ratio Lower Upper
93 100
63 88.5 71.5 107.3
95 32.0 25.8 38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 955
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	18478
Ion	Ratio	Lower	Upper
136	100		
137	11.4	9.1	13.7
54	12.4	9.8	14.8
68	8.3	6.4	9.6



#8
Nitrobenzene-d5
Concen: 1.686 ng
RT: 8.845 min Scan# 802
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35

Tgt Ion:	82	Resp:	25808
Ion	Ratio	Lower	Upper
82	100		
128	35.7	29.7	44.5
54	64.5	53.0	79.6









#9
Naphthalene
Concen: 1.679 ng
RT: 10.532 min Scan# 960
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion	128	129	127
Resp:	85545	10.9	12.6
Ion Ratio	100	10.9	12.6
Lower		9.0	10.5
Upper		13.6	15.7



#10
Hexachlorobutadiene
Concen: 1.540 ng
RT: 10.831 min Scan# 988
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35

Tgt Ion	225	223	227
Resp:	12572	0.0	64.5
Ion Ratio	100	0.0	64.5
Lower		0.0	51.4
Upper		0.0	77.0









#11
2-Methylnaphthalene-d10
Concen: 1.664 ng
RT: 12.071 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 43281
Ion Ratio Lower Upper
152 100
151 20.8 16.7 25.1



#12
2-Methylnaphthalene
Concen: 1.675 ng
RT: 12.147 min Scan# 1221
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35

Tgt Ion:142 Resp: 53951
Ion Ratio Lower Upper
142 100
141 87.5 70.4 105.6
115 36.4 29.8 44.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.329 min Scan# 1504
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 9335
Ion Ratio Lower Upper
164 100
162 103.4 81.4 122.0
160 50.8 40.6 61.0



#14
2,4,6-Tribromophenol
Concen: 1.255 ng
RT: 15.815 min Scan# 1639
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35

Tgt Ion:330 Resp: 4968
Ion Ratio Lower Upper
330 100
332 95.6 76.6 115.0
141 66.4 54.6 81.8









#15
2-Fluorobiphenyl
Concen: 1.691 ng
RT: 12.950 min Scan# 1375
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.5	28.3	42.5
170	24.8	19.8	29.6



#16
Acenaphthylene
Concen: 1.696 ng
RT: 14.040 min Scan# 1477
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.6	23.4
153	12.9	10.6	15.8









#17
Acenaphthene
Concen: 1.718 ng
RT: 14.393 min Scan# 1510
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 52549
Ion Ratio Lower Upper
154 100
153 108.8 88.2 132.2
152 56.9 46.9 70.3



#18
Fluorene
Concen: 1.714 ng
RT: 15.377 min Scan# 1602
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35

Tgt Ion:166 Resp: 65352
Ion Ratio Lower Upper
166 100
165 98.1 78.8 118.2
167 13.2 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.069 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN034688.D

Acq: 24 Oct 2024 11:35

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:188 Resp: 19334

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.3 8.9 13.3



#20

4,6-Dinitro-2-methylphenol

Concen: 1.734 ng

RT: 15.452 min Scan# 1609

Delta R.T. 0.000 min

Lab File: BN034688.D

Acq: 24 Oct 2024 11:35




Tgt Ion:198 Resp: 4583


Ion Ratio Lower Upper

198 100

51 48.7 91.8 137.8#

105 35.4 44.3 66.5#









#21
4-Bromophenyl-phenylether
Concen: 1.615 ng
RT: 16.275 min Scan# 1676
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion	Ratio	Lower	Upper
248	100		
250	101.1	80.9	121.3
141	47.4	40.5	60.7



#22
Hexachlorobenzene
Concen: 1.569 ng
RT: 16.374 min Scan# 1684
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35

Tgt Ion	Ratio	Lower	Upper
284	100		
142	55.6	43.6	65.4
249	33.9	26.6	40.0









#23
Atrazine
Concen: 1.516 ng
RT: 16.548 min Scan# 1698
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:200 Resp: 14024
Ion Ratio Lower Upper
200 100
173 24.5 19.7 29.5
215 44.3 35.7 53.5



#24
Pentachlorophenol
Concen: 1.467 ng
RT: 16.721 min Scan# 1712
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35

Tgt Ion:266 Resp: 6756
Ion Ratio Lower Upper
266 100
264 62.2 49.4 74.2
268 63.5 50.5 75.7









#25
Phenanthrene
Concen: 1.673 ng
RT: 17.106 min Scan# 1743
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 96240
Ion Ratio Lower Upper
178 100
176 19.2 15.4 23.2
179 15.2 12.2 18.4



#26
Anthracene
Concen: 1.695 ng
RT: 17.193 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35

Tgt Ion:178 Resp: 87194
Ion Ratio Lower Upper
178 100
176 18.7 14.9 22.3
179 15.1 12.2 18.2





#27

Fluoranthene-d10

Concen: 1.553 ng

RT: 19.094 min Scan# 1982

Delta R.T. 0.000 min

Lab File: BN034688.D

Acq: 24 Oct 2024 11:35

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:212 Resp: 73515

Ion Ratio Lower Upper

212 100

106 22.2 17.8 26.8

104 12.6 10.2 15.4



#28

Fluoranthene

Concen: 1.598 ng

RT: 19.127 min Scan# 1989

Delta R.T. 0.000 min

Lab File: BN034688.D

Acq: 24 Oct 2024 11:35




Tgt Ion:202 Resp: 103093


Ion Ratio Lower Upper

202 100

101 16.0 12.8 19.2

203 17.2 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.250 min Scan# 2421

Delta R.T. 0.000 min

Lab File: BN034688.D

Acq: 24 Oct 2024 11:35

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:240 Resp: 11924

Ion Ratio Lower Upper

240 100

120 18.2 14.2 21.2

236 28.6 23.1 34.7



#30

Pyrene

Concen: 1.936 ng

RT: 19.489 min Scan# 2067

Delta R.T. 0.005 min

Lab File: BN034688.D

Acq: 24 Oct 2024 11:35




Tgt Ion:202 Resp: 103194


Ion Ratio Lower Upper

202 100

200 21.0 16.6 25.0

203 17.7 14.2 21.4









#31
Terphenyl-d14
Concen: 1.788 ng
RT: 19.703 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:244 Resp: 41261
Ion Ratio Lower Upper
244 100
212 10.5 8.6 13.0
122 28.7 22.3 33.5



#32
Benzo(a)anthracene
Concen: 1.693 ng
RT: 21.232 min Scan# 2419
Delta R.T. 0.000 min
Lab File: BN034688.D
Acq: 24 Oct 2024 11:35

Tgt Ion:228 Resp: 76954
Ion Ratio Lower Upper
228 100
226 27.4 22.0 33.0
229 19.1 15.6 23.4





#33

Chrysene

Concen: 1.741 ng

RT: 21.286 min Scan# 2425

Delta R.T. 0.000 min

Lab File: BN034688.D

Acq: 24 Oct 2024 11:35

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:228 Resp: 78358

Ion Ratio Lower Upper

228 100

226 30.4 24.1 36.1

229 19.1 15.8 23.8



#34

Bis(2-ethylhexyl)phthalate

Concen: 1.354 ng

RT: 21.196 min Scan# 2415

Delta R.T. 0.000 min

Lab File: BN034688.D

Acq: 24 Oct 2024 11:35




Tgt Ion:149 Resp: 41892


Ion Ratio Lower Upper

149 100

167 24.7 19.7 29.5

279 1.7 1.8 2.6#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.467 min Scan# 2913

Delta R.T. 0.000 min

Lab File: BN034688.D

Acq: 24 Oct 2024 11:35

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:264 Resp: 10103

Ion Ratio Lower Upper

264 100

260 26.9 21.9 32.9

265 75.7 60.6 91.0



#36

Indeno(1,2,3-cd)pyrene

Concen: 1.682 ng

RT: 25.697 min Scan# 3676

Delta R.T. 0.003 min

Lab File: BN034688.D

Acq: 24 Oct 2024 11:35




Tgt Ion:276 Resp: 65482


Ion Ratio Lower Upper

276 100

138 43.4 35.1 52.7

277 24.3 19.7 29.5





#37

Benzo(b)fluoranthene

Concen: 1.880 ng

RT: 22.803 min Scan# 2

Delta R.T. 0.000 min

Lab File: BN034688.D

Acq: 24 Oct 2024 11:35

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:252 Resp: 68644

Ion Ratio Lower Upper

252 100

253 22.3 20.0 30.0

125 18.9 20.9 31.3#



#38

Benzo(k)fluoranthene

Concen: 1.942 ng

RT: 22.850 min Scan# 2702

Delta R.T. 0.000 min

Lab File: BN034688.D

Acq: 24 Oct 2024 11:35




Tgt Ion:252 Resp: 69183


Ion Ratio Lower Upper

252 100

253 22.1 20.6 31.0

125 18.8 21.6 32.4#





#39

Benzo(a)pyrene

Concen: 1.802 ng

RT: 23.370 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN034688.D

Acq: 24 Oct 2024 11:35

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:252 Resp: 55896

Ion Ratio Lower Upper

252 100

253 22.8 22.2 33.4

125 21.7 26.0 39.0#



#40

Dibenzo(a,h)anthracene

Concen: 1.696 ng

RT: 25.715 min Scan# 3682

Delta R.T. 0.000 min

Lab File: BN034688.D

Acq: 24 Oct 2024 11:35




Tgt Ion:278 Resp: 51867


Ion Ratio Lower Upper

278 100

139 29.7 25.4 38.2

279 24.2 22.4 33.6





#41

Benzo(g,h,i)perylene

Concen: 1.661 ng

RT: 26.373 min Scan# 3907

Delta R.T. 0.000 min

Lab File: BN034688.D


Acq: 24 Oct 2024 11:35

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:276 Resp: 56042

Ion Ratio Lower Upper

276 100

277 23.8 21.3 31.9

138 37.5 31.9 47.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
 Data File : BN034689.D
 Acq On : 24 Oct 2024 12:11
 Operator : RC/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Oct 24 12:44:24 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 24 12:41:40 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.712	152	6787	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	19859	0.400	ng	0.00
13) Acenaphthene-d10	14.329	164	10274	0.400	ng	0.00
19) Phenanthrene-d10	17.069	188	21008	0.400	ng	0.00
29) Chrysene-d12	21.250	240	13427	0.400	ng	0.00
35) Perylene-d12	23.467	264	11211	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.322	112	63436	3.054	ng	0.00
5) Phenol-d6	6.882	99	85211	3.208	ng	0.00
8) Nitrobenzene-d5	8.845	82	53485	3.251	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	88547	3.167	ng	0.00
14) 2,4,6-Tribromophenol	15.815	330	11167	2.564	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	126562	3.125	ng	0.00
27) Fluoranthene-d10	19.094	212	153423	2.983	ng	0.00
31) Terphenyl-d14	19.703	244	88081	3.389	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.285	88	27324	3.306	ng	99
3) n-Nitrosodimethylamine	3.581	42	41310	4.134	ng	# 95
6) bis(2-Chloroethyl)ether	7.142	93	67465	3.316	ng	99
9) Naphthalene	10.532	128	173884	3.175	ng	99
10) Hexachlorobutadiene	10.831	225	25389	2.893	ng	# 99
12) 2-Methylnaphthalene	12.147	142	110779	3.200	ng	99
16) Acenaphthylene	14.040	152	161452	3.270	ng	99
17) Acenaphthene	14.393	154	109212	3.243	ng	98
18) Fluorene	15.377	166	136703	3.257	ng	99
20) 4,6-Dinitro-2-methylph...	15.452	198	10505	3.657	ng	# 44
21) 4-Bromophenyl-phenylether	16.275	248	35596	3.098	ng	96
22) Hexachlorobenzene	16.374	284	38020	2.981	ng	100
23) Atrazine	16.548	200	30485	3.032	ng	99
24) Pentachlorophenol	16.721	266	15457	3.088	ng	99
25) Phenanthrene	17.106	178	200178	3.203	ng	100
26) Anthracene	17.193	178	185283	3.315	ng	100
28) Fluoranthene	19.127	202	216788	3.093	ng	100
30) Pyrene	19.484	202	219062	3.649	ng	100
32) Benzo(a)anthracene	21.232	228	166793	3.259	ng	100
33) Chrysene	21.286	228	164703	3.250	ng	100
34) Bis(2-ethylhexyl)phtha...	21.196	149	98247	2.819	ng	# 99
36) Indeno(1,2,3-cd)pyrene	25.697	276	136074	3.150	ng	99
37) Benzo(b)fluoranthene	22.806	252	144801	3.574	ng	# 88
38) Benzo(k)fluoranthene	22.847	252	143101	3.620	ng	# 87
39) Benzo(a)pyrene	23.370	252	118654	3.448	ng	# 82
40) Dibenzo(a,h)anthracene	25.715	278	107743	3.175	ng	93
41) Benzo(g,h,i)perylene	26.373	276	115668	3.089	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034689.D
Acq On : 24 Oct 2024 12:11
Operator : RC/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Oct 24 12:44:24 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 12:41:40 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.712 min Scan# 60
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 6787
Ion Ratio Lower Upper
152 100
150 156.1 122.7 184.1
115 63.3 51.0 76.6



#2
1,4-Dioxane
Concen: 3.306 ng
RT: 3.285 min Scan# 56
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11

Tgt Ion: 88 Resp: 27324
Ion Ratio Lower Upper
88 100
43 39.4 32.1 48.1
58 92.6 73.7 110.5









#3
n-Nitrosodimethylamine
Concen: 4.134 ng
RT: 3.581 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 41310
Ion Ratio Lower Upper
42 100
74 94.5 77.2 115.8
44 8.2 13.4 20.2#



#4
2-Fluorophenol
Concen: 3.054 ng
RT: 5.322 min Scan# 338
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11

Tgt Ion: 112 Resp: 63436
Ion Ratio Lower Upper
112 100
64 65.8 53.0 79.6
63 34.8 27.6 41.4









#5
Phenol-d6
Concen: 3.208 ng
RT: 6.882 min Scan# 511
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 85211
Ion Ratio Lower Upper
99 100
42 26.8 21.8 32.8
71 32.9 26.6 40.0



#6
bis(2-Chloroethyl)ether
Concen: 3.316 ng
RT: 7.142 min Scan# 590
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11

Tgt Ion: 93 Resp: 67465
Ion Ratio Lower Upper
93 100
63 88.4 71.5 107.3
95 32.0 25.8 38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 955
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	19859
Ion	Ratio	Lower	Upper
136	100		
137	11.6	9.1	13.7
54	12.4	9.8	14.8
68	8.2	6.4	9.6



#8
Nitrobenzene-d5
Concen: 3.251 ng
RT: 8.845 min Scan# 802
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11

Tgt Ion:	82	Resp:	53485
Ion	Ratio	Lower	Upper
82	100		
128	35.5	29.7	44.5
54	64.9	53.0	79.6





#9

Naphthalene

Concen: 3.175 ng

RT: 10.532 min Scan# 960

Delta R.T. 0.000 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:128 Resp: 173884

Ion Ratio Lower Upper

128 100

129 10.9 9.0 13.6

127 12.7 10.5 15.7



#10

Hexachlorobutadiene

Concen: 2.893 ng

RT: 10.831 min Scan# 988

Delta R.T. 0.000 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11




Tgt Ion:225 Resp: 25389


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 63.7 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 3.167 ng
RT: 12.071 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 88547
Ion Ratio Lower Upper
152 100
151 20.7 16.7 25.1



#12
2-Methylnaphthalene
Concen: 3.200 ng
RT: 12.147 min Scan# 1221
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11

Tgt Ion:142 Resp: 110779
Ion Ratio Lower Upper
142 100
141 88.2 70.4 105.6
115 36.5 29.8 44.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.329 min Scan# 1504
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion	Ratio	Lower	Upper
164	100		
162	104.0	81.4	122.0
160	51.5	40.6	61.0



#14
2,4,6-Tribromophenol
Concen: 2.564 ng
RT: 15.815 min Scan# 1639
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11

Tgt Ion	Ratio	Lower	Upper
330	100		
332	96.2	76.6	115.0
141	67.0	54.6	81.8





#15

2-Fluorobiphenyl

Concen: 3.125 ng

RT: 12.950 min Scan# 1375

Delta R.T. 0.000 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:172 Resp: 126562

Ion Ratio Lower Upper

172 100

171 35.4 28.3 42.5

170 24.5 19.8 29.6



#16

Acenaphthylene

Concen: 3.270 ng

RT: 14.040 min Scan# 1477

Delta R.T. 0.000 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11




Tgt Ion:152 Resp: 161452


Ion Ratio Lower Upper

152 100

151 19.3 15.6 23.4

153 12.9 10.6 15.8









#17
Acenaphthene
Concen: 3.243 ng
RT: 14.393 min Scan# 1510
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 109212
Ion Ratio Lower Upper
154 100
153 108.5 88.2 132.2
152 55.7 46.9 70.3



#18
Fluorene
Concen: 3.257 ng
RT: 15.377 min Scan# 1602
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11

Tgt Ion:166 Resp: 136703
Ion Ratio Lower Upper
166 100
165 97.2 78.8 118.2
167 13.1 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.069 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:188 Resp: 21008

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.3 8.9 13.3



#20

4,6-Dinitro-2-methylphenol

Concen: 3.657 ng

RT: 15.452 min Scan# 1609

Delta R.T. 0.000 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11




Tgt Ion:198 Resp: 10505


Ion Ratio Lower Upper

198 100

51 40.4 91.8 137.8#

105 33.3 44.3 66.5#









#21
4-Bromophenyl-phenylether
Concen: 3.098 ng
RT: 16.275 min Scan# 1676
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion	Ratio	Lower	Upper
248	100		
250	103.5	80.9	121.3
141	46.6	40.5	60.7



#22
Hexachlorobenzene
Concen: 2.981 ng
RT: 16.374 min Scan# 1684
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11

Tgt Ion	Ratio	Lower	Upper
284	100		
142	55.0	43.6	65.4
249	33.4	26.6	40.0









#23
Atrazine
Concen: 3.032 ng
RT: 16.548 min Scan# 1698
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 30485
Ion Ratio Lower Upper
200 100
173 23.2 19.7 29.5
215 44.2 35.7 53.5



#24
Pentachlorophenol
Concen: 3.088 ng
RT: 16.721 min Scan# 1712
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11

Tgt Ion:266 Resp: 15457
Ion Ratio Lower Upper
266 100
264 62.8 49.4 74.2
268 64.0 50.5 75.7









#25
Phenanthrene
Concen: 3.203 ng
RT: 17.106 min Scan# 1743
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 200178
Ion Ratio Lower Upper
178 100
176 19.2 15.4 23.2
179 15.0 12.2 18.4



#26
Anthracene
Concen: 3.315 ng
RT: 17.193 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11

Tgt Ion:178 Resp: 185283
Ion Ratio Lower Upper
178 100
176 18.7 14.9 22.3
179 15.2 12.2 18.2





#27

Fluoranthene-d10

Concen: 2.983 ng

RT: 19.094 min Scan# 1982

Delta R.T. 0.000 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:212 Resp: 153423

Ion Ratio Lower Upper

212 100

106 22.2 17.8 26.8

104 12.6 10.2 15.4



#28

Fluoranthene

Concen: 3.093 ng

RT: 19.127 min Scan# 1989

Delta R.T. 0.000 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11




Tgt Ion:202 Resp: 216788


Ion Ratio Lower Upper

202 100

101 15.9 12.8 19.2

203 17.1 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.250 min Scan# 2421

Delta R.T. 0.000 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:240 Resp: 13427

Ion Ratio Lower Upper

240 100

120 17.2 14.2 21.2

236 28.9 23.1 34.7



#30

Pyrene

Concen: 3.649 ng

RT: 19.484 min Scan# 2066

Delta R.T. 0.000 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11




Tgt Ion:202 Resp: 219062


Ion Ratio Lower Upper

202 100

200 21.0 16.6 25.0

203 17.6 14.2 21.4









#31
Terphenyl-d14
Concen: 3.389 ng
RT: 19.703 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion	Ratio	Lower	Upper
244	100		
212	10.2	8.6	13.0
122	26.4	22.3	33.5



#32
Benzo(a)anthracene
Concen: 3.259 ng
RT: 21.232 min Scan# 2419
Delta R.T. 0.000 min
Lab File: BN034689.D
Acq: 24 Oct 2024 12:11

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.8	22.0	33.0
229	19.4	15.6	23.4





#33

Chrysene

Concen: 3.250 ng

RT: 21.286 min Scan# 2425

Delta R.T. 0.000 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:228 Resp: 164703

Ion Ratio Lower Upper

228 100

226 30.1 24.1 36.1

229 19.4 15.8 23.8



#34

Bis(2-ethylhexyl)phthalate

Concen: 2.819 ng

RT: 21.196 min Scan# 2415

Delta R.T. 0.000 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11




Tgt Ion:149 Resp: 98247


Ion Ratio Lower Upper

149 100

167 25.0 19.7 29.5

279 1.8 1.8 2.6#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.467 min Scan# 2913

Delta R.T. 0.000 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:264 Resp: 11211

Ion Ratio Lower Upper

264 100

260 27.2 21.9 32.9

265 70.8 60.6 91.0



#36

Indeno(1,2,3-cd)pyrene

Concen: 3.150 ng

RT: 25.697 min Scan# 3676

Delta R.T. 0.003 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11




Tgt Ion:276 Resp: 136074


Ion Ratio Lower Upper

276 100

138 43.2 35.1 52.7

277 24.6 19.7 29.5





#37

Benzo(b)fluoranthene

Concen: 3.574 ng

RT: 22.806 min Scan# 2

Delta R.T. 0.003 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:252 Resp: 144801

Ion Ratio Lower Upper

252 100

253 22.0 20.0 30.0

125 17.3 20.9 31.3#



#38

Benzo(k)fluoranthene

Concen: 3.620 ng

RT: 22.847 min Scan# 2701

Delta R.T. -0.003 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11




Tgt Ion:252 Resp: 143101


Ion Ratio Lower Upper

252 100

253 21.8 20.6 31.0

125 17.7 21.6 32.4#





#39

Benzo(a)pyrene

Concen: 3.448 ng

RT: 23.370 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:252 Resp: 118654

Ion Ratio Lower Upper

252 100

253 21.8 22.2 33.4#

125 19.5 26.0 39.0#



#40

Dibenzo(a,h)anthracene

Concen: 3.175 ng

RT: 25.715 min Scan# 3682

Delta R.T. 0.000 min

Lab File: BN034689.D

Acq: 24 Oct 2024 12:11




Tgt Ion:278 Resp: 107743


Ion Ratio Lower Upper

278 100

139 29.1 25.4 38.2

279 23.4 22.4 33.6





#41

Benzo(g,h,i)perylene

Concen: 3.089 ng

RT: 26.373 min Scan# 3907

Delta R.T. 0.000 min

Lab File: BN034689.D


Acq: 24 Oct 2024 12:11

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:276 Resp: 115668

Ion Ratio Lower Upper

276 100

277 23.9 21.3 31.9

138 37.1 31.9 47.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
 Data File : BN034690.D
 Acq On : 24 Oct 2024 12:48
 Operator : RC/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Oct 24 13:13:27 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 24 12:41:40 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.712	152	7073	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	20368	0.400	ng	0.00
13) Acenaphthene-d10	14.325	164	10267	0.400	ng	0.00
19) Phenanthrene-d10	17.064	188	20306	0.400	ng	# 0.00
29) Chrysene-d12	21.248	240	12752	0.400	ng	0.00
35) Perylene-d12	23.464	264	10972	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.322	112	104062	4.808	ng	0.00
5) Phenol-d6	6.882	99	141189	5.100	ng	0.00
8) Nitrobenzene-d5	8.846	82	88506	5.246	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	144128	5.026	ng	0.00
14) 2,4,6-Tribromophenol	15.810	330	18762	4.310	ng	0.00
15) 2-Fluorobiphenyl	12.957	172	205465	5.077	ng	0.00
27) Fluoranthene-d10	19.097	212	231016	4.647	ng	0.00
31) Terphenyl-d14	19.705	244	130770	5.298	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.285	88	42348	4.917	ng	99
3) n-Nitrosodimethylamine	3.581	42	66202	6.357	ng	# 96
6) bis(2-Chloroethyl)ether	7.142	93	107896	5.090	ng	99
9) Naphthalene	10.532	128	280398	4.992	ng	99
10) Hexachlorobutadiene	10.831	225	40523	4.502	ng	# 99
12) 2-Methylnaphthalene	12.147	142	179479	5.055	ng	99
16) Acenaphthylene	14.047	152	264846	5.368	ng	99
17) Acenaphthene	14.390	154	175767	5.223	ng	97
18) Fluorene	15.373	166	213452	5.090	ng	99
20) 4,6-Dinitro-2-methylph...	15.448	198	17103	6.160	ng	# 61
21) 4-Bromophenyl-phenylether	16.269	248	55507	4.998	ng	# 81
22) Hexachlorobenzene	16.381	284	58013	4.706	ng	99
23) Atrazine	16.542	200	45354	4.667	ng	# 92
24) Pentachlorophenol	16.716	266	25315	5.233	ng	99
25) Phenanthrene	17.101	178	308331	5.103	ng	99
26) Anthracene	17.200	178	287167	5.315	ng	100
28) Fluoranthene	19.124	202	322212	4.756	ng	100
30) Pyrene	19.487	202	326483	5.726	ng	100
32) Benzo(a)anthracene	21.230	228	250242	5.148	ng	100
33) Chrysene	21.283	228	244106	5.072	ng	100
34) Bis(2-ethylhexyl)phtha...	21.194	149	164213	4.961	ng	99
36) Indeno(1,2,3-cd)pyrene	25.695	276	222798	5.271	ng	99
37) Benzo(b)fluoranthene	22.801	252	221458	5.585	ng	# 88
38) Benzo(k)fluoranthene	22.847	252	221207	5.717	ng	# 86
39) Benzo(a)pyrene	23.371	252	185096	5.496	ng	# 82
40) Dibenzo(a,h)anthracene	25.716	278	176660	5.319	ng	93
41) Benzo(g,h,i)perylene	26.370	276	189041	5.158	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034690.D
Acq On : 24 Oct 2024 12:48
Operator : RC/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Oct 24 13:13:27 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 12:41:40 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.712 min Scan# 60
Delta R.T. 0.000 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 7073
Ion Ratio Lower Upper
152 100
150 154.5 122.7 184.1
115 62.9 51.0 76.6



#2
1,4-Dioxane
Concen: 4.917 ng
RT: 3.285 min Scan# 56
Delta R.T. 0.000 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48

Tgt Ion: 88 Resp: 42348
Ion Ratio Lower Upper
88 100
43 39.3 32.1 48.1
58 92.9 73.7 110.5









#3
n-Nitrosodimethylamine
Concen: 6.357 ng
RT: 3.581 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 66202
Ion Ratio Lower Upper
42 100
74 95.6 77.2 115.8
44 7.1 13.4 20.2#



#4
2-Fluorophenol
Concen: 4.808 ng
RT: 5.322 min Scan# 338
Delta R.T. 0.000 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48

Tgt Ion: 112 Resp: 104062
Ion Ratio Lower Upper
112 100
64 66.0 53.0 79.6
63 34.7 27.6 41.4









#5
Phenol-d6
Concen: 5.100 ng
RT: 6.882 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 141189
Ion Ratio Lower Upper
99 100
42 26.5 21.8 32.8
71 32.9 26.6 40.0



#6
bis(2-Chloroethyl)ether
Concen: 5.090 ng
RT: 7.142 min Scan# 590
Delta R.T. 0.000 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48

Tgt Ion: 93 Resp: 107896
Ion Ratio Lower Upper
93 100
63 88.8 71.5 107.3
95 32.1 25.8 38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	20368
Ion Ratio	Lower	Upper	
136	100		
137	11.6	9.1	13.7
54	12.6	9.8	14.8
68	8.0	6.4	9.6



#8
Nitrobenzene-d5
Concen: 5.246 ng
RT: 8.846 min Scan# 802
Delta R.T. 0.000 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48

Tgt Ion:	82	Resp:	88506
Ion Ratio	Lower	Upper	
82	100		
128	36.8	29.7	44.5
54	63.7	53.0	79.6





#9

Naphthalene

Concen: 4.992 ng

RT: 10.532 min Scan# 960

Delta R.T. 0.000 min

Lab File: BN034690.D

Acq: 24 Oct 2024 12:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:128 Resp: 280398

Ion Ratio Lower Upper

128 100

129 10.9 9.0 13.6

127 12.7 10.5 15.7



#10

Hexachlorobutadiene

Concen: 4.502 ng

RT: 10.831 min Scan# 988

Delta R.T. 0.000 min

Lab File: BN034690.D

Acq: 24 Oct 2024 12:48




Tgt Ion:225 Resp: 40523


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 63.3 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 5.026 ng
RT: 12.071 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 144128
Ion Ratio Lower Upper
152 100
151 20.6 16.7 25.1



#12
2-Methylnaphthalene
Concen: 5.055 ng
RT: 12.147 min Scan# 1221
Delta R.T. 0.000 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48

Tgt Ion:142 Resp: 179479
Ion Ratio Lower Upper
142 100
141 88.5 70.4 105.6
115 36.6 29.8 44.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.325 min Scan# 1504
Delta R.T. -0.004 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 10267
Ion Ratio Lower Upper
164 100
162 105.4 81.4 122.0
160 52.4 40.6 61.0



#14
2,4,6-Tribromophenol
Concen: 4.310 ng
RT: 15.810 min Scan# 1638
Delta R.T. -0.005 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48

Tgt Ion:330 Resp: 18762
Ion Ratio Lower Upper
330 100
332 95.4 76.6 115.0
141 65.7 54.6 81.8









#15
2-Fluorobiphenyl
Concen: 5.077 ng
RT: 12.957 min Scan# 1375
Delta R.T. 0.007 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:172 Resp: 205465
Ion Ratio Lower Upper
172 100
171 34.2 28.3 42.5
170 22.8 19.8 29.6



#16
Acenaphthylene
Concen: 5.368 ng
RT: 14.047 min Scan# 1477
Delta R.T. 0.007 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48

Tgt Ion:152 Resp: 264846
Ion Ratio Lower Upper
152 100
151 19.3 15.6 23.4
153 12.9 10.6 15.8









#17
Acenaphthene
Concen: 5.223 ng
RT: 14.390 min Scan# 1510
Delta R.T. -0.004 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 175767
Ion Ratio Lower Upper
154 100
153 108.2 88.2 132.2
152 54.7 46.9 70.3



#18
Fluorene
Concen: 5.090 ng
RT: 15.373 min Scan# 1601
Delta R.T. -0.004 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48

Tgt Ion:166 Resp: 213452
Ion Ratio Lower Upper
166 100
165 97.4 78.8 118.2
167 13.3 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.064 min Scan# 11

Delta R.T. -0.005 min

Lab File: BN034690.D

Acq: 24 Oct 2024 12:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:188 Resp: 20306

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 17.1 8.9 13.3#



#20

4,6-Dinitro-2-methylphenol

Concen: 6.160 ng

RT: 15.448 min Scan# 1608

Delta R.T. -0.004 min

Lab File: BN034690.D

Acq: 24 Oct 2024 12:48




Tgt Ion:198 Resp: 17103


Ion Ratio Lower Upper

198 100

51 58.7 91.8 137.8#

105 47.2 44.3 66.5









#21
4-Bromophenyl-phenylether
Concen: 4.998 ng
RT: 16.269 min Scan# 1675
Delta R.T. -0.005 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:248 Resp: 55507
Ion Ratio Lower Upper
248 100
250 95.0 80.9 121.3
141 81.9 40.5 60.7#



#22
Hexachlorobenzene
Concen: 4.706 ng
RT: 16.381 min Scan# 1684
Delta R.T. 0.007 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48

Tgt Ion:284 Resp: 58013
Ion Ratio Lower Upper
284 100
142 55.3 43.6 65.4
249 33.7 26.6 40.0





#23

Atrazine

Concen: 4.667 ng

RT: 16.542 min Scan# 1

Delta R.T. -0.005 min

Lab File: BN034690.D

Acq: 24 Oct 2024 12:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:200 Resp: 45354

Ion Ratio Lower Upper

200 100

173 31.6 19.7 29.5#

215 41.9 35.7 53.5



#24

Pentachlorophenol

Concen: 5.233 ng

RT: 16.716 min Scan# 1711

Delta R.T. -0.005 min

Lab File: BN034690.D

Acq: 24 Oct 2024 12:48




Tgt Ion:266 Resp: 25315


Ion Ratio Lower Upper

266 100

264 63.1 49.4 74.2

268 63.1 50.5 75.7









#25
Phenanthrene
Concen: 5.103 ng
RT: 17.101 min Scan# 1742
Delta R.T. -0.005 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:178 Resp: 308331
Ion Ratio Lower Upper
178 100
176 19.1 15.4 23.2
179 15.0 12.2 18.4



#26
Anthracene
Concen: 5.315 ng
RT: 17.200 min Scan# 1750
Delta R.T. 0.007 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48

Tgt Ion:178 Resp: 287167
Ion Ratio Lower Upper
178 100
176 18.7 14.9 22.3
179 15.1 12.2 18.2





#27

Fluoranthene-d10

Concen: 4.647 ng

RT: 19.097 min Scan# 1982

Delta R.T. 0.002 min

Lab File: BN034690.D

Acq: 24 Oct 2024 12:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:212 Resp: 231016

Ion Ratio Lower Upper

212 100

106 22.3 17.8 26.8

104 12.6 10.2 15.4



#28

Fluoranthene

Concen: 4.756 ng

RT: 19.124 min Scan# 1987

Delta R.T. -0.002 min

Lab File: BN034690.D

Acq: 24 Oct 2024 12:48




Tgt Ion:202 Resp: 322212


Ion Ratio Lower Upper

202 100

101 15.9 12.8 19.2

203 17.2 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.248 min Scan# 2419

Delta R.T. -0.002 min

Lab File: BN034690.D

Acq: 24 Oct 2024 12:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:240 Resp: 12752

Ion Ratio Lower Upper

240 100

120 19.6 14.2 21.2

236 29.0 23.1 34.7



#30

Pyrene

Concen: 5.726 ng

RT: 19.487 min Scan# 2065

Delta R.T. 0.002 min

Lab File: BN034690.D

Acq: 24 Oct 2024 12:48




Tgt Ion:202 Resp: 326483


Ion Ratio Lower Upper

202 100

200 21.0 16.6 25.0

203 17.6 14.2 21.4









#31
Terphenyl-d14
Concen: 5.298 ng
RT: 19.705 min Scan# 2112
Delta R.T. 0.002 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:244 Resp: 130770
Ion Ratio Lower Upper
244 100
212 10.1 8.6 13.0
122 24.5 22.3 33.5



#32
Benzo(a)anthracene
Concen: 5.148 ng
RT: 21.230 min Scan# 2417
Delta R.T. -0.002 min
Lab File: BN034690.D
Acq: 24 Oct 2024 12:48

Tgt Ion:228 Resp: 250242
Ion Ratio Lower Upper
228 100
226 27.6 22.0 33.0
229 19.1 15.6 23.4





#33

Chrysene

Concen: 5.072 ng

RT: 21.283 min Scan# 24

Delta R.T. -0.002 min

Lab File: BN034690.D

Acq: 24 Oct 2024 12:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:228 Resp: 244106

Ion Ratio Lower Upper

228 100

226 30.2 24.1 36.1

229 19.4 15.8 23.8



#34

Bis(2-ethylhexyl)phthalate

Concen: 4.961 ng

RT: 21.194 min Scan# 2413

Delta R.T. -0.002 min

Lab File: BN034690.D

Acq: 24 Oct 2024 12:48




Tgt Ion:149 Resp: 164213


Ion Ratio Lower Upper

149 100

167 25.1 19.7 29.5

279 1.8 1.8 2.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.464 min Scan# 2911

Delta R.T. -0.002 min

Lab File: BN034690.D


Acq: 24 Oct 2024 12:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0






Tgt Ion:264 Resp: 10972

Ion Ratio Lower Upper

264 100

260 26.6 21.9 32.9

265 75.9 60.6 91.0



#36

Indeno(1,2,3-cd)pyrene

Concen: 5.271 ng

RT: 25.695 min Scan# 3674

Delta R.T. 0.001 min

Lab File: BN034690.D

Acq: 24 Oct 2024 12:48




Tgt Ion:276 Resp: 222798


Ion Ratio Lower Upper

276 100

138 43.1 35.1 52.7

277 24.7 19.7 29.5





#37

Benzo(b)fluoranthene

Concen: 5.585 ng

RT: 22.801 min Scan# 2

Delta R.T. -0.002 min

Lab File: BN034690.D

Acq: 24 Oct 2024 12:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:252 Resp: 221458

Ion Ratio Lower Upper

252 100

253 21.9 20.0 30.0

125 17.2 20.9 31.3#



#38

Benzo(k)fluoranthene

Concen: 5.717 ng

RT: 22.847 min Scan# 2700

Delta R.T. -0.002 min

Lab File: BN034690.D

Acq: 24 Oct 2024 12:48




Tgt Ion:252 Resp: 221207


Ion Ratio Lower Upper

252 100

253 21.6 20.6 31.0

125 17.1 21.6 32.4#





#39

Benzo(a)pyrene

Concen: 5.496 ng

RT: 23.371 min Scan# 21

Delta R.T. 0.004 min

Lab File: BN034690.D

Acq: 24 Oct 2024 12:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:252 Resp: 185096

Ion Ratio Lower Upper

252 100

253 21.8 22.2 33.4#

125 19.1 26.0 39.0#



#40

Dibenzo(a,h)anthracene

Concen: 5.319 ng

RT: 25.716 min Scan# 3681

Delta R.T. 0.001 min

Lab File: BN034690.D

Acq: 24 Oct 2024 12:48




Tgt Ion:278 Resp: 176660


Ion Ratio Lower Upper

278 100

139 28.9 25.4 38.2

279 23.4 22.4 33.6





#41

Benzo(g,h,i)perylene

Concen: 5.158 ng

RT: 26.370 min Scan# 3907

Delta R.T. -0.002 min

Lab File: BN034690.D


Acq: 24 Oct 2024 12:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:276 Resp: 189041

Ion Ratio Lower Upper

276 100

277 23.7 21.3 31.9

138 36.9 31.9 47.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
 Data File : BN034691.D
 Acq On : 24 Oct 2024 14:14
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

ICVBN102424

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024

Quant Time: Oct 24 14:43:51 2024

Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Oct 24 13:17:02 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.705	152	6674	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	19848	0.400	ng	0.00
13) Acenaphthene-d10	14.329	164	10372	0.400	ng	0.00
19) Phenanthrene-d10	17.069	188	21015	0.400	ng	0.00
29) Chrysene-d12	21.250	240	12401	0.400	ng	0.00
35) Perylene-d12	23.466	264	10594	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.314	112	7964	0.401	ng	0.00
5) Phenol-d6	6.882	99	10366	0.396	ng	0.00
8) Nitrobenzene-d5	8.845	82	5758	0.346	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	9741	0.357	ng	0.00
14) 2,4,6-Tribromophenol	15.815	330	1099	0.343	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	14035	0.341	ng	0.00
27) Fluoranthene-d10	19.094	212	16566	0.350	ng	0.00
31) Terphenyl-d14	19.703	244	9400	0.371	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.285	88	3154	0.355	ng	100
3) n-Nitrosodimethylamine	3.581	42	4576	0.354	ng	# 97
6) bis(2-Chloroethyl)ether	7.141	93	7850	0.369	ng	99
9) Naphthalene	10.532	128	19666	0.358	ng	100
10) Hexachlorobutadiene	10.831	225	2964	0.360	ng	# 99
12) 2-Methylnaphthalene	12.143	142	12146	0.356	ng	98
16) Acenaphthylene	14.040	152	16384	0.329	ng	99
17) Acenaphthene	14.382	154	11510	0.335	ng	99
18) Fluorene	15.377	166	14579	0.344	ng	99
20) 4,6-Dinitro-2-methylph...	15.451	198	947	0.324	ng	88
21) 4-Bromophenyl-phenylether	16.275	248	3884	0.347	ng	98
22) Hexachlorobenzene	16.374	284	4355	0.358	ng	98
23) Atrazine	16.548	200	2928	0.333	ng	99
24) Pentachlorophenol	16.721	266	1463	0.340	ng	98
25) Phenanthrene	17.106	178	22550	0.357	ng	100
26) Anthracene	17.193	178	19340	0.345	ng	100
28) Fluoranthene	19.127	202	22945	0.348	ng	100
30) Pyrene	19.489	202	23297	0.366	ng	100
32) Benzo(a)anthracene	21.232	228	16543	0.349	ng	99
33) Chrysene	21.286	228	17696	0.361	ng	99
34) Bis(2-ethylhexyl)phtha...	21.196	149	8661m	0.322	ng	
36) Indeno(1,2,3-cd)pyrene	25.697	276	14157	0.347	ng	100
37) Benzo(b)fluoranthene	22.803	252	15350	0.363	ng	99
38) Benzo(k)fluoranthene	22.850	252	14742	0.353	ng	99
39) Benzo(a)pyrene	23.373	252	12054	0.355	ng	99
40) Dibenzo(a,h)anthracene	25.718	278	11142	0.346	ng	99
41) Benzo(g,h,i)perylene	26.373	276	12576	0.354	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034691.D
Acq On : 24 Oct 2024 14:14
Operator : RC/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

ICVBN102424

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024


Quant Time: Oct 24 14:43:51 2024


Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Oct 24 13:17:02 2024


Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.705 min Scan# 60
Delta R.T. -0.007 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424



Tgt Ion:152 Resp: 6674
Ion Ratio Lower Upper
152 100
150 157.2 122.7 184.1
115 65.2 51.0 76.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#2
1,4-Dioxane
Concen: 0.355 ng
RT: 3.285 min Scan# 56
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion: 88 Resp: 3154
Ion Ratio Lower Upper
88 100
43 39.5 32.1 48.1
58 91.9 73.7 110.5









#3
n-Nitrosodimethylamine
Concen: 0.354 ng
RT: 3.581 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Instrument :
BNA_N
ClientSampleId :
ICVBN102424

Tgt Ion: 42 Resp: 4570
Ion Ratio Lower Upper
42 100
74 98.6 77.2 115.8
44 20.5 13.4 20.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#4
2-Fluorophenol
Concen: 0.401 ng
RT: 5.314 min Scan# 337
Delta R.T. -0.007 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion:112 Resp: 7964
Ion Ratio Lower Upper
112 100
64 66.0 53.0 79.6
63 35.2 27.6 41.4









#5
Phenol-d6
Concen: 0.396 ng
RT: 6.882 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Instrument :
BNA_N
ClientSampleId :
ICVBN102424

Tgt Ion: 99 Resp: 10360
Ion Ratio Lower Upper
99 100
42 27.0 21.8 32.8
71 33.3 26.6 40.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#6
bis(2-Chloroethyl)ether
Concen: 0.369 ng
RT: 7.141 min Scan# 590
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion: 93 Resp: 7850
Ion Ratio Lower Upper
93 100
63 87.8 71.5 107.3
95 32.2 25.8 38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 955
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Instrument :
BNA_N
ClientSampleId :
ICVBN102424

Tgt Ion:	136	Resp:	1984
Ion Ratio	Lower	Upper	
136	100		
137	11.3	9.1	13.7
54	11.4	9.8	14.8
68	7.5	6.4	9.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#8
Nitrobenzene-d5
Concen: 0.346 ng
RT: 8.845 min Scan# 802
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion:	82	Resp:	5758
Ion Ratio	Lower	Upper	
82	100		
128	39.9	29.7	44.5
54	63.8	53.0	79.6



#9
Naphthalene

Concen: 0.358 ng

RT: 10.532 min Scan# 960

Delta R.T. -0.000 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14

Instrument :

BNA_N

ClientSampleId :





ICVBN102424

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.4	9.0	13.6
127	12.8	10.5	15.7

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#10

Hexachlorobutadiene

Concen: 0.360 ng




RT: 10.831 min Scan# 988


Delta R.T. -0.000 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14


Tgt Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.4	51.4	77.0





#11
2-Methylnaphthalene-d10
Concen: 0.357 ng
RT: 12.071 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424




Tgt Ion:152 Resp: 974
Ion Ratio Lower Upper
152 100
151 20.7 16.7 25.1

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024




#12
2-Methylnaphthalene
Concen: 0.356 ng
RT: 12.143 min Scan# 1220
Delta R.T. -0.004 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Tgt Ion:142 Resp: 12146
Ion Ratio Lower Upper
142 100
141 89.4 70.4 105.6
115 38.7 29.8 44.6





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.329 min Scan# 1504
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424




Tgt Ion:164 Resp: 10372
Ion Ratio Lower Upper
164 100
162 101.1 81.4 122.0
160 50.2 40.6 61.0

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024




#14
2,4,6-Tribromophenol
Concen: 0.343 ng
RT: 15.815 min Scan# 1639
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14



Tgt Ion:330 Resp: 1099
Ion Ratio Lower Upper
330 100
332 96.5 76.6 115.0
141 66.2 54.6 81.8





#15

2-Fluorobiphenyl

Concen: 0.341 ng

RT: 12.950 min Scan# 1375

Delta R.T. -0.000 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14

Instrument :

BNA_N

ClientSampleId :





ICVBN102424

Tgt Ion:	172	Resp:	1403
Ion Ratio	Lower	Upper	
172	100		
171	35.0	28.3	42.5
170	24.5	19.8	29.6

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#16

Acenaphthylene

Concen: 0.329 ng




RT: 14.040 min Scan# 1477


Delta R.T. -0.000 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14


Tgt Ion:	152	Resp:	16384
Ion Ratio	Lower	Upper	
152	100		
151	19.0	15.6	23.4
153	13.1	10.6	15.8





#17
Acenaphthene
Concen: 0.335 ng
RT: 14.382 min Scan# 1510
Delta R.T. -0.011 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424



Tgt Ion:154 Resp: 11510
Ion Ratio Lower Upper
154 100
153 111.6 88.2 132.2
152 58.6 46.9 70.3




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#18
Fluorene
Concen: 0.344 ng
RT: 15.377 min Scan# 1602
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion:166 Resp: 14579
Ion Ratio Lower Upper
166 100
165 99.0 78.8 118.2
167 13.5 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.069 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14

Instrument :

BNA_N

ClientSampleId :





ICVBN102424

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	11.5	8.9	13.3

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#20

4,6-Dinitro-2-methylphenol

Concen: 0.324 ng




RT: 15.451 min Scan# 1609


Delta R.T. -0.001 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14

Tgt Ion	Ratio	Lower	Upper
198	100		
51	100.2	91.8	137.8
105	48.7	44.3	66.5





#21
4-Bromophenyl-phenylether
Concen: 0.347 ng
RT: 16.275 min Scan# 1676
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Instrument :

BNA_N





ClientSampleId :

ICVBN102424

Tgt Ion	Ratio	Lower	Upper
248	100		
250	103.8	80.9	121.3
141	50.3	40.5	60.7




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#22
Hexachlorobenzene
Concen: 0.358 ng
RT: 16.374 min Scan# 1684
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14


Tgt Ion	Ratio	Lower	Upper
284	100		
142	55.6	43.6	65.4
249	34.2	26.6	40.0





#23
Atrazine
Concen: 0.333 ng
RT: 16.548 min Scan# 1698
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424



Tgt Ion:200 Resp: 2928
Ion Ratio Lower Upper
200 100
173 25.5 19.7 29.5
215 45.1 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#24
Pentachlorophenol
Concen: 0.340 ng
RT: 16.721 min Scan# 1712
Delta R.T. 0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14


Tgt Ion:266 Resp: 1463
Ion Ratio Lower Upper
266 100
264 61.6 49.4 74.2
268 61.0 50.5 75.7





#25
Phenanthrene
Concen: 0.357 ng
RT: 17.106 min Scan# 1743
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424



Tgt Ion:178 Resp: 22550
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.2
179 15.4 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#26
Anthracene
Concen: 0.345 ng
RT: 17.193 min Scan# 1750
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion:178 Resp: 19340
Ion Ratio Lower Upper
178 100
176 18.5 14.9 22.3
179 15.1 12.2 18.2





#27

Fluoranthene-d10

Concen: 0.350 ng

RT: 19.094 min Scan# 1982

Delta R.T. -0.000 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14

Instrument :

BNA_N

ClientSampleId :





ICVBN102424

Tgt Ion:	212	Resp:	1656
Ion Ratio	Lower	Upper	
212	100		
106	22.2	17.8	26.8
104	13.1	10.2	15.4

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#28

Fluoranthene

Concen: 0.348 ng




RT: 19.127 min Scan# 1989


Delta R.T. -0.000 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14

Tgt Ion:	202	Resp:	22945
Ion Ratio	Lower	Upper	
202	100		
101	16.0	12.8	19.2
203	17.1	13.5	20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.250 min Scan# 2421

Delta R.T. -0.000 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14

Instrument :

BNA_N





ClientSampleId :

ICVBN102424

Tgt Ion:	240	Resp:	1240
Ion Ratio	Lower	Upper	
240	100		
120	18.6	14.2	21.2
236	28.6	23.1	34.7

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#30

Pyrene

Concen: 0.366 ng




RT: 19.489 min Scan# 2067


Delta R.T. 0.005 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14


Tgt Ion:	202	Resp:	23297
Ion Ratio	Lower	Upper	
202	100		
200	20.9	16.6	25.0
203	17.7	14.2	21.4





#31
Terphenyl-d14
Concen: 0.371 ng
RT: 19.703 min Scan# 2113
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424



Tgt Ion:244 Resp: 9400
Ion Ratio Lower Upper
244 100
212 11.1 8.6 13.0
122 27.4 22.3 33.5




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#32
Benzo(a)anthracene
Concen: 0.349 ng
RT: 21.232 min Scan# 2419
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion:228 Resp: 16543
Ion Ratio Lower Upper
228 100
226 28.1 22.0 33.0
229 19.1 15.6 23.4





#33

Chrysene

Concen: 0.361 ng

RT: 21.286 min Scan# 2425

Delta R.T. -0.000 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14





Instrument :

BNA_N

ClientSampleId :

ICVBN102424

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.6	24.1	36.1
229	19.6	15.8	23.8

Manual Integrations
APPROVEDReviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024

#34

Bis(2-ethylhexyl)phthalate

Concen: 0.322 ng m




RT: 21.196 min Scan# 2415


Delta R.T. 0.000 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14


Tgt Ion	Ratio	Lower	Upper
149	100		
167	23.9	19.7	29.5
279	2.2	1.8	2.6





#35
Perylene-d12
Concen: 0.400 ng
RT: 23.466 min Scan# 2913
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424



Tgt Ion:264 Resp: 10594
Ion Ratio Lower Upper
264 100
260 27.4 21.9 32.9
265 72.9 60.6 91.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.347 ng
RT: 25.697 min Scan# 3676
Delta R.T. 0.003 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion:276 Resp: 14157
Ion Ratio Lower Upper
276 100
138 43.9 35.1 52.7
277 25.2 19.7 29.5





#37

Benzo(b)fluoranthene

Concen: 0.363 ng

RT: 22.803 min Scan# 20

Delta R.T. -0.000 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14

Instrument :

BNA_N

ClientSampleId :





ICVBN102424

Tgt Ion:	252	Resp:	15350
Ion Ratio	Lower	Upper	
252	100		
253	24.9	20.0	30.0
125	25.2	20.9	31.3

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#38

Benzo(k)fluoranthene

Concen: 0.353 ng




RT: 22.850 min Scan# 2702


Delta R.T. -0.000 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14

Tgt Ion:	252	Resp:	14742
Ion Ratio	Lower	Upper	
252	100		
253	25.6	20.6	31.0
125	26.3	21.6	32.4





#39

Benzo(a)pyrene

Concen: 0.355 ng

RT: 23.373 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14

Instrument :

BNA_N

ClientSampleId :





ICVBN102424

Tgt Ion:	252	Resp:	12054
Ion Ratio	Lower	Upper	
252	100		
253	26.9	22.2	33.4
125	32.0	26.0	39.0

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#40

Dibenzo(a,h)anthracene

Concen: 0.346 ng




RT: 25.718 min Scan# 3683


Delta R.T. 0.003 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14

Tgt Ion:	278	Resp:	11142
Ion Ratio	Lower	Upper	
278	100		
139	33.0	25.4	38.2
279	28.2	22.4	33.6





#41

Benzo(g,h,i)perylene

Concen: 0.354 ng

RT: 26.373 min Scan# 3907

Delta R.T. -0.000 min

Lab File: BN034691.D


Acq: 24 Oct 2024 14:14

Instrument :

BNA_N

ClientSampleId :

ICVBN102424



Tgt Ion:276 Resp: 12570

Ion Ratio Lower Upper

276 100

277 25.5 21.3 31.9



138 40.2 31.9 47.9

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034691.D
Acq On : 24 Oct 2024 14:14
Operator : RC/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN102424

Quant Time: Oct 24 14:43:51 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.532	0.473	11.1	103	0.00
3	n-Nitrosodimethylamine	0.775	0.686	11.5	103	0.00
4 S	2-Fluorophenol	1.189	1.193	-0.3	118	0.00
5 S	Phenol-d6	1.568	1.553	1.0	118	0.00
6	bis(2-Chloroethyl)ether	1.274	1.176	7.7	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
8 S	Nitrobenzene-d5	0.336	0.290	13.7	106	0.00
9	Naphthalene	1.107	0.991	10.5	108	0.00
10	Hexachlorobutadiene	0.166	0.149	10.2	107	0.00
11 SURR	2-Methylnaphthalene-d10	0.550	0.491	10.7	109	0.00
12	2-Methylnaphthalene	0.687	0.612	10.9	109	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	111	0.00
14 S	2,4,6-Tribromophenol	0.124	0.106	14.5	123	0.00
15 S	2-Fluorobiphenyl	1.590	1.353	14.9	109	0.00
16	Acenaphthylene	1.921	1.580	17.8	111	0.00
17	Acenaphthene	1.324	1.110	16.2	110	-0.01
18	Fluorene	1.634	1.406	14.0	113	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
20	4,6-Dinitro-2-methylphenol	0.056	0.045	19.6	122	0.00
21	4-Bromophenyl-phenylether	0.213	0.185	13.1	112	0.00
22	Hexachlorobenzene	0.231	0.207	10.4	113	0.00
23	Atrazine	0.168	0.139	17.3	115	0.00
24	Pentachlorophenol	0.082	0.070	14.6	126	0.00
25	Phenanthrene	1.202	1.073	10.7	116	0.00
26	Anthracene	1.067	0.920	13.8	114	0.00
27 SURR	Fluoranthene-d10	0.901	0.788	12.5	116	0.00
28	Fluoranthene	1.253	1.092	12.8	116	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	117	0.00
30	Pyrene	2.055	1.879	8.6	117	0.00
31 S	Terphenyl-d14	0.817	0.758	7.2	120	0.00
32	Benzo(a)anthracene	1.528	1.334	12.7	113	0.00
33	Chrysene	1.581	1.427	9.7	116	0.00
34	Bis(2-ethylhexyl)phthalate	0.867	0.698	19.5	113	0.00
35 I	Perylene-d12	1.000	1.000	0.0	110	0.00
36	Indeno(1,2,3-cd)pyrene	1.541	1.336	13.3	104	0.00
37	Benzo(b)fluoranthene	1.597	1.449	9.3	108	0.00
38	Benzo(k)fluoranthene	1.576	1.392	11.7	111	0.00
39 C	Benzo(a)pyrene	1.282	1.138	11.2	109	0.00
40	Dibenzo(a,h)anthracene	1.215	1.052	13.4	104	0.00
41	Benzo(g,h,i)perylene	1.340	1.187	11.4	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034691.D
Acq On : 24 Oct 2024 14:14
Operator : RC/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN102424

Quant Time: Oct 24 14:43:51 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	106	0.00
2	1,4-Dioxane	0.400	0.355	11.3	103	0.00
3	n-Nitrosodimethylamine	0.400	0.354	11.5	103	0.00
4 S	2-Fluorophenol	0.400	0.401	-0.3	118	0.00
5 S	Phenol-d6	0.400	0.396	1.0	118	0.00
6	bis(2-Chloroethyl)ether	0.400	0.369	7.8	109	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	107	0.00
8 S	Nitrobenzene-d5	0.400	0.346	13.5	106	0.00
9	Naphthalene	0.400	0.358	10.5	108	0.00
10	Hexachlorobutadiene	0.400	0.360	10.0	107	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.357	10.8	109	0.00
12	2-Methylnaphthalene	0.400	0.356	11.0	109	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	111	0.00
14 S	2,4,6-Tribromophenol	0.400	0.343	14.2	123	0.00
15 S	2-Fluorobiphenyl	0.400	0.341	14.8	109	0.00
16	Acenaphthylene	0.400	0.329	17.8	111	0.00
17	Acenaphthene	0.400	0.335	16.3	110	-0.01
18	Fluorene	0.400	0.344	14.0	113	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	116	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.324	19.0	122	0.00
21	4-Bromophenyl-phenylether	0.400	0.347	13.3	112	0.00
22	Hexachlorobenzene	0.400	0.358	10.5	113	0.00
23	Atrazine	0.400	0.333	16.8	115	0.00
24	Pentachlorophenol	0.400	0.340	15.0	126	0.00
25	Phenanthrene	0.400	0.357	10.8	116	0.00
26	Anthracene	0.400	0.345	13.8	114	0.00
27 SURR	Fluoranthene-d10	0.400	0.350	12.5	116	0.00
28	Fluoranthene	0.400	0.348	13.0	116	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	117	0.00
30	Pyrene	0.400	0.366	8.5	117	0.00
31 S	Terphenyl-d14	0.400	0.371	7.3	120	0.00
32	Benzo(a)anthracene	0.400	0.349	12.8	113	0.00
33	Chrysene	0.400	0.361	9.8	116	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.322	19.5	113	0.00
35 I	Perylene-d12	0.400	0.400	0.0	110	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.347	13.3	104	0.00
37	Benzo(b)fluoranthene	0.400	0.363	9.3	108	0.00
38	Benzo(k)fluoranthene	0.400	0.353	11.8	111	0.00
39 C	Benzo(a)pyrene	0.400	0.355	11.3	109	0.00
40	Dibenzo(a,h)anthracene	0.400	0.346	13.5	104	0.00
41	Benzo(g,h,i)perylene	0.400	0.354	11.5	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG No.: P4440

Instrument ID: BNA_N Calibration Date/Time: 10/24/2024 21:01

Lab File ID: BN034701.D Init. Calib. Date(s): 10/24/2024 10/24/2024

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 09:11 12:48

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.550	0.473		-14.0	20.0
Fluoranthene-d10	0.901	0.768		-14.8	20.0
2-Fluorophenol	1.189	1.172		-1.4	20.0
Phenol-d6	1.568	1.616		3.1	20.0
Nitrobenzene-d5	0.336	0.290		-13.7	20.0
2-Fluorobiphenyl	1.590	1.425		-10.4	20.0
2,4,6-Tribromophenol	0.124	0.096		-22.6	20.0
Terphenyl-d14	0.817	0.690		-15.5	20.0
1,4-Dioxane	0.532	0.489		-8.1	20.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
 Data File : BN034701.D
 Acq On : 24 Oct 2024 21:01
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Oct 25 02:07:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 24 13:17:02 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.712	152	5885	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	17350	0.400	ng	0.00
13) Acenaphthene-d10	14.325	164	8078	0.400	ng	0.00
19) Phenanthrene-d10	17.064	188	14577	0.400	ng	# 0.00
29) Chrysene-d12	21.248	240	8841	0.400	ng	0.00
35) Perylene-d12	23.464	264	8336	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.322	112	6898	0.394	ng	0.00
5) Phenol-d6	6.882	99	9511	0.412	ng	0.00
8) Nitrobenzene-d5	8.845	82	5024	0.345	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	8204	0.344	ng	0.00
14) 2,4,6-Tribromophenol	15.810	330	779	0.312	ng	0.00
15) 2-Fluorobiphenyl	12.957	172	11511	0.359	ng	0.00
27) Fluoranthene-d10	19.092	212	11190	0.341	ng	0.00
31) Terphenyl-d14	19.700	244	6100	0.338	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.285	88	2877	0.367	ng	98
3) n-Nitrosodimethylamine	3.581	42	4327	0.380	ng	# 93
6) bis(2-Chloroethyl)ether	7.142	93	6834	0.365	ng	98
9) Naphthalene	10.532	128	17319	0.361	ng	100
10) Hexachlorobutadiene	10.831	225	2605	0.362	ng	# 100
12) 2-Methylnaphthalene	12.143	142	10362	0.348	ng	98
16) Acenaphthylene	14.047	152	13092	0.337	ng	99
17) Acenaphthene	14.389	154	9042	0.338	ng	98
18) Fluorene	15.373	166	10979	0.333	ng	99
20) 4,6-Dinitro-2-methylph...	15.448	198	621	0.306	ng	92
21) 4-Bromophenyl-phenylether	16.269	248	2820	0.363	ng	# 89
22) Hexachlorobenzene	16.381	284	3143	0.373	ng	98
23) Atrazine	16.542	200	1901	0.311	ng	# 93
24) Pentachlorophenol	16.716	266	1008	0.338	ng	97
25) Phenanthrene	17.101	178	15738	0.359	ng	100
26) Anthracene	17.200	178	13426	0.345	ng	99
28) Fluoranthene	19.124	202	15690	0.343	ng	99
30) Pyrene	19.487	202	15827	0.348	ng	99
32) Benzo(a)anthracene	21.230	228	11854	0.351	ng	99
33) Chrysene	21.283	228	12923	0.370	ng	100
34) Bis(2-ethylhexyl)phtha...	21.194	149	6855	0.358	ng	98
36) Indeno(1,2,3-cd)pyrene	25.692	276	12128	0.378	ng	97
37) Benzo(b)fluoranthene	22.801	252	12003	0.361	ng	98
38) Benzo(k)fluoranthene	22.845	252	11895	0.362	ng	98
39) Benzo(a)pyrene	23.368	252	9650	0.361	ng	97
40) Dibenzo(a,h)anthracene	25.713	278	9542	0.377	ng	97
41) Benzo(g,h,i)perylene	26.365	276	10879	0.390	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034701.D
Acq On : 24 Oct 2024 21:01
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Oct 25 02:07:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.712 min Scan# 669
Delta R.T. 0.000 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:152 Resp: 5885
Ion Ratio Lower Upper
152 100
150 153.2 122.7 184.1
115 63.1 51.0 76.6




#2
1,4-Dioxane
Concen: 0.367 ng
RT: 3.285 min Scan# 56
Delta R.T. 0.000 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01

Tgt Ion: 88 Resp: 2877
Ion Ratio Lower Upper
88 100
43 40.9 32.1 48.1
58 89.8 73.7 110.5



Abundance Scan 97 (3.581 min): BN034686.D\data.ms (-93)



#3

n-Nitrosodimethylamine

Concen: 0.380 ng

RT: 3.581 min Scan# 91

Delta R.T. 0.000 min

Lab File: BN034701.D

Acq: 24 Oct 2024 21:01

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4

Tgt Ion: 42 Resp: 4327


Ion Ratio Lower Upper

42 100


74 91.2 77.2 115.8

44 9.1 13.4 20.2#


Abundance Scan 97 (3.581 min): BN034701.D\data.ms



Abundance




Abundance Scan 97 (3.581 min): BN034701.D\data.ms (-76)



Time-->

Abundance Scan 338 (5.322 min): BN034686.D\data.ms (-33)



#4

2-Fluorophenol

Concen: 0.394 ng

RT: 5.322 min Scan# 338

Delta R.T. 0.000 min

Lab File: BN034701.D

Acq: 24 Oct 2024 21:01

Tgt Ion: 112 Resp: 6898


Ion Ratio Lower Upper

112 100


64 67.5 53.0 79.6

63 35.6 27.6 41.4

Abundance Scan 338 (5.322 min): BN034701.D\data.ms





Abundance



Time-->

Abundance Scan 338 (5.322 min): BN034701.D\data.ms (-31)









#5
Phenol-d6
Concen: 0.412 ng
RT: 6.882 min Scan# 511
Delta R.T. 0.000 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 9511
Ion Ratio Lower Upper
99 100
42 29.2 21.8 32.8
71 34.1 26.6 40.0



#6
bis(2-Chloroethyl)ether
Concen: 0.365 ng
RT: 7.142 min Scan# 590
Delta R.T. 0.000 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01

Tgt Ion: 93 Resp: 6834
Ion Ratio Lower Upper
93 100
63 90.8 71.5 107.3
95 33.1 25.8 38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	17350
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.1	13.7
54	12.1	9.8	14.8
68	7.7	6.4	9.6



#8
Nitrobenzene-d5
Concen: 0.345 ng
RT: 8.845 min Scan# 802
Delta R.T. 0.000 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01

Tgt Ion:	82	Resp:	5024
Ion	Ratio	Lower	Upper
82	100		
128	37.0	29.7	44.5
54	65.8	53.0	79.6









#9
Naphthalene
Concen: 0.361 ng
RT: 10.532 min Scan# 960
Delta R.T. 0.000 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.4	9.0	13.6
127	13.2	10.5	15.7



#10
Hexachlorobutadiene
Concen: 0.362 ng
RT: 10.831 min Scan# 988
Delta R.T. 0.000 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01

Tgt Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.0	51.4	77.0









#11
2-Methylnaphthalene-d10
Concen: 0.344 ng
RT: 12.071 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 8204
Ion Ratio Lower Upper
152 100
151 21.0 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.348 ng
RT: 12.143 min Scan# 1220
Delta R.T. -0.004 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01

Tgt Ion:142 Resp: 10362
Ion Ratio Lower Upper
142 100
141 89.3 70.4 105.6
115 38.7 29.8 44.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.325 min Scan# 1504
Delta R.T. -0.004 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 8078
Ion Ratio Lower Upper
164 100
162 103.8 81.4 122.0
160 54.0 40.6 61.0



#14
2,4,6-Tribromophenol
Concen: 0.312 ng
RT: 15.810 min Scan# 1638
Delta R.T. -0.005 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01

Tgt Ion:330 Resp: 779
Ion Ratio Lower Upper
330 100
332 94.7 76.6 115.0
141 71.4 54.6 81.8









#15
2-Fluorobiphenyl
Concen: 0.359 ng
RT: 12.957 min Scan# 11
Delta R.T. 0.007 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	172	Resp:	11511
Ion Ratio	Lower	Upper	
172	100		
171	34.0	28.3	42.5
170	23.3	19.8	29.6



#16
Acenaphthylene
Concen: 0.337 ng
RT: 14.047 min Scan# 1477
Delta R.T. 0.007 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01

Tgt Ion:	152	Resp:	13092
Ion Ratio	Lower	Upper	
152	100		
151	19.1	15.6	23.4
153	12.8	10.6	15.8









#17
Acenaphthene
Concen: 0.338 ng
RT: 14.389 min Scan# 1509
Delta R.T. -0.004 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 9042
Ion Ratio Lower Upper
154 100
153 111.5 88.2 132.2
152 61.5 46.9 70.3



#18
Fluorene
Concen: 0.333 ng
RT: 15.373 min Scan# 1601
Delta R.T. -0.004 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01

Tgt Ion:166 Resp: 10979
Ion Ratio Lower Upper
166 100
165 97.2 78.8 118.2
167 13.2 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.064 min Scan# 11
Delta R.T. -0.005 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 14577
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 15.2 8.9 13.3#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.306 ng
RT: 15.448 min Scan# 1608
Delta R.T. -0.004 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01

Tgt Ion:198 Resp: 621
Ion Ratio Lower Upper
198 100
51 124.4 91.8 137.8
105 59.8 44.3 66.5









#21
4-Bromophenyl-phenylether
Concen: 0.363 ng
RT: 16.269 min Scan# 1675
Delta R.T. -0.005 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 2820
Ion Ratio Lower Upper
248 100
250 96.0 80.9 121.3
141 67.5 40.5 60.7#



#22
Hexachlorobenzene
Concen: 0.373 ng
RT: 16.381 min Scan# 1684
Delta R.T. 0.007 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01

Tgt Ion:284 Resp: 3143
Ion Ratio Lower Upper
284 100
142 56.7 43.6 65.4
249 34.2 26.6 40.0





#23

Atrazine

Concen: 0.311 ng

RT: 16.542 min Scan# 1

Delta R.T. -0.005 min

Lab File: BN034701.D

Acq: 24 Oct 2024 21:01

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:200 Resp: 1901

Ion Ratio Lower Upper

200 100

173 32.8 19.7 29.5#

215 46.2 35.7 53.5



#24

Pentachlorophenol

Concen: 0.338 ng

RT: 16.716 min Scan# 1711

Delta R.T. -0.005 min

Lab File: BN034701.D

Acq: 24 Oct 2024 21:01




Tgt Ion:266 Resp: 1008


Ion Ratio Lower Upper

266 100

264 63.5 49.4 74.2

268 65.7 50.5 75.7









#25
Phenanthrene
Concen: 0.359 ng
RT: 17.101 min Scan# 1742
Delta R.T. -0.005 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 15738
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.2
179 15.2 12.2 18.4



#26
Anthracene
Concen: 0.345 ng
RT: 17.200 min Scan# 1750
Delta R.T. 0.007 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01

Tgt Ion:178 Resp: 13426
Ion Ratio Lower Upper
178 100
176 18.3 14.9 22.3
179 15.0 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.341 ng
RT: 19.092 min Scan# 1982
Delta R.T. -0.002 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
212	100		
106	22.7	17.8	26.8
104	13.4	10.2	15.4



#28
Fluoranthene
Concen: 0.343 ng
RT: 19.124 min Scan# 1987
Delta R.T. -0.002 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01

Tgt Ion	Ratio	Lower	Upper
202	100		
101	16.8	12.8	19.2
203	17.0	13.5	20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.248 min Scan# 2419
Delta R.T. -0.002 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
240	100		
120	20.3	14.2	21.2
236	28.6	23.1	34.7



#30
Pyrene
Concen: 0.348 ng
RT: 19.487 min Scan# 2065
Delta R.T. 0.002 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01

Tgt Ion	Ratio	Lower	Upper
202	100		
200	21.0	16.6	25.0
203	17.6	14.2	21.4









#31
Terphenyl-d14
Concen: 0.338 ng
RT: 19.700 min Scan# 2113
Delta R.T. -0.002 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
244	100		
212	11.5	8.6	13.0
122	30.3	22.3	33.5



#32
Benzo(a)anthracene
Concen: 0.351 ng
RT: 21.230 min Scan# 2417
Delta R.T. -0.002 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.9	22.0	33.0
229	19.9	15.6	23.4





#33

Chrysene

Concen: 0.370 ng

RT: 21.283 min Scan# 2425

Delta R.T. -0.002 min

Lab File: BN034701.D

Acq: 24 Oct 2024 21:01

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:228 Resp: 12923

Ion Ratio Lower Upper

228 100

226 29.9 24.1 36.1

229 19.9 15.8 23.8



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.358 ng

RT: 21.194 min Scan# 2413

Delta R.T. -0.002 min

Lab File: BN034701.D

Acq: 24 Oct 2024 21:01




Tgt Ion:149 Resp: 6855


Ion Ratio Lower Upper

149 100

167 23.7 19.7 29.5

279 1.9 1.8 2.6









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.464 min Scan# 2913
Delta R.T. -0.002 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:264 Resp: 8336
Ion Ratio Lower Upper
264 100
260 27.6 21.9 32.9
265 77.6 60.6 91.0



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.378 ng
RT: 25.692 min Scan# 3671
Delta R.T. -0.002 min
Lab File: BN034701.D
Acq: 24 Oct 2024 21:01

Tgt Ion:276 Resp: 12128
Ion Ratio Lower Upper
276 100
138 46.6 35.1 52.7
277 24.4 19.7 29.5





#37

Benzo(b)fluoranthene

Concen: 0.361 ng

RT: 22.801 min Scan# 2697

Delta R.T. -0.002 min

Lab File: BN034701.D

Acq: 24 Oct 2024 21:01

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:252 Resp: 12003

Ion Ratio Lower Upper

252 100

253 25.1 20.0 30.0

125 27.7 20.9 31.3



#38

Benzo(k)fluoranthene

Concen: 0.362 ng

RT: 22.845 min Scan# 2697

Delta R.T. -0.005 min

Lab File: BN034701.D

Acq: 24 Oct 2024 21:01




Tgt Ion:252 Resp: 11895


Ion Ratio Lower Upper

252 100

253 25.8 20.6 31.0

125 29.4 21.6 32.4





#39

Benzo(a)pyrene

Concen: 0.361 ng

RT: 23.368 min Scan# 21

Delta R.T. 0.001 min

Lab File: BN034701.D

Acq: 24 Oct 2024 21:01

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:252 Resp: 9650

Ion Ratio Lower Upper

252 100

253 28.0 22.2 33.4

125 35.0 26.0 39.0



#40

Dibenzo(a,h)anthracene

Concen: 0.377 ng

RT: 25.713 min Scan# 3678

Delta R.T. -0.002 min

Lab File: BN034701.D

Acq: 24 Oct 2024 21:01




Tgt Ion:278 Resp: 9542


Ion Ratio Lower Upper

278 100

139 34.2 25.4 38.2

279 27.4 22.4 33.6





#41

Benzo(g,h,i)perylene

Concen: 0.390 ng

RT: 26.365 min Scan# 3901

Delta R.T. -0.008 min

Lab File: BN034701.D


Acq: 24 Oct 2024 21:01

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:276 Resp: 10879

Ion Ratio Lower Upper

276 100

277 25.9 21.3 31.9

138 42.0 31.9 47.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034701.D
Acq On : 24 Oct 2024 21:01
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 25 02:07:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.532	0.489	8.1	94	0.00
3	n-Nitrosodimethylamine	0.775	0.735	5.2	97	0.00
4 S	2-Fluorophenol	1.189	1.172	1.4	102	0.00
5 S	Phenol-d6	1.568	1.616	-3.1	109	0.00
6	bis(2-Chloroethyl)ether	1.274	1.161	8.9	95	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
8 S	Nitrobenzene-d5	0.336	0.290	13.7	92	0.00
9	Naphthalene	1.107	0.998	9.8	95	0.00
10	Hexachlorobutadiene	0.166	0.150	9.6	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.550	0.473	14.0	92	0.00
12	2-Methylnaphthalene	0.687	0.597	13.1	93	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
14 S	2,4,6-Tribromophenol	0.124	0.096	22.6	87	0.00
15 S	2-Fluorobiphenyl	1.590	1.425	10.4	90	0.00
16	Acenaphthylene	1.921	1.621	15.6	89	0.00
17	Acenaphthene	1.324	1.119	15.5	87	0.00
18	Fluorene	1.634	1.359	16.8	85	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	80	0.00
20	4,6-Dinitro-2-methylphenol	0.056	0.043	23.2	80	0.00
21	4-Bromophenyl-phenylether	0.213	0.193	9.4	81	0.00
22	Hexachlorobenzene	0.231	0.216	6.5	82	0.00
23	Atrazine	0.168	0.130	22.6	75	0.00
24	Pentachlorophenol	0.082	0.069	15.9	87	0.00
25	Phenanthrene	1.202	1.080	10.1	81	0.00
26	Anthracene	1.067	0.921	13.7	79	0.00
27 SURR	Fluoranthene-d10	0.901	0.768	14.8	79	0.00
28	Fluoranthene	1.253	1.076	14.1	79	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
30	Pyrene	2.055	1.790	12.9	79	0.00
31 S	Terphenyl-d14	0.817	0.690	15.5	78	0.00
32	Benzo(a)anthracene	1.528	1.341	12.2	81	0.00
33	Chrysene	1.581	1.462	7.5	84	0.00
34	Bis(2-ethylhexyl)phthalate	0.867	0.775	10.6	89	0.00
35 I	Perylene-d12	1.000	1.000	0.0	87	0.00
36	Indeno(1,2,3-cd)pyrene	1.541	1.455	5.6	89	0.00
37	Benzo(b)fluoranthene	1.597	1.440	9.8	84	0.00
38	Benzo(k)fluoranthene	1.576	1.427	9.5	89	0.00
39 C	Benzo(a)pyrene	1.282	1.158	9.7	87	0.00
40	Dibenzo(a,h)anthracene	1.215	1.145	5.8	89	0.00
41	Benzo(g,h,i)perylene	1.340	1.305	2.6	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034701.D
Acq On : 24 Oct 2024 21:01
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 25 02:07:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	93	0.00
2	1,4-Dioxane	0.400	0.367	8.3	94	0.00
3	n-Nitrosodimethylamine	0.400	0.380	5.0	97	0.00
4 S	2-Fluorophenol	0.400	0.394	1.5	102	0.00
5 S	Phenol-d6	0.400	0.412	-3.0	109	0.00
6	bis(2-Chloroethyl)ether	0.400	0.365	8.8	95	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	94	0.00
8 S	Nitrobenzene-d5	0.400	0.345	13.8	92	0.00
9	Naphthalene	0.400	0.361	9.8	95	0.00
10	Hexachlorobutadiene	0.400	0.362	9.5	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.344	14.0	92	0.00
12	2-Methylnaphthalene	0.400	0.348	13.0	93	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	87	0.00
14 S	2,4,6-Tribromophenol	0.400	0.312	22.0	87	0.00
15 S	2-Fluorobiphenyl	0.400	0.359	10.3	90	0.00
16	Acenaphthylene	0.400	0.337	15.8	89	0.00
17	Acenaphthene	0.400	0.338	15.5	87	0.00
18	Fluorene	0.400	0.333	16.8	85	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	80	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.306	23.5	80	0.00
21	4-Bromophenyl-phenylether	0.400	0.363	9.3	81	0.00
22	Hexachlorobenzene	0.400	0.373	6.8	82	0.00
23	Atrazine	0.400	0.311	22.3	75	0.00
24	Pentachlorophenol	0.400	0.338	15.5	87	0.00
25	Phenanthrene	0.400	0.359	10.3	81	0.00
26	Anthracene	0.400	0.345	13.8	79	0.00
27 SURR	Fluoranthene-d10	0.400	0.341	14.8	79	0.00
28	Fluoranthene	0.400	0.343	14.2	79	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	83	0.00
30	Pyrene	0.400	0.348	13.0	79	0.00
31 S	Terphenyl-d14	0.400	0.338	15.5	78	0.00
32	Benzo(a)anthracene	0.400	0.351	12.3	81	0.00
33	Chrysene	0.400	0.370	7.5	84	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.358	10.5	89	0.00
35 I	Perylene-d12	0.400	0.400	0.0	87	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.378	5.5	89	0.00
37	Benzo(b)fluoranthene	0.400	0.361	9.8	84	0.00
38	Benzo(k)fluoranthene	0.400	0.362	9.5	89	0.00
39 C	Benzo(a)pyrene	0.400	0.361	9.8	87	0.00
40	Dibenzo(a,h)anthracene	0.400	0.377	5.8	89	0.00
41	Benzo(g,h,i)perylene	0.400	0.390	2.5	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG No.: P4440

Instrument ID: BNA_N Calibration Date/Time: 10/25/2024 07:51

Lab File ID: BN034719.D Init. Calib. Date(s): 10/24/2024 10/24/2024

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 09:11 12:48

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.550	0.486		-11.6	50.0
Fluoranthene-d10	0.901	0.942		4.6	50.0
2-Fluorophenol	1.189	1.294		8.8	50.0
Phenol-d6	1.568	1.736		10.7	50.0
Nitrobenzene-d5	0.336	0.300		-10.7	50.0
2-Fluorobiphenyl	1.590	1.367		-14.0	50.0
2,4,6-Tribromophenol	0.124	0.113		-8.9	50.0
Terphenyl-d14	0.817	0.696		-14.8	50.0
1,4-Dioxane	0.532	0.475		-10.7	50.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
 Data File : BN034719.D
 Acq On : 25 Oct 2024 07:51
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Oct 25 08:18:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 24 13:17:02 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.712	152	6027	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	17907	0.400	ng	0.00
13) Acenaphthene-d10	14.318	164	8880	0.400	ng	-0.01
19) Phenanthrene-d10	17.057	188	17802	0.400	ng	#-0.01
29) Chrysene-d12	21.241	240	14770	0.400	ng	# 0.00
35) Perylene-d12	23.461	264	15941	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.322	112	7797	0.435	ng	0.00
5) Phenol-d6	6.882	99	10465	0.443	ng	0.00
8) Nitrobenzene-d5	8.845	82	5374	0.358	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	8705	0.354	ng	0.00
14) 2,4,6-Tribromophenol	15.815	330	1006	0.366	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	12140	0.344	ng	0.00
27) Fluoranthene-d10	19.094	212	16769	0.418	ng	0.00
31) Terphenyl-d14	19.698	244	10278	0.341	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.285	88	2863	0.357	ng	97
3) n-Nitrosodimethylamine	3.588	42	4215	0.361	ng	96
6) bis(2-Chloroethyl)ether	7.141	93	7199	0.375	ng	99
9) Naphthalene	10.532	128	18267	0.369	ng	100
10) Hexachlorobutadiene	10.821	225	2610	0.351	ng	# 98
12) 2-Methylnaphthalene	12.143	142	10986	0.357	ng	97
16) Acenaphthylene	14.040	152	14421	0.338	ng	100
17) Acenaphthene	14.382	154	10024	0.341	ng	98
18) Fluorene	15.377	166	12279	0.339	ng	100
20) 4,6-Dinitro-2-methylph...	15.441	198	746	0.301	ng	# 54
21) 4-Bromophenyl-phenylether	16.262	248	3313	0.349	ng	# 63
22) Hexachlorobenzene	16.374	284	3421	0.332	ng	95
23) Atrazine	16.535	200	2697	0.362	ng	# 90
24) Pentachlorophenol	16.709	266	1325	0.364	ng	98
25) Phenanthrene	17.106	178	18834	0.352	ng	99
26) Anthracene	17.193	178	17465	0.368	ng	99
28) Fluoranthene	19.122	202	23020	0.413	ng	99
30) Pyrene	19.484	202	24493	0.323	ng	99
32) Benzo(a)anthracene	21.232	228	22023	0.390	ng	99
33) Chrysene	21.277	228	20725	0.355	ng	98
34) Bis(2-ethylhexyl)phtha...	21.187	149	18018	0.563	ng	98
36) Indeno(1,2,3-cd)pyrene	25.688	276	23125	0.377	ng	95
37) Benzo(b)fluoranthene	22.797	252	21412	0.336	ng	# 34
38) Benzo(k)fluoranthene	22.841	252	20953	0.334	ng	# 32
39) Benzo(a)pyrene	23.364	252	19025	0.372	ng	# 20
40) Dibenzo(a,h)anthracene	25.709	278	17877	0.369	ng	# 58
41) Benzo(g,h,i)perylene	26.364	276	19691	0.369	ng	# 83


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034719.D
Acq On : 25 Oct 2024 07:51
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Oct 25 08:18:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.712 min Scan# 60
Delta R.T. 0.000 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:152 Resp: 6027
Ion Ratio Lower Upper
152 100
150 154.5 122.7 184.1
115 65.7 51.0 76.6




#2
1,4-Dioxane
Concen: 0.357 ng
RT: 3.285 min Scan# 56
Delta R.T. -0.000 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Tgt Ion: 88 Resp: 2863
Ion Ratio Lower Upper
88 100
43 45.3 32.1 48.1
58 92.2 73.7 110.5



Abundance Scan 97 (3.581 min): BN034686.D\data.ms (-93)



#3

n-Nitrosodimethylamine

Concen: 0.361 ng

RT: 3.588 min Scan# 91

Delta R.T. 0.007 min

Lab File: BN034719.D

Acq: 25 Oct 2024 07:51

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion: 42 Resp: 4215


Ion Ratio Lower Upper

42 100


74 100.4 77.2 115.8

44 14.9 13.4 20.2


Abundance Scan 98 (3.588 min): BN034719.D\data.ms




Abundance



Abundance Scan 98 (3.588 min): BN034719.D\data.ms (-76)



Abundance Scan 338 (5.322 min): BN034686.D\data.ms (-33)



#4

2-Fluorophenol

Concen: 0.435 ng

RT: 5.322 min Scan# 338

Delta R.T. -0.000 min

Lab File: BN034719.D

Acq: 25 Oct 2024 07:51

Tgt Ion: 112 Resp: 7797


Ion Ratio Lower Upper

112 100


64 68.9 53.0 79.6

63 36.4 27.6 41.4


Abundance Scan 338 (5.322 min): BN034719.D\data.ms




Abundance



Abundance Scan 338 (5.322 min): BN034719.D\data.ms (-31)









#5
Phenol-d6
Concen: 0.443 ng
RT: 6.882 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 10465
Ion Ratio Lower Upper
99 100
42 26.7 21.8 32.8
71 33.4 26.6 40.0



#6
bis(2-Chloroethyl)ether
Concen: 0.375 ng
RT: 7.141 min Scan# 590
Delta R.T. -0.000 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Tgt Ion: 93 Resp: 7199
Ion Ratio Lower Upper
93 100
63 88.1 71.5 107.3
95 32.2 25.8 38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	17907
Ion	Ratio	Lower	Upper
136	100		
137	11.6	9.1	13.7
54	12.3	9.8	14.8
68	7.8	6.4	9.6



#8
Nitrobenzene-d5
Concen: 0.358 ng
RT: 8.845 min Scan# 802
Delta R.T. -0.000 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Tgt Ion:	82	Resp:	5374
Ion	Ratio	Lower	Upper
82	100		
128	39.0	29.7	44.5
54	66.9	53.0	79.6









#9
Naphthalene
Concen: 0.369 ng
RT: 10.532 min Scan# 90
Delta R.T. -0.000 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	128	Resp:	18267
Ion	Ratio	Lower	Upper
128	100		
129	11.5	9.0	13.6
127	13.2	10.5	15.7



#10
Hexachlorobutadiene
Concen: 0.351 ng
RT: 10.821 min Scan# 987
Delta R.T. -0.011 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Tgt Ion:	225	Resp:	2610
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.0	51.4	77.0









#11
2-Methylnaphthalene-d10
Concen: 0.354 ng
RT: 12.071 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 8705
Ion Ratio Lower Upper
152 100
151 20.9 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.357 ng
RT: 12.143 min Scan# 1220
Delta R.T. -0.004 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Tgt Ion:142 Resp: 10986
Ion Ratio Lower Upper
142 100
141 90.6 70.4 105.6
115 40.0 29.8 44.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.318 min Scan# 1503
Delta R.T. -0.011 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 8880
Ion Ratio Lower Upper
164 100
162 107.0 81.4 122.0
160 53.0 40.6 61.0



#14
2,4,6-Tribromophenol
Concen: 0.366 ng
RT: 15.815 min Scan# 1639
Delta R.T. -0.000 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Tgt Ion:330 Resp: 1006
Ion Ratio Lower Upper
330 100
332 97.7 76.6 115.0
141 68.0 54.6 81.8









#15
2-Fluorobiphenyl
Concen: 0.344 ng
RT: 12.950 min Scan# 1375
Delta R.T. -0.000 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	172	Resp:	12140
Ion Ratio	Lower	Upper	
172	100		
171	35.1	28.3	42.5
170	24.4	19.8	29.6



#16
Acenaphthylene
Concen: 0.338 ng
RT: 14.040 min Scan# 1477
Delta R.T. -0.000 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Tgt Ion:	152	Resp:	14421
Ion Ratio	Lower	Upper	
152	100		
151	19.5	15.6	23.4
153	13.2	10.6	15.8









#17
Acenaphthene
Concen: 0.341 ng
RT: 14.382 min Scan# 1509
Delta R.T. -0.011 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 10024
Ion Ratio Lower Upper
154 100
153 111.6 88.2 132.2
152 56.1 46.9 70.3



#18
Fluorene
Concen: 0.339 ng
RT: 15.377 min Scan# 1602
Delta R.T. -0.000 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Tgt Ion:166 Resp: 12279
Ion Ratio Lower Upper
166 100
165 98.8 78.8 118.2
167 13.6 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.057 min Scan# 11
Delta R.T. -0.012 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 17802
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 22.1 8.9 13.3#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.301 ng
RT: 15.441 min Scan# 1608
Delta R.T. -0.011 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Tgt Ion:198 Resp: 746
Ion Ratio Lower Upper
198 100
51 165.1 91.8 137.8#
105 89.0 44.3 66.5#









#21
4-Bromophenyl-phenylether
Concen: 0.349 ng
RT: 16.262 min Scan# 1675
Delta R.T. -0.012 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 3313
Ion Ratio Lower Upper
248 100
250 92.6 80.9 121.3
141 115.7 40.5 60.7#



#22
Hexachlorobenzene
Concen: 0.332 ng
RT: 16.374 min Scan# 1684
Delta R.T. -0.000 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Tgt Ion:284 Resp: 3421
Ion Ratio Lower Upper
284 100
142 58.6 43.6 65.4
249 35.2 26.6 40.0









#23
Atrazine
Concen: 0.362 ng
RT: 16.535 min Scan# 11
Delta R.T. -0.012 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:200 Resp: 2697
Ion Ratio Lower Upper
200 100
173 37.0 19.7 29.5#
215 43.4 35.7 53.5



#24
Pentachlorophenol
Concen: 0.364 ng
RT: 16.709 min Scan# 1711
Delta R.T. -0.012 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Tgt Ion:266 Resp: 1325
Ion Ratio Lower Upper
266 100
264 63.3 49.4 74.2
268 65.1 50.5 75.7









#25
Phenanthrene
Concen: 0.352 ng
RT: 17.106 min Scan# 1743
Delta R.T. -0.000 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 18834
Ion Ratio Lower Upper
178 100
176 19.9 15.4 23.2
179 15.6 12.2 18.4



#26
Anthracene
Concen: 0.368 ng
RT: 17.193 min Scan# 1750
Delta R.T. -0.000 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Tgt Ion:178 Resp: 17465
Ion Ratio Lower Upper
178 100
176 18.8 14.9 22.3
179 15.5 12.2 18.2





#27

Fluoranthene-d10

Concen: 0.418 ng

RT: 19.094 min Scan# 1982

Delta R.T. -0.000 min

Lab File: BN034719.D

Acq: 25 Oct 2024 07:51

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:212 Resp: 16769

Ion Ratio Lower Upper

212 100

106 23.4 17.8 26.8

104 13.5 10.2 15.4



#28

Fluoranthene

Concen: 0.413 ng

RT: 19.122 min Scan# 1988

Delta R.T. -0.005 min

Lab File: BN034719.D

Acq: 25 Oct 2024 07:51




Tgt Ion:202 Resp: 23020


Ion Ratio Lower Upper

202 100

101 16.8 12.8 19.2

203 17.3 13.5 20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.241 min Scan# 2421
Delta R.T. -0.009 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 14770
Ion Ratio Lower Upper
240 100
120 34.7 14.2 21.2#
236 30.4 23.1 34.7



#30
Pyrene
Concen: 0.323 ng
RT: 19.484 min Scan# 2066
Delta R.T. -0.000 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Tgt Ion:202 Resp: 24493
Ion Ratio Lower Upper
202 100
200 21.3 16.6 25.0
203 18.0 14.2 21.4









#31
Terphenyl-d14
Concen: 0.341 ng
RT: 19.698 min Scan# 2112
Delta R.T. -0.005 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:244 Resp: 10278
Ion Ratio Lower Upper
244 100
212 10.9 8.6 13.0
122 34.1 22.3 33.5#



#32
Benzo(a)anthracene
Concen: 0.390 ng
RT: 21.232 min Scan# 2419
Delta R.T. -0.000 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Tgt Ion:228 Resp: 22023
Ion Ratio Lower Upper
228 100
226 27.2 22.0 33.0
229 20.4 15.6 23.4









#33
Chrysene
Concen: 0.355 ng
RT: 21.277 min Scan# 2424
Delta R.T. -0.009 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.9	24.1	36.1
229	20.1	15.8	23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.563 ng
RT: 21.187 min Scan# 2414
Delta R.T. -0.009 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Tgt Ion	Ratio	Lower	Upper
149	100		
167	23.5	19.7	29.5
279	2.0	1.8	2.6









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.461 min Scan# 2911
Delta R.T. -0.006 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:264 Resp: 15941
Ion Ratio Lower Upper
264 100
260 27.7 21.9 32.9
265 57.4 60.6 91.0#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.377 ng
RT: 25.688 min Scan# 3673
Delta R.T. -0.006 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Tgt Ion:276 Resp: 23125
Ion Ratio Lower Upper
276 100
138 48.9 35.1 52.7
277 24.2 19.7 29.5









#37
Benzo(b)fluoranthene
Concen: 0.336 ng
RT: 22.797 min Scan# 2684
Delta R.T. -0.006 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 21412
Ion Ratio Lower Upper
252 100
253 25.3 20.0 30.0
125 92.0 20.9 31.3#



#38
Benzo(k)fluoranthene
Concen: 0.334 ng
RT: 22.841 min Scan# 2699
Delta R.T. -0.009 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Tgt Ion:252 Resp: 20953
Ion Ratio Lower Upper
252 100
253 25.5 20.6 31.0
125 95.3 21.6 32.4#





#39

Benzo(a)pyrene

Concen: 0.372 ng

RT: 23.364 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN034719.D

Acq: 25 Oct 2024 07:51

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:252 Resp: 19025

Ion Ratio Lower Upper

252 100

253 26.4 22.2 33.4

125 114.5 26.0 39.0#



#40

Dibenzo(a,h)anthracene

Concen: 0.369 ng

RT: 25.709 min Scan# 3680

Delta R.T. -0.006 min

Lab File: BN034719.D

Acq: 25 Oct 2024 07:51




Tgt Ion:278 Resp: 17877


Ion Ratio Lower Upper

278 100

139 74.3 25.4 38.2#


279 26.4 22.4 33.6







#41
Benzo(g,h,i)perylene
Concen: 0.369 ng
RT: 26.364 min Scan# 3904
Delta R.T. -0.009 min
Lab File: BN034719.D
Acq: 25 Oct 2024 07:51

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:276 Resp: 19691
Ion Ratio Lower Upper
276 100
277 25.3 21.3 31.9
138 56.2 31.9 47.9#



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034719.D
Acq On : 25 Oct 2024 07:51
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 25 08:18:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.532	0.475	10.7	93	0.00
3	n-Nitrosodimethylamine	0.775	0.699	9.8	95	0.00
4 S	2-Fluorophenol	1.189	1.294	-8.8	115	0.00
5 S	Phenol-d6	1.568	1.736	-10.7	119	0.00
6	bis(2-Chloroethyl)ether	1.274	1.194	6.3	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
8 S	Nitrobenzene-d5	0.336	0.300	10.7	99	0.00
9	Naphthalene	1.107	1.020	7.9	100	0.00
10	Hexachlorobutadiene	0.166	0.146	12.0	95	-0.01
11 SURR	2-Methylnaphthalene-d10	0.550	0.486	11.6	98	0.00
12	2-Methylnaphthalene	0.687	0.614	10.6	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	95	-0.01
14 S	2,4,6-Tribromophenol	0.124	0.113	8.9	112	0.00
15 S	2-Fluorobiphenyl	1.590	1.367	14.0	95	0.00
16	Acenaphthylene	1.921	1.624	15.5	98	0.00
17	Acenaphthene	1.324	1.129	14.7	96	-0.01
18	Fluorene	1.634	1.383	15.4	95	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	98	-0.01
20	4,6-Dinitro-2-methylphenol	0.056	0.042	25.0	96	-0.01
21	4-Bromophenyl-phenylether	0.213	0.186	12.7	96	-0.01
22	Hexachlorobenzene	0.231	0.192	16.9	89	0.00
23	Atrazine	0.168	0.151	10.1	106	-0.01
24	Pentachlorophenol	0.082	0.074	9.8	114	-0.01
25	Phenanthrene	1.202	1.058	12.0	97	0.00
26	Anthracene	1.067	0.981	8.1	103	0.00
27 SURR	Fluoranthene-d10	0.901	0.942	-4.6	118	0.00
28	Fluoranthene	1.253	1.293	-3.2	116	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	139	0.00
30	Pyrene	2.055	1.658	19.3	123	0.00
31 S	Terphenyl-d14	0.817	0.696	14.8	131	0.00
32	Benzo(a)anthracene	1.528	1.491	2.4	151#	0.00
33	Chrysene	1.581	1.403	11.3	135	0.00
34	Bis(2-ethylhexyl)phthalate	0.867	1.220	-40.7#	235#	0.00
35 I	Perylene-d12	1.000	1.000	0.0	166#	0.00
36	Indeno(1,2,3-cd)pyrene	1.541	1.451	5.8	170#	0.00
37	Benzo(b)fluoranthene	1.597	1.343	15.9	151#	0.00
38	Benzo(k)fluoranthene	1.576	1.314	16.6	158#	0.00
39 C	Benzo(a)pyrene	1.282	1.193	6.9	171#	0.00
40	Dibenzo(a,h)anthracene	1.215	1.121	7.7	167#	0.00
41	Benzo(g,h,i)perylene	1.340	1.235	7.8	163#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034719.D
Acq On : 25 Oct 2024 07:51
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 25 08:18:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	95	0.00
2	1,4-Dioxane	0.400	0.357	10.8	93	0.00
3	n-Nitrosodimethylamine	0.400	0.361	9.8	95	0.00
4 S	2-Fluorophenol	0.400	0.435	-8.7	115	0.00
5 S	Phenol-d6	0.400	0.443	-10.7	119	0.00
6	bis(2-Chloroethyl)ether	0.400	0.375	6.3	100	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	97	0.00
8 S	Nitrobenzene-d5	0.400	0.358	10.5	99	0.00
9	Naphthalene	0.400	0.369	7.8	100	0.00
10	Hexachlorobutadiene	0.400	0.351	12.3	95	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.354	11.5	98	0.00
12	2-Methylnaphthalene	0.400	0.357	10.8	99	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	95	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.366	8.5	112	0.00
15 S	2-Fluorobiphenyl	0.400	0.344	14.0	95	0.00
16	Acenaphthylene	0.400	0.338	15.5	98	0.00
17	Acenaphthene	0.400	0.341	14.8	96	-0.01
18	Fluorene	0.400	0.339	15.3	95	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	98	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.301	24.8	96	-0.01
21	4-Bromophenyl-phenylether	0.400	0.349	12.8	96	-0.01
22	Hexachlorobenzene	0.400	0.332	17.0	89	0.00
23	Atrazine	0.400	0.362	9.5	106	-0.01
24	Pentachlorophenol	0.400	0.364	9.0	114	-0.01
25	Phenanthrene	0.400	0.352	12.0	97	0.00
26	Anthracene	0.400	0.368	8.0	103	0.00
27 SURR	Fluoranthene-d10	0.400	0.418	-4.5	118	0.00
28	Fluoranthene	0.400	0.413	-3.2	116	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	139	0.00
30	Pyrene	0.400	0.323	19.3	123	0.00
31 S	Terphenyl-d14	0.400	0.341	14.8	131	0.00
32	Benzo(a)anthracene	0.400	0.390	2.5	151	0.00
33	Chrysene	0.400	0.355	11.3	135	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.563	-40.7#	235	0.00
35 I	Perylene-d12	0.400	0.400	0.0	166	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.377	5.8	170	0.00
37	Benzo(b)fluoranthene	0.400	0.336	16.0	151	0.00
38	Benzo(k)fluoranthene	0.400	0.334	16.5	158	0.00
39 C	Benzo(a)pyrene	0.400	0.372	7.0	171	0.00
40	Dibenzo(a,h)anthracene	0.400	0.369	7.8	167	0.00
41	Benzo(g,h,i)perylene	0.400	0.369	7.8	163	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG No.: P4440

Instrument ID: BNA_N Calibration Date/Time: 10/25/2024 09:46

Lab File ID: BN034721.D Init. Calib. Date(s): 10/24/2024 10/24/2024

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 09:11 12:48

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.550	0.491		-10.7	20.0
Fluoranthene-d10	0.901	0.834		-7.4	20.0
2-Fluorophenol	1.189	1.280		7.7	20.0
Phenol-d6	1.568	1.733		10.5	20.0
Nitrobenzene-d5	0.336	0.300		-10.7	20.0
2-Fluorobiphenyl	1.590	1.396		-12.2	20.0
2,4,6-Tribromophenol	0.124	0.119		-4.0	20.0
Terphenyl-d14	0.817	0.690		-15.5	20.0
1,4-Dioxane	0.532	0.460		-13.5	20.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102524\
Data File : BN034721.D
Acq On : 25 Oct 2024 09:46
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Oct 25 10:44:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.705	152	6565	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	19566	0.400	ng	0.00
13) Acenaphthene-d10	14.325	164	9604	0.400	ng	0.00
19) Phenanthrene-d10	17.064	188	17577	0.400	ng	0.00
29) Chrysene-d12	21.248	240	13068	0.400	ng	0.00
35) Perylene-d12	23.459	264	14018	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.314	112	8404	0.431	ng	0.00
5) Phenol-d6	6.882	99	11379	0.442	ng	0.00
8) Nitrobenzene-d5	8.845	82	5869	0.358	ng	0.00
11) 2-Methylnaphthalene-d10	12.067	152	9607	0.357	ng	0.00
14) 2,4,6-Tribromophenol	15.810	330	1142	0.384	ng	0.00
15) 2-Fluorobiphenyl	12.946	172	13407	0.351	ng	0.00
27) Fluoranthene-d10	19.092	212	14664	0.370	ng	0.00
31) Terphenyl-d14	19.700	244	9017	0.338	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.285	88	3018	0.345	ng	97
3) n-Nitrosodimethylamine	3.581	42	4242	0.334	ng	# 91
6) bis(2-Chloroethyl)ether	7.142	93	7827	0.374	ng	99
9) Naphthalene	10.522	128	19690	0.364	ng	99
10) Hexachlorobutadiene	10.821	225	2852	0.351	ng	# 99
12) 2-Methylnaphthalene	12.143	142	12083	0.360	ng	99
16) Acenaphthylene	14.037	152	15625	0.339	ng	99
17) Acenaphthene	14.389	154	10877	0.342	ng	99
18) Fluorene	15.373	166	12726	0.324	ng	99
20) 4,6-Dinitro-2-methylph...	15.448	198	824	0.337	ng	98
21) 4-Bromophenyl-phenylether	16.269	248	3420	0.365	ng	96
22) Hexachlorobenzene	16.368	284	3624	0.357	ng	98
23) Atrazine	16.542	200	2675	0.363	ng	98
24) Pentachlorophenol	16.716	266	1357	0.378	ng	98
25) Phenanthrene	17.101	178	19007	0.360	ng	100
26) Anthracene	17.188	178	16769	0.358	ng	100
28) Fluoranthene	19.120	202	20290	0.368	ng	99
30) Pyrene	19.482	202	21115	0.315	ng	99
32) Benzo(a)anthracene	21.230	228	18695	0.374	ng	99
33) Chrysene	21.283	228	18028	0.349	ng	99
34) Bis(2-ethylhexyl)phtha...	21.194	149	15075	0.532	ng	100
36) Indeno(1,2,3-cd)pyrene	25.683	276	19938	0.369	ng	99
37) Benzo(b)fluoranthene	22.798	252	18466	0.330	ng	# 86
38) Benzo(k)fluoranthene	22.842	252	17979	0.326	ng	# 84
39) Benzo(a)pyrene	23.362	252	16295	0.363	ng	# 83
40) Dibenzo(a,h)anthracene	25.707	278	15708	0.369	ng	# 84
41) Benzo(g,h,i)perylene	26.359	276	17122	0.365	ng	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102524\
 Data File : BN034721.D
 Acq On : 25 Oct 2024 09:46
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Oct 25 10:44:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 24 13:17:02 2024
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.705 min Scan# 669
Delta R.T. -0.007 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:152 Resp: 6565
Ion Ratio Lower Upper
152 100
150 157.4 122.7 184.1
115 67.9 51.0 76.6




#2
1,4-Dioxane
Concen: 0.345 ng
RT: 3.285 min Scan# 56
Delta R.T. 0.000 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46

Tgt Ion: 88 Resp: 3018
Ion Ratio Lower Upper
88 100
43 42.9 32.1 48.1
58 94.3 73.7 110.5



Abundance Scan 97 (3.581 min): BN034686.D\data.ms (-93)



#3

n-Nitrosodimethylamine

Concen: 0.334 ng

RT: 3.581 min Scan# 91

Delta R.T. 0.000 min

Lab File: BN034721.D

Acq: 25 Oct 2024 09:46

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4



Tgt Ion: 42 Resp: 4242

Ion Ratio Lower Upper


42 100

74 104.9 77.2 115.8


44 11.5 13.4 20.2#



Abundance Scan 97 (3.581 min): BN034721.D\data.ms (-76)



Abundance Scan 338 (5.322 min): BN034686.D\data.ms (-33)



#4

2-Fluorophenol

Concen: 0.431 ng

RT: 5.314 min Scan# 337

Delta R.T. -0.007 min

Lab File: BN034721.D

Acq: 25 Oct 2024 09:46



Tgt Ion: 112 Resp: 8404

Ion Ratio Lower Upper


112 100


64 67.2 53.0 79.6

63 34.9 27.6 41.4



Abundance Scan 337 (5.314 min): BN034721.D\data.ms (-31)









#5
Phenol-d6
Concen: 0.442 ng
RT: 6.882 min Scan# 511
Delta R.T. 0.000 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 11379
Ion Ratio Lower Upper
99 100
42 26.5 21.8 32.8
71 33.6 26.6 40.0



#6
bis(2-Chloroethyl)ether
Concen: 0.374 ng
RT: 7.142 min Scan# 590
Delta R.T. 0.000 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46

Tgt Ion: 93 Resp: 7827
Ion Ratio Lower Upper
93 100
63 88.1 71.5 107.3
95 32.2 25.8 38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	19566
Ion	Ratio	Lower	Upper
136	100		
137	11.4	9.1	13.7
54	11.0	9.8	14.8
68	7.4	6.4	9.6



#8
Nitrobenzene-d5
Concen: 0.358 ng
RT: 8.845 min Scan# 802
Delta R.T. 0.000 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46

Tgt Ion:	82	Resp:	5869
Ion	Ratio	Lower	Upper
82	100		
128	40.5	29.7	44.5
54	64.0	53.0	79.6





#9

Naphthalene

Concen: 0.364 ng

RT: 10.522 min Scan# 91

Delta R.T. -0.011 min

Lab File: BN034721.D

Acq: 25 Oct 2024 09:46

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:128 Resp: 19690

Ion Ratio Lower Upper

128 100

129 11.6 9.0 13.6

127 13.7 10.5 15.7



#10

Hexachlorobutadiene

Concen: 0.351 ng

RT: 10.821 min Scan# 987

Delta R.T. -0.011 min

Lab File: BN034721.D

Acq: 25 Oct 2024 09:46




Tgt Ion:225 Resp: 2852


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 65.4 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 0.357 ng
RT: 12.067 min Scan# 11
Delta R.T. -0.004 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 9607
Ion Ratio Lower Upper
152 100
151 20.8 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.360 ng
RT: 12.143 min Scan# 1220
Delta R.T. -0.004 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46

Tgt Ion:142 Resp: 12083
Ion Ratio Lower Upper
142 100
141 88.8 70.4 105.6
115 38.7 29.8 44.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.325 min Scan# 1504
Delta R.T. -0.004 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 9604
Ion Ratio Lower Upper
164 100
162 104.7 81.4 122.0
160 50.2 40.6 61.0



#14
2,4,6-Tribromophenol
Concen: 0.384 ng
RT: 15.810 min Scan# 1638
Delta R.T. -0.005 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46

Tgt Ion:330 Resp: 1142
Ion Ratio Lower Upper
330 100
332 97.1 76.6 115.0
141 63.8 54.6 81.8









#15
2-Fluorobiphenyl
Concen: 0.351 ng
RT: 12.946 min Scan# 1375
Delta R.T. -0.004 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	172	Resp:	13407
Ion Ratio	Lower	Upper	
172	100		
171	35.7	28.3	42.5
170	24.6	19.8	29.6



#16
Acenaphthylene
Concen: 0.339 ng
RT: 14.037 min Scan# 1476
Delta R.T. -0.004 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46

Tgt Ion:	152	Resp:	15625
Ion Ratio	Lower	Upper	
152	100		
151	19.3	15.6	23.4
153	12.8	10.6	15.8









#17
Acenaphthene
Concen: 0.342 ng
RT: 14.389 min Scan# 1510
Delta R.T. -0.004 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 10877
Ion Ratio Lower Upper
154 100
153 110.3 88.2 132.2
152 56.4 46.9 70.3



#18
Fluorene
Concen: 0.324 ng
RT: 15.373 min Scan# 1601
Delta R.T. -0.004 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46

Tgt Ion:166 Resp: 12726
Ion Ratio Lower Upper
166 100
165 99.3 78.8 118.2
167 13.5 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.064 min Scan# 11
Delta R.T. -0.005 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	12.9	8.9	13.3



#20
4,6-Dinitro-2-methylphenol
Concen: 0.337 ng
RT: 15.448 min Scan# 1608
Delta R.T. -0.004 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46

Tgt Ion	Ratio	Lower	Upper
198	100		
51	117.2	91.8	137.8
105	56.4	44.3	66.5









#21
4-Bromophenyl-phenylether
Concen: 0.365 ng
RT: 16.269 min Scan# 1675
Delta R.T. -0.005 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
248	100		
250	99.7	80.9	121.3
141	57.6	40.5	60.7



#22
Hexachlorobenzene
Concen: 0.357 ng
RT: 16.368 min Scan# 1683
Delta R.T. -0.005 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46

Tgt Ion	Ratio	Lower	Upper
284	100		
142	56.3	43.6	65.4
249	34.7	26.6	40.0









#23
Atrazine
Concen: 0.363 ng
RT: 16.542 min Scan# 11
Delta R.T. -0.005 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
200	100		
173	26.5	19.7	29.5
215	44.4	35.7	53.5



#24
Pentachlorophenol
Concen: 0.378 ng
RT: 16.716 min Scan# 1711
Delta R.T. -0.005 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46

Tgt Ion	Ratio	Lower	Upper
266	100		
264	64.4	49.4	74.2
268	64.0	50.5	75.7









#25
Phenanthrene
Concen: 0.360 ng
RT: 17.101 min Scan# 1742
Delta R.T. -0.005 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 19007
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.2
179 15.4 12.2 18.4



#26
Anthracene
Concen: 0.358 ng
RT: 17.188 min Scan# 1749
Delta R.T. -0.005 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46

Tgt Ion:178 Resp: 16769
Ion Ratio Lower Upper
178 100
176 18.6 14.9 22.3
179 15.0 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.370 ng
RT: 19.092 min Scan# 1980
Delta R.T. -0.002 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
212	100		
106	23.1	17.8	26.8
104	13.0	10.2	15.4



#28
Fluoranthene
Concen: 0.368 ng
RT: 19.120 min Scan# 1986
Delta R.T. -0.007 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46

Tgt Ion	Ratio	Lower	Upper
202	100		
101	16.6	12.8	19.2
203	17.0	13.5	20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.248 min Scan# 2419
Delta R.T. -0.002 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
240	100		
120	17.9	14.2	21.2
236	29.0	23.1	34.7



#30
Pyrene
Concen: 0.315 ng
RT: 19.482 min Scan# 2064
Delta R.T. -0.002 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46

Tgt Ion	Ratio	Lower	Upper
202	100		
200	21.2	16.6	25.0
203	18.0	14.2	21.4









#31
Terphenyl-d14
Concen: 0.338 ng
RT: 19.700 min Scan# 2113
Delta R.T. -0.002 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 9017
Ion Ratio Lower Upper
244 100
212 10.5 8.6 13.0
122 25.6 22.3 33.5



#32
Benzo(a)anthracene
Concen: 0.374 ng
RT: 21.230 min Scan# 2417
Delta R.T. -0.002 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46

Tgt Ion:228 Resp: 18695
Ion Ratio Lower Upper
228 100
226 27.1 22.0 33.0
229 19.8 15.6 23.4





#33

Chrysene

Concen: 0.349 ng

RT: 21.283 min Scan# 2425

Delta R.T. -0.002 min

Lab File: BN034721.D

Acq: 25 Oct 2024 09:46

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:228 Resp: 18028

Ion Ratio Lower Upper

228 100

226 29.8 24.1 36.1

229 20.4 15.8 23.8



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.532 ng

RT: 21.194 min Scan# 2413

Delta R.T. -0.002 min

Lab File: BN034721.D

Acq: 25 Oct 2024 09:46




Tgt Ion:149 Resp: 15075


Ion Ratio Lower Upper

149 100

167 24.5 19.7 29.5

279 2.0 1.8 2.6









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.459 min Scan# 2913
Delta R.T. -0.008 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:264 Resp: 14018
Ion Ratio Lower Upper
264 100
260 26.6 21.9 32.9
265 65.0 60.6 91.0



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.369 ng
RT: 25.683 min Scan# 3668
Delta R.T. -0.011 min
Lab File: BN034721.D
Acq: 25 Oct 2024 09:46

Tgt Ion:276 Resp: 19938
Ion Ratio Lower Upper
276 100
138 45.0 35.1 52.7
277 24.8 19.7 29.5





#37

Benzo(b)fluoranthene

Concen: 0.330 ng

RT: 22.798 min Scan# 2681

Delta R.T. -0.005 min

Lab File: BN034721.D

Acq: 25 Oct 2024 09:46

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:252 Resp: 18466

Ion Ratio Lower Upper

252 100

253 25.4 20.0 30.0

125 40.1 20.9 31.3#



#38

Benzo(k)fluoranthene

Concen: 0.326 ng

RT: 22.842 min Scan# 2696

Delta R.T. -0.008 min

Lab File: BN034721.D

Acq: 25 Oct 2024 09:46




Tgt Ion:252 Resp: 17979


Ion Ratio Lower Upper

252 100

253 25.2 20.6 31.0

125 42.9 21.6 32.4#





#39

Benzo(a)pyrene

Concen: 0.363 ng

RT: 23.362 min Scan# 21

Delta R.T. -0.005 min

Lab File: BN034721.D

Acq: 25 Oct 2024 09:46

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:252 Resp: 16295

Ion Ratio Lower Upper

252 100

253 26.8 22.2 33.4

125 49.4 26.0 39.0#



#40

Dibenzo(a,h)anthracene

Concen: 0.369 ng

RT: 25.707 min Scan# 3676

Delta R.T. -0.008 min

Lab File: BN034721.D

Acq: 25 Oct 2024 09:46




Tgt Ion:278 Resp: 15708


Ion Ratio Lower Upper

278 100

139 47.3 25.4 38.2#

279 27.2 22.4 33.6





#41

Benzo(g,h,i)perylene

Concen: 0.365 ng

RT: 26.359 min Scan# 3899

Delta R.T. -0.014 min

Lab File: BN034721.D


Acq: 25 Oct 2024 09:46

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:276 Resp: 17122

Ion Ratio Lower Upper

276 100

277 25.6 21.3 31.9

138 46.9 31.9 47.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102524\
Data File : BN034721.D
Acq On : 25 Oct 2024 09:46
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 25 10:44:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.532	0.460	13.5	98	0.00
3	n-Nitrosodimethylamine	0.775	0.646	16.6	95	0.00
4 S	2-Fluorophenol	1.189	1.280	-7.7	124	0.00
5 S	Phenol-d6	1.568	1.733	-10.5	130	0.00
6	bis(2-Chloroethyl)ether	1.274	1.192	6.4	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
8 S	Nitrobenzene-d5	0.336	0.300	10.7	108	0.00
9	Naphthalene	1.107	1.006	9.1	108	-0.01
10	Hexachlorobutadiene	0.166	0.146	12.0	103	-0.01
11 SURR	2-Methylnaphthalene-d10	0.550	0.491	10.7	108	0.00
12	2-Methylnaphthalene	0.687	0.618	10.0	109	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.124	0.119	4.0	127	0.00
15 S	2-Fluorobiphenyl	1.590	1.396	12.2	104	0.00
16	Acenaphthylene	1.921	1.627	15.3	106	0.00
17	Acenaphthene	1.324	1.133	14.4	104	0.00
18	Fluorene	1.634	1.325	18.9	98	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4,6-Dinitro-2-methylphenol	0.056	0.047	16.1	106	0.00
21	4-Bromophenyl-phenylether	0.213	0.195	8.5	99	0.00
22	Hexachlorobenzene	0.231	0.206	10.8	94	0.00
23	Atrazine	0.168	0.152	9.5	105	0.00
24	Pentachlorophenol	0.082	0.077	6.1	117	0.00
25	Phenanthrene	1.202	1.081	10.1	98	0.00
26	Anthracene	1.067	0.954	10.6	99	0.00
27 SURR	Fluoranthene-d10	0.901	0.834	7.4	103	0.00
28	Fluoranthene	1.253	1.154	7.9	102	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	123	0.00
30	Pyrene	2.055	1.616	21.4	106	0.00
31 S	Terphenyl-d14	0.817	0.690	15.5	115	0.00
32	Benzo(a)anthracene	1.528	1.431	6.3	128	0.00
33	Chrysene	1.581	1.380	12.7	118	0.00
34	Bis(2-ethylhexyl)phthalate	0.867	1.154	-33.1#	196#	0.00
35 I	Perylene-d12	1.000	1.000	0.0	146	0.00
36	Indeno(1,2,3-cd)pyrene	1.541	1.422	7.7	147	-0.01
37	Benzo(b)fluoranthene	1.597	1.317	17.5	130	0.00
38	Benzo(k)fluoranthene	1.576	1.283	18.6	135	0.00
39 C	Benzo(a)pyrene	1.282	1.162	9.4	147	0.00
40	Dibenzo(a,h)anthracene	1.215	1.121	7.7	147	0.00
41	Benzo(g,h,i)perylene	1.340	1.221	8.9	141	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102524\
Data File : BN034721.D
Acq On : 25 Oct 2024 09:46
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 25 10:44:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	104	0.00
2	1,4-Dioxane	0.400	0.345	13.8	98	0.00
3	n-Nitrosodimethylamine	0.400	0.334	16.5	95	0.00
4 S	2-Fluorophenol	0.400	0.431	-7.7	124	0.00
5 S	Phenol-d6	0.400	0.442	-10.5	130	0.00
6	bis(2-Chloroethyl)ether	0.400	0.374	6.5	109	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	106	0.00
8 S	Nitrobenzene-d5	0.400	0.358	10.5	108	0.00
9	Naphthalene	0.400	0.364	9.0	108	-0.01
10	Hexachlorobutadiene	0.400	0.351	12.3	103	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.357	10.8	108	0.00
12	2-Methylnaphthalene	0.400	0.360	10.0	109	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.400	0.384	4.0	127	0.00
15 S	2-Fluorobiphenyl	0.400	0.351	12.3	104	0.00
16	Acenaphthylene	0.400	0.339	15.3	106	0.00
17	Acenaphthene	0.400	0.342	14.5	104	0.00
18	Fluorene	0.400	0.324	19.0	98	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	97	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.337	15.8	106	0.00
21	4-Bromophenyl-phenylether	0.400	0.365	8.8	99	0.00
22	Hexachlorobenzene	0.400	0.357	10.8	94	0.00
23	Atrazine	0.400	0.363	9.3	105	0.00
24	Pentachlorophenol	0.400	0.378	5.5	117	0.00
25	Phenanthrene	0.400	0.360	10.0	98	0.00
26	Anthracene	0.400	0.358	10.5	99	0.00
27 SURR	Fluoranthene-d10	0.400	0.370	7.5	103	0.00
28	Fluoranthene	0.400	0.368	8.0	102	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	123	0.00
30	Pyrene	0.400	0.315	21.3	106	0.00
31 S	Terphenyl-d14	0.400	0.338	15.5	115	0.00
32	Benzo(a)anthracene	0.400	0.374	6.5	128	0.00
33	Chrysene	0.400	0.349	12.8	118	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.532	-33.0#	196	0.00
35 I	Perylene-d12	0.400	0.400	0.0	146	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.369	7.8	147	-0.01
37	Benzo(b)fluoranthene	0.400	0.330	17.5	130	0.00
38	Benzo(k)fluoranthene	0.400	0.326	18.5	135	0.00
39 C	Benzo(a)pyrene	0.400	0.363	9.3	147	0.00
40	Dibenzo(a,h)anthracene	0.400	0.369	7.8	147	0.00
41	Benzo(g,h,i)perylene	0.400	0.365	8.8	141	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG No.: P4440

Instrument ID: BNA_N Calibration Date/Time: 10/25/2024 20:41

Lab File ID: BN034732.D Init. Calib. Date(s): 10/24/2024 10/24/2024

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 09:11 12:48

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.550	0.499		-9.3	50.0
Fluoranthene-d10	0.901	0.778		-13.7	50.0
2-Fluorophenol	1.189	1.284		8.0	50.0
Phenol-d6	1.568	1.673		6.7	50.0
Nitrobenzene-d5	0.336	0.311		-7.4	50.0
2-Fluorobiphenyl	1.590	1.375		-13.5	50.0
2,4,6-Tribromophenol	0.124	0.129		4.0	50.0
Terphenyl-d14	0.817	0.831		1.7	50.0
1,4-Dioxane	0.532	0.481		-9.6	50.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102524\
 Data File : BN034732.D
 Acq On : 25 Oct 2024 20:41
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Oct 25 22:43:59 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 25 22:11:36 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.654	152	8703	0.400	ng	-0.06
7) Naphthalene-d8	10.415	136	25971	0.400	ng	#-0.06
13) Acenaphthene-d10	14.272	164	13484	0.400	ng	-0.06
19) Phenanthrene-d10	17.026	188	25916	0.400	ng	-0.04
29) Chrysene-d12	21.212	240	13865	0.400	ng	-0.04
35) Perylene-d12	23.412	264	12945	0.400	ng	-0.05
System Monitoring Compounds						
4) 2-Fluorophenol	5.264	112	11176	0.432	ng	-0.06
5) Phenol-d6	6.824	99	14560	0.427	ng	-0.06
8) Nitrobenzene-d5	8.781	82	8064	0.370	ng	-0.06
11) 2-Methylnaphthalene-d10	12.014	152	12963	0.363	ng	-0.06
14) 2,4,6-Tribromophenol	15.773	330	1738	0.417	ng	-0.04
15) 2-Fluorobiphenyl	12.903	172	18535	0.346	ng	-0.05
27) Fluoranthene-d10	19.055	212	20172	0.346	ng	-0.04
31) Terphenyl-d14	19.663	244	11526	0.407	ng	-0.04
Target Compounds						
2) 1,4-Dioxane	3.242	88	4182	0.361	ng	99
3) n-Nitrosodimethylamine	3.538	42	5887	0.349	ng	# 90
6) bis(2-Chloroethyl)ether	7.084	93	10361	0.374	ng	98
9) Naphthalene	10.468	128	26161	0.364	ng	100
10) Hexachlorobutadiene	10.767	225	3881	0.360	ng	# 99
12) 2-Methylnaphthalene	12.090	142	16236	0.364	ng	100
16) Acenaphthylene	13.994	152	21659	0.334	ng	99
17) Acenaphthene	14.336	154	15327	0.343	ng	99
18) Fluorene	15.330	166	18945	0.344	ng	99
20) 4,6-Dinitro-2-methylph...	15.405	198	1194	0.331	ng	# 84
21) 4-Bromophenyl-phenylether	16.232	248	5089	0.368	ng	93
22) Hexachlorobenzene	16.331	284	5412	0.361	ng	99
23) Atrazine	16.505	200	4155	0.383	ng	99
24) Pentachlorophenol	16.679	266	2169	0.409	ng	100
25) Phenanthrene	17.064	178	28002	0.360	ng	100
26) Anthracene	17.150	178	24377	0.353	ng	100
28) Fluoranthene	19.087	202	28255	0.348	ng	99
30) Pyrene	19.445	202	28410	0.399	ng	100
32) Benzo(a)anthracene	21.194	228	19350	0.365	ng	100
33) Chrysene	21.248	228	19781	0.361	ng	100
34) Bis(2-ethylhexyl)phtha...	21.158	149	15745	0.524	ng	99
36) Indeno(1,2,3-cd)pyrene	25.619	276	17779	0.357	ng	99
37) Benzo(b)fluoranthene	22.754	252	18264	0.353	ng	100
38) Benzo(k)fluoranthene	22.801	252	17879	0.351	ng	99
39) Benzo(a)pyrene	23.315	252	14600	0.352	ng	97
40) Dibenzo(a,h)anthracene	25.634	278	14173	0.360	ng	97
41) Benzo(g,h,i)perylene	26.286	276	15303	0.353	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102524\
Data File : BN034732.D
Acq On : 25 Oct 2024 20:41
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Oct 25 22:43:59 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 25 22:11:36 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.654 min Scan# 60
Delta R.T. -0.058 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:152 Resp: 8703
Ion Ratio Lower Upper
152 100
150 153.7 122.7 184.1
115 62.2 51.0 76.6




#2
1,4-Dioxane
Concen: 0.361 ng
RT: 3.242 min Scan# 50
Delta R.T. -0.043 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41

Tgt Ion: 88 Resp: 4182
Ion Ratio Lower Upper
88 100
43 40.9 32.1 48.1
58 90.9 73.7 110.5



Abundance Scan 97 (3.581 min): BN034686.D\data.ms (-93)



#3

n-Nitrosodimethylamine

Concen: 0.349 ng

RT: 3.538 min Scan# 91

Delta R.T. -0.043 min

Lab File: BN034732.D

Acq: 25 Oct 2024 20:41

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion: 42 Resp: 5887



Ion Ratio Lower Upper

42 100


74 104.8 77.2 115.8

44 10.2 13.4 20.2#


Abundance Scan 91 (3.538 min): BN034732.D\data.ms



Abundance Scan 91 (3.538 min): BN034732.D\data.ms (-76)



Abundance Scan 338 (5.322 min): BN034686.D\data.ms (-33)



#4

2-Fluorophenol

Concen: 0.432 ng

RT: 5.264 min Scan# 330

Delta R.T. -0.058 min

Lab File: BN034732.D

Acq: 25 Oct 2024 20:41

Tgt Ion: 112 Resp: 11176



Ion Ratio Lower Upper

112 100


64 65.2 53.0 79.6


63 34.2 27.6 41.4

Abundance Scan 330 (5.264 min): BN034732.D\data.ms



Abundance Scan 330 (5.264 min): BN034732.D\data.ms (-31)









#5
Phenol-d6
Concen: 0.427 ng
RT: 6.824 min Scan# 54
Delta R.T. -0.058 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 14560
Ion Ratio Lower Upper
99 100
42 26.2 21.8 32.8
71 32.7 26.6 40.0



#6
bis(2-Chloroethyl)ether
Concen: 0.374 ng
RT: 7.084 min Scan# 582
Delta R.T. -0.058 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41

Tgt Ion: 93 Resp: 10361
Ion Ratio Lower Upper
93 100
63 87.6 71.5 107.3
95 32.5 25.8 38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.415 min Scan# 949
Delta R.T. -0.064 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	25971
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.1	13.7
54	14.9	9.8	14.8#
68	9.7	6.4	9.6#



#8
Nitrobenzene-d5
Concen: 0.370 ng
RT: 8.781 min Scan# 796
Delta R.T. -0.064 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41

Tgt Ion:	82	Resp:	8064
Ion	Ratio	Lower	Upper
82	100		
128	34.0	29.7	44.5
54	68.8	53.0	79.6









#9
Naphthalene
Concen: 0.364 ng
RT: 10.468 min Scan# 91
Delta R.T. -0.064 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:128 Resp: 26161
Ion Ratio Lower Upper
128 100
129 11.1 9.0 13.6
127 13.1 10.5 15.7



#10
Hexachlorobutadiene
Concen: 0.360 ng
RT: 10.767 min Scan# 982
Delta R.T. -0.064 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41

Tgt Ion:225 Resp: 3881
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.1 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 0.363 ng
RT: 12.014 min Scan# 11
Delta R.T. -0.057 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 12963
Ion Ratio Lower Upper
152 100
151 20.7 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.364 ng
RT: 12.090 min Scan# 1206
Delta R.T. -0.057 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41

Tgt Ion:142 Resp: 16236
Ion Ratio Lower Upper
142 100
141 88.0 70.4 105.6
115 36.7 29.8 44.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.272 min Scan# 1498
Delta R.T. -0.057 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion	Ratio	Lower	Upper
164	100		
162	106.7	81.4	122.0
160	54.1	40.6	61.0



#14
2,4,6-Tribromophenol
Concen: 0.417 ng
RT: 15.773 min Scan# 1635
Delta R.T. -0.043 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41

Tgt Ion	Ratio	Lower	Upper
330	100		
332	95.5	76.6	115.0
141	63.9	54.6	81.8









#15
2-Fluorobiphenyl
Concen: 0.346 ng
RT: 12.903 min Scan# 11
Delta R.T. -0.046 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	172	Resp:	18535
Ion Ratio	Lower	Upper	
172	100		
171	33.6	28.3	42.5
170	23.3	19.8	29.6



#16
Acenaphthylene
Concen: 0.334 ng
RT: 13.994 min Scan# 1472
Delta R.T. -0.046 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41

Tgt Ion:	152	Resp:	21659
Ion Ratio	Lower	Upper	
152	100		
151	19.1	15.6	23.4
153	12.6	10.6	15.8









#17
Acenaphthene
Concen: 0.343 ng
RT: 14.336 min Scan# 1510
Delta R.T. -0.057 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 15327
Ion Ratio Lower Upper
154 100
153 109.6 88.2 132.2
152 56.5 46.9 70.3



#18
Fluorene
Concen: 0.344 ng
RT: 15.330 min Scan# 1597
Delta R.T. -0.046 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41

Tgt Ion:166 Resp: 18945
Ion Ratio Lower Upper
166 100
165 97.6 78.8 118.2
167 13.2 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.026 min Scan# 11
Delta R.T. -0.043 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 25916
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.6 8.9 13.3



#20
4,6-Dinitro-2-methylphenol
Concen: 0.331 ng
RT: 15.405 min Scan# 1604
Delta R.T. -0.047 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41

Tgt Ion:198 Resp: 1194
Ion Ratio Lower Upper
198 100
51 89.8 91.8 137.8#
105 54.4 44.3 66.5









#21
4-Bromophenyl-phenylether
Concen: 0.368 ng
RT: 16.232 min Scan# 1676
Delta R.T. -0.043 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion	Ratio	Lower	Upper
248	100		
250	106.0	80.9	121.3
141	43.4	40.5	60.7



#22
Hexachlorobenzene
Concen: 0.361 ng
RT: 16.331 min Scan# 1680
Delta R.T. -0.043 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41

Tgt Ion	Ratio	Lower	Upper
284	100		
142	54.4	43.6	65.4
249	34.0	26.6	40.0









#23
Atrazine
Concen: 0.383 ng
RT: 16.505 min Scan# 1
Delta R.T. -0.043 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:200 Resp: 4155
Ion Ratio Lower Upper
200 100
173 25.9 19.7 29.5
215 45.2 35.7 53.5



#24
Pentachlorophenol
Concen: 0.409 ng
RT: 16.679 min Scan# 1708
Delta R.T. -0.042 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41

Tgt Ion:266 Resp: 2169
Ion Ratio Lower Upper
266 100
264 62.4 49.4 74.2
268 63.2 50.5 75.7









#25
Phenanthrene
Concen: 0.360 ng
RT: 17.064 min Scan# 1743
Delta R.T. -0.043 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 28002
Ion Ratio Lower Upper
178 100
176 19.1 15.4 23.2
179 15.1 12.2 18.4



#26
Anthracene
Concen: 0.353 ng
RT: 17.150 min Scan# 1746
Delta R.T. -0.043 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41

Tgt Ion:178 Resp: 24377
Ion Ratio Lower Upper
178 100
176 18.5 14.9 22.3
179 15.1 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.346 ng
RT: 19.055 min Scan# 1972
Delta R.T. -0.039 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion	Ratio	Lower	Upper
212	100		
106	22.7	17.8	26.8
104	12.9	10.2	15.4



#28
Fluoranthene
Concen: 0.348 ng
RT: 19.087 min Scan# 1979
Delta R.T. -0.039 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41

Tgt Ion	Ratio	Lower	Upper
202	100		
101	16.2	12.8	19.2
203	17.2	13.5	20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.212 min Scan# 2415
Delta R.T. -0.038 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion	Ratio	Lower	Upper
240	100		
120	20.3	14.2	21.2
236	29.0	23.1	34.7



#30
Pyrene
Concen: 0.399 ng
RT: 19.445 min Scan# 2056
Delta R.T. -0.039 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41

Tgt Ion	Ratio	Lower	Upper
202	100		
200	21.1	16.6	25.0
203	17.8	14.2	21.4









#31
Terphenyl-d14
Concen: 0.407 ng
RT: 19.663 min Scan# 2113
Delta R.T. -0.039 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:244 Resp: 11526
Ion Ratio Lower Upper
244 100
212 11.0 8.6 13.0
122 29.3 22.3 33.5



#32
Benzo(a)anthracene
Concen: 0.365 ng
RT: 21.194 min Scan# 2413
Delta R.T. -0.038 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41

Tgt Ion:228 Resp: 19350
Ion Ratio Lower Upper
228 100
226 27.8 22.0 33.0
229 19.4 15.6 23.4





#33

Chrysene

Concen: 0.361 ng

RT: 21.248 min Scan# 2419

Delta R.T. -0.038 min

Lab File: BN034732.D

Acq: 25 Oct 2024 20:41

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:228 Resp: 19781

Ion Ratio Lower Upper

228 100

226 30.2 24.1 36.1

229 19.6 15.8 23.8



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.524 ng

RT: 21.158 min Scan# 2409

Delta R.T. -0.038 min

Lab File: BN034732.D

Acq: 25 Oct 2024 20:41




Tgt Ion:149 Resp: 15745


Ion Ratio Lower Upper

149 100

167 24.9 19.7 29.5

279 1.8 1.8 2.6









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.412 min Scan# 21
Delta R.T. -0.055 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:264 Resp: 12945
Ion Ratio Lower Upper
264 100
260 27.0 21.9 32.9
265 62.7 60.6 91.0



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.357 ng
RT: 25.619 min Scan# 3648
Delta R.T. -0.075 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41

Tgt Ion:276 Resp: 17779
Ion Ratio Lower Upper
276 100
138 44.6 35.1 52.7
277 24.4 19.7 29.5





#37

Benzo(b)fluoranthene

Concen: 0.353 ng

RT: 22.754 min Scan# 2668

Delta R.T. -0.049 min

Lab File: BN034732.D

Acq: 25 Oct 2024 20:41

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:252 Resp: 18264

Ion Ratio Lower Upper

252 100

253 24.9 20.0 30.0

125 25.7 20.9 31.3



#38

Benzo(k)fluoranthene

Concen: 0.351 ng

RT: 22.801 min Scan# 2684

Delta R.T. -0.049 min

Lab File: BN034732.D

Acq: 25 Oct 2024 20:41




Tgt Ion:252 Resp: 17879


Ion Ratio Lower Upper

252 100

253 25.1 20.6 31.0

125 27.6 21.6 32.4









#39
Benzo(a)pyrene
Concen: 0.352 ng
RT: 23.315 min Scan# 21
Delta R.T. -0.052 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	252	Resp:	14600
Ion Ratio	Lower	Upper	
252	100		
253	26.0	22.2	33.4
125	31.4	26.0	39.0



#40
Dibenzo(a,h)anthracene
Concen: 0.360 ng
RT: 25.634 min Scan# 3653
Delta R.T. -0.081 min
Lab File: BN034732.D
Acq: 25 Oct 2024 20:41

Tgt Ion:	278	Resp:	14173
Ion Ratio	Lower	Upper	
278	100		
139	34.8	25.4	38.2
279	27.5	22.4	33.6





#41

Benzo(g,h,i)perylene

Concen: 0.353 ng

RT: 26.286 min Scan# 31

Delta R.T. -0.087 min

Lab File: BN034732.D


Acq: 25 Oct 2024 20:41

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:276 Resp: 15303

Ion Ratio Lower Upper

276 100

277 24.7 21.3 31.9

138 38.8 31.9 47.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102524\
Data File : BN034732.D
Acq On : 25 Oct 2024 20:41
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 25 22:43:59 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 25 22:11:36 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	138	-0.06
2	1,4-Dioxane	0.532	0.481	9.6	136	-0.04
3	n-Nitrosodimethylamine	0.775	0.676	12.8	132	-0.04
4 S	2-Fluorophenol	1.189	1.284	-8.0	165#	-0.06
5 S	Phenol-d6	1.568	1.673	-6.7	166#	-0.06
6	bis(2-Chloroethyl)ether	1.274	1.191	6.5	144	-0.06
7 I	Naphthalene-d8	1.000	1.000	0.0	140	-0.06
8 S	Nitrobenzene-d5	0.336	0.311	7.4	148	-0.06
9	Naphthalene	1.107	1.007	9.0	143	-0.06
10	Hexachlorobutadiene	0.166	0.149	10.2	141	-0.06
11 SURR	2-Methylnaphthalene-d10	0.550	0.499	9.3	146	-0.06
12	2-Methylnaphthalene	0.687	0.625	9.0	146	-0.06
13 I	Acenaphthene-d10	1.000	1.000	0.0	145	-0.06
14 S	2,4,6-Tribromophenol	0.124	0.129	-4.0	194#	-0.04
15 S	2-Fluorobiphenyl	1.590	1.375	13.5	144	-0.05
16	Acenaphthylene	1.921	1.606	16.4	147	-0.05
17	Acenaphthene	1.324	1.137	14.1	147	-0.06
18	Fluorene	1.634	1.405	14.0	147	-0.05
19 I	Phenanthrene-d10	1.000	1.000	0.0	143	-0.04
20	4,6-Dinitro-2-methylphenol	0.056	0.046	17.9	154#	-0.05
21	4-Bromophenyl-phenylether	0.213	0.196	8.0	147	-0.04
22	Hexachlorobenzene	0.231	0.209	9.5	141	-0.04
23	Atrazine	0.168	0.160	4.8	163#	-0.04
24	Pentachlorophenol	0.082	0.084	-2.4	186#	-0.04
25	Phenanthrene	1.202	1.080	10.1	145	-0.04
26	Anthracene	1.067	0.941	11.8	144	-0.04
27 SURR	Fluoranthene-d10	0.901	0.778	13.7	142	-0.04
28	Fluoranthene	1.253	1.090	13.0	142	-0.04
29 I	Chrysene-d12	1.000	1.000	0.0	130	-0.04
30	Pyrene	2.055	2.049	0.3	142	-0.04
31 S	Terphenyl-d14	0.817	0.831	-1.7	147	-0.04
32	Benzo(a)anthracene	1.528	1.396	8.6	133	-0.04
33	Chrysene	1.581	1.427	9.7	129	-0.04
34	Bis(2-ethylhexyl)phthalate	0.867	1.136	-31.0#	205#	-0.04
35 I	Perylene-d12	1.000	1.000	0.0	135	-0.05
36	Indeno(1,2,3-cd)pyrene	1.541	1.373	10.9	131	-0.08
37	Benzo(b)fluoranthene	1.597	1.411	11.6	129	-0.05
38	Benzo(k)fluoranthene	1.576	1.381	12.4	134	-0.05
39 C	Benzo(a)pyrene	1.282	1.128	12.0	131	-0.05
40	Dibenzo(a,h)anthracene	1.215	1.095	9.9	133	-0.08
41	Benzo(g,h,i)perylene	1.340	1.182	11.8	126	-0.09

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102524\
Data File : BN034732.D
Acq On : 25 Oct 2024 20:41
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 25 22:43:59 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 25 22:11:36 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	138	-0.06
2	1,4-Dioxane	0.400	0.361	9.8	136	-0.04
3	n-Nitrosodimethylamine	0.400	0.349	12.8	132	-0.04
4 S	2-Fluorophenol	0.400	0.432	-8.0	165	-0.06
5 S	Phenol-d6	0.400	0.427	-6.7	166	-0.06
6	bis(2-Chloroethyl)ether	0.400	0.374	6.5	144	-0.06
7 I	Naphthalene-d8	0.400	0.400	0.0	140	-0.06
8 S	Nitrobenzene-d5	0.400	0.370	7.5	148	-0.06
9	Naphthalene	0.400	0.364	9.0	143	-0.06
10	Hexachlorobutadiene	0.400	0.360	10.0	141	-0.06
11 SURR	2-Methylnaphthalene-d10	0.400	0.363	9.3	146	-0.06
12	2-Methylnaphthalene	0.400	0.364	9.0	146	-0.06
13 I	Acenaphthene-d10	0.400	0.400	0.0	145	-0.06
14 S	2,4,6-Tribromophenol	0.400	0.417	-4.2	194	-0.04
15 S	2-Fluorobiphenyl	0.400	0.346	13.5	144	-0.05
16	Acenaphthylene	0.400	0.334	16.5	147	-0.05
17	Acenaphthene	0.400	0.343	14.2	147	-0.06
18	Fluorene	0.400	0.344	14.0	147	-0.05
19 I	Phenanthrene-d10	0.400	0.400	0.0	143	-0.04
20	4,6-Dinitro-2-methylphenol	0.400	0.331	17.3	154	-0.05
21	4-Bromophenyl-phenylether	0.400	0.368	8.0	147	-0.04
22	Hexachlorobenzene	0.400	0.361	9.8	141	-0.04
23	Atrazine	0.400	0.383	4.3	163	-0.04
24	Pentachlorophenol	0.400	0.409	-2.2	186	-0.04
25	Phenanthrene	0.400	0.360	10.0	145	-0.04
26	Anthracene	0.400	0.353	11.8	144	-0.04
27 SURR	Fluoranthene-d10	0.400	0.346	13.5	142	-0.04
28	Fluoranthene	0.400	0.348	13.0	142	-0.04
29 I	Chrysene-d12	0.400	0.400	0.0	130	-0.04
30	Pyrene	0.400	0.399	0.3	142	-0.04
31 S	Terphenyl-d14	0.400	0.407	-1.7	147	-0.04
32	Benzo(a)anthracene	0.400	0.365	8.8	133	-0.04
33	Chrysene	0.400	0.361	9.8	129	-0.04
34	Bis(2-ethylhexyl)phthalate	0.400	0.524	-31.0#	205	-0.04
35 I	Perylene-d12	0.400	0.400	0.0	135	-0.05
36	Indeno(1,2,3-cd)pyrene	0.400	0.357	10.8	131	-0.08
37	Benzo(b)fluoranthene	0.400	0.353	11.8	129	-0.05
38	Benzo(k)fluoranthene	0.400	0.351	12.3	134	-0.05
39 C	Benzo(a)pyrene	0.400	0.352	12.0	131	-0.05
40	Dibenzo(a,h)anthracene	0.400	0.360	10.0	133	-0.08
41	Benzo(g,h,i)perylene	0.400	0.353	11.8	126	-0.09

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



QC SAMPLE DATA



- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034683.D
Acq On : 24 Oct 2024 07:59
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Oct 24 13:17:02 2024






AutoFind: Scans 2490, 2491, 2492; Background Corrected with Scan 2483

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	66.1	282114	PASS
68	69	0.00	2	1.7	3908	PASS
69	198	0.00	100	54.4	232312	PASS
70	69	0.00	2	0.5	1275	PASS
127	198	10	80	60.3	257365	PASS
197	198	0.00	2	0.2	1012	PASS
198	198	100	100	100.0	426901	PASS
199	198	5	9	6.5	27752	PASS
275	198	10	60	20.6	88117	PASS
365	198	1	100	2.5	10758	PASS
441	198	0.01	100	7.8	33160	PASS
442	442	50	100	100.0	210005	PASS
443	442	15	24	18.6	39008	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034683.D
Acq On : 24 Oct 2024 07:59
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 24 12:50:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 23 02:40:52 2024
Response via : Initial Calibration



TIC: BN034683.D\data.ms

(70) Pentachlorophenol (C)

16.716min (-0.010) 17805.12 ng




response 375874

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	61.45
264.00	61.60	62.35
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034683.D
Acq On : 24 Oct 2024 07:59
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 24 12:50:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 23 02:40:52 2024
Response via : Initial Calibration



TIC: BN034683.D\data.ms

(77) Benzidine

19.316min (+ 0.002) 0.00 ng

response 3022437

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.04
183.00	13.20	11.45
0.00	0.00	0.00

DDT Breakdown



Date	Instrument Name	DFTPP Data File
10/24/2024	BNA_N	<u>BN034683.D</u>
Compound Name	Response	Retention Time
DDT	1329547	20.563
DDD	11879	20.175
DDE	111	19.622
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
11990	1341537	0.89

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034700.D
Acq On : 24 Oct 2024 20:22
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Oct 24 13:17:02 2024






AutoFind: Scans 2490, 2491, 2492; Background Corrected with Scan 2483

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	73.1	153072	PASS
68	69	0.00	2	1.6	1977	PASS
69	198	0.00	100	59.5	124685	PASS
70	69	0.00	2	0.5	613	PASS
127	198	10	80	65.2	136648	PASS
197	198	0.00	2	0.5	1148	PASS
198	198	100	100	100.0	209429	PASS
199	198	5	9	6.5	13571	PASS
275	198	10	60	20.7	43333	PASS
365	198	1	100	2.6	5435	PASS
441	198	0.01	100	7.2	15057	PASS
442	442	50	100	100.0	88755	PASS
443	442	15	24	21.6	19165	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034700.D
Acq On : 24 Oct 2024 20:22
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 25 02:24:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 25 02:24:36 2024
Response via : Initial Calibration



TIC: BN034700.D\data.ms

(70) Pentachlorophenol (C)

16.722min (0.000) 17028.21 ng




response 171184

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.80
264.00	61.60	64.51
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034700.D
Acq On : 24 Oct 2024 20:22
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 25 02:24:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 25 02:24:36 2024
Response via : Initial Calibration



TIC: BN034700.D\data.ms

(77) Benzidine

19.316min (0.000) 0.00 ng

response 1621578

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.38
183.00	13.20	11.24
0.00	0.00	0.00

Instrument :
BNA_N
ClientSampleId :
DFTPP

DDT Breakdown



Date	Instrument Name	DFTPP Data File
10/24/2024	BNA_N	<u>BN034700.D</u>
Compound Name	Response	Retention Time
DDT	622621	20.557
DDD	5755	20.169
DDE	0	19.622
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
5755	628376	0.92

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102524\
Data File : BN034720.D
Acq On : 25 Oct 2024 09:07
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Oct 25 22:11:36 2024






AutoFind: Scans 2489, 2490, 2491; Background Corrected with Scan 2482

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	66.6	199713	PASS
68	69	0.00	2	1.6	2696	PASS
69	198	0.00	100	56.4	169054	PASS
70	69	0.00	2	0.4	594	PASS
127	198	10	80	61.5	184363	PASS
197	198	0.00	2	0.3	764	PASS
198	198	100	100	100.0	299904	PASS
199	198	5	9	6.8	20312	PASS
275	198	10	60	21.0	62957	PASS
365	198	1	100	2.8	8432	PASS
441	198	0.01	100	7.4	22267	PASS
442	442	50	100	100.0	145949	PASS
443	442	15	24	19.4	28343	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102524\
Data File : BN034720.D
Acq On : 25 Oct 2024 09:07
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 25 14:47:07 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 25 02:24:36 2024
Response via : Initial Calibration



TIC: BN034720.D\data.ms

(70) Pentachlorophenol (C)

16.716min (-0.010) 16177.27 ng




response 249213

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.29
264.00	61.60	65.12
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102524\
Data File : BN034720.D
Acq On : 25 Oct 2024 09:07
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 25 14:47:07 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 25 02:24:36 2024
Response via : Initial Calibration



TIC: BN034720.D\data.ms

(77) Benzidine

19.316min (+ 0.002) 0.00 ng

response 2170006

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.53
183.00	13.20	11.36
0.00	0.00	0.00

Instrument :
BNA_N
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/25/2024	BNA_N	<u>BN034720.D</u>
Compound Name	Response	Retention Time
DDT	888185	20.557
DDD	14092	20.169
DDE	315	19.61
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
14407	902592	1.60

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	
Project:	CTO WE13	Date Received:	
Client Sample ID:	PB164282BL	SDG No.:	P4440
Lab Sample ID:	PB164282BL	Matrix:	Water
Analytical Method:	SW8270SIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN034705.D	1	10/21/24 08:45	10/24/24 23:26	PB164282

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.33		30 - 150		83%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.34		30 - 150		86%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.34		55 - 111		85%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.37		53 - 106		91%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.39		58 - 132		97%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	6450	7.712				
1146-65-2	Naphthalene-d8	17700	10.479				
15067-26-2	Acenaphthene-d10	7660	14.329				
1517-22-2	Phenanthrene-d10	13700	17.069				
1719-03-5	Chrysene-d12	7480	21.241				
1520-96-3	Perylene-d12	6740	23.461				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034705.D
Acq On : 24 Oct 2024 23:26
Operator : RC/JU
Sample : PB164282BL
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB164282BL

Quant Time: Oct 25 02:08:37 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.712	152	6454	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	17746	0.400	ng	0.00
13) Acenaphthene-d10	14.329	164	7655	0.400	ng	0.00
19) Phenanthrene-d10	17.069	188	13670	0.400	ng	0.00
29) Chrysene-d12	21.241	240	7475	0.400	ng	# 0.00
35) Perylene-d12	23.461	264	6736	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.322	112	5757	0.300	ng	0.00
5) Phenol-d6	6.882	99	7731	0.306	ng	0.00
8) Nitrobenzene-d5	8.845	82	5072	0.341	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	8137	0.333	ng	0.00
14) 2,4,6-Tribromophenol	15.815	330	504	0.213	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	11147	0.366	ng	0.00
27) Fluoranthene-d10	19.094	212	10547	0.343	ng	0.00
31) Terphenyl-d14	19.698	244	5913	0.387	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034705.D
Acq On : 24 Oct 2024 23:26
Operator : RC/JU
Sample : PB164282BL
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB164282BL

Quant Time: Oct 25 02:08:37 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.712 min Scan# 669
Delta R.T. 0.000 min
Lab File: BN034705.D
Acq: 24 Oct 2024 23:26




Instrument :
BNA_N
ClientSampleId :
PB164282BL


Tgt Ion:152 Resp: 6454
Ion Ratio Lower Upper
152 100
150 151.1 122.7 184.1
115 61.7 51.0 76.6



#4
2-Fluorophenol
Concen: 0.300 ng
RT: 5.322 min Scan# 338
Delta R.T. -0.000 min
Lab File: BN034705.D
Acq: 24 Oct 2024 23:26

Tgt Ion:112 Resp: 5757
Ion Ratio Lower Upper
112 100
64 66.7 53.0 79.6
63 35.6 27.6 41.4









#5
Phenol-d6
Concen: 0.306 ng
RT: 6.882 min Scan# 554
Delta R.T. -0.000 min
Lab File: BN034705.D
Acq: 24 Oct 2024 23:26




Instrument :
BNA_N
ClientSampleId :
PB164282BL


Tgt Ion: 99 Resp: 7731
Ion Ratio Lower Upper
99 100
42 28.2 21.8 32.8
71 34.0 26.6 40.0



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 955
Delta R.T. -0.000 min
Lab File: BN034705.D
Acq: 24 Oct 2024 23:26

Tgt Ion: 136 Resp: 17746
Ion Ratio Lower Upper
136 100
137 11.3 9.1 13.7
54 12.2 9.8 14.8
68 8.0 6.4 9.6









#8
Nitrobenzene-d5
Concen: 0.341 ng
RT: 8.845 min Scan# 802
Delta R.T. -0.000 min
Lab File: BN034705.D
Acq: 24 Oct 2024 23:26




Instrument :
BNA_N
ClientSampleId :
PB164282BL


Tgt Ion:	82	Resp:	5072
Ion	Ratio	Lower	Upper
82	100		
128	37.0	29.7	44.5
54	65.3	53.0	79.6



#11
2-Methylnaphthalene-d10
Concen: 0.333 ng
RT: 12.071 min Scan# 1201
Delta R.T. -0.000 min
Lab File: BN034705.D
Acq: 24 Oct 2024 23:26

Tgt Ion:	152	Resp:	8137
Ion	Ratio	Lower	Upper
152	100		
151	20.8	16.7	25.1









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.329 min Scan# 1504
Delta R.T. -0.000 min
Lab File: BN034705.D
Acq: 24 Oct 2024 23:26




Instrument :
BNA_N
ClientSampleId :
PB164282BL


Tgt Ion:164 Resp: 7655
Ion Ratio Lower Upper
164 100
162 101.7 81.4 122.0
160 51.9 40.6 61.0



#14
2,4,6-Tribromophenol
Concen: 0.213 ng
RT: 15.815 min Scan# 1639
Delta R.T. -0.000 min
Lab File: BN034705.D
Acq: 24 Oct 2024 23:26

Tgt Ion:330 Resp: 504
Ion Ratio Lower Upper
330 100
332 94.6 76.6 115.0
141 68.3 54.6 81.8









#15
2-Fluorobiphenyl
Concen: 0.366 ng
RT: 12.950 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN034705.D
Acq: 24 Oct 2024 23:26




Instrument :
BNA_N
ClientSampleId :
PB164282BL


Tgt Ion:	172	Resp:	11147
Ion Ratio	Lower	Upper	
172	100		
171	35.4	28.3	42.5
170	24.6	19.8	29.6



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.069 min Scan# 1740
Delta R.T. -0.000 min
Lab File: BN034705.D
Acq: 24 Oct 2024 23:26

Tgt Ion:	188	Resp:	13670
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	10.4	8.9	13.3









#27
Fluoranthene-d10
Concen: 0.343 ng
RT: 19.094 min Scan# 1982
Delta R.T. -0.000 min
Lab File: BN034705.D
Acq: 24 Oct 2024 23:26




Instrument :
BNA_N
ClientSampleId :
PB164282BL


Tgt Ion	Ratio	Lower	Upper
212	100		
106	23.6	17.8	26.8
104	13.5	10.2	15.4



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.241 min Scan# 2420
Delta R.T. -0.009 min
Lab File: BN034705.D
Acq: 24 Oct 2024 23:26


Tgt Ion	Ratio	Lower	Upper
240	100		
120	28.0	14.2	21.2#
236	30.9	23.1	34.7








#31
Terphenyl-d14
Concen: 0.387 ng
RT: 19.698 min Scan# 2112
Delta R.T. -0.005 min
Lab File: BN034705.D
Acq: 24 Oct 2024 23:26

Instrument :
BNA_N
ClientSampleId :
PB164282BL






Tgt Ion:244 Resp: 5913
Ion Ratio Lower Upper
244 100
212 12.1 8.6 13.0
122 30.7 22.3 33.5



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.461 min Scan# 2911
Delta R.T. -0.006 min
Lab File: BN034705.D
Acq: 24 Oct 2024 23:26

Tgt Ion:264 Resp: 6736
Ion Ratio Lower Upper
264 100
260 27.7 21.9 32.9
265 93.2 60.6 91.0#



Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	CTO WE13		Date Received:	
Client Sample ID:	PB164282BS		SDG No.:	P4440
Lab Sample ID:	PB164282BS		Matrix:	Water
Analytical Method:	SW8270SIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN034706.D	1	10/21/24 08:45	10/25/24 00:02	PB164282

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.30		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.50		30 - 150		125%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.34		30 - 150		84%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.34		55 - 111		84%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.36		53 - 106		90%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.36		58 - 132		91%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	6280	7.712				
1146-65-2	Naphthalene-d8	17400	10.479				
15067-26-2	Acenaphthene-d10	7560	14.318				
1517-22-2	Phenanthrene-d10	13500	17.056				
1719-03-5	Chrysene-d12	7660	21.241				
1520-96-3	Perylene-d12	6460	23.461				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
 Data File : BN034706.D
 Acq On : 25 Oct 2024 00:02
 Operator : RC/JU
 Sample : PB164282BS
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

PB164282BS

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024

Quant Time: Oct 25 02:08:50 2024

Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Oct 24 13:17:02 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.712	152	6277	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	17435	0.400	ng	0.00
13) Acenaphthene-d10	14.318	164	7557	0.400	ng	-0.01
19) Phenanthrene-d10	17.056	188	13505	0.400	ng	#-0.01
29) Chrysene-d12	21.241	240	7661	0.400	ng	# 0.00
35) Perylene-d12	23.461	264	6455	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.321	112	5611	0.301	ng	0.00
5) Phenol-d6	6.881	99	7614	0.309	ng	0.00
8) Nitrobenzene-d5	8.845	82	4924	0.337	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	11948m	0.498	ng	0.00
14) 2,4,6-Tribromophenol	15.815	330	563	0.241	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	10877	0.362	ng	0.00
27) Fluoranthene-d10	19.094	212	10197	0.335	ng	0.00
31) Terphenyl-d14	19.703	244	5692	0.364	ng	0.00
Target Compounds					Qvalue	
2) 1,4-Dioxane	3.285	88	2475	0.296	ng	# 46
3) n-Nitrosodimethylamine	3.581	42	4401	0.362	ng	96
6) bis(2-Chloroethyl)ether	7.141	93	7177	0.359	ng	99
9) Naphthalene	10.532	128	16863	0.350	ng	100
10) Hexachlorobutadiene	10.831	225	2523	0.349	ng	# 98
12) 2-Methylnaphthalene	12.143	142	10035	0.335	ng	99
16) Acenaphthylene	14.040	152	12822	0.353	ng	100
17) Acenaphthene	14.382	154	8528	0.341	ng	98
18) Fluorene	15.377	166	9962	0.323	ng	99
20) 4,6-Dinitro-2-methylph...	15.441	198	538	0.286	ng	# 56
21) 4-Bromophenyl-phenylether	16.262	248	2561	0.356	ng	# 61
22) Hexachlorobenzene	16.374	284	2810	0.360	ng	97
23) Atrazine	16.535	200	1814	0.321	ng	# 88
24) Pentachlorophenol	16.721	266	1431	0.518	ng	98
25) Phenanthrene	17.106	178	14856	0.366	ng	100
26) Anthracene	17.193	178	13024	0.362	ng	100
28) Fluoranthene	19.122	202	14119	0.334	ng	99
30) Pyrene	19.484	202	14340	0.364	ng	99
32) Benzo(a)anthracene	21.232	228	10619	0.363	ng	99
33) Chrysene	21.277	228	11699	0.386	ng	98
34) Bis(2-ethylhexyl)phtha...	21.187	149	5670	0.341	ng	# 100
36) Indeno(1,2,3-cd)pyrene	25.688	276	10806	0.435	ng	98
37) Benzo(b)fluoranthene	22.797	252	10399	0.403	ng	94
38) Benzo(k)fluoranthene	22.841	252	10151	0.399	ng	96
39) Benzo(a)pyrene	23.364	252	8689	0.420	ng	95
40) Dibenzo(a,h)anthracene	25.706	278	8269	0.422	ng	97
41) Benzo(g,h,i)perylene	26.361	276	8830	0.408	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034706.D
Acq On : 25 Oct 2024 00:02
Operator : RC/JU
Sample : PB164282BS
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

PB164282BS


Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024


Quant Time: Oct 25 02:08:50 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.712 min Scan# 669
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS



Tgt Ion: 152 Resp: 627
Ion Ratio Lower Upper
152 100
150 153.7 122.7 184.1
115 62.4 51.0 76.6




Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024




#2
1,4-Dioxane
Concen: 0.296 ng
RT: 3.285 min Scan# 56
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02


Tgt Ion: 88 Resp: 2475
Ion Ratio Lower Upper
88 100
43 124.6 32.1 48.1#
58 109.9 73.7 110.5




Abundance Scan 97 (3.581 min): BN034686.D\data.ms (-93)



Abundance Scan 97 (3.581 min): BN034706.D\data.ms



Abundance Scan 97 (3.581 min): BN034706.D\data.ms (-76)



#3

n-Nitrosodimethylamine

Concen: 0.362 ng

RT: 3.581 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN034706.D

Acq: 25 Oct 2024 00:02

Tgt Ion: 42 Resp: 440

Ion Ratio Lower Upper

42 100

74 92.1 77.2 115.8

44 14.7 13.4 20.2

Instrument :

BNA_N

ClientSampleId :

PB164282BS


Manual Integrations

APPROVED


Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024


Abundance




Abundance Scan 338 (5.322 min): BN034686.D\data.ms (-33)



Abundance Scan 338 (5.321 min): BN034706.D\data.ms



Abundance Scan 338 (5.321 min): BN034706.D\data.ms (-31)



#4

2-Fluorophenol

Concen: 0.301 ng

RT: 5.321 min Scan# 338

Delta R.T. -0.000 min

Lab File: BN034706.D

Acq: 25 Oct 2024 00:02

Tgt Ion:112 Resp: 5611


Ion Ratio Lower Upper


112 100

64 66.5 53.0 79.6

63 35.5 27.6 41.4


Abundance





#5
Phenol-d6
Concen: 0.309 ng
RT: 6.881 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS



Tgt Ion: 99 Resp: 7614
Ion Ratio Lower Upper
99 100
42 28.9 21.8 32.8
71 33.7 26.6 40.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#6
bis(2-Chloroethyl)ether
Concen: 0.359 ng
RT: 7.141 min Scan# 590
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion: 93 Resp: 7177
Ion Ratio Lower Upper
93 100
63 90.6 71.5 107.3
95 31.9 25.8 38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Instrument :
BNA_N
ClientSampleId :
PB164282BS

Tgt Ion:	136	Resp:	1743
Ion	Ratio	Lower	Upper
136	100		
137	11.4	9.1	13.7
54	11.4	9.8	14.8
68	7.4	6.4	9.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#8
Nitrobenzene-d5
Concen: 0.337 ng
RT: 8.845 min Scan# 802
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion:	82	Resp:	4924
Ion	Ratio	Lower	Upper
82	100		
128	36.9	29.7	44.5
54	64.3	53.0	79.6









#9
Naphthalene
Concen: 0.350 ng
RT: 10.532 min Scan# 960
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Instrument :
BNA_N
ClientSampleId :
PB164282BS

Tgt Ion:128 Resp: 1686
Ion Ratio Lower Upper
128 100
129 11.4 9.0 13.6
127 13.1 10.5 15.7




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#10
Hexachlorobutadiene
Concen: 0.349 ng
RT: 10.831 min Scan# 988
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion:225 Resp: 2523
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.0 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 0.498 ng m
RT: 12.071 min Scan# 1194
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Instrument :
BNA_N
ClientSampleId :
PB164282BS

Tgt Ion:152 Resp: 1194
Ion Ratio Lower Upper
152 100
151 14.0 16.7 25.1




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#12
2-Methylnaphthalene
Concen: 0.335 ng
RT: 12.143 min Scan# 1220
Delta R.T. -0.004 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02


Tgt Ion:142 Resp: 10035
Ion Ratio Lower Upper
142 100
141 88.4 70.4 105.6
115 38.2 29.8 44.6





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.318 min Scan# 1503
Delta R.T. -0.011 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS



Tgt Ion:164 Resp: 755
Ion Ratio Lower Upper
164 100
162 106.1 81.4 122.0
160 55.4 40.6 61.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#14
2,4,6-Tribromophenol
Concen: 0.241 ng
RT: 15.815 min Scan# 1639
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion:330 Resp: 563
Ion Ratio Lower Upper
330 100
332 94.7 76.6 115.0
141 73.7 54.6 81.8









#15
2-Fluorobiphenyl
Concen: 0.362 ng
RT: 12.950 min Scan# 1375
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Instrument :
BNA_N
ClientSampleId :
PB164282BS

Tgt Ion	Ratio	Lower	Upper
172	100		
171	34.6	28.3	42.5
170	23.9	19.8	29.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#16
Acenaphthylene
Concen: 0.353 ng
RT: 14.040 min Scan# 1477
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02


Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.6	23.4
153	13.1	10.6	15.8





#17
Acenaphthene
Concen: 0.341 ng
RT: 14.382 min Scan# 1509
Delta R.T. -0.011 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS



Tgt Ion:154 Resp: 8528
Ion Ratio Lower Upper
154 100
153 112.4 88.2 132.2
152 60.6 46.9 70.3




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#18
Fluorene
Concen: 0.323 ng
RT: 15.377 min Scan# 1602
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion:166 Resp: 9962
Ion Ratio Lower Upper
166 100
165 97.3 78.8 118.2
167 13.6 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.056 min Scan# 11

Delta R.T. -0.013 min

Lab File: BN034706.D

Acq: 25 Oct 2024 00:02

Instrument :

BNA_N

ClientSampleId :





PB164282BS

Tgt Ion:	188	Resp:	1350
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	22.2	8.9	13.3

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#20

4,6-Dinitro-2-methylphenol

Concen: 0.286 ng




RT: 15.441 min Scan# 1608


Delta R.T. -0.011 min

Lab File: BN034706.D

Acq: 25 Oct 2024 00:02

Tgt Ion:	198	Resp:	538
Ion Ratio	Lower	Upper	
198	100		
51	160.2	91.8	137.8#
105	91.1	44.3	66.5#






#21
4-Bromophenyl-phenylether
Concen: 0.356 ng
RT: 16.262 min Scan# 1675
Delta R.T. -0.013 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Instrument :

BNA_N

ClientSampleId :




PB164282BS




Tgt Ion:248 Resp: 256
Ion Ratio Lower Upper
248 100
250 88.5 80.9 121.3
141 113.9 40.5 60.7

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024




#22
Hexachlorobenzene
Concen: 0.360 ng
RT: 16.374 min Scan# 1684
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Tgt Ion:284 Resp: 2810
Ion Ratio Lower Upper
284 100
142 57.3 43.6 65.4
249 34.1 26.6 40.0





#23
Atrazine
Concen: 0.321 ng
RT: 16.535 min Scan# 1697
Delta R.T. -0.013 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS



Tgt Ion:200 Resp: 1814
Ion Ratio Lower Upper
200 100
173 40.1 19.7 29.5
215 43.2 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#24
Pentachlorophenol
Concen: 0.518 ng
RT: 16.721 min Scan# 1712
Delta R.T. 0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02


Tgt Ion:266 Resp: 1431
Ion Ratio Lower Upper
266 100
264 63.0 49.4 74.2
268 60.7 50.5 75.7





#25
Phenanthrene
Concen: 0.366 ng
RT: 17.106 min Scan# 1743
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS



Tgt Ion:178 Resp: 14850
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.2
179 15.3 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#26
Anthracene
Concen: 0.362 ng
RT: 17.193 min Scan# 1750
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02


Tgt Ion:178 Resp: 13024
Ion Ratio Lower Upper
178 100
176 18.8 14.9 22.3
179 15.3 12.2 18.2





#27
Fluoranthene-d10
Concen: 0.335 ng
RT: 19.094 min Scan# 1982
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS




Tgt Ion:212 Resp: 1019
Ion Ratio Lower Upper
212 100
106 23.4 17.8 26.8
104 13.7 10.2 15.4

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024




#28
Fluoranthene
Concen: 0.334 ng
RT: 19.122 min Scan# 1988
Delta R.T. -0.005 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02



Tgt Ion:202 Resp: 14119
Ion Ratio Lower Upper
202 100
101 16.4 12.8 19.2
203 17.0 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.241 min Scan# 2421

Delta R.T. -0.009 min

Lab File: BN034706.D

Acq: 25 Oct 2024 00:02

Instrument :

BNA_N

ClientSampleId :

PB164282BS





Tgt Ion:	240	Resp:	766
Ion Ratio	Lower	Upper	
240	100		
120	28.4	14.2	21.2
236	30.8	23.1	34.7

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#30

Pyrene

Concen: 0.364 ng




RT: 19.484 min Scan# 2066


Delta R.T. -0.000 min

Lab File: BN034706.D

Acq: 25 Oct 2024 00:02


Tgt Ion:	202	Resp:	14340
Ion Ratio	Lower	Upper	
202	100		
200	21.0	16.6	25.0
203	17.5	14.2	21.4





#31
Terphenyl-d14
Concen: 0.364 ng
RT: 19.703 min Scan# 2113
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS



Tgt Ion:244 Resp: 5692
Ion Ratio Lower Upper
244 100
212 11.7 8.6 13.0
122 27.9 22.3 33.5




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#32
Benzo(a)anthracene
Concen: 0.363 ng
RT: 21.232 min Scan# 2419
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion:228 Resp: 10619
Ion Ratio Lower Upper
228 100
226 27.1 22.0 33.0
229 20.2 15.6 23.4





#33

Chrysene

Concen: 0.386 ng

RT: 21.277 min Scan# 2425

Delta R.T. -0.009 min

Lab File: BN034706.D

Acq: 25 Oct 2024 00:02

Instrument :

BNA_N

ClientSampleId :

PB164282BS

Tgt Ion:228 Resp: 1169

Ion Ratio Lower Upper

228 100

226 31.3 24.1 36.1





229 19.3 15.8 23.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.341 ng

RT: 21.187 min Scan# 2414

Delta R.T. -0.009 min

Lab File: BN034706.D

Acq: 25 Oct 2024 00:02




Tgt Ion:149 Resp: 5670


Ion Ratio Lower Upper

149 100

167 24.5 19.7 29.5

279 1.7 1.8 2.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.461 min Scan# 2911

Delta R.T. -0.006 min

Lab File: BN034706.D


Acq: 25 Oct 2024 00:02

Instrument :

BNA_N

ClientSampleId :

PB164282BS



Tgt Ion:264 Resp: 645

Ion Ratio Lower Upper

264 100

260 28.1 21.9 32.9




265 95.5 60.6 91.0

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.435 ng

RT: 25.688 min Scan# 3673

Delta R.T. -0.006 min

Lab File: BN034706.D

Acq: 25 Oct 2024 00:02




Tgt Ion:276 Resp: 10806


Ion Ratio Lower Upper

276 100

138 46.0 35.1 52.7

277 24.3 19.7 29.5





#37

Benzo(b)fluoranthene

Concen: 0.403 ng

RT: 22.797 min Scan# 2684

Delta R.T. -0.006 min

Lab File: BN034706.D

Acq: 25 Oct 2024 00:02

Instrument :

BNA_N

ClientSampleId :

PB164282BS





Tgt Ion	252	253	125
Resp:	1039		
Ratio	100	26.8	30.7
Lower		20.0	20.9
Upper		30.0	31.3

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#38

Benzo(k)fluoranthene

Concen: 0.399 ng




RT: 22.841 min Scan# 2699


Delta R.T. -0.009 min

Lab File: BN034706.D

Acq: 25 Oct 2024 00:02

Tgt Ion	252	253	125
Resp:	10151		
Ratio	100	26.4	30.7
Lower		20.6	21.6
Upper		31.0	32.4





#39

Benzo(a)pyrene

Concen: 0.420 ng

RT: 23.364 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN034706.D

Acq: 25 Oct 2024 00:02

Instrument :

BNA_N

ClientSampleId :

PB164282BS





Tgt Ion:	252	Resp:	868
Ion Ratio	Lower	Upper	
252	100		
253	28.5	22.2	33.4
125	37.5	26.0	39.0

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#40

Dibenzo(a,h)anthracene

Concen: 0.422 ng




RT: 25.706 min Scan# 3679


Delta R.T. -0.009 min

Lab File: BN034706.D

Acq: 25 Oct 2024 00:02


Tgt Ion:	278	Resp:	8269
Ion Ratio	Lower	Upper	
278	100		
139	34.5	25.4	38.2
279	27.9	22.4	33.6





#41
Benzo(g,h,i)perylene
Concen: 0.408 ng
RT: 26.361 min Scan# 3903
Delta R.T. -0.012 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02



Instrument :
BNA_N
ClientSampleId :
PB164282BS



Tgt Ion:276 Resp: 8830
Ion Ratio Lower Upper
276 100
277 27.0 21.3 31.9
138 42.2 31.9 47.9

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:		
Project:	CTO WE13		Date Received:		
Client Sample ID:	PB164282BSD		SDG No.:	P4440	
Lab Sample ID:	PB164282BSD		Matrix:	Water	
Analytical Method:	SW8270SIM		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN034711.D	1	10/21/24 08:45	10/25/24 03:02	PB164282

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.31		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.50		30 - 150		125%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.33		30 - 150		83%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.34		55 - 111		84%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.36		53 - 106		89%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.37		58 - 132		93%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	5970	7.712				
1146-65-2	Naphthalene-d8	16600	10.479				
15067-26-2	Acenaphthene-d10	7140	14.329				
1517-22-2	Phenanthrene-d10	13000	17.057				
1719-03-5	Chrysene-d12	7030	21.241				
1520-96-3	Perylene-d12	5870	23.458				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
 Data File : BN034711.D
 Acq On : 25 Oct 2024 03:02
 Operator : RC/JU
 Sample : PB164282BSD
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

PB164282BSD

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024

Quant Time: Oct 25 04:20:43 2024

Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Thu Oct 24 13:17:02 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.712	152	5974	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	16564	0.400	ng	0.00
13) Acenaphthene-d10	14.329	164	7143	0.400	ng	0.00
19) Phenanthrene-d10	17.057	188	13002	0.400	ng	#-0.01
29) Chrysene-d12	21.241	240	7027	0.400	ng	# 0.00
35) Perylene-d12	23.458	264	5873	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.322	112	5283	0.297	ng	0.00
5) Phenol-d6	6.882	99	7080	0.302	ng	0.00
8) Nitrobenzene-d5	8.845	82	4660	0.335	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	11354m	0.499	ng	0.00
14) 2,4,6-Tribromophenol	15.815	330	521	0.236	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	10106	0.356	ng	0.00
27) Fluoranthene-d10	19.094	212	9709	0.332	ng	0.00
31) Terphenyl-d14	19.703	244	5323	0.371	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.285	88	2435	0.306	ng	# 53
3) n-Nitrosodimethylamine	3.581	42	4293	0.371	ng	96
6) bis(2-Chloroethyl)ether	7.142	93	6795	0.357	ng	99
9) Naphthalene	10.532	128	15953	0.348	ng	99
10) Hexachlorobutadiene	10.831	225	2383	0.347	ng	# 97
12) 2-Methylnaphthalene	12.143	142	9409	0.331	ng	98
16) Acenaphthylene	14.040	152	12085	0.352	ng	99
17) Acenaphthene	14.382	154	7997	0.338	ng	98
18) Fluorene	15.377	166	9417	0.323	ng	100
20) 4,6-Dinitro-2-methylph...	15.441	198	532	0.294	ng	# 60
21) 4-Bromophenyl-phenylether	16.262	248	2358	0.340	ng	# 62
22) Hexachlorobenzene	16.374	284	2677	0.356	ng	98
23) Atrazine	16.535	200	1782	0.327	ng	# 89
24) Pentachlorophenol	16.721	266	1302	0.490	ng	99
25) Phenanthrene	17.106	178	14207	0.364	ng	100
26) Anthracene	17.193	178	12435	0.359	ng	100
28) Fluoranthene	19.122	202	13425	0.329	ng	100
30) Pyrene	19.484	202	13610	0.377	ng	99
32) Benzo(a)anthracene	21.232	228	9893	0.369	ng	99
33) Chrysene	21.277	228	11013	0.396	ng	98
34) Bis(2-ethylhexyl)phtha...	21.187	149	5309	0.349	ng	# 96
36) Indeno(1,2,3-cd)pyrene	25.683	276	9709	0.429	ng	96
37) Benzo(b)fluoranthene	22.797	252	9630	0.411	ng	94
38) Benzo(k)fluoranthene	22.841	252	9277m	0.401	ng	
39) Benzo(a)pyrene	23.367	252	8133	0.432	ng	# 93
40) Dibenzo(a,h)anthracene	25.703	278	7436	0.417	ng	96
41) Benzo(g,h,i)perylene	26.364	276	7954	0.404	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
BNA_N
ClientSampleId :
PB164282BSD

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.712 min Scan# 60
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD



Tgt Ion: 152 Resp: 5974
Ion Ratio Lower Upper
152 100
150 152.2 122.7 184.1
115 63.6 51.0 76.6




Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024




#2
1,4-Dioxane
Concen: 0.306 ng
RT: 3.285 min Scan# 56
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02


Tgt Ion: 88 Resp: 2435
Ion Ratio Lower Upper
88 100
43 114.0 32.1 48.1#
58 107.2 73.7 110.5




Abundance Scan 97 (3.581 min): BN034686.D\data.ms (-93)



Abundance Scan 97 (3.581 min): BN034711.D\data.ms



Abundance Scan 97 (3.581 min): BN034711.D\data.ms (-76)



#3

n-Nitrosodimethylamine

Concen: 0.371 ng

RT: 3.581 min Scan# 91

Delta R.T. 0.000 min

Lab File: BN034711.D

Acq: 25 Oct 2024 03:02

Tgt Ion: 42 Resp: 429

Ion Ratio Lower Upper

42 100

74 92.4 77.2 115.8

44 15.0 13.4 20.2

Instrument :

BNA_N

ClientSampleId :


PB164282BSD

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 10/25/2024


Supervised By :mohammad ahmed 10/26/2024




Abundance Scan 338 (5.322 min): BN034686.D\data.ms (-33)



Abundance Scan 338 (5.322 min): BN034711.D\data.ms



Abundance Scan 338 (5.322 min): BN034711.D\data.ms (-31)



#4

2-Fluorophenol

Concen: 0.297 ng

RT: 5.322 min Scan# 338

Delta R.T. 0.000 min

Lab File: BN034711.D

Acq: 25 Oct 2024 03:02


Tgt Ion:112 Resp: 5283


Ion Ratio Lower Upper

112 100

64 67.3 53.0 79.6

63 35.9 27.6 41.4









#5
Phenol-d6
Concen: 0.302 ng
RT: 6.882 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Instrument :
BNA_N
ClientSampleId :
PB164282BSD

Tgt Ion: 99 Resp: 7080
Ion Ratio Lower Upper
99 100
42 30.9 21.8 32.8
71 33.8 26.6 40.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#6
bis(2-Chloroethyl)ether
Concen: 0.357 ng
RT: 7.142 min Scan# 590
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02


Tgt Ion: 93 Resp: 6795
Ion Ratio Lower Upper
93 100
63 90.5 71.5 107.3
95 32.1 25.8 38.6





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD



Tgt Ion: 136 Resp: 1656
Ion Ratio Lower Upper
136 100
137 11.6 9.1 13.7
54 11.1 9.8 14.8
68 7.5 6.4 9.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#8
Nitrobenzene-d5
Concen: 0.335 ng
RT: 8.845 min Scan# 802
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Tgt Ion: 82 Resp: 4660
Ion Ratio Lower Upper
82 100
128 37.7 29.7 44.5
54 64.1 53.0 79.6









#9
Naphthalene
Concen: 0.348 ng
RT: 10.532 min Scan# 960
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Instrument :
BNA_N
ClientSampleId :
PB164282BSD

Tgt Ion:128 Resp: 1595
Ion Ratio Lower Upper
128 100
129 11.4 9.0 13.6
127 13.4 10.5 15.7




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#10
Hexachlorobutadiene
Concen: 0.347 ng
RT: 10.831 min Scan# 988
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02


Tgt Ion:225 Resp: 2383
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 62.1 51.4 77.0





#11
2-Methylnaphthalene-d10
Concen: 0.499 ng m
RT: 12.071 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD




Tgt Ion:152 Resp: 11354
Ion Ratio Lower Upper
152 100
151 14.1 16.7 25.1

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024




#12
2-Methylnaphthalene
Concen: 0.331 ng
RT: 12.143 min Scan# 1220
Delta R.T. -0.004 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Tgt Ion:142 Resp: 9409
Ion Ratio Lower Upper
142 100
141 87.4 70.4 105.6
115 39.3 29.8 44.6





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.329 min Scan# 1504
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD




Tgt Ion:164 Resp: 714
Ion Ratio Lower Upper
164 100
162 100.9 81.4 122.0
160 51.4 40.6 61.0

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024




#14
2,4,6-Tribromophenol
Concen: 0.236 ng
RT: 15.815 min Scan# 1639
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02



Tgt Ion:330 Resp: 521
Ion Ratio Lower Upper
330 100
332 93.5 76.6 115.0
141 77.7 54.6 81.8









#15
2-Fluorobiphenyl
Concen: 0.356 ng
RT: 12.950 min Scan# 1375
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Instrument :
BNA_N
ClientSampleId :
PB164282BSD

Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.5	28.3	42.5
170	24.7	19.8	29.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#16
Acenaphthylene
Concen: 0.352 ng
RT: 14.040 min Scan# 1477
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02


Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.6	23.4
153	12.9	10.6	15.8





#17
Acenaphthene
Concen: 0.338 ng
RT: 14.382 min Scan# 1510
Delta R.T. -0.011 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD



Tgt Ion:154 Resp: 799
Ion Ratio Lower Upper
154 100
153 111.9 88.2 132.2
152 60.7 46.9 70.3




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#18
Fluorene
Concen: 0.323 ng
RT: 15.377 min Scan# 1602
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Tgt Ion:166 Resp: 9417
Ion Ratio Lower Upper
166 100
165 98.4 78.8 118.2
167 13.2 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.057 min Scan# 11

Delta R.T. -0.012 min

Lab File: BN034711.D

Acq: 25 Oct 2024 03:02





Instrument :

BNA_N

ClientSampleId :

PB164282BSD

Tgt Ion:	188	Resp:	1300
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	23.0	8.9	13.3

Manual Integrations
APPROVEDReviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024

#20

4,6-Dinitro-2-methylphenol

Concen: 0.294 ng




RT: 15.441 min Scan# 1608


Delta R.T. -0.011 min

Lab File: BN034711.D

Acq: 25 Oct 2024 03:02


Tgt Ion:	198	Resp:	532
Ion Ratio	Lower	Upper	
198	100		
51	156.1	91.8	137.8#
105	88.2	44.3	66.5#





#21
4-Bromophenyl-phenylether
Concen: 0.340 ng
RT: 16.262 min Scan# 1675
Delta R.T. -0.012 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD



Tgt Ion:248 Resp: 235
Ion Ratio Lower Upper
248 100
250 91.6 80.9 121.3
141 116.5 40.5 60.7




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#22
Hexachlorobenzene
Concen: 0.356 ng
RT: 16.374 min Scan# 1684
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02


Tgt Ion:284 Resp: 2677
Ion Ratio Lower Upper
284 100
142 56.9 43.6 65.4
249 33.8 26.6 40.0





#23
Atrazine
Concen: 0.327 ng
RT: 16.535 min Scan# 1697
Delta R.T. -0.012 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD



Tgt Ion:200 Resp: 178
Ion Ratio Lower Upper
200 100
173 36.2 19.7 29.5
215 41.4 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#24
Pentachlorophenol
Concen: 0.490 ng
RT: 16.721 min Scan# 1712
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02


Tgt Ion:266 Resp: 1302
Ion Ratio Lower Upper
266 100
264 62.4 49.4 74.2
268 62.1 50.5 75.7





#25
Phenanthrene
Concen: 0.364 ng
RT: 17.106 min Scan# 1743
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD



Tgt Ion:178 Resp: 1420
Ion Ratio Lower Upper
178 100
176 19.5 15.4 23.2
179 15.4 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#26
Anthracene
Concen: 0.359 ng
RT: 17.193 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02


Tgt Ion:178 Resp: 12435
Ion Ratio Lower Upper
178 100
176 18.5 14.9 22.3
179 15.3 12.2 18.2





#27
Fluoranthene-d10
Concen: 0.332 ng
RT: 19.094 min Scan# 1982
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD




Tgt Ion:212 Resp: 9709
Ion Ratio Lower Upper
212 100
106 23.6 17.8 26.8
104 13.8 10.2 15.4

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024




#28
Fluoranthene
Concen: 0.329 ng
RT: 19.122 min Scan# 1988
Delta R.T. -0.005 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02



Tgt Ion:202 Resp: 13425
Ion Ratio Lower Upper
202 100
101 16.0 12.8 19.2
203 17.1 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.241 min Scan# 2421

Delta R.T. -0.009 min

Lab File: BN034711.D

Acq: 25 Oct 2024 03:02

Instrument :

BNA_N

ClientSampleId :





PB164282BSD

Tgt Ion:	240	Resp:	702
Ion Ratio	Lower	Upper	
240	100		
120	30.8	14.2	21.2
236	31.0	23.1	34.7

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#30

Pyrene

Concen: 0.377 ng




RT: 19.484 min Scan# 2066


Delta R.T. 0.000 min

Lab File: BN034711.D

Acq: 25 Oct 2024 03:02


Tgt Ion:	202	Resp:	13610
Ion Ratio	Lower	Upper	
202	100		
200	21.0	16.6	25.0
203	17.5	14.2	21.4





#31
Terphenyl-d14
Concen: 0.371 ng
RT: 19.703 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD



Tgt Ion:244 Resp: 532
Ion Ratio Lower Upper
244 100
212 11.8 8.6 13.0
122 28.3 22.3 33.5




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#32
Benzo(a)anthracene
Concen: 0.369 ng
RT: 21.232 min Scan# 2419
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02


Tgt Ion:228 Resp: 9893
Ion Ratio Lower Upper
228 100
226 27.4 22.0 33.0
229 20.1 15.6 23.4





#33
Chrysene
Concen: 0.396 ng
RT: 21.277 min Scan# 2424
Delta R.T. -0.009 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD



Tgt Ion:228 Resp: 1101
Ion Ratio Lower Upper
228 100
226 31.8 24.1 36.1
229 19.5 15.8 23.8




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.349 ng
RT: 21.187 min Scan# 2414
Delta R.T. -0.009 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Tgt Ion:149 Resp: 5309
Ion Ratio Lower Upper
149 100
167 22.6 19.7 29.5
279 1.6 1.8 2.6#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.458 min Scan# 2910

Delta R.T. -0.009 min

Lab File: BN034711.D


Acq: 25 Oct 2024 03:02

Instrument :

BNA_N

ClientSampleId :

PB164282BSD






Tgt Ion:	264	Resp:	587
Ion Ratio	Lower	Upper	
264	100		
260	28.4	21.9	32.9
265	103.3	60.6	91.0

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#36

Indeno(1,2,3-cd)pyrene


Concen: 0.429 ng

RT: 25.683 min Scan# 3671



Delta R.T. -0.012 min


Lab File: BN034711.D

Acq: 25 Oct 2024 03:02



Tgt Ion:	276	Resp:	9709
Ion Ratio	Lower	Upper	
276	100		
138	47.6	35.1	52.7
277	24.0	19.7	29.5





#37

Benzo(b)fluoranthene

Concen: 0.411 ng

RT: 22.797 min Scan# 2684

Delta R.T. -0.006 min

Lab File: BN034711.D

Acq: 25 Oct 2024 03:02

Instrument :

BNA_N

ClientSampleId :

PB164282BSD





Tgt Ion:252	Resp:	9630
Ion Ratio	Lower	Upper
252	100	
253	26.4	20.0 30.0
125	31.1	20.9 31.3

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#38

Benzo(k)fluoranthene

Concen: 0.401 ng m




RT: 22.841 min Scan# 2699


Delta R.T. -0.009 min

Lab File: BN034711.D

Acq: 25 Oct 2024 03:02

Tgt Ion:252	Resp:	9277
Ion Ratio	Lower	Upper
252	100	
253	27.3	20.6 31.0
125	32.7	21.6 32.4





#39

Benzo(a)pyrene

Concen: 0.432 ng

RT: 23.367 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN034711.D

Acq: 25 Oct 2024 03:02

Instrument :

BNA_N

ClientSampleId :





PB164282BSD

Tgt Ion:	252	Resp:	813
Ion Ratio	Lower	Upper	
252	100		
253	28.6	22.2	33.4
125	39.1	26.0	39.0

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024

Supervised By :mohammad ahmed 10/26/2024



#40

Dibenzo(a,h)anthracene

Concen: 0.417 ng




RT: 25.703 min Scan# 3678


Delta R.T. -0.012 min

Lab File: BN034711.D

Acq: 25 Oct 2024 03:02


Tgt Ion:	278	Resp:	7436
Ion Ratio	Lower	Upper	
278	100		
139	35.2	25.4	38.2
279	29.0	22.4	33.6





#41
Benzo(g,h,i)perylene
Concen: 0.404 ng
RT: 26.364 min Scan# 3904
Delta R.T. -0.009 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02



Instrument :
BNA_N
ClientSampleId :
PB164282BSD



Tgt Ion:276 Resp: 7954
Ion Ratio Lower Upper
276 100
277 26.7 21.3 31.9
138 43.3 31.9 47.9

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/25/2024
Supervised By :mohammad ahmed 10/26/2024



Manual Integration Report

Sequence:	BN102424	Instrument	BNA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICC0.8	BN034687.D	Benzo(b)fluoranthene	yogesh	10/25/2024 11:48:33 PM	mohammad	10/26/2024 4:13:48 AM	Peak Integrated by Software
SSTDICV0.4	BN034691.D	Bis(2-ethylhexyl)phthalate	yogesh	10/25/2024 11:48:35 PM	mohammad	10/26/2024 4:13:48 AM	Peak Integrated by Software
PB164282BS	BN034706.D	2-Methylnaphthalene-d10	yogesh	10/25/2024 11:48:47 PM	mohammad	10/26/2024 4:13:48 AM	Peak Integrated by Software
PB164282BSD	BN034711.D	2-Methylnaphthalene-d10	yogesh	10/25/2024 11:49:03 PM	mohammad	10/26/2024 4:13:48 AM	Peak Integrated by Software
PB164282BSD	BN034711.D	Benzo(k)fluoranthene	yogesh	10/25/2024 11:49:03 PM	mohammad	10/26/2024 4:13:48 AM	Peak Integrated by Software



Manual Integration Report

Sequence:	BN102524	Instrument	BNA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN102424

Review By	yogesh	Review On	10/25/2024 11:49:23 PM		
Supervise By	mohammad	Supervise On	10/26/2024 4:13:48 AM		
SubDirectory	BN102424	HP Acquire Method	BNA_N	HP Processing Method	bn102424
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597			
CCC		SP6601			
Internal Standard/PEM		SP6527			
ICV/I.BLK		SP6548			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN034683.D	24 Oct 2024 07:59	RC/JU	Ok
2	SSTDICC0.1	BN034684.D	24 Oct 2024 09:11	RC/JU	Ok
3	SSTDICC0.2	BN034685.D	24 Oct 2024 09:47	RC/JU	Ok
4	SSTDICCC0.4	BN034686.D	24 Oct 2024 10:23	RC/JU	Ok
5	SSTDICC0.8	BN034687.D	24 Oct 2024 10:59	RC/JU	Ok,M
6	SSTDICC1.6	BN034688.D	24 Oct 2024 11:35	RC/JU	Ok
7	SSTDICC3.2	BN034689.D	24 Oct 2024 12:11	RC/JU	Ok
8	SSTDICC5.0	BN034690.D	24 Oct 2024 12:48	RC/JU	Ok
9	SSTDICV0.4	BN034691.D	24 Oct 2024 14:14	RC/JU	Ok,M
10	PB164229BL	BN034692.D	24 Oct 2024 14:50	RC/JU	Ok
11	P4428-01	BN034693.D	24 Oct 2024 15:28	RC/JU	Ok
12	PB164229BS	BN034694.D	24 Oct 2024 16:05	RC/JU	Ok
13	P4428-03	BN034695.D	24 Oct 2024 16:41	RC/JU	Ok
14	P4428-04	BN034696.D	24 Oct 2024 17:18	RC/JU	Ok
15	P4475-04	BN034697.D	24 Oct 2024 17:54	RC/JU	Ok
16	P4475-06	BN034698.D	24 Oct 2024 18:30	RC/JU	Ok
17	SSTDCCC0.4	BN034699.D	24 Oct 2024 19:06	RC/JU	Ok
18	DFTPP	BN034700.D	24 Oct 2024 20:22	RC/JU	Ok
19	SSTDCCC0.4	BN034701.D	24 Oct 2024 21:01	RC/JU	Ok
20	PB164230BL	BN034702.D	24 Oct 2024 21:38	RC/JU	Ok
21	PB164323BL	BN034703.D	24 Oct 2024 22:14	RC/JU	Ok

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN102424

Review By	yogesh	Review On	10/25/2024 11:49:23 PM
Supervise By	mohammad	Supervise On	10/26/2024 4:13:48 AM
SubDirectory	BN102424	HP Acquire Method	BNA_N
		HP Processing Method	bn102424
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597		
CCC	SP6601		
Internal Standard/PEM	SP6527		
ICV/I.BLK	SP6548		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PB164229BSD	BN034704.D	24 Oct 2024 22:50	RC/JU	Ok
23	PB164282BL	BN034705.D	24 Oct 2024 23:26	RC/JU	Ok
24	PB164282BS	BN034706.D	25 Oct 2024 00:02	RC/JU	Ok,M
25	PB164230BS	BN034707.D	25 Oct 2024 00:38	RC/JU	Ok,M
26	PB164323BS	BN034708.D	25 Oct 2024 01:14	RC/JU	Ok,M
27	PB164323BSD	BN034709.D	25 Oct 2024 01:50	RC/JU	Ok,M
28	P4428-02	BN034710.D	25 Oct 2024 02:26	RC/JU	Ok
29	PB164282BSD	BN034711.D	25 Oct 2024 03:02	RC/JU	Ok,M
30	P4438-03	BN034712.D	25 Oct 2024 03:38	RC/JU	ReRun
31	P4438-04	BN034713.D	25 Oct 2024 04:14	RC/JU	Ok
32	P4438-05	BN034714.D	25 Oct 2024 04:51	RC/JU	Ok
33	P4438-06	BN034715.D	25 Oct 2024 05:27	RC/JU	Ok
34	P4438-07	BN034716.D	25 Oct 2024 06:03	RC/JU	Ok
35	P4440-01	BN034717.D	25 Oct 2024 06:39	RC/JU	ReRun
36	P4432-01	BN034718.D	25 Oct 2024 07:15	RC/JU	Not Ok
37	SSTDCCC0.4	BN034719.D	25 Oct 2024 07:51	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN102524

Review By	yogesh	Review On	10/26/2024 12:01:15 AM		
Supervise By	mohammad	Supervise On	10/26/2024 4:14:54 AM		
SubDirectory	BN102524	HP Acquire Method	BNA_N	HP Processing Method	bn102424
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597			
CCC		SP6601			
Internal Standard/PEM		SP6527			
ICV/I.BLK		SP6548			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN034720.D	25 Oct 2024 09:07	RC/JU	Ok
2	SSTDCCC0.4	BN034721.D	25 Oct 2024 09:46	RC/JU	Ok
3	PB164323BL	BN034722.D	25 Oct 2024 10:22	RC/JU	Ok
4	P4438-03RE	BN034723.D	25 Oct 2024 10:58	RC/JU	Confirms
5	P4440-01RE	BN034724.D	25 Oct 2024 11:34	RC/JU	Confirms
6	PB164409BL	BN034725.D	25 Oct 2024 16:28	RC/JU	Not Ok
7	PB164409BL	BN034726.D	25 Oct 2024 17:04	RC/JU	Ok
8	PB164409BS	BN034727.D	25 Oct 2024 17:40	RC/JU	Ok,M
9	PB164409BSD	BN034728.D	25 Oct 2024 18:16	RC/JU	Ok,M
10	P4550-02	BN034729.D	25 Oct 2024 18:53	RC/JU	ReRun
11	P4550-05	BN034730.D	25 Oct 2024 19:29	RC/JU	Ok,M
12	P4550-09	BN034731.D	25 Oct 2024 20:05	RC/JU	Ok
13	SSTDCCC0.4	BN034732.D	25 Oct 2024 20:41	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN102424

Review By	yogesh	Review On	10/25/2024 11:49:23 PM
Supervise By	mohammad	Supervise On	10/26/2024 4:13:48 AM
SubDirectory	BN102424	HP Acquire Method	BNA_N
		HP Processing Method	bn102424

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597
CCC	SP6601
Internal Standard/PEM	SP6527
ICV/I.BLK	SP6548
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN034683.D	24 Oct 2024 07:59		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN034684.D	24 Oct 2024 09:11	Compound#20,24 removed from 0.1 ppm	RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN034685.D	24 Oct 2024 09:47		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN034686.D	24 Oct 2024 10:23	Calibration Good for DOD & Non-DOD both	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN034687.D	24 Oct 2024 10:59		RC/JU	Ok,M
6	SSTDICC1.6	SSTDICC1.6	BN034688.D	24 Oct 2024 11:35		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN034689.D	24 Oct 2024 12:11		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN034690.D	24 Oct 2024 12:48		RC/JU	Ok
9	SSTDICV0.4	ICVBN102424	BN034691.D	24 Oct 2024 14:14		RC/JU	Ok,M
10	PB164229BL	PB164229BL	BN034692.D	24 Oct 2024 14:50	USE For Only P4428	RC/JU	Ok
11	P4428-01	RW7-SP100-20241016	BN034693.D	24 Oct 2024 15:28		RC/JU	Ok
12	PB164229BS	PB164229BS	BN034694.D	24 Oct 2024 16:05	USE For Only P4428	RC/JU	Ok
13	P4428-03	RW7-SP302-20241016	BN034695.D	24 Oct 2024 16:41		RC/JU	Ok
14	P4428-04	RW7-SP303-20241016	BN034696.D	24 Oct 2024 17:18		RC/JU	Ok
15	P4475-04	BP-VPB-190-GW-238-2	BN034697.D	24 Oct 2024 17:54		RC/JU	Ok
16	P4475-06	BP-VPB-190-GW-278-2	BN034698.D	24 Oct 2024 18:30		RC/JU	Ok
17	SSTDCCC0.4	SSTDCCC0.4EC	BN034699.D	24 Oct 2024 19:06		RC/JU	Ok

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN102424

Review By	yogesh	Review On	10/25/2024 11:49:23 PM
Supervise By	mohammad	Supervise On	10/26/2024 4:13:48 AM
SubDirectory	BN102424	HP Acquire Method	BNA_N
		HP Processing Method	bn102424

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597
CCC	SP6601
Internal Standard/PEM	SP6527
ICV/I.BLK	SP6548
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	DFTPP	DFTPP	BN034700.D	24 Oct 2024 20:22		RC/JU	Ok
19	SSTDCCC0.4	SSTDCCC0.4	BN034701.D	24 Oct 2024 21:01		RC/JU	Ok
20	PB164230BL	PB164230BL	BN034702.D	24 Oct 2024 21:38	USE For Only P4432	RC/JU	Ok
21	PB164323BL	PB164323BL	BN034703.D	24 Oct 2024 22:14		RC/JU	Ok
22	PB164229BSD	PB164229BSD	BN034704.D	24 Oct 2024 22:50		RC/JU	Ok
23	PB164282BL	PB164282BL	BN034705.D	24 Oct 2024 23:26		RC/JU	Ok
24	PB164282BS	PB164282BS	BN034706.D	25 Oct 2024 00:02		RC/JU	Ok,M
25	PB164230BS	PB164230BS	BN034707.D	25 Oct 2024 00:38	USE For Only P4432	RC/JU	Ok,M
26	PB164323BS	PB164323BS	BN034708.D	25 Oct 2024 01:14		RC/JU	Ok,M
27	PB164323BSD	PB164323BSD	BN034709.D	25 Oct 2024 01:50		RC/JU	Ok,M
28	P4428-02	RW7-SP201-20241016	BN034710.D	25 Oct 2024 02:26		RC/JU	Ok
29	PB164282BSD	PB164282BSD	BN034711.D	25 Oct 2024 03:02		RC/JU	Ok,M
30	P4438-03	VPB190-HYD-2024101	BN034712.D	25 Oct 2024 03:38	Surrogate Fail	RC/JU	ReRun
31	P4438-04	BP-VPB-190-GW-58-60	BN034713.D	25 Oct 2024 04:14		RC/JU	Ok
32	P4438-05	BP-VPB-190-GW-98-10	BN034714.D	25 Oct 2024 04:51		RC/JU	Ok
33	P4438-06	BP-VPB-190-GW-158-1	BN034715.D	25 Oct 2024 05:27		RC/JU	Ok
34	P4438-07	BP-VPB-190-GW-198-2	BN034716.D	25 Oct 2024 06:03		RC/JU	Ok
35	P4440-01	BPOW6-11-HYD-20241	BN034717.D	25 Oct 2024 06:39	Surrogate Fail	RC/JU	ReRun
36	P4432-01	LOT-4078A	BN034718.D	25 Oct 2024 07:15	Need Straight Run	RC/JU	Not Ok

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN102424

Review By	yogesh	Review On	10/25/2024 11:49:23 PM				
Supervise By	mohammad	Supervise On	10/26/2024 4:13:48 AM				
SubDirectory	BN102424	HP Acquire Method	BNA_N	HP Processing Method	bn102424		
STD. NAME		STD REF.#					
Tune/Reschk		SP6573					
Initial Calibration Stds		SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597					
CCC		SP6601					
Internal Standard/PEM		SP6527					
ICV/I.BLK		SP6548					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							
37	SSTDCCC0.4	SSTDCCC0.4EC	BN034719.D	25 Oct 2024 07:51		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN102524

Review By	yogesh	Review On	10/26/2024 12:01:15 AM
Supervise By	mohammad	Supervise On	10/26/2024 4:14:54 AM
SubDirectory	BN102524	HP Acquire Method	BNA_N
		HP Processing Method	bn102424

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597
CCC	SP6601
Internal Standard/PEM	SP6527
ICV/I.BLK	SP6548
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN034720.D	25 Oct 2024 09:07		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN034721.D	25 Oct 2024 09:46		RC/JU	Ok
3	PB164323BL	PB164323BL	BN034722.D	25 Oct 2024 10:22		RC/JU	Ok
4	P4438-03RE	VPB190-HYD-2024101	BN034723.D	25 Oct 2024 10:58	Surrogate Fail	RC/JU	Confirms
5	P4440-01RE	BPOW6-11-HYD-20241	BN034724.D	25 Oct 2024 11:34	Surrogate Fail	RC/JU	Confirms
6	PB164409BL	PB164409BL	BN034725.D	25 Oct 2024 16:28	Analyzed for contamination check	RC/JU	Not Ok
7	PB164409BL	PB164409BL	BN034726.D	25 Oct 2024 17:04		RC/JU	Ok
8	PB164409BS	PB164409BS	BN034727.D	25 Oct 2024 17:40		RC/JU	Ok,M
9	PB164409BSD	PB164409BSD	BN034728.D	25 Oct 2024 18:16		RC/JU	Ok,M
10	P4550-02	VPB190-HYD-2024102	BN034729.D	25 Oct 2024 18:53	Surrogate Fail	RC/JU	ReRun
11	P4550-05	BP-VPB-190-GW-318-3	BN034730.D	25 Oct 2024 19:29		RC/JU	Ok,M
12	P4550-09	BP-VPB-190-GW-358-3	BN034731.D	25 Oct 2024 20:05		RC/JU	Ok
13	SSTDCCC0.4	SSTDCCC0.4EC	BN034732.D	25 Oct 2024 20:41		RC/JU	Ok

M : Manual Integration

SOP ID: M3510C,3580A-Extraction SVOC-20

Clean Up SOP #: N/A

Matrix : Water

Welgh By: N/A **Extraction By:** RS

Balance check: N/A **Filter By:** RS

Balance ID: N/A **pH Meter ID:** N/A

pH Strip Lot#: E3574 **Hood ID:** 4,6,7

Extraction Method: ☒ Separatory Funnel ☐ Continous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet

Extraction Start Date : 10/21/2024

Extraction Start Time : 08:45

Extraction End Date : 10/21/2024

Extraction End Time : 13:45

Concentration By: EH

Supervisor By : rajesh

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	0.4 PPM	SP6606
Surrogate	1.0ML	0.4 PPM	SP6634
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3817
Baked Na2SO4	N/A	EP2551
10N NaoH	N/A	EP2550
H2SO4 1:1	N/A	EP2548
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

1.5 ML Vial lot# 2210673. pH Adjusted<2 with 1:1 H2SO4 &>11 with 10 N NaOH.

KD Bath ID: Water bath -01 **Envap ID:** NEVAP-02

KD Bath Temperature: 60 °C **Envap Temperature:** 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Recelved By/Location
10/21/24	RP (Ext. Lab)	Relsvoc
13:50	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction SVOC-20

Concentration Date: 10/21/2024

Sample ID	Client Sample ID	Test	g / (ml)	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB164282BL	SBLK282	SVOC-SIMGrou p1	1000	6	RUPESH	rajesh	1			SEP-10
PB164282BS	SLCS282	SVOC-SIMGrou p1	1000	6	RUPESH	rajesh	1			11
PB164282BS D	SLCSD282	SVOC-SIMGrou p1	1000	6	RUPESH	rajesh	1			12
P4438-03	VPB190-HYD-20241016	SVOC-SIMGrou p1	970	6	RUPESH	rajesh	1	C		13
P4438-04	BP-VPB-190-GW-58-60	SVOC-SIMGrou p1	740	6	RUPESH	rajesh	1	C	Muddy	14
P4438-05	BP-VPB-190-GW-98-100	SVOC-SIMGrou p1	700	6	RUPESH	rajesh	1	C		15
P4438-06	BP-VPB-190-158-160	SVOC-SIMGrou p1	960	6	RUPESH	rajesh	1	C		16
P4438-07	BP-VPB-190-GW-198-200	SVOC-SIMGrou p1	780	6	RUPESH	rajesh	1	C	Muddy	17
P4440-01	BPOW6-11-HYD-20241016	SVOC-SIMGrou p1	1000	6	RUPESH	rajesh	1	C		18

* Extracts relinquished on the same date as received.

WORKLIST(Hardcopy Internal Chain)

WorkList Name : P4438 WorkList ID : 184626 Department : Extraction Date : 10-21-2024 08:41:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4438-03	VPB190-HYD-20241016	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	K51	10/16/2024	8270-Modified
P4438-04	BP-VPB-190-GW-58-60	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	K51	10/15/2024	8270-Modified
P4438-05	BP-VPB-190-GW-98-100	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	K51	10/16/2024	8270-Modified
P4438-06	BP-VPB-190-158-160	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	K51	10/17/2024	8270-Modified
P4438-07	BP-VPB-190-GW-198-200	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	K51	10/17/2024	8270-Modified
P4440-01	BPOW6-11-HYD-20241016	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	K51	10/16/2024	8270-Modified

Date/Time 10/21/24 8:43
 Raw Sample Received by: R. (S. 10/21/24)
 Raw Sample Relinquished by: R. (S. 10/21/24)

Date/Time 10/21/24 8:10
 Raw Sample Received by: R. (S. 10/21/24)
 Raw Sample Relinquished by: R. (S. 10/21/24)

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

Prep Standard - Chemical Standard Summary

Order ID : P4440
Test : SVOC-SIMGroup1
Prepbatch ID : PB164282,
Sequence ID/Qc Batch ID: BN102224,BN102424,BN102524,

Standard ID :
EP2548,EP2550,EP2551,SP6527,SP6547,SP6548,SP6573,SP6596,SP6597,SP6598,SP6599,SP6600,SP6601,SP6602,SP6603,SP6606,SP6634,

Chemical ID :
E3551,E3657,E3746,E3759,E3768,E3786,E3788,E3817,M5178,S10103,S10247,S10782,S10977,S 11003,S11011,S11097,S11494,S11566,S11766,S11771,S12029,S12077,S12096,S12105,S12112,S12113,S12117,S12126,S12453,W3112,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2548	10/16/2024	04/16/2025	Rajesh Parikh	None	None	RUPESHKUMAR SHAH 10/16/2024

FROM 1000.00000ml of W3112 + 4000.00000ml of M5178 = Final Quantity: 2000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1874	10 N SODIUM HYDROXIDE SOLN	EP2550	10/16/2024	04/16/2025	Rajesh Parikh	Extraction_SC ALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 10/16/2024

FROM 1000.00000ml of W3112 + 400.00000gram of E3657 = Final Quantity: 1000.000 ml

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2551	10/18/2024	01/03/2025	Rajesh Parikh	Extraction_SC ALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 10/18/2024

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3493	Internal Standard 0.4 PPM	SP6527	06/10/2024	12/05/2024	Jagrut Upadhyay	None	None	mohammad ahmed 07/05/2024

FROM 0.10000ml of S12029 + 4.90000ml of E3759 = Final Quantity: 5.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3355	8270-SIM MDL-3.2PPM CALIBRATION STOCK SOL- 2ND	SP6547	07/08/2024	11/21/2024	Jagrut Upadhyay	None	None	mohammad ahmed
SOURCE FROM 0.00630ml of S10977 + 0.01280ml of S11003 + 0.03200ml of S10782 + 0.03200ml of S11766 + 0.06400ml of S11566 + 0.06400ml of S12096 + 0.06400ml of S12117 + 19.72490ml of E3746 = Final Quantity: 20.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3356	8270-SIM MDL-0.4PPM CALIBRATION SOL ICV-2ND	SP6548	07/08/2024	11/21/2024	Jagrut Upadhyay	None	None	mohammad ahmed
SOURCE FROM 0.87500ml of E3746 + 0.01000ml of SP6527 + 0.12500ml of SP6547 = Final Quantity: 1.010 ml								

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3895	50 ug/ml DFTTP 8270E	SP6573	07/15/2024	01/08/2025	Rahul Chavli	None	None	Yogesh Patel
								07/17/2024

FROM 1.00000ml of S10247 + 19.00000ml of E3768 = Final Quantity: 20.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3339	8270 sim calibration stock 10ppm (CPI)	SP6596	08/09/2024	11/21/2024	Jagrut Upadhyay	None	None	Yogesh Patel
								08/21/2024

FROM 0.02500ml of S12113 + 0.03350ml of S10103 + 0.05000ml of S11494 + 0.10000ml of S12112 + 0.12500ml of S10782 + 0.25000ml of S11097 + 0.25000ml of S12077 + 24.16650ml of E3786 = Final Quantity: 25.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3361	8270-SIM MDL-5PPM CALIBRATION SOLUTION	SP6597	08/09/2024	11/21/2024	Jagrut Upadhyay	None	None	Yogesh Patel
08/21/2024								

FROM 0.50000ml of E3786 + 0.01000ml of SP6527 + 0.50000ml of SP6596 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3341	8270-SIM MDL-3.2PPM CALIBRATION SOLUTION	SP6598	08/09/2024	11/21/2024	Jagrut Upadhyay	None	None	Yogesh Patel
08/21/2024								

FROM 0.68000ml of E3786 + 0.01000ml of SP6527 + 0.32000ml of SP6596 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3344	8270-SIM MDL-1.6PPM CALIBRATION SOLUTION	SP6599	08/09/2024	11/21/2024	Jagrut Upadhyay	None	None	Yogesh Patel
								08/21/2024

FROM 0.84000ml of E3786 + 0.01000ml of SP6527 + 0.16000ml of SP6596 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3342	8270-SIM MDL-0.8PPM CALIBRATION SOLUTION	SP6600	08/09/2024	11/21/2024	Jagrut Upadhyay	None	None	Yogesh Patel
								08/21/2024

FROM 0.92000ml of E3786 + 0.01000ml of SP6527 + 0.08000ml of SP6596 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3343	8270-SIM MDL-0.4PPM CALIBRATION SOLUTION	SP6601	08/09/2024	11/21/2024	Jagrut Upadhyay	None	None	Yogesh Patel
								08/21/2024

FROM 0.96000ml of E3786 + 0.01000ml of SP6527 + 0.04000ml of SP6596 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3345	8270-SIM MDL-0.2PPM CALIBRATION SOLUTION	SP6602	08/09/2024	11/21/2024	Jagrut Upadhyay	None	None	Yogesh Patel
								08/21/2024

FROM 0.50000ml of E3786 + 0.01000ml of SP6527 + 0.50000ml of SP6601 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3346	8270-SIM MDL-0.1PPM CALIBRATION SOLUTION	SP6603	08/09/2024	11/21/2024	Jagrut Upadhyay	None	None	Yogesh Patel
								08/21/2024

FROM 0.75000ml of E3786 + 0.01000ml of SP6527 + 0.25000ml of SP6601 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3492	8270-SIM-Spike 0.4 PPM	SP6606	08/20/2024	02/12/2025	Rahul Chavli	None	None	mohammad ahmed
								08/21/2024

FROM 0.00160ml of S11011 + 0.02000ml of S11771 + 0.04000ml of S12105 + 0.04000ml of S12126 + 0.04000ml of S12453 + 99.85840ml of E3788 = Final Quantity: 100.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3491	8270-SIM-Surrogate 0.4 PPM	SP6634	09/20/2024	11/21/2024	Rahul Chavli	None	None	Yogesh Patel
FROM 0.00400ml of S10977 + 0.00800ml of S11003 + 0.02000ml of S10782 + 99.96800ml of E3788 = Final Quantity: 100.000 ml								

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	01/03/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-5 / Sodium Hydroxide Pellets 2.5 Kg, Pk of 4	23B1556310	12/31/2025	12/04/2023 / Rajesh	12/01/2023 / Rajesh	E3657

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24C0162011	11/25/2024	05/25/2024 / Rajesh	05/08/2024 / Rajesh	E3746

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24D1962005	12/08/2024	06/08/2024 / Rajesh	05/31/2024 / Rajesh	E3759

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24E2462004	01/08/2025	07/08/2024 / Rajesh	06/21/2024 / Rajesh	E3768

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24F1062004	02/01/2025	08/01/2024 / Rajesh	07/16/2024 / Rajesh	E3786

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	23H1462005	04/23/2025	08/13/2024 / Rajesh	08/13/2024 / Rajesh	E3788

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24H2762011	04/09/2025	10/09/2024 / Rajesh	10/09/2024 / Rajesh	E3817

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	0000281827	03/29/2026	05/25/2022 / william	04/05/2022 / william	M5178

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-112090-04 / CLP Acid Surrogate Solution, 7500 mg/L, 1ml	440246	02/08/2025	08/08/2024 / Jagrut	12/09/2021 / Christian	S10103

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31615 / SV Mixture, GC/MS Tuning Mixture, CH2Cl2, 1mL,	A0182667	01/15/2025	07/15/2024 / Rahul	03/18/2022 / Christian	S10247

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	33913 / SOM01.0 SIM Analysis Standard (Surrogate), 2000 PPM	A0186160	11/21/2024	05/21/2024 / Jagrut	09/07/2022 / Christian	S10782

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0188108	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S10977

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH ₂ Cl ₂ ,5ml	A0189418	11/30/2024	05/31/2024 / Jagrut	12/28/2022 / Christian	S11003

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555872 / Custom Standard, pentachlorophenol Std [CS 5328-5]	A0193449	02/20/2025	08/20/2024 / yogesh	01/13/2023 / Christian	S11011

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-110381-01 / 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1ml	495831	02/08/2025	08/08/2024 / Jagrut	02/07/2023 / Christian	S11097

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110094-02 / CLP Base/Neutral Surrogate Solution, 5000 mg/L, 1ml	506889	02/08/2025	08/08/2024 / Jagrut	08/11/2023 / Yogesh	S11494

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0201940	12/05/2024	06/05/2024 / Rahul	09/18/2023 / Kiran	S11566

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	12/14/2024	06/14/2024 / Rahul	11/21/2023 / Rahul	S11766

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	02/20/2025	08/20/2024 / Rahul	11/21/2023 / Rahul	S11771

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0201320	12/05/2024	06/05/2024 / Rahul	12/21/2023 / Rahul	S12029

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110816-01 / Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL, (Maximum Expiration: 180 Days)	414127	02/08/2025	08/08/2024 / Jagrut	01/31/2024 / Rahul	S12077

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	12/05/2024	06/05/2024 / Rahul	02/05/2024 / Rahul	S12096

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0207706	02/12/2025	08/12/2024 / Rahul	02/05/2024 / Rahul	S12105

[CS 4978-2]

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-010223-01 / 1,4-Dioxane Solution, 2,000mg/L, 1ml	454157	01/09/2025	07/09/2024 / Jagrut	03/08/2024 / Rahul	S12112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-010223-01 / 1,4-Dioxane Solution, 2,000mg/L, 1ml	454157	02/09/2025	08/09/2024 / Jagrut	03/08/2024 / Rahul	S12113

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	12/05/2024	06/05/2024 / Rahul	03/15/2024 / Rahul	S12117

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	02/12/2025	08/12/2024 / Rahul	03/15/2024 / Rahul	S12126

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	02/12/2025	08/12/2024 / Rahul	07/23/2024 / RAHUL	S12453

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.: **Storage:** **Solvent:** **Exp. Date:** **Description:**
Z-112090 440246 $\leq -10^{\circ}\text{C}$ Methylene Chloride 2/16/2026 CLP Acid Surrogate Solution, 7,500 mg/L, 1 mL
-04

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
2-chlorophenol-d ₄	93951-73-6	99.3	248.12.7P	7487 $\pm$ 17.2
2-fluorophenol	367-12-4	99.8	10.7.3.3P	7513 $\pm$ 17.26
phenol-d ₆	13127-88-3	99.9	949.120.8P	7481 $\pm$ 17.19
2,4,6-tribromophenol	118-79-6	99.8	12.1.6P	7469 $\pm$ 17.17

Received on

02/25/21

by
CG

S9236
to

S9240

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Erica Castiglione

Certified By:

Erica Castiglione
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Receivalon
02/07/23 by CG

Sheet 511096
to
511099

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 4

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110381-01	495831	≤ -10 °C	Methylene Chloride	10/30/2027	Method 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acenaphthene	83-32-9	99.9	13.1.5P	1003 ± 17.27
acenaphthylene	208-96-8	97.6	14.290.1P	999.8 ± 17.22
aniline	62-53-3	99.9	64.7.1P	995 ± 17.13
anthracene	120-12-7	99.5	15.7.1P	1001 ± 17.24
azobenzene	103-33-3	98.1	252.7.2P	999.1 ± 17.21
benzo[a]anthracene	56-55-3	100	16.7.3P	1001 ± 17.24
benzo[b]fluoranthene	205-99-2	99.8	17.421.3P	1001 ± 19.91
benzo[k]fluoranthene	207-08-9	98.9	18.421.4P	1001 ± 17.92
benzo[ghi]perylene	191-24-2	93	19.286.4P	999.6 ± 19.88
benzo[a]pyrene	50-32-8	97	20.286.2P	999.1 ± 26.35
benzyl alcohol	100-51-6	99.9	65.18.1P	1001 ± 17.24
bis(2-chloroethoxy)methane	111-91-1	99.1	31.3.15P	999.7 ± 17.89
bis(2-chloroethyl)ether	111-44-4	99.8	32.7.1P	1001 ± 17.23
bis(2-chloro-1-methylethyl) ether	108-60-1	99.5	34.3.13P	999.5 ± 17.89
bis(2-ethylhexyl)adipate	103-23-1	99.5	874.7.1P	999.5 ± 17.21
bis(2-ethylhexyl)phthalate	117-81-7	99.4	33.29.1P	998.8 ± 19.86
4-bromophenyl phenyl ether	101-55-3	99.4	35.7.1P	999.1 ± 17.2
butyl benzyl phthalate	85-68-7	98.4	36.1.6P	984.7 ± 19.58
carbazole	86-74-8	99.4	239.7.2P	1000 ± 17.22

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Brianna Smith

Certified By: _____

Brianna Smith
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certificate of Analysis

Page 4 of 4

Catalog No.: Z-110381-01

Lot No.: 495831

Expiration Date: 10/30/2027

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2,4-trichlorobenzene	120-82-1	99.6	54.29.1P	1000 ± 17.22
2,4,5-trichlorophenol	95-95-4	96.5	121.7.1.1P	1000 ± 17.22
2,4,6-trichlorophenol	88-06-2	99.6	113.7.1P	1002 ± 17.25

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By:



Briana Smith
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received on
03/10/22
by
CG
S10242
to
S10247

Catalog No. : 31615 **Lot No.:** A0182667
Description : GC/MS Tuning Mixture
GC/MS Tuning Mixture 1,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2025 **Storage:** 10°C or colder
Handling: Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Pentachlorophenol CAS # 87-86-5 (Lot 211229RSR) Purity 99%	1,003.6 µg/mL	+/- 5.8897 µg/mL +/- 45.7132 µg/mL +/- 66.0037 µg/mL	Gravimetric Unstressed Stressed
2	DFTPP (Decafluorotriphenylphosphine) CAS # 5074-71-5 (Lot Q117-147) Purity 95%	1,006.6 µg/mL	+/- 5.9074 µg/mL +/- 45.8508 µg/mL +/- 66.2023 µg/mL	Gravimetric Unstressed Stressed
3	Benzidine CAS # 92-87-5 (Lot 211228JLM) Purity 99%	1,008.4 µg/mL	+/- 5.9179 µg/mL +/- 45.9318 µg/mL +/- 66.3193 µg/mL	Gravimetric Unstressed Stressed
4	4,4'-DDT CAS # 50-29-3 (Lot 210916JLM) Purity 99%	1,007.6 µg/mL	+/- 5.9132 µg/mL +/- 45.8954 µg/mL +/- 66.2667 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 08-Mar-2022

Balance: B345965662

Marlene Cowan - Operations Tech I

Date Passed: 10-Mar-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received on
09/07/22
by
CG
S10778
to
S10782

Catalog No. : 33913 **Lot No.:** A0186160

Description : SOM01.0 SIM Analysis Standard

SOM01.0 SIM Analysis Standard 2000µg/mL, Methylene chloride, 1mL /ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2028 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	2-Methylnaphthalene-d10	2,015.0 µg/mL	+/- 11.8254 µg/mL	Gravimetric
	CAS # 7297-45-2 (Lot EF-135)		+/- 90.7728 µg/mL	Unstressed
	Purity 96%		+/- 100.7207 µg/mL	Stressed
2	Fluoranthene-d10	2,007.0 µg/mL	+/- 11.7782 µg/mL	Gravimetric
	CAS # 93951-69-0 (Lot PR-20668)		+/- 90.4107 µg/mL	Unstressed
	Purity 99%		+/- 100.3188 µg/mL	Stressed
Solvent: Methylene chloride				
CAS # 75-09-2				
Purity 99%				

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 09-Jun-2022

Balance: B442140311

Marlina Cowan - Operations Tech II ARM QC

Date Passed: 13-Jun-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by
CG on
12/28/22
SI0951
FO
SI0980

Catalog No. : 31087 **Lot No.:** A0188108
Description : Acid Surrogate Mix (4/89 SOW)
Acid Surrogate 10,000µg/mL, Methanol, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : August 31, 2030 **Storage:** 10°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	2-Fluorophenol CAS # 367-12-4 Purity 99% (Lot STBF3761V)	10,088.5 µg/mL	+/- 58.6554 µg/mL Gravimetric +/- 294.4162 µg/mL Unstressed +/- 357.2628 µg/mL Stressed
2	Phenol-d6 CAS # 13127-88-3 Purity 99% (Lot PR-31262)	10,043.3 µg/mL	+/- 58.3923 µg/mL Gravimetric +/- 293.0957 µg/mL Unstressed +/- 355.6603 µg/mL Stressed
3	2,4,6-Tribromophenol CAS # 118-79-6 Purity 99% (Lot MKCJ7664)	10,010.0 µg/mL	+/- 58.1990 µg/mL Gravimetric +/- 292.1253 µg/mL Unstressed +/- 354.4829 µg/mL Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)


Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Morgan Craighead - Mix Technician

Date Mixed: 02-Aug-2022 **Balance:** 1127510105


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 05-Aug-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 **Lot No.:** A0189418
Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : August 31, 2028 **Storage:** 10°C or colder
Handling: Sonicate prior to use. **Ship:** Ambient

Received by
CG on
12/28/22
\$10981
to
S11010

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	Nitrobenzene-d5 CAS # 4165-60-0 (Lot PR-29940A) Purity 99%	5,009.8 µg/mL	+/- 29.1271 µg/mL Gravimetric +/- 225.6421 µg/mL Unstressed +/- 250.3778 µg/mL Stressed
2	2-Fluorobiphenyl CAS # 321-60-8 (Lot 00021384) Purity 99%	5,026.6 µg/mL	+/- 29.2250 µg/mL Gravimetric +/- 226.4003 µg/mL Unstressed +/- 251.2191 µg/mL Stressed
3	p-Terphenyl-d14 CAS # 1718-51-0 (Lot PR-30504) Purity 99%	5,027.3 µg/mL	+/- 29.2289 µg/mL Gravimetric +/- 226.4304 µg/mL Unstressed +/- 251.2524 µg/mL Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:


330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


John Friedline - Operations Technician I

Date Mixed: 09-Sep-2022

Balance: 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 13-Sep-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555872 **Lot No.:** A0193449
Description : Custom Pentachlorophenol Standard
Custom Pentachlorophenol Standard 25,000µg/mL, Methanol,
1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : January 31, 2026 **Storage:** 10°C or colder
Ship: Ambient

Received on
01/13/23
by
CG
S11011
+0
S11015

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP221012	99%	25,050.0 µg/mL	+/- 778.6378

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer

Russ Bookhamer - Operations Technician I

Date Mixed: 11-Jan-2023

Balance: B442140311

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %
COMMENTS		
QC: PhC Irma Belmares		

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3



Certificate of Analysis

Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	$\leq 0.005 \%$	$< 0.005 \%$	PASS
Chloride	$\leq 0.005 \%$	0.002 %	PASS
Heavy Metals	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Iron	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Magnesium	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Mercury	$\leq 0.1 \text{ ppm}$	$< 0.1 \text{ ppm}$	PASS
Nickel	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Nitrogen Compounds	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Phosphate	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Potassium	$\leq 0.02 \%$	$< 0.02 \%$	PASS
Purity	$\geq 97.0 \%$	99.2 %	PASS
Sodium Carbonate	$\leq 1.0 \%$	0.5 %	PASS
Sulfate	$\leq 0.003 \%$	$< 0.003 \%$	PASS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.


This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.



Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24C0162011
Manufactured Date: 2024-01-04
Expiration Date: 2025-04-04
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	2
Assay (CH ₂ Cl ₂) (by GC, exclusive of preservative, corrected for water)	≥ 99.8 %	100.0 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	≤ 0.02 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC
Manufacturer source batch: MG24A04224

E 3746

Ken Koehnlein
Sr. Manager, Quality Assurance

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24D1962005
Manufactured Date: 2024-03-16
Expiration Date: 2025-06-15
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	8
Assay (CH ₂ Cl ₂) (by GC, exclusive of preservative, corrected for water)	≥ 99.8 %	99.9 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Titration Acid (μeq/g)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	≤ 0.02 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC
Manufacturer source batch: MG24C16563

E 3759


Jamie Croak
Director Quality Operations, Bioscience Production

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24E2462004
Manufactured Date: 2024-04-10
Expiration Date: 2025-07-10
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	3
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8 \%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	$\leq 1.0 \text{ ppm}$	0.1 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	$\leq 10 \text{ ppm}$	5 ppm
Water (by KF, coulometric)	$\leq 0.02 \%$	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC
Manufacturer source batch: MG24D10725

E 3768

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087, U.S.A. Phone 610.386.1700

Page 1 of 1

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24F1062004
Manufactured Date: 2024-04-15
Expiration Date: 2025-07-15
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	7
Assay (CH ₂ Cl ₂) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8\%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Titration Acid (μ eq/g)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02\%$	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC
Manufacturer source batch: MG24D15750

E 3786

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087, U.S.A. Phone 610.386.1700

Page 1 of 1

Acetone

BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

Avantor™



Material No.: 9254-03
Batch No.: 23H1462005
Manufactured Date: 2023-07-26
Expiration Date: 2026-07-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	≥ 99.4 %	99.7 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	≤ 0.3	0.1
Titration Base (μeq/g)	≤ 0.6	< 0.1
Water (H ₂ O)	≤ 0.5 %	0.3 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 8/13/24

E 3788

Ken Koehnlein
Sr. Manager, Quality Assurance

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24H2762011
Manufactured Date: 2024-06-05
Expiration Date: 2025-09-04
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	2
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	5
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8\%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02\%$	$< 0.01\%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3817

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Hydrochloric Acid, 36.5–38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis



Material No.: 9530-33
Batch No.: 0000281827
Manufactured Date: 2021/03/30
Retest Date: 2026/03/29
Revision No: 1

Certificate of Analysis

Test	Specification	Result
ACS – Assay (as HCl) (by acid–base titrn)	36.5 – 38.0 %	37.6
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	1
ACS – Specific Gravity at 60°/60°F	1.185 – 1.192	1.189
ACS – Bromide (Br)	<= 0.005 %	< 0.005
ACS – Extractable Organic Substances	<= 5 ppm	< 1
ACS – Free Chlorine (as Cl ₂)	<= 0.5 ppm	< 0.5
Phosphate (PO ₄)	<= 0.05 ppm	< 0.03
Sulfate (SO ₄)	<= 0.5 ppm	< 0.3
Sulfite (SO ₃)	<= 0.8 ppm	0.3
Ammonium (NH ₄)	<= 3 ppm	< 1
Trace Impurities – Arsenic (As)	<= 0.010 ppm	< 0.003
Trace Impurities – Aluminum (Al)	<= 10.0 ppb	0.5
Arsenic and Antimony (as As)	<= 5 ppb	< 3
Trace Impurities – Barium (Ba)	<= 1.0 ppb	< 0.2
Trace Impurities – Beryllium (Be)	<= 1.0 ppb	< 0.2
Trace Impurities – Bismuth (Bi)	<= 10.0 ppb	< 1.0
Trace Impurities – Boron (B)	<= 20.0 ppb	< 5.0
Trace Impurities – Cadmium (Cd)	<= 1.0 ppb	< 0.3
Trace Impurities – Calcium (Ca)	<= 50.0 ppb	15.0
Trace Impurities – Chromium (Cr)	<= 1.0 ppb	< 0.4
Trace Impurities – Cobalt (Co)	<= 1.0 ppb	< 0.3
Trace Impurities – Copper (Cu)	<= 1.0 ppb	< 0.1
Trace Impurities – Gallium (Ga)	<= 1.0 ppb	< 0.2

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Test	Specification	Result
Trace Impurities – Germanium (Ge)	<= 3.0 ppb	< 2.0
Trace Impurities – Gold (Au)	<= 4.0 ppb	3.0
Heavy Metals (as Pb)	<= 100 ppb	< 50
Trace Impurities – Iron (Fe)	<= 15.0 ppb	1.0
Trace Impurities – Lead (Pb)	<= 1.0 ppb	< 0.5
Trace Impurities – Lithium (Li)	<= 1.0 ppb	< 0.2
Trace Impurities – Magnesium (Mg)	<= 10.0 ppb	< 0.4
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	< 0.4
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	0.2
Trace Impurities – Molybdenum (Mo)	<= 10.0 ppb	< 5.0
Trace Impurities – Nickel (Ni)	<= 4.0 ppb	< 0.3
Trace Impurities – Niobium (Nb)	<= 1.0 ppb	< 0.2
Trace Impurities – Potassium (K)	<= 9.0 ppb	< 2.0
Trace Impurities – Selenium (Se), For Information Only	ppb	1.0
Trace Impurities – Silicon (Si)	<= 100.0 ppb	18.0
Trace Impurities – Silver (Ag)	<= 1.0 ppb	< 0.3
Trace Impurities – Sodium (Na)	<= 100.0 ppb	< 5.0
Trace Impurities – Strontium (Sr)	<= 1.0 ppb	< 0.2
Trace Impurities – Tantalum (Ta)	<= 1.0 ppb	< 0.9
Trace Impurities – Thallium (Tl)	<= 5.0 ppb	< 2.0
Trace Impurities – Tin (Sn)	<= 5.0 ppb	< 0.8
Trace Impurities – Titanium (Ti)	<= 1.0 ppb	< 0.2
Trace Impurities – Vanadium (V)	<= 1.0 ppb	< 0.2
Trace Impurities – Zinc (Zn)	<= 5.0 ppb	0.4
Trace Impurities – Zirconium (Zr)	<= 1.0 ppb	< 0.1

For Laboratory, Research or Manufacturing Use

Product Information (not specifications):

Appearance (clear, fuming liquid)

Meets ACS Specifications

Country of Origin: US

Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



5580 Skylane Blvd
Santa Rosa, CA 95403
(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Page 1 of 1

Catalog No.: Lot No.:	Storage:	Solvent:	Exp. Date:	Rev	Description:
Z-110094-02 506889	≤ -10 °C	Methylene Chloride	7/25/2028	0	CLP Base/Neutral Surrogate Solution, 5,000 mg/L, 1 ml
Compound		CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2-dichlorobenzene-d ₄		2199-69-1	99.7	247.29.3P	5035 ± 28.02
2-fluorobiphenyl		321-60-8	99.69	8.286.1.1P	4999 ± 103.66
nitrobenzene-d5		4165-60-0	99.67	7.9.3P	4988 ± 27.32
p-terphenyl-d14		1718-51-0	99.3	9.120.8P	5005 ± 27.85

511494 } Y.P.
↓ 08/11/2023
511498

*Not a certified value

Certified By: _____
Clint Tipton
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 555223

Lot No.: A0201940

Description:

Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size:

2 mL

Expiration Date:

September 30, 2025

Handling:

This product is photosensitive.

Pkg Amt: > 1 mL

Storage: 10°C or colder

Ship: Ambient

511539
↓
51568
7.6
0911

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S230321RSR	99%	1,001.0 µg/mL	+/- 22.9799
2	Atrazine	1912-24-9	5FYWL	99%	1,010.0 µg/mL	+/- 23.1865
3	Benzidine	92-87-5	S221205RSR	99%	1,008.0 µg/mL	+/- 23.1406
4	epsilon-Caprolactam	105-60-2	I16X016	99%	1,008.0 µg/mL	+/- 23.1406

Solvent:

Methylene chloride

CAS # 75-09-2

Purity 99%

REVIEWED

By Jennifer Pollino at 7:10 am, Sep 13, 2023

Date Mixed: 13-Sep-2023

Balance: B345965662

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853

Lot No.: A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2028

Storage: 0°C or colder

Ship: Ambient

511749
↓
511794 } RC / 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023

Balance Serial # B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453
Description : 1,4-dioxane
1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2028 **Storage:** 0°C or colder
Ship: Ambient

S11749
↓
S11794 } RC / 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023

Balance Serial # B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206 **Lot No.:** A0201320

Description : SV Internal Standard Mix 2mg/ml
SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2029 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S12013
↓
S12042 } RC
12/26/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,017.0 µg/mL	+/- 90.8469
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,011.3 µg/mL	+/- 90.5917
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,008.6 µg/mL	+/- 90.4685
4	Phenanthrene-d10	1517-22-2	PR-32303	99%	2,019.4 µg/mL	+/- 90.9550
5	Chrysene-d12	1719-03-5	PR-32210	99%	2,013.7 µg/mL	+/- 90.6968
6	Perylene-d12	1520-96-3	PR-33205	99%	2,012.7 µg/mL	+/- 90.6517

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Peter Robbins - Operations Technician I

Date Mixed: 23-Aug-2023

Balance Serial # B345965662

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 25-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110816-01	414127	≤ -10 °C	Methylene Chloride	6/21/2025	Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
atrazine	1912-24-9	99.5	337.7.3P	997 ± 5.81
benzidine	92-87-5	99.9	124.18.6.2P	991.8 ± 5.77
caprolactam	105-60-2	99.9	271.1.6P	999 ± 5.82

S12075 } RC
↓
S12079 } 02/01/24

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____

Shane Overcash
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

512082 } RC/
↓
512111 } 02/22/24

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0207706

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

512082 } RC/
↓
512111 } 02/22/24

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,001.0 µg/mL	+/- 29.424320
2	Acetophenone	98-86-2	STBH8205	99%	1,004.0 µg/mL	+/- 29.512504
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,005.0 µg/mL	+/- 29.541899
4	Benzoic acid	65-85-0	MKCR2694	99%	1,003.0 µg/mL	+/- 29.483110
5	Biphenyl	92-52-4	MKCL6515	99%	1,006.0 µg/mL	+/- 29.571294

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

John Friedline - Operations Technician I

Date Mixed: 12-Feb-2024

Balance: B345965662



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-020223-01	454157	≤ -10 °C	P/T Methanol	6/10/2026	1,4-Dioxane Solution, 2000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,4-dioxane	123-91-1	100	223.1.3P	1997 ± 57.08

512112 } RC/
↓
912116 } 03/08/24

*Not a certified value

Certified By: _____

Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:	
Z-020223-01	454157	≤ -10 °C	P/T Methanol	6/10/2026	1,4-Dioxane Solution, 2000 mg/L, 1 mL	
Compound			CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,4-dioxane			123-91-1	100	223.1.3P	1997 ± 57.08

512112 } RC/
↓
912116 } 03/08/24

*Not a certified value

Certified By: _____

Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

512117 } RC/
↓ 03/18/24
512146

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988	1
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998	2
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200	3
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573	4
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180	5
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574	6
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837	7
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290	8
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829	9
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937	10
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505	11
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838	12
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909	13
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442	14
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512	15
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230	16
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876	17
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887	18
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265	
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422	
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968	
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998	
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634	
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221	
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240	
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553	
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251	
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220	
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674	
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979	
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130	
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857	
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220	
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958	
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230	
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553	

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCKK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

512117 } RC/
↓ 03/18/24
512146 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988	1
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998	2
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200	3
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573	4
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180	5
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574	6
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837	7
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290	8
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829	9
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937	10
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505	11
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838	12
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909	13
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442	14
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512	15
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230	16
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876	17
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887	18
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265	
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422	
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968	
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998	
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634	
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221	
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240	
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553	
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251	
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220	
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674	
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979	
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130	
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857	
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220	
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958	
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230	
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553	

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCKK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1
Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓
S12508 } 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



SHIPPING DOCUMENTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18

CHEMTECH CHAIN OF CUSTODY RECORD		284 Sheffield Street, Mountainside, NJ 07092 (908) 789-8900 Fax: (908) 78-8922 www.chemtech.net		Chemtech Project Number: P4440/41										
		COC Number:												
CLIENT INFORMATION		PROJECT INFORMATION		BILLING INFORMATION										
COMPANY: Tetra Tech		PROJECT NAME: NWIRP Bethpage		BILL TO: SEE CONTRACT PO#										
ADDRESS: 4433 Corporation Lane Suite 300		PROJECT #: 112G08005-WE13 LOCATION: VPB-196		ADDRESS:										
CITY: Virginia Beach STATE: VA ZIP: 23462		PROJECT MANAGER: Ernie Wu		CITY: STATE: ZIP:										
ATTENTION: Ernie Wu		E-MAIL: ernie.wu@tetrattech.com		ATTENTION: PHONE:										
PHONE: 757-466-4901 FAX: 757-461-4148		PHONE: 757-466-4901 FAX: 757-461-4148												
DATA TURNAROUND INFORMATION		DATA DELIVERABLE INFORMATION		ANALYSIS										
FAX: _____ 2 & 10 _____ DAYS* HARD COPY: _____ 2 & 10 _____ DAYS* EDD _____ 2 & 10 _____ DAYS* * TO BE APPROVED BY CHEMTECH STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS		☐ RESULTS ONLY ☐ USEPA CLP ☐ RESULTS + QC ☐ New York State ASP "B" ☐ New Jersey REDUCED ☐ New York State ASP "A" ☐ New Jersey CLP ☐ Other _____ ☐ EDD Format _____		VOC(SW846-8260B) 1,4 Dioxane (8270 SIM) 1,4 Dioxane (522) <table border="1" style="border-collapse: collapse; text-align: center; width: 100%;"> <tr> <td>1</td><td>2</td><td>3</td><td>4</td><td>5</td><td>6</td><td>7</td><td>8</td><td>9</td> </tr> </table>		1	2	3	4	5	6	7	8	9
1	2	3	4	5	6	7	8	9						
				PRESERVATIVES										
				COMMENTS										
CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE	SAMPLE COLLECTION	# of Bottles									
			COMP GRAB	DATE TIME										
1.	BPOW6-11-HYD-20241016	AQ		10/16/24 10:00	5									
2.	BPOW6-11-TB-20241016	QA		10/16/24 9:00	2									
3.														
4.														
5.														
6.														
7.														
8.														
9.														
10.														

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY					
RELINQUISHED BY SAMPLER	DATE/TIME	RECEIVED BY	Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant ☐ Cooler Temp 2.3° MeOH extraction requires an additional 4oz. Jar for percent solid Comments: Standard TAT		
1.	10/17/24 1700	1.			
RELINQUISHED BY	DATE/TIME	RECEIVED BY			
2.	9:15 10-18-24	2.			
RELINQUISHED BY	DATE/TIME	RECEIVED FOR LAB BY			
3.		3.			

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight
 CHEMTECH: ☐ Picked Up ☐ Overnight

Shipment Complete
☐ YES ☐ NO

WHITE - CHEMTECH COPYFOR RETURN TO CLIENT YELLOW - CHEMTECH COPY PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (L-A-B)	L2219
Maine	2024021
Maryland	296
New Hampshire	255423
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18

LOGIN REPORT/SAMPLE TRANSFER

Order ID : P4440	TETR06	Order Date : 10/18/2024 10:22:00 AM	Project Mgr :
Client Name : Tetra Tech NUS, Inc.		Project Name : CTO WE13	Report Type : Level 4
Client Contact : Ernie Wu		Receive DateTime : 10/18/2024 9:15:00 AM	EDD Type : ADAPT
Invoice Name : Tetra Tech NUS, Inc.		Purchase Order :	Hard Copy Date :
Invoice Contact : Ernie Wu			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
P4440-01	BPOW6-11-HYD-20241016	Water	10/16/2024	10:00	VOCMS Group1		8260-Low	10 Bus. Days	
P4440-02	BPOW6-11-TB-20241016	Water	10/16/2024	09:00	VOCMS Group1		8260-Low	10 Bus. Days	

Relinquished By : 

Date / Time : 10/16/24 1120

Received By : 

Date / Time : 10/16/27 11:20 Ref # 4

Storage Area : VOA Refridgerator Room

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
 Data File : BN034687.D
 Acq On : 24 Oct 2024 10:59
 Operator : RC/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Oct 24 12:43:27 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 24 12:41:40 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.712	152	6279	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	18298	0.400	ng	0.00
13) Acenaphthene-d10	14.325	164	8788	0.400	ng	0.00
19) Phenanthrene-d10	17.063	188	18528	0.400	ng	# 0.00
29) Chrysene-d12	21.247	240	12066	0.400	ng	# 0.00
35) Perylene-d12	23.467	264	10772	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.321	112	15922	0.829	ng	0.00
5) Phenol-d6	6.881	99	20820	0.847	ng	0.00
8) Nitrobenzene-d5	8.845	82	12949	0.854	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	21347	0.829	ng	0.00
14) 2,4,6-Tribromophenol	15.810	330	2227	0.598	ng	0.00
15) 2-Fluorobiphenyl	12.957	172	29939	0.864	ng	0.00
27) Fluoranthene-d10	19.096	212	36834	0.812	ng	0.00
31) Terphenyl-d14	19.705	244	20915	0.896	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.285	88	6845	0.895	ng	98
3) n-Nitrosodimethylamine	3.581	42	10574	1.144	ng	# 97
6) bis(2-Chloroethyl)ether	7.141	93	17336	0.921	ng	99
9) Naphthalene	10.532	128	43252	0.857	ng	99
10) Hexachlorobutadiene	10.831	225	6528	0.807	ng	# 99
12) 2-Methylnaphthalene	12.146	142	26668	0.836	ng	100
16) Acenaphthylene	14.047	152	35850	0.849	ng	99
17) Acenaphthene	14.389	154	25224	0.876	ng	99
18) Fluorene	15.373	166	31002	0.864	ng	100
20) 4,6-Dinitro-2-methylph...	15.448	198	2025	0.799	ng	# 77
21) 4-Bromophenyl-phenylether	16.269	248	8290	0.818	ng	# 80
22) Hexachlorobenzene	16.381	284	9007	0.801	ng	98
23) Atrazine	16.542	200	6777	0.764	ng	# 93
24) Pentachlorophenol	16.716	266	3059	0.693	ng	99
25) Phenanthrene	17.101	178	47181	0.856	ng	100
26) Anthracene	17.200	178	41660	0.845	ng	100
28) Fluoranthene	19.129	202	51141	0.827	ng	100
30) Pyrene	19.486	202	51944	0.963	ng	100
32) Benzo(a)anthracene	21.238	228	39570	0.860	ng	99
33) Chrysene	21.283	228	41243	0.906	ng	99
34) Bis(2-ethylhexyl)phtha...	21.203	149	20342	0.650	ng	# 99
36) Indeno(1,2,3-cd)pyrene	25.698	276	34379	0.828	ng	100
37) Benzo(b)fluoranthene	22.850	252	37118	0.954	ng	93
38) Benzo(k)fluoranthene	22.850	252	37118	0.977	ng	# 91
39) Benzo(a)pyrene	23.371	252	29490	0.892	ng	# 88
40) Dibenzo(a,h)anthracene	25.718	278	26850	0.823	ng	96
41) Benzo(g,h,i)perylene	26.373	276	29665	0.824	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034687.D
Acq On : 24 Oct 2024 10:59
Operator : RC/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Quant Time: Oct 24 12:43:27 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 12:41:40 2024
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.712 min Scan# 669
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8





Tgt Ion:152 Resp: 6279
Ion Ratio Lower Upper
152 100
150 154.4 122.7 184.1
115 63.2 51.0 76.6




#2
1,4-Dioxane
Concen: 0.895 ng
RT: 3.285 min Scan# 56
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59



Tgt Ion: 88 Resp: 6845
Ion Ratio Lower Upper
88 100
43 41.3 32.1 48.1
58 93.9 73.7 110.5









#3
 n-Nitrosodimethylamine
 Concen: 1.144 ng
 RT: 3.581 min Scan# 91
 Delta R.T. -0.000 min
 Lab File: BN034687.D
 Acq: 24 Oct 2024 10:59




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion: 42 Resp: 10574
 Ion Ratio Lower Upper
 42 100
 74 95.9 77.2 115.8
 44 9.3 13.4 20.2#



#4
 2-Fluorophenol
 Concen: 0.829 ng
 RT: 5.321 min Scan# 338
 Delta R.T. -0.000 min
 Lab File: BN034687.D
 Acq: 24 Oct 2024 10:59

Tgt Ion: 112 Resp: 15922
 Ion Ratio Lower Upper
 112 100
 64 66.1 53.0 79.6
 63 34.5 27.6 41.4









#5
Phenol-d6
Concen: 0.847 ng
RT: 6.881 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 99 Resp: 20820
Ion Ratio Lower Upper
99 100
42 27.0 21.8 32.8
71 33.1 26.6 40.0



#6
bis(2-Chloroethyl)ether
Concen: 0.921 ng
RT: 7.141 min Scan# 590
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion: 93 Resp: 17336
Ion Ratio Lower Upper
93 100
63 88.1 71.5 107.3
95 32.1 25.8 38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 955
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	18298
Ion Ratio	Lower	Upper	
136	100		
137	11.3	9.1	13.7
54	12.4	9.8	14.8
68	8.2	6.4	9.6



#8
Nitrobenzene-d5
Concen: 0.854 ng
RT: 8.845 min Scan# 802
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion:	82	Resp:	12949
Ion Ratio	Lower	Upper	
82	100		
128	36.2	29.7	44.5
54	65.0	53.0	79.6









#9
Naphthalene
Concen: 0.857 ng
RT: 10.532 min Scan# 960
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:128 Resp: 43252
Ion Ratio Lower Upper
128 100
129 11.1 9.0 13.6
127 12.8 10.5 15.7



#10
Hexachlorobutadiene
Concen: 0.807 ng
RT: 10.831 min Scan# 988
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion:225 Resp: 6528
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.3 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 0.829 ng
RT: 12.071 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 21347
Ion Ratio Lower Upper
152 100
151 20.9 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.836 ng
RT: 12.146 min Scan# 1221
Delta R.T. -0.000 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion:142 Resp: 26668
Ion Ratio Lower Upper
142 100
141 88.0 70.4 105.6
115 37.3 29.8 44.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.325 min Scan# 1503
Delta R.T. -0.004 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:164 Resp: 8788
Ion Ratio Lower Upper
164 100
162 106.2 81.4 122.0
160 53.4 40.6 61.0



#14
2,4,6-Tribromophenol
Concen: 0.598 ng
RT: 15.810 min Scan# 1638
Delta R.T. -0.006 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion:330 Resp: 2227
Ion Ratio Lower Upper
330 100
332 96.3 76.6 115.0
141 67.4 54.6 81.8









#15
2-Fluorobiphenyl
Concen: 0.864 ng
RT: 12.957 min Scan# 1375
Delta R.T. 0.007 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	172	Resp:	29939
Ion Ratio	Lower	Upper	
172	100		
171	34.3	28.3	42.5
170	23.0	19.8	29.6



#16
Acenaphthylene
Concen: 0.849 ng
RT: 14.047 min Scan# 1477
Delta R.T. 0.007 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion:	152	Resp:	35850
Ion Ratio	Lower	Upper	
152	100		
151	19.1	15.6	23.4
153	12.8	10.6	15.8









#17
Acenaphthene
Concen: 0.876 ng
RT: 14.389 min Scan# 1509
Delta R.T. -0.004 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:154 Resp: 25224
Ion Ratio Lower Upper
154 100
153 109.5 88.2 132.2
152 57.2 46.9 70.3



#18
Fluorene
Concen: 0.864 ng
RT: 15.373 min Scan# 1601
Delta R.T. -0.004 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion:166 Resp: 31002
Ion Ratio Lower Upper
166 100
165 98.5 78.8 118.2
167 13.3 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.063 min Scan# 1

Delta R.T. -0.006 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:188 Resp: 18528

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 15.9 8.9 13.3#



#20

4,6-Dinitro-2-methylphenol

Concen: 0.799 ng

RT: 15.448 min Scan# 1608

Delta R.T. -0.004 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59




Tgt Ion:198 Resp: 2025


Ion Ratio Lower Upper

198 100

51 79.0 91.8 137.8#


105 52.7 44.3 66.5








#21
4-Bromophenyl-phenylether
Concen: 0.818 ng
RT: 16.269 min Scan# 1675
Delta R.T. -0.006 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8





Tgt Ion:248 Resp: 8290
Ion Ratio Lower Upper
248 100
250 93.8 80.9 121.3
141 81.7 40.5 60.7#




#22
Hexachlorobenzene
Concen: 0.801 ng
RT: 16.381 min Scan# 1684
Delta R.T. 0.007 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59



Tgt Ion:284 Resp: 9007
Ion Ratio Lower Upper
284 100
142 56.0 43.6 65.4
249 34.1 26.6 40.0









#23
Atrazine
Concen: 0.764 ng
RT: 16.542 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:200 Resp: 6777
Ion Ratio Lower Upper
200 100
173 32.2 19.7 29.5#
215 42.9 35.7 53.5



#24
Pentachlorophenol
Concen: 0.693 ng
RT: 16.716 min Scan# 1711
Delta R.T. -0.006 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion:266 Resp: 3059
Ion Ratio Lower Upper
266 100
264 61.0 49.4 74.2
268 62.3 50.5 75.7









#25
Phenanthrene
Concen: 0.856 ng
RT: 17.101 min Scan# 1742
Delta R.T. -0.006 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:178 Resp: 47181
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.2
179 15.1 12.2 18.4



#26
Anthracene
Concen: 0.845 ng
RT: 17.200 min Scan# 1750
Delta R.T. 0.007 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion:178 Resp: 41660
Ion Ratio Lower Upper
178 100
176 18.8 14.9 22.3
179 15.2 12.2 18.2





#27

Fluoranthene-d10

Concen: 0.812 ng

RT: 19.096 min Scan# 1981

Delta R.T. 0.002 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:212 Resp: 36834

Ion Ratio Lower Upper

212 100

106 22.1 17.8 26.8

104 12.5 10.2 15.4



#28

Fluoranthene

Concen: 0.827 ng

RT: 19.129 min Scan# 1988

Delta R.T. 0.002 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59




Tgt Ion:202 Resp: 51141


Ion Ratio Lower Upper

202 100

101 16.0 12.8 19.2

203 17.2 13.5 20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.247 min Scan# 2419
Delta R.T. -0.003 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	240	Resp:	12066
Ion Ratio	Lower	Upper	
240	100		
120	24.6	14.2	21.2#
236	30.0	23.1	34.7



#30
Pyrene
Concen: 0.963 ng
RT: 19.486 min Scan# 2065
Delta R.T. 0.002 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion:	202	Resp:	51944
Ion Ratio	Lower	Upper	
202	100		
200	21.0	16.6	25.0
203	17.5	14.2	21.4









#31
 Terphenyl-d14
 Concen: 0.896 ng
 RT: 19.705 min Scan# 2112
 Delta R.T. 0.002 min
 Lab File: BN034687.D
 Acq: 24 Oct 2024 10:59




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:	244	Resp:	20915
Ion Ratio	Lower	Upper	
244	100		
212	10.6	8.6	13.0
122	26.3	22.3	33.5



#32
 Benzo(a)anthracene
 Concen: 0.860 ng
 RT: 21.238 min Scan# 2418
 Delta R.T. 0.006 min
 Lab File: BN034687.D
 Acq: 24 Oct 2024 10:59

Tgt Ion:	228	Resp:	39570
Ion Ratio	Lower	Upper	
228	100		
226	26.6	22.0	33.0
229	20.0	15.6	23.4









#33
Chrysene
Concen: 0.906 ng
RT: 21.283 min Scan# 2423
Delta R.T. -0.002 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:228 Resp: 41243
Ion Ratio Lower Upper
228 100
226 30.8 24.1 36.1
229 19.1 15.8 23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.650 ng
RT: 21.203 min Scan# 2414
Delta R.T. 0.006 min
Lab File: BN034687.D
Acq: 24 Oct 2024 10:59

Tgt Ion:149 Resp: 20342
Ion Ratio Lower Upper
149 100
167 25.0 19.7 29.5
279 1.7 1.8 2.6#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.467 min Scan# 2913

Delta R.T. 0.001 min

Lab File: BN034687.D


Acq: 24 Oct 2024 10:59

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8






Tgt Ion:264 Resp: 10772

Ion Ratio Lower Upper

264 100

260 27.0 21.9 32.9

265 70.3 60.6 91.0



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.828 ng

RT: 25.698 min Scan# 3673

Delta R.T. 0.004 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59




Tgt Ion:276 Resp: 34379


Ion Ratio Lower Upper

276 100

138 43.6 35.1 52.7

277 24.4 19.7 29.5





#37

Benzo(b)fluoranthene

Concen: 0.954 ng

RT: 22.850 min Scan# 2699

Delta R.T. 0.047 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:252 Resp: 37118

Ion Ratio Lower Upper

252 100

253 22.8 20.0 30.0

125 20.9 20.9 31.3



#38

Benzo(k)fluoranthene

Concen: 0.977 ng

RT: 22.850 min Scan# 2699

Delta R.T. 0.001 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59




Tgt Ion:252 Resp: 37118


Ion Ratio Lower Upper

252 100

253 22.8 20.6 31.0

125 20.9 21.6 32.4#





#39

Benzo(a)pyrene

Concen: 0.892 ng

RT: 23.371 min Scan# 21

Delta R.T. 0.004 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:252 Resp: 29490

Ion Ratio Lower Upper

252 100

253 23.4 22.2 33.4

125 23.9 26.0 39.0#



#40

Dibenzo(a,h)anthracene

Concen: 0.823 ng

RT: 25.718 min Scan# 3680

Delta R.T. 0.004 min

Lab File: BN034687.D

Acq: 24 Oct 2024 10:59




Tgt Ion:278 Resp: 26850


Ion Ratio Lower Upper

278 100

139 30.3 25.4 38.2


279 25.0 22.4 33.6







#41
 Benzo(g,h,i)perylene
 Concen: 0.824 ng
 RT: 26.373 min Scan# 3904
 Delta R.T. 0.001 min
 Lab File: BN034687.D
 Acq: 24 Oct 2024 10:59

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8



Tgt Ion:	276	Resp:	29665
Ion	Ratio	Lower	Upper
276	100		
277	24.2	21.3	31.9
138	38.6	31.9	47.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
 Data File : BN034691.D
 Acq On : 24 Oct 2024 14:14
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102424

Quant Time: Oct 24 14:43:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 24 13:17:02 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.705	152	6674	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	19848	0.400	ng	0.00
13) Acenaphthene-d10	14.329	164	10372	0.400	ng	0.00
19) Phenanthrene-d10	17.069	188	21015	0.400	ng	0.00
29) Chrysene-d12	21.250	240	12401	0.400	ng	0.00
35) Perylene-d12	23.466	264	10594	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.314	112	7964	0.401	ng	0.00
5) Phenol-d6	6.882	99	10366	0.396	ng	0.00
8) Nitrobenzene-d5	8.845	82	5758	0.346	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	9741	0.357	ng	0.00
14) 2,4,6-Tribromophenol	15.815	330	1099	0.343	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	14035	0.341	ng	0.00
27) Fluoranthene-d10	19.094	212	16566	0.350	ng	0.00
31) Terphenyl-d14	19.703	244	9400	0.371	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.285	88	3154	0.355	ng	100
3) n-Nitrosodimethylamine	3.581	42	4576	0.354	ng	# 97
6) bis(2-Chloroethyl)ether	7.141	93	7850	0.369	ng	99
9) Naphthalene	10.532	128	19666	0.358	ng	100
10) Hexachlorobutadiene	10.831	225	2964	0.360	ng	# 99
12) 2-Methylnaphthalene	12.143	142	12146	0.356	ng	98
16) Acenaphthylene	14.040	152	16384	0.329	ng	99
17) Acenaphthene	14.382	154	11510	0.335	ng	99
18) Fluorene	15.377	166	14579	0.344	ng	99
20) 4,6-Dinitro-2-methylph...	15.451	198	947	0.324	ng	88
21) 4-Bromophenyl-phenylether	16.275	248	3884	0.347	ng	98
22) Hexachlorobenzene	16.374	284	4355	0.358	ng	98
23) Atrazine	16.548	200	2928	0.333	ng	99
24) Pentachlorophenol	16.721	266	1463	0.340	ng	98
25) Phenanthrene	17.106	178	22550	0.357	ng	100
26) Anthracene	17.193	178	19340	0.345	ng	100
28) Fluoranthene	19.127	202	22945	0.348	ng	100
30) Pyrene	19.489	202	23297	0.366	ng	100
32) Benzo(a)anthracene	21.232	228	16543	0.349	ng	99
33) Chrysene	21.286	228	17696	0.361	ng	99
34) Bis(2-ethylhexyl)phtha...	21.196	149	8375	0.312	ng	100
36) Indeno(1,2,3-cd)pyrene	25.697	276	14157	0.347	ng	100
37) Benzo(b)fluoranthene	22.803	252	15350	0.363	ng	99
38) Benzo(k)fluoranthene	22.850	252	14742	0.353	ng	99
39) Benzo(a)pyrene	23.373	252	12054	0.355	ng	99
40) Dibenzo(a,h)anthracene	25.718	278	11142	0.346	ng	99
41) Benzo(g,h,i)perylene	26.373	276	12576	0.354	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034691.D
Acq On : 24 Oct 2024 14:14
Operator : RC/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN102424


Quant Time: Oct 24 14:43:51 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.705 min Scan# 6
 Delta R.T. -0.007 min
 Lab File: BN034691.D
 Acq: 24 Oct 2024 14:14


Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102424





Tgt Ion:152 Resp: 6674
 Ion Ratio Lower Upper
 152 100
 150 157.2 122.7 184.1
 115 65.2 51.0 76.6




#2
 1,4-Dioxane
 Concen: 0.355 ng
 RT: 3.285 min Scan# 56
 Delta R.T. -0.000 min
 Lab File: BN034691.D
 Acq: 24 Oct 2024 14:14



Tgt Ion: 88 Resp: 3154
 Ion Ratio Lower Upper
 88 100
 43 39.5 32.1 48.1
 58 91.9 73.7 110.5









#3
n-Nitrosodimethylamine
Concen: 0.354 ng
RT: 3.581 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424


Tgt Ion: 42 Resp: 4576
Ion Ratio Lower Upper
42 100
74 98.6 77.2 115.8
44 20.5 13.4 20.2#



#4
2-Fluorophenol
Concen: 0.401 ng
RT: 5.314 min Scan# 337
Delta R.T. -0.007 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14


Tgt Ion: 112 Resp: 7964
Ion Ratio Lower Upper
112 100
64 66.0 53.0 79.6
63 35.2 27.6 41.4








#5
Phenol-d6
Concen: 0.396 ng
RT: 6.882 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14


Instrument :
BNA_N
ClientSampleId :
ICVBN102424





Tgt Ion: 99 Resp: 10366
Ion Ratio Lower Upper
99 100
42 27.0 21.8 32.8
71 33.3 26.6 40.0




#6
bis(2-Chloroethyl)ether
Concen: 0.369 ng
RT: 7.141 min Scan# 590
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14



Tgt Ion: 93 Resp: 7850
Ion Ratio Lower Upper
93 100
63 87.8 71.5 107.3
95 32.2 25.8 38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 955
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424


Tgt Ion:	136	Resp:	19848
Ion Ratio	Lower	Upper	
136	100		
137	11.3	9.1	13.7
54	11.4	9.8	14.8
68	7.5	6.4	9.6



#8
Nitrobenzene-d5
Concen: 0.346 ng
RT: 8.845 min Scan# 802
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion:	82	Resp:	5758
Ion Ratio	Lower	Upper	
82	100		
128	39.9	29.7	44.5
54	63.8	53.0	79.6









#9
Naphthalene
Concen: 0.358 ng
RT: 10.532 min Scan# 901
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424


Tgt Ion:128 Resp: 19666
Ion Ratio Lower Upper
128 100
129 11.4 9.0 13.6
127 12.8 10.5 15.7



#10
Hexachlorobutadiene
Concen: 0.360 ng
RT: 10.831 min Scan# 988
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion:225 Resp: 2964
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.4 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 0.357 ng
RT: 12.071 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424


Tgt Ion:152 Resp: 9741
Ion Ratio Lower Upper
152 100
151 20.7 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.356 ng
RT: 12.143 min Scan# 1220
Delta R.T. -0.004 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14


Tgt Ion:142 Resp: 12146
Ion Ratio Lower Upper
142 100
141 89.4 70.4 105.6
115 38.7 29.8 44.6








#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.329 min Scan# 1504
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Instrument :
BNA_N
ClientSampleId :
ICVBN102424







Tgt Ion:164 Resp: 10372
Ion Ratio Lower Upper
164 100
162 101.1 81.4 122.0
160 50.2 40.6 61.0



#14
2,4,6-Tribromophenol
Concen: 0.343 ng
RT: 15.815 min Scan# 1639
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion:330 Resp: 1099
Ion Ratio Lower Upper
330 100
332 96.5 76.6 115.0
141 66.2 54.6 81.8









#15
2-Fluorobiphenyl
Concen: 0.341 ng
RT: 12.950 min Scan# 14035
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424


Tgt Ion:	172	Resp:	14035
Ion Ratio	Lower	Upper	
172	100		
171	35.0	28.3	42.5
170	24.5	19.8	29.6



#16
Acenaphthylene
Concen: 0.329 ng
RT: 14.040 min Scan# 1477
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion:	152	Resp:	16384
Ion Ratio	Lower	Upper	
152	100		
151	19.0	15.6	23.4
153	13.1	10.6	15.8









#17
Acenaphthene
Concen: 0.335 ng
RT: 14.382 min Scan# 1509
Delta R.T. -0.011 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424


Tgt Ion:154 Resp: 11510
Ion Ratio Lower Upper
154 100
153 111.6 88.2 132.2
152 58.6 46.9 70.3



#18
Fluorene
Concen: 0.344 ng
RT: 15.377 min Scan# 1602
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion:166 Resp: 14579
Ion Ratio Lower Upper
166 100
165 99.0 78.8 118.2
167 13.5 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.069 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14

Instrument :

BNA_N

ClientSampleId :

ICVBN102424





Tgt Ion:188 Resp: 21015

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.5 8.9 13.3



#20

4,6-Dinitro-2-methylphenol

Concen: 0.324 ng

RT: 15.451 min Scan# 1609

Delta R.T. -0.001 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14




Tgt Ion:198 Resp: 947


Ion Ratio Lower Upper

198 100

51 100.2 91.8 137.8

105 48.7 44.3 66.5









#21
4-Bromophenyl-phenylether
Concen: 0.347 ng
RT: 16.275 min Scan# 1676
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424


Tgt Ion:248 Resp: 3884
Ion Ratio Lower Upper
248 100
250 103.8 80.9 121.3
141 50.3 40.5 60.7



#22
Hexachlorobenzene
Concen: 0.358 ng
RT: 16.374 min Scan# 1684
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion:284 Resp: 4355
Ion Ratio Lower Upper
284 100
142 55.6 43.6 65.4
249 34.2 26.6 40.0









#23
Atrazine
Concen: 0.333 ng
RT: 16.548 min Scan# 1698
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424


Tgt Ion:200 Resp: 2928
Ion Ratio Lower Upper
200 100
173 25.5 19.7 29.5
215 45.1 35.7 53.5



#24
Pentachlorophenol
Concen: 0.340 ng
RT: 16.721 min Scan# 1712
Delta R.T. 0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion:266 Resp: 1463
Ion Ratio Lower Upper
266 100
264 61.6 49.4 74.2
268 61.0 50.5 75.7









#25
Phenanthrene
Concen: 0.357 ng
RT: 17.106 min Scan# 1743
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424


Tgt Ion:178 Resp: 22550
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.2
179 15.4 12.2 18.4



#26
Anthracene
Concen: 0.345 ng
RT: 17.193 min Scan# 1750
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion:178 Resp: 19340
Ion Ratio Lower Upper
178 100
176 18.5 14.9 22.3
179 15.1 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.350 ng
RT: 19.094 min Scan# 1982
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424


Tgt Ion:212 Resp: 16566
Ion Ratio Lower Upper
212 100
106 22.2 17.8 26.8
104 13.1 10.2 15.4



#28
Fluoranthene
Concen: 0.348 ng
RT: 19.127 min Scan# 1989
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion:202 Resp: 22945
Ion Ratio Lower Upper
202 100
101 16.0 12.8 19.2
203 17.1 13.5 20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.250 min Scan# 2421
Delta R.T. -0.000 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14




Instrument :
BNA_N
ClientSampleId :
ICVBN102424


Tgt Ion:240 Resp: 12401
Ion Ratio Lower Upper
240 100
120 18.6 14.2 21.2
236 28.6 23.1 34.7



#30
Pyrene
Concen: 0.366 ng
RT: 19.489 min Scan# 2067
Delta R.T. 0.005 min
Lab File: BN034691.D
Acq: 24 Oct 2024 14:14

Tgt Ion:202 Resp: 23297
Ion Ratio Lower Upper
202 100
200 20.9 16.6 25.0
203 17.7 14.2 21.4









#31
 Terphenyl-d14
 Concen: 0.371 ng
 RT: 19.703 min Scan# 2113
 Delta R.T. -0.000 min
 Lab File: BN034691.D
 Acq: 24 Oct 2024 14:14




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102424


Tgt Ion:244 Resp: 9400
 Ion Ratio Lower Upper
 244 100
 212 11.1 8.6 13.0
 122 27.4 22.3 33.5



#32
 Benzo(a)anthracene
 Concen: 0.349 ng
 RT: 21.232 min Scan# 2419
 Delta R.T. -0.000 min
 Lab File: BN034691.D
 Acq: 24 Oct 2024 14:14

Tgt Ion:228 Resp: 16543
 Ion Ratio Lower Upper
 228 100
 226 28.1 22.0 33.0
 229 19.1 15.6 23.4









#33
 Chrysene
 Concen: 0.361 ng
 RT: 21.286 min Scan# 2425
 Delta R.T. -0.000 min
 Lab File: BN034691.D
 Acq: 24 Oct 2024 14:14




Instrument : BNA_N
 ClientSampleId : ICVBN102424


Tgt Ion:	228	Resp:	17696
Ion Ratio	Lower	Upper	
228	100		
226	30.6	24.1	36.1
229	19.6	15.8	23.8



#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.312 ng
 RT: 21.196 min Scan# 2415
 Delta R.T. 0.000 min
 Lab File: BN034691.D
 Acq: 24 Oct 2024 14:14


Tgt Ion:	149	Resp:	8375
Ion Ratio	Lower	Upper	
149	100		
167	24.7	19.7	29.5
279	2.3	1.8	2.6








#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.466 min Scan# 2913
 Delta R.T. -0.000 min
 Lab File: BN034691.D
 Acq: 24 Oct 2024 14:14

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102424



Tgt Ion: 264 Resp: 10594




Ion	Ratio	Lower	Upper
264	100		
260	27.4	21.9	32.9
265	72.9	60.6	91.0




#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.347 ng
 RT: 25.697 min Scan# 3676
 Delta R.T. 0.003 min
 Lab File: BN034691.D
 Acq: 24 Oct 2024 14:14

Tgt Ion: 276 Resp: 14157

Ion	Ratio	Lower	Upper
276	100		
138	43.9	35.1	52.7
277	25.2	19.7	29.5





#37

Benzo(b)fluoranthene

Concen: 0.363 ng

RT: 22.803 min Scan# 2

Delta R.T. -0.000 min

Lab File: BN034691.D


Acq: 24 Oct 2024 14:14

Instrument :

BNA_N

ClientSampleId :

ICVBN102424





Tgt Ion:252 Resp: 15350

Ion Ratio Lower Upper

252 100

253 24.9 20.0 30.0

125 25.2 20.9 31.3



#38

Benzo(k)fluoranthene

Concen: 0.353 ng

RT: 22.850 min Scan# 2702

Delta R.T. -0.000 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14




Tgt Ion:252 Resp: 14742


Ion Ratio Lower Upper

252 100

253 25.6 20.6 31.0

125 26.3 21.6 32.4





#39

Benzo(a)pyrene

Concen: 0.355 ng

RT: 23.373 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14

Instrument :

BNA_N

ClientSampleId :

ICVBN102424





Tgt Ion:252 Resp: 12054

Ion Ratio Lower Upper

252 100

253 26.9 22.2 33.4

125 32.0 26.0 39.0



#40

Dibenzo(a,h)anthracene

Concen: 0.346 ng

RT: 25.718 min Scan# 3683

Delta R.T. 0.003 min

Lab File: BN034691.D

Acq: 24 Oct 2024 14:14




Tgt Ion:278 Resp: 11142


Ion Ratio Lower Upper

278 100

139 33.0 25.4 38.2


279 28.2 22.4 33.6







#41
 Benzo(g,h,i)perylene
 Concen: 0.354 ng
 RT: 26.373 min Scan# 30
 Delta R.T. -0.000 min
 Lab File: BN034691.D
 Acq: 24 Oct 2024 14:14

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102424



Tgt Ion:	276	Resp:	12576
Ion Ratio	Lower	Upper	
276	100		
277	25.5	21.3	31.9
138	40.2	31.9	47.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
 Data File : BN034706.D
 Acq On : 25 Oct 2024 00:02
 Operator : RC/JU
 Sample : PB164282BS
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB164282BS

Quant Time: Oct 25 02:08:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 24 13:17:02 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.712	152	6277	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	17435	0.400	ng	0.00
13) Acenaphthene-d10	14.318	164	7557	0.400	ng	-0.01
19) Phenanthrene-d10	17.056	188	13505	0.400	ng	#-0.01
29) Chrysene-d12	21.241	240	7661	0.400	ng	# 0.00
35) Perylene-d12	23.461	264	6455	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.321	112	5611	0.301	ng	0.00
5) Phenol-d6	6.881	99	7614	0.309	ng	0.00
8) Nitrobenzene-d5	8.845	82	4924	0.337	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	8357	0.349	ng	0.00
14) 2,4,6-Tribromophenol	15.815	330	563	0.241	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	10877	0.362	ng	0.00
27) Fluoranthene-d10	19.094	212	10197	0.335	ng	0.00
31) Terphenyl-d14	19.703	244	5692	0.364	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.285	88	2475	0.296	ng	# 46
3) n-Nitrosodimethylamine	3.581	42	4401	0.362	ng	96
6) bis(2-Chloroethyl)ether	7.141	93	7177	0.359	ng	99
9) Naphthalene	10.532	128	16863	0.350	ng	100
10) Hexachlorobutadiene	10.831	225	2523	0.349	ng	# 98
12) 2-Methylnaphthalene	12.143	142	10035	0.335	ng	99
16) Acenaphthylene	14.040	152	12822	0.353	ng	100
17) Acenaphthene	14.382	154	8528	0.341	ng	98
18) Fluorene	15.377	166	9962	0.323	ng	99
20) 4,6-Dinitro-2-methylph...	15.441	198	538	0.286	ng	# 56
21) 4-Bromophenyl-phenylether	16.262	248	2561	0.356	ng	# 61
22) Hexachlorobenzene	16.374	284	2810	0.360	ng	97
23) Atrazine	16.535	200	1814	0.321	ng	# 88
24) Pentachlorophenol	16.721	266	1431	0.518	ng	98
25) Phenanthrene	17.106	178	14856	0.366	ng	100
26) Anthracene	17.193	178	13024	0.362	ng	100
28) Fluoranthene	19.122	202	14119	0.334	ng	99
30) Pyrene	19.484	202	14340	0.364	ng	99
32) Benzo(a)anthracene	21.232	228	10619	0.363	ng	99
33) Chrysene	21.277	228	11699	0.386	ng	98
34) Bis(2-ethylhexyl)phtha...	21.187	149	5670	0.341	ng	# 100
36) Indeno(1,2,3-cd)pyrene	25.688	276	10806	0.435	ng	98
37) Benzo(b)fluoranthene	22.797	252	10399	0.403	ng	94
38) Benzo(k)fluoranthene	22.841	252	10151	0.399	ng	96
39) Benzo(a)pyrene	23.364	252	8689	0.420	ng	95
40) Dibenzo(a,h)anthracene	25.706	278	8269	0.422	ng	97
41) Benzo(g,h,i)perylene	26.361	276	8830	0.408	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034706.D
Acq On : 25 Oct 2024 00:02
Operator : RC/JU
Sample : PB164282BS
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB164282BS


Quant Time: Oct 25 02:08:50 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.712 min Scan# 669
 Delta R.T. -0.000 min
 Lab File: BN034706.D
 Acq: 25 Oct 2024 00:02

Instrument :
 BNA_N
 ClientSampleId :
 PB164282BS



Tgt Ion:152 Resp: 6277




Ion	Ratio	Lower	Upper
152	100		
150	153.7	122.7	184.1
115	62.4	51.0	76.6




#2
 1,4-Dioxane
 Concen: 0.296 ng
 RT: 3.285 min Scan# 56
 Delta R.T. -0.000 min
 Lab File: BN034706.D
 Acq: 25 Oct 2024 00:02

Tgt Ion: 88 Resp: 2475

Ion	Ratio	Lower	Upper
88	100		
43	124.6	32.1	48.1
58	109.9	73.7	110.5









#3
 n-Nitrosodimethylamine
 Concen: 0.362 ng
 RT: 3.581 min Scan# 91
 Delta R.T. -0.000 min
 Lab File: BN034706.D
 Acq: 25 Oct 2024 00:02




Instrument :
 BNA_N
 ClientSampleId :
 PB164282BS


Tgt Ion: 42 Resp: 4401
 Ion Ratio Lower Upper
 42 100
 74 92.1 77.2 115.8
 44 14.7 13.4 20.2



#4
 2-Fluorophenol
 Concen: 0.301 ng
 RT: 5.321 min Scan# 338
 Delta R.T. -0.000 min
 Lab File: BN034706.D
 Acq: 25 Oct 2024 00:02

Tgt Ion: 112 Resp: 5611
 Ion Ratio Lower Upper
 112 100
 64 66.5 53.0 79.6
 63 35.5 27.6 41.4









#5
Phenol-d6
Concen: 0.309 ng
RT: 6.881 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS


Tgt Ion: 99 Resp: 7614
Ion Ratio Lower Upper
99 100
42 28.9 21.8 32.8
71 33.7 26.6 40.0



#6
bis(2-Chloroethyl)ether
Concen: 0.359 ng
RT: 7.141 min Scan# 590
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion: 93 Resp: 7177
Ion Ratio Lower Upper
93 100
63 90.6 71.5 107.3
95 31.9 25.8 38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS


Tgt Ion:	136	Resp:	17435
Ion	Ratio	Lower	Upper
136	100		
137	11.4	9.1	13.7
54	11.4	9.8	14.8
68	7.4	6.4	9.6



#8
Nitrobenzene-d5
Concen: 0.337 ng
RT: 8.845 min Scan# 802
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion:	82	Resp:	4924
Ion	Ratio	Lower	Upper
82	100		
128	36.9	29.7	44.5
54	64.3	53.0	79.6









#9
Naphthalene
Concen: 0.350 ng
RT: 10.532 min Scan# 960
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS


Tgt Ion:128 Resp: 16863
Ion Ratio Lower Upper
128 100
129 11.4 9.0 13.6
127 13.1 10.5 15.7



#10
Hexachlorobutadiene
Concen: 0.349 ng
RT: 10.831 min Scan# 988
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion:225 Resp: 2523
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.0 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 0.349 ng
RT: 12.071 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS


Tgt Ion:152 Resp: 8357
Ion Ratio Lower Upper
152 100
151 20.0 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.335 ng
RT: 12.143 min Scan# 1220
Delta R.T. -0.004 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion:142 Resp: 10035
Ion Ratio Lower Upper
142 100
141 88.4 70.4 105.6
115 38.2 29.8 44.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.318 min Scan# 1503
Delta R.T. -0.011 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS


Tgt Ion:164 Resp: 7557
Ion Ratio Lower Upper
164 100
162 106.1 81.4 122.0
160 55.4 40.6 61.0



#14
2,4,6-Tribromophenol
Concen: 0.241 ng
RT: 15.815 min Scan# 1639
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion:330 Resp: 563
Ion Ratio Lower Upper
330 100
332 94.7 76.6 115.0
141 73.7 54.6 81.8









#15
2-Fluorobiphenyl
Concen: 0.362 ng
RT: 12.950 min Scan# 1375
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS


Tgt Ion:	172	Resp:	10877
Ion Ratio	Lower	Upper	
172	100		
171	34.6	28.3	42.5
170	23.9	19.8	29.6



#16
Acenaphthylene
Concen: 0.353 ng
RT: 14.040 min Scan# 1477
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02


Tgt Ion:	152	Resp:	12822
Ion Ratio	Lower	Upper	
152	100		
151	19.5	15.6	23.4
153	13.1	10.6	15.8








#17
Acenaphthene
Concen: 0.341 ng
RT: 14.382 min Scan# 1509
Delta R.T. -0.011 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Instrument :
BNA_N
ClientSampleId :
PB164282BS







Tgt Ion:154 Resp: 8528
Ion Ratio Lower Upper
154 100
153 112.4 88.2 132.2
152 60.6 46.9 70.3



#18
Fluorene
Concen: 0.323 ng
RT: 15.377 min Scan# 1602
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion:166 Resp: 9962
Ion Ratio Lower Upper
166 100
165 97.3 78.8 118.2
167 13.6 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.056 min Scan# 11
Delta R.T. -0.013 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS


Tgt Ion:188 Resp: 13505
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 22.2 8.9 13.3#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.286 ng
RT: 15.441 min Scan# 1608
Delta R.T. -0.011 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion:198 Resp: 538
Ion Ratio Lower Upper
198 100
51 160.2 91.8 137.8#
105 91.1 44.3 66.5#









#21
4-Bromophenyl-phenylether
Concen: 0.356 ng
RT: 16.262 min Scan# 1675
Delta R.T. -0.013 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS


Tgt Ion:248 Resp: 2561
Ion Ratio Lower Upper
248 100
250 88.5 80.9 121.3
141 113.9 40.5 60.7#



#22
Hexachlorobenzene
Concen: 0.360 ng
RT: 16.374 min Scan# 1684
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion:284 Resp: 2810
Ion Ratio Lower Upper
284 100
142 57.3 43.6 65.4
249 34.1 26.6 40.0









#23
Atrazine
Concen: 0.321 ng
RT: 16.535 min Scan# 1697
Delta R.T. -0.013 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS


Tgt Ion:200 Resp: 1814
Ion Ratio Lower Upper
200 100
173 40.1 19.7 29.5#
215 43.2 35.7 53.5



#24
Pentachlorophenol
Concen: 0.518 ng
RT: 16.721 min Scan# 1712
Delta R.T. 0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion:266 Resp: 1431
Ion Ratio Lower Upper
266 100
264 63.0 49.4 74.2
268 60.7 50.5 75.7









#25
 Phenanthrene
 Concen: 0.366 ng
 RT: 17.106 min Scan# 1743
 Delta R.T. -0.000 min
 Lab File: BN034706.D
 Acq: 25 Oct 2024 00:02




Instrument :
 BNA_N
 ClientSampleId :
 PB164282BS


Tgt Ion:178	Resp:	14856
Ion Ratio	Lower	Upper
178	100	
176	19.3	15.4 23.2
179	15.3	12.2 18.4



#26
 Anthracene
 Concen: 0.362 ng
 RT: 17.193 min Scan# 1750
 Delta R.T. -0.000 min
 Lab File: BN034706.D
 Acq: 25 Oct 2024 00:02

Tgt Ion:178	Resp:	13024
Ion Ratio	Lower	Upper
178	100	
176	18.8	14.9 22.3
179	15.3	12.2 18.2









#27
Fluoranthene-d10
Concen: 0.335 ng
RT: 19.094 min Scan# 1982
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS


Tgt Ion:212 Resp: 10197
Ion Ratio Lower Upper
212 100
106 23.4 17.8 26.8
104 13.7 10.2 15.4



#28
Fluoranthene
Concen: 0.334 ng
RT: 19.122 min Scan# 1988
Delta R.T. -0.005 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion:202 Resp: 14119
Ion Ratio Lower Upper
202 100
101 16.4 12.8 19.2
203 17.0 13.5 20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.241 min Scan# 2421
Delta R.T. -0.009 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS


Tgt Ion:	240	Resp:	7661
Ion Ratio	Lower	Upper	
240	100		
120	28.4	14.2	21.2#
236	30.8	23.1	34.7



#30
Pyrene
Concen: 0.364 ng
RT: 19.484 min Scan# 2066
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion:	202	Resp:	14340
Ion Ratio	Lower	Upper	
202	100		
200	21.0	16.6	25.0
203	17.5	14.2	21.4









#31
Terphenyl-d14
Concen: 0.364 ng
RT: 19.703 min Scan# 2113
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02




Instrument :
BNA_N
ClientSampleId :
PB164282BS


Tgt Ion:244 Resp: 5692
Ion Ratio Lower Upper
244 100
212 11.7 8.6 13.0
122 27.9 22.3 33.5



#32
Benzo(a)anthracene
Concen: 0.363 ng
RT: 21.232 min Scan# 2419
Delta R.T. -0.000 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion:228 Resp: 10619
Ion Ratio Lower Upper
228 100
226 27.1 22.0 33.0
229 20.2 15.6 23.4









#33
 Chrysene
 Concen: 0.386 ng
 RT: 21.277 min Scan# 2425
 Delta R.T. -0.009 min
 Lab File: BN034706.D
 Acq: 25 Oct 2024 00:02




Instrument :
 BNA_N
 ClientSampleId :
 PB164282BS


Tgt Ion:	228	Resp:	11699
Ion Ratio	Lower	Upper	
228	100		
226	31.3	24.1	36.1
229	19.3	15.8	23.8



#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.341 ng
 RT: 21.187 min Scan# 2414
 Delta R.T. -0.009 min
 Lab File: BN034706.D
 Acq: 25 Oct 2024 00:02


Tgt Ion:	149	Resp:	5670
Ion Ratio	Lower	Upper	
149	100		
167	24.5	19.7	29.5
279	1.7	1.8	2.6








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.461 min Scan# 2911
Delta R.T. -0.006 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Instrument :
BNA_N
ClientSampleId :
PB164282BS







Tgt Ion:264 Resp: 6455
Ion Ratio Lower Upper
264 100
260 28.1 21.9 32.9
265 95.5 60.6 91.0#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.435 ng
RT: 25.688 min Scan# 3673
Delta R.T. -0.006 min
Lab File: BN034706.D
Acq: 25 Oct 2024 00:02

Tgt Ion:276 Resp: 10806
Ion Ratio Lower Upper
276 100
138 46.0 35.1 52.7
277 24.3 19.7 29.5





#37

Benzo(b)fluoranthene

Concen: 0.403 ng

RT: 22.797 min Scan# 2684

Delta R.T. -0.006 min

Lab File: BN034706.D


Acq: 25 Oct 2024 00:02

Instrument :

BNA_N

ClientSampleId :

PB164282BS





Tgt Ion:252 Resp: 10399

Ion Ratio Lower Upper

252 100

253 26.8 20.0 30.0

125 30.7 20.9 31.3



#38

Benzo(k)fluoranthene

Concen: 0.399 ng

RT: 22.841 min Scan# 2699

Delta R.T. -0.009 min

Lab File: BN034706.D

Acq: 25 Oct 2024 00:02




Tgt Ion:252 Resp: 10151


Ion Ratio Lower Upper

252 100

253 26.4 20.6 31.0

125 30.7 21.6 32.4





#39

Benzo(a)pyrene

Concen: 0.420 ng

RT: 23.364 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN034706.D

Acq: 25 Oct 2024 00:02

Instrument :

BNA_N

ClientSampleId :

PB164282BS





Tgt Ion:252 Resp: 8689

Ion Ratio Lower Upper

252 100

253 28.5 22.2 33.4

125 37.5 26.0 39.0



#40

Dibenzo(a,h)anthracene

Concen: 0.422 ng

RT: 25.706 min Scan# 3679

Delta R.T. -0.009 min

Lab File: BN034706.D

Acq: 25 Oct 2024 00:02




Tgt Ion:278 Resp: 8269


Ion Ratio Lower Upper

278 100

139 34.5 25.4 38.2

279 27.9 22.4 33.6





#41

Benzo(g,h,i)perylene

Concen: 0.408 ng

RT: 26.361 min Scan# 3903

Delta R.T. -0.012 min

Lab File: BN034706.D


Acq: 25 Oct 2024 00:02

Instrument :

BNA_N

ClientSampleId :

PB164282BS





Tgt Ion:276 Resp: 8830

Ion Ratio Lower Upper

276 100

277 27.0 21.3 31.9

138 42.2 31.9 47.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
 Data File : BN034711.D
 Acq On : 25 Oct 2024 03:02
 Operator : RC/JU
 Sample : PB164282BSD
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB164282BSD

Quant Time: Oct 25 04:20:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 24 13:17:02 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.712	152	5974	0.400	ng	0.00
7) Naphthalene-d8	10.479	136	16564	0.400	ng	0.00
13) Acenaphthene-d10	14.329	164	7143	0.400	ng	0.00
19) Phenanthrene-d10	17.057	188	13002	0.400	ng	#-0.01
29) Chrysene-d12	21.241	240	7027	0.400	ng	# 0.00
35) Perylene-d12	23.458	264	5873	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.322	112	5283	0.297	ng	0.00
5) Phenol-d6	6.882	99	7080	0.302	ng	0.00
8) Nitrobenzene-d5	8.845	82	4660	0.335	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	7875	0.346	ng	0.00
14) 2,4,6-Tribromophenol	15.815	330	521	0.236	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	10106	0.356	ng	0.00
27) Fluoranthene-d10	19.094	212	9709	0.332	ng	0.00
31) Terphenyl-d14	19.703	244	5323	0.371	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.285	88	2435	0.306	ng	# 53
3) n-Nitrosodimethylamine	3.581	42	4293	0.371	ng	96
6) bis(2-Chloroethyl)ether	7.142	93	6795	0.357	ng	99
9) Naphthalene	10.532	128	15953	0.348	ng	99
10) Hexachlorobutadiene	10.831	225	2383	0.347	ng	# 97
12) 2-Methylnaphthalene	12.143	142	9409	0.331	ng	98
16) Acenaphthylene	14.040	152	12085	0.352	ng	99
17) Acenaphthene	14.382	154	7997	0.338	ng	98
18) Fluorene	15.377	166	9417	0.323	ng	100
20) 4,6-Dinitro-2-methylph...	15.441	198	532	0.294	ng	# 60
21) 4-Bromophenyl-phenylether	16.262	248	2358	0.340	ng	# 62
22) Hexachlorobenzene	16.374	284	2677	0.356	ng	98
23) Atrazine	16.535	200	1782	0.327	ng	# 89
24) Pentachlorophenol	16.721	266	1302	0.490	ng	99
25) Phenanthrene	17.106	178	14207	0.364	ng	100
26) Anthracene	17.193	178	12435	0.359	ng	100
28) Fluoranthene	19.122	202	13425	0.329	ng	100
30) Pyrene	19.484	202	13610	0.377	ng	99
32) Benzo(a)anthracene	21.232	228	9893	0.369	ng	99
33) Chrysene	21.277	228	11013	0.396	ng	98
34) Bis(2-ethylhexyl)phtha...	21.187	149	5309	0.349	ng	# 96
36) Indeno(1,2,3-cd)pyrene	25.683	276	9709	0.429	ng	96
37) Benzo(b)fluoranthene	22.797	252	9630	0.411	ng	94
38) Benzo(k)fluoranthene	22.797	252	9630	0.416	ng	95
39) Benzo(a)pyrene	23.367	252	8133	0.432	ng	# 93
40) Dibenzo(a,h)anthracene	25.703	278	7436	0.417	ng	96
41) Benzo(g,h,i)perylene	26.364	276	7954	0.404	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102424\
Data File : BN034711.D
Acq On : 25 Oct 2024 03:02
Operator : RC/JU
Sample : PB164282BSD
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB164282BSD

Quant Time: Oct 25 04:20:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 24 13:17:02 2024
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.712 min Scan# 669
 Delta R.T. 0.000 min
 Lab File: BN034711.D
 Acq: 25 Oct 2024 03:02




Instrument :
 BNA_N
 ClientSampleId :
 PB164282BSD


Tgt Ion:152 Resp: 5974
 Ion Ratio Lower Upper
 152 100
 150 152.2 122.7 184.1
 115 63.6 51.0 76.6



#2
 1,4-Dioxane
 Concen: 0.306 ng
 RT: 3.285 min Scan# 56
 Delta R.T. 0.000 min
 Lab File: BN034711.D
 Acq: 25 Oct 2024 03:02

Tgt Ion: 88 Resp: 2435
 Ion Ratio Lower Upper
 88 100
 43 114.0 32.1 48.1#
 58 107.2 73.7 110.5









#3
n-Nitrosodimethylamine
Concen: 0.371 ng
RT: 3.581 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD


Tgt Ion: 42 Resp: 4293
Ion Ratio Lower Upper
42 100
74 92.4 77.2 115.8
44 15.0 13.4 20.2



#4
2-Fluorophenol
Concen: 0.297 ng
RT: 5.322 min Scan# 338
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Tgt Ion: 112 Resp: 5283
Ion Ratio Lower Upper
112 100
64 67.3 53.0 79.6
63 35.9 27.6 41.4









#5
Phenol-d6
Concen: 0.302 ng
RT: 6.882 min Scan# 511
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD


Tgt Ion:	99	Resp:	7080
Ion	Ratio	Lower	Upper
99	100		
42	30.9	21.8	32.8
71	33.8	26.6	40.0



#6
bis(2-Chloroethyl)ether
Concen: 0.357 ng
RT: 7.142 min Scan# 590
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Tgt Ion:	93	Resp:	6795
Ion	Ratio	Lower	Upper
93	100		
63	90.5	71.5	107.3
95	32.1	25.8	38.6









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.479 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD


Tgt Ion:	136	Resp:	16564
Ion	Ratio	Lower	Upper
136	100		
137	11.6	9.1	13.7
54	11.1	9.8	14.8
68	7.5	6.4	9.6



#8
Nitrobenzene-d5
Concen: 0.335 ng
RT: 8.845 min Scan# 802
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Tgt Ion:	82	Resp:	4660
Ion	Ratio	Lower	Upper
82	100		
128	37.7	29.7	44.5
54	64.1	53.0	79.6









#9
Naphthalene
Concen: 0.348 ng
RT: 10.532 min Scan# 960
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD


Tgt Ion:128 Resp: 15953
Ion Ratio Lower Upper
128 100
129 11.4 9.0 13.6
127 13.4 10.5 15.7



#10
Hexachlorobutadiene
Concen: 0.347 ng
RT: 10.831 min Scan# 988
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Tgt Ion:225 Resp: 2383
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 62.1 51.4 77.0









#11
2-Methylnaphthalene-d10
Concen: 0.346 ng
RT: 12.071 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD


Tgt Ion:152 Resp: 7875
Ion Ratio Lower Upper
152 100
151 20.3 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.331 ng
RT: 12.143 min Scan# 1220
Delta R.T. -0.004 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Tgt Ion:142 Resp: 9409
Ion Ratio Lower Upper
142 100
141 87.4 70.4 105.6
115 39.3 29.8 44.6





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.329 min Scan# 1504

Delta R.T. 0.000 min

Lab File: BN034711.D

Acq: 25 Oct 2024 03:02

Instrument :

BNA_N

ClientSampleId :

PB164282BSD





Tgt Ion:164 Resp: 7143

Ion Ratio Lower Upper

164 100

162 100.9 81.4 122.0

160 51.4 40.6 61.0



#14

2,4,6-Tribromophenol

Concen: 0.236 ng

RT: 15.815 min Scan# 1639

Delta R.T. 0.000 min

Lab File: BN034711.D

Acq: 25 Oct 2024 03:02




Tgt Ion:330 Resp: 521


Ion Ratio Lower Upper

330 100

332 93.5 76.6 115.0

141 77.7 54.6 81.8









#15
2-Fluorobiphenyl
Concen: 0.356 ng
RT: 12.950 min Scan# 1375
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD


Tgt Ion:172 Resp: 10106
Ion Ratio Lower Upper
172 100
171 35.5 28.3 42.5
170 24.7 19.8 29.6



#16
Acenaphthylene
Concen: 0.352 ng
RT: 14.040 min Scan# 1477
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Tgt Ion:152 Resp: 12085
Ion Ratio Lower Upper
152 100
151 19.1 15.6 23.4
153 12.9 10.6 15.8









#17
Acenaphthene
Concen: 0.338 ng
RT: 14.382 min Scan# 1509
Delta R.T. -0.011 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD


Tgt Ion:154 Resp: 7997
Ion Ratio Lower Upper
154 100
153 111.9 88.2 132.2
152 60.7 46.9 70.3



#18
Fluorene
Concen: 0.323 ng
RT: 15.377 min Scan# 1602
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Tgt Ion:166 Resp: 9417
Ion Ratio Lower Upper
166 100
165 98.4 78.8 118.2
167 13.2 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.057 min Scan# 11
Delta R.T. -0.012 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD


Tgt Ion:188 Resp: 13002
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 23.0 8.9 13.3#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.294 ng
RT: 15.441 min Scan# 1608
Delta R.T. -0.011 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Tgt Ion:198 Resp: 532
Ion Ratio Lower Upper
198 100
51 156.1 91.8 137.8#
105 88.2 44.3 66.5#









#21
4-Bromophenyl-phenylether
Concen: 0.340 ng
RT: 16.262 min Scan# 1675
Delta R.T. -0.012 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD


Tgt Ion:248 Resp: 2358
Ion Ratio Lower Upper
248 100
250 91.6 80.9 121.3
141 116.5 40.5 60.7#



#22
Hexachlorobenzene
Concen: 0.356 ng
RT: 16.374 min Scan# 1684
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Tgt Ion:284 Resp: 2677
Ion Ratio Lower Upper
284 100
142 56.9 43.6 65.4
249 33.8 26.6 40.0





#23

Atrazine

Concen: 0.327 ng

RT: 16.535 min Scan# 1

Delta R.T. -0.012 min

Lab File: BN034711.D

Acq: 25 Oct 2024 03:02

Instrument :

BNA_N

ClientSampleId :

PB164282BSD





Tgt Ion:200 Resp: 1782

Ion Ratio Lower Upper

200 100

173 36.2 19.7 29.5#

215 41.4 35.7 53.5



#24

Pentachlorophenol

Concen: 0.490 ng

RT: 16.721 min Scan# 1712

Delta R.T. 0.000 min

Lab File: BN034711.D

Acq: 25 Oct 2024 03:02




Tgt Ion:266 Resp: 1302


Ion Ratio Lower Upper

266 100

264 62.4 49.4 74.2

268 62.1 50.5 75.7









#25
Phenanthrene
Concen: 0.364 ng
RT: 17.106 min Scan# 1743
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD


Tgt Ion:178 Resp: 14207
Ion Ratio Lower Upper
178 100
176 19.5 15.4 23.2
179 15.4 12.2 18.4



#26
Anthracene
Concen: 0.359 ng
RT: 17.193 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Tgt Ion:178 Resp: 12435
Ion Ratio Lower Upper
178 100
176 18.5 14.9 22.3
179 15.3 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.332 ng
RT: 19.094 min Scan# 1982
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD


Tgt Ion:212 Resp: 9709
Ion Ratio Lower Upper
212 100
106 23.6 17.8 26.8
104 13.8 10.2 15.4



#28
Fluoranthene
Concen: 0.329 ng
RT: 19.122 min Scan# 1988
Delta R.T. -0.005 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Tgt Ion:202 Resp: 13425
Ion Ratio Lower Upper
202 100
101 16.0 12.8 19.2
203 17.1 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.241 min Scan# 2421

Delta R.T. -0.009 min

Lab File: BN034711.D

Acq: 25 Oct 2024 03:02

Instrument :

BNA_N

ClientSampleId :

PB164282BSD





Tgt Ion:240 Resp: 7027

Ion Ratio Lower Upper

240 100

120 30.8 14.2 21.2#

236 31.0 23.1 34.7



#30

Pyrene

Concen: 0.377 ng

RT: 19.484 min Scan# 2066

Delta R.T. 0.000 min

Lab File: BN034711.D

Acq: 25 Oct 2024 03:02




Tgt Ion:202 Resp: 13610


Ion Ratio Lower Upper

202 100

200 21.0 16.6 25.0

203 17.5 14.2 21.4









#31
 Terphenyl-d14
 Concen: 0.371 ng
 RT: 19.703 min Scan# 2113
 Delta R.T. 0.000 min
 Lab File: BN034711.D
 Acq: 25 Oct 2024 03:02




Instrument :
 BNA_N
 ClientSampleId :
 PB164282BSD


Tgt Ion:244	Resp:	5323
Ion Ratio	Lower	Upper
244	100	
212	11.8	8.6 13.0
122	28.3	22.3 33.5



#32
 Benzo(a)anthracene
 Concen: 0.369 ng
 RT: 21.232 min Scan# 2419
 Delta R.T. 0.000 min
 Lab File: BN034711.D
 Acq: 25 Oct 2024 03:02

Tgt Ion:228	Resp:	9893
Ion Ratio	Lower	Upper
228	100	
226	27.4	22.0 33.0
229	20.1	15.6 23.4









#33
Chrysene
Concen: 0.396 ng
RT: 21.277 min Scan# 2425
Delta R.T. -0.009 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD


Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.8	24.1	36.1
229	19.5	15.8	23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.349 ng
RT: 21.187 min Scan# 2414
Delta R.T. -0.009 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Tgt Ion	Ratio	Lower	Upper
149	100		
167	22.6	19.7	29.5
279	1.6	1.8	2.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.458 min Scan# 2910

Delta R.T. -0.009 min

Lab File: BN034711.D


Acq: 25 Oct 2024 03:02

Instrument :

BNA_N

ClientSampleId :

PB164282BSD






Tgt Ion:264 Resp: 5873

Ion Ratio Lower Upper

264 100

260 28.4 21.9 32.9

265 103.3 60.6 91.0#



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.429 ng

RT: 25.683 min Scan# 3671

Delta R.T. -0.012 min

Lab File: BN034711.D

Acq: 25 Oct 2024 03:02




Tgt Ion:276 Resp: 9709


Ion Ratio Lower Upper

276 100

138 47.6 35.1 52.7

277 24.0 19.7 29.5









#37
Benzo(b)fluoranthene
Concen: 0.411 ng
RT: 22.797 min Scan# 2684
Delta R.T. -0.006 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD


Tgt Ion:252 Resp: 9630
Ion Ratio Lower Upper
252 100
253 26.4 20.0 30.0
125 31.1 20.9 31.3



#38
Benzo(k)fluoranthene
Concen: 0.416 ng
RT: 22.797 min Scan# 2684
Delta R.T. -0.053 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02

Tgt Ion:252 Resp: 9630
Ion Ratio Lower Upper
252 100
253 26.4 20.6 31.0
125 31.1 21.6 32.4









#39
Benzo(a)pyrene
Concen: 0.432 ng
RT: 23.367 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02




Instrument :
BNA_N
ClientSampleId :
PB164282BSD


Tgt Ion:252 Resp: 8133
Ion Ratio Lower Upper
252 100
253 28.6 22.2 33.4
125 39.1 26.0 39.0#



#40
Dibenzo(a,h)anthracene
Concen: 0.417 ng
RT: 25.703 min Scan# 3678
Delta R.T. -0.012 min
Lab File: BN034711.D
Acq: 25 Oct 2024 03:02


Tgt Ion:278 Resp: 7436
Ion Ratio Lower Upper
278 100
139 35.2 25.4 38.2
279 29.0 22.4 33.6





#41
 Benzo(g,h,i)perylene
 Concen: 0.404 ng
 RT: 26.364 min Scan# 3904
 Delta R.T. -0.009 min
 Lab File: BN034711.D
 Acq: 25 Oct 2024 03:02

Instrument :
 BNA_N
 ClientSampleId :
 PB164282BSD



Tgt Ion:	276	Resp:	7954
Ion	Ratio	Lower	Upper
276	100		
277	26.7	21.3	31.9
138	43.3	31.9	47.9

