

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG No.: P4440

Instrument ID: BNA_N Calibration Date/Time: 10/25/2024 07:51

Lab File ID: BN034719.D Init. Calib. Date(s): 10/24/2024 10/24/2024

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 09:11 12:48

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.550	0.486		-11.6	50.0
Fluoranthene-d10	0.901	0.942		4.6	50.0
2-Fluorophenol	1.189	1.294		8.8	50.0
Phenol-d6	1.568	1.736		10.7	50.0
Nitrobenzene-d5	0.336	0.300		-10.7	50.0
2-Fluorobiphenyl	1.590	1.367		-14.0	50.0
2,4,6-Tribromophenol	0.124	0.113		-8.9	50.0
Terphenyl-d14	0.817	0.696		-14.8	50.0
1,4-Dioxane	0.532	0.475		-10.7	50.0

All other compounds must meet a minimum RRF of 0.010.