

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG No.: P4440

Instrument ID: MSVOA_N Calibration Date(s): 09/30/2024 09/30/2024

Heated Purge: (Y/N) N Calibration Time(s): 12:25 14:48

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:	RRF100 = VN084213.D		RRF050 = VN084214.D		RRF020 = VN084215.D		RRF010 = VN084216.D		RRF005 = VN084217.D		RRF001 = VN084218.D	
COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	RRF	% RSD				
Chloromethane	0.619	0.671	0.677	0.648	0.747	0.690	0.675	6.4				
Vinyl Chloride	0.611	0.667	0.665	0.641	0.724	0.617	0.654	6.4				
Bromomethane	0.368	0.432	0.432	0.424	0.501		0.431	10.9				
Chloroethane	0.375	0.419	0.430	0.433	0.522	0.632	0.468	19.9				
Trichlorofluoromethane	0.953	1.060	1.047	0.959	1.113	0.978	1.019	6.4				
1,1,2-Trichlorotrifluoroethane	0.542	0.590	0.603	0.532	0.633	0.602	0.584	6.7				
1,1-Dichloroethene	0.538	0.587	0.596	0.533	0.631	0.493	0.563	8.9				
Acetone	0.299	0.342	0.337	0.298	0.336	0.299	0.318	6.8				
Carbon Disulfide	1.588	1.723	1.746	1.650	1.908	2.080	1.782	10.2				
Methyl tert-butyl Ether	1.825	2.033	2.035	1.802	2.049	1.656	1.900	8.6				
Methylene Chloride	0.594	0.655	0.661	0.618	0.669	0.686	0.647	5.3				
trans-1,2-Dichloroethene	0.555	0.622	0.619	0.570	0.627	0.546	0.590	6.2				
1,1-Dichloroethane	1.075	1.193	1.163	1.084	1.226	1.046	1.131	6.4				
2-Butanone	0.395	0.452	0.465	0.419	0.467	0.404	0.434	7.3				
Carbon Tetrachloride	0.515	0.549	0.553	0.508	0.545	0.482	0.525	5.4				
cis-1,2-Dichloroethene	0.670	0.741	0.741	0.655	0.767	0.703	0.713	6.2				
Chloroform	1.083	1.204	1.205	1.117	1.259	1.181	1.175	5.5				
1,1,1-Trichloroethane	0.997	1.102	1.089	1.018	1.154	0.972	1.055	6.7				
Methylcyclohexane	0.567	0.583	0.577	0.507	0.534	0.428	0.533	11				
Benzene	1.434	1.553	1.559	1.410	1.574	1.421	1.492	5.2				
1,2-Dichloroethane	0.480	0.528	0.524	0.498	0.544	0.460	0.506	6.3				
Trichloroethene	0.334	0.362	0.361	0.329	0.379	0.325	0.348	6.3				
1,2-Dichloropropane	0.346	0.375	0.380	0.339	0.388	0.289	0.353	10.4				
Bromodichloromethane	0.521	0.557	0.556	0.497	0.570	0.475	0.529	7.2				
4-Methyl-2-Pentanone	0.449	0.508	0.510	0.468	0.484	0.387	0.468	9.9				
Toluene	0.904	0.970	0.963	0.861	0.920	0.840	0.910	5.8				
t-1,3-Dichloropropene	0.562	0.590	0.573	0.517	0.564	0.430	0.539	10.9				
cis-1,3-Dichloropropene	0.592	0.638	0.608	0.569	0.606	0.469	0.580	10.2				
1,1,2-Trichloroethane	0.317	0.349	0.347	0.316	0.340	0.288	0.326	7.3				
2-Hexanone	0.341	0.387	0.387	0.340	0.357	0.285	0.349	10.8				

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG No.: P4440
Instrument ID: MSVOA_N Calibration Date(s): 09/30/2024 09/30/2024
Heated Purge: (Y/N) N Calibration Time(s): 12:25 14:48
GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF100 = VN084213.D		RRF050 = VN084214.D		RRF020 = VN084215.D			
		RRF010 = VN084216.D		RRF005 = VN084217.D		RRF001 = VN084218.D			
COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	RRF	% RSD	
Dibromochloromethane	0.384	0.418	0.409	0.374	0.396	0.330	0.385	8.1	
Tetrachloroethene	0.323	0.359	0.351	0.338	0.373	0.346	0.348	5	
Chlorobenzene	1.068	1.170	1.143	1.069	1.173	1.028	1.109	5.5	
Ethyl Benzene	1.988	2.121	2.028	1.840	2.030	1.756	1.961	6.9	
m/p-Xylenes	0.741	0.806	0.779	0.695	0.729	0.600	0.725	10	
o-Xylene	0.714	0.774	0.738	0.666	0.734	0.491	0.686	14.9	
Styrene	1.234	1.312	1.238	1.112	1.159	0.918	1.162	11.9	
Bromoform	0.282	0.315	0.303	0.260	0.288	0.239	0.281	10	
Isopropylbenzene	3.737	4.132	4.055	3.677	3.864	3.428	3.815	6.8	
1,1,2,2-Tetrachloroethane	1.001	1.183	1.187	1.127	1.291	1.095	1.147	8.5	
1,3-Dichlorobenzene	1.624	1.787	1.780	1.646	1.843	1.679	1.727	5.1	
1,4-Dichlorobenzene	1.628	1.784	1.734	1.638	1.889	1.789	1.744	5.7	
1,2-Dichlorobenzene	1.555	1.710	1.740	1.613	1.769	1.770	1.693	5.3	
1,2-Dichloroethane-d4	0.673	0.737	0.764	0.712	0.821		0.741	7.5	
Dibromofluoromethane	0.308	0.324	0.341	0.316	0.359		0.330	6.1	
Toluene-d8	1.189	1.243	1.242	1.148	1.242		1.213	3.5	
4-Bromofluorobenzene	0.447	0.452	0.449	0.406	0.456		0.442	4.6	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.