

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG No.: P4440

Instrument ID: MSVOA\_N Calibration Date/Time: 10/21/2024 20:04

Lab File ID: VN084448.D Init. Calib. Date(s): 09/30/2024 09/30/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:25 14:48

GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND                       | RRF   | RRF050 | MIN<br>RRF | %D     | MAX%D |
|--------------------------------|-------|--------|------------|--------|-------|
| Chloromethane                  | 0.675 | 0.575  | 0.1        | -14.81 | 50    |
| Vinyl Chloride                 | 0.654 | 0.580  |            | -11.31 | 50    |
| Bromomethane                   | 0.431 | 0.354  |            | -17.86 | 50    |
| Chloroethane                   | 0.468 | 0.379  |            | -19.02 | 50    |
| Trichlorofluoromethane         | 1.019 | 0.994  |            | -2.45  | 50    |
| 1,1,2-Trichlorotrifluoroethane | 0.584 | 0.582  |            | -0.34  | 50    |
| 1,1-Dichloroethene             | 0.563 | 0.556  |            | -1.24  | 50    |
| Acetone                        | 0.318 | 0.255  |            | -19.81 | 50    |
| Carbon Disulfide               | 1.782 | 1.354  |            | -24.02 | 50    |
| Methyl tert-butyl Ether        | 1.900 | 1.972  |            | 3.79   | 50    |
| Methylene Chloride             | 0.647 | 0.658  |            | 1.7    | 50    |
| trans-1,2-Dichloroethene       | 0.590 | 0.582  |            | -1.36  | 50    |
| 1,1-Dichloroethane             | 1.131 | 1.157  | 0.1        | 2.3    | 50    |
| 2-Butanone                     | 0.434 | 0.406  |            | -6.45  | 50    |
| Carbon Tetrachloride           | 0.525 | 0.534  |            | 1.71   | 50    |
| cis-1,2-Dichloroethene         | 0.713 | 0.722  |            | 1.26   | 50    |
| Chloroform                     | 1.175 | 1.226  |            | 4.34   | 50    |
| 1,1,1-Trichloroethane          | 1.055 | 1.082  |            | 2.56   | 50    |
| Methylcyclohexane              | 0.533 | 0.493  |            | -7.51  | 50    |
| Benzene                        | 1.492 | 1.520  |            | 1.88   | 50    |
| 1,2-Dichloroethane             | 0.506 | 0.507  |            | 0.2    | 50    |
| Trichloroethene                | 0.348 | 0.353  |            | 1.44   | 50    |
| 1,2-Dichloropropane            | 0.353 | 0.379  |            | 7.36   | 50    |
| Bromodichloromethane           | 0.529 | 0.566  |            | 6.99   | 50    |
| 4-Methyl-2-Pentanone           | 0.468 | 0.503  |            | 7.48   | 50    |
| Toluene                        | 0.910 | 0.953  |            | 4.72   | 50    |
| t-1,3-Dichloropropene          | 0.539 | 0.551  |            | 2.23   | 50    |
| cis-1,3-Dichloropropene        | 0.580 | 0.604  |            | 4.14   | 50    |
| 1,1,2-Trichloroethane          | 0.326 | 0.361  |            | 10.74  | 50    |
| 2-Hexanone                     | 0.349 | 0.364  |            | 4.3    | 50    |
| Dibromochloromethane           | 0.385 | 0.434  |            | 12.73  | 50    |
| Tetrachloroethene              | 0.348 | 0.341  |            | -2.01  | 50    |
| Chlorobenzene                  | 1.109 | 1.111  | 0.3        | 0.18   | 50    |
| Ethyl Benzene                  | 1.961 | 2.021  |            | 3.06   | 50    |
| m/p-Xylenes                    | 0.725 | 0.758  |            | 4.55   | 50    |
| o-Xylene                       | 0.686 | 0.734  |            | 7      | 50    |
| Styrene                        | 1.162 | 1.285  |            | 10.59  | 50    |
| Bromoform                      | 0.281 | 0.306  | 0.1        | 8.9    | 50    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG No.: P4440

Instrument ID: MSVOA\_N Calibration Date/Time: 10/21/2024 20:04

Lab File ID: VN084448.D Init. Calib. Date(s): 09/30/2024 09/30/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:25 14:48

GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND                  | RRF   | RRF050 | MIN<br>RRF | %D    | MAX%D |
|---------------------------|-------|--------|------------|-------|-------|
| Isopropylbenzene          | 3.815 | 3.702  |            | -2.96 | 50    |
| 1,1,2,2-Tetrachloroethane | 1.147 | 1.151  | 0.3        | 0.35  | 50    |
| 1,3-Dichlorobenzene       | 1.727 | 1.669  |            | -3.36 | 50    |
| 1,4-Dichlorobenzene       | 1.744 | 1.676  |            | -3.9  | 50    |
| 1,2-Dichlorobenzene       | 1.693 | 1.649  |            | -2.6  | 50    |
| 1,2-Dichloroethane-d4     | 0.741 | 0.710  |            | -4.18 | 50    |
| Dibromofluoromethane      | 0.330 | 0.343  |            | 3.94  | 50    |
| Toluene-d8                | 1.213 | 1.259  |            | 3.79  | 50    |
| 4-Bromofluorobenzene      | 0.442 | 0.461  |            | 4.3   | 50    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.