



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Cover Page

Order ID : P4440

Project ID : CTO WE13

Client : Tetra Tech NUS, Inc.

Lab Sample Number

P4440-01
P4440-02

Client Sample Number

BPOW6-11-HYD-20241016
BPOW6-11-TB-20241016

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 10/23/2024

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: CTO WE13

Project Manager : Ernie Wu

Chemtech Project # P4440

Test Name: VOCMS Group1

A. Number of Samples and Date of Receipt:

2 Water samples were received on 10/18/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1 and VOCMS Group1. This data package contains results for VOCMS Group1.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_N were done using GC column RXI-624SIL MS 30m 0.25mm 1.4 um. Cat#13868. The analysis of VOCMS Group1 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration File ID VN084428.D met the requirements except for 1,1,2-Trichloroethane, 2-Hexanone, 4-Methyl-2-Pentanone are failing high but no positive hit in associate sample therefore no corrective action taken while Dibromochloromethane failing marginally high and Carbon Disulfide is failing marginally low therefore no corrective action taken.

The Tuning criteria met requirements.

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)."



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092
NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: P4440

MATRIX: Water

METHOD: 8260D

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements. The Initial Calibration met the requirements . The Continuous Calibration File ID VN084428.D met the requirements except for 1,1,2-Trichloroethane,2-Hexanone,4-Methyl-2-Pentanone are failing high but no positive hit in associate sample therefore no corrective action taken while Dibromochloromethane failing marginally high and Carbon Disulfide is failing marginally low therefore no corrective action taken.		✓	
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.			✓
9.	Internal Standard Area/Retention Time Shift Meet Criteria Comments:			✓
10.	Analysis Holding Time Met If not met, list number of days exceeded for each sample:			✓

ADDITIONAL COMMENTS:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)."

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: P4440

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 10/23/2024

LAB CHRONICLE

OrderID:	P4440	OrderDate:	10/18/2024 10:22:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	CTO WE13
Contact:	Ernie Wu	Location:	K51,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P4440-01	BPOW6-11-HYD-2024 1016	Water	VOCMS Group1	8260-Low	10/16/24		10/21/24	10/18/24
P4440-02	BPOW6-11-TB-20241 016	Water	VOCMS Group1	8260-Low	10/16/24		10/21/24	10/18/24

Hit Summary Sheet
SW-846

SDG No.: P4440

Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:	BPOW6-11-HYD-20241016								
P4440-01	BPOW6-11-HYD-2	Water	Chloroform	1.20		0.26	0.50	1.00	ug/L
P4440-01	BPOW6-11-HYD-2	Water	Bromodichloromethane	2.80		0.24	0.50	1.00	ug/L
P4440-01	BPOW6-11-HYD-2	Water	Dibromochloromethane	4.50		0.18	0.50	1.00	ug/L
P4440-01	BPOW6-11-HYD-2	Water	Bromoform	2.70		0.21	0.50	1.00	ug/L
Total Voc :				11.2					
Total Concentration:				11.2					



QC SUMMARY

Surrogate Summary

SDG No.: P4440

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4440-01	BPOW6-11-HYD-20241016	1,2-Dichloroethane-d4	50	51.3	103	81	118
		Dibromofluoromethane	50	51.5	103	80	119
		Toluene-d8	50	49.7	99	89	112
		4-Bromofluorobenzene	50	46.9	94	85	114
P4440-02	BPOW6-11-TB-20241016	1,2-Dichloroethane-d4	50	47.5	95	81	118
		Dibromofluoromethane	50	50.9	102	80	119
		Toluene-d8	50	49.1	98	89	112
		4-Bromofluorobenzene	50	45.6	91	85	114
VN1021WBL01	VN1021WBL01	1,2-Dichloroethane-d4	50	48.6	97	81	118
		Dibromofluoromethane	50	51.1	102	80	119
		Toluene-d8	50	48.8	98	89	112
		4-Bromofluorobenzene	50	44.7	89	85	114
VN1021WBS01	VN1021WBS01	1,2-Dichloroethane-d4	50	50.5	101	81	118
		Dibromofluoromethane	50	54.3	109	80	119
		Toluene-d8	50	54.0	108	89	112
		4-Bromofluorobenzene	50	52.5	105	85	114
VN1021WBSD0	VN1021WBSD01	1,2-Dichloroethane-d4	50	51.1	102	81	118
		Dibromofluoromethane	50	52.8	106	80	119
		Toluene-d8	50	51.8	104	89	112
		4-Bromofluorobenzene	50	52.5	105	85	114

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4440

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VN084433.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1021WBS01	Chloromethane	20	16.1	ug/L	81			50	139	
	Vinyl chloride	20	16.7	ug/L	84			58	137	
	Bromomethane	20	16.5	ug/L	83			53	141	
	Chloroethane	20	15.0	ug/L	75			60	138	
	Trichlorofluoromethane	20	18.1	ug/L	91			65	141	
	1,1,2-Trichlorotrifluoroethane	20	18.1	ug/L	91			70	136	
	1,1-Dichloroethene	20	17.6	ug/L	88			71	131	
	Acetone	100	87.5	ug/L	88			39	160	
	Carbon disulfide	20	14.2	ug/L	71			64	133	
	Methyl tert-butyl Ether	20	18.1	ug/L	91			71	124	
	Methylene Chloride	20	18.6	ug/L	93			74	124	
	trans-1,2-Dichloroethene	20	17.4	ug/L	87			75	124	
	1,1-Dichloroethane	20	19.2	ug/L	96			77	125	
	2-Butanone	100	88.7	ug/L	89			56	143	
	Carbon Tetrachloride	20	18.9	ug/L	95			72	136	
	cis-1,2-Dichloroethene	20	18.2	ug/L	91			78	123	
	Chloroform	20	19.0	ug/L	95			79	124	
	1,1,1-Trichloroethane	20	18.7	ug/L	94			74	131	
	Methylcyclohexane	20	16.2	ug/L	81			72	132	
	Benzene	20	19.0	ug/L	95			79	120	
	1,2-Dichloroethane	20	18.9	ug/L	95			73	128	
	Trichloroethene	20	19.2	ug/L	96			79	123	
	1,2-Dichloropropane	20	19.6	ug/L	98			78	122	
	Bromodichloromethane	20	19.5	ug/L	98			79	125	
	4-Methyl-2-Pentanone	100	96.8	ug/L	97			67	130	
	Toluene	20	19.6	ug/L	98			80	121	
	t-1,3-Dichloropropene	20	18.5	ug/L	93			73	127	
	cis-1,3-Dichloropropene	20	18.9	ug/L	95			75	124	
	1,1,2-Trichloroethane	20	20.4	ug/L	102			80	119	
	2-Hexanone	100	96.1	ug/L	96			57	139	
	Dibromochloromethane	20	20.1	ug/L	101			74	126	
	Tetrachloroethene	20	18.9	ug/L	95			74	129	
	Chlorobenzene	20	19.2	ug/L	96			82	118	
	Ethyl Benzene	20	18.5	ug/L	93			79	121	
	m/p-Xylenes	40	38.5	ug/L	96			80	121	
	o-Xylene	20	20.0	ug/L	100			78	122	
	Styrene	20	19.5	ug/L	98			78	123	
	Bromoform	20	19.5	ug/L	98			66	130	
	Isopropylbenzene	20	18.2	ug/L	91			72	131	
	1,1,2,2-Tetrachloroethane	20	19.4	ug/L	97			71	121	
	1,3-Dichlorobenzene	20	18.6	ug/L	93			80	119	
	1,4-Dichlorobenzene	20	18.8	ug/L	94			79	118	
	1,2-Dichlorobenzene	20	18.3	ug/L	92			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: **P4440**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **SW8260-Low**

Datafile : VN084434.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1021WBSD01	Chloromethane	20	16.3	ug/L	81	0		50	139	20
	Vinyl chloride	20	17.1	ug/L	86	2		58	137	20
	Bromomethane	20	17.0	ug/L	85	2		53	141	20
	Chloroethane	20	16.0	ug/L	80	6		60	138	20
	Trichlorofluoromethane	20	19.0	ug/L	95	4		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	19.3	ug/L	97	6		70	136	20
	1,1-Dichloroethene	20	17.7	ug/L	89	1		71	131	20
	Acetone	100	91.6	ug/L	92	4		39	160	20
	Carbon disulfide	20	14.5	ug/L	73	3		64	133	20
	Methyl tert-butyl Ether	20	19.8	ug/L	99	8		71	124	20
	Methylene Chloride	20	20.2	ug/L	101	8		74	124	20
	trans-1,2-Dichloroethene	20	18.4	ug/L	92	6		75	124	20
	1,1-Dichloroethane	20	19.6	ug/L	98	2		77	125	20
	2-Butanone	100	96.8	ug/L	97	9		56	143	20
	Carbon Tetrachloride	20	19.4	ug/L	97	2		72	136	20
	cis-1,2-Dichloroethene	20	19.2	ug/L	96	5		78	123	20
	Chloroform	20	20.1	ug/L	101	6		79	124	20
	1,1,1-Trichloroethane	20	19.8	ug/L	99	5		74	131	20
	Methylcyclohexane	20	17.0	ug/L	85	5		72	132	20
	Benzene	20	19.6	ug/L	98	3		79	120	20
	1,2-Dichloroethane	20	20.6	ug/L	103	8		73	128	20
	Trichloroethene	20	19.5	ug/L	98	2		79	123	20
	1,2-Dichloropropane	20	20.9	ug/L	104	6		78	122	20
	Bromodichloromethane	20	20.9	ug/L	104	6		79	125	20
	4-Methyl-2-Pentanone	100	100	ug/L	100	3		67	130	20
	Toluene	20	20.3	ug/L	102	4		80	121	20
	t-1,3-Dichloropropene	20	19.7	ug/L	99	6		73	127	20
	cis-1,3-Dichloropropene	20	20.0	ug/L	100	5		75	124	20
	1,1,2-Trichloroethane	20	22.5	ug/L	113	10		80	119	20
	2-Hexanone	100	100	ug/L	100	4		57	139	20
	Dibromochloromethane	20	21.8	ug/L	109	8		74	126	20
	Tetrachloroethene	20	19.3	ug/L	97	2		74	129	20
	Chlorobenzene	20	19.5	ug/L	98	2		82	118	20
	Ethyl Benzene	20	18.4	ug/L	92	1		79	121	20
	m/p-Xylenes	40	38.8	ug/L	97	1		80	121	20
	o-Xylene	20	18.8	ug/L	94	6		78	122	20
	Styrene	20	19.8	ug/L	99	1		78	123	20
	Bromoform	20	20.7	ug/L	104	6		66	130	20
	Isopropylbenzene	20	18.4	ug/L	92	1		72	131	20
	1,1,2,2-Tetrachloroethane	20	20.6	ug/L	103	6		71	121	20
	1,3-Dichlorobenzene	20	19.3	ug/L	97	4		80	119	20
	1,4-Dichlorobenzene	20	19.4	ug/L	97	3		79	118	20
	1,2-Dichlorobenzene	20	18.8	ug/L	94	2		80	119	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VN1021WBL01

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: P4440

SAS No.: P4440 SDG NO.: P4440

Lab File ID: VN084430.D

Lab Sample ID: VN1021WBL01

Date Analyzed: 10/21/2024

Time Analyzed: 12:36

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VN1021WBS01	VN1021WBS01	VN084433.D	10/21/2024
VN1021WBSD01	VN1021WBSD01	VN084434.D	10/21/2024
BPOW6-11-TB-20241016	P4440-02	VN084437.D	10/21/2024
BPOW6-11-HYD-20241016	P4440-01	VN084443.D	10/21/2024

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG NO.: P4440
Lab File ID: VN084211.D BFB Injection Date: 09/30/2024
Instrument ID: MSVOA_N BFB Injection Time: 09:24
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.9
75	30.0 - 60.0% of mass 95	53.9
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.6 (0.8) 1
174	50.0 - 100.0% of mass 95	71
175	5.0 - 9.0% of mass 174	5.5 (7.8) 1
176	95.0 - 101.0% of mass 174	69.7 (98.2) 1
177	5.0 - 9.0% of mass 176	4.9 (7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC100	VSTDICC100	VN084213.D	09/30/2024	12:25
VSTDICCC050	VSTDICCC050	VN084214.D	09/30/2024	12:49
VSTDICC020	VSTDICC020	VN084215.D	09/30/2024	13:13
VSTDICC010	VSTDICC010	VN084216.D	09/30/2024	13:37
VSTDICC005	VSTDICC005	VN084217.D	09/30/2024	14:00
VSTDICC001	VSTDICC001	VN084218.D	09/30/2024	14:48



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG NO.: P4440
Lab File ID: VN084427.D BFB Injection Date: 10/21/2024
Instrument ID: MSVOA_N BFB Injection Time: 11:14
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	17
75	30.0 - 60.0% of mass 95	51.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.6
173	Less than 2.0% of mass 174	0.7 (0.9) 1
174	50.0 - 100.0% of mass 95	72.7
175	5.0 - 9.0% of mass 174	5.1 (7) 1
176	95.0 - 101.0% of mass 174	70 (96.3) 1
177	5.0 - 9.0% of mass 176	4.6 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN084428.D	10/21/2024	11:38
VN1021WBL01	VN1021WBL01	VN084430.D	10/21/2024	12:36
VN1021WBS01	VN1021WBS01	VN084433.D	10/21/2024	13:48
VN1021WBSD01	VN1021WBSD01	VN084434.D	10/21/2024	14:22
BPOW6-11-TB-20241016	P4440-02	VN084437.D	10/21/2024	15:41
BPOW6-11-HYD-20241016	P4440-01	VN084443.D	10/21/2024	18:04
VSTDCCC050EC	VSTDCCC050	VN084448.D	10/21/2024	20:04

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG NO.: P4440
Lab File ID: VN084428.D Date Analyzed: 10/21/2024
Instrument ID: MSVOA_N Time Analyzed: 11:38
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	157355	8.22	262800	9.10	244635	11.87
UPPER LIMIT	314710	8.724	525600	9.6	489270	12.365
LOWER LIMIT	78677.5	7.724	131400	8.6	122318	11.365
EPA SAMPLE NO.						
BPOW6-11-HYD-20241016	160653	8.22	286612	9.10	255046	11.87
BPOW6-11-TB-20241016	177910	8.22	306096	9.10	264394	11.87
VN1021WBL01	167468	8.22	295479	9.10	254088	11.87
VN1021WBS01	189913	8.22	322919	9.10	284161	11.87
VN1021WBSD01	163086	8.22	278516	9.10	255670	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG NO.: P4440
 Lab File ID: VN084428.D Date Analyzed: 10/21/2024
 Instrument ID: MSVOA_N Time Analyzed: 11:38
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	127638	13.788				
UPPER LIMIT	255276	14.288				
LOWER LIMIT	63819	13.288				
EPA SAMPLE NO.						
BPOW6-11-HYD-20241016	101220	13.79				
BPOW6-11-TB-20241016	103232	13.79				
VN1021WBL01	99954	13.79				
VN1021WBS01	138325	13.79				
VN1021WBSD01	124118	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	10/16/24		
Project:	CTO WE13			Date Received:	10/18/24		
Client Sample ID:	BPOW6-11-HYD-20241016			SDG No.:	P4440		
Lab Sample ID:	P4440-01			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	RXI-624	ID :	0.25	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084443.D	1		10/21/24 18:04	VN102124

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	1.20		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	2.80		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	10/16/24		
Project:	CTO WE13			Date Received:	10/18/24		
Client Sample ID:	BPOW6-11-HYD-20241016			SDG No.:	P4440		
Lab Sample ID:	P4440-01			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	RXI-624	ID :	0.25	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084443.D	1		10/21/24 18:04	VN102124

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	4.50		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	2.70		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.3		81 - 118		103%	SPK: 50
1868-53-7	Dibromofluoromethane	51.5		80 - 119		103%	SPK: 50
2037-26-5	Toluene-d8	49.7		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.8		85 - 114		94%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	161000	8.224				
540-36-3	1,4-Difluorobenzene	287000	9.1				
3114-55-4	Chlorobenzene-d5	255000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	101000	13.788				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/16/24	
Project:	CTO WE13		Date Received:	10/18/24	
Client Sample ID:	BPOW6-11-HYD-20241016		SDG No.:	P4440	
Lab Sample ID:	P4440-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084443.D	1		10/21/24 18:04	VN102124

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084443.D
Acq On : 21 Oct 2024 18:04
Operator : JC\MD
Sample : P4440-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-HYD-20241016

Quant Time: Oct 22 01:39:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.224	168	160653	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	286612	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	255046	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	101220	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	122301	51.344	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	102.680%	
35) Dibromofluoromethane	8.165	113	97309	51.499	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	103.000%	
50) Toluene-d8	10.565	98	345786	49.741	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	99.480%	
62) 4-Bromofluorobenzene	12.847	95	118648	46.848	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	93.700%	
Target Compounds						
					Qvalue	
30) Chloroform	7.971	83	4593	1.217	ug/l	91
47) Bromodichloromethane	9.888	83	8525	2.811	ug/l #	92
60) Dibromochloromethane	11.359	129	9837	4.454	ug/l	96
71) Bromoform	12.576	173	3930	2.741	ug/l #	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084443.D
Acq On : 21 Oct 2024 18:04
Operator : JC\MD
Sample : P4440-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-HYD-20241016


Quant Time: Oct 22 01:39:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084443.D
Acq: 21 Oct 2024 18:04

Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-HYD-20241016







Tgt Ion:168 Resp: 160653
Ion Ratio Lower Upper
168 100
99 64.4 54.2 81.2



#30
Chloroform
Concen: 1.217 ug/l
RT: 7.971 min Scan# 1023
Delta R.T. 0.006 min
Lab File: VN084443.D
Acq: 21 Oct 2024 18:04

Tgt Ion: 83 Resp: 4593
Ion Ratio Lower Upper
83 100
85 56.6 51.0 76.4





#33

1,2-Dichloroethane-d4

Concen: 51.344 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VN084443.D

Acq: 21 Oct 2024 18:04

Instrument :

MSVOA_N

ClientSampleId :





BPOW6-11-HYD-20241016

Tgt Ion: 65 Resp: 122301

Ion Ratio Lower Upper

65 100

67 51.2 0.0 102.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.000 min

Lab File: VN084443.D

Acq: 21 Oct 2024 18:04




Tgt Ion: 114 Resp: 286612


Ion Ratio Lower Upper

114 100

63 22.0 0.0 43.8

88 16.5 0.0 31.6





#35

Dibromofluoromethane

Concen: 51.499 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN084443.D

Acq: 21 Oct 2024 18:04

Instrument :

MSVOA_N

ClientSampleId :

BPOW6-11-HYD-20241016





Tgt Ion:113 Resp: 97309

Ion Ratio Lower Upper

113 100

111 102.7 83.3 124.9

192 17.9 13.5 20.3



#47

Bromodichloromethane

Concen: 2.811 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. 0.000 min

Lab File: VN084443.D

Acq: 21 Oct 2024 18:04




Tgt Ion: 83 Resp: 8525


Ion Ratio Lower Upper

83 100

85 59.2 51.5 77.3

127 14.7 5.7 8.5#









#50
Toluene-d8
Concen: 49.741 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.000 min
Lab File: VN084443.D
Acq: 21 Oct 2024 18:04




Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-HYD-20241016


Tgt Ion: 98 Resp: 345786
Ion Ratio Lower Upper
98 100
100 64.3 52.7 79.1



#60
Dibromochloromethane
Concen: 4.454 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN084443.D
Acq: 21 Oct 2024 18:04

Tgt Ion: 129 Resp: 9837
Ion Ratio Lower Upper
129 100
127 72.8 38.3 114.9





#62

4-Bromofluorobenzene

Concen: 46.848 ug/l

RT: 12.847 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN084443.D

Acq: 21 Oct 2024 18:04

Instrument :

MSVOA_N

ClientSampleId :

BPOW6-11-HYD-20241016





Tgt Ion: 95 Resp: 118648

Ion Ratio Lower Upper

95 100

174 73.4 0.0 145.2

176 70.6 0.0 140.0



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VN084443.D

Acq: 21 Oct 2024 18:04




Tgt Ion: 117 Resp: 255046


Ion Ratio Lower Upper

117 100

82 57.0 47.2 70.8

119 31.2 25.4 38.0









#71
Bromoform
Concen: 2.741 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. 0.000 min
Lab File: VN084443.D
Acq: 21 Oct 2024 18:04




Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-HYD-20241016

Tgt Ion:173 Resp: 3930
Ion Ratio Lower Upper
173 100
175 50.9 23.5 70.6
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN084443.D
Acq: 21 Oct 2024 18:04

Tgt Ion:152 Resp: 101220
Ion Ratio Lower Upper
152 100
115 63.4 31.3 93.9
150 158.9 0.0 349.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084443.D
Acq On : 21 Oct 2024 18:04
Operator : JC\MD
Sample : P4440-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-HYD-20241016

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M

Title : SW846 8260

Signal : TIC: VN084443.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	8.165	1047	1056	1061	rBV	130085	324610	35.45%	6.749%
2	8.224	1061	1066	1084	rVB	236808	509203	55.62%	10.586%
3	8.577	1117	1126	1139	rBV	149857	331899	36.25%	6.900%
4	9.100	1207	1215	1228	rBV	335628	677977	74.05%	14.095%
5	9.888	1340	1349	1359	rVB2	11775	23107	2.52%	0.480%
6	10.565	1456	1464	1476	rBV	505946	915566	100.00%	19.034%
7	11.359	1593	1599	1606	rVB	17328	29150	3.18%	0.606%
8	11.865	1677	1685	1698	rBV	447460	787406	86.00%	16.370%
9	12.577	1796	1806	1812	rVB4	8216	13981	1.53%	0.291%
10	12.847	1844	1852	1862	rVB	341837	564003	61.60%	11.726%
11	13.788	2005	2012	2022	rBV	379909	633135	69.15%	13.163%




Sum of corrected areas: 4810037

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084443.D
Acq On : 21 Oct 2024 18:04
Operator : JC\MD
Sample : P4440-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-HYD-20241016

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084443.D
Acq On : 21 Oct 2024 18:04
Operator : JC\MD
Sample : P4440-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-HYD-20241016

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084443.D
Acq On : 21 Oct 2024 18:04
Operator : JC\MD
Sample : P4440-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-HYD-20241016

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	10/16/24	
Project:	CTO WE13			Date Received:	10/18/24	
Client Sample ID:	BPOW6-11-TB-20241016			SDG No.:	P4440	
Lab Sample ID:	P4440-02			Matrix:	Water	
Analytical Method:	SW8260			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID :	0.25	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084437.D	1		10/21/24 15:41	VN102124

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	10/16/24		
Project:	CTO WE13			Date Received:	10/18/24		
Client Sample ID:	BPOW6-11-TB-20241016			SDG No.:	P4440		
Lab Sample ID:	P4440-02			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	RXI-624	ID :	0.25	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084437.D	1		10/21/24 15:41	VN102124

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	47.5		81 - 118		95%	SPK: 50
1868-53-7	Dibromofluoromethane	50.9		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	49.1		89 - 112		98%	SPK: 50
460-00-4	4-Bromofluorobenzene	45.6		85 - 114		91%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	178000	8.224				
540-36-3	1,4-Difluorobenzene	306000	9.1				
3114-55-4	Chlorobenzene-d5	264000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	103000	13.794				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/16/24	
Project:	CTO WE13		Date Received:	10/18/24	
Client Sample ID:	BPOW6-11-TB-20241016		SDG No.:	P4440	
Lab Sample ID:	P4440-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084437.D	1		10/21/24 15:41	VN102124

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084437.D
 Acq On : 21 Oct 2024 15:41
 Operator : JC\MD
 Sample : P4440-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BPOW6-11-TB-20241016

Quant Time: Oct 22 01:35:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	177910	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	306096	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	264394	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	103232	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	125403	47.540	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.080%
35) Dibromofluoromethane	8.165	113	102688	50.887	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.780%
50) Toluene-d8	10.565	98	364855	49.143	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.280%
62) 4-Bromofluorobenzene	12.847	95	123415	45.629	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.260%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084437.D
Acq On : 21 Oct 2024 15:41
Operator : JC\MD
Sample : P4440-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-TB-20241016


Quant Time: Oct 22 01:35:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084437.D
Acq: 21 Oct 2024 15:41

Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-TB-20241016







Tgt Ion:168 Resp: 177910
Ion Ratio Lower Upper
168 100
99 61.1 54.2 81.2



#33
1,2-Dichloroethane-d4
Concen: 47.540 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. -0.000 min
Lab File: VN084437.D
Acq: 21 Oct 2024 15:41

Tgt Ion: 65 Resp: 125403
Ion Ratio Lower Upper
65 100
67 50.4 0.0 102.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN084437.D

Acq: 21 Oct 2024 15:41

Instrument :

MSVOA_N

ClientSampleId :

BPOW6-11-TB-20241016





Tgt Ion:114 Resp: 306096

Ion Ratio Lower Upper

114 100

63 20.8 0.0 43.8

88 16.2 0.0 31.6



#35

Dibromofluoromethane

Concen: 50.887 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN084437.D

Acq: 21 Oct 2024 15:41




Tgt Ion:113 Resp: 102688


Ion Ratio Lower Upper

113 100

111 101.7 83.3 124.9

192 17.6 13.5 20.3









#50
Toluene-d8
Concen: 49.143 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VN084437.D
Acq: 21 Oct 2024 15:41




Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-TB-20241016


Tgt Ion: 98 Resp: 364855
Ion Ratio Lower Upper
98 100
100 64.5 52.7 79.1



#62
4-Bromofluorobenzene
Concen: 45.629 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN084437.D
Acq: 21 Oct 2024 15:41

Tgt Ion: 95 Resp: 123415
Ion Ratio Lower Upper
95 100
174 74.5 0.0 145.2
176 70.5 0.0 140.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1
Delta R.T. -0.000 min
Lab File: VN084437.D
Acq: 21 Oct 2024 15:41




Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-TB-20241016

Tgt Ion	Ratio	Lower	Upper
117	100		
82	56.5	47.2	70.8
119	31.3	25.4	38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.006 min
Lab File: VN084437.D
Acq: 21 Oct 2024 15:41

Tgt Ion	Ratio	Lower	Upper
152	100		
115	63.2	31.3	93.9
150	159.8	0.0	349.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084437.D
Acq On : 21 Oct 2024 15:41
Operator : JC\MD
Sample : P4440-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-TB-20241016

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M

Title : SW846 8260

Signal : TIC: VN084437.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	8.165	1044	1056	1061	rBV2	138202	340777	35.37%	6.848%
2	8.224	1061	1066	1079	rVB	254783	552058	57.30%	11.094%
3	8.577	1117	1126	1138	rBV2	154140	342365	35.53%	6.880%
4	9.100	1205	1215	1227	rBV	360817	726308	75.38%	14.595%
5	10.565	1454	1464	1478	rBV	527899	963505	100.00%	19.362%
6	11.865	1677	1685	1694	rVB	469512	814585	84.54%	16.369%
7	12.847	1845	1852	1862	rVB	352487	586486	60.87%	11.786%
8	13.788	2005	2012	2024	rBV	376487	650211	67.48%	13.066%




Sum of corrected areas: 4976295

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084437.D
Acq On : 21 Oct 2024 15:41
Operator : JC\MD
Sample : P4440-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-TB-20241016

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084437.D
Acq On : 21 Oct 2024 15:41
Operator : JC\MD
Sample : P4440-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-TB-20241016

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084437.D
Acq On : 21 Oct 2024 15:41
Operator : JC\MD
Sample : P4440-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPOW6-11-TB-20241016

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc



CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG No.: P4440
 Instrument ID: MSVOA_N Calibration Date(s): 09/30/2024 09/30/2024
 Heated Purge: (Y/N) N Calibration Time(s): 12:25 14:48
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:	RRF100 = VN084213.D			RRF050 = VN084214.D		RRF020 = VN084215.D		
	RRF010 = VN084216.D			RRF005 = VN084217.D		RRF001 = VN084218.D		
COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	RRF	% RSD
Chloromethane	0.619	0.671	0.677	0.648	0.747	0.690	0.675	6.4
Vinyl Chloride	0.611	0.667	0.665	0.641	0.724	0.617	0.654	6.4
Bromomethane	0.368	0.432	0.432	0.424	0.501		0.431	10.9
Chloroethane	0.375	0.419	0.430	0.433	0.522	0.632	0.468	19.9
Trichlorofluoromethane	0.953	1.060	1.047	0.959	1.113	0.978	1.019	6.4
1,1,2-Trichlorotrifluoroethane	0.542	0.590	0.603	0.532	0.633	0.602	0.584	6.7
1,1-Dichloroethene	0.538	0.587	0.596	0.533	0.631	0.493	0.563	8.9
Acetone	0.299	0.342	0.337	0.298	0.336	0.299	0.318	6.8
Carbon Disulfide	1.588	1.723	1.746	1.650	1.908	2.080	1.782	10.2
Methyl tert-butyl Ether	1.825	2.033	2.035	1.802	2.049	1.656	1.900	8.6
Methylene Chloride	0.594	0.655	0.661	0.618	0.669	0.686	0.647	5.3
trans-1,2-Dichloroethene	0.555	0.622	0.619	0.570	0.627	0.546	0.590	6.2
1,1-Dichloroethane	1.075	1.193	1.163	1.084	1.226	1.046	1.131	6.4
2-Butanone	0.395	0.452	0.465	0.419	0.467	0.404	0.434	7.3
Carbon Tetrachloride	0.515	0.549	0.553	0.508	0.545	0.482	0.525	5.4
cis-1,2-Dichloroethene	0.670	0.741	0.741	0.655	0.767	0.703	0.713	6.2
Chloroform	1.083	1.204	1.205	1.117	1.259	1.181	1.175	5.5
1,1,1-Trichloroethane	0.997	1.102	1.089	1.018	1.154	0.972	1.055	6.7
Methylcyclohexane	0.567	0.583	0.577	0.507	0.534	0.428	0.533	11
Benzene	1.434	1.553	1.559	1.410	1.574	1.421	1.492	5.2
1,2-Dichloroethane	0.480	0.528	0.524	0.498	0.544	0.460	0.506	6.3
Trichloroethene	0.334	0.362	0.361	0.329	0.379	0.325	0.348	6.3
1,2-Dichloropropane	0.346	0.375	0.380	0.339	0.388	0.289	0.353	10.4
Bromodichloromethane	0.521	0.557	0.556	0.497	0.570	0.475	0.529	7.2
4-Methyl-2-Pentanone	0.449	0.508	0.510	0.468	0.484	0.387	0.468	9.9
Toluene	0.904	0.970	0.963	0.861	0.920	0.840	0.910	5.8
t-1,3-Dichloropropene	0.562	0.590	0.573	0.517	0.564	0.430	0.539	10.9
cis-1,3-Dichloropropene	0.592	0.638	0.608	0.569	0.606	0.469	0.580	10.2
1,1,2-Trichloroethane	0.317	0.349	0.347	0.316	0.340	0.288	0.326	7.3
2-Hexanone	0.341	0.387	0.387	0.340	0.357	0.285	0.349	10.8

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG No.: P4440
 Instrument ID: MSVOA_N Calibration Date(s): 09/30/2024 09/30/2024
 Heated Purge: (Y/N) N Calibration Time(s): 12:25 14:48
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF100 = VN084213.D		RRF050 = VN084214.D		RRF020 = VN084215.D			
		RRF010 = VN084216.D		RRF005 = VN084217.D		RRF001 = VN084218.D			
COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	RRF	% RSD	
Dibromochloromethane	0.384	0.418	0.409	0.374	0.396	0.330	0.385	8.1	
Tetrachloroethene	0.323	0.359	0.351	0.338	0.373	0.346	0.348	5	
Chlorobenzene	1.068	1.170	1.143	1.069	1.173	1.028	1.109	5.5	
Ethyl Benzene	1.988	2.121	2.028	1.840	2.030	1.756	1.961	6.9	
m/p-Xylenes	0.741	0.806	0.779	0.695	0.729	0.600	0.725	10	
o-Xylene	0.714	0.774	0.738	0.666	0.734	0.491	0.686	14.9	
Styrene	1.234	1.312	1.238	1.112	1.159	0.918	1.162	11.9	
Bromoform	0.282	0.315	0.303	0.260	0.288	0.239	0.281	10	
Isopropylbenzene	3.737	4.132	4.055	3.677	3.864	3.428	3.815	6.8	
1,1,2,2-Tetrachloroethane	1.001	1.183	1.187	1.127	1.291	1.095	1.147	8.5	
1,3-Dichlorobenzene	1.624	1.787	1.780	1.646	1.843	1.679	1.727	5.1	
1,4-Dichlorobenzene	1.628	1.784	1.734	1.638	1.889	1.789	1.744	5.7	
1,2-Dichlorobenzene	1.555	1.710	1.740	1.613	1.769	1.770	1.693	5.3	
1,2-Dichloroethane-d4	0.673	0.737	0.764	0.712	0.821		0.741	7.5	
Dibromofluoromethane	0.308	0.324	0.341	0.316	0.359		0.330	6.1	
Toluene-d8	1.189	1.243	1.242	1.148	1.242		1.213	3.5	
4-Bromofluorobenzene	0.447	0.452	0.449	0.406	0.456		0.442	4.6	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\
 Method File : 82N093024W.M
 Title : SW846 8260
 Last Update : Tue Oct 01 07:11:01 2024
 Response Via : Initial Calibration

Calibration Files

1 =VN084218.D 5 =VN084217.D 10 =VN084216.D 20 =VN084215.D 50 =VN084214.D 100 =VN084213.

D

	Compound	1	5	10	20	50	100	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.600	0.607	0.562	0.600	0.613	0.567	0.591	3.67
3) P	Chloromethane	0.690	0.747	0.648	0.677	0.671	0.619	0.675	6.39
4) C	Vinyl Chloride	0.617	0.724	0.641	0.665	0.667	0.611	0.654	6.37#
5) T	Bromomethane	0.501	0.424	0.432	0.432	0.368	0.431	0.431	10.93
6) T	Chloroethane	0.632	0.522	0.433	0.430	0.419	0.375	0.468	19.90
7) T	Trichlorofluor...	0.978	1.113	0.959	1.047	1.060	0.953	1.019	6.36
8) T	Diethyl Ether	0.363	0.398	0.352	0.408	0.390	0.355	0.378	6.36
9) T	1,1,2-Trichlor...	0.602	0.633	0.532	0.603	0.590	0.542	0.584	6.68
10) T	Methyl Iodide	0.764	0.688	0.769	0.806	0.731	0.752	0.752	5.87
11) T	Tert butyl alc...	0.143	0.116	0.128	0.123	0.101	0.122	0.122	12.49
12) CM	1,1-Dichloroet...	0.493	0.631	0.533	0.596	0.587	0.538	0.563	8.94#
13) T	Acrolein	0.162	0.130	0.135	0.142	0.131	0.140	0.140	9.47
14) T	Allyl chloride	0.888	1.097	0.896	1.086	0.995	0.905	0.978	9.85
15) T	Acrylonitrile	0.306	0.329	0.300	0.324	0.316	0.277	0.309	6.10
16) T	Acetone	0.299	0.336	0.298	0.337	0.342	0.299	0.318	6.80
17) T	Carbon Disulfide	2.080	1.908	1.650	1.746	1.723	1.588	1.782	10.18
18) T	Methyl Acetate	0.700	0.765	0.691	0.694	0.707	0.607	0.694	7.27
19) T	Methyl tert-bu...	1.656	2.049	1.802	2.035	2.033	1.825	1.900	8.57
20) T	Methylene Chlo...	0.686	0.669	0.618	0.661	0.655	0.594	0.647	5.29
21) T	trans-1,2-Dich...	0.546	0.627	0.570	0.619	0.622	0.555	0.590	6.21
22) T	Diisopropyl ether	1.756	2.207	1.900	2.170	2.164	1.938	2.022	9.10
23) T	Vinyl Acetate	1.196	1.585	1.411	1.655	1.648	1.505	1.500	11.67
24) P	1,1-Dichloroet...	1.046	1.226	1.084	1.163	1.193	1.075	1.131	6.43
25) T	2-Butanone	0.404	0.467	0.419	0.465	0.452	0.395	0.434	7.35
26) T	2,2-Dichloropr...	0.904	1.086	0.983	1.078	1.082	0.985	1.020	7.31
27) T	cis-1,2-Dichlo...	0.703	0.767	0.655	0.741	0.741	0.670	0.713	6.19
28) T	Bromochloromet...	0.535	0.619	0.423	0.486	0.494	0.472	0.505	13.20
29) T	Tetrahydrofuran	0.249	0.290	0.266	0.281	0.283	0.236	0.268	8.01
30) C	Chloroform	1.181	1.259	1.116	1.205	1.204	1.083	1.175	5.49#
31) T	Cyclohexane	1.321	1.047	1.084	1.068	0.970	1.098	1.098	12.01
32) T	1,1,1-Trichlor...	0.972	1.154	1.018	1.089	1.102	0.997	1.055	6.68
33) S	1,2-Dichloroet...	0.821	0.712	0.764	0.737	0.673	0.741	0.741	7.51
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...	0.359	0.316	0.341	0.324	0.308	0.330	0.330	6.14
36) T	1,1-Dichloropr...	0.417	0.504	0.468	0.501	0.509	0.474	0.479	7.25
37) T	Ethyl Acetate	0.447	0.560	0.488	0.496	0.503	0.452	0.491	8.34
38) T	Carbon Tetrach...	0.482	0.545	0.508	0.553	0.549	0.515	0.525	5.41
39) T	Methylcyclohexane	0.428	0.534	0.507	0.577	0.583	0.567	0.533	11.04
40) TM	Benzene	1.421	1.574	1.410	1.559	1.553	1.434	1.492	5.18
41) T	Methacrylonitrile	0.257	0.273	0.251	0.284	0.287	0.247	0.267	6.50
42) TM	1,2-Dichloroet...	0.460	0.544	0.498	0.524	0.528	0.480	0.506	6.32
43) T	Isopropyl Acetate	1.312	0.973	0.857	0.930	0.898	0.804	0.962	18.82
44) TM	Trichloroethene	0.325	0.379	0.329	0.361	0.362	0.334	0.348	6.32
45) C	1,2-Dichloropr...	0.289	0.388	0.339	0.380	0.375	0.346	0.353	10.44#
46) T	Dibromomethane	0.197	0.250	0.226	0.258	0.260	0.238	0.238	9.99
47) T	Bromodichlorom...	0.475	0.570	0.497	0.556	0.557	0.521	0.529	7.17
48) T	Methyl methacr...	0.314	0.408	0.371	0.411	0.418	0.376	0.383	10.14
49) T	1,4-Dioxane	0.007	0.007	0.007	0.007	0.007	0.006	0.007	5.48
50) S	Toluene-d8	1.242	1.148	1.242	1.243	1.189	1.213	1.213	3.54
51) T	4-Methyl-2-Pen...	0.387	0.484	0.468	0.510	0.508	0.449	0.468	9.88
52) CM	Toluene	0.840	0.920	0.861	0.963	0.970	0.904	0.910	5.79#
53) T	t-1,3-Dichloro...	0.430	0.564	0.517	0.573	0.590	0.562	0.539	10.92
54) T	cis-1,3-Dichlo...	0.469	0.606	0.569	0.608	0.638	0.592	0.580	10.18
55) T	1,1,2-Trichlor...	0.288	0.340	0.316	0.347	0.349	0.317	0.326	7.33

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\										
Method File : 82N093024W.M										
56)	T	Ethyl methacry...	0.414	0.548	0.513	0.587	0.602	0.556	0.537	12.63
57)	T	1,3-Dichloropr...	0.533	0.623	0.551	0.618	0.616	0.555	0.583	6.95
58)	T	2-Chloroethyl ...	0.213	0.267	0.233	0.232	0.280	0.268	0.249	10.63
59)	T	2-Hexanone	0.285	0.357	0.340	0.387	0.387	0.341	0.349	10.80
60)	T	Dibromochlorom...	0.330	0.396	0.374	0.409	0.418	0.384	0.385	8.12
61)	T	1,2-Dibromoethane	0.309	0.361	0.315	0.362	0.358	0.328	0.339	7.18
62)	S	4-Bromofluorob...	0.456	0.406	0.449	0.452	0.447	0.442		4.62
-----ISTD-----										
63)	I	Chlorobenzene-d5								
64)	T	Tetrachloroethene	0.346	0.373	0.338	0.351	0.359	0.323	0.348	4.98
65)	PM	Chlorobenzene	1.028	1.173	1.069	1.143	1.170	1.068	1.109	5.52
66)	T	1,1,1,2-Tetrac...	0.334	0.409	0.366	0.409	0.404	0.369	0.382	8.01
67)	C	Ethyl Benzene	1.756	2.030	1.840	2.028	2.121	1.988	1.961	6.94#
68)	T	m/p-Xylenes	0.600	0.729	0.695	0.779	0.806	0.741	0.725	10.00
69)	T	o-Xylene	0.491	0.734	0.666	0.738	0.774	0.714	0.686	14.87
70)	T	Styrene	0.918	1.159	1.112	1.238	1.312	1.234	1.162	11.87
71)	P	Bromoform	0.239	0.288	0.260	0.303	0.315	0.282	0.281	9.98
-----ISTD-----										
72)	I	1,4-Dichlorobenzen...								
73)	T	Isopropylbenzene	3.428	3.864	3.677	4.055	4.132	3.737	3.815	6.78
74)	T	N-amyl acetate	1.583	1.758	1.705	1.827	1.857	1.636	1.728	6.20
75)	P	1,1,2,2-Tetrac...	1.095	1.291	1.127	1.187	1.183	1.001	1.147	8.52
76)	T	1,2,3-Trichlor...	0.895	1.251	0.999	1.003	1.125	0.851	1.021	14.47
77)	T	Bromobenzene	0.891	0.982	0.896	0.955	0.947	0.851	0.920	5.33
78)	T	n-propylbenzene	3.455	4.544	4.401	4.719	4.906	4.498	4.421	11.44
79)	T	2-Chlorotoluene	2.640	2.948	2.737	2.891	3.013	2.696	2.821	5.34
80)	T	1,3,5-Trimethy...	2.369	3.209	3.077	3.358	3.491	3.170	3.112	12.62
81)	T	trans-1,4-Dich...	0.408	0.428	0.435	0.451	0.407	0.426		4.34
82)	T	4-Chlorotoluene	2.537	2.968	2.777	2.962	3.029	2.765	2.840	6.46
83)	T	tert-Butylbenzene	2.407	2.844	2.681	2.891	3.027	2.748	2.766	7.68
84)	T	1,2,4-Trimethy...	2.334	3.209	3.102	3.449	3.522	3.195	3.135	13.53
85)	T	sec-Butylbenzene	2.842	3.641	3.715	4.021	4.143	3.787	3.691	12.39
86)	T	p-Isopropyltol...	2.204	3.030	2.995	3.361	3.509	3.215	3.052	15.04
87)	T	1,3-Dichlorobe...	1.679	1.843	1.646	1.780	1.787	1.624	1.727	5.13
88)	T	1,4-Dichlorobe...	1.789	1.889	1.638	1.734	1.784	1.628	1.744	5.71
89)	T	n-Butylbenzene	2.386	2.689	2.567	2.961	3.092	2.912	2.768	9.63
90)	T	Hexachloroethane	0.564	0.678	0.608	0.637	0.639	0.590	0.620	6.51
91)	T	1,2-Dichlorobe...	1.770	1.769	1.613	1.740	1.710	1.555	1.693	5.28
92)	T	1,2-Dibromo-3-...	0.207	0.271	0.242	0.253	0.238	0.209	0.237	10.44
93)	T	1,2,4-Trichlor...	0.657	0.839	0.792	0.923	0.933	0.859	0.834	12.15
94)	T	Hexachlorobuta...	0.422	0.460	0.401	0.445	0.397	0.370	0.416	7.98
95)	T	Naphthalene	2.467	2.702	2.657	2.935	3.079	2.747	2.764	7.79
96)	T	1,2,3-Trichlor...	0.751	0.822	0.820	0.903	0.896	0.816	0.835	6.83

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084213.D
 Acq On : 30 Sep 2024 12:25
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Quant Time: Oct 01 06:48:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 10/01/2024
 Supervised By : Mahesh Dadoda 10/01/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	192402	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	324582	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	288467	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	142880	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.577	65	259018	90.797	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	181.600%#
35) Dibromofluoromethane	8.165	113	200166	93.542	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	187.080%#
50) Toluene-d8	10.565	98	772146	98.079	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	196.160%#
62) 4-Bromofluorobenzene	12.847	95	289867	101.065	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	202.140%#

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.124	85	218015	95.796	ug/l	94
3) Chloromethane	2.359	50	238104	91.628	ug/l	98
4) Vinyl Chloride	2.512	62	234942	93.338	ug/l	99
5) Bromomethane	2.936	94	141604	85.288	ug/l	94
6) Chloroethane	3.112	64	144394	80.099	ug/l	92
7) Trichlorofluoromethane	3.495	101	366907	93.608	ug/l	98
8) Diethyl Ether	3.953	74	136436	93.854	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.365	101	208443	92.811	ug/l	99
10) Methyl Iodide	4.589	142	281269	97.248	ug/l	97
11) Tert butyl alcohol	5.518	59	194750	414.018	ug/l	99
12) 1,1-Dichloroethene	4.336	96	206863	95.491	ug/l	95
13) Acrolein	4.177	56	251811	466.304	ug/l	99
14) Allyl chloride	5.018	41	348108	92.526	ug/l	99
15) Acrylonitrile	5.712	53	533270	448.931	ug/l	99
16) Acetone	4.424	43	576139	470.138	ug/l	99
17) Carbon Disulfide	4.712	76	611042	89.090	ug/l	100
18) Methyl Acetate	5.018	43	233708	87.499	ug/l	99
19) Methyl tert-butyl Ether	5.794	73	702366	96.061	ug/l	99
20) Methylene Chloride	5.277	84	228706	91.809	ug/l	91
21) trans-1,2-Dichloroethene	5.783	96	213704	94.160	ug/l	94
22) Diisopropyl ether	6.671	45	745614	95.811	ug/l	98
23) Vinyl Acetate	6.600	43	2895429	498.979	ug/l	99
24) 1,1-Dichloroethane	6.565	63	413523	95.021	ug/l	99
25) 2-Butanone	7.483	43	759354	454.946	ug/l	100
26) 2,2-Dichloropropane	7.488	77	378893	96.560	ug/l	99
27) cis-1,2-Dichloroethene	7.483	96	257696	93.962	ug/l	99
28) Bromochloromethane	7.812	49	181546	93.497	ug/l	100
29) Tetrahydrofuran	7.836	42	453342	440.124	ug/l	99
30) Chloroform	7.965	83	416595	92.156	ug/l	100
31) Cyclohexane	8.253	56	373445	88.380	ug/l	99
32) 1,1,1-Trichloroethane	8.165	97	383566	94.476	ug/l	98
36) 1,1-Dichloropropene	8.371	75	307616	98.969	ug/l	99
37) Ethyl Acetate	7.559	43	293597	92.073	ug/l	100
38) Carbon Tetrachloride	8.359	117	334398	98.077	ug/l	98
39) Methylcyclohexane	9.600	83	368299	106.466	ug/l	99
40) Benzene	8.606	78	931076	96.145	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084213.D
 Acq On : 30 Sep 2024 12:25
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/01/2024
 Supervised By : Mahesh Dadoda 10/01/2024

Quant Time: Oct 01 06:48:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	160270	92.581	ug/l	97
42) 1,2-Dichloroethane	8.671	62	311563	94.900	ug/l	100
43) Isopropyl Acetate	8.688	43	521814	83.581	ug/l	100
44) Trichloroethene	9.347	130	216678	95.798	ug/l	97
45) 1,2-Dichloropropane	9.618	63	224724	98.095	ug/l	97
46) Dibromomethane	9.706	93	154739	100.091	ug/l	99
47) Bromodichloromethane	9.888	83	338406	98.526	ug/l	99
48) Methyl methacrylate	9.677	41	244061	98.163	ug/l	99
49) 1,4-Dioxane	9.694	88	82854	1810.110	ug/l	99
51) 4-Methyl-2-Pentanone	10.441	43	1456154	479.699	ug/l	100
52) Toluene	10.629	92	586544	99.324	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	365099	104.280	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	384030	101.975	ug/l	98
55) 1,1,2-Trichloroethane	11.018	97	205711	97.170	ug/l	97
56) Ethyl methacrylate	10.871	69	360916	103.608	ug/l	99
57) 1,3-Dichloropropane	11.165	76	360601	95.326	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	870416	538.640	ug/l	99
59) 2-Hexanone	11.194	43	1105736	487.363	ug/l	99
60) Dibromochloromethane	11.359	129	249435	99.721	ug/l	98
61) 1,2-Dibromoethane	11.465	107	212705	96.694	ug/l	100
64) Tetrachloroethene	11.100	164	186069	92.608	ug/l	99
65) Chlorobenzene	11.888	112	616423	96.376	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.959	131	212695	96.546	ug/l	99
67) Ethyl Benzene	11.959	91	1146786	101.385	ug/l	98
68) m/p-Xylenes	12.071	106	855381	204.437	ug/l	100
69) o-Xylene	12.394	106	412098	104.101	ug/l	99
70) Styrene	12.412	104	711855	106.165	ug/l	99
71) Bromoform	12.576	173	162862	100.412	ug/l #	97
73) Isopropylbenzene	12.694	105	1067798	97.936	ug/l	100
74) N-amyl acetate	12.494	43	467641	94.716	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.935	83	286161	87.285	ug/l	98
76) 1,2,3-Trichloropropane	12.994	75	243293m	85.433	ug/l	
77) Bromobenzene	12.982	156	243075	92.425	ug/l	100
78) n-propylbenzene	13.035	91	1285399	101.754	ug/l	99
79) 2-Chlorotoluene	13.123	91	770341	95.567	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	905999	101.869	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	116289	95.579	ug/l	98
82) 4-Chlorotoluene	13.218	91	790045	97.362	ug/l	100
83) tert-Butylbenzene	13.435	119	785331	99.349	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	913037	101.908	ug/l	99
85) sec-Butylbenzene	13.612	105	1082055	102.580	ug/l	100
86) p-Isopropyltoluene	13.729	119	918730	105.328	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	463998	94.046	ug/l	100
88) 1,4-Dichlorobenzene	13.812	146	465087	93.346	ug/l	99
89) n-Butylbenzene	14.053	91	832034	105.197	ug/l	99
90) Hexachloroethane	14.329	117	168712	95.289	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	444319	91.843	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	59837	88.446	ug/l	98
93) 1,2,4-Trichlorobenzene	15.388	180	245477	103.017	ug/l	99
94) Hexachlorobutadiene	15.500	225	105801	89.034	ug/l	98
95) Naphthalene	15.635	128	784876	99.356	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	233126	97.736	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084213.D
Acq On : 30 Sep 2024 12:25
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024

Quant Time: Oct 01 06:48:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\  
Data File : VN084213.D  
Acq On    : 30 Sep 2024  12:25  
Operator  : JC\MD  
Sample    : VSTDICC100  
Misc      : 5.0mL/MSVOA_N/WATER  
ALS Vial  : 4   Sample Multiplier: 1
```


Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024


Quant Time: Oct 01 06:48:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion: 168 Resp: 192402
Ion Ratio Lower Upper
168 100
99 64.4 54.2 81.2




Manual Integrations
APPROVED

Reviewed By : John Carlone 10/01/2024
Supervised By : Mahesh Dadoda 10/01/2024




#2
Dichlorodifluoromethane
Concen: 95.796 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 85 Resp: 218015
Ion Ratio Lower Upper
85 100
87 31.6 17.6 52.9



Abundance Scan 69 (2.359 min): VN084214.D\data.ms (-63)



#3

Chloromethane

Concen: 91.628 ug/l

RT: 2.359 min Scan# 69

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 50 Resp: 238104

Ion Ratio Lower Upper

50 100


52 34.7 26.8 40.2

Manual Integrations



APPROVED

Reviewed By :John Carlone 10/01/2024


Supervised By :Mahesh Dadoda 10/01/2024



Abundance Scan 69 (2.359 min): VN084213.D\data.ms (-18)



Abundance Scan 96 (2.518 min): VN084214.D\data.ms (-88)



#4

Vinyl Chloride

Concen: 93.338 ug/l

RT: 2.512 min Scan# 95

Delta R.T. -0.006 min

Lab File: VN084213.D


Acq: 30 Sep 2024 12:25

Tgt Ion: 62 Resp: 234942



Ion Ratio Lower Upper


62 100

64 32.2 26.2 39.4



Abundance Scan 95 (2.512 min): VN084213.D\data.ms (-45)





#5

Bromomethane

Concen: 85.288 ug/l

RT: 2.936 min Scan# 16

Delta R.T. -0.018 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 94 Resp: 141604

Ion Ratio Lower Upper

94 100





96 97.2 72.9 109.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#6

Chloroethane

Concen: 80.099 ug/l

RT: 3.112 min Scan# 197

Delta R.T. -0.006 min

Lab File: VN084213.D




Acq: 30 Sep 2024 12:25


Tgt Ion: 64 Resp: 144394

Ion Ratio Lower Upper

64 100

66 33.7 23.5 35.3





#7

Trichlorofluoromethane

Concen: 93.608 ug/l

RT: 3.495 min Scan# 26

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:101 Resp: 366907

Ion Ratio Lower Upper

101 100





103 66.3 51.8 77.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#8

Diethyl Ether

Concen: 93.854 ug/l

RT: 3.953 min Scan# 340

Delta R.T. -0.006 min

Lab File: VN084213.D




Acq: 30 Sep 2024 12:25


Tgt Ion: 74 Resp: 136436

Ion Ratio Lower Upper

74 100


45 96.5 47.7 143.1





#9
1,1,2-Trichlorotrifluoroethane
Concen: 92.811 ug/l
RT: 4.365 min Scan# 411
Delta R.T. -0.006 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100



Tgt Ion:101 Resp: 208443
Ion Ratio Lower Upper
101 100
85 46.2 36.2 54.4
151 74.1 59.3 88.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#10
Methyl Iodide
Concen: 97.248 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.006 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion:142 Resp: 281269
Ion Ratio Lower Upper
142 100
127 46.5 35.3 52.9
141 14.9 11.0 16.6





#11

Tert butyl alcohol

Concen: 414.018 ug/l

RT: 5.518 min Scan# 607

Delta R.T. -0.006 min

Lab File: VN084213.D


Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100



Tgt Ion: 59 Resp: 194756

Ion Ratio Lower Upper

59 100




57 9.9 8.2 12.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#12

1,1-Dichloroethene

Concen: 95.491 ug/l

RT: 4.336 min Scan# 405

Delta R.T. -0.006 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion: 96 Resp: 206863


Ion Ratio Lower Upper

96 100

61 169.2 128.6 192.8

98 63.7 51.7 77.5





#13

Acrolein

Concen: 466.304 ug/l

RT: 4.177 min Scan# 37

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 56 Resp: 251811

Ion Ratio Lower Upper

56 100





55 70.4 57.2 85.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#14

Allyl chloride

Concen: 92.526 ug/l

RT: 5.018 min Scan# 521

Delta R.T. -0.006 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion: 41 Resp: 348108


Ion Ratio Lower Upper

41 100

39 73.7 59.0 88.6

76 36.4 29.8 44.6





#15

Acrylonitrile

Concen: 448.931 ug/l

RT: 5.712 min Scan# 63

Delta R.T. -0.006 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC100

Tgt Ion:	53	Resp:	533270
Ion	Ratio	Lower	Upper
53	100		
52	81.8	66.2	99.4
51	35.6	28.8	43.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#16

Acetone

Concen: 470.138 ug/l




RT: 4.424 min Scan# 420


Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Tgt Ion:	43	Resp:	576139
Ion	Ratio	Lower	Upper
43	100		
58	31.3	24.4	36.6





#17

Carbon Disulfide

Concen: 89.090 ug/l

RT: 4.712 min Scan# 469

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 76 Resp: 611042

Ion Ratio Lower Upper

76 100





78 9.2 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#18

Methyl Acetate

Concen: 87.499 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN084213.D




Acq: 30 Sep 2024 12:25


Tgt Ion: 43 Resp: 233708

Ion Ratio Lower Upper

43 100

74 23.4 19.2 28.8





#19

Methyl tert-butyl Ether

Concen: 96.061 ug/l

RT: 5.794 min Scan# 65

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 73 Resp: 702366

Ion Ratio Lower Upper

73 100





57 21.8 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#20

Methylene Chloride

Concen: 91.809 ug/l

RT: 5.277 min Scan# 565

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Tgt Ion: 84 Resp: 228706




Ion Ratio Lower Upper


84 100

49 119.6 107.2 160.8

51 36.6 31.8 47.8

86 62.5 52.9 79.3





#21

trans-1,2-Dichloroethene

Concen: 94.160 ug/l

RT: 5.783 min Scan# 65

Delta R.T. -0.006 min

Lab File: VN084213.D


Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100



Tgt Ion: 96 Resp: 213704

Ion Ratio Lower Upper

96 100

61 145.8 108.4 162.6




98 63.2 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#22

Diisopropyl ether

Concen: 95.811 ug/l

RT: 6.671 min Scan# 802

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Tgt Ion: 45 Resp: 745614




Ion Ratio Lower Upper


45 100

43 50.8 42.3 63.5

87 27.4 21.8 32.6

59 11.3 9.6 14.4





#23

Vinyl Acetate

Concen: 498.979 ug/l

RT: 6.600 min Scan# 790

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 43 Resp: 2895429

Ion Ratio Lower Upper

43 100





86 10.4 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#24

1,1-Dichloroethane

Concen: 95.021 ug/l

RT: 6.565 min Scan# 784

Delta R.T. -0.006 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion: 63 Resp: 413523


Ion Ratio Lower Upper

63 100

98 7.3 3.4 10.2

100 4.4 2.0 5.8





#25

2-Butanone

Concen: 454.946 ug/l

RT: 7.483 min Scan# 940

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 43 Resp: 759354

Ion Ratio Lower Upper

43 100





72 26.6 21.4 32.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#26

2,2-Dichloropropane

Concen: 96.560 ug/l

RT: 7.488 min Scan# 941

Delta R.T. 0.000 min

Lab File: VN084213.D




Acq: 30 Sep 2024 12:25


Tgt Ion: 77 Resp: 378893

Ion Ratio Lower Upper

77 100

97 21.7 10.8 32.3









#27
cis-1,2-Dichloroethene
Concen: 93.962 ug/l
RT: 7.483 min Scan# 941
Delta R.T. -0.006 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 96 Resp: 257696
Ion Ratio Lower Upper
96 100
61 151.7 0.0 301.8
98 63.8 0.0 128.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#28
Bromochloromethane
Concen: 93.497 ug/l
RT: 7.812 min Scan# 996
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 49 Resp: 181546
Ion Ratio Lower Upper
49 100
129 1.7 0.0 3.6
130 72.1 58.0 87.0









#29
Tetrahydrofuran
Concen: 440.124 ug/l
RT: 7.836 min Scan# 1000
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion:	42	Resp:	453342
Ion	Ratio	Lower	Upper
42	100		
72	44.0	34.6	51.8
71	41.2	32.2	48.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#30
Chloroform
Concen: 92.156 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion:	83	Resp:	416595
Ion	Ratio	Lower	Upper
83	100		
85	63.5	51.0	76.4





#31

Cyclohexane

Concen: 88.380 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 56 Resp: 373445

Ion Ratio Lower Upper

56 100

69 31.2 23.8 35.6





84 85.7 68.8 103.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#32

1,1,1-Trichloroethane

Concen: 94.476 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.006 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion: 97 Resp: 383566


Ion Ratio Lower Upper

97 100

99 63.3 51.2 76.8

61 45.2 37.5 56.3









#33
1,2-Dichloroethane-d4
Concen: 90.797 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 65 Resp: 259018
Ion Ratio Lower Upper
65 100
67 51.9 0.0 102.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion:114 Resp: 324582
Ion Ratio Lower Upper
114 100
63 22.0 0.0 43.8
88 17.2 0.0 31.6





#35

Dibromofluoromethane

Concen: 93.542 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:113 Resp: 200166

Ion Ratio Lower Upper

113 100

111 101.6 83.3 124.9





192 17.8 13.5 20.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#36

1,1-Dichloropropene

Concen: 98.969 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion: 75 Resp: 307616


Ion Ratio Lower Upper

75 100

110 33.5 16.9 50.7

77 30.5 24.7 37.1





#37

Ethyl Acetate

Concen: 92.073 ug/l

RT: 7.559 min Scan# 953

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC100

Tgt Ion:	43	Resp:	293597
Ion	Ratio	Lower	Upper
43	100		
61	13.3	10.7	16.1
70	10.7	8.8	13.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#38

Carbon Tetrachloride

Concen: 98.077 ug/l




RT: 8.359 min Scan# 1089


Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Tgt Ion:	117	Resp:	334398
Ion	Ratio	Lower	Upper
117	100		
119	95.6	78.7	118.1
121	29.8	24.4	36.6





#39

Methylcyclohexane

Concen: 106.466 ug/l

RT: 9.600 min Scan# 13

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 83 Resp: 368299

Ion Ratio Lower Upper

83 100

55 79.2 62.8 94.2





98 48.5 38.8 58.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#40

Benzene

Concen: 96.145 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VN084213.D




Acq: 30 Sep 2024 12:25


Tgt Ion: 78 Resp: 931076

Ion Ratio Lower Upper

78 100

77 23.7 19.1 28.7





#41

Methacrylonitrile

Concen: 92.581 ug/l

RT: 7.777 min Scan# 990

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 41 Resp: 160270

Ion Ratio Lower Upper

41 100

39 56.6 43.3 64.9

67 72.4 56.1 84.1





52 30.4 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#42

1,2-Dichloroethane

Concen: 94.900 ug/l

RT: 8.671 min Scan# 1142

Delta R.T. 0.000 min

Lab File: VN084213.D




Acq: 30 Sep 2024 12:25


Tgt Ion: 62 Resp: 311563

Ion Ratio Lower Upper

62 100

98 9.2 0.0 18.2





#43

Isopropyl Acetate

Concen: 83.581 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC100

Tgt Ion:	43	Resp:	521814
Ion	Ratio	Lower	Upper
43	100		
61	25.4	20.4	30.6
87	12.6	10.3	15.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#44

Trichloroethene

Concen: 95.798 ug/l




RT: 9.347 min Scan# 1257


Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Tgt Ion:	130	Resp:	216678
Ion	Ratio	Lower	Upper
130	100		
95	104.9	0.0	203.0





#45

1,2-Dichloropropane

Concen: 98.095 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 63 Resp: 224724

Ion Ratio Lower Upper

63 100





65 32.2 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#46

Dibromomethane

Concen: 100.091 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion: 93 Resp: 154739


Ion Ratio Lower Upper

93 100

95 82.2 65.0 97.4

174 89.5 71.8 107.8





#47

Bromodichloromethane

Concen: 98.526 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC100

Tgt Ion:	83	Resp:	338406
Ion	Ratio	Lower	Upper
83	100		
85	64.1	51.5	77.3
127	8.4	5.7	8.5

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#48

Methyl methacrylate

Concen: 98.163 ug/l




RT: 9.677 min Scan# 1313


Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Tgt Ion:	41	Resp:	244061
Ion	Ratio	Lower	Upper
41	100		
69	81.9	67.1	100.7
39	57.1	45.8	68.6





#49

1,4-Dioxane

Concen: 1810.110 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 88 Resp: 82854

Ion Ratio Lower Upper

88 100

43 30.8 25.3 37.9





58 70.9 56.4 84.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#50

Toluene-d8

Concen: 98.079 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN084213.D




Acq: 30 Sep 2024 12:25


Tgt Ion: 98 Resp: 772146

Ion Ratio Lower Upper

98 100

100 65.5 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 479.699 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 43 Resp: 1456154
Ion Ratio Lower Upper
43 100
58 39.3 31.6 47.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#52
Toluene
Concen: 99.324 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 92 Resp: 586544
Ion Ratio Lower Upper
92 100
91 171.0 137.3 205.9





#53

t-1,3-Dichloropropene

Concen: 104.280 ug/l

RT: 10.835 min Scan# 1510

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 75 Resp: 365099

Ion Ratio Lower Upper

75 100





77 32.3 25.4 38.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#54

cis-1,3-Dichloropropene

Concen: 101.975 ug/l

RT: 10.312 min Scan# 1421

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion: 75 Resp: 384030


Ion Ratio Lower Upper

75 100

77 32.5 24.2 36.2

39 49.2 38.7 58.1





#55

1,1,2-Trichloroethane

Concen: 97.170 ug/l

RT: 11.018 min Scan# 1540

Delta R.T. 0.006 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 97 Resp: 205711

Ion Ratio Lower Upper

97 100

83 83.8 71.0 106.4

85 54.4 41.7 62.5





99 62.2 50.1 75.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#56

Ethyl methacrylate

Concen: 103.608 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion: 69 Resp: 360916


Ion Ratio Lower Upper

69 100

41 67.4 53.3 79.9

39 36.1 28.8 43.2





#57

1,3-Dichloropropane

Concen: 95.326 ug/l

RT: 11.165 min Scan# 1565

Delta R.T. 0.006 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 76 Resp: 360601

Ion Ratio Lower Upper

76 100





78 32.6 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#58

2-Chloroethyl Vinyl ether

Concen: 538.640 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN084213.D




Acq: 30 Sep 2024 12:25


Tgt Ion: 63 Resp: 870416

Ion Ratio Lower Upper

63 100

106 27.5 21.8 32.6





#59

2-Hexanone

Concen: 487.363 ug/l

RT: 11.194 min Scan# 1571

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 1105736

Ion Ratio Lower Upper

43 100





58 55.0 27.3 81.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#60

Dibromochloromethane

Concen: 99.721 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. 0.000 min

Lab File: VN084213.D




Acq: 30 Sep 2024 12:25


Tgt Ion:129 Resp: 249435

Ion Ratio Lower Upper

129 100

127 78.1 38.3 114.9









#61
1,2-Dibromoethane
Concen: 96.694 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 107 Resp: 212705
Ion Ratio Lower Upper
107 100
109 94.0 75.4 113.0




Manual Integrations
APPROVED


Reviewed By : John Carlone 10/01/2024
Supervised By : Mahesh Dadoda 10/01/2024



#62
4-Bromofluorobenzene
Concen: 101.065 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 95 Resp: 289867
Ion Ratio Lower Upper
95 100
174 71.4 0.0 145.2
176 68.2 0.0 140.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion:117 Resp: 288467

Ion Ratio Lower Upper

117 100

82 59.7 47.2 70.8





119 32.5 25.4 38.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#64

Tetrachloroethene

Concen: 92.608 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Tgt Ion:164 Resp: 186069




Ion Ratio Lower Upper


164 100

166 127.4 101.0 151.4

129 102.4 80.8 121.2

131 95.7 78.1 117.1





#65

Chlorobenzene

Concen: 96.376 ug/l

RT: 11.888 min Scan# 1689

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:112 Resp: 616423

Ion Ratio Lower Upper

112 100





114 32.0 27.0 40.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#66

1,1,1,2-Tetrachloroethane

Concen: 96.546 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion:131 Resp: 212695


Ion Ratio Lower Upper

131 100

133 94.6 47.9 143.7

119 64.8 32.8 98.3





#67

Ethyl Benzene

Concen: 101.385 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 91 Resp: 1146786

Ion Ratio Lower Upper

91 100





106 29.5 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#68

m/p-Xylenes

Concen: 204.437 ug/l

RT: 12.071 min Scan# 1720

Delta R.T. 0.000 min

Lab File: VN084213.D




Acq: 30 Sep 2024 12:25


Tgt Ion:106 Resp: 855381

Ion Ratio Lower Upper

106 100

91 208.5 167.1 250.7





#69

o-Xylene

Concen: 104.101 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:106 Resp: 412098

Ion Ratio Lower Upper

106 100





91 222.0 110.2 330.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#70

Styrene

Concen: 106.165 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion:104 Resp: 711855


Ion Ratio Lower Upper

104 100

78 52.5 42.7 64.1

103 54.7 43.6 65.4





#71

Bromoform

Concen: 100.412 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion:	173	Resp:	162862
Ion	Ratio	Lower	Upper
173	100		
175	49.3	23.5	70.6
254	0.0	0.0	0.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.788 min Scan# 2012


Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Tgt Ion:	152	Resp:	142880
Ion	Ratio	Lower	Upper
152	100		
115	63.0	31.3	93.9
150	188.1	0.0	349.8





#73

Isopropylbenzene

Concen: 97.936 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion:105 Resp: 1067798

Ion Ratio Lower Upper

105 100





120 25.9 12.9 38.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#74

N-ethyl acetate

Concen: 94.716 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Tgt Ion: 43 Resp: 467641




Ion Ratio Lower Upper


43 100

70 42.8 34.4 51.6

55 24.2 19.6 29.4

61 24.6 19.6 29.4





#75

1,1,2,2-Tetrachloroethane

Concen: 87.285 ug/l

RT: 12.935 min Scan# 18

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion:	83	Resp:	286161
Ion	Ratio	Lower	Upper
83	100		
131	10.2	4.9	14.7
85	65.7	31.9	95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#76

1,2,3-Trichloropropane

Concen: 85.433 ug/l m




RT: 12.994 min Scan# 1877


Delta R.T. 0.006 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Tgt Ion:	75	Resp:	243293
Ion	Ratio	Lower	Upper
75	100		
77	229.8	96.9	290.7





#77

Bromobenzene

Concen: 92.425 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion:	156	Resp:	243075
Ion	Ratio	Lower	Upper
156	100		
77	230.0	115.1	345.3
158	98.3	48.9	146.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#78

n-propylbenzene

Concen: 101.754 ug/l




RT: 13.035 min Scan# 1884


Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Tgt Ion:	91	Resp:	1285399
Ion	Ratio	Lower	Upper
91	100		
120	21.9	10.8	32.4





#79

2-Chlorotoluene

Concen: 95.567 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion: 91 Resp: 770341

Ion Ratio Lower Upper

91 100





126 31.4 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#80

1,3,5-Trimethylbenzene

Concen: 101.869 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VN084213.D




Acq: 30 Sep 2024 12:25


Tgt Ion:105 Resp: 905999

Ion Ratio Lower Upper

105 100

120 47.3 23.7 71.1





#81

trans-1,4-Dichloro-2-butene

Concen: 95.579 ug/l

RT: 12.735 min Scan# 1833

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 75 Resp: 116289

Ion Ratio Lower Upper

75 100

53 95.1 78.0 117.0





89 44.4 35.5 53.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#82

4-Chlorotoluene

Concen: 97.362 ug/l

RT: 13.218 min Scan# 1915

Delta R.T. 0.000 min

Lab File: VN084213.D




Acq: 30 Sep 2024 12:25


Tgt Ion: 91 Resp: 790045

Ion Ratio Lower Upper

91 100

126 31.2 15.7 47.1





#83

tert-Butylbenzene

Concen: 99.349 ug/l

RT: 13.435 min Scan# 19

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:119 Resp: 785331

Ion Ratio Lower Upper

119 100

91 67.3 34.1 102.3





134 25.2 12.8 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#84

1,2,4-Trimethylbenzene

Concen: 101.908 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.000 min

Lab File: VN084213.D




Acq: 30 Sep 2024 12:25


Tgt Ion:105 Resp: 913037

Ion Ratio Lower Upper

105 100

120 43.7 22.3 66.8





#85

sec-Butylbenzene

Concen: 102.580 ug/l

RT: 13.612 min Scan# 19

Delta R.T. -0.006 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion:105 Resp: 1082055

Ion Ratio Lower Upper

105 100





134 19.4 9.8 29.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#86

p-Isopropyltoluene

Concen: 105.328 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion:119 Resp: 918730


Ion Ratio Lower Upper

119 100

134 25.2 12.6 37.8

91 25.1 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 94.046 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100

Tgt Ion:146 Resp: 463998

Ion Ratio Lower Upper

146 100

111 43.1 21.8 65.3





148 63.9 31.9 95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#88

1,4-Dichlorobenzene

Concen: 93.346 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion:146 Resp: 465087


Ion Ratio Lower Upper

146 100

111 41.7 21.2 63.6

148 64.4 31.9 95.9





#89

n-Butylbenzene

Concen: 105.197 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC100

Tgt Ion:	91	Resp:	832034
Ion	Ratio	Lower	Upper
91	100		
92	53.8	26.6	79.7
134	24.5	12.3	36.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#90

Hexachloroethane

Concen: 95.289 ug/l




RT: 14.329 min Scan# 2104


Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Tgt Ion:	117	Resp:	168712
Ion	Ratio	Lower	Upper
117	100		
201	67.4	34.4	103.2





#91

1,2-Dichlorobenzene

Concen: 91.843 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. 0.000 min

Lab File: VN084213.D


Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC100



Tgt Ion:	146	Resp:	444319
Ion	Ratio	Lower	Upper
146	100		
111	44.4	22.0	66.0
148	63.1	32.1	96.3

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#92

1,2-Dibromo-3-Chloropropane

Concen: 88.446 ug/l




RT: 14.717 min Scan# 2170


Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Tgt Ion:	75	Resp:	59837
Ion	Ratio	Lower	Upper
75	100		
155	77.6	38.1	114.5
157	96.3	49.6	149.0





#93

1,2,4-Trichlorobenzene

Concen: 103.017 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. 0.000 min

Lab File: VN084213.D


Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC100



Tgt Ion:	180	Resp:	245477
Ion	Ratio	Lower	Upper
180	100		
182	94.6	47.0	141.0
145	33.9	16.4	49.4

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#94

Hexachlorobutadiene


Concen: 89.034 ug/l

RT: 15.500 min Scan# 2303



Delta R.T. 0.000 min


Lab File: VN084213.D

Acq: 30 Sep 2024 12:25



Tgt Ion:	225	Resp:	105801
Ion	Ratio	Lower	Upper
225	100		
223	62.0	32.4	97.2
227	64.0	32.4	97.0





#95

Naphthalene

Concen: 99.356 ug/l

RT: 15.635 min Scan# 2326

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion:128 Resp: 784876

Ion Ratio Lower Upper

128 100

127 12.7 10.5 15.7





129 10.7 8.7 13.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#96

1,2,3-Trichlorobenzene

Concen: 97.736 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion:180 Resp: 233126

Ion Ratio Lower Upper

180 100

182 95.8 48.2 144.6

145 35.9 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084214.D
 Acq On : 30 Sep 2024 12:49
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/01/2024
 Supervised By : Mahesh Dadoda 10/01/2024

Quant Time: Oct 01 06:48:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	175379	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	304127	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	264820	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	127701	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	129300	49.725	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.440%
35) Dibromofluoromethane	8.165	113	98531	49.143	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.280%
50) Toluene-d8	10.565	98	377886	51.228	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.460%
62) 4-Bromofluorobenzene	12.847	95	137388	51.124	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.240%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	107530	51.835	ug/l	100
3) Chloromethane	2.359	50	117689	49.686	ug/l	100
4) Vinyl Chloride	2.518	62	117031	51.007	ug/l	100
5) Bromomethane	2.953	94	75848	50.118	ug/l	100
6) Chloroethane	3.118	64	73440	44.693	ug/l	100
7) Trichlorofluoromethane	3.495	101	185982	52.055	ug/l	100
8) Diethyl Ether	3.959	74	68385	51.608	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.371	101	103552	50.583	ug/l	100
10) Methyl Iodide	4.595	142	141310	53.600	ug/l	100
11) Tert butyl alcohol	5.524	59	107986	251.850	ug/l	100
12) 1,1-Dichloroethene	4.342	96	102934	52.128	ug/l	100
13) Acrolein	4.177	56	124940	253.821	ug/l	100
14) Allyl chloride	5.024	41	174518	50.889	ug/l	100
15) Acrylonitrile	5.718	53	277346	256.145	ug/l	100
16) Acetone	4.424	43	299515	268.132	ug/l	100
17) Carbon Disulfide	4.712	76	302183	48.335	ug/l	100
18) Methyl Acetate	5.018	43	124007	50.934	ug/l	100
19) Methyl tert-butyl Ether	5.794	73	356478	53.487	ug/l	100
20) Methylene Chloride	5.277	84	114918	50.609	ug/l	100
21) trans-1,2-Dichloroethene	5.789	96	109048	52.711	ug/l	100
22) Diisopropyl ether	6.671	45	379516	53.501	ug/l	100
23) Vinyl Acetate	6.600	43	1444909	273.176	ug/l	100
24) 1,1-Dichloroethane	6.571	63	209222	52.742	ug/l	100
25) 2-Butanone	7.483	43	396369	260.524	ug/l	100
26) 2,2-Dichloropropane	7.488	77	189745	53.050	ug/l	100
27) cis-1,2-Dichloroethene	7.488	96	129873	51.951	ug/l	100
28) Bromochloromethane	7.812	49	86584	48.919	ug/l	100
29) Tetrahydrofuran	7.835	42	248328	264.488	ug/l	100
30) Chloroform	7.965	83	211178	51.249	ug/l	100
31) Cyclohexane	8.253	56	187282	48.624	ug/l	100
32) 1,1,1-Trichloroethane	8.171	97	193315	52.237	ug/l	100
36) 1,1-Dichloropropene	8.371	75	154855	53.172	ug/l	100
37) Ethyl Acetate	7.559	43	153024	51.216	ug/l	100
38) Carbon Tetrachloride	8.359	117	167086	52.301	ug/l	100
39) Methylcyclohexane	9.600	83	177433	54.741	ug/l	100
40) Benzene	8.606	78	472170	52.037	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084214.D
 Acq On : 30 Sep 2024 12:49
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/01/2024
 Supervised By : Mahesh Dadoda 10/01/2024

Quant Time: Oct 01 06:48:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	87365	53.861	ug/l	100
42) 1,2-Dichloroethane	8.671	62	160606	52.210	ug/l	100
43) Isopropyl Acetate	8.688	43	272992	46.667	ug/l	100
44) Trichloroethene	9.347	130	110119	51.961	ug/l	100
45) 1,2-Dichloropropane	9.618	63	114143	53.176	ug/l	100
46) Dibromomethane	9.706	93	79007	54.542	ug/l	100
47) Bromodichloromethane	9.888	83	169263	52.595	ug/l	100
48) Methyl methacrylate	9.677	41	127042	54.534	ug/l	100
49) 1,4-Dioxane	9.694	88	44512	1037.858	ug/l	100
51) 4-Methyl-2-Pentanone	10.441	43	772927	271.750	ug/l	100
52) Toluene	10.629	92	295099	53.332	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	179453	54.703	ug/l	100
54) cis-1,3-Dichloropropene	10.312	75	193952	54.966	ug/l	100
55) 1,1,2-Trichloroethane	11.012	97	106288	53.583	ug/l	100
56) Ethyl methacrylate	10.871	69	183059	56.085	ug/l	100
57) 1,3-Dichloropropane	11.159	76	187372	52.864	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	426364	281.593	ug/l	100
59) 2-Hexanone	11.194	43	588361	276.767	ug/l	100
60) Dibromochloromethane	11.359	129	126992	54.184	ug/l	100
61) 1,2-Dibromoethane	11.465	107	108817	52.794	ug/l	100
64) Tetrachloroethene	11.100	164	94966	51.486	ug/l	100
65) Chlorobenzene	11.888	112	309849	52.770	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	106981	52.897	ug/l	100
67) Ethyl Benzene	11.959	91	561770	54.100	ug/l	100
68) m/p-Xylenes	12.071	106	427131	111.201	ug/l	100
69) o-Xylene	12.394	106	204854	56.370	ug/l	100
70) Styrene	12.412	104	347383	56.434	ug/l	100
71) Bromoform	12.576	173	83487	56.070	ug/l #	100
73) Isopropylbenzene	12.694	105	527625	54.145	ug/l	100
74) N-ethyl acetate	12.494	43	237131	53.738	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.935	83	151079	51.560	ug/l	100
76) 1,2,3-Trichloropropane	12.988	75	143676m	56.449	ug/l	
77) Bromobenzene	12.982	156	120918	51.442	ug/l	100
78) n-propylbenzene	13.035	91	626503	55.490	ug/l	100
79) 2-Chlorotoluene	13.123	91	384795	53.411	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	445807	56.084	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	57572	52.943	ug/l	100
82) 4-Chlorotoluene	13.217	91	386854	53.341	ug/l	100
83) tert-Butylbenzene	13.435	119	386532	54.711	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	449720	56.161	ug/l	100
85) sec-Butylbenzene	13.617	105	529076	56.119	ug/l	100
86) p-Isopropyltoluene	13.729	119	448148	57.485	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	228219	51.755	ug/l	100
88) 1,4-Dichlorobenzene	13.812	146	227835	51.163	ug/l	100
89) n-Butylbenzene	14.053	91	394831	55.854	ug/l	100
90) Hexachloroethane	14.329	117	81626	51.582	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	218378	50.506	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	30379	50.241	ug/l	100
93) 1,2,4-Trichlorobenzene	15.388	180	119164	55.953	ug/l	100
94) Hexachlorobutadiene	15.500	225	50680	47.718	ug/l	100
95) Naphthalene	15.635	128	393177	55.688	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	114483	53.701	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084214.D
Acq On : 30 Sep 2024 12:49
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024

Quant Time: Oct 01 06:48:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084214.D
Acq On : 30 Sep 2024 12:49
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Quant Time: Oct 01 06:48:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 10/01/2024
Supervised By : Mahesh Dadoda 10/01/2024









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:168 Resp: 175379
Ion Ratio Lower Upper
168 100
99 67.7 54.2 81.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#2
Dichlorodifluoromethane
Concen: 51.835 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 85 Resp: 107530
Ion Ratio Lower Upper
85 100
87 35.3 17.6 52.9





#3

Chloromethane

Concen: 49.686 ug/l

RT: 2.359 min Scan# 69

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 50 Resp: 117689

Ion Ratio Lower Upper

50 100





52 33.5 26.8 40.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#4

Vinyl Chloride

Concen: 51.007 ug/l

RT: 2.518 min Scan# 96

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion: 62 Resp: 117031

Ion Ratio Lower Upper

62 100

64 32.8 26.2 39.4





#5

Bromomethane

Concen: 50.118 ug/l

RT: 2.953 min Scan# 170

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 94 Resp: 75848

Ion Ratio Lower Upper

94 100





96 91.1 72.9 109.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#6

Chloroethane

Concen: 44.693 ug/l

RT: 3.118 min Scan# 198

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion: 64 Resp: 73440

Ion Ratio Lower Upper

64 100

66 29.4 23.5 35.3





#7

Trichlorofluoromethane

Concen: 52.055 ug/l

RT: 3.495 min Scan# 262

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:101 Resp: 185982

Ion Ratio Lower Upper

101 100





103 64.7 51.8 77.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#8

Diethyl Ether

Concen: 51.608 ug/l

RT: 3.959 min Scan# 341

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion: 74 Resp: 68385

Ion Ratio Lower Upper

74 100

45 95.4 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.583 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion	Ratio	Lower	Upper
101	100		
85	45.3	36.2	54.4
151	74.1	59.3	88.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#10
Methyl Iodide
Concen: 53.600 ug/l
RT: 4.595 min Scan# 449
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion	Ratio	Lower	Upper
142	100		
127	44.1	35.3	52.9
141	13.8	11.0	16.6





#11

Tert butyl alcohol

Concen: 251.850 ug/l

RT: 5.524 min Scan# 607

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 59 Resp: 107986

Ion Ratio Lower Upper

59 100





57 10.2 8.2 12.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#12

1,1-Dichloroethene

Concen: 52.128 ug/l

RT: 4.342 min Scan# 406

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49




Tgt Ion: 96 Resp: 102934


Ion Ratio Lower Upper

96 100

61 160.7 128.6 192.8

98 64.6 51.7 77.5





#13

Acrolein

Concen: 253.821 ug/l

RT: 4.177 min Scan# 378

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 56 Resp: 124940

Ion Ratio Lower Upper

56 100





55 71.5 57.2 85.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#14

Allyl chloride

Concen: 50.889 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49




Tgt Ion: 41 Resp: 174518


Ion Ratio Lower Upper

41 100

39 73.8 59.0 88.6

76 37.2 29.8 44.6





#15

Acrylonitrile

Concen: 256.145 ug/l

RT: 5.718 min Scan# 640

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion:	53	Resp:	277346
Ion	Ratio	Lower	Upper
53	100		
52	82.8	66.2	99.4
51	36.0	28.8	43.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#16

Acetone

Concen: 268.132 ug/l




RT: 4.424 min Scan# 420


Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion:	43	Resp:	299515
Ion	Ratio	Lower	Upper
43	100		
58	30.5	24.4	36.6





#17

Carbon Disulfide

Concen: 48.335 ug/l

RT: 4.712 min Scan# 469

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 302183

Ion Ratio Lower Upper

76 100





78 9.1 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#18

Methyl Acetate

Concen: 50.934 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion: 43 Resp: 124007

Ion Ratio Lower Upper

43 100

74 24.0 19.2 28.8





#19

Methyl tert-butyl Ether

Concen: 53.487 ug/l

RT: 5.794 min Scan# 653

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 73 Resp: 356478

Ion Ratio Lower Upper

73 100





57 22.1 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#20

Methylene Chloride

Concen: 50.609 ug/l

RT: 5.277 min Scan# 565

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion: 84 Resp: 114918




Ion Ratio Lower Upper


84 100

49 134.0 107.2 160.8

51 39.8 31.8 47.8

86 66.1 52.9 79.3





#21

trans-1,2-Dichloroethene

Concen: 52.711 ug/l

RT: 5.789 min Scan# 652

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 96 Resp: 109048

Ion Ratio Lower Upper

96 100

61 135.5 108.4 162.6





98 62.3 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#22

Diisopropyl ether

Concen: 53.501 ug/l

RT: 6.671 min Scan# 802

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion: 45 Resp: 379516




Ion Ratio Lower Upper


45 100

43 52.9 42.3 63.5

87 27.2 21.8 32.6

59 12.0 9.6 14.4





#23

Vinyl Acetate

Concen: 273.176 ug/l

RT: 6.600 min Scan# 790

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 1444909

Ion Ratio Lower Upper

43 100





86 10.7 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#24

1,1-Dichloroethane

Concen: 52.742 ug/l

RT: 6.571 min Scan# 785

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49




Tgt Ion: 63 Resp: 209222


Ion Ratio Lower Upper

63 100

98 6.8 3.4 10.2

100 3.9 2.0 5.8





#25

2-Butanone

Concen: 260.524 ug/l

RT: 7.483 min Scan# 940

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion: 43 Resp: 396369

Ion Ratio Lower Upper

43 100

72 26.8 21.4 32.2

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#26

2,2-Dichloropropane

Concen: 53.050 ug/l

RT: 7.488 min Scan# 941

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion: 77 Resp: 189745

Ion Ratio Lower Upper

77 100

97 21.5 10.8 32.3









#27
cis-1,2-Dichloroethene
Concen: 51.951 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:	96	Resp:	129873
Ion	Ratio	Lower	Upper
96	100		
61	150.9	0.0	301.8
98	64.1	0.0	128.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#28
Bromochloromethane
Concen: 48.919 ug/l
RT: 7.812 min Scan# 996
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion:	49	Resp:	86584
Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	3.6
130	72.5	58.0	87.0









#29
Tetrahydrofuran
Concen: 264.488 ug/l
RT: 7.835 min Scan# 1000
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 42 Resp: 248328
Ion Ratio Lower Upper
42 100
72 43.2 34.6 51.8
71 40.3 32.2 48.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#30
Chloroform
Concen: 51.249 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 83 Resp: 211178
Ion Ratio Lower Upper
83 100
85 63.7 51.0 76.4





#31

Cyclohexane

Concen: 48.624 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 56 Resp: 187282

Ion Ratio Lower Upper

56 100

69 29.7 23.8 35.6





84 86.0 68.8 103.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#32

1,1,1-Trichloroethane

Concen: 52.237 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49




Tgt Ion: 97 Resp: 193315


Ion Ratio Lower Upper

97 100

99 64.0 51.2 76.8

61 46.9 37.5 56.3









#33
1,2-Dichloroethane-d4
Concen: 49.725 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 65 Resp: 129300
Ion Ratio Lower Upper
65 100
67 51.0 0.0 102.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion:114 Resp: 304127
Ion Ratio Lower Upper
114 100
63 21.9 0.0 43.8
88 15.8 0.0 31.6





#35

Dibromofluoromethane

Concen: 49.143 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 113 Resp: 98531

Ion Ratio Lower Upper

113 100

111 104.1 83.3 124.9





192 16.9 13.5 20.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#36

1,1-Dichloropropene

Concen: 53.172 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49




Tgt Ion: 75 Resp: 154855


Ion Ratio Lower Upper

75 100

110 33.8 16.9 50.7

77 30.9 24.7 37.1





#37

Ethyl Acetate

Concen: 51.216 ug/l

RT: 7.559 min Scan# 953

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N





ClientSampleId :

VSTDICCC050

Tgt Ion:	43	Resp:	153024
Ion	Ratio	Lower	Upper
43	100		
61	13.4	10.7	16.1
70	11.0	8.8	13.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#38

Carbon Tetrachloride

Concen: 52.301 ug/l




RT: 8.359 min Scan# 1089


Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion:	117	Resp:	167086
Ion	Ratio	Lower	Upper
117	100		
119	98.4	78.7	118.1
121	30.5	24.4	36.6





#39

Methylcyclohexane

Concen: 54.741 ug/l

RT: 9.600 min Scan# 1300

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 83 Resp: 177433

Ion Ratio Lower Upper

83 100

55 78.5 62.8 94.2





98 48.5 38.8 58.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#40

Benzene

Concen: 52.037 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion: 78 Resp: 472170

Ion Ratio Lower Upper

78 100

77 23.9 19.1 28.7





#41

Methacrylonitrile

Concen: 53.861 ug/l

RT: 7.777 min Scan# 990

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion:	41	Resp:	87365
Ion	Ratio	Lower	Upper
41	100		
39	54.1	43.3	64.9
67	70.1	56.1	84.1
52	30.6	24.5	36.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#42

1,2-Dichloroethane

Concen: 52.210 ug/l




RT: 8.671 min Scan# 1142


Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion:	62	Resp:	160606
Ion	Ratio	Lower	Upper
62	100		
98	9.1	0.0	18.2





#43

Isopropyl Acetate

Concen: 46.667 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion:	43	Resp:	272992
Ion	Ratio	Lower	Upper
43	100		
61	25.5	20.4	30.6
87	12.9	10.3	15.5

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#44

Trichloroethene

Concen: 51.961 ug/l




RT: 9.347 min Scan# 1257


Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion:	130	Resp:	110119
Ion	Ratio	Lower	Upper
130	100		
95	101.5	0.0	203.0





#45

1,2-Dichloropropane

Concen: 53.176 ug/l

RT: 9.618 min Scan# 13

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 63 Resp: 114143

Ion Ratio Lower Upper

63 100





65 30.8 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#46

Dibromomethane

Concen: 54.542 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49




Tgt Ion: 93 Resp: 79007


Ion Ratio Lower Upper

93 100

95 81.2 65.0 97.4

174 89.8 71.8 107.8





#47

Bromodichloromethane

Concen: 52.595 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N





ClientSampleId :

VSTDICCC050

Tgt Ion:	83	Resp:	169263
Ion	Ratio	Lower	Upper
83	100		
85	64.4	51.5	77.3
127	7.1	5.7	8.5

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#48

Methyl methacrylate

Concen: 54.534 ug/l




RT: 9.677 min Scan# 1313


Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion:	41	Resp:	127042
Ion	Ratio	Lower	Upper
41	100		
69	83.9	67.1	100.7
39	57.2	45.8	68.6





#49

1,4-Dioxane

Concen: 1037.858 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.000 min

Lab File: VN084214.D


Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050



Tgt Ion: 88 Resp: 44512

Ion Ratio Lower Upper

88 100

43 31.6 25.3 37.9




58 70.5 56.4 84.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#50

Toluene-d8

Concen: 51.228 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion: 98 Resp: 377886

Ion Ratio Lower Upper

98 100

100 65.9 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 271.750 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 43 Resp: 772927
Ion Ratio Lower Upper
43 100
58 39.5 31.6 47.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#52
Toluene
Concen: 53.332 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 92 Resp: 295099
Ion Ratio Lower Upper
92 100
91 171.6 137.3 205.9





#53

t-1,3-Dichloropropene

Concen: 54.703 ug/l

RT: 10.835 min Scan# 1510

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 75 Resp: 179453

Ion Ratio Lower Upper

75 100





77 31.7 25.4 38.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#54

cis-1,3-Dichloropropene

Concen: 54.966 ug/l

RT: 10.312 min Scan# 1421

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49




Tgt Ion: 75 Resp: 193952


Ion Ratio Lower Upper

75 100

77 30.2 24.2 36.2

39 48.4 38.7 58.1





#55

1,1,2-Trichloroethane

Concen: 53.583 ug/l

RT: 11.012 min Scan# 1540

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 97 Resp: 106288

Ion Ratio Lower Upper

97 100

83 88.7 71.0 106.4

85 52.1 41.7 62.5





99 62.6 50.1 75.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#56

Ethyl methacrylate

Concen: 56.085 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49




Tgt Ion: 69 Resp: 183059


Ion Ratio Lower Upper

69 100

41 66.6 53.3 79.9

39 36.0 28.8 43.2





#57

1,3-Dichloropropane

Concen: 52.864 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 187372

Ion Ratio Lower Upper

76 100





78 32.2 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#58

2-Chloroethyl Vinyl ether

Concen: 281.593 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion: 63 Resp: 426364

Ion Ratio Lower Upper

63 100

106 27.2 21.8 32.6





#59

2-Hexanone

Concen: 276.767 ug/l

RT: 11.194 min Scan# 1571

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 588361

Ion Ratio Lower Upper

43 100





58 54.5 27.3 81.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#60

Dibromochloromethane

Concen: 54.184 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion:129 Resp: 126992

Ion Ratio Lower Upper

129 100

127 76.6 38.3 114.9









#61
1,2-Dibromoethane
Concen: 52.794 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 107 Resp: 108817
Ion Ratio Lower Upper
107 100
109 94.2 75.4 113.0




Manual Integrations
APPROVED


Reviewed By : John Carlone 10/01/2024
Supervised By : Mahesh Dadoda 10/01/2024



#62
4-Bromofluorobenzene
Concen: 51.124 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 95 Resp: 137388
Ion Ratio Lower Upper
95 100
174 72.6 0.0 145.2
176 70.0 0.0 140.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:117 Resp: 264820

Ion Ratio Lower Upper

117 100

82 59.0 47.2 70.8





119 31.7 25.4 38.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#64

Tetrachloroethene

Concen: 51.486 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion:164 Resp: 94966




Ion Ratio Lower Upper


164 100

166 126.2 101.0 151.4

129 101.0 80.8 121.2

131 97.6 78.1 117.1





#65

Chlorobenzene

Concen: 52.770 ug/l

RT: 11.888 min Scan# 1689

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:112 Resp: 309849

Ion Ratio Lower Upper

112 100





114 33.8 27.0 40.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#66

1,1,1,2-Tetrachloroethane

Concen: 52.897 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49




Tgt Ion:131 Resp: 106981


Ion Ratio Lower Upper

131 100

133 95.8 47.9 143.7

119 65.5 32.8 98.3





#67

Ethyl Benzene

Concen: 54.100 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 561770

Ion Ratio Lower Upper

91 100





106 30.6 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#68

m/p-Xylenes

Concen: 111.201 ug/l

RT: 12.071 min Scan# 1720

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion:106 Resp: 427131

Ion Ratio Lower Upper

106 100

91 208.9 167.1 250.7





#69

o-Xylene

Concen: 56.370 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:106 Resp: 204854

Ion Ratio Lower Upper

106 100





91 220.4 110.2 330.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#70

Styrene

Concen: 56.434 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49




Tgt Ion:104 Resp: 347383


Ion Ratio Lower Upper

104 100

78 53.4 42.7 64.1

103 54.5 43.6 65.4





#71

Bromoform

Concen: 56.070 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion:	173	Resp:	83487
Ion Ratio	Lower	Upper	
173	100		
175	47.1	23.5	70.6
254	0.0	0.0	0.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.788 min Scan# 2012


Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion:	152	Resp:	127701
Ion Ratio	Lower	Upper	
152	100		
115	62.6	31.3	93.9
150	174.9	0.0	349.8





#73

Isopropylbenzene

Concen: 54.145 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:105 Resp: 527625

Ion Ratio Lower Upper

105 100





120 25.8 12.9 38.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#74

N-ethyl acetate

Concen: 53.738 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion: 43 Resp: 237131




Ion Ratio Lower Upper


43 100

70 43.0 34.4 51.6

55 24.5 19.6 29.4

61 24.5 19.6 29.4





#75

1,1,2,2-Tetrachloroethane

Concen: 51.560 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion:	83	Resp:	151079
Ion	Ratio	Lower	Upper
83	100		
131	9.8	4.9	14.7
85	63.8	31.9	95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#76

1,2,3-Trichloropropane

Concen: 56.449 ug/l m




RT: 12.988 min Scan# 1876


Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion:	75	Resp:	143676
Ion	Ratio	Lower	Upper
75	100		
77	193.8	96.9	290.7





#77

Bromobenzene

Concen: 51.442 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. 0.000 min

Lab File: VN084214.D


Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :




VSTDICCC050



Tgt Ion:	156	Resp:	120918
Ion	Ratio	Lower	Upper
156	100		
77	230.2	115.1	345.3
158	97.9	48.9	146.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#78

n-propylbenzene

Concen: 55.490 ug/l




RT: 13.035 min Scan# 1884


Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion:	91	Resp:	626503
Ion	Ratio	Lower	Upper
91	100		
120	21.6	10.8	32.4





#79

2-Chlorotoluene

Concen: 53.411 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 384795

Ion Ratio Lower Upper

91 100





126 31.4 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#80

1,3,5-Trimethylbenzene

Concen: 56.084 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion:105 Resp: 445807

Ion Ratio Lower Upper

105 100

120 47.4 23.7 71.1









#81
trans-1,4-Dichloro-2-butene
Concen: 52.943 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion	Ratio	Lower	Upper
75	100		
53	97.5	78.0	117.0
89	44.4	35.5	53.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#82
4-Chlorotoluene
Concen: 53.341 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion	Ratio	Lower	Upper
91	100		
126	31.4	15.7	47.1





#83

tert-Butylbenzene

Concen: 54.711 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N





ClientSampleId :

VSTDICCC050

Tgt Ion:	119	Resp:	386532
Ion Ratio	Lower	Upper	
119	100		
91	68.2	34.1	102.3
134	25.6	12.8	38.4

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#84

1,2,4-Trimethylbenzene

Concen: 56.161 ug/l




RT: 13.482 min Scan# 1960


Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion:	105	Resp:	449720
Ion Ratio	Lower	Upper	
105	100		
120	44.5	22.3	66.8





#85

sec-Butylbenzene

Concen: 56.119 ug/l

RT: 13.617 min Scan# 1983

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:105 Resp: 529076

Ion Ratio Lower Upper

105 100





134 19.6 9.8 29.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#86

p-Isopropyltoluene

Concen: 57.485 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49




Tgt Ion:119 Resp: 448148


Ion Ratio Lower Upper

119 100

134 25.2 12.6 37.8

91 24.8 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 51.755 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N





ClientSampleId :

VSTDICCC050

Tgt Ion:	146	Resp:	228219
Ion	Ratio	Lower	Upper
146	100		
111	43.5	21.8	65.3
148	63.8	31.9	95.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#88

1,4-Dichlorobenzene

Concen: 51.163 ug/l




RT: 13.812 min Scan# 2016


Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion:	146	Resp:	227835
Ion	Ratio	Lower	Upper
146	100		
111	42.4	21.2	63.6
148	63.9	31.9	95.9





#89

n-Butylbenzene

Concen: 55.854 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N





ClientSampleId :

VSTDICCC050

Tgt Ion:	91	Resp:	394831
Ion	Ratio	Lower	Upper
91	100		
92	53.1	26.6	79.7
134	24.6	12.3	36.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#90

Hexachloroethane

Concen: 51.582 ug/l




RT: 14.329 min Scan# 2104


Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion:	117	Resp:	81626
Ion	Ratio	Lower	Upper
117	100		
201	68.8	34.4	103.2





#91

1,2-Dichlorobenzene

Concen: 50.506 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N





ClientSampleId :

VSTDICCC050

Tgt Ion:	146	Resp:	218378
Ion	Ratio	Lower	Upper
146	100		
111	44.0	22.0	66.0
148	64.2	32.1	96.3

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#92

1,2-Dibromo-3-Chloropropane

Concen: 50.241 ug/l




RT: 14.717 min Scan# 2170


Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion:	75	Resp:	30379
Ion	Ratio	Lower	Upper
75	100		
155	76.3	38.1	114.5
157	99.3	49.6	149.0





#93

1,2,4-Trichlorobenzene

Concen: 55.953 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. 0.000 min

Lab File: VN084214.D


Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050






Tgt Ion:	180	Resp:	119164
Ion	Ratio	Lower	Upper
180	100		
182	94.0	47.0	141.0
145	32.9	16.4	49.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#94

Hexachlorobutadiene


Concen: 47.718 ug/l

RT: 15.500 min Scan# 2303



Delta R.T. 0.000 min


Lab File: VN084214.D

Acq: 30 Sep 2024 12:49



Tgt Ion:	225	Resp:	50680
Ion	Ratio	Lower	Upper
225	100		
223	64.8	32.4	97.2
227	64.7	32.4	97.0





#95

Naphthalene

Concen: 55.688 ug/l

RT: 15.635 min Scan# 2326

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:	128	Resp:	393177
Ion	Ratio	Lower	Upper
128	100		
127	13.1	10.5	15.7
129	10.9	8.7	13.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#96

1,2,3-Trichlorobenzene

Concen: 53.701 ug/l




RT: 15.835 min Scan# 2360

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion:	180	Resp:	114483
Ion	Ratio	Lower	Upper
180	100		
182	96.4	48.2	144.6
145	34.6	17.3	51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084215.D
 Acq On : 30 Sep 2024 13:13
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/01/2024
 Supervised By : Mahesh
 Dadoda 10/01/2024

Quant Time: Oct 01 06:48:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	178087	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	307298	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	270754	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	127763	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	54409	20.606	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	41.220%#		
35) Dibromofluoromethane	8.171	113	41903	20.684	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	41.360%#		
50) Toluene-d8	10.565	98	152668	20.483	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	40.960%#		
62) 4-Bromofluorobenzene	12.847	95	55201	20.329	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	40.660%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	42709	20.275	ug/l	94
3) Chloromethane	2.359	50	48218	20.047	ug/l	99
4) Vinyl Chloride	2.518	62	47375	20.334	ug/l	96
5) Bromomethane	2.942	94	30751	20.010	ug/l	97
6) Chloroethane	3.118	64	30650	18.369	ug/l	96
7) Trichlorofluoromethane	3.495	101	74579	20.557	ug/l	95
8) Diethyl Ether	3.959	74	29090	21.619	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.377	101	42958	20.665	ug/l	99
10) Methyl Iodide	4.589	142	54769	20.458	ug/l	97
11) Tert butyl alcohol	5.518	59	45501	104.506	ug/l	98
12) 1,1-Dichloroethene	4.336	96	42450	21.171	ug/l	95
13) Acrolein	4.183	56	48261	96.553	ug/l	98
14) Allyl chloride	5.024	41	77338	22.208	ug/l	92
15) Acrylonitrile	5.718	53	115451	105.004	ug/l	99
16) Acetone	4.430	43	119948	105.747	ug/l	98
17) Carbon Disulfide	4.706	76	124366	19.590	ug/l	99
18) Methyl Acetate	5.030	43	49436	19.996	ug/l	99
19) Methyl tert-butyl Ether	5.794	73	144972	21.421	ug/l	100
20) Methylene Chloride	5.271	84	47091	20.423	ug/l	94
21) trans-1,2-Dichloroethene	5.783	96	44064	20.976	ug/l	92
22) Diisopropyl ether	6.671	45	154575	21.460	ug/l	98
23) Vinyl Acetate	6.600	43	589404	109.739	ug/l	99
24) 1,1-Dichloroethane	6.565	63	82821	20.561	ug/l	98
25) 2-Butanone	7.482	43	165762	107.295	ug/l	100
26) 2,2-Dichloropropane	7.494	77	76824	21.152	ug/l	100
27) cis-1,2-Dichloroethene	7.488	96	52758	20.783	ug/l	99
28) Bromochloromethane	7.812	49	34591	19.246	ug/l	99
29) Tetrahydrofuran	7.835	42	100146	105.041	ug/l	99
30) Chloroform	7.965	83	85848	20.517	ug/l	99
31) Cyclohexane	8.253	56	77216	19.743	ug/l	98
32) 1,1,1-Trichloroethane	8.171	97	77552	20.637	ug/l	97
36) 1,1-Dichloropropene	8.371	75	61556	20.918	ug/l	99
37) Ethyl Acetate	7.559	43	61023	20.213	ug/l	99
38) Carbon Tetrachloride	8.359	117	67913	21.039	ug/l	95
39) Methylcyclohexane	9.600	83	70956	21.665	ug/l	99
40) Benzene	8.606	78	191581	20.896	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084215.D
 Acq On : 30 Sep 2024 13:13
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024

Quant Time: Oct 01 06:48:50 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M

Quant Title : SW846 8260

QLast Update : Tue Oct 01 06:46:32 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.782	41	34943	21.320	ug/l	99
42) 1,2-Dichloroethane	8.671	62	64453	20.736	ug/l	99
43) Isopropyl Acetate	8.688	43	114278	19.334	ug/l	99
44) Trichloroethene	9.347	130	44412	20.740	ug/l	98
45) 1,2-Dichloropropane	9.618	63	46651	21.509	ug/l	100
46) Dibromomethane	9.706	93	31672	21.639	ug/l	97
47) Bromodichloromethane	9.888	83	68343	21.017	ug/l	96
48) Methyl methacrylate	9.682	41	50498	21.453	ug/l	98
49) 1,4-Dioxane	9.700	88	18411	424.848	ug/l #	93
51) 4-Methyl-2-Pentanone	10.441	43	313636	109.132	ug/l	99
52) Toluene	10.629	92	118412	21.179	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	70403	21.240	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	74778	20.973	ug/l	96
55) 1,1,2-Trichloroethane	11.018	97	42631	21.270	ug/l	98
56) Ethyl methacrylate	10.871	69	72159	21.880	ug/l	99
57) 1,3-Dichloropropane	11.159	76	75915	21.197	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.159	63	142843	93.367	ug/l	99
59) 2-Hexanone	11.194	43	237694	110.658	ug/l	100
60) Dibromochloromethane	11.359	129	50326	21.251	ug/l	97
61) 1,2-Dibromoethane	11.470	107	44554	21.393	ug/l	98
64) Tetrachloroethene	11.100	164	38055	20.179	ug/l	97
65) Chlorobenzene	11.888	112	123787	20.620	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.959	131	44285	21.417	ug/l	96
67) Ethyl Benzene	11.959	91	219635	20.688	ug/l	98
68) m/p-Xylenes	12.070	106	168826	42.989	ug/l	98
69) o-Xylene	12.394	106	79896	21.503	ug/l	98
70) Styrene	12.412	104	134066	21.302	ug/l	98
71) Bromoform	12.576	173	32824	21.561	ug/l #	99
73) Isopropylbenzene	12.694	105	207217	21.254	ug/l	99
74) N-amyl acetate	12.494	43	93392	21.154	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.935	83	60656	20.690	ug/l	98
76) 1,2,3-Trichloropropane	12.994	75	51246m	20.124	ug/l	
77) Bromobenzene	12.976	156	48803	20.752	ug/l	98
78) n-propylbenzene	13.035	91	241164	21.350	ug/l	98
79) 2-Chlorotoluene	13.123	91	147738	20.497	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	171600	21.577	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	22222	20.426	ug/l	92
82) 4-Chlorotoluene	13.217	91	151387	20.864	ug/l	100
83) tert-Butylbenzene	13.435	119	147727	20.899	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	176283	22.004	ug/l	99
85) sec-Butylbenzene	13.611	105	205477	21.784	ug/l	99
86) p-Isopropyltoluene	13.729	119	171766	22.022	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	90970	20.620	ug/l	99
88) 1,4-Dichlorobenzene	13.811	146	88628	19.893	ug/l	96
89) n-Butylbenzene	14.053	91	151334	21.398	ug/l	100
90) Hexachloroethane	14.335	117	32554	20.562	ug/l	95
91) 1,2-Dichlorobenzene	14.106	146	88942	20.560	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	12914	21.347	ug/l	96
93) 1,2,4-Trichlorobenzene	15.388	180	47149	22.128	ug/l	99
94) Hexachlorobutadiene	15.500	225	22742	21.402	ug/l	97
95) Naphthalene	15.641	128	149988	21.233	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	46153	21.639	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084215.D
Acq On : 30 Sep 2024 13:13
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh
Dadoda 10/01/2024

Quant Time: Oct 01 06:48:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084215.D
Acq On : 30 Sep 2024 13:13
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh
Dadoda 10/01/2024


Quant Time: Oct 01 06:48:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
Client Sample Id :
VSTDIC020



Tgt Ion: 168 Resp: 178087
Ion Ratio Lower Upper
168 100
99 65.9 54.2 81.2




Manual Integrations
APPROVED

Reviewed By : John Carlone 10/01/2024
Supervised By : Mahesh
Dadoda 10/01/2024




#2
Dichlorodifluoromethane
Concen: 20.275 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 85 Resp: 42709
Ion Ratio Lower Upper
85 100
87 31.8 17.6 52.9



Abundance Scan 69 (2.359 min): VN084214.D\data.ms (-63)



#3

Chloromethane

Concen: 20.047 ug/l

RT: 2.359 min Scan# 69

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 50 Resp: 48218

Ion Ratio Lower Upper

50 100

52 33.0 26.8 40.2

Manual Integrations


APPROVED

Reviewed By :John Carlone 10/01/2024


Supervised By :Mahesh

Dadoda 10/01/2024


Abundance Scan 69 (2.359 min): VN084215.D\data.ms




Abundance Scan 69 (2.359 min): VN084215.D\data.ms (-18)



Abundance



Abundance Scan 96 (2.518 min): VN084214.D\data.ms (-88)



#4

Vinyl Chloride

Concen: 20.334 ug/l

RT: 2.518 min Scan# 96

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13


Tgt Ion: 62 Resp: 47375

Ion Ratio Lower Upper


62 100

64 30.6 26.2 39.4


Abundance Scan 96 (2.518 min): VN084215.D\data.ms




Abundance Scan 96 (2.518 min): VN084215.D\data.ms (-45)



Abundance





#5

Bromomethane

Concen: 20.010 ug/l

RT: 2.942 min Scan# 16

Delta R.T. -0.012 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 94 Resp: 30751

Ion Ratio Lower Upper

94 100

96 94.3 72.9 109.3





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#6

Chloroethane

Concen: 18.369 ug/l

RT: 3.118 min Scan# 198

Delta R.T. -0.000 min

Lab File: VN084215.D




Acq: 30 Sep 2024 13:13


Tgt Ion: 64 Resp: 30650

Ion Ratio Lower Upper

64 100

66 31.8 23.5 35.3





#7

Trichlorofluoromethane

Concen: 20.557 ug/l

RT: 3.495 min Scan# 262

Delta R.T. -0.000 min

Lab File: VN084215.D


Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020






Tgt Ion: 101 Resp: 74579
Ion Ratio Lower Upper
101 100
103 61.2 51.8 77.6

Manual Integrations
APPROVED

Reviewed By : John Carlone 10/01/2024

Supervised By : Mahesh

Dadoda 10/01/2024



#8

Diethyl Ether


Concen: 21.619 ug/l

RT: 3.959 min Scan# 341



Delta R.T. -0.000 min


Lab File: VN084215.D

Acq: 30 Sep 2024 13:13



Tgt Ion: 74 Resp: 29090
Ion Ratio Lower Upper
74 100
45 92.8 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.665 ug/l
RT: 4.377 min Scan# 411
Delta R.T. 0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:101 Resp: 42958
Ion Ratio Lower Upper
101 100
85 45.9 36.2 54.4
151 74.9 59.3 88.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh
Dadoda 10/01/2024



#10
Methyl Iodide
Concen: 20.458 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion:142 Resp: 54769
Ion Ratio Lower Upper
142 100
127 46.7 35.3 52.9
141 13.7 11.0 16.6





#11

Tert butyl alcohol

Concen: 104.506 ug/l

RT: 5.518 min Scan# 607

Delta R.T. -0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 59 Resp: 45501

Ion Ratio Lower Upper

59 100

57 9.4 8.2 12.2





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#12

1,1-Dichloroethene

Concen: 21.171 ug/l

RT: 4.336 min Scan# 405

Delta R.T. -0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion: 96 Resp: 42450


Ion Ratio Lower Upper

96 100

61 166.5 128.6 192.8

98 60.0 51.7 77.5





#13

Acrolein

Concen: 96.553 ug/l

RT: 4.183 min Scan# 37

Delta R.T. 0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 56 Resp: 48261

Ion Ratio Lower Upper

56 100

55 69.7 57.2 85.8





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#14

Allyl chloride

Concen: 22.208 ug/l

RT: 5.024 min Scan# 522

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion: 41 Resp: 77338


Ion Ratio Lower Upper

41 100

39 66.9 59.0 88.6

76 33.5 29.8 44.6





#15

Acrylonitrile

Concen: 105.004 ug/l

RT: 5.718 min Scan# 640

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:	53	Resp:	115451
Ion	Ratio	Lower	Upper
53	100		
52	81.7	66.2	99.4
51	35.2	28.8	43.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#16

Acetone

Concen: 105.747 ug/l




RT: 4.430 min Scan# 421


Delta R.T. 0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Tgt Ion:	43	Resp:	119948
Ion	Ratio	Lower	Upper
43	100		
58	29.5	24.4	36.6





#17

Carbon Disulfide

Concen: 19.590 ug/l

RT: 4.706 min Scan# 46

Delta R.T. -0.006 min

Lab File: VN084215.D


Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020



Tgt Ion: 76 Resp: 124366

Ion Ratio Lower Upper

76 100

78 8.6 7.3 10.9




Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#18

Methyl Acetate

Concen: 19.996 ug/l

RT: 5.030 min Scan# 523

Delta R.T. 0.012 min

Lab File: VN084215.D




Acq: 30 Sep 2024 13:13


Tgt Ion: 43 Resp: 49436

Ion Ratio Lower Upper

43 100

74 24.3 19.2 28.8





#19

Methyl tert-butyl Ether

Concen: 21.421 ug/l

RT: 5.794 min Scan# 65

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampled :

VSTDIC020

Tgt Ion: 73 Resp: 144972

Ion Ratio Lower Upper

73 100

57 22.1 17.7 26.5





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#20

Methylene Chloride

Concen: 20.423 ug/l

RT: 5.271 min Scan# 564

Delta R.T. -0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Tgt Ion: 84 Resp: 47091




Ion Ratio Lower Upper


84 100

49 122.3 107.2 160.8

51 38.6 31.8 47.8

86 65.8 52.9 79.3





#21

trans-1,2-Dichloroethene

Concen: 20.976 ug/l

RT: 5.783 min Scan# 65

Delta R.T. -0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 96 Resp: 44064

Ion Ratio Lower Upper

96 100

61 149.0 108.4 162.6

98 63.9 49.8 74.8





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#22

Diisopropyl ether

Concen: 21.460 ug/l

RT: 6.671 min Scan# 802

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Tgt Ion: 45 Resp: 154575




Ion Ratio Lower Upper


45 100

43 51.7 42.3 63.5

87 28.8 21.8 32.6

59 10.5 9.6 14.4





#23

Vinyl Acetate

Concen: 109.739 ug/l

RT: 6.600 min Scan# 790

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 43 Resp: 589404

Ion Ratio Lower Upper

43 100

86 10.2 8.6 12.8





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#24

1,1-Dichloroethane

Concen: 20.561 ug/l

RT: 6.565 min Scan# 784

Delta R.T. -0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion: 63 Resp: 82821


Ion Ratio Lower Upper

63 100

98 6.8 3.4 10.2

100 5.3 2.0 5.8





#25

2-Butanone

Concen: 107.295 ug/l

RT: 7.482 min Scan# 940

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 43 Resp: 165762

Ion Ratio Lower Upper

43 100

72 26.7 21.4 32.2





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#26

2,2-Dichloropropane

Concen: 21.152 ug/l

RT: 7.494 min Scan# 942

Delta R.T. 0.006 min

Lab File: VN084215.D




Acq: 30 Sep 2024 13:13


Tgt Ion: 77 Resp: 76824

Ion Ratio Lower Upper

77 100


97 21.3 10.8 32.3





#27
cis-1,2-Dichloroethene
Concen: 20.783 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 96 Resp: 52758
Ion Ratio Lower Upper
96 100
61 149.6 0.0 301.8
98 63.6 0.0 128.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh
Dadoda 10/01/2024



#28
Bromochloromethane
Concen: 19.246 ug/l
RT: 7.812 min Scan# 996
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 49 Resp: 34591
Ion Ratio Lower Upper
49 100
129 1.6 0.0 3.6
130 71.9 58.0 87.0





#29

Tetrahydrofuran

Concen: 105.041 ug/l

RT: 7.835 min Scan# 1000

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 42 Resp: 100146

Ion Ratio Lower Upper

42 100

72 44.7 34.6 51.8

71 40.3 32.2 48.4





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#30

Chloroform

Concen: 20.517 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. -0.000 min

Lab File: VN084215.D




Acq: 30 Sep 2024 13:13


Tgt Ion: 83 Resp: 85848

Ion Ratio Lower Upper

83 100

85 62.7 51.0 76.4





#31

Cyclohexane

Concen: 19.743 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 56 Resp: 77216

Ion Ratio Lower Upper

56 100

69 31.9 23.8 35.6

84 87.6 68.8 103.2





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#32

1,1,1-Trichloroethane

Concen: 20.637 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion: 97 Resp: 77552


Ion Ratio Lower Upper

97 100

99 68.3 51.2 76.8

61 46.9 37.5 56.3









#33
1,2-Dichloroethane-d4
Concen: 20.606 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 65 Resp: 54409
Ion Ratio Lower Upper
65 100
67 49.7 0.0 102.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh
Dadoda 10/01/2024



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion:114 Resp: 307298
Ion Ratio Lower Upper
114 100
63 21.8 0.0 43.8
88 16.6 0.0 31.6





#35

Dibromofluoromethane

Concen: 20.684 ug/l

RT: 8.171 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 113 Resp: 41903

Ion Ratio Lower Upper

113 100

111 100.2 83.3 124.9

192 16.8 13.5 20.3





Manual Integrations

APPROVED

Reviewed By : John Carlone 10/01/2024

Supervised By : Mahesh

Dadoda 10/01/2024



#36

1,1-Dichloropropene

Concen: 20.918 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion: 75 Resp: 61556


Ion Ratio Lower Upper

75 100

110 33.1 16.9 50.7

77 31.1 24.7 37.1





#37

Ethyl Acetate

Concen: 20.213 ug/l

RT: 7.559 min Scan# 953

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion:	43	Resp:	61023
Ion	Ratio	Lower	Upper
43	100		
61	14.1	10.7	16.1
70	11.2	8.8	13.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#38

Carbon Tetrachloride

Concen: 21.039 ug/l




RT: 8.359 min Scan# 1089


Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Tgt Ion:	117	Resp:	67913
Ion	Ratio	Lower	Upper
117	100		
119	93.0	78.7	118.1
121	29.2	24.4	36.6





#39

Methylcyclohexane

Concen: 21.665 ug/l

RT: 9.600 min Scan# 1300

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 83 Resp: 70956

Ion Ratio Lower Upper

83 100

55 77.8 62.8 94.2

98 47.9 38.8 58.2





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#40

Benzene

Concen: 20.896 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN084215.D




Acq: 30 Sep 2024 13:13


Tgt Ion: 78 Resp: 191581

Ion Ratio Lower Upper

78 100

77 22.5 19.1 28.7





#41

Methacrylonitrile

Concen: 21.320 ug/l

RT: 7.782 min Scan# 991

Delta R.T. 0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:	41	Resp:	34943
Ion	Ratio	Lower	Upper
41	100		
39	53.8	43.3	64.9
67	71.5	56.1	84.1
52	29.7	24.5	36.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#42

1,2-Dichloroethane

Concen: 20.736 ug/l




RT: 8.671 min Scan# 1142


Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Tgt Ion:	62	Resp:	64453
Ion	Ratio	Lower	Upper
62	100		
98	9.5	0.0	18.2





#43

Isopropyl Acetate

Concen: 19.334 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 43 Resp: 114278

Ion Ratio Lower Upper

43 100

61 24.7 20.4 30.6

87 12.7 10.3 15.5





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#44

Trichloroethene

Concen: 20.740 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. -0.000 min

Lab File: VN084215.D




Acq: 30 Sep 2024 13:13


Tgt Ion:130 Resp: 44412

Ion Ratio Lower Upper

130 100

95 103.2 0.0 203.0





#45

1,2-Dichloropropane

Concen: 21.509 ug/l

RT: 9.618 min Scan# 13

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 63 Resp: 46651

Ion Ratio Lower Upper

63 100

65 30.9 24.6 37.0





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#46

Dibromomethane

Concen: 21.639 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion: 93 Resp: 31672


Ion Ratio Lower Upper

93 100

95 83.8 65.0 97.4

174 92.0 71.8 107.8





#47

Bromodichloromethane

Concen: 21.017 ug/l

RT: 9.888 min Scan# 13

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 83 Resp: 68343

Ion Ratio Lower Upper

83 100

85 61.4 51.5 77.3

127 8.0 5.7 8.5





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#48

Methyl methacrylate

Concen: 21.453 ug/l

RT: 9.682 min Scan# 1314

Delta R.T. 0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion: 41 Resp: 50498


Ion Ratio Lower Upper

41 100

69 82.3 67.1 100.7

39 59.0 45.8 68.6





#49

1,4-Dioxane

Concen: 424.848 ug/l

RT: 9.700 min Scan# 1316

Delta R.T. 0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion: 88 Resp: 18411

Ion Ratio Lower Upper

88 100

43 22.7 25.3 37.9

58 72.9 56.4 84.6





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#50

Toluene-d8

Concen: 20.483 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN084215.D




Acq: 30 Sep 2024 13:13


Tgt Ion: 98 Resp: 152668

Ion Ratio Lower Upper

98 100

100 67.0 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 109.132 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 43 Resp: 313636
Ion Ratio Lower Upper
43 100
58 38.9 31.6 47.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh
Dadoda 10/01/2024



#52
Toluene
Concen: 21.179 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 92 Resp: 118412
Ion Ratio Lower Upper
92 100
91 171.0 137.3 205.9





#53

t-1,3-Dichloropropene

Concen: 21.240 ug/l

RT: 10.835 min Scan# 1510

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 75 Resp: 70403

Ion Ratio Lower Upper

75 100

77 32.2 25.4 38.0





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#54

cis-1,3-Dichloropropene

Concen: 20.973 ug/l

RT: 10.312 min Scan# 1421

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion: 75 Resp: 74778


Ion Ratio Lower Upper

75 100

77 32.6 24.2 36.2

39 51.4 38.7 58.1





#55

1,1,2-Trichloroethane

Concen: 21.270 ug/l

RT: 11.018 min Scan# 1540

Delta R.T. 0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 97 Resp: 42631

Ion Ratio Lower Upper

97 100

83 86.8 71.0 106.4

85 53.6 41.7 62.5

99 63.9 50.1 75.1





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#56

Ethyl methacrylate

Concen: 21.880 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion: 69 Resp: 72159


Ion Ratio Lower Upper

69 100

41 67.2 53.3 79.9

39 36.6 28.8 43.2





#57

1,3-Dichloropropane

Concen: 21.197 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 76 Resp: 75915

Ion Ratio Lower Upper

76 100

78 31.2 25.8 38.6





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#58

2-Chloroethyl Vinyl ether

Concen: 93.367 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN084215.D




Acq: 30 Sep 2024 13:13


Tgt Ion: 63 Resp: 142843

Ion Ratio Lower Upper

63 100

106 27.5 21.8 32.6





#59

2-Hexanone

Concen: 110.658 ug/l

RT: 11.194 min Scan# 1571

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 237694

Ion Ratio Lower Upper

43 100

58 54.4 27.3 81.8





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#60

Dibromochloromethane

Concen: 21.251 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. -0.000 min

Lab File: VN084215.D




Acq: 30 Sep 2024 13:13


Tgt Ion:129 Resp: 50326

Ion Ratio Lower Upper

129 100


127 78.8 38.3 114.9





#61
1,2-Dibromoethane
Concen: 21.393 ug/l
RT: 11.470 min Scan# 1617
Delta R.T. 0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020



Tgt Ion: 107 Resp: 44554
Ion Ratio Lower Upper
107 100
109 92.5 75.4 113.0




Manual Integrations
APPROVED


Reviewed By : John Carlone 10/01/2024
Supervised By : Mahesh
Dadoda 10/01/2024



#62
4-Bromofluorobenzene
Concen: 20.329 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 95 Resp: 55201
Ion Ratio Lower Upper
95 100
174 73.2 0.0 145.2
176 68.9 0.0 140.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:117 Resp: 270754

Ion Ratio Lower Upper

117 100

82 57.7 47.2 70.8

119 31.2 25.4 38.0





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#64

Tetrachloroethene

Concen: 20.179 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Tgt Ion:164 Resp: 38055




Ion Ratio Lower Upper


164 100

166 126.4 101.0 151.4

129 106.4 80.8 121.2

131 94.2 78.1 117.1





#65

Chlorobenzene

Concen: 20.620 ug/l

RT: 11.888 min Scan# 1689

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:112 Resp: 123787

Ion Ratio Lower Upper

112 100

114 33.3 27.0 40.6





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#66

1,1,1,2-Tetrachloroethane

Concen: 21.417 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion:131 Resp: 44285


Ion Ratio Lower Upper

131 100

133 92.4 47.9 143.7

119 62.0 32.8 98.3





#67

Ethyl Benzene

Concen: 20.688 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 219635

Ion Ratio Lower Upper

91 100

106 29.7 24.5 36.7





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#68

m/p-Xylenes

Concen: 42.989 ug/l

RT: 12.070 min Scan# 1720

Delta R.T. -0.000 min

Lab File: VN084215.D




Acq: 30 Sep 2024 13:13


Tgt Ion:106 Resp: 168826

Ion Ratio Lower Upper

106 100

91 206.4 167.1 250.7





#69

o-Xylene

Concen: 21.503 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion:106 Resp: 79896

Ion Ratio Lower Upper

106 100

91 223.3 110.2 330.6





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#70

Styrene

Concen: 21.302 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion:104 Resp: 134066


Ion Ratio Lower Upper

104 100

78 51.8 42.7 64.1

103 55.2 43.6 65.4





#71

Bromoform

Concen: 21.561 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion:	173	Resp:	32824
Ion	Ratio	Lower	Upper
173	100		
175	47.8	23.5	70.6
254	0.0	0.0	0.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.788 min Scan# 2012


Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Tgt Ion:	152	Resp:	127763
Ion	Ratio	Lower	Upper
152	100		
115	62.3	31.3	93.9
150	164.5	0.0	349.8





#73

Isopropylbenzene

Concen: 21.254 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020

Tgt Ion:105 Resp: 207217

Ion Ratio Lower Upper

105 100

120 25.0 12.9 38.7





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#74

N-ethyl acetate

Concen: 21.154 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Tgt Ion: 43 Resp: 93392




Ion Ratio Lower Upper


43 100

70 42.8 34.4 51.6

55 25.0 19.6 29.4

61 24.6 19.6 29.4





#75

1,1,2,2-Tetrachloroethane

Concen: 20.690 ug/l

RT: 12.935 min Scan# 18

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:	83	Resp:	60656
Ion	Ratio	Lower	Upper
83	100		
131	10.6	4.9	14.7
85	65.1	31.9	95.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#76

1,2,3-Trichloropropane

Concen: 20.124 ug/l m




RT: 12.994 min Scan# 1877


Delta R.T. 0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Tgt Ion:	75	Resp:	51246
Ion	Ratio	Lower	Upper
75	100		
77	216.2	96.9	290.7





#77

Bromobenzene

Concen: 20.752 ug/l

RT: 12.976 min Scan# 1875

Delta R.T. -0.006 min

Lab File: VN084215.D


Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020



Tgt Ion:156 Resp: 48803

Ion Ratio Lower Upper

156 100

77 227.1 115.1 345.3

158 96.5 48.9 146.8




Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#78

n-propylbenzene

Concen: 21.350 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. -0.000 min

Lab File: VN084215.D




Acq: 30 Sep 2024 13:13


Tgt Ion: 91 Resp: 241164

Ion Ratio Lower Upper

91 100

120 22.4 10.8 32.4





#79

2-Chlorotoluene

Concen: 20.497 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 147738

Ion Ratio Lower Upper

91 100

126 32.0 15.7 47.1





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#80

1,3,5-Trimethylbenzene

Concen: 21.577 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN084215.D




Acq: 30 Sep 2024 13:13


Tgt Ion:105 Resp: 171600

Ion Ratio Lower Upper

105 100

120 47.5 23.7 71.1





#81

trans-1,4-Dichloro-2-butene

Concen: 20.426 ug/l

RT: 12.735 min Scan# 1833

Delta R.T. -0.000 min

Lab File: VN084215.D


Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020






Tgt Ion:	75	Resp:	22222
Ion	Ratio	Lower	Upper
75	100		
53	105.8	78.0	117.0
89	48.1	35.5	53.3

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#82

4-Chlorotoluene

Concen: 20.864 ug/l




RT: 13.217 min Scan# 1915


Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Tgt Ion:	91	Resp:	151387
Ion	Ratio	Lower	Upper
91	100		
126	31.6	15.7	47.1





#83

tert-Butylbenzene

Concen: 20.899 ug/l

RT: 13.435 min Scan# 19

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:119 Resp: 147727

Ion Ratio Lower Upper

119 100

91 68.9 34.1 102.3

134 25.9 12.8 38.4





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#84

1,2,4-Trimethylbenzene

Concen: 22.004 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. -0.000 min

Lab File: VN084215.D




Acq: 30 Sep 2024 13:13


Tgt Ion:105 Resp: 176283

Ion Ratio Lower Upper

105 100

120 43.6 22.3 66.8





#85

sec-Butylbenzene

Concen: 21.784 ug/l

RT: 13.611 min Scan# 19

Delta R.T. -0.006 min

Lab File: VN084215.D


Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020



Tgt Ion:105 Resp: 205477

Ion Ratio Lower Upper

105 100

134 19.4 9.8 29.4




Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#86

p-Isopropyltoluene

Concen: 22.022 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion:119 Resp: 171766


Ion Ratio Lower Upper

119 100

134 25.9 12.6 37.8

91 24.7 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 20.620 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion:	146	Resp:	90970
Ion	Ratio	Lower	Upper
146	100		
111	42.4	21.8	65.3
148	63.4	31.9	95.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#88

1,4-Dichlorobenzene

Concen: 19.893 ug/l




RT: 13.811 min Scan# 2016


Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Tgt Ion:	146	Resp:	88628
Ion	Ratio	Lower	Upper
146	100		
111	45.2	21.2	63.6
148	66.2	31.9	95.9





#89

n-Butylbenzene

Concen: 21.398 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion:	91	Resp:	151334
Ion	Ratio	Lower	Upper
91	100		
92	53.1	26.6	79.7
134	23.9	12.3	36.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#90

Hexachloroethane

Concen: 20.562 ug/l




RT: 14.335 min Scan# 2105


Delta R.T. 0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Tgt Ion:	117	Resp:	32554
Ion	Ratio	Lower	Upper
117	100		
201	65.1	34.4	103.2





#91

1,2-Dichlorobenzene

Concen: 20.560 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:	146	Resp:	88942
Ion Ratio	Lower	Upper	
146	100		
111	42.8	22.0	66.0
148	63.6	32.1	96.3

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#92

1,2-Dibromo-3-Chloropropane

Concen: 21.347 ug/l




RT: 14.717 min Scan# 2170


Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Tgt Ion:	75	Resp:	12914
Ion Ratio	Lower	Upper	
75	100		
155	76.9	38.1	114.5
157	93.4	49.6	149.0





#93

1,2,4-Trichlorobenzene

Concen: 22.128 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. -0.000 min

Lab File: VN084215.D


Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020






Tgt Ion:	180	Resp:	47149
Ion Ratio	Lower	Upper	
180	100		
182	94.5	47.0	141.0
145	32.3	16.4	49.4

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#94

Hexachlorobutadiene


Concen: 21.402 ug/l

RT: 15.500 min Scan# 2303



Delta R.T. -0.000 min


Lab File: VN084215.D

Acq: 30 Sep 2024 13:13



Tgt Ion:	225	Resp:	22742
Ion Ratio	Lower	Upper	
225	100		
223	62.8	32.4	97.2
227	62.7	32.4	97.0





#95

Naphthalene

Concen: 21.233 ug/l

RT: 15.641 min Scan# 2326

Delta R.T. 0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020

Tgt Ion:128 Resp: 149988

Ion Ratio Lower Upper

128 100

127 12.8 10.5 15.7

129 10.7 8.7 13.1





Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh

Dadoda 10/01/2024



#96

1,2,3-Trichlorobenzene

Concen: 21.639 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion:180 Resp: 46153

Ion Ratio Lower Upper

180 100

182 90.7 48.2 144.6

145 35.1 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084216.D
 Acq On : 30 Sep 2024 13:37
 Operator : JC\MD
 Sample : VSTDICC010
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC010

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/01/2024
 Supervised By : Mahesh Dadoda 10/01/2024

Quant Time: Oct 01 06:49:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	176243	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	304981	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	262325	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	119657	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.582	65	25083	9.599	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	19.200%#
35) Dibromofluoromethane	8.165	113	19289	9.594	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	19.180%#
50) Toluene-d8	10.565	98	70000	9.463	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	18.920%#
62) 4-Bromofluorobenzene	12.847	95	24756	9.186	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	18.380%#

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.130	85	19806	9.501	ug/l	96
3) Chloromethane	2.359	50	22840	9.595	ug/l	97
4) Vinyl Chloride	2.518	62	22584	9.795	ug/l	97
5) Bromomethane	2.965	94	14957	9.835	ug/l	96
6) Chloroethane	3.124	64	15254	9.238	ug/l	89
7) Trichlorofluoromethane	3.494	101	33820	9.420	ug/l	99
8) Diethyl Ether	3.965	74	12422	9.328	ug/l	96
9) 1,1,2-Trichlorotrifluo...	4.377	101	18747	9.113	ug/l	98
10) Methyl Iodide	4.594	142	24268	9.160	ug/l	96
11) Tert butyl alcohol	5.512	59	20490	47.553	ug/l #	86
12) 1,1-Dichloroethene	4.341	96	18798	9.473	ug/l	86
13) Acrolein	4.183	56	22976	46.448	ug/l	98
14) Allyl chloride	5.024	41	31571	9.161	ug/l	98
15) Acrylonitrile	5.718	53	52809	48.533	ug/l	99
16) Acetone	4.424	43	52518	46.785	ug/l	96
17) Carbon Disulfide	4.718	76	58158	9.257	ug/l	99
18) Methyl Acetate	5.024	43	24366	9.959	ug/l	95
19) Methyl tert-butyl Ether	5.794	73	63531	9.486	ug/l	94
20) Methylene Chloride	5.277	84	21801	9.554	ug/l	87
21) trans-1,2-Dichloroethene	5.788	96	20107	9.672	ug/l	85
22) Diisopropyl ether	6.671	45	66959	9.393	ug/l	97
23) Vinyl Acetate	6.606	43	248594	46.769	ug/l	98
24) 1,1-Dichloroethane	6.571	63	38215	9.586	ug/l	99
25) 2-Butanone	7.482	43	73877	48.320	ug/l	99
26) 2,2-Dichloropropane	7.488	77	34655	9.641	ug/l	98
27) cis-1,2-Dichloroethene	7.488	96	23095	9.193	ug/l	98
28) Bromochloromethane	7.812	49	14904	8.379	ug/l	100
29) Tetrahydrofuran	7.835	42	46917	49.725	ug/l	99
30) Chloroform	7.965	83	39355	9.504	ug/l	90
31) Cyclohexane	8.253	56	36916	9.538	ug/l	92
32) 1,1,1-Trichloroethane	8.171	97	35869	9.645	ug/l #	87
36) 1,1-Dichloropropene	8.377	75	28525	9.767	ug/l	99
37) Ethyl Acetate	7.565	43	29754	9.931	ug/l #	81
38) Carbon Tetrachloride	8.365	117	30964	9.665	ug/l	92
39) Methylcyclohexane	9.600	83	30932	9.516	ug/l	95
40) Benzene	8.606	78	86029	9.454	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084216.D
 Acq On : 30 Sep 2024 13:37
 Operator : JC\MD
 Sample : VSTDICC010
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC010

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/01/2024
 Supervised By : Mahesh Dadoda 10/01/2024

Quant Time: Oct 01 06:49:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	15332	9.426	ug/l	97
42) 1,2-Dichloroethane	8.671	62	30402	9.855	ug/l	97
43) Isopropyl Acetate	8.688	43	52246	8.906	ug/l	99
44) Trichloroethene	9.347	130	20092	9.454	ug/l	99
45) 1,2-Dichloropropane	9.618	63	20703	9.618	ug/l	96
46) Dibromomethane	9.706	93	13770	9.479	ug/l	95
47) Bromodichloromethane	9.888	83	30289	9.385	ug/l #	98
48) Methyl methacrylate	9.682	41	22612	9.679	ug/l	98
49) 1,4-Dioxane	9.694	88	8627	200.587	ug/l #	78
51) 4-Methyl-2-Pentanone	10.441	43	142632	50.007	ug/l	99
52) Toluene	10.629	92	52537	9.468	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	31555	9.592	ug/l	93
54) cis-1,3-Dichloropropene	10.312	75	34684	9.802	ug/l	96
55) 1,1,2-Trichloroethane	11.018	97	19245	9.675	ug/l	97
56) Ethyl methacrylate	10.870	69	31262	9.551	ug/l	97
57) 1,3-Dichloropropane	11.165	76	33613	9.457	ug/l	97
58) 2-Chloroethyl Vinyl ether	10.159	63	71035	46.784	ug/l	99
59) 2-Hexanone	11.194	43	103615	48.604	ug/l	100
60) Dibromochloromethane	11.359	129	22833	9.715	ug/l	99
61) 1,2-Dibromoethane	11.470	107	19220	9.299	ug/l	95
64) Tetrachloroethene	11.100	164	17743	9.711	ug/l	97
65) Chlorobenzene	11.894	112	56099	9.645	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	19227	9.597	ug/l	97
67) Ethyl Benzene	11.964	91	96533	9.385	ug/l	97
68) m/p-Xylenes	12.070	106	72910	19.162	ug/l	100
69) o-Xylene	12.394	106	34965	9.713	ug/l	100
70) Styrene	12.412	104	58367	9.572	ug/l	99
71) Bromoform	12.576	173	13618	9.233	ug/l #	96
73) Isopropylbenzene	12.694	105	88000	9.638	ug/l	99
74) N-ethyl acetate	12.494	43	40801	9.868	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	26966	9.822	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	23907m	10.024	ug/l	
77) Bromobenzene	12.982	156	21439	9.734	ug/l	99
78) n-propylbenzene	13.035	91	105334	9.957	ug/l	99
79) 2-Chlorotoluene	13.123	91	65511	9.704	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	73637	9.887	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.729	75	10237	10.047	ug/l	94
82) 4-Chlorotoluene	13.223	91	66449	9.778	ug/l	99
83) tert-Butylbenzene	13.435	119	64152	9.691	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	74245	9.895	ug/l	99
85) sec-Butylbenzene	13.611	105	88913	10.065	ug/l	100
86) p-Isopropyltoluene	13.729	119	71675	9.812	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	39402	9.536	ug/l	98
88) 1,4-Dichlorobenzene	13.811	146	39191	9.392	ug/l	97
89) n-Butylbenzene	14.053	91	61422	9.273	ug/l	99
90) Hexachloroethane	14.335	117	14560	9.820	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	38594	9.526	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	5799	10.235	ug/l	97
93) 1,2,4-Trichlorobenzene	15.388	180	18960	9.501	ug/l	99
94) Hexachlorobutadiene	15.500	225	9592	9.638	ug/l	95
95) Naphthalene	15.635	128	63596	9.613	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	19629	9.826	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084216.D
Acq On : 30 Sep 2024 13:37
Operator : JC\MD
Sample : VSTDICC010
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024

Quant Time: Oct 01 06:49:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084216.D
Acq On : 30 Sep 2024 13:37
Operator : JC\MD
Sample : VSTDIC010
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010

Manual Integrations
APPROVED

Reviewed By : John Carlone 10/01/2024
Supervised By : Mahesh Dadoda 10/01/2024


Quant Time: Oct 01 06:49:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 10
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010



Tgt Ion: 168 Resp: 176243
Ion Ratio Lower Upper
168 100
99 66.2 54.2 81.2




Manual Integrations
APPROVED

Reviewed By : John Carlone 10/01/2024
Supervised By : Mahesh Dadoda 10/01/2024




#2
Dichlorodifluoromethane
Concen: 9.501 ug/l
RT: 2.130 min Scan# 30
Delta R.T. 0.006 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 85 Resp: 19806
Ion Ratio Lower Upper
85 100
87 32.8 17.6 52.9



Abundance Scan 69 (2.359 min): VN084214.D\data.ms (-63)



#3

Chloromethane

Concen: 9.595 ug/l

RT: 2.359 min Scan# 69

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 50 Resp: 22840

Ion Ratio Lower Upper

50 100


52 31.7 26.8 40.2

Manual Integrations



APPROVED

Reviewed By :John Carlone 10/01/2024


Supervised By :Mahesh Dadoda 10/01/2024



Abundance Scan 69 (2.359 min): VN084216.D\data.ms (-18)



Abundance Scan 96 (2.518 min): VN084214.D\data.ms (-88)



#4

Vinyl Chloride

Concen: 9.795 ug/l

RT: 2.518 min Scan# 96

Delta R.T. -0.000 min

Lab File: VN084216.D


Acq: 30 Sep 2024 13:37

Tgt Ion: 62 Resp: 22584



Ion Ratio Lower Upper


62 100

64 30.9 26.2 39.4



Abundance Scan 96 (2.518 min): VN084216.D\data.ms (-45)





#5

Bromomethane

Concen: 9.835 ug/l

RT: 2.965 min Scan# 171

Delta R.T. 0.012 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 94 Resp: 14957

Ion Ratio Lower Upper

94 100





96 86.8 72.9 109.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#6

Chloroethane

Concen: 9.238 ug/l

RT: 3.124 min Scan# 199

Delta R.T. 0.006 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37


Tgt Ion: 64 Resp: 15254

Ion Ratio Lower Upper

64 100

66 35.0 23.5 35.3





#7

Trichlorofluoromethane

Concen: 9.420 ug/l

RT: 3.494 min Scan# 262

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion:101 Resp: 33820

Ion Ratio Lower Upper

101 100





103 64.2 51.8 77.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#8

Diethyl Ether

Concen: 9.328 ug/l

RT: 3.965 min Scan# 342

Delta R.T. 0.006 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37


Tgt Ion: 74 Resp: 12422

Ion Ratio Lower Upper

74 100

45 99.0 47.7 143.1





#9

1,1,2-Trichlorotrifluoroethane

Concen: 9.113 ug/l

RT: 4.377 min Scan# 411

Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC010

Tgt Ion:	101	Resp:	18747
Ion Ratio	Lower	Upper	
101	100		
85	47.3	36.2	54.4
151	75.1	59.3	88.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#10

Methyl Iodide

Concen: 9.160 ug/l




RT: 4.594 min Scan# 449


Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion:	142	Resp:	24268
Ion Ratio	Lower	Upper	
142	100		
127	46.4	35.3	52.9
141	15.9	11.0	16.6





#11

Tert butyl alcohol

Concen: 47.553 ug/l

RT: 5.512 min Scan# 607

Delta R.T. -0.012 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37

Tgt Ion: 59 Resp: 20496

Ion Ratio Lower Upper

59 100

57 5.1 8.2 12.2



Instrument :

MSVOA_N

ClientSampleId :


VSTDICC010

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#12

1,1-Dichloroethene

Concen: 9.473 ug/l

RT: 4.341 min Scan# 406

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion: 96 Resp: 18798


Ion Ratio Lower Upper

96 100

61 186.2 128.6 192.8

98 66.5 51.7 77.5





#13

Acrolein

Concen: 46.448 ug/l

RT: 4.183 min Scan# 37

Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 56 Resp: 22976

Ion Ratio Lower Upper

56 100





55 72.8 57.2 85.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#14

Allyl chloride

Concen: 9.161 ug/l

RT: 5.024 min Scan# 522

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion: 41 Resp: 31571


Ion Ratio Lower Upper

41 100

39 71.4 59.0 88.6

76 36.1 29.8 44.6





#15

Acrylonitrile

Concen: 48.533 ug/l

RT: 5.718 min Scan# 640

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC010

Tgt Ion:	53	Resp:	52809
Ion	Ratio	Lower	Upper
53	100		
52	82.2	66.2	99.4
51	35.4	28.8	43.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#16

Acetone

Concen: 46.785 ug/l




RT: 4.424 min Scan# 420


Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion:	43	Resp:	52518
Ion	Ratio	Lower	Upper
43	100		
58	32.8	24.4	36.6





#17

Carbon Disulfide

Concen: 9.257 ug/l

RT: 4.718 min Scan# 47

Delta R.T. 0.006 min

Lab File: VN084216.D


Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010



Tgt Ion: 76 Resp: 58158

Ion Ratio Lower Upper

76 100




78 8.6 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#18

Methyl Acetate

Concen: 9.959 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37


Tgt Ion: 43 Resp: 24366

Ion Ratio Lower Upper

43 100

74 21.3 19.2 28.8





#19

Methyl tert-butyl Ether

Concen: 9.486 ug/l

RT: 5.794 min Scan# 65

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 73 Resp: 63531

Ion Ratio Lower Upper

73 100





57 19.3 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#20

Methylene Chloride

Concen: 9.554 ug/l

RT: 5.277 min Scan# 565

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion: 84 Resp: 21801




Ion Ratio Lower Upper


84 100

49 115.7 107.2 160.8

51 36.0 31.8 47.8

86 57.2 52.9 79.3





#21

trans-1,2-Dichloroethene

Concen: 9.672 ug/l

RT: 5.788 min Scan# 65

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 96 Resp: 20107

Ion Ratio Lower Upper

96 100

61 153.1 108.4 162.6





98 73.7 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#22

Diisopropyl ether

Concen: 9.393 ug/l

RT: 6.671 min Scan# 802

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion: 45 Resp: 66959




Ion Ratio Lower Upper


45 100

43 55.7 42.3 63.5

87 25.9 21.8 32.6

59 11.9 9.6 14.4





#23

Vinyl Acetate

Concen: 46.769 ug/l

RT: 6.606 min Scan# 790

Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 43 Resp: 248594

Ion Ratio Lower Upper

43 100





86 9.8 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#24

1,1-Dichloroethane

Concen: 9.586 ug/l

RT: 6.571 min Scan# 785

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion: 63 Resp: 38215


Ion Ratio Lower Upper

63 100

98 7.3 3.4 10.2

100 3.8 2.0 5.8





#25

2-Butanone

Concen: 48.320 ug/l

RT: 7.482 min Scan# 940

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 43 Resp: 73877

Ion Ratio Lower Upper

43 100





72 27.5 21.4 32.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#26

2,2-Dichloropropane

Concen: 9.641 ug/l

RT: 7.488 min Scan# 941

Delta R.T. -0.000 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37


Tgt Ion: 77 Resp: 34655

Ion Ratio Lower Upper

77 100

97 22.7 10.8 32.3









#27
cis-1,2-Dichloroethene
Concen: 9.193 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Tgt Ion:	96	Resp:	23095
Ion	Ratio	Lower	Upper
96	100		
61	152.9	0.0	301.8
98	66.0	0.0	128.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#28
Bromochloromethane
Concen: 8.379 ug/l
RT: 7.812 min Scan# 996
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion:	49	Resp:	14904
Ion	Ratio	Lower	Upper
49	100		
129	0.4	0.0	3.6
130	72.3	58.0	87.0





#29

Tetrahydrofuran

Concen: 49.725 ug/l

RT: 7.835 min Scan# 10

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC010

Tgt Ion:	42	Resp:	46917
Ion	Ratio	Lower	Upper
42	100		
72	42.7	34.6	51.8
71	39.5	32.2	48.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#30

Chloroform

Concen: 9.504 ug/l




RT: 7.965 min Scan# 1022


Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion:	83	Resp:	39355
Ion	Ratio	Lower	Upper
83	100		
85	71.7	51.0	76.4





#31

Cyclohexane

Concen: 9.538 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 56 Resp: 36916

Ion Ratio Lower Upper

56 100

69 33.8 23.8 35.6





84 79.0 68.8 103.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#32

1,1,1-Trichloroethane

Concen: 9.645 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion: 97 Resp: 35869


Ion Ratio Lower Upper

97 100

99 46.3 51.2 76.8#

61 46.7 37.5 56.3









#33
1,2-Dichloroethane-d4
Concen: 9.599 ug/l
RT: 8.582 min Scan# 1126
Delta R.T. 0.006 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Tgt Ion: 65 Resp: 25083
Ion Ratio Lower Upper
65 100
67 52.7 0.0 102.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion:114 Resp: 304981
Ion Ratio Lower Upper
114 100
63 21.5 0.0 43.8
88 16.7 0.0 31.6





#35

Dibromofluoromethane

Concen: 9.594 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion:113 Resp: 19289

Ion Ratio Lower Upper

113 100

111 108.7 83.3 124.9





192 17.7 13.5 20.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#36

1,1-Dichloropropene

Concen: 9.767 ug/l

RT: 8.377 min Scan# 1092

Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion: 75 Resp: 28525


Ion Ratio Lower Upper

75 100

110 32.3 16.9 50.7

77 31.0 24.7 37.1





#37

Ethyl Acetate

Concen: 9.931 ug/l

RT: 7.565 min Scan# 95

Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC010

Tgt Ion:	43	Resp:	29754
Ion	Ratio	Lower	Upper
43	100		
61	0.0	10.7	16.1
70	10.2	8.8	13.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#38

Carbon Tetrachloride

Concen: 9.665 ug/l




RT: 8.365 min Scan# 1090


Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion:	117	Resp:	30964
Ion	Ratio	Lower	Upper
117	100		
119	88.4	78.7	118.1
121	32.4	24.4	36.6





#39

Methylcyclohexane

Concen: 9.516 ug/l

RT: 9.600 min Scan# 13

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 83 Resp: 30932

Ion Ratio Lower Upper

83 100

55 84.9 62.8 94.2





98 48.4 38.8 58.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#40

Benzene

Concen: 9.454 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37


Tgt Ion: 78 Resp: 86029

Ion Ratio Lower Upper

78 100

77 22.8 19.1 28.7





#41

Methacrylonitrile

Concen: 9.426 ug/l

RT: 7.777 min Scan# 990

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC010

Tgt Ion:	41	Resp:	15332
Ion	Ratio	Lower	Upper
41	100		
39	55.8	43.3	64.9
67	73.0	56.1	84.1
52	30.4	24.5	36.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#42

1,2-Dichloroethane

Concen: 9.855 ug/l




RT: 8.671 min Scan# 1142


Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion:	62	Resp:	30402
Ion	Ratio	Lower	Upper
62	100		
98	8.1	0.0	18.2





#43

Isopropyl Acetate

Concen: 8.906 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 43 Resp: 52246

Ion Ratio Lower Upper

43 100

61 25.1 20.4 30.6





87 12.3 10.3 15.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#44

Trichloroethene

Concen: 9.454 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. -0.000 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37


Tgt Ion:130 Resp: 20092

Ion Ratio Lower Upper

130 100

95 102.9 0.0 203.0





#45

1,2-Dichloropropane

Concen: 9.618 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 63 Resp: 20703

Ion Ratio Lower Upper

63 100





65 33.0 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#46

Dibromomethane

Concen: 9.479 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion: 93 Resp: 13770


Ion Ratio Lower Upper

93 100

95 87.2 65.0 97.4

174 92.7 71.8 107.8





#47

Bromodichloromethane

Concen: 9.385 ug/l

RT: 9.888 min Scan# 13

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC010

Tgt Ion:	83	Resp:	30289
Ion	Ratio	Lower	Upper
83	100		
85	65.7	51.5	77.3
127	9.6	5.7	8.5

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#48

Methyl methacrylate

Concen: 9.679 ug/l




RT: 9.682 min Scan# 1314


Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion:	41	Resp:	22612
Ion	Ratio	Lower	Upper
41	100		
69	82.7	67.1	100.7
39	55.6	45.8	68.6





#49

1,4-Dioxane

Concen: 200.587 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 88 Resp: 8627

Ion Ratio Lower Upper

88 100

43 0.0 25.3 37.9





58 76.0 56.4 84.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#50

Toluene-d8

Concen: 9.463 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37


Tgt Ion: 98 Resp: 70000

Ion Ratio Lower Upper

98 100

100 66.0 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 50.007 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Tgt Ion: 43 Resp: 142632
Ion Ratio Lower Upper
43 100
58 38.9 31.6 47.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#52
Toluene
Concen: 9.468 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 92 Resp: 52537
Ion Ratio Lower Upper
92 100
91 170.1 137.3 205.9









#53
t-1,3-Dichloropropene
Concen: 9.592 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Tgt Ion: 75 Resp: 31555
Ion Ratio Lower Upper
75 100
77 35.8 25.4 38.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#54
cis-1,3-Dichloropropene
Concen: 9.802 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 75 Resp: 34684
Ion Ratio Lower Upper
75 100
77 33.1 24.2 36.2
39 46.6 38.7 58.1





#55

1,1,2-Trichloroethane

Concen: 9.675 ug/l

RT: 11.018 min Scan# 1540

Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 97 Resp: 19245

Ion Ratio Lower Upper

97 100

83 85.7 71.0 106.4

85 53.5 41.7 62.5





99 59.9 50.1 75.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#56

Ethyl methacrylate

Concen: 9.551 ug/l

RT: 10.870 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion: 69 Resp: 31262


Ion Ratio Lower Upper

69 100

41 68.2 53.3 79.9

39 38.1 28.8 43.2





#57

1,3-Dichloropropane

Concen: 9.457 ug/l

RT: 11.165 min Scan# 1565

Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 76 Resp: 33613

Ion Ratio Lower Upper

76 100





78 33.7 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#58

2-Chloroethyl Vinyl ether

Concen: 46.784 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37


Tgt Ion: 63 Resp: 71035

Ion Ratio Lower Upper

63 100

106 27.9 21.8 32.6





#59

2-Hexanone

Concen: 48.604 ug/l

RT: 11.194 min Scan# 1571

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 43 Resp: 103615

Ion Ratio Lower Upper

43 100





58 54.2 27.3 81.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#60

Dibromochloromethane

Concen: 9.715 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. -0.000 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37


Tgt Ion:129 Resp: 22833

Ion Ratio Lower Upper

129 100

127 75.9 38.3 114.9









#61
1,2-Dibromoethane
Concen: 9.299 ug/l
RT: 11.470 min Scan# 1617
Delta R.T. 0.006 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Tgt Ion: 107 Resp: 19220
Ion Ratio Lower Upper
107 100
109 98.6 75.4 113.0




Manual Integrations
APPROVED


Reviewed By : John Carlone 10/01/2024
Supervised By : Mahesh Dadoda 10/01/2024



#62
4-Bromofluorobenzene
Concen: 9.186 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 95 Resp: 24756
Ion Ratio Lower Upper
95 100
174 72.9 0.0 145.2
176 70.5 0.0 140.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 16

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion:117 Resp: 262325

Ion Ratio Lower Upper

117 100

82 60.4 47.2 70.8





119 33.9 25.4 38.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#64

Tetrachloroethene

Concen: 9.711 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion:164 Resp: 17743




Ion Ratio Lower Upper


164 100

166 124.4 101.0 151.4

129 98.0 80.8 121.2

131 94.2 78.1 117.1





#65

Chlorobenzene

Concen: 9.645 ug/l

RT: 11.894 min Scan# 1689

Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion:112 Resp: 56099

Ion Ratio Lower Upper

112 100





114 32.6 27.0 40.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#66

1,1,1,2-Tetrachloroethane

Concen: 9.597 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion:131 Resp: 19227


Ion Ratio Lower Upper

131 100

133 100.5 47.9 143.7

119 64.6 32.8 98.3





#67

Ethyl Benzene

Concen: 9.385 ug/l

RT: 11.964 min Scan# 1701

Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 91 Resp: 96533

Ion Ratio Lower Upper

91 100





106 29.2 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#68

m/p-Xylenes

Concen: 19.162 ug/l

RT: 12.070 min Scan# 1720

Delta R.T. -0.000 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37


Tgt Ion:106 Resp: 72910

Ion Ratio Lower Upper

106 100

91 208.7 167.1 250.7





#69

o-Xylene

Concen: 9.713 ug/l

RT: 12.394 min Scan# 1775

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion:106 Resp: 34965

Ion Ratio Lower Upper

106 100





91 220.8 110.2 330.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#70

Styrene

Concen: 9.572 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion:104 Resp: 58367


Ion Ratio Lower Upper

104 100

78 52.7 42.7 64.1

103 53.4 43.6 65.4





#71

Bromoform

Concen: 9.233 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC010

Tgt Ion:	173	Resp:	13618
Ion Ratio	Lower	Upper	
173	100		
175	49.9	23.5	70.6
254	0.0	0.0	0.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.788 min Scan# 2012


Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion:	152	Resp:	119657
Ion Ratio	Lower	Upper	
152	100		
115	62.4	31.3	93.9
150	160.7	0.0	349.8





#73

Isopropylbenzene

Concen: 9.638 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion:105 Resp: 88000

Ion Ratio Lower Upper

105 100





120 26.2 12.9 38.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#74

N-ethyl acetate

Concen: 9.868 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion: 43 Resp: 40801




Ion Ratio Lower Upper


43 100

70 43.4 34.4 51.6

55 24.2 19.6 29.4

61 23.4 19.6 29.4





#75

1,1,2,2-Tetrachloroethane

Concen: 9.822 ug/l

RT: 12.935 min Scan# 18

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC010

Tgt Ion:	83	Resp:	26966
Ion	Ratio	Lower	Upper
83	100		
131	9.8	4.9	14.7
85	64.7	31.9	95.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#76

1,2,3-Trichloropropane

Concen: 10.024 ug/l m




RT: 12.994 min Scan# 1877


Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion:	75	Resp:	23907
Ion	Ratio	Lower	Upper
75	100		
77	205.3	96.9	290.7





#77

Bromobenzene

Concen: 9.734 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. -0.000 min

Lab File: VN084216.D


Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC010



Tgt Ion:	156	Resp:	21439
Ion Ratio	Lower	Upper	
156	100		
77	229.0	115.1	345.3
158	95.3	48.9	146.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#78

n-propylbenzene

Concen: 9.957 ug/l




RT: 13.035 min Scan# 1884


Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion:	91	Resp:	105334
Ion Ratio	Lower	Upper	
91	100		
120	22.1	10.8	32.4





#79

2-Chlorotoluene

Concen: 9.704 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion: 91 Resp: 65511

Ion Ratio Lower Upper

91 100





126 32.0 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#80

1,3,5-Trimethylbenzene

Concen: 9.887 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37


Tgt Ion:105 Resp: 73637

Ion Ratio Lower Upper

105 100

120 47.5 23.7 71.1





#81

trans-1,4-Dichloro-2-butene

Concen: 10.047 ug/l

RT: 12.729 min Scan# 18

Delta R.T. -0.006 min

Lab File: VN084216.D


Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010






Tgt Ion:	75	Resp:	10237
Ion	Ratio	Lower	Upper
75	100		
53	89.4	78.0	117.0
89	44.7	35.5	53.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#82

4-Chlorotoluene

Concen: 9.778 ug/l




RT: 13.223 min Scan# 1916


Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion:	91	Resp:	66449
Ion	Ratio	Lower	Upper
91	100		
126	31.0	15.7	47.1





#83

tert-Butylbenzene

Concen: 9.691 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC010

Tgt Ion:	119	Resp:	64152
Ion	Ratio	Lower	Upper
119	100		
91	67.9	34.1	102.3
134	26.6	12.8	38.4

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#84

1,2,4-Trimethylbenzene

Concen: 9.895 ug/l




RT: 13.482 min Scan# 1960


Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion:	105	Resp:	74245
Ion	Ratio	Lower	Upper
105	100		
120	45.2	22.3	66.8





#85

sec-Butylbenzene

Concen: 10.065 ug/l

RT: 13.611 min Scan# 1983

Delta R.T. -0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010

Tgt Ion:105 Resp: 88913

Ion Ratio Lower Upper

105 100





134 19.4 9.8 29.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#86

p-Isopropyltoluene

Concen: 9.812 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion:119 Resp: 71675


Ion Ratio Lower Upper

119 100

134 24.2 12.6 37.8

91 24.7 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 9.536 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC010

Tgt Ion:	146	Resp:	39402
Ion Ratio	Lower	Upper	
146	100		
111	43.1	21.8	65.3
148	65.7	31.9	95.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#88

1,4-Dichlorobenzene

Concen: 9.392 ug/l




RT: 13.811 min Scan# 2016


Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion:	146	Resp:	39191
Ion Ratio	Lower	Upper	
146	100		
111	45.4	21.2	63.6
148	65.5	31.9	95.9





#89

n-Butylbenzene

Concen: 9.273 ug/l

RT: 14.053 min Scan# 20

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC010

Tgt Ion:	91	Resp:	61422
Ion	Ratio	Lower	Upper
91	100		
92	53.4	26.6	79.7
134	23.2	12.3	36.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#90

Hexachloroethane

Concen: 9.820 ug/l




RT: 14.335 min Scan# 2105


Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37


Tgt Ion:	117	Resp:	14560
Ion	Ratio	Lower	Upper
117	100		
201	68.1	34.4	103.2





#91
1,2-Dichlorobenzene
Concen: 9.526 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010



Tgt Ion	Ratio	Lower	Upper
146	100		
111	43.1	22.0	66.0
148	63.1	32.1	96.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 10.235 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion	Ratio	Lower	Upper
75	100		
155	73.5	38.1	114.5
157	96.2	49.6	149.0





#93

1,2,4-Trichlorobenzene

Concen: 9.501 ug/l

RT: 15.388 min Scan# 2284

Delta R.T. -0.000 min

Lab File: VN084216.D


Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC010



Tgt Ion:	180	Resp:	18960
Ion Ratio	Lower	Upper	
180	100		
182	93.4	47.0	141.0
145	31.7	16.4	49.4

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#94

Hexachlorobutadiene


Concen: 9.638 ug/l

RT: 15.500 min Scan# 2303



Delta R.T. -0.000 min


Lab File: VN084216.D

Acq: 30 Sep 2024 13:37



Tgt Ion:	225	Resp:	9592
Ion Ratio	Lower	Upper	
225	100		
223	59.4	32.4	97.2
227	66.8	32.4	97.0





#95

Naphthalene

Concen: 9.613 ug/l

RT: 15.635 min Scan# 2326

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion:	128	Resp:	63596
Ion	Ratio	Lower	Upper
128	100		
127	12.9	10.5	15.7
129	10.7	8.7	13.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#96

1,2,3-Trichlorobenzene

Concen: 9.826 ug/l




RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion:	180	Resp:	19629
Ion	Ratio	Lower	Upper
180	100		
182	93.5	48.2	144.6
145	36.4	17.3	51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084217.D
 Acq On : 30 Sep 2024 14:00
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/01/2024
 Supervised By : Mahesh Dadoda 10/01/2024

Quant Time: Oct 01 06:49:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	161547	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	288067	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	245710	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	112491	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	13262	5.537	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	11.080%#
35) Dibromofluoromethane	8.171	113	10333	5.441	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	10.880%#
50) Toluene-d8	10.565	98	35782	5.121	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	10.240%#
62) 4-Bromofluorobenzene	12.847	95	13132	5.159	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	10.320%#

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.130	85	9807	5.132	ug/l	94
3) Chloromethane	2.359	50	12064	5.529	ug/l	92
4) Vinyl Chloride	2.512	62	11701	5.536	ug/l	96
5) Bromomethane	2.947	94	8091	5.804	ug/l	97
6) Chloroethane	3.118	64	8433	5.571	ug/l	94
7) Trichlorofluoromethane	3.495	101	17988	5.466	ug/l	90
8) Diethyl Ether	3.953	74	6430	5.268	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.365	101	10220m	5.420	ug/l	
10) Methyl Iodide	4.589	142	12344	5.083	ug/l #	97
11) Tert butyl alcohol	5.512	59	11537	29.211	ug/l #	91
12) 1,1-Dichloroethene	4.342	96	10191	5.603	ug/l	95
13) Acrolein	4.177	56	13122	28.940	ug/l	99
14) Allyl chloride	5.030	41	17724	5.611	ug/l	90
15) Acrylonitrile	5.718	53	26537	26.607	ug/l	99
16) Acetone	4.430	43	27153	26.389	ug/l	95
17) Carbon Disulfide	4.712	76	30816	5.351	ug/l	97
18) Methyl Acetate	5.024	43	12354	5.509	ug/l #	72
19) Methyl tert-butyl Ether	5.800	73	33101	5.392	ug/l	97
20) Methylene Chloride	5.265	84	10809	5.168	ug/l	96
21) trans-1,2-Dichloroethene	5.789	96	10126m	5.314	ug/l	
22) Diisopropyl ether	6.671	45	35654	5.457	ug/l	99
23) Vinyl Acetate	6.606	43	128057m	26.284	ug/l	
24) 1,1-Dichloroethane	6.565	63	19800	5.419	ug/l	95
25) 2-Butanone	7.482	43	37739	26.929	ug/l	99
26) 2,2-Dichloropropane	7.488	77	17552	5.327	ug/l	98
27) cis-1,2-Dichloroethene	7.488	96	12388	5.380	ug/l	99
28) Bromochloromethane	7.812	49	9996	6.131	ug/l #	95
29) Tetrahydrofuran	7.841	42	23456	27.121	ug/l	97
30) Chloroform	7.965	83	20341	5.359	ug/l	92
31) Cyclohexane	8.259	56	21337	6.014	ug/l	99
32) 1,1,1-Trichloroethane	8.171	97	18635	5.467	ug/l #	53
36) 1,1-Dichloropropene	8.371	75	14527	5.266	ug/l	99
37) Ethyl Acetate	7.565	43	16138	5.702	ug/l	97
38) Carbon Tetrachloride	8.359	117	15700	5.188	ug/l	95
39) Methylcyclohexane	9.594	83	15377	5.009	ug/l	94
40) Benzene	8.606	78	45343	5.276	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084217.D
 Acq On : 30 Sep 2024 14:00
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/01/2024
 Supervised By : Mahesh Dadoda 10/01/2024

Quant Time: Oct 01 06:49:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.782	41	7878	5.128	ug/l	93
42) 1,2-Dichloroethane	8.677	62	15661	5.375	ug/l	99
43) Isopropyl Acetate	8.682	43	28029m	5.059	ug/l	
44) Trichloroethene	9.347	130	10921	5.440	ug/l	87
45) 1,2-Dichloropropane	9.618	63	11181	5.499	ug/l	95
46) Dibromomethane	9.706	93	7207	5.253	ug/l	93
47) Bromodichloromethane	9.888	83	16407	5.382	ug/l #	97
48) Methyl methacrylate	9.676	41	11766	5.332	ug/l	97
49) 1,4-Dioxane	9.688	88	4121	101.444	ug/l #	79
51) 4-Methyl-2-Pentanone	10.441	43	69730	25.883	ug/l	99
52) Toluene	10.629	92	26489	5.054	ug/l	96
53) t-1,3-Dichloropropene	10.835	75	16242	5.227	ug/l	94
54) cis-1,3-Dichloropropene	10.312	75	17448	5.220	ug/l	98
55) 1,1,2-Trichloroethane	11.018	97	9805	5.219	ug/l	89
56) Ethyl methacrylate	10.876	69	15794	5.109	ug/l	96
57) 1,3-Dichloropropane	11.159	76	17954	5.348	ug/l	96
58) 2-Chloroethyl Vinyl ether	10.159	63	38417	26.787	ug/l	99
59) 2-Hexanone	11.194	43	51488	25.570	ug/l	98
60) Dibromochloromethane	11.359	129	11412	5.141	ug/l	96
61) 1,2-Dibromoethane	11.470	107	10394	5.324	ug/l	98
64) Tetrachloroethene	11.106	164	9167	5.356	ug/l	90
65) Chlorobenzene	11.888	112	28818	5.290	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.959	131	10054	5.358	ug/l	97
67) Ethyl Benzene	11.965	91	49891	5.178	ug/l	97
68) m/p-Xylenes	12.065	106	35820	10.051	ug/l	98
69) o-Xylene	12.400	106	18041	5.350	ug/l	96
70) Styrene	12.412	104	28472	4.985	ug/l	99
71) Bromoform	12.582	173	7067	5.115	ug/l #	95
73) Isopropylbenzene	12.694	105	43468	5.064	ug/l	98
74) N-amyl acetate	12.494	43	19775	5.087	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.941	83	14517	5.624	ug/l	94
76) 1,2,3-Trichloropropane	12.994	75	14072m	6.276	ug/l	
77) Bromobenzene	12.976	156	11049	5.336	ug/l	99
78) n-propylbenzene	13.035	91	51120	5.140	ug/l	99
79) 2-Chlorotoluene	13.123	91	33161	5.225	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	36100	5.156	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.735	75	4595	4.797	ug/l #	79
82) 4-Chlorotoluene	13.223	91	33382	5.225	ug/l	99
83) tert-Butylbenzene	13.435	119	31990	5.140	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	36099	5.118	ug/l	100
85) sec-Butylbenzene	13.612	105	40954	4.931	ug/l	98
86) p-Isopropyltoluene	13.729	119	34083	4.963	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	20730	5.337	ug/l	98
88) 1,4-Dichlorobenzene	13.812	146	21249	5.417	ug/l	96
89) n-Butylbenzene	14.059	91	30253	4.858	ug/l	99
90) Hexachloroethane	14.329	117	7628	5.472	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	19905	5.226	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.723	75	3046	5.719	ug/l	92
93) 1,2,4-Trichlorobenzene	15.394	180	9439	5.031	ug/l	95
94) Hexachlorobutadiene	15.500	225	5174	5.530	ug/l	89
95) Naphthalene	15.641	128	30391	4.886	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	9245	4.923	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084217.D
Acq On : 30 Sep 2024 14:00
Operator : JC\MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024

Quant Time: Oct 01 06:49:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084217.D
Acq On : 30 Sep 2024 14:00
Operator : JC\MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By : John Carlone 10/01/2024
Supervised By : Mahesh Dadoda 10/01/2024


Quant Time: Oct 01 06:49:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 10
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005



Tgt Ion: 168 Resp: 161547
Ion Ratio Lower Upper
168 100
99 66.3 54.2 81.2




Manual Integrations
APPROVED

Reviewed By : John Carlone 10/01/2024
Supervised By : Mahesh Dadoda 10/01/2024




#2
Dichlorodifluoromethane
Concen: 5.132 ug/l
RT: 2.130 min Scan# 30
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 85 Resp: 9807
Ion Ratio Lower Upper
85 100
87 31.8 17.6 52.9



Abundance Scan 69 (2.359 min): VN084214.D\data.ms (-63)



#3

Chloromethane

Concen: 5.529 ug/l

RT: 2.359 min Scan# 69

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 50 Resp: 12064

Ion Ratio Lower Upper

50 100


52 28.9 26.8 40.2

Manual Integrations


APPROVED

Reviewed By :John Carlone 10/01/2024


Supervised By :Mahesh Dadoda 10/01/2024




Abundance



Abundance Scan 69 (2.359 min): VN084217.D\data.ms (-18)



Abundance Scan 96 (2.518 min): VN084214.D\data.ms (-88)



#4

Vinyl Chloride

Concen: 5.536 ug/l

RT: 2.512 min Scan# 95

Delta R.T. -0.006 min

Lab File: VN084217.D


Acq: 30 Sep 2024 14:00

Tgt Ion: 62 Resp: 11701


Ion Ratio Lower Upper

62 100


64 35.0 26.2 39.4




Abundance



Abundance Scan 95 (2.512 min): VN084217.D\data.ms (-45)





#5

Bromomethane

Concen: 5.804 ug/l

RT: 2.947 min Scan# 16

Delta R.T. -0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion: 94 Resp: 8091

Ion Ratio Lower Upper

94 100





96 88.4 72.9 109.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#6

Chloroethane

Concen: 5.571 ug/l

RT: 3.118 min Scan# 198

Delta R.T. -0.000 min

Lab File: VN084217.D




Acq: 30 Sep 2024 14:00


Tgt Ion: 64 Resp: 8433

Ion Ratio Lower Upper

64 100

66 32.8 23.5 35.3





#7

Trichlorofluoromethane

Concen: 5.466 ug/l

RT: 3.495 min Scan# 26

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:101 Resp: 17988

Ion Ratio Lower Upper

101 100





103 72.8 51.8 77.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#8

Diethyl Ether

Concen: 5.268 ug/l

RT: 3.953 min Scan# 340

Delta R.T. -0.006 min

Lab File: VN084217.D




Acq: 30 Sep 2024 14:00


Tgt Ion: 74 Resp: 6430

Ion Ratio Lower Upper

74 100

45 96.9 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 5.420 ug/l m
RT: 4.365 min Scan# 411
Delta R.T. -0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion	Ratio	Lower	Upper
101	100		
85	46.1	36.2	54.4
151	71.5	59.3	88.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#10
Methyl Iodide
Concen: 5.083 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion	Ratio	Lower	Upper
142	100		
127	44.9	35.3	52.9
141	18.0	11.0	16.6





#11

Tert butyl alcohol

Concen: 29.211 ug/l

RT: 5.512 min Scan# 607

Delta R.T. -0.012 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 59 Resp: 11537

Ion Ratio Lower Upper

59 100





57 7.0 8.2 12.2#

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#12

1,1-Dichloroethene

Concen: 5.603 ug/l

RT: 4.342 min Scan# 406

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00




Tgt Ion: 96 Resp: 10191


Ion Ratio Lower Upper

96 100

61 164.1 128.6 192.8

98 74.5 51.7 77.5





#13

Acrolein

Concen: 28.940 ug/l

RT: 4.177 min Scan# 37

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion: 56 Resp: 13122

Ion Ratio Lower Upper

56 100





55 70.5 57.2 85.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#14

Allyl chloride

Concen: 5.611 ug/l

RT: 5.030 min Scan# 523

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00




Tgt Ion: 41 Resp: 17724


Ion Ratio Lower Upper

41 100

39 62.5 59.0 88.6

76 35.6 29.8 44.6





#15

Acrylonitrile

Concen: 26.607 ug/l

RT: 5.718 min Scan# 640

Delta R.T. -0.000 min

Lab File: VN084217.D


Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :




VSTDICC005



Tgt Ion	53	Resp	26537
Ion	Ratio	Lower	Upper
53	100		
52	82.1	66.2	99.4
51	35.6	28.8	43.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#16

Acetone


Concen: 26.389 ug/l

RT: 4.430 min Scan# 421



Delta R.T. 0.006 min


Lab File: VN084217.D

Acq: 30 Sep 2024 14:00



Tgt Ion	43	Resp	27153
Ion	Ratio	Lower	Upper
43	100		
58	27.9	24.4	36.6





#17

Carbon Disulfide

Concen: 5.351 ug/l

RT: 4.712 min Scan# 469

Delta R.T. -0.000 min

Lab File: VN084217.D


Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005



Tgt Ion: 76 Resp: 30816

Ion Ratio Lower Upper

76 100




78 7.9 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#18

Methyl Acetate

Concen: 5.509 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN084217.D




Acq: 30 Sep 2024 14:00


Tgt Ion: 43 Resp: 12354

Ion Ratio Lower Upper

43 100

74 9.9 19.2 28.8#





#19

Methyl tert-butyl Ether

Concen: 5.392 ug/l

RT: 5.800 min Scan# 65

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 73 Resp: 33101

Ion Ratio Lower Upper

73 100





57 23.3 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#20

Methylene Chloride

Concen: 5.168 ug/l

RT: 5.265 min Scan# 563

Delta R.T. -0.012 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion: 84 Resp: 10809




Ion Ratio Lower Upper


84 100

49 127.8 107.2 160.8

51 38.2 31.8 47.8

86 63.0 52.9 79.3





#21

trans-1,2-Dichloroethene

Concen: 5.314 ug/l m

RT: 5.789 min Scan# 652

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 96 Resp: 10126

Ion Ratio Lower Upper

96 100

61 139.8 108.4 162.6





98 67.5 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#22

Diisopropyl ether

Concen: 5.457 ug/l

RT: 6.671 min Scan# 802

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion: 45 Resp: 35654




Ion Ratio Lower Upper


45 100

43 52.7 42.3 63.5

87 28.1 21.8 32.6

59 10.5 9.6 14.4





#23

Vinyl Acetate

Concen: 26.284 ug/l m

RT: 6.606 min Scan# 790

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 43 Resp: 128057

Ion Ratio Lower Upper

43 100





86 11.2 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#24

1,1-Dichloroethane

Concen: 5.419 ug/l

RT: 6.565 min Scan# 784

Delta R.T. -0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00




Tgt Ion: 63 Resp: 19800


Ion Ratio Lower Upper

63 100

98 8.9 3.4 10.2

100 4.5 2.0 5.8





#25

2-Butanone

Concen: 26.929 ug/l

RT: 7.482 min Scan# 940

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion: 43 Resp: 37739

Ion Ratio Lower Upper

43 100

72 26.4 21.4 32.2

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#26

2,2-Dichloropropane

Concen: 5.327 ug/l

RT: 7.488 min Scan# 941

Delta R.T. -0.000 min

Lab File: VN084217.D




Acq: 30 Sep 2024 14:00


Tgt Ion: 77 Resp: 17552

Ion Ratio Lower Upper

77 100

97 22.4 10.8 32.3





#27

cis-1,2-Dichloroethene

Concen: 5.380 ug/l

RT: 7.488 min Scan# 941

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 96 Resp: 12388

Ion Ratio Lower Upper

96 100

61 148.3 0.0 301.8





98 64.1 0.0 128.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#28

Bromochloromethane

Concen: 6.131 ug/l

RT: 7.812 min Scan# 996

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00




Tgt Ion: 49 Resp: 9996


Ion Ratio Lower Upper

49 100

129 0.0 0.0 3.6

130 68.3 58.0 87.0





#29

Tetrahydrofuran

Concen: 27.121 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC005

Tgt Ion:	42	Resp:	23456
Ion	Ratio	Lower	Upper
42	100		
72	40.3	34.6	51.8
71	40.1	32.2	48.4

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#30

Chloroform

Concen: 5.359 ug/l




RT: 7.965 min Scan# 1022


Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion:	83	Resp:	20341
Ion	Ratio	Lower	Upper
83	100		
85	70.0	51.0	76.4





#31

Cyclohexane

Concen: 6.014 ug/l

RT: 8.259 min Scan# 1071

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC005

Tgt Ion:	56	Resp:	21337
Ion	Ratio	Lower	Upper
56	100		
69	29.5	23.8	35.6
84	85.5	68.8	103.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#32

1,1,1-Trichloroethane

Concen: 5.467 ug/l




RT: 8.171 min Scan# 1057


Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion:	97	Resp:	18635
Ion	Ratio	Lower	Upper
97	100		
99	0.0	51.2	76.8#
61	47.1	37.5	56.3









#33
1,2-Dichloroethane-d4
Concen: 5.537 ug/l
RT: 8.571 min Scan# 1126
Delta R.T. -0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion: 65 Resp: 13262
Ion Ratio Lower Upper
65 100
67 51.7 0.0 102.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion:114 Resp: 288067
Ion Ratio Lower Upper
114 100
63 21.9 0.0 43.8
88 15.7 0.0 31.6





#35

Dibromofluoromethane

Concen: 5.441 ug/l

RT: 8.171 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC005

Tgt Ion:	113	Resp:	10333
Ion	Ratio	Lower	Upper
113	100		
111	105.9	83.3	124.9
192	16.6	13.5	20.3

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#36

1,1-Dichloropropene

Concen: 5.266 ug/l




RT: 8.371 min Scan# 1091


Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion:	75	Resp:	14527
Ion	Ratio	Lower	Upper
75	100		
110	34.3	16.9	50.7
77	30.0	24.7	37.1





#37

Ethyl Acetate

Concen: 5.702 ug/l

RT: 7.565 min Scan# 95

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC005

Tgt Ion:	43	Resp:	16138
Ion	Ratio	Lower	Upper
43	100		
61	12.7	10.7	16.1
70	9.0	8.8	13.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#38

Carbon Tetrachloride

Concen: 5.188 ug/l




RT: 8.359 min Scan# 1089


Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion:	117	Resp:	15700
Ion	Ratio	Lower	Upper
117	100		
119	93.4	78.7	118.1
121	32.3	24.4	36.6





#39

Methylcyclohexane

Concen: 5.009 ug/l

RT: 9.594 min Scan# 12

Delta R.T. -0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 83 Resp: 15377

Ion Ratio Lower Upper

83 100

55 71.9 62.8 94.2





98 46.9 38.8 58.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#40

Benzene

Concen: 5.276 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN084217.D




Acq: 30 Sep 2024 14:00


Tgt Ion: 78 Resp: 45343

Ion Ratio Lower Upper

78 100

77 23.3 19.1 28.7





#41

Methacrylonitrile

Concen: 5.128 ug/l

RT: 7.782 min Scan# 991

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC005

Tgt Ion:	41	Resp:	7878
Ion	Ratio	Lower	Upper
41	100		
39	63.1	43.3	64.9
67	74.5	56.1	84.1
52	28.7	24.5	36.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#42

1,2-Dichloroethane

Concen: 5.375 ug/l




RT: 8.677 min Scan# 1143


Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion:	62	Resp:	15661
Ion	Ratio	Lower	Upper
62	100		
98	9.3	0.0	18.2





#43

Isopropyl Acetate

Concen: 5.059 ug/l m

RT: 8.682 min Scan# 1145

Delta R.T. -0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 43 Resp: 28029

Ion Ratio Lower Upper

43 100

61 24.0 20.4 30.6





87 11.7 10.3 15.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#44

Trichloroethene

Concen: 5.440 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. -0.000 min

Lab File: VN084217.D




Acq: 30 Sep 2024 14:00


Tgt Ion:130 Resp: 10921

Ion Ratio Lower Upper

130 100

95 88.8 0.0 203.0





#45

1,2-Dichloropropane

Concen: 5.499 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 63 Resp: 11181

Ion Ratio Lower Upper

63 100





65 33.3 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#46

Dibromomethane

Concen: 5.253 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00




Tgt Ion: 93 Resp: 7207


Ion Ratio Lower Upper

93 100

95 88.6 65.0 97.4

174 95.8 71.8 107.8





#47

Bromodichloromethane

Concen: 5.382 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC005

Tgt Ion:	83	Resp:	16407
Ion	Ratio	Lower	Upper
83	100		
85	66.6	51.5	77.3
127	8.5	5.7	8.5

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#48

Methyl methacrylate

Concen: 5.332 ug/l




RT: 9.676 min Scan# 1313


Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion:	41	Resp:	11766
Ion	Ratio	Lower	Upper
41	100		
69	81.4	67.1	100.7
39	60.1	45.8	68.6





#49

1,4-Dioxane

Concen: 101.444 ug/l

RT: 9.688 min Scan# 1316

Delta R.T. -0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 88 Resp: 4121

Ion Ratio Lower Upper

88 100

43 0.0 25.3 37.9





58 74.7 56.4 84.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#50

Toluene-d8

Concen: 5.121 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN084217.D




Acq: 30 Sep 2024 14:00


Tgt Ion: 98 Resp: 35782

Ion Ratio Lower Upper

98 100

100 66.5 52.7 79.1





#51

4-Methyl-2-Pentanone

Concen: 25.883 ug/l

RT: 10.441 min Scan# 1443

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion: 43 Resp: 69730

Ion Ratio Lower Upper

43 100





58 39.0 31.6 47.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#52

Toluene

Concen: 5.054 ug/l

RT: 10.629 min Scan# 1475

Delta R.T. -0.000 min

Lab File: VN084217.D




Acq: 30 Sep 2024 14:00


Tgt Ion: 92 Resp: 26489

Ion Ratio Lower Upper

92 100

91 176.6 137.3 205.9





#53

t-1,3-Dichloropropene

Concen: 5.227 ug/l

RT: 10.835 min Scan# 1510

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 75 Resp: 16242

Ion Ratio Lower Upper

75 100





77 34.9 25.4 38.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#54

cis-1,3-Dichloropropene

Concen: 5.220 ug/l

RT: 10.312 min Scan# 1421

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00




Tgt Ion: 75 Resp: 17448


Ion Ratio Lower Upper

75 100

77 32.2 24.2 36.2

39 47.8 38.7 58.1





#55

1,1,2-Trichloroethane

Concen: 5.219 ug/l

RT: 11.018 min Scan# 15

Delta R.T. 0.006 min

Lab File: VN084217.D


Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005



Tgt Ion: 97 Resp: 9805

Ion Ratio Lower Upper

97 100

83 100.1 71.0 106.4

85 58.2 41.7 62.5




99 70.8 50.1 75.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#56

Ethyl methacrylate


Concen: 5.109 ug/l

RT: 10.876 min Scan# 1517

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00





Tgt Ion: 69 Resp: 15794


Ion Ratio Lower Upper

69 100

41 65.1 53.3 79.9

39 31.7 28.8 43.2





#57

1,3-Dichloropropane

Concen: 5.348 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 76 Resp: 17954

Ion Ratio Lower Upper

76 100





78 29.8 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#58

2-Chloroethyl Vinyl ether

Concen: 26.787 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN084217.D




Acq: 30 Sep 2024 14:00


Tgt Ion: 63 Resp: 38417

Ion Ratio Lower Upper

63 100

106 27.8 21.8 32.6





#59

2-Hexanone

Concen: 25.570 ug/l

RT: 11.194 min Scan# 1571

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 43 Resp: 51488

Ion Ratio Lower Upper

43 100





58 52.7 27.3 81.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#60

Dibromochloromethane

Concen: 5.141 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. -0.000 min

Lab File: VN084217.D




Acq: 30 Sep 2024 14:00


Tgt Ion:129 Resp: 11412

Ion Ratio Lower Upper

129 100

127 80.4 38.3 114.9









#61
1,2-Dibromoethane
Concen: 5.324 ug/l
RT: 11.470 min Scan# 1617
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Tgt Ion: 107 Resp: 10394
Ion Ratio Lower Upper
107 100
109 92.6 75.4 113.0




Manual Integrations
APPROVED


Reviewed By : John Carlone 10/01/2024
Supervised By : Mahesh Dadoda 10/01/2024



#62
4-Bromofluorobenzene
Concen: 5.159 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 95 Resp: 13132
Ion Ratio Lower Upper
95 100
174 70.1 0.0 145.2
176 64.1 0.0 140.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion:117 Resp: 245710

Ion Ratio Lower Upper

117 100

82 60.4 47.2 70.8





119 32.4 25.4 38.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#64

Tetrachloroethene

Concen: 5.356 ug/l

RT: 11.106 min Scan# 1556

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion:164 Resp: 9167




Ion Ratio Lower Upper


164 100

166 109.1 101.0 151.4

129 90.7 80.8 121.2

131 94.5 78.1 117.1





#65

Chlorobenzene

Concen: 5.290 ug/l

RT: 11.888 min Scan# 1689

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:112 Resp: 28818

Ion Ratio Lower Upper

112 100





114 33.0 27.0 40.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#66

1,1,1,2-Tetrachloroethane

Concen: 5.358 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00




Tgt Ion:131 Resp: 10054


Ion Ratio Lower Upper

131 100

133 91.7 47.9 143.7

119 64.8 32.8 98.3





#67

Ethyl Benzene

Concen: 5.178 ug/l

RT: 11.965 min Scan# 1719

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion: 91 Resp: 49891

Ion Ratio Lower Upper

91 100





106 28.7 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#68

m/p-Xylenes

Concen: 10.051 ug/l

RT: 12.065 min Scan# 1719

Delta R.T. -0.006 min

Lab File: VN084217.D




Acq: 30 Sep 2024 14:00


Tgt Ion:106 Resp: 35820

Ion Ratio Lower Upper

106 100

91 212.5 167.1 250.7





#69

o-Xylene

Concen: 5.350 ug/l

RT: 12.400 min Scan# 1775

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:106 Resp: 18041

Ion Ratio Lower Upper

106 100





91 214.1 110.2 330.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#70

Styrene

Concen: 4.985 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00




Tgt Ion:104 Resp: 28472


Ion Ratio Lower Upper

104 100

78 52.7 42.7 64.1

103 54.5 43.6 65.4





#71

Bromoform

Concen: 5.115 ug/l

RT: 12.582 min Scan# 1806

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion:	173	Resp:	7067
Ion Ratio	Lower	Upper	
173	100		
175	50.2	23.5	70.6
254	0.0	0.0	0.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.788 min Scan# 2012


Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion:	152	Resp:	112491
Ion Ratio	Lower	Upper	
152	100		
115	62.1	31.3	93.9
150	158.0	0.0	349.8





#73

Isopropylbenzene

Concen: 5.064 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion:105 Resp: 43468

Ion Ratio Lower Upper

105 100





120 26.7 12.9 38.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#74

N-ethyl acetate

Concen: 5.087 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion: 43 Resp: 19775




Ion Ratio Lower Upper


43 100

70 43.5 34.4 51.6

55 26.3 19.6 29.4

61 24.8 19.6 29.4





#75

1,1,2,2-Tetrachloroethane

Concen: 5.624 ug/l

RT: 12.941 min Scan# 1868

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC005

Tgt Ion:	83	Resp:	14517
Ion	Ratio	Lower	Upper
83	100		
131	9.2	4.9	14.7
85	59.0	31.9	95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#76

1,2,3-Trichloropropane

Concen: 6.276 ug/l m




RT: 12.994 min Scan# 1877


Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion:	75	Resp:	14072
Ion	Ratio	Lower	Upper
75	100		
77	180.3	96.9	290.7





#77

Bromobenzene

Concen: 5.336 ug/l

RT: 12.976 min Scan# 1875

Delta R.T. -0.006 min

Lab File: VN084217.D


Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC005



Tgt Ion:	156	Resp:	11049
Ion	Ratio	Lower	Upper
156	100		
77	229.7	115.1	345.3
158	96.2	48.9	146.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#78

n-propylbenzene

Concen: 5.140 ug/l




RT: 13.035 min Scan# 1884


Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion:	91	Resp:	51120
Ion	Ratio	Lower	Upper
91	100		
120	22.3	10.8	32.4





#79

2-Chlorotoluene

Concen: 5.225 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion: 91 Resp: 33161

Ion Ratio Lower Upper

91 100





126 31.0 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#80

1,3,5-Trimethylbenzene

Concen: 5.156 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN084217.D




Acq: 30 Sep 2024 14:00


Tgt Ion:105 Resp: 36100

Ion Ratio Lower Upper

105 100

120 49.2 23.7 71.1





#81

trans-1,4-Dichloro-2-butene

Concen: 4.797 ug/l

RT: 12.735 min Scan# 1833

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:	75	Resp:	4595
Ion	Ratio	Lower	Upper
75	100		
53	120.1	78.0	117.0
89	34.5	35.5	53.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#82

4-Chlorotoluene

Concen: 5.225 ug/l




RT: 13.223 min Scan# 1916


Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion:	91	Resp:	33382
Ion	Ratio	Lower	Upper
91	100		
126	32.1	15.7	47.1





#83

tert-Butylbenzene

Concen: 5.140 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC005

Tgt Ion:	119	Resp:	31990
Ion	Ratio	Lower	Upper
119	100		
91	68.5	34.1	102.3
134	26.4	12.8	38.4

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#84

1,2,4-Trimethylbenzene

Concen: 5.118 ug/l




RT: 13.482 min Scan# 1960


Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion:	105	Resp:	36099
Ion	Ratio	Lower	Upper
105	100		
120	44.2	22.3	66.8





#85

sec-Butylbenzene

Concen: 4.931 ug/l

RT: 13.612 min Scan# 1983

Delta R.T. -0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005

Tgt Ion:105 Resp: 40954

Ion Ratio Lower Upper

105 100





134 20.6 9.8 29.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#86

p-Isopropyltoluene

Concen: 4.963 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00




Tgt Ion:119 Resp: 34083


Ion Ratio Lower Upper

119 100

134 25.2 12.6 37.8

91 26.0 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 5.337 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:	146	Resp:	20730
Ion	Ratio	Lower	Upper
146	100		
111	45.0	21.8	65.3
148	64.7	31.9	95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#88

1,4-Dichlorobenzene

Concen: 5.417 ug/l




RT: 13.812 min Scan# 2016


Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion:	146	Resp:	21249
Ion	Ratio	Lower	Upper
146	100		
111	47.9	21.2	63.6
148	64.5	31.9	95.9





#89

n-Butylbenzene

Concen: 4.858 ug/l

RT: 14.059 min Scan# 2057

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC005

Tgt Ion:	91	Resp:	30253
Ion	Ratio	Lower	Upper
91	100		
92	53.8	26.6	79.7
134	24.9	12.3	36.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#90

Hexachloroethane

Concen: 5.472 ug/l




RT: 14.329 min Scan# 2104


Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00


Tgt Ion:	117	Resp:	7628
Ion	Ratio	Lower	Upper
117	100		
201	67.4	34.4	103.2





#91
1,2-Dichlorobenzene
Concen: 5.226 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005



Tgt Ion: 146 Resp: 19905
Ion Ratio Lower Upper
146 100
111 44.8 22.0 66.0
148 66.8 32.1 96.3




Manual Integrations
APPROVED


Reviewed By : John Carlone 10/01/2024
Supervised By : Mahesh Dadoda 10/01/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 5.719 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 75 Resp: 3046
Ion Ratio Lower Upper
75 100
155 65.1 38.1 114.5
157 95.3 49.6 149.0





#93

1,2,4-Trichlorobenzene

Concen: 5.031 ug/l

RT: 15.394 min Scan# 2284

Delta R.T. 0.006 min

Lab File: VN084217.D


Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005






Tgt Ion:	180	Resp:	9439
Ion	Ratio	Lower	Upper
180	100		
182	97.8	47.0	141.0
145	37.6	16.4	49.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#94

Hexachlorobutadiene


Concen: 5.530 ug/l

RT: 15.500 min Scan# 2303



Delta R.T. -0.000 min


Lab File: VN084217.D

Acq: 30 Sep 2024 14:00



Tgt Ion:	225	Resp:	5174
Ion	Ratio	Lower	Upper
225	100		
223	56.0	32.4	97.2
227	56.4	32.4	97.0





#95

Naphthalene

Concen: 4.886 ug/l

RT: 15.641 min Scan# 23

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion:	128	Resp:	30391
Ion Ratio	Lower	Upper	
128	100		
127	13.2	10.5	15.7
129	10.4	8.7	13.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#96

1,2,3-Trichlorobenzene

Concen: 4.923 ug/l




RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion:	180	Resp:	9245
Ion Ratio	Lower	Upper	
180	100		
182	94.9	48.2	144.6
145	36.0	17.3	51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084218.D
 Acq On : 30 Sep 2024 14:48
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/01/2024
 Supervised By : Mahesh Dadoda 10/01/2024

Quant Time: Oct 01 06:49:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	160573	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	290146	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	244376	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	101963	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount 50.000	Range 74 - 125		Recovery =	0.000	%#	
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount 50.000	Range 75 - 124		Recovery =	0.000	%#	
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount 50.000	Range 86 - 113		Recovery =	0.000	%#	
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount 50.000	Range 77 - 121		Recovery =	0.000	%#	
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	1928	1.015	ug/l	83
3) Chloromethane	2.365	50	2217	1.022	ug/l	99
4) Vinyl Chloride	2.512	62	1981	0.943	ug/l	89
6) Chloroethane	3.124	64	2029	1.349	ug/l	97
7) Trichlorofluoromethane	3.500	101	3140	0.960	ug/l	94
8) Diethyl Ether	3.965	74	1167	0.962	ug/l	88
9) 1,1,2-Trichlorotrifluo...	4.371	101	1934	1.032	ug/l #	63
12) 1,1-Dichloroethene	4.353	96	1584	0.876	ug/l #	50
14) Allyl chloride	5.018	41	2852m	0.908	ug/l	
15) Acrylonitrile	5.736	53	4920	4.963	ug/l #	81
16) Acetone	4.436	43	4799	4.692	ug/l	90
17) Carbon Disulfide	4.718	76	6680	1.167	ug/l	100
18) Methyl Acetate	5.030	43	2249	1.009	ug/l #	64
19) Methyl tert-butyl Ether	5.806	73	5319	0.872	ug/l #	83
20) Methylene Chloride	5.277	84	2203	1.060	ug/l #	83
21) trans-1,2-Dichloroethene	5.783	96	1753	0.925	ug/l	86
22) Diisopropyl ether	6.671	45	5639	0.868	ug/l #	96
23) Vinyl Acetate	6.606	43	19212m	3.967	ug/l	
24) 1,1-Dichloroethane	6.571	63	3358	0.925	ug/l #	83
25) 2-Butanone	7.488	43	6488	4.658	ug/l	91
26) 2,2-Dichloropropane	7.494	77	2902	0.886	ug/l	96
27) cis-1,2-Dichloroethene	7.483	96	2259	0.987	ug/l	92
28) Bromochloromethane	7.824	49	1718	1.060	ug/l #	87
29) Tetrahydrofuran	7.841	42	4006	4.660	ug/l	92
30) Chloroform	7.971	83	3793	1.005	ug/l	99
32) 1,1,1-Trichloroethane	8.171	97	3120	0.921	ug/l #	53
36) 1,1-Dichloropropene	8.365	75	2420	0.871	ug/l	98
37) Ethyl Acetate	7.559	43	2596	0.911	ug/l #	78
38) Carbon Tetrachloride	8.359	117	2795	0.917	ug/l	90
39) Methylcyclohexane	9.600	83	2486	0.804	ug/l #	88
40) Benzene	8.606	78	8245	0.952	ug/l	94
41) Methacrylonitrile	7.782	41	1490	0.963	ug/l #	66
42) 1,2-Dichloroethane	8.671	62	2669	0.909	ug/l	88
43) Isopropyl Acetate	8.688	43	7614m	1.364	ug/l	
44) Trichloroethene	9.347	130	1885	0.932	ug/l	87
45) 1,2-Dichloropropane	9.624	63	1676	0.818	ug/l #	44

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084218.D
 Acq On : 30 Sep 2024 14:48
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024

Quant Time: Oct 01 06:49:39 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M

Quant Title : SW846 8260

QLast Update : Tue Oct 01 06:46:32 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	9.706	93	1144	0.828	ug/l	87
47) Bromodichloromethane	9.894	83	2754	0.897	ug/l #	93
48) Methyl methacrylate	9.676	41	1824	0.821	ug/l	93
49) 1,4-Dioxane	9.694	88	800	19.552	ug/l #	78
51) 4-Methyl-2-Pentanone	10.447	43	11218	4.134	ug/l	97
52) Toluene	10.629	92	4875	0.923	ug/l	91
53) t-1,3-Dichloropropene	10.829	75	2493	0.797	ug/l	98
54) cis-1,3-Dichloropropene	10.306	75	2720	0.808	ug/l	95
55) 1,1,2-Trichloroethane	11.018	97	1669	0.882	ug/l	90
56) Ethyl methacrylate	10.876	69	2402	0.771	ug/l #	77
57) 1,3-Dichloropropane	11.159	76	3092	0.914	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	6179	4.278	ug/l	98
59) 2-Hexanone	11.194	43	8282	4.084	ug/l	96
60) Dibromochloromethane	11.353	129	1916	0.857	ug/l	91
61) 1,2-Dibromoethane	11.465	107	1795	0.913	ug/l	99
64) Tetrachloroethene	11.106	164	1690	0.993	ug/l #	87
65) Chlorobenzene	11.888	112	5025	0.927	ug/l #	86
66) 1,1,1,2-Tetrachloroethane	11.959	131	1632	0.874	ug/l #	66
67) Ethyl Benzene	11.965	91	8582	0.896	ug/l #	88
68) m/p-Xylenes	12.070	106	5869	1.656	ug/l	95
69) o-Xylene	12.400	106	2398	0.715	ug/l	83
70) Styrene	12.412	104	4489	0.790	ug/l	94
71) Bromoform	12.576	173	1168	0.850	ug/l #	94
73) Isopropylbenzene	12.688	105	6991	0.899	ug/l	97
74) N-amyl acetate	12.494	43	3228	0.916	ug/l #	88
75) 1,1,2,2-Tetrachloroethane	12.941	83	2233	0.954	ug/l #	95
76) 1,2,3-Trichloropropane	12.994	75	1826m	0.899	ug/l	
77) Bromobenzene	12.976	156	1818	0.969	ug/l	98
78) n-propylbenzene	13.035	91	7045	0.781	ug/l	95
79) 2-Chlorotoluene	13.129	91	5383	0.936	ug/l	92
80) 1,3,5-Trimethylbenzene	13.170	105	4830	0.761	ug/l	95
82) 4-Chlorotoluene	13.223	91	5174	0.893	ug/l	100
83) tert-Butylbenzene	13.435	119	4909	0.870	ug/l	88
84) 1,2,4-Trimethylbenzene	13.482	105	4760	0.744	ug/l	100
85) sec-Butylbenzene	13.612	105	5795	0.770	ug/l	99
86) p-Isopropyltoluene	13.723	119	4495	0.722	ug/l	98
87) 1,3-Dichlorobenzene	13.729	146	3424	0.972	ug/l	96
88) 1,4-Dichlorobenzene	13.812	146	3648m	1.026	ug/l	
89) n-Butylbenzene	14.053	91	4866	0.862	ug/l	97
90) Hexachloroethane	14.335	117	1151	0.911	ug/l	82
91) 1,2-Dichlorobenzene	14.100	146	3610	1.046	ug/l	92
92) 1,2-Dibromo-3-Chloropr...	14.711	75	423	0.876	ug/l	65
93) 1,2,4-Trichlorobenzene	15.400	180	1340	0.788	ug/l	89
94) Hexachlorobutadiene	15.494	225	861	1.015	ug/l	90
95) Naphthalene	15.641	128	5031	0.892	ug/l	97
96) 1,2,3-Trichlorobenzene	15.847	180	1531	0.899	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084218.D
Acq On : 30 Sep 2024 14:48
Operator : JC\MD
Sample : VSTDIC001
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Manual Integrations
APPROVED

Reviewed By : John Carlone 10/01/2024
Supervised By : Mahesh Dadoda 10/01/2024


Quant Time: Oct 01 06:49:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001



Tgt Ion: 168 Resp: 160573
Ion Ratio Lower Upper
168 100
99 68.1 54.2 81.2




Manual Integrations
APPROVED


Reviewed By : John Carlone 10/01/2024
Supervised By : Mahesh Dadoda 10/01/2024



#2
Dichlorodifluoromethane
Concen: 1.015 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 85 Resp: 1928
Ion Ratio Lower Upper
85 100
87 25.3 17.6 52.9





#3

Chloromethane

Concen: 1.022 ug/l

RT: 2.365 min Scan# 70

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC001

Tgt Ion:	50	Resp:	2217
Ion	Ratio	Lower	Upper
50	100		
52	34.2	26.8	40.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#4

Vinyl Chloride

Concen: 0.943 ug/l




RT: 2.512 min Scan# 95


Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48


Tgt Ion:	62	Resp:	1981
Ion	Ratio	Lower	Upper
62	100		
64	39.0	26.2	39.4





#6
Chloroethane
Concen: 1.349 ug/l
RT: 3.124 min Scan# 198
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 64 Resp: 2029
Ion Ratio Lower Upper
64 100
66 31.0 23.5 35.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#7
Trichlorofluoromethane
Concen: 0.960 ug/l
RT: 3.500 min Scan# 263
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion:101 Resp: 3140
Ion Ratio Lower Upper
101 100
103 60.0 51.8 77.6





#8

Diethyl Ether

Concen: 0.962 ug/l

RT: 3.965 min Scan# 341

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 74 Resp: 1167

Ion Ratio Lower Upper

74 100





45 83.9 47.7 143.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#9

1,1,2-Trichlorotrifluoroethane

Concen: 1.032 ug/l

RT: 4.371 min Scan# 411

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion:101 Resp: 1934


Ion Ratio Lower Upper

101 100

85 25.4 36.2 54.4#

151 38.8 59.3 88.9#





#12

1,1-Dichloroethene

Concen: 0.876 ug/l

RT: 4.353 min Scan# 406

Delta R.T. 0.012 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion: 96 Resp: 1584

Ion Ratio Lower Upper

96 100

61 88.5 128.6 192.8#

98 34.3 51.7 77.5#

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#14

Allyl chloride

Concen: 0.908 ug/l m

RT: 5.018 min Scan# 521

Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion: 41 Resp: 2852


Ion Ratio Lower Upper

41 100

39 68.0 59.0 88.6

76 24.5 29.8 44.6#





#15

Acrylonitrile

Concen: 4.963 ug/l

RT: 5.736 min Scan# 640

Delta R.T. 0.018 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001





Tgt Ion:	53	Resp:	4920
Ion	Ratio	Lower	Upper
53	100		
52	82.8	66.2	99.4
51	0.0	28.8	43.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#16

Acetone

Concen: 4.692 ug/l




RT: 4.436 min Scan# 422


Delta R.T. 0.012 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion:	43	Resp:	4799
Ion	Ratio	Lower	Upper
43	100		
58	24.9	24.4	36.6





#17

Carbon Disulfide

Concen: 1.167 ug/l

RT: 4.718 min Scan# 471

Delta R.T. 0.006 min

Lab File: VN084218.D


Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001



Tgt Ion: 76 Resp: 6680

Ion Ratio Lower Upper

76 100




78 9.1 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#18

Methyl Acetate

Concen: 1.009 ug/l

RT: 5.030 min Scan# 523

Delta R.T. 0.012 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion: 43 Resp: 2249

Ion Ratio Lower Upper

43 100

74 6.4 19.2 28.8#





#19

Methyl tert-butyl Ether

Concen: 0.872 ug/l

RT: 5.806 min Scan# 653

Delta R.T. 0.012 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 73 Resp: 5319

Ion Ratio Lower Upper

73 100





57 13.9 17.7 26.5#

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#20

Methylene Chloride

Concen: 1.060 ug/l

RT: 5.277 min Scan# 565

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion: 84 Resp: 2203




Ion Ratio Lower Upper


84 100

49 100.6 107.2 160.8#

51 43.3 31.8 47.8

86 64.5 52.9 79.3





#21

trans-1,2-Dichloroethene

Concen: 0.925 ug/l

RT: 5.783 min Scan# 65

Delta R.T. -0.006 min

Lab File: VN084218.D


Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001



Tgt Ion: 96 Resp: 1753

Ion Ratio Lower Upper

96 100

61 115.2 108.4 162.6




98 55.5 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#22

Diisopropyl ether


Concen: 0.868 ug/l

RT: 6.671 min Scan# 802

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48



Tgt Ion: 45 Resp: 5639



Ion Ratio Lower Upper


45 100

43 52.8 42.3 63.5

87 22.2 21.8 32.6

59 16.4 9.6 14.4#





#23

Vinyl Acetate

Concen: 3.967 ug/l m

RT: 6.606 min Scan# 791

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 43 Resp: 19212

Ion Ratio Lower Upper

43 100





86 13.5 8.6 12.8#

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#24

1,1-Dichloroethane

Concen: 0.925 ug/l

RT: 6.571 min Scan# 785

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion: 63 Resp: 3358


Ion Ratio Lower Upper

63 100

98 0.0 3.4 10.2#

100 0.0 2.0 5.8#





#25

2-Butanone

Concen: 4.658 ug/l

RT: 7.488 min Scan# 941

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion: 43 Resp: 6488

Ion Ratio Lower Upper

43 100

72 22.1 21.4 32.2

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC001

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#26

2,2-Dichloropropane

Concen: 0.886 ug/l

RT: 7.494 min Scan# 942

Delta R.T. 0.006 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion: 77 Resp: 2902

Ion Ratio Lower Upper

77 100

97 19.7 10.8 32.3









#27
cis-1,2-Dichloroethene
Concen: 0.987 ug/l
RT: 7.483 min Scan# 941
Delta R.T. -0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001

Tgt Ion	Ratio	Lower	Upper
96	100		
61	140.3	0.0	301.8
98	59.5	0.0	128.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#28
Bromochloromethane
Concen: 1.060 ug/l
RT: 7.824 min Scan# 998
Delta R.T. 0.012 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	3.6
130	61.3	58.0	87.0





#29

Tetrahydrofuran

Concen: 4.660 ug/l

RT: 7.841 min Scan# 10

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion:	42	Resp:	4006
Ion	Ratio	Lower	Upper
42	100		
72	40.5	34.6	51.8
71	33.4	32.2	48.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#30

Chloroform

Concen: 1.005 ug/l




RT: 7.971 min Scan# 1023


Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48


Tgt Ion:	83	Resp:	3793
Ion	Ratio	Lower	Upper
83	100		
85	62.8	51.0	76.4





#32
1,1,1-Trichloroethane
Concen: 0.921 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001



Tgt Ion: 97 Resp: 3120
Ion Ratio Lower Upper
97 100
99 0.0 51.2 76.8#
61 46.5 37.5 56.3

Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024




#33
1,2-Dichloroethane-d4
Concen: 0.000 ug/l
Expected RT: 8.58 min

Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 65
Sig Exp Ratio
65 100
67 51.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 12

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion:114 Resp: 290146

Ion Ratio Lower Upper

114 100

63 22.3 0.0 43.8





88 16.9 0.0 31.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#35

Dibromofluoromethane

Concen: 0.000 ug/l

Expected RT: 8.16 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48


Tgt Ion: 113

Sig Exp Ratio

113 100


111 104.1

192 16.9



Ion 110.80 (110.50 to 111.50): VN084218.D\data.ms

Ion 191.70 (191.40 to 192.40): VN084218.D\data.ms



#36

1,1-Dichloropropene

Concen: 0.871 ug/l

RT: 8.365 min Scan# 1091

Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 75 Resp: 2420

Ion Ratio Lower Upper

75 100

110 32.6 16.9 50.7





77 31.9 24.7 37.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#37

Ethyl Acetate

Concen: 0.911 ug/l

RT: 7.559 min Scan# 953

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion: 43 Resp: 2596


Ion Ratio Lower Upper

43 100

61 4.8 10.7 16.1#

70 2.4 8.8 13.2#





#38

Carbon Tetrachloride

Concen: 0.917 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. -0.000 min

Lab File: VN084218.D


Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :




VSTDIC001



Tgt Ion:	117	Resp:	2795
Ion	Ratio	Lower	Upper
117	100		
119	86.1	78.7	118.1
121	31.6	24.4	36.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#39

Methylcyclohexane


Concen: 0.804 ug/l

RT: 9.600 min Scan# 1300



Delta R.T. -0.000 min


Lab File: VN084218.D

Acq: 30 Sep 2024 14:48



Tgt Ion:	83	Resp:	2486
Ion	Ratio	Lower	Upper
83	100		
55	87.2	62.8	94.2
98	59.0	38.8	58.2





#40

Benzene

Concen: 0.952 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 78 Resp: 8245

Ion Ratio Lower Upper

78 100





77 21.1 19.1 28.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#41

Methacrylonitrile

Concen: 0.963 ug/l

RT: 7.782 min Scan# 991

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion: 41 Resp: 1490




Ion Ratio Lower Upper


41 100

39 21.5 43.3 64.9#

67 56.9 56.1 84.1

52 0.0 24.5 36.7#





#42

1,2-Dichloroethane

Concen: 0.909 ug/l

RT: 8.671 min Scan# 1142

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 62 Resp: 2669

Ion Ratio Lower Upper

62 100





98 4.9 0.0 18.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#43

Isopropyl Acetate

Concen: 1.364 ug/l m

RT: 8.688 min Scan# 1145

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion: 43 Resp: 7614


Ion Ratio Lower Upper

43 100

61 14.3 20.4 30.6#

87 7.3 10.3 15.5#





#44

Trichloroethene

Concen: 0.932 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion:130 Resp: 1885

Ion Ratio Lower Upper

130 100





95 87.9 0.0 203.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#45

1,2-Dichloropropane

Concen: 0.818 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. 0.006 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion: 63 Resp: 1676

Ion Ratio Lower Upper

63 100

65 0.0 24.6 37.0#





#46

Dibromomethane

Concen: 0.828 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion: 93 Resp: 1144

Ion Ratio Lower Upper

93 100

95 95.1 65.0 97.4





174 100.3 71.8 107.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#47

Bromodichloromethane

Concen: 0.897 ug/l

RT: 9.894 min Scan# 1350

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion: 83 Resp: 2754


Ion Ratio Lower Upper

83 100

85 68.7 51.5 77.3

127 0.0 5.7 8.5#





#48

Methyl methacrylate

Concen: 0.821 ug/l

RT: 9.676 min Scan# 1313

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC001

Tgt Ion:	41	Resp:	1824
Ion	Ratio	Lower	Upper
41	100		
69	86.5	67.1	100.7
39	66.4	45.8	68.6

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#49

1,4-Dioxane

Concen: 19.552 ug/l




RT: 9.694 min Scan# 1316


Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion:	88	Resp:	800
Ion	Ratio	Lower	Upper
88	100		
43	0.0	25.3	37.9
58	75.9	56.4	84.6





#50

Toluene-d8

Concen: 0.000 ug/l

Expected RT: 10.56 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion: 98

Sig Exp Ratio

98 100

100 65.9

Instrument :

MSVOA_N

ClientSampleId :



VSTDICC001

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#51

4-Methyl-2-Pentanone

Concen: 4.134 ug/l

RT: 10.447 min Scan# 1444

Delta R.T. 0.006 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion: 43 Resp: 11218

Ion Ratio Lower Upper

43 100

58 37.4 31.6 47.4





#52

Toluene

Concen: 0.923 ug/l

RT: 10.629 min Scan# 1475

Delta R.T. -0.000 min

Lab File: VN084218.D


Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001



Tgt Ion: 92 Resp: 4875

Ion Ratio Lower Upper

92 100




91 159.8 137.3 205.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#53

t-1,3-Dichloropropene

Concen: 0.797 ug/l

RT: 10.829 min Scan# 1509

Delta R.T. -0.006 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion: 75 Resp: 2493

Ion Ratio Lower Upper

75 100

77 32.7 25.4 38.0





#54

cis-1,3-Dichloropropene

Concen: 0.808 ug/l

RT: 10.306 min Scan# 14

Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 75 Resp: 2720

Ion Ratio Lower Upper

75 100

77 34.6 24.2 36.2





39 46.3 38.7 58.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#55

1,1,2-Trichloroethane

Concen: 0.882 ug/l

RT: 11.018 min Scan# 1541

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion: 97 Resp: 1669




Ion Ratio Lower Upper


97 100

83 75.7 71.0 106.4

85 57.4 41.7 62.5

99 68.1 50.1 75.1





#56

Ethyl methacrylate

Concen: 0.771 ug/l

RT: 10.876 min Scan# 1516

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC001

Tgt	Ion	69	Resp:	2402
Ion	Ratio	Lower	Upper	
69	100			
41	64.2	53.3	79.9	
39	0.0	28.8	43.2	

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#57

1,3-Dichloropropane

Concen: 0.914 ug/l




RT: 11.159 min Scan# 1565


Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt	Ion	76	Resp:	3092
Ion	Ratio	Lower	Upper	
76	100			
78	32.2	25.8	38.6	





#58

2-Chloroethyl Vinyl ether

Concen: 4.278 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 63 Resp: 6179

Ion Ratio Lower Upper

63 100





106 28.2 21.8 32.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#59

2-Hexanone

Concen: 4.084 ug/l

RT: 11.194 min Scan# 1571

Delta R.T. -0.000 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion: 43 Resp: 8282

Ion Ratio Lower Upper

43 100

58 51.4 27.3 81.8





#60

Dibromochloromethane

Concen: 0.857 ug/l

RT: 11.353 min Scan# 1599

Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion:129 Resp: 1916

Ion Ratio Lower Upper

129 100





127 69.2 38.3 114.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#61

1,2-Dibromoethane

Concen: 0.913 ug/l

RT: 11.465 min Scan# 1617

Delta R.T. -0.000 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion:107 Resp: 1795

Ion Ratio Lower Upper

107 100

109 94.9 75.4 113.0





#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
Expected RT: 12.85 min

Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion	95
Sig	Exp Ratio
95	100
174	72.6
176	70.0

Instrument :

MSVOA_N



ClientSampleId :

VSTDICC001

Manual Integrations




APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion	117	Resp	244376
Ion Ratio	Lower	Upper	
117	100		
82	60.8	47.2	70.8
119	34.0	25.4	38.0





#64

Tetrachloroethene

Concen: 0.993 ug/l

RT: 11.106 min Scan# 1555

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion:164 Resp: 1690

Ion Ratio Lower Upper

164 100

166 92.5 101.0 151.4#

129 103.5 80.8 121.2





131 95.1 78.1 117.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#65

Chlorobenzene

Concen: 0.927 ug/l

RT: 11.888 min Scan# 1689

Delta R.T. -0.000 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion:112 Resp: 5025

Ion Ratio Lower Upper

112 100

114 41.6 27.0 40.6#





#66

1,1,1,2-Tetrachloroethane

Concen: 0.874 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:	131	Resp:	1632
Ion Ratio	Lower	Upper	
131	100		
133	96.9	47.9	143.7
119	0.0	32.8	98.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#67

Ethyl Benzene

Concen: 0.896 ug/l




RT: 11.965 min Scan# 1702


Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion:	91	Resp:	8582
Ion Ratio	Lower	Upper	
91	100		
106	24.3	24.5	36.7





#68

m/p-Xylenes

Concen: 1.656 ug/l

RT: 12.070 min Scan# 17

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion:106 Resp: 5869

Ion Ratio Lower Upper

106 100





91 201.7 167.1 250.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#69

o-Xylene

Concen: 0.715 ug/l

RT: 12.400 min Scan# 1776

Delta R.T. 0.006 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion:106 Resp: 2398

Ion Ratio Lower Upper

106 100

91 248.5 110.2 330.6





#70

Styrene

Concen: 0.790 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC001

Tgt Ion:	104	Resp:	4489
Ion Ratio	Lower	Upper	
104	100		
78	47.1	42.7	64.1
103	52.2	43.6	65.4

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#71

Bromoform

Concen: 0.850 ug/l




RT: 12.576 min Scan# 1806


Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion:	173	Resp:	1168
Ion Ratio	Lower	Upper	
173	100		
175	43.1	23.5	70.6
254	0.0	0.0	0.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2012

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion:152 Resp: 101963

Ion Ratio Lower Upper

152 100

115 62.6 31.3 93.9





150 159.0 0.0 349.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#73

Isopropylbenzene

Concen: 0.899 ug/l

RT: 12.688 min Scan# 1825

Delta R.T. -0.006 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion:105 Resp: 6991

Ion Ratio Lower Upper

105 100

120 24.1 12.9 38.7





#74

N-amy1 acetate

Concen: 0.916 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:	43	Resp:	3228
Ion	Ratio	Lower	Upper
43	100		
70	35.2	34.4	51.6
55	21.8	19.6	29.4
61	14.8	19.6	29.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#75

1,1,2,2-Tetrachloroethane

Concen: 0.954 ug/l




RT: 12.941 min Scan# 1868


Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion:	83	Resp:	2233
Ion	Ratio	Lower	Upper
83	100		
131	0.0	4.9	14.7
85	65.3	31.9	95.7





#76

1,2,3-Trichloropropane

Concen: 0.899 ug/l m

RT: 12.994 min Scan# 18

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 75 Resp: 1826

Ion Ratio Lower Upper

75 100





77 231.3 96.9 290.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#77

Bromobenzene

Concen: 0.969 ug/l

RT: 12.976 min Scan# 1874

Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion:156 Resp: 1818


Ion Ratio Lower Upper

156 100

77 232.3 115.1 345.3

158 92.7 48.9 146.8





#78

n-propylbenzene

Concen: 0.781 ug/l

RT: 13.035 min Scan# 18

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 91 Resp: 7045

Ion Ratio Lower Upper

91 100





120 24.2 10.8 32.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#79

2-Chlorotoluene

Concen: 0.936 ug/l

RT: 13.129 min Scan# 1900

Delta R.T. 0.006 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion: 91 Resp: 5383

Ion Ratio Lower Upper

91 100

126 27.2 15.7 47.1





#80

1,3,5-Trimethylbenzene

Concen: 0.761 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 105 Resp: 4830

Ion Ratio Lower Upper

105 100





120 50.6 23.7 71.1

Manual Integrations

APPROVED

Reviewed By : John Carlone 10/01/2024

Supervised By : Mahesh Dadoda 10/01/2024



#82

4-Chlorotoluene

Concen: 0.893 ug/l

RT: 13.223 min Scan# 1916

Delta R.T. 0.006 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion: 91 Resp: 5174

Ion Ratio Lower Upper

91 100

126 31.3 15.7 47.1





#83

tert-Butylbenzene

Concen: 0.870 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N





ClientSampleId :

VSTDIC001

Tgt Ion:	119	Resp:	4909
Ion	Ratio	Lower	Upper
119	100		
91	59.0	34.1	102.3
134	18.3	12.8	38.4

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#84

1,2,4-Trimethylbenzene

Concen: 0.744 ug/l




RT: 13.482 min Scan# 1960


Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion:	105	Resp:	4760
Ion	Ratio	Lower	Upper
105	100		
120	44.7	22.3	66.8





#85

sec-Butylbenzene

Concen: 0.770 ug/l

RT: 13.612 min Scan# 19

Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001

Tgt Ion:105 Resp: 5795

Ion Ratio Lower Upper

105 100





134 20.1 9.8 29.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#86

p-Isopropyltoluene

Concen: 0.722 ug/l

RT: 13.723 min Scan# 2001

Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion:119 Resp: 4495


Ion Ratio Lower Upper

119 100

134 24.2 12.6 37.8

91 23.4 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 0.972 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC001

Tgt Ion:	146	Resp:	3424
Ion	Ratio	Lower	Upper
146	100		
111	45.5	21.8	65.3
148	67.0	31.9	95.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#88

1,4-Dichlorobenzene

Concen: 1.026 ug/l m




RT: 13.812 min Scan# 2016


Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion:	146	Resp:	3648
Ion	Ratio	Lower	Upper
146	100		
111	42.7	21.2	63.6
148	62.9	31.9	95.9





#89

n-Butylbenzene

Concen: 0.862 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N





ClientSampleId :

VSTDICC001

Tgt Ion:	91	Resp:	4866
Ion	Ratio	Lower	Upper
91	100		
92	52.1	26.6	79.7
134	21.4	12.3	36.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#90

Hexachloroethane

Concen: 0.911 ug/l




RT: 14.335 min Scan# 2105


Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion:	117	Resp:	1151
Ion	Ratio	Lower	Upper
117	100		
201	54.3	34.4	103.2





#91

1,2-Dichlorobenzene

Concen: 1.046 ug/l

RT: 14.100 min Scan# 2066

Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:	146	Resp:	3610
Ion Ratio	Lower	Upper	
146	100		
111	39.9	22.0	66.0
148	57.1	32.1	96.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#92

1,2-Dibromo-3-Chloropropane

Concen: 0.876 ug/l




RT: 14.711 min Scan# 2169


Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion:	75	Resp:	423
Ion Ratio	Lower	Upper	
75	100		
155	51.3	38.1	114.5
157	59.1	49.6	149.0





#93

1,2,4-Trichlorobenzene

Concen: 0.788 ug/l

RT: 15.400 min Scan# 2284

Delta R.T. 0.012 min

Lab File: VN084218.D


Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001






Tgt Ion:	180	Resp:	1340
Ion	Ratio	Lower	Upper
180	100		
182	103.0	47.0	141.0
145	41.3	16.4	49.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#94

Hexachlorobutadiene


Concen: 1.015 ug/l

RT: 15.494 min Scan# 2302



Delta R.T. -0.006 min


Lab File: VN084218.D

Acq: 30 Sep 2024 14:48



Tgt Ion:	225	Resp:	861
Ion	Ratio	Lower	Upper
225	100		
223	52.7	32.4	97.2
227	60.7	32.4	97.0





#95

Naphthalene

Concen: 0.892 ug/l

RT: 15.641 min Scan# 23

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion:	128	Resp:	5031
Ion Ratio	Lower	Upper	
128	100		
127	11.5	10.5	15.7
129	11.3	8.7	13.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#96

1,2,3-Trichlorobenzene

Concen: 0.899 ug/l




RT: 15.847 min Scan# 2362

Delta R.T. 0.012 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion:	180	Resp:	1531
Ion Ratio	Lower	Upper	
180	100		
182	101.0	48.2	144.6
145	37.5	17.3	51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084220.D
 Acq On : 30 Sep 2024 15:36
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN093024

Quant Time: Oct 01 07:14:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 10/01/2024
 Supervised By : Mahesh Dadoda 10/01/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	174241	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	300916	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	263364	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	130851	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.577	65	130630	50.564	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.120%
35) Dibromofluoromethane	8.165	113	100200	50.508	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.020%
50) Toluene-d8	10.565	98	387775	53.129	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.260%
62) 4-Bromofluorobenzene	12.847	95	140850	52.971	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.940%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.124	85	108641	52.712	ug/l	95
3) Chloromethane	2.359	50	121140	51.477	ug/l	97
4) Vinyl Chloride	2.512	62	116792	51.235	ug/l	99
5) Bromomethane	2.947	94	71654	47.656	ug/l	99
6) Chloroethane	3.118	64	73658	45.119	ug/l	94
7) Trichlorofluoromethane	3.495	101	188482	53.099	ug/l	98
8) Diethyl Ether	3.959	74	70360	53.445	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.371	101	104895	51.573	ug/l	99
10) Methyl Iodide	4.589	142	141430	53.996	ug/l	92
11) Tert butyl alcohol	5.524	59	109545	257.154	ug/l	99
12) 1,1-Dichloroethene	4.342	96	104103	53.064	ug/l	95
13) Acrolein	4.177	56	125971	257.587	ug/l	97
14) Allyl chloride	5.024	41	174077	51.092	ug/l	100
15) Acrylonitrile	5.718	53	293962	273.264	ug/l	99
16) Acetone	4.430	43	308020	277.547	ug/l	97
17) Carbon Disulfide	4.712	76	302450	48.693	ug/l	100
18) Methyl Acetate	5.024	43	130494	53.948	ug/l	98
19) Methyl tert-butyl Ether	5.800	73	360019	54.371	ug/l	96
20) Methylene Chloride	5.277	84	117572	52.116	ug/l	89
21) trans-1,2-Dichloroethene	5.783	96	108894	52.980	ug/l	97
22) Diisopropyl ether	6.671	45	384647	54.579	ug/l	99
23) Vinyl Acetate	6.600	43	1492704	285.569	ug/l	99
24) 1,1-Dichloroethane	6.565	63	211544	53.676	ug/l	98
25) 2-Butanone	7.482	43	426860	282.397	ug/l	98
26) 2,2-Dichloropropane	7.488	77	189858	53.428	ug/l	99
27) cis-1,2-Dichloroethene	7.488	96	128843	51.876	ug/l	98
28) Bromochloromethane	7.818	49	84632	48.129	ug/l	97
29) Tetrahydrofuran	7.835	42	262996	281.940	ug/l	99
30) Chloroform	7.965	83	213634	52.184	ug/l	95
31) Cyclohexane	8.259	56	190602	49.809	ug/l	100
32) 1,1,1-Trichloroethane	8.171	97	196319	53.395	ug/l	99
36) 1,1-Dichloropropene	8.371	75	155365	53.917	ug/l	99
37) Ethyl Acetate	7.559	43	159764	54.043	ug/l	99
38) Carbon Tetrachloride	8.359	117	169403	53.593	ug/l	94
39) Methylcyclohexane	9.600	83	181676	56.648	ug/l	98
40) Benzene	8.606	78	478538	53.301	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084220.D
 Acq On : 30 Sep 2024 15:36
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN093024

Manual Integrations
 APPROVED

Reviewed By :John Carlone 10/01/2024
 Supervised By :Mahesh Dadoda 10/01/2024

Quant Time: Oct 01 07:14:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	89407	55.708	ug/l	98
42) 1,2-Dichloroethane	8.671	62	164758	54.131	ug/l	99
43) Isopropyl Acetate	8.688	43	278668	48.126	ug/l	100
44) Trichloroethene	9.353	130	109810	52.368	ug/l	100
45) 1,2-Dichloropropane	9.618	63	116161	54.694	ug/l	99
46) Dibromomethane	9.706	93	80284	56.015	ug/l	100
47) Bromodichloromethane	9.888	83	173015	54.335	ug/l	98
48) Methyl methacrylate	9.682	41	132287	57.391	ug/l	98
49) 1,4-Dioxane	9.694	88	47169	1111.545	ug/l	99
51) 4-Methyl-2-Pentanone	10.441	43	808630	287.336	ug/l	100
52) Toluene	10.629	92	296651	54.185	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	183153	56.426	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	193475	55.416	ug/l	98
55) 1,1,2-Trichloroethane	11.018	97	108550	55.307	ug/l	98
56) Ethyl methacrylate	10.876	69	189121	58.561	ug/l	100
57) 1,3-Dichloropropane	11.165	76	191966	54.738	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	415454	277.315	ug/l	98
59) 2-Hexanone	11.194	43	621000	295.238	ug/l	100
60) Dibromochloromethane	11.359	129	127786	55.105	ug/l	99
61) 1,2-Dibromoethane	11.470	107	111244	54.548	ug/l	98
64) Tetrachloroethene	11.100	164	96460	52.585	ug/l	99
65) Chlorobenzene	11.888	112	312728	53.555	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.959	131	111388	55.381	ug/l	97
67) Ethyl Benzene	11.965	91	573790	55.563	ug/l	99
68) m/p-Xylenes	12.070	106	434515	113.748	ug/l	99
69) o-Xylene	12.400	106	209737	58.032	ug/l	99
70) Styrene	12.412	104	355892	58.136	ug/l	99
71) Bromoform	12.576	173	84787	57.258	ug/l #	99
73) Isopropylbenzene	12.694	105	535240	53.604	ug/l	100
74) N-amyl acetate	12.494	43	249223	55.118	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	157322	52.398	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	134836m	50.475	ug/l	
77) Bromobenzene	12.976	156	124463	51.676	ug/l	98
78) n-propylbenzene	13.035	91	638464	55.188	ug/l	99
79) 2-Chlorotoluene	13.123	91	390502	52.898	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	456943	56.101	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	58987	52.939	ug/l	99
82) 4-Chlorotoluene	13.217	91	397246	53.455	ug/l	100
83) tert-Butylbenzene	13.435	119	390221	53.903	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	461123	56.199	ug/l	98
85) sec-Butylbenzene	13.617	105	536685	55.556	ug/l	99
86) p-Isopropyltoluene	13.729	119	452245	56.614	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	234190	51.830	ug/l	99
88) 1,4-Dichlorobenzene	13.811	146	234070	51.298	ug/l	99
89) n-Butylbenzene	14.053	91	396001	54.671	ug/l	100
90) Hexachloroethane	14.329	117	82363	50.795	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	226320	51.082	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	32296	52.125	ug/l	96
93) 1,2,4-Trichlorobenzene	15.394	180	114157	52.311	ug/l	99
94) Hexachlorobutadiene	15.500	225	50461	46.368	ug/l	97
95) Naphthalene	15.641	128	380549	52.602	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	111086	50.853	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084220.D
Acq On : 30 Sep 2024 15:36
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICV\VN093024

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024

Quant Time: Oct 01 07:14:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024

Manual Integrations APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024



Tgt Ion: 168 Resp: 174241
Ion Ratio Lower Upper
168 100
99 66.5 54.2 81.2




Manual Integrations
APPROVED


Reviewed By : John Carlone 10/01/2024
Supervised By : Mahesh Dadoda 10/01/2024



#2
Dichlorodifluoromethane
Concen: 52.712 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 85 Resp: 108641
Ion Ratio Lower Upper
85 100
87 32.5 17.6 52.9





#3

Chloromethane

Concen: 51.477 ug/l

RT: 2.359 min Scan# 69

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 50 Resp: 121140

Ion Ratio Lower Upper

50 100





52 31.9 26.8 40.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#4

Vinyl Chloride

Concen: 51.235 ug/l

RT: 2.512 min Scan# 95

Delta R.T. -0.006 min

Lab File: VN084220.D




Acq: 30 Sep 2024 15:36


Tgt Ion: 62 Resp: 116792

Ion Ratio Lower Upper

62 100

64 33.3 26.2 39.4





#5

Bromomethane

Concen: 47.656 ug/l

RT: 2.947 min Scan# 16

Delta R.T. -0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 94 Resp: 71654

Ion Ratio Lower Upper

94 100





96 91.7 72.9 109.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#6

Chloroethane

Concen: 45.119 ug/l

RT: 3.118 min Scan# 198

Delta R.T. -0.000 min

Lab File: VN084220.D




Acq: 30 Sep 2024 15:36


Tgt Ion: 64 Resp: 73658

Ion Ratio Lower Upper

64 100

66 32.7 23.5 35.3





#7

Trichlorofluoromethane

Concen: 53.099 ug/l

RT: 3.495 min Scan# 262

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion:101 Resp: 188482

Ion Ratio Lower Upper

101 100





103 63.4 51.8 77.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#8

Diethyl Ether

Concen: 53.445 ug/l

RT: 3.959 min Scan# 341

Delta R.T. -0.000 min

Lab File: VN084220.D




Acq: 30 Sep 2024 15:36


Tgt Ion: 74 Resp: 70360

Ion Ratio Lower Upper

74 100

45 97.0 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 51.573 ug/l
RT: 4.371 min Scan# 411
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024

Tgt Ion:101 Resp: 104895
Ion Ratio Lower Upper
101 100
85 46.0 36.2 54.4
151 74.0 59.3 88.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#10
Methyl Iodide
Concen: 53.996 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.006 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion:142 Resp: 141430
Ion Ratio Lower Upper
142 100
127 50.3 35.3 52.9
141 14.7 11.0 16.6





#11

Tert butyl alcohol

Concen: 257.154 ug/l

RT: 5.524 min Scan# 607

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 59 Resp: 109545

Ion Ratio Lower Upper

59 100





57 10.6 8.2 12.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#12

1,1-Dichloroethene

Concen: 53.064 ug/l

RT: 4.342 min Scan# 406

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion: 96 Resp: 104103


Ion Ratio Lower Upper

96 100

61 166.3 128.6 192.8

98 60.4 51.7 77.5





#13

Acrolein

Concen: 257.587 ug/l

RT: 4.177 min Scan# 378

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 56 Resp: 125971

Ion Ratio Lower Upper

56 100





55 69.4 57.2 85.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#14

Allyl chloride

Concen: 51.092 ug/l

RT: 5.024 min Scan# 522

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion: 41 Resp: 174077


Ion Ratio Lower Upper

41 100

39 73.7 59.0 88.6

76 37.1 29.8 44.6





#15

Acrylonitrile

Concen: 273.264 ug/l

RT: 5.718 min Scan# 640

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024





Tgt Ion:	53	Resp:	293962
Ion	Ratio	Lower	Upper
53	100		
52	81.3	66.2	99.4
51	35.7	28.8	43.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#16

Acetone

Concen: 277.547 ug/l




RT: 4.430 min Scan# 421


Delta R.T. 0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Tgt Ion:	43	Resp:	308020
Ion	Ratio	Lower	Upper
43	100		
58	32.2	24.4	36.6





#17

Carbon Disulfide

Concen: 48.693 ug/l

RT: 4.712 min Scan# 469

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 76 Resp: 302450

Ion Ratio Lower Upper

76 100





78 9.1 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#18

Methyl Acetate

Concen: 53.948 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN084220.D




Acq: 30 Sep 2024 15:36


Tgt Ion: 43 Resp: 130494

Ion Ratio Lower Upper

43 100

74 23.0 19.2 28.8





#19

Methyl tert-butyl Ether

Concen: 54.371 ug/l

RT: 5.800 min Scan# 65

Delta R.T. 0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 73 Resp: 360019

Ion Ratio Lower Upper

73 100





57 20.3 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#20

Methylene Chloride

Concen: 52.116 ug/l

RT: 5.277 min Scan# 565

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Tgt Ion: 84 Resp: 117572




Ion Ratio Lower Upper


84 100

49 120.2 107.2 160.8

51 36.9 31.8 47.8

86 56.5 52.9 79.3





#21

trans-1,2-Dichloroethene

Concen: 52.980 ug/l

RT: 5.783 min Scan# 65

Delta R.T. -0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 96 Resp: 108894

Ion Ratio Lower Upper

96 100

61 139.8 108.4 162.6





98 62.9 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#22

Diisopropyl ether

Concen: 54.579 ug/l

RT: 6.671 min Scan# 802

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Tgt Ion: 45 Resp: 384647




Ion Ratio Lower Upper


45 100

43 53.5 42.3 63.5

87 27.0 21.8 32.6

59 11.3 9.6 14.4





#23

Vinyl Acetate

Concen: 285.569 ug/l

RT: 6.600 min Scan# 790

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 43 Resp: 1492704

Ion Ratio Lower Upper

43 100





86 10.5 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#24

1,1-Dichloroethane

Concen: 53.676 ug/l

RT: 6.565 min Scan# 784

Delta R.T. -0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion: 63 Resp: 211544


Ion Ratio Lower Upper

63 100

98 6.4 3.4 10.2

100 4.9 2.0 5.8





#25

2-Butanone

Concen: 282.397 ug/l

RT: 7.482 min Scan# 940

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Tgt Ion: 43 Resp: 426860

Ion Ratio Lower Upper

43 100

72 25.9 21.4 32.2

Instrument :

MSVOA_N

ClientSampleId :





ICVVN093024

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#26

2,2-Dichloropropane

Concen: 53.428 ug/l

RT: 7.488 min Scan# 941

Delta R.T. -0.000 min

Lab File: VN084220.D




Acq: 30 Sep 2024 15:36


Tgt Ion: 77 Resp: 189858

Ion Ratio Lower Upper

77 100

97 22.1 10.8 32.3









#27
cis-1,2-Dichloroethene
Concen: 51.876 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024

Tgt Ion:	96	Resp:	128843
Ion	Ratio	Lower	Upper
96	100		
61	154.8	0.0	301.8
98	64.2	0.0	128.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#28
Bromochloromethane
Concen: 48.129 ug/l
RT: 7.818 min Scan# 997
Delta R.T. 0.006 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion:	49	Resp:	84632
Ion	Ratio	Lower	Upper
49	100		
129	1.6	0.0	3.6
130	70.0	58.0	87.0





#29

Tetrahydrofuran

Concen: 281.940 ug/l

RT: 7.835 min Scan# 1000

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 42 Resp: 262996

Ion Ratio Lower Upper

42 100

72 42.8 34.6 51.8





71 40.0 32.2 48.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#30

Chloroform

Concen: 52.184 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. -0.000 min

Lab File: VN084220.D




Acq: 30 Sep 2024 15:36


Tgt Ion: 83 Resp: 213634

Ion Ratio Lower Upper

83 100

85 67.5 51.0 76.4





#31

Cyclohexane

Concen: 49.809 ug/l

RT: 8.259 min Scan# 1071

Delta R.T. 0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 56 Resp: 190602

Ion Ratio Lower Upper

56 100

69 30.1 23.8 35.6





84 85.8 68.8 103.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#32

1,1,1-Trichloroethane

Concen: 53.395 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion: 97 Resp: 196319


Ion Ratio Lower Upper

97 100

99 64.5 51.2 76.8

61 47.6 37.5 56.3









#33
1,2-Dichloroethane-d4
Concen: 50.564 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024

Tgt Ion: 65 Resp: 130630
Ion Ratio Lower Upper
65 100
67 53.0 0.0 102.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion:114 Resp: 300916
Ion Ratio Lower Upper
114 100
63 20.6 0.0 43.8
88 17.0 0.0 31.6





#35

Dibromofluoromethane

Concen: 50.508 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICNVN093024

Tgt Ion:113 Resp: 100200

Ion Ratio Lower Upper

113 100

111 102.6 83.3 124.9





192 17.3 13.5 20.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#36

1,1-Dichloropropene

Concen: 53.917 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion: 75 Resp: 155365


Ion Ratio Lower Upper

75 100

110 33.2 16.9 50.7

77 30.7 24.7 37.1





#37

Ethyl Acetate

Concen: 54.043 ug/l

RT: 7.559 min Scan# 953

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N





ClientSampleId :

ICVVN093024

Tgt Ion:	43	Resp:	159764
Ion	Ratio	Lower	Upper
43	100		
61	13.5	10.7	16.1
70	10.0	8.8	13.2

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#38

Carbon Tetrachloride

Concen: 53.593 ug/l




RT: 8.359 min Scan# 1089


Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Tgt Ion:	117	Resp:	169403
Ion	Ratio	Lower	Upper
117	100		
119	90.5	78.7	118.1
121	29.9	24.4	36.6





#39

Methylcyclohexane

Concen: 56.648 ug/l

RT: 9.600 min Scan# 13

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 83 Resp: 181676

Ion Ratio Lower Upper

83 100

55 76.1 62.8 94.2





98 48.6 38.8 58.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#40

Benzene

Concen: 53.301 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN084220.D




Acq: 30 Sep 2024 15:36


Tgt Ion: 78 Resp: 478538

Ion Ratio Lower Upper

78 100

77 23.3 19.1 28.7





#41

Methacrylonitrile

Concen: 55.708 ug/l

RT: 7.777 min Scan# 990

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N





ClientSampleId :

ICVVN093024

Tgt Ion:	41	Resp:	89407
Ion	Ratio	Lower	Upper
41	100		
39	56.2	43.3	64.9
67	70.8	56.1	84.1
52	30.2	24.5	36.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#42

1,2-Dichloroethane

Concen: 54.131 ug/l




RT: 8.671 min Scan# 1142


Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Tgt Ion:	62	Resp:	164758
Ion	Ratio	Lower	Upper
62	100		
98	8.6	0.0	18.2





#43

Isopropyl Acetate

Concen: 48.126 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 43 Resp: 278668

Ion Ratio Lower Upper

43 100

61 25.8 20.4 30.6





87 12.9 10.3 15.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#44

Trichloroethene

Concen: 52.368 ug/l

RT: 9.353 min Scan# 1258

Delta R.T. 0.006 min

Lab File: VN084220.D




Acq: 30 Sep 2024 15:36


Tgt Ion:130 Resp: 109810

Ion Ratio Lower Upper

130 100

95 101.8 0.0 203.0





#45

1,2-Dichloropropane

Concen: 54.694 ug/l

RT: 9.618 min Scan# 13

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 63 Resp: 116161

Ion Ratio Lower Upper

63 100





65 30.4 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#46

Dibromomethane

Concen: 56.015 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion: 93 Resp: 80284


Ion Ratio Lower Upper

93 100

95 81.5 65.0 97.4

174 89.6 71.8 107.8





#47

Bromodichloromethane

Concen: 54.335 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N





ClientSampleId :

ICVVN093024

Tgt Ion:	83	Resp:	173015
Ion	Ratio	Lower	Upper
83	100		
85	62.5	51.5	77.3
127	7.9	5.7	8.5

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#48

Methyl methacrylate

Concen: 57.391 ug/l




RT: 9.682 min Scan# 1314


Delta R.T. 0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Tgt Ion:	41	Resp:	132287
Ion	Ratio	Lower	Upper
41	100		
69	81.0	67.1	100.7
39	56.9	45.8	68.6





#49

1,4-Dioxane

Concen: 1111.545 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. -0.000 min

Lab File: VN084220.D


Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024



Tgt Ion: 88 Resp: 47169

Ion Ratio Lower Upper

88 100

43 31.2 25.3 37.9




58 71.7 56.4 84.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#50

Toluene-d8


Concen: 53.129 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36






Tgt Ion: 98 Resp: 387775

Ion Ratio Lower Upper

98 100

100 65.1 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 287.336 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024

Tgt Ion: 43 Resp: 808630
Ion Ratio Lower Upper
43 100
58 39.5 31.6 47.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#52
Toluene
Concen: 54.185 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 92 Resp: 296651
Ion Ratio Lower Upper
92 100
91 173.0 137.3 205.9





#53

t-1,3-Dichloropropene

Concen: 56.426 ug/l

RT: 10.835 min Scan# 1510

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 75 Resp: 183153

Ion Ratio Lower Upper

75 100





77 32.1 25.4 38.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#54

cis-1,3-Dichloropropene

Concen: 55.416 ug/l

RT: 10.312 min Scan# 1421

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion: 75 Resp: 193475


Ion Ratio Lower Upper

75 100

77 31.6 24.2 36.2

39 49.4 38.7 58.1





#55

1,1,2-Trichloroethane

Concen: 55.307 ug/l

RT: 11.018 min Scan# 1515

Delta R.T. 0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 97 Resp: 108550

Ion Ratio Lower Upper

97 100

83 85.0 71.0 106.4

85 53.6 41.7 62.5





99 62.4 50.1 75.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#56

Ethyl methacrylate

Concen: 58.561 ug/l

RT: 10.876 min Scan# 1517

Delta R.T. 0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion: 69 Resp: 189121


Ion Ratio Lower Upper

69 100

41 66.4 53.3 79.9

39 35.9 28.8 43.2





#57

1,3-Dichloropropane

Concen: 54.738 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 76 Resp: 191966

Ion Ratio Lower Upper

76 100





78 32.0 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#58

2-Chloroethyl Vinyl ether

Concen: 277.315 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN084220.D




Acq: 30 Sep 2024 15:36


Tgt Ion: 63 Resp: 415454

Ion Ratio Lower Upper

63 100

106 28.0 21.8 32.6





#59

2-Hexanone

Concen: 295.238 ug/l

RT: 11.194 min Scan# 1571

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 43 Resp: 621000

Ion Ratio Lower Upper

43 100





58 54.4 27.3 81.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#60

Dibromochloromethane

Concen: 55.105 ug/l

RT: 11.359 min Scan# 1599

Delta R.T. -0.000 min

Lab File: VN084220.D




Acq: 30 Sep 2024 15:36


Tgt Ion:129 Resp: 127786

Ion Ratio Lower Upper

129 100

127 77.9 38.3 114.9









#61
1,2-Dibromoethane
Concen: 54.548 ug/l
RT: 11.470 min Scan# 1617
Delta R.T. 0.006 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024

Tgt Ion: 107 Resp: 111244
Ion Ratio Lower Upper
107 100
109 92.5 75.4 113.0




Manual Integrations
APPROVED


Reviewed By : John Carlone 10/01/2024
Supervised By : Mahesh Dadoda 10/01/2024



#62
4-Bromofluorobenzene
Concen: 52.971 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36


Tgt Ion: 95 Resp: 140850
Ion Ratio Lower Upper
95 100
174 70.8 0.0 145.2
176 69.2 0.0 140.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024



Tgt Ion:117 Resp: 263364
Ion Ratio Lower Upper
117 100
82 61.2 47.2 70.8
119 32.9 25.4 38.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#64
Tetrachloroethene
Concen: 52.585 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion:164 Resp: 96460
Ion Ratio Lower Upper
164 100
166 126.7 101.0 151.4
129 102.6 80.8 121.2
131 98.1 78.1 117.1





#65

Chlorobenzene

Concen: 53.555 ug/l

RT: 11.888 min Scan# 1689

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion:112 Resp: 312728

Ion Ratio Lower Upper

112 100





114 32.0 27.0 40.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#66

1,1,1,2-Tetrachloroethane

Concen: 55.381 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion:131 Resp: 111388


Ion Ratio Lower Upper

131 100

133 93.4 47.9 143.7

119 62.9 32.8 98.3





#67

Ethyl Benzene

Concen: 55.563 ug/l

RT: 11.965 min Scan# 1701

Delta R.T. 0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 91 Resp: 573796

Ion Ratio Lower Upper

91 100





106 30.3 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#68

m/p-Xylenes

Concen: 113.748 ug/l

RT: 12.070 min Scan# 1720

Delta R.T. -0.000 min

Lab File: VN084220.D




Acq: 30 Sep 2024 15:36


Tgt Ion:106 Resp: 434515

Ion Ratio Lower Upper

106 100

91 208.1 167.1 250.7





#69

o-Xylene

Concen: 58.032 ug/l

RT: 12.400 min Scan# 1775

Delta R.T. 0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion:106 Resp: 209737

Ion Ratio Lower Upper

106 100





91 219.4 110.2 330.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#70

Styrene

Concen: 58.136 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion:104 Resp: 355892


Ion Ratio Lower Upper

104 100

78 52.7 42.7 64.1

103 55.1 43.6 65.4





#71

Bromoform

Concen: 57.258 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N





ClientSampleId :

ICVVN093024

Tgt Ion:	173	Resp:	84787
Ion	Ratio	Lower	Upper
173	100		
175	48.0	23.5	70.6
254	0.0	0.0	0.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.788 min Scan# 2012


Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Tgt Ion:	152	Resp:	130851
Ion	Ratio	Lower	Upper
152	100		
115	62.2	31.3	93.9
150	174.1	0.0	349.8





#73

Isopropylbenzene

Concen: 53.604 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion:105 Resp: 535240

Ion Ratio Lower Upper

105 100





120 25.9 12.9 38.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#74

N-ethyl acetate

Concen: 55.118 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Tgt Ion: 43 Resp: 249223




Ion Ratio Lower Upper


43 100

70 42.5 34.4 51.6

55 24.2 19.6 29.4

61 23.8 19.6 29.4





#75

1,1,2,2-Tetrachloroethane

Concen: 52.398 ug/l

RT: 12.935 min Scan# 18

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :





ICVVN093024

Tgt Ion:	83	Resp:	157322
Ion	Ratio	Lower	Upper
83	100		
131	10.0	4.9	14.7
85	63.6	31.9	95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#76

1,2,3-Trichloropropane

Concen: 50.475 ug/l m




RT: 12.994 min Scan# 1877


Delta R.T. 0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Tgt Ion:	75	Resp:	134836
Ion	Ratio	Lower	Upper
75	100		
77	208.3	96.9	290.7





#77

Bromobenzene

Concen: 51.676 ug/l

RT: 12.976 min Scan# 1875

Delta R.T. -0.006 min

Lab File: VN084220.D


Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :




ICVVN093024



Tgt Ion:	156	Resp:	124463
Ion	Ratio	Lower	Upper
156	100		
77	225.7	115.1	345.3
158	96.8	48.9	146.8

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#78

n-propylbenzene

Concen: 55.188 ug/l




RT: 13.035 min Scan# 1884


Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Tgt Ion:	91	Resp:	638464
Ion	Ratio	Lower	Upper
91	100		
120	21.9	10.8	32.4





#79

2-Chlorotoluene

Concen: 52.898 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 91 Resp: 390502

Ion Ratio Lower Upper

91 100





126 31.2 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#80

1,3,5-Trimethylbenzene

Concen: 56.101 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN084220.D




Acq: 30 Sep 2024 15:36


Tgt Ion:105 Resp: 456943

Ion Ratio Lower Upper

105 100

120 47.0 23.7 71.1





#81

trans-1,4-Dichloro-2-butene

Concen: 52.939 ug/l

RT: 12.735 min Scan# 1833

Delta R.T. -0.000 min

Lab File: VN084220.D


Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024



Tgt Ion: 75 Resp: 58987

Ion Ratio Lower Upper

75 100

53 98.1 78.0 117.0




89 45.7 35.5 53.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#82

4-Chlorotoluene

Concen: 53.455 ug/l

RT: 13.217 min Scan# 1915

Delta R.T. -0.000 min

Lab File: VN084220.D




Acq: 30 Sep 2024 15:36


Tgt Ion: 91 Resp: 397246

Ion Ratio Lower Upper

91 100

126 31.3 15.7 47.1





#83

tert-Butylbenzene

Concen: 53.903 ug/l

RT: 13.435 min Scan# 19

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N





ClientSampleId :

ICVVN093024

Tgt Ion:	119	Resp:	390221
Ion	Ratio	Lower	Upper
119	100		
91	68.6	34.1	102.3
134	23.9	12.8	38.4

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#84

1,2,4-Trimethylbenzene

Concen: 56.199 ug/l




RT: 13.482 min Scan# 1960


Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Tgt Ion:	105	Resp:	461123
Ion	Ratio	Lower	Upper
105	100		
120	43.4	22.3	66.8





#85

sec-Butylbenzene

Concen: 55.556 ug/l

RT: 13.617 min Scan# 1983

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion:105 Resp: 536685

Ion Ratio Lower Upper

105 100





134 19.2 9.8 29.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024

Supervised By :Mahesh Dadoda 10/01/2024



#86

p-Isopropyltoluene

Concen: 56.614 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion:119 Resp: 452245


Ion Ratio Lower Upper

119 100

134 25.3 12.6 37.8

91 24.9 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 51.830 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN084220.D


Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :




ICVVN093024



Tgt Ion:	146	Resp:	234190
Ion	Ratio	Lower	Upper
146	100		
111	42.7	21.8	65.3
148	63.9	31.9	95.7

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#88

1,4-Dichlorobenzene


Concen: 51.298 ug/l

RT: 13.811 min Scan# 2016



Delta R.T. -0.000 min


Lab File: VN084220.D

Acq: 30 Sep 2024 15:36



Tgt Ion:	146	Resp:	234070
Ion	Ratio	Lower	Upper
146	100		
111	42.2	21.2	63.6
148	63.2	31.9	95.9





#89

n-Butylbenzene

Concen: 54.671 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N





ClientSampleId :

ICVVN093024

Tgt Ion:	91	Resp:	396001
Ion	Ratio	Lower	Upper
91	100		
92	53.1	26.6	79.7
134	24.7	12.3	36.9

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#90

Hexachloroethane

Concen: 50.795 ug/l




RT: 14.329 min Scan# 2104


Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Tgt Ion:	117	Resp:	82363
Ion	Ratio	Lower	Upper
117	100		
201	68.4	34.4	103.2





#91

1,2-Dichlorobenzene

Concen: 51.082 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N





ClientSampleId :

ICVVN093024

Tgt Ion:	146	Resp:	226320
Ion	Ratio	Lower	Upper
146	100		
111	43.9	22.0	66.0
148	64.7	32.1	96.3

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#92

1,2-Dibromo-3-Chloropropane

Concen: 52.125 ug/l




RT: 14.717 min Scan# 2170


Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Tgt Ion:	75	Resp:	32296
Ion	Ratio	Lower	Upper
75	100		
155	74.4	38.1	114.5
157	94.7	49.6	149.0





#93

1,2,4-Trichlorobenzene

Concen: 52.311 ug/l

RT: 15.394 min Scan# 2284

Delta R.T. 0.006 min

Lab File: VN084220.D


Acq: 30 Sep 2024 15:36

Instrument :




MSVOA_N

ClientSampleId :

ICVVN093024



Tgt Ion:	180	Resp:	114157
Ion Ratio	Lower	Upper	
180	100		
182	95.1	47.0	141.0
145	33.8	16.4	49.4

Manual Integrations
APPROVEDReviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024

#94

Hexachlorobutadiene


Concen: 46.368 ug/l

RT: 15.500 min Scan# 2303



Delta R.T. -0.000 min


Lab File: VN084220.D

Acq: 30 Sep 2024 15:36



Tgt Ion:	225	Resp:	50461
Ion Ratio	Lower	Upper	
225	100		
223	61.5	32.4	97.2
227	62.7	32.4	97.0





#95

Naphthalene

Concen: 52.602 ug/l

RT: 15.641 min Scan# 2326

Delta R.T. 0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :





ICVVN093024

Tgt Ion:	128	Resp:	380549
Ion	Ratio	Lower	Upper
128	100		
127	12.6	10.5	15.7
129	11.0	8.7	13.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/01/2024
Supervised By :Mahesh Dadoda 10/01/2024



#96

1,2,3-Trichlorobenzene

Concen: 50.853 ug/l




RT: 15.835 min Scan# 2360

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Tgt Ion:	180	Resp:	111086
Ion	Ratio	Lower	Upper
180	100		
182	95.5	48.2	144.6
145	34.9	17.3	51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084220.D
 Acq On : 30 Sep 2024 15:36
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN093024

Quant Time: Oct 01 07:14:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.591	0.624	-5.6	101	0.00
3 P	Chloromethane	0.675	0.695	-3.0	103	0.00
4 C	Vinyl Chloride	0.654	0.670	-2.4#	100	0.00
5 T	Bromomethane	0.431	0.411	4.6	94	0.00
6 T	Chloroethane	0.468	0.423	9.6	100	0.00
7 T	Trichlorofluoromethane	1.019	1.082	-6.2	101	0.00
8 T	Diethyl Ether	0.378	0.404	-6.9	103	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.584	0.602	-3.1	101	0.00
10 T	Methyl Iodide	0.752	0.812	-8.0	100	0.00
11 T	Tert butyl alcohol	0.122	0.126	-3.3	101	0.00
12 CM	1,1-Dichloroethene	0.563	0.597	-6.0#	101	0.00
13 T	Acrolein	0.140	0.145	-3.6	101	0.00
14 T	Allyl chloride	0.978	0.999	-2.1	100	0.00
15 T	Acrylonitrile	0.309	0.337	-9.1	106	0.00
16 T	Acetone	0.318	0.354	-11.3	103	0.00
17 T	Carbon Disulfide	1.782	1.736	2.6	100	0.00
18 T	Methyl Acetate	0.694	0.749	-7.9	105	0.00
19 T	Methyl tert-butyl Ether	1.900	2.066	-8.7	101	0.00
20 T	Methylene Chloride	0.647	0.675	-4.3	102	0.00
21 T	trans-1,2-Dichloroethene	0.590	0.625	-5.9	100	0.00
22 T	Diisopropyl ether	2.022	2.208	-9.2	101	0.00
23 T	Vinyl Acetate	1.500	1.713	-14.2	103	0.00
24 P	1,1-Dichloroethane	1.131	1.214	-7.3	101	0.00
25 T	2-Butanone	0.434	0.490	-12.9	108	0.00
26 T	2,2-Dichloropropane	1.020	1.090	-6.9	100	0.00
27 T	cis-1,2-Dichloroethene	0.713	0.739	-3.6	99	0.00
28 T	Bromochloromethane	0.505	0.486	3.8	98	0.00
29 T	Tetrahydrofuran	0.268	0.302	-12.7	106	0.00
30 C	Chloroform	1.175	1.226	-4.3#	101	0.00
31 T	Cyclohexane	1.098	1.094	0.4	102	0.00
32 T	1,1,1-Trichloroethane	1.055	1.127	-6.8	102	0.00
33 S	1,2-Dichloroethane-d4	0.741	0.750	-1.2	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.330	0.333	-0.9	102	0.00
36 T	1,1-Dichloropropene	0.479	0.516	-7.7	100	0.00
37 T	Ethyl Acetate	0.491	0.531	-8.1	104	0.00
38 T	Carbon Tetrachloride	0.525	0.563	-7.2	101	0.00
39 T	Methylcyclohexane	0.533	0.604	-13.3	102	0.00
40 TM	Benzene	1.492	1.590	-6.6	101	0.00
41 T	Methacrylonitrile	0.267	0.297	-11.2	102	0.00
42 TM	1,2-Dichloroethane	0.506	0.548	-8.3	103	0.00
43 T	Isopropyl Acetate	0.962	0.926	3.7	102	0.00
44 TM	Trichloroethene	0.348	0.365	-4.9	100	0.00
45 C	1,2-Dichloropropane	0.353	0.386	-9.3#	102	0.00
46 T	Dibromomethane	0.238	0.267	-12.2	102	0.00
47 T	Bromodichloromethane	0.529	0.575	-8.7	102	0.00
48 T	Methyl methacrylate	0.383	0.440	-14.9	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084220.D
 Acq On : 30 Sep 2024 15:36
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN093024

Quant Time: Oct 01 07:14:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.008	-14.3	106	0.00
50 S	Toluene-d8	1.213	1.289	-6.3	103	0.00
51 T	4-Methyl-2-Pentanone	0.468	0.537	-14.7	105	0.00
52 CM	Toluene	0.910	0.986	-8.4#	101	0.00
53 T	t-1,3-Dichloropropene	0.539	0.609	-13.0	102	0.00
54 T	cis-1,3-Dichloropropene	0.580	0.643	-10.9	100	0.00
55 T	1,1,2-Trichloroethane	0.326	0.361	-10.7	102	0.00
56 T	Ethyl methacrylate	0.537	0.628	-16.9	103	0.00
57 T	1,3-Dichloropropane	0.583	0.638	-9.4	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.249	0.276	-10.8	97	0.00
59 T	2-Hexanone	0.349	0.413	-18.3	106	0.00
60 T	Dibromochloromethane	0.385	0.425	-10.4	101	0.00
61 T	1,2-Dibromoethane	0.339	0.370	-9.1	102	0.00
62 S	4-Bromofluorobenzene	0.442	0.468	-5.9	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.348	0.366	-5.2	102	0.00
65 PM	Chlorobenzene	1.109	1.187	-7.0	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.382	0.423	-10.7	104	0.00
67 C	Ethyl Benzene	1.961	2.179	-11.1#	102	0.00
68 T	m/p-Xylenes	0.725	0.825	-13.8	102	0.00
69 T	o-Xylene	0.686	0.796	-16.0	102	0.00
70 T	Styrene	1.162	1.351	-16.3	102	0.00
71 P	Bromoform	0.281	0.322	-14.6	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.815	4.090	-7.2	101	0.00
74 T	N-amyl acetate	1.728	1.905	-10.2	105	0.00
75 P	1,1,2,2-Tetrachloroethane	1.147	1.202	-4.8	104	0.00
76 T	1,2,3-Trichloropropane	1.021	1.030	-0.9	94	0.00
77 T	Bromobenzene	0.920	0.951	-3.4	103	0.00
78 T	n-propylbenzene	4.421	4.879	-10.4	102	0.00
79 T	2-Chlorotoluene	2.821	2.984	-5.8	101	0.00
80 T	1,3,5-Trimethylbenzene	3.112	3.492	-12.2	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.426	0.451	-5.9	102	0.00
82 T	4-Chlorotoluene	2.840	3.036	-6.9	103	0.00
83 T	tert-Butylbenzene	2.766	2.982	-7.8	101	0.00
84 T	1,2,4-Trimethylbenzene	3.135	3.524	-12.4	103	0.00
85 T	sec-Butylbenzene	3.691	4.101	-11.1	101	0.00
86 T	p-Isopropyltoluene	3.052	3.456	-13.2	101	0.00
87 T	1,3-Dichlorobenzene	1.727	1.790	-3.6	103	0.00
88 T	1,4-Dichlorobenzene	1.744	1.789	-2.6	103	0.00
89 T	n-Butylbenzene	2.768	3.026	-9.3	100	0.00
90 T	Hexachloroethane	0.620	0.629	-1.5	101	0.00
91 T	1,2-Dichlorobenzene	1.693	1.730	-2.2	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.237	0.247	-4.2	106	0.00
93 T	1,2,4-Trichlorobenzene	0.834	0.872	-4.6	96	0.00
94 T	Hexachlorobutadiene	0.416	0.386	7.2	100	0.00
95 T	Naphthalene	2.764	2.908	-5.2	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084220.D
Acq On : 30 Sep 2024 15:36
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024

Quant Time: Oct 01 07:14:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.835	0.849	-1.7	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084220.D
Acq On : 30 Sep 2024 15:36
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024

Quant Time: Oct 01 07:14:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	50.000	52.712	-5.4	101	0.00
3 P	Chloromethane	50.000	51.477	-3.0	103	0.00
4 C	Vinyl Chloride	50.000	51.235	-2.5#	100	0.00
5 T	Bromomethane	50.000	47.656	4.7	94	0.00
6 T	Chloroethane	50.000	45.119	9.8	100	0.00
7 T	Trichlorofluoromethane	50.000	53.099	-6.2	101	0.00
8 T	Diethyl Ether	50.000	53.445	-6.9	103	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.573	-3.1	101	0.00
10 T	Methyl Iodide	50.000	53.996	-8.0	100	0.00
11 T	Tert butyl alcohol	250.000	257.154	-2.9	101	0.00
12 CM	1,1-Dichloroethene	50.000	53.064	-6.1#	101	0.00
13 T	Acrolein	250.000	257.587	-3.0	101	0.00
14 T	Allyl chloride	50.000	51.092	-2.2	100	0.00
15 T	Acrylonitrile	250.000	273.264	-9.3	106	0.00
16 T	Acetone	250.000	277.547	-11.0	103	0.00
17 T	Carbon Disulfide	50.000	48.693	2.6	100	0.00
18 T	Methyl Acetate	50.000	53.948	-7.9	105	0.00
19 T	Methyl tert-butyl Ether	50.000	54.371	-8.7	101	0.00
20 T	Methylene Chloride	50.000	52.116	-4.2	102	0.00
21 T	trans-1,2-Dichloroethene	50.000	52.980	-6.0	100	0.00
22 T	Diisopropyl ether	50.000	54.579	-9.2	101	0.00
23 T	Vinyl Acetate	250.000	285.569	-14.2	103	0.00
24 P	1,1-Dichloroethane	50.000	53.676	-7.4	101	0.00
25 T	2-Butanone	250.000	282.397	-13.0	108	0.00
26 T	2,2-Dichloropropane	50.000	53.428	-6.9	100	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.876	-3.8	99	0.00
28 T	Bromochloromethane	50.000	48.129	3.7	98	0.00
29 T	Tetrahydrofuran	250.000	281.940	-12.8	106	0.00
30 C	Chloroform	50.000	52.184	-4.4#	101	0.00
31 T	Cyclohexane	50.000	49.809	0.4	102	0.00
32 T	1,1,1-Trichloroethane	50.000	53.395	-6.8	102	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.564	-1.1	101	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	99	0.00
35 S	Dibromofluoromethane	50.000	50.508	-1.0	102	0.00
36 T	1,1-Dichloropropene	50.000	53.917	-7.8	100	0.00
37 T	Ethyl Acetate	50.000	54.043	-8.1	104	0.00
38 T	Carbon Tetrachloride	50.000	53.593	-7.2	101	0.00
39 T	Methylcyclohexane	50.000	56.648	-13.3	102	0.00
40 TM	Benzene	50.000	53.301	-6.6	101	0.00
41 T	Methacrylonitrile	50.000	55.708	-11.4	102	0.00
42 TM	1,2-Dichloroethane	50.000	54.131	-8.3	103	0.00
43 T	Isopropyl Acetate	50.000	48.126	3.7	102	0.00
44 TM	Trichloroethene	50.000	52.368	-4.7	100	0.00
45 C	1,2-Dichloropropane	50.000	54.694	-9.4#	102	0.00
46 T	Dibromomethane	50.000	56.015	-12.0	102	0.00
47 T	Bromodichloromethane	50.000	54.335	-8.7	102	0.00
48 T	Methyl methacrylate	50.000	57.391	-14.8	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084220.D
 Acq On : 30 Sep 2024 15:36
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN093024

Quant Time: Oct 01 07:14:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1111.545	-11.2	106	0.00
50 S	Toluene-d8	50.000	53.129	-6.3	103	0.00
51 T	4-Methyl-2-Pentanone	250.000	287.336	-14.9	105	0.00
52 CM	Toluene	50.000	54.185	-8.4#	101	0.00
53 T	t-1,3-Dichloropropene	50.000	56.426	-12.9	102	0.00
54 T	cis-1,3-Dichloropropene	50.000	55.416	-10.8	100	0.00
55 T	1,1,2-Trichloroethane	50.000	55.307	-10.6	102	0.00
56 T	Ethyl methacrylate	50.000	58.561	-17.1	103	0.00
57 T	1,3-Dichloropropane	50.000	54.738	-9.5	102	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	277.315	-10.9	97	0.00
59 T	2-Hexanone	250.000	295.238	-18.1	106	0.00
60 T	Dibromochloromethane	50.000	55.105	-10.2	101	0.00
61 T	1,2-Dibromoethane	50.000	54.548	-9.1	102	0.00
62 S	4-Bromofluorobenzene	50.000	52.971	-5.9	103	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	99	0.00
64 T	Tetrachloroethene	50.000	52.585	-5.2	102	0.00
65 PM	Chlorobenzene	50.000	53.555	-7.1	101	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	55.381	-10.8	104	0.00
67 C	Ethyl Benzene	50.000	55.563	-11.1#	102	0.00
68 T	m/p-Xylenes	100.000	113.748	-13.7	102	0.00
69 T	o-Xylene	50.000	58.032	-16.1	102	0.00
70 T	Styrene	50.000	58.136	-16.3	102	0.00
71 P	Bromoform	50.000	57.258	-14.5	102	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	102	0.00
73 T	Isopropylbenzene	50.000	53.604	-7.2	101	0.00
74 T	N-amyl acetate	50.000	55.118	-10.2	105	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	52.398	-4.8	104	0.00
76 T	1,2,3-Trichloropropane	50.000	50.475	-1.0	94	0.00
77 T	Bromobenzene	50.000	51.676	-3.4	103	0.00
78 T	n-propylbenzene	50.000	55.188	-10.4	102	0.00
79 T	2-Chlorotoluene	50.000	52.898	-5.8	101	0.00
80 T	1,3,5-Trimethylbenzene	50.000	56.101	-12.2	102	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	52.939	-5.9	102	0.00
82 T	4-Chlorotoluene	50.000	53.455	-6.9	103	0.00
83 T	tert-Butylbenzene	50.000	53.903	-7.8	101	0.00
84 T	1,2,4-Trimethylbenzene	50.000	56.199	-12.4	103	0.00
85 T	sec-Butylbenzene	50.000	55.556	-11.1	101	0.00
86 T	p-Isopropyltoluene	50.000	56.614	-13.2	101	0.00
87 T	1,3-Dichlorobenzene	50.000	51.830	-3.7	103	0.00
88 T	1,4-Dichlorobenzene	50.000	51.298	-2.6	103	0.00
89 T	n-Butylbenzene	50.000	54.671	-9.3	100	0.00
90 T	Hexachloroethane	50.000	50.795	-1.6	101	0.00
91 T	1,2-Dichlorobenzene	50.000	51.082	-2.2	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	52.125	-4.3	106	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.311	-4.6	96	0.00
94 T	Hexachlorobutadiene	50.000	46.368	7.3	100	0.00
95 T	Naphthalene	50.000	52.602	-5.2	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084220.D
Acq On : 30 Sep 2024 15:36
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024

Quant Time: Oct 01 07:14:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.853	-1.7	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG No.: P4440

Instrument ID: MSVOA_N Calibration Date/Time: 10/21/2024 11:38

Lab File ID: VN084428.D Init. Calib. Date(s): 09/30/2024 09/30/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:25 14:48

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.675	0.602	0.1	-10.81	20
Vinyl Chloride	0.654	0.597		-8.72	20
Bromomethane	0.431	0.370		-14.15	20
Chloroethane	0.468	0.401		-14.32	20
Trichlorofluoromethane	1.019	1.044		2.45	20
1,1,2-Trichlorotrifluoroethane	0.584	0.603		3.25	20
1,1-Dichloroethene	0.563	0.566		0.53	20
Acetone	0.318	0.317		-0.31	20
Carbon Disulfide	1.782	1.424		-20.09	20
Methyl tert-butyl Ether	1.900	2.031		6.89	20
Methylene Chloride	0.647	0.671		3.71	20
trans-1,2-Dichloroethene	0.590	0.593		0.51	20
1,1-Dichloroethane	1.131	1.189	0.1	5.13	20
2-Butanone	0.434	0.467		7.6	20
Carbon Tetrachloride	0.525	0.569		8.38	20
cis-1,2-Dichloroethene	0.713	0.752		5.47	20
Chloroform	1.175	1.280		8.94	20
1,1,1-Trichloroethane	1.055	1.115		5.69	20
Methylcyclohexane	0.533	0.522		-2.06	20
Benzene	1.492	1.635		9.58	20
1,2-Dichloroethane	0.506	0.549		8.5	20
Trichloroethene	0.348	0.375		7.76	20
1,2-Dichloropropane	0.353	0.401		13.6	20
Bromodichloromethane	0.529	0.606		14.56	20
4-Methyl-2-Pentanone	0.468	0.576		23.08	20
Toluene	0.910	1.023		12.42	20
t-1,3-Dichloropropene	0.539	0.601		11.5	20
cis-1,3-Dichloropropene	0.580	0.651		12.24	20
1,1,2-Trichloroethane	0.326	0.397		21.78	20
2-Hexanone	0.349	0.426		22.06	20
Dibromochloromethane	0.385	0.466		21.04	20
Tetrachloroethene	0.348	0.366		5.17	20
Chlorobenzene	1.109	1.181	0.3	6.49	20
Ethyl Benzene	1.961	2.070		5.56	20
m/p-Xylenes	0.725	0.807		11.31	20
o-Xylene	0.686	0.765		11.52	20
Styrene	1.162	1.354		16.52	20
Bromoform	0.281	0.332	0.1	18.15	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG No.: P4440

Instrument ID: MSVOA_N Calibration Date/Time: 10/21/2024 11:38

Lab File ID: VN084428.D Init. Calib. Date(s): 09/30/2024 09/30/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:25 14:48

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.815	3.797		-0.47	20
1,1,2,2-Tetrachloroethane	1.147	1.198	0.3	4.45	20
1,3-Dichlorobenzene	1.727	1.778		2.95	20
1,4-Dichlorobenzene	1.744	1.755		0.63	20
1,2-Dichlorobenzene	1.693	1.727		2.01	20
1,2-Dichloroethane-d4	0.741	0.791		6.75	20
Dibromofluoromethane	0.330	0.387		17.27	20
Toluene-d8	1.213	1.431		17.97	20
4-Bromofluorobenzene	0.442	0.522		18.1	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084428.D
 Acq On : 21 Oct 2024 11:38
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :John Carlone 10/22/2024
 Supervised By :Mahesh Dadoda 10/22/2024

Quant Time: Oct 22 01:29:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	157355	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	262800	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	244635	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	127638	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	124506	53.365	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.740%			
35) Dibromofluoromethane	8.165	113	101751	58.729	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 117.460%			
50) Toluene-d8	10.565	98	376165	59.014	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 118.020%#			
62) 4-Bromofluorobenzene	12.847	95	137246	59.102	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 118.200%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	83308	44.759	ug/l	98
3) Chloromethane	2.359	50	94679	44.550	ug/l	98
4) Vinyl Chloride	2.512	62	93996	45.660	ug/l	97
5) Bromomethane	2.936	94	58206	42.866	ug/l	93
6) Chloroethane	3.106	64	63113	42.808	ug/l	94
7) Trichlorofluoromethane	3.489	101	164207	51.224	ug/l	99
8) Diethyl Ether	3.959	74	59329	49.902	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.365	101	94908	51.670	ug/l	99
10) Methyl Iodide	4.583	142	116312	49.171	ug/l	96
11) Tert butyl alcohol	5.530	59	94486	245.606	ug/l	100
12) 1,1-Dichloroethene	4.342	96	89116	50.300	ug/l	98
13) Acrolein	4.177	56	85169	192.843	ug/l	100
14) Allyl chloride	5.018	41	134288	43.643	ug/l	98
15) Acrylonitrile	5.718	53	275307	283.386	ug/l	99
16) Acetone	4.430	43	249463	248.905	ug/l	100
17) Carbon Disulfide	4.706	76	224021	39.937	ug/l	99
18) Methyl Acetate	5.024	43	143956	65.900	ug/l	99
19) Methyl tert-butyl Ether	5.794	73	319663	53.457	ug/l	98
20) Methylene Chloride	5.271	84	105509	51.787	ug/l	88
21) trans-1,2-Dichloroethene	5.783	96	93324	50.278	ug/l	94
22) Diisopropyl ether	6.671	45	336994	52.949	ug/l	99
23) Vinyl Acetate	6.606	43	1190800	252.258	ug/l	100
24) 1,1-Dichloroethane	6.571	63	187107	52.570	ug/l	98
25) 2-Butanone	7.483	43	367222	269.013	ug/l	99
26) 2,2-Dichloropropane	7.488	77	159617	49.738	ug/l	97
27) cis-1,2-Dichloroethene	7.488	96	118267	52.728	ug/l	96
28) Bromochloromethane	7.812	49	83905	52.836	ug/l	93
29) Tetrahydrofuran	7.841	42	238619	283.258	ug/l	98
30) Chloroform	7.965	83	201405	54.476	ug/l	98
31) Cyclohexane	8.259	56	150876	43.659	ug/l	98
32) 1,1,1-Trichloroethane	8.165	97	175417	52.830	ug/l	98
36) 1,1-Dichloropropene	8.371	75	130914	52.021	ug/l	98
37) Ethyl Acetate	7.559	43	137280	53.172	ug/l	98
38) Carbon Tetrachloride	8.359	117	149628	54.202	ug/l	97
39) Methylcyclohexane	9.600	83	137117	48.955	ug/l	97
40) Benzene	8.606	78	429647	54.796	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084428.D
 Acq On : 21 Oct 2024 11:38
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024

Quant Time: Oct 22 01:29:31 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M

Quant Title : SW846 8260

QLast Update : Tue Oct 01 07:11:01 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	79086	56.424	ug/l	96
42) 1,2-Dichloroethane	8.671	62	144178	54.240	ug/l	99
43) Isopropyl Acetate	8.688	43	233616	46.197	ug/l	98
44) Trichloroethene	9.353	130	98494	53.784	ug/l	100
45) 1,2-Dichloropropane	9.618	63	105490	56.873	ug/l	96
46) Dibromomethane	9.706	93	74735	59.706	ug/l	98
47) Bromodichloromethane	9.888	83	159166	57.235	ug/l #	96
48) Methyl methacrylate	9.682	41	112649	55.960	ug/l	99
49) 1,4-Dioxane	9.694	88	46681	1259.594	ug/l	97
51) 4-Methyl-2-Pentanone	10.441	43	757152	308.066	ug/l	99
52) Toluene	10.629	92	268859	56.231	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	157928	55.712	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	171021	56.089	ug/l	96
55) 1,1,2-Trichloroethane	11.018	97	104404	60.910	ug/l	97
56) Ethyl methacrylate	10.871	69	164383	58.283	ug/l	96
57) 1,3-Dichloropropane	11.165	76	176631	57.670	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	397738	303.996	ug/l	97
59) 2-Hexanone	11.194	43	560104	304.908	ug/l	97
60) Dibromochloromethane	11.359	129	122440	60.458	ug/l	99
61) 1,2-Dibromoethane	11.471	107	102008	57.274	ug/l	100
64) Tetrachloroethene	11.100	164	89495	52.523	ug/l	97
65) Chlorobenzene	11.888	112	289014	53.283	ug/l	93
66) 1,1,1,2-Tetrachloroethane	11.959	131	103631	55.468	ug/l	98
67) Ethyl Benzene	11.965	91	506469	52.799	ug/l	99
68) m/p-Xylenes	12.071	106	394932	111.301	ug/l	99
69) o-Xylene	12.394	106	187215	55.767	ug/l	100
70) Styrene	12.412	104	331204	58.246	ug/l	98
71) Bromoform	12.576	173	81253	59.072	ug/l #	100
73) Isopropylbenzene	12.694	105	484656	49.760	ug/l	100
74) N-amyl acetate	12.494	43	204351	46.332	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.935	83	152956	52.226	ug/l	98
76) 1,2,3-Trichloropropane	12.994	75	127281m	48.846	ug/l	
77) Bromobenzene	12.976	156	118952	50.631	ug/l	96
78) n-propylbenzene	13.035	91	576662	51.101	ug/l	98
79) 2-Chlorotoluene	13.123	91	354683	49.256	ug/l	98
80) 1,3,5-Trimethylbenzene	13.170	105	416797	52.460	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	58513	53.835	ug/l	90
82) 4-Chlorotoluene	13.218	91	364438	50.275	ug/l	99
83) tert-Butylbenzene	13.435	119	352099	49.862	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	421969	52.722	ug/l	99
85) sec-Butylbenzene	13.617	105	492135	52.226	ug/l	99
86) p-Isopropyltoluene	13.729	119	412249	52.906	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	226954	51.493	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	223977	50.322	ug/l	99
89) n-Butylbenzene	14.053	91	349709	49.495	ug/l	99
90) Hexachloroethane	14.329	117	76242	48.204	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	220370	50.991	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	30011	49.657	ug/l	97
93) 1,2,4-Trichlorobenzene	15.388	180	109356	51.373	ug/l	99
94) Hexachlorobutadiene	15.500	225	50783	47.838	ug/l	96
95) Naphthalene	15.641	128	361412	51.214	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	110846	52.021	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084428.D
Acq On : 21 Oct 2024 11:38
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024

Quant Time: Oct 22 01:29:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084428.D
Acq On : 21 Oct 2024 11:38
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Manual Integrations

Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024


Quant Time: Oct 22 01:29:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:168 Resp: 15735
Ion Ratio Lower Upper
168 100
99 61.7 54.2 81.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#2
Dichlorodifluoromethane
Concen: 44.759 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38


Tgt Ion: 85 Resp: 83308
Ion Ratio Lower Upper
85 100
87 34.4 17.6 52.9





#3
Chloromethane
Concen: 44.550 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 50 Resp: 94679
Ion Ratio Lower Upper
50 100
52 32.3 26.8 40.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#4
Vinyl Chloride
Concen: 45.660 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 62 Resp: 93996
Ion Ratio Lower Upper
62 100
64 30.9 26.2 39.4









#5
Bromomethane
Concen: 42.866 ug/l
RT: 2.936 min Scan# 11
Delta R.T. -0.018 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 94 Resp: 58200
Ion Ratio Lower Upper
94 100
96 98.2 72.9 109.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#6
Chloroethane
Concen: 42.808 ug/l
RT: 3.106 min Scan# 196
Delta R.T. -0.012 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38


Tgt Ion: 64 Resp: 63113
Ion Ratio Lower Upper
64 100
66 32.4 23.5 35.3





#7
Trichlorofluoromethane
Concen: 51.224 ug/l
RT: 3.489 min Scan# 262
Delta R.T. -0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:101 Resp: 16420
Ion Ratio Lower Upper
101 100
103 64.0 51.8 77.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#8
Diethyl Ether
Concen: 49.902 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 74 Resp: 59329
Ion Ratio Lower Upper
74 100
45 95.6 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 51.670 ug/l
RT: 4.365 min Scan# 411
Delta R.T. -0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 94908
Ion Ratio Lower Upper
101 100
85 44.2 36.2 54.4
151 75.0 59.3 88.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#10
Methyl Iodide
Concen: 49.171 ug/l
RT: 4.583 min Scan# 447
Delta R.T. -0.012 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion:142 Resp: 116312
Ion Ratio Lower Upper
142 100
127 47.5 35.3 52.9
141 14.4 11.0 16.6





#11

Tert butyl alcohol

Concen: 245.606 ug/l

RT: 5.530 min Scan# 607

Delta R.T. 0.006 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 59 Resp: 9448

Ion Ratio Lower Upper

59 100





57 10.2 8.2 12.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#12

1,1-Dichloroethene

Concen: 50.300 ug/l

RT: 4.342 min Scan# 406

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38




Tgt Ion: 96 Resp: 89116


Ion Ratio Lower Upper

96 100

61 162.3 128.6 192.8

98 60.6 51.7 77.5





#13

Acrolein

Concen: 192.843 ug/l

RT: 4.177 min Scan# 31

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 56 Resp: 85169

Ion Ratio Lower Upper

56 100





55 71.1 57.2 85.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#14

Allyl chloride

Concen: 43.643 ug/l

RT: 5.018 min Scan# 521

Delta R.T. -0.006 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38




Tgt Ion: 41 Resp: 134288


Ion Ratio Lower Upper

41 100

39 73.1 59.0 88.6


76 39.4 29.8 44.6





#15
Acrylonitrile
Concen: 283.386 ug/l
RT: 5.718 min Scan# 640
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 53 Resp: 27530
Ion Ratio Lower Upper
53 100
52 82.1 66.2 99.4
51 35.2 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#16
Acetone
Concen: 248.905 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38


Tgt Ion: 43 Resp: 249463
Ion Ratio Lower Upper
43 100
58 30.4 24.4 36.6





#17
Carbon Disulfide
Concen: 39.937 ug/l
RT: 4.706 min Scan# 468
Delta R.T. -0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 76 Resp: 22402
Ion Ratio Lower Upper
76 100
78 9.4 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#18
Methyl Acetate
Concen: 65.900 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 43 Resp: 143956
Ion Ratio Lower Upper
43 100
74 24.6 19.2 28.8









#19
Methyl tert-butyl Ether
Concen: 53.457 ug/l
RT: 5.794 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 73 Resp: 31966
Ion Ratio Lower Upper
73 100
57 23.3 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#20
Methylene Chloride
Concen: 51.787 ug/l
RT: 5.271 min Scan# 564
Delta R.T. -0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 84 Resp: 105509
Ion Ratio Lower Upper
84 100
49 112.4 107.2 160.8
51 37.8 31.8 47.8
86 61.9 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 50.278 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 96 Resp: 93324
Ion Ratio Lower Upper
96 100
61 126.8 108.4 162.6
98 59.2 49.8 74.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#22
Diisopropyl ether
Concen: 52.949 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 45 Resp: 336994
Ion Ratio Lower Upper
45 100
43 52.6 42.3 63.5
87 26.8 21.8 32.6
59 12.8 9.6 14.4









#23
Vinyl Acetate
Concen: 252.258 ug/l
RT: 6.606 min Scan# 791
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 1190800
Ion Ratio Lower Upper
43 100
86 10.9 8.6 12.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#24
1,1-Dichloroethane
Concen: 52.570 ug/l
RT: 6.571 min Scan# 785
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38


Tgt Ion: 63 Resp: 187107
Ion Ratio Lower Upper
63 100
98 6.2 3.4 10.2
100 4.9 2.0 5.8





#25
2-Butanone
Concen: 269.013 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 43 Resp: 36722
Ion Ratio Lower Upper
43 100
72 27.4 21.4 32.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024




#26
2,2-Dichloropropane
Concen: 49.738 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Tgt Ion: 77 Resp: 159617
Ion Ratio Lower Upper
77 100
97 23.0 10.8 32.3





#27
cis-1,2-Dichloroethene
Concen: 52.728 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 96 Resp: 11826
Ion Ratio Lower Upper
96 100
61 144.9 0.0 301.8
98 65.0 0.0 128.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#28
Bromochloromethane
Concen: 52.836 ug/l
RT: 7.812 min Scan# 996
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 49 Resp: 83905
Ion Ratio Lower Upper
49 100
129 1.7 0.0 3.6
130 78.3 58.0 87.0





#29

Tetrahydrofuran

Concen: 283.258 ug/l

RT: 7.841 min Scan# 1001

Delta R.T. 0.006 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 42 Resp: 238619

Ion Ratio Lower Upper

42 100

72 45.2 34.6 51.8





71 41.3 32.2 48.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#30

Chloroform

Concen: 54.476 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.000 min

Lab File: VN084428.D




Acq: 21 Oct 2024 11:38


Tgt Ion: 83 Resp: 201405

Ion Ratio Lower Upper

83 100


85 65.0 51.0 76.4





#31
Cyclohexane
Concen: 43.659 ug/l
RT: 8.259 min Scan# 1071
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 150870
Ion Ratio Lower Upper
56 100
69 27.6 23.8 35.6
84 84.5 68.8 103.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#32
1,1,1-Trichloroethane
Concen: 52.830 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 97 Resp: 175417
Ion Ratio Lower Upper
97 100
99 65.5 51.2 76.8
61 46.3 37.5 56.3





#33

1,2-Dichloroethane-d4

Concen: 53.365 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 124500

Ion Ratio Lower Upper

65 100





67 53.5 0.0 102.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38




Tgt Ion:114 Resp: 262800


Ion Ratio Lower Upper

114 100

63 22.4 0.0 43.8

88 16.4 0.0 31.6





#35

Dibromofluoromethane

Concen: 58.729 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:113 Resp: 10175

Ion Ratio Lower Upper

113 100

111 102.8 83.3 124.9





192 17.8 13.5 20.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#36

1,1-Dichloropropene

Concen: 52.021 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38




Tgt Ion: 75 Resp: 130914


Ion Ratio Lower Upper

75 100

110 34.9 16.9 50.7

77 31.6 24.7 37.1





#37

Ethyl Acetate

Concen: 53.172 ug/l

RT: 7.559 min Scan# 953

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 137280

Ion Ratio Lower Upper

43 100

61 14.4 10.7 16.1





70 11.2 8.8 13.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#38

Carbon Tetrachloride

Concen: 54.202 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38




Tgt Ion:117 Resp: 149628


Ion Ratio Lower Upper

117 100

119 95.2 78.7 118.1

121 28.3 24.4 36.6





#39

Methylcyclohexane

Concen: 48.955 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 13711

Ion Ratio Lower Upper

83 100

55 74.6 62.8 94.2





98 49.4 38.8 58.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#40

Benzene

Concen: 54.796 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VN084428.D




Acq: 21 Oct 2024 11:38


Tgt Ion: 78 Resp: 429647

Ion Ratio Lower Upper

78 100


77 23.3 19.1 28.7





#41
Methacrylonitrile
Concen: 56.424 ug/l
RT: 7.777 min Scan# 990
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 79086
Ion Ratio Lower Upper
41 100
39 56.4 43.3 64.9
67 75.1 56.1 84.1
52 30.9 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#42
1,2-Dichloroethane
Concen: 54.240 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38


Tgt Ion: 62 Resp: 144178
Ion Ratio Lower Upper
62 100
98 9.6 0.0 18.2





#43
Isopropyl Acetate
Concen: 46.197 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 233610
Ion Ratio Lower Upper
43 100
61 26.8 20.4 30.6
87 13.4 10.3 15.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#44
Trichloroethene
Concen: 53.784 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion:130 Resp: 98494
Ion Ratio Lower Upper
130 100
95 102.0 0.0 203.0





#45

1,2-Dichloropropane

Concen: 56.873 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 105490

Ion Ratio Lower Upper

63 100





65 32.8 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#46

Dibromomethane

Concen: 59.706 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38




Tgt Ion: 93 Resp: 74735


Ion Ratio Lower Upper

93 100

95 80.7 65.0 97.4

174 93.5 71.8 107.8





#47

Bromodichloromethane

Concen: 57.235 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 159160

Ion Ratio Lower Upper

83 100

85 61.4 51.5 77.3





127 8.9 5.7 8.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#48

Methyl methacrylate

Concen: 55.960 ug/l

RT: 9.682 min Scan# 1314

Delta R.T. 0.006 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38




Tgt Ion: 41 Resp: 112649


Ion Ratio Lower Upper

41 100

69 84.7 67.1 100.7

39 55.9 45.8 68.6





#49

1,4-Dioxane

Concen: 1259.594 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 88 Resp: 4668

Ion Ratio Lower Upper

88 100

43 28.2 25.3 37.9





58 71.6 56.4 84.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#50

Toluene-d8

Concen: 59.014 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN084428.D




Acq: 21 Oct 2024 11:38


Tgt Ion: 98 Resp: 376165

Ion Ratio Lower Upper

98 100

100 64.8 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 308.066 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 757152
Ion Ratio Lower Upper
43 100
58 39.9 31.6 47.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#52
Toluene
Concen: 56.231 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38


Tgt Ion: 92 Resp: 268859
Ion Ratio Lower Upper
92 100
91 170.0 137.3 205.9





#53
t-1,3-Dichloropropene
Concen: 55.712 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 157928
Ion Ratio Lower Upper
75 100
77 32.5 25.4 38.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#54
cis-1,3-Dichloropropene
Concen: 56.089 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38


Tgt Ion: 75 Resp: 171021
Ion Ratio Lower Upper
75 100
77 30.9 24.2 36.2
39 44.8 38.7 58.1





#55
1,1,2-Trichloroethane
Concen: 60.910 ug/l
RT: 11.018 min Scan# 1540
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 97 Resp: 104404
Ion Ratio Lower Upper
97 100
83 83.3 71.0 106.4
85 52.3 41.7 62.5
99 60.4 50.1 75.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#56
Ethyl methacrylate
Concen: 58.283 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 69 Resp: 164383
Ion Ratio Lower Upper
69 100
41 63.2 53.3 79.9
39 34.8 28.8 43.2





#57

1,3-Dichloropropane

Concen: 57.670 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.006 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 17663

Ion Ratio Lower Upper

76 100





78 32.5 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#58

2-Chloroethyl Vinyl ether

Concen: 303.996 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN084428.D




Acq: 21 Oct 2024 11:38


Tgt Ion: 63 Resp: 397738

Ion Ratio Lower Upper

63 100


106 28.9 21.8 32.6





#59
2-Hexanone
Concen: 304.908 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 560104
Ion Ratio Lower Upper
43 100
58 56.4 27.3 81.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#60
Dibromochloromethane
Concen: 60.458 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38


Tgt Ion:129 Resp: 122440
Ion Ratio Lower Upper
129 100
127 75.8 38.3 114.9





#61
1,2-Dibromoethane
Concen: 57.274 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:107 Resp: 102008
Ion Ratio Lower Upper
107 100
109 94.4 75.4 113.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#62
4-Bromofluorobenzene
Concen: 59.102 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 95 Resp: 137246
Ion Ratio Lower Upper
95 100
174 76.1 0.0 145.2
176 71.4 0.0 140.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:117 Resp: 24463

Ion Ratio Lower Upper

117 100

82 56.4 47.2 70.8





119 30.2 25.4 38.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#64

Tetrachloroethene

Concen: 52.523 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Tgt Ion:164 Resp: 89495




Ion Ratio Lower Upper


164 100

166 126.0 101.0 151.4

129 95.5 80.8 121.2


131 94.8 78.1 117.1





#65
Chlorobenzene
Concen: 53.283 ug/l
RT: 11.888 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:112 Resp: 289014
Ion Ratio Lower Upper
112 100
114 29.8 27.0 40.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024




#66
1,1,1,2-Tetrachloroethane
Concen: 55.468 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38



Tgt Ion:131 Resp: 103631
Ion Ratio Lower Upper
131 100
133 93.7 47.9 143.7
119 65.0 32.8 98.3









#67
Ethyl Benzene
Concen: 52.799 ug/l
RT: 11.965 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 506469
Ion Ratio Lower Upper
91 100
106 31.0 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#68
m/p-Xylenes
Concen: 111.301 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion:106 Resp: 394932
Ion Ratio Lower Upper
106 100
91 206.8 167.1 250.7









#69
o-Xylene
Concen: 55.767 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 187215
Ion Ratio Lower Upper
106 100
91 219.7 110.2 330.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#70
Styrene
Concen: 58.246 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38


Tgt Ion:104 Resp: 331204
Ion Ratio Lower Upper
104 100
78 50.7 42.7 64.1
103 54.3 43.6 65.4





#71
Bromoform
Concen: 59.072 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:173 Resp: 8125
Ion Ratio Lower Upper
173 100
175 47.1 23.5 70.6
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Tgt Ion:152 Resp: 127638
Ion Ratio Lower Upper
152 100
115 61.0 31.3 93.9
150 173.0 0.0 349.8





#73
Isopropylbenzene
Concen: 49.760 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:105 Resp: 484650
Ion Ratio Lower Upper
105 100
120 25.6 12.9 38.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#74
N-ethyl acetate
Concen: 46.332 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38


Tgt Ion: 43 Resp: 204351
Ion Ratio Lower Upper
43 100
70 44.9 34.4 51.6
55 26.9 19.6 29.4
61 25.1 19.6 29.4





#75
1,1,2,2-Tetrachloroethane
Concen: 52.226 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion: 83 Resp: 152950
Ion Ratio Lower Upper
83 100
131 10.5 4.9 14.7
85 65.3 31.9 95.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024




#76
1,2,3-Trichloropropane
Concen: 48.846 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Tgt Ion: 75 Resp: 127281
Ion Ratio Lower Upper
75 100
77 206.1 96.9 290.7





#77
Bromobenzene
Concen: 50.631 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:156 Resp: 118952
Ion Ratio Lower Upper
156 100
77 220.5 115.1 345.3
158 97.0 48.9 146.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#78
n-propylbenzene
Concen: 51.101 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 91 Resp: 576662
Ion Ratio Lower Upper
91 100
120 22.6 10.8 32.4





#79

2-Chlorotoluene

Concen: 49.256 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 35468

Ion Ratio Lower Upper

91 100





126 32.7 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#80

1,3,5-Trimethylbenzene

Concen: 52.460 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VN084428.D




Acq: 21 Oct 2024 11:38


Tgt Ion:105 Resp: 416797

Ion Ratio Lower Upper

105 100


120 47.7 23.7 71.1





#81
trans-1,4-Dichloro-2-butene
Concen: 53.835 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 5851
Ion Ratio Lower Upper
75 100
53 86.4 78.0 117.0
89 40.2 35.5 53.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#82
4-Chlorotoluene
Concen: 50.275 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 91 Resp: 364438
Ion Ratio Lower Upper
91 100
126 31.9 15.7 47.1





#83

tert-Butylbenzene

Concen: 49.862 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:119 Resp: 352099

Ion Ratio Lower Upper

119 100

91 68.1 34.1 102.3





134 26.4 12.8 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#84

1,2,4-Trimethylbenzene

Concen: 52.722 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.000 min

Lab File: VN084428.D




Acq: 21 Oct 2024 11:38


Tgt Ion:105 Resp: 421969

Ion Ratio Lower Upper

105 100


120 44.9 22.3 66.8





#85
 sec-Butylbenzene
 Concen: 52.226 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. 0.000 min
 Lab File: VN084428.D
 Acq: 21 Oct 2024 11:38




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050




Tgt Ion:105 Resp: 492135
 Ion Ratio Lower Upper
 105 100
 134 19.2 9.8 29.4

Manual Integrations
 APPROVED



Reviewed By :John Carlone 10/22/2024
 Supervised By :Mahesh Dadoda 10/22/2024




#86
 p-Isopropyltoluene
 Concen: 52.906 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN084428.D
 Acq: 21 Oct 2024 11:38




Tgt Ion:119 Resp: 412249
 Ion Ratio Lower Upper
 119 100
 134 26.0 12.6 37.8
 91 24.7 12.4 37.2





#87
1,3-Dichlorobenzene
Concen: 51.493 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050




Tgt Ion:146 Resp: 226954
Ion Ratio Lower Upper
146 100
111 42.6 21.8 65.3
148 63.9 31.9 95.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024




#88
1,4-Dichlorobenzene
Concen: 50.322 ug/l
RT: 13.812 min Scan# 2016
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38



Tgt Ion:146 Resp: 223977
Ion Ratio Lower Upper
146 100
111 41.7 21.2 63.6
148 63.1 31.9 95.9





#89

n-Butylbenzene

Concen: 49.495 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 349709

Ion Ratio Lower Upper

91 100

92 52.7 26.6 79.7





134 24.9 12.3 36.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#90

Hexachloroethane

Concen: 48.204 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. 0.000 min

Lab File: VN084428.D




Acq: 21 Oct 2024 11:38


Tgt Ion:117 Resp: 76242

Ion Ratio Lower Upper

117 100


201 69.3 34.4 103.2





#91
1,2-Dichlorobenzene
Concen: 50.991 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050



Tgt Ion:146 Resp: 220370
Ion Ratio Lower Upper
146 100
111 44.0 22.0 66.0
148 64.4 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 49.657 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 75 Resp: 30011
Ion Ratio Lower Upper
75 100
155 79.3 38.1 114.5
157 102.7 49.6 149.0





#93

1,2,4-Trichlorobenzene

Concen: 51.373 ug/l

RT: 15.388 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050

Tgt Ion:180 Resp: 109350

Ion Ratio Lower Upper

180 100

182 94.6 47.0 141.0





145 32.7 16.4 49.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#94

Hexachlorobutadiene

Concen: 47.838 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38




Tgt Ion:225 Resp: 50783


Ion Ratio Lower Upper

225 100

223 60.9 32.4 97.2

227 62.5 32.4 97.0









#95
Naphthalene
Concen: 51.214 ug/l
RT: 15.641 min Scan# 2326
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion:128 Resp: 36141
Ion Ratio Lower Upper
128 100
127 13.1 10.5 15.7
129 11.2 8.7 13.1




Manual Integrations
APPROVED

Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#96
1,2,3-Trichlorobenzene
Concen: 52.021 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion:180 Resp: 110846
Ion Ratio Lower Upper
180 100
182 95.2 48.2 144.6
145 34.9 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084428.D
Acq On : 21 Oct 2024 11:38
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 22 01:29:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.591	0.529	10.5	77	0.00
3 P	Chloromethane	0.675	0.602	10.8	80	0.00
4 C	Vinyl Chloride	0.654	0.597	8.7#	80	0.00
5 T	Bromomethane	0.431	0.370	14.2	77	-0.02
6 T	Chloroethane	0.468	0.401	14.3	86	-0.01
7 T	Trichlorofluoromethane	1.019	1.044	-2.5	88	0.00
8 T	Diethyl Ether	0.378	0.377	0.3	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.584	0.603	-3.3	92	0.00
10 T	Methyl Iodide	0.752	0.739	1.7	82	-0.01
11 T	Tert butyl alcohol	0.122	0.120	1.6	87	0.00
12 CM	1,1-Dichloroethene	0.563	0.566	-0.5#	87	0.00
13 T	Acrolein	0.140	0.108	22.9	68	0.00
14 T	Allyl chloride	0.978	0.853	12.8	77	0.00
15 T	Acrylonitrile	0.309	0.350	-13.3	99	0.00
16 T	Acetone	0.318	0.317	0.3	83	0.00
17 T	Carbon Disulfide	1.782	1.424	20.1	74	0.00
18 T	Methyl Acetate	0.694	0.915	-31.8#	116	0.00
19 T	Methyl tert-butyl Ether	1.900	2.031	-6.9	90	0.00
20 T	Methylene Chloride	0.647	0.671	-3.7	92	0.00
21 T	trans-1,2-Dichloroethene	0.590	0.593	-0.5	86	0.00
22 T	Diisopropyl ether	2.022	2.142	-5.9	89	0.00
23 T	Vinyl Acetate	1.500	1.514	-0.9	82	0.00
24 P	1,1-Dichloroethane	1.131	1.189	-5.1	89	0.00
25 T	2-Butanone	0.434	0.467	-7.6	93	0.00
26 T	2,2-Dichloropropane	1.020	1.014	0.6	84	0.00
27 T	cis-1,2-Dichloroethene	0.713	0.752	-5.5	91	0.00
28 T	Bromochloromethane	0.505	0.533	-5.5	97	0.00
29 T	Tetrahydrofuran	0.268	0.303	-13.1	96	0.00
30 C	Chloroform	1.175	1.280	-8.9#	95	0.00
31 T	Cyclohexane	1.098	0.959	12.7	81	0.00
32 T	1,1,1-Trichloroethane	1.055	1.115	-5.7	91	0.00
33 S	1,2-Dichloroethane-d4	0.741	0.791	-6.7	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	86	0.00
35 S	Dibromofluoromethane	0.330	0.387	-17.3	103	0.00
36 T	1,1-Dichloropropene	0.479	0.498	-4.0	85	0.00
37 T	Ethyl Acetate	0.491	0.522	-6.3	90	0.00
38 T	Carbon Tetrachloride	0.525	0.569	-8.4	90	0.00
39 T	Methylcyclohexane	0.533	0.522	2.1	77	0.00
40 TM	Benzene	1.492	1.635	-9.6	91	0.00
41 T	Methacrylonitrile	0.267	0.301	-12.7	91	0.00
42 TM	1,2-Dichloroethane	0.506	0.549	-8.5	90	0.00
43 T	Isopropyl Acetate	0.962	0.889	7.6	86	0.00
44 TM	Trichloroethene	0.348	0.375	-7.8	89	0.00
45 C	1,2-Dichloropropane	0.353	0.401	-13.6#	92	0.00
46 T	Dibromomethane	0.238	0.284	-19.3	95	0.00
47 T	Bromodichloromethane	0.529	0.606	-14.6	94	0.00
48 T	Methyl methacrylate	0.383	0.429	-12.0	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084428.D
 Acq On : 21 Oct 2024 11:38
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 22 01:29:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.009	-28.6#	105	0.00
50 S	Toluene-d8	1.213	1.431	-18.0	100	0.00
51 T	4-Methyl-2-Pentanone	0.468	0.576	-23.1	98	0.00
52 CM	Toluene	0.910	1.023	-12.4#	91	0.00
53 T	t-1,3-Dichloropropene	0.539	0.601	-11.5	88	0.00
54 T	cis-1,3-Dichloropropene	0.580	0.651	-12.2	88	0.00
55 T	1,1,2-Trichloroethane	0.326	0.397	-21.8	98	0.00
56 T	Ethyl methacrylate	0.537	0.626	-16.6	90	0.00
57 T	1,3-Dichloropropane	0.583	0.672	-15.3	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.249	0.303	-21.7	93	0.00
59 T	2-Hexanone	0.349	0.426	-22.1	95	0.00
60 T	Dibromochloromethane	0.385	0.466	-21.0	96	0.00
61 T	1,2-Dibromoethane	0.339	0.388	-14.5	94	0.00
62 S	4-Bromofluorobenzene	0.442	0.522	-18.1	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.348	0.366	-5.2	94	0.00
65 PM	Chlorobenzene	1.109	1.181	-6.5	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.382	0.424	-11.0	97	0.00
67 C	Ethyl Benzene	1.961	2.070	-5.6#	90	0.00
68 T	m/p-Xylenes	0.725	0.807	-11.3	92	0.00
69 T	o-Xylene	0.686	0.765	-11.5	91	0.00
70 T	Styrene	1.162	1.354	-16.5	95	0.00
71 P	Bromoform	0.281	0.332	-18.1	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.815	3.797	0.5	92	0.00
74 T	N-amyl acetate	1.728	1.601	7.3	86	0.00
75 P	1,1,2,2-Tetrachloroethane	1.147	1.198	-4.4	101	0.00
76 T	1,2,3-Trichloropropane	1.021	0.997	2.4	89	0.00
77 T	Bromobenzene	0.920	0.932	-1.3	98	0.00
78 T	n-propylbenzene	4.421	4.518	-2.2	92	0.00
79 T	2-Chlorotoluene	2.821	2.779	1.5	92	0.00
80 T	1,3,5-Trimethylbenzene	3.112	3.265	-4.9	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.426	0.458	-7.5	102	0.00
82 T	4-Chlorotoluene	2.840	2.855	-0.5	94	0.00
83 T	tert-Butylbenzene	2.766	2.759	0.3	91	0.00
84 T	1,2,4-Trimethylbenzene	3.135	3.306	-5.5	94	0.00
85 T	sec-Butylbenzene	3.691	3.856	-4.5	93	0.00
86 T	p-Isopropyltoluene	3.052	3.230	-5.8	92	0.00
87 T	1,3-Dichlorobenzene	1.727	1.778	-3.0	99	0.00
88 T	1,4-Dichlorobenzene	1.744	1.755	-0.6	98	0.00
89 T	n-Butylbenzene	2.768	2.740	1.0	89	0.00
90 T	Hexachloroethane	0.620	0.597	3.7	93	0.00
91 T	1,2-Dichlorobenzene	1.693	1.727	-2.0	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.237	0.235	0.8	99	0.00
93 T	1,2,4-Trichlorobenzene	0.834	0.857	-2.8	92	0.00
94 T	Hexachlorobutadiene	0.416	0.398	4.3	100	0.00
95 T	Naphthalene	2.764	2.832	-2.5	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084428.D
Acq On : 21 Oct 2024 11:38
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 22 01:29:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.835	0.868	-4.0	97	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084428.D
 Acq On : 21 Oct 2024 11:38
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 22 01:29:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	50.000	44.759	10.5	77	0.00
3 P	Chloromethane	50.000	44.550	10.9	80	0.00
4 C	Vinyl Chloride	50.000	45.660	8.7#	80	0.00
5 T	Bromomethane	50.000	42.866	14.3	77	-0.02
6 T	Chloroethane	50.000	42.808	14.4	86	-0.01
7 T	Trichlorofluoromethane	50.000	51.224	-2.4	88	0.00
8 T	Diethyl Ether	50.000	49.902	0.2	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.670	-3.3	92	0.00
10 T	Methyl Iodide	50.000	49.171	1.7	82	-0.01
11 T	Tert butyl alcohol	250.000	245.606	1.8	87	0.00
12 CM	1,1-Dichloroethene	50.000	50.300	-0.6#	87	0.00
13 T	Acrolein	250.000	192.843	22.9	68	0.00
14 T	Allyl chloride	50.000	43.643	12.7	77	0.00
15 T	Acrylonitrile	250.000	283.386	-13.4	99	0.00
16 T	Acetone	250.000	248.905	0.4	83	0.00
17 T	Carbon Disulfide	50.000	39.937	20.1	74	0.00
18 T	Methyl Acetate	50.000	65.900	-31.8#	116	0.00
19 T	Methyl tert-butyl Ether	50.000	53.457	-6.9	90	0.00
20 T	Methylene Chloride	50.000	51.787	-3.6	92	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.278	-0.6	86	0.00
22 T	Diisopropyl ether	50.000	52.949	-5.9	89	0.00
23 T	Vinyl Acetate	250.000	252.258	-0.9	82	0.00
24 P	1,1-Dichloroethane	50.000	52.570	-5.1	89	0.00
25 T	2-Butanone	250.000	269.013	-7.6	93	0.00
26 T	2,2-Dichloropropane	50.000	49.738	0.5	84	0.00
27 T	cis-1,2-Dichloroethene	50.000	52.728	-5.5	91	0.00
28 T	Bromochloromethane	50.000	52.836	-5.7	97	0.00
29 T	Tetrahydrofuran	250.000	283.258	-13.3	96	0.00
30 C	Chloroform	50.000	54.476	-9.0#	95	0.00
31 T	Cyclohexane	50.000	43.659	12.7	81	0.00
32 T	1,1,1-Trichloroethane	50.000	52.830	-5.7	91	0.00
33 S	1,2-Dichloroethane-d4	50.000	53.365	-6.7	96	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	86	0.00
35 S	Dibromofluoromethane	50.000	58.729	-17.5	103	0.00
36 T	1,1-Dichloropropene	50.000	52.021	-4.0	85	0.00
37 T	Ethyl Acetate	50.000	53.172	-6.3	90	0.00
38 T	Carbon Tetrachloride	50.000	54.202	-8.4	90	0.00
39 T	Methylcyclohexane	50.000	48.955	2.1	77	0.00
40 TM	Benzene	50.000	54.796	-9.6	91	0.00
41 T	Methacrylonitrile	50.000	56.424	-12.8	91	0.00
42 TM	1,2-Dichloroethane	50.000	54.240	-8.5	90	0.00
43 T	Isopropyl Acetate	50.000	46.197	7.6	86	0.00
44 TM	Trichloroethene	50.000	53.784	-7.6	89	0.00
45 C	1,2-Dichloropropane	50.000	56.873	-13.7#	92	0.00
46 T	Dibromomethane	50.000	59.706	-19.4	95	0.00
47 T	Bromodichloromethane	50.000	57.235	-14.5	94	0.00
48 T	Methyl methacrylate	50.000	55.960	-11.9	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084428.D
 Acq On : 21 Oct 2024 11:38
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 22 01:29:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1259.594	-26.0#	105	0.00
50 S	Toluene-d8	50.000	59.014	-18.0	100	0.00
51 T	4-Methyl-2-Pentanone	250.000	308.066	-23.2	98	0.00
52 CM	Toluene	50.000	56.231	-12.5#	91	0.00
53 T	t-1,3-Dichloropropene	50.000	55.712	-11.4	88	0.00
54 T	cis-1,3-Dichloropropene	50.000	56.089	-12.2	88	0.00
55 T	1,1,2-Trichloroethane	50.000	60.910	-21.8	98	0.00
56 T	Ethyl methacrylate	50.000	58.283	-16.6	90	0.00
57 T	1,3-Dichloropropane	50.000	57.670	-15.3	94	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	303.996	-21.6	93	0.00
59 T	2-Hexanone	250.000	304.908	-22.0	95	0.00
60 T	Dibromochloromethane	50.000	60.458	-20.9	96	0.00
61 T	1,2-Dibromoethane	50.000	57.274	-14.5	94	0.00
62 S	4-Bromofluorobenzene	50.000	59.102	-18.2	100	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	92	0.00
64 T	Tetrachloroethene	50.000	52.523	-5.0	94	0.00
65 PM	Chlorobenzene	50.000	53.283	-6.6	93	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	55.468	-10.9	97	0.00
67 C	Ethyl Benzene	50.000	52.799	-5.6#	90	0.00
68 T	m/p-Xylenes	100.000	111.301	-11.3	92	0.00
69 T	o-Xylene	50.000	55.767	-11.5	91	0.00
70 T	Styrene	50.000	58.246	-16.5	95	0.00
71 P	Bromoform	50.000	59.072	-18.1	97	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	100	0.00
73 T	Isopropylbenzene	50.000	49.760	0.5	92	0.00
74 T	N-amyl acetate	50.000	46.332	7.3	86	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	52.226	-4.5	101	0.00
76 T	1,2,3-Trichloropropane	50.000	48.846	2.3	89	0.00
77 T	Bromobenzene	50.000	50.631	-1.3	98	0.00
78 T	n-propylbenzene	50.000	51.101	-2.2	92	0.00
79 T	2-Chlorotoluene	50.000	49.256	1.5	92	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.460	-4.9	93	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	53.835	-7.7	102	0.00
82 T	4-Chlorotoluene	50.000	50.275	-0.5	94	0.00
83 T	tert-Butylbenzene	50.000	49.862	0.3	91	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.722	-5.4	94	0.00
85 T	sec-Butylbenzene	50.000	52.226	-4.5	93	0.00
86 T	p-Isopropyltoluene	50.000	52.906	-5.8	92	0.00
87 T	1,3-Dichlorobenzene	50.000	51.493	-3.0	99	0.00
88 T	1,4-Dichlorobenzene	50.000	50.322	-0.6	98	0.00
89 T	n-Butylbenzene	50.000	49.495	1.0	89	0.00
90 T	Hexachloroethane	50.000	48.204	3.6	93	0.00
91 T	1,2-Dichlorobenzene	50.000	50.991	-2.0	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.657	0.7	99	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.373	-2.7	92	0.00
94 T	Hexachlorobutadiene	50.000	47.838	4.3	100	0.00
95 T	Naphthalene	50.000	51.214	-2.4	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084428.D
Acq On : 21 Oct 2024 11:38
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 22 01:29:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.021	-4.0	97	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG No.: P4440

Instrument ID: MSVOA_N Calibration Date/Time: 10/21/2024 20:04

Lab File ID: VN084448.D Init. Calib. Date(s): 09/30/2024 09/30/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:25 14:48

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.675	0.575	0.1	-14.81	50
Vinyl Chloride	0.654	0.580		-11.31	50
Bromomethane	0.431	0.354		-17.86	50
Chloroethane	0.468	0.379		-19.02	50
Trichlorofluoromethane	1.019	0.994		-2.45	50
1,1,2-Trichlorotrifluoroethane	0.584	0.582		-0.34	50
1,1-Dichloroethene	0.563	0.556		-1.24	50
Acetone	0.318	0.255		-19.81	50
Carbon Disulfide	1.782	1.354		-24.02	50
Methyl tert-butyl Ether	1.900	1.972		3.79	50
Methylene Chloride	0.647	0.658		1.7	50
trans-1,2-Dichloroethene	0.590	0.582		-1.36	50
1,1-Dichloroethane	1.131	1.157	0.1	2.3	50
2-Butanone	0.434	0.406		-6.45	50
Carbon Tetrachloride	0.525	0.534		1.71	50
cis-1,2-Dichloroethene	0.713	0.722		1.26	50
Chloroform	1.175	1.226		4.34	50
1,1,1-Trichloroethane	1.055	1.082		2.56	50
Methylcyclohexane	0.533	0.493		-7.51	50
Benzene	1.492	1.520		1.88	50
1,2-Dichloroethane	0.506	0.507		0.2	50
Trichloroethene	0.348	0.353		1.44	50
1,2-Dichloropropane	0.353	0.379		7.36	50
Bromodichloromethane	0.529	0.566		6.99	50
4-Methyl-2-Pentanone	0.468	0.503		7.48	50
Toluene	0.910	0.953		4.72	50
t-1,3-Dichloropropene	0.539	0.551		2.23	50
cis-1,3-Dichloropropene	0.580	0.604		4.14	50
1,1,2-Trichloroethane	0.326	0.361		10.74	50
2-Hexanone	0.349	0.364		4.3	50
Dibromochloromethane	0.385	0.434		12.73	50
Tetrachloroethene	0.348	0.341		-2.01	50
Chlorobenzene	1.109	1.111	0.3	0.18	50
Ethyl Benzene	1.961	2.021		3.06	50
m/p-Xylenes	0.725	0.758		4.55	50
o-Xylene	0.686	0.734		7	50
Styrene	1.162	1.285		10.59	50
Bromoform	0.281	0.306	0.1	8.9	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG No.: P4440

Instrument ID: MSVOA_N Calibration Date/Time: 10/21/2024 20:04

Lab File ID: VN084448.D Init. Calib. Date(s): 09/30/2024 09/30/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:25 14:48

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.815	3.702		-2.96	50
1,1,2,2-Tetrachloroethane	1.147	1.151	0.3	0.35	50
1,3-Dichlorobenzene	1.727	1.669		-3.36	50
1,4-Dichlorobenzene	1.744	1.676		-3.9	50
1,2-Dichlorobenzene	1.693	1.649		-2.6	50
1,2-Dichloroethane-d4	0.741	0.710		-4.18	50
Dibromofluoromethane	0.330	0.343		3.94	50
Toluene-d8	1.213	1.259		3.79	50
4-Bromofluorobenzene	0.442	0.461		4.3	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084448.D
 Acq On : 21 Oct 2024 20:04
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/22/2024
 Supervised By : Mahesh Dadoda 10/22/2024

Quant Time: Oct 22 01:42:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	164267	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	281823	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	256750	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	130386	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	116690	47.911	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 95.820%			
35) Dibromofluoromethane	8.165	113	96698	52.045	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.100%			
50) Toluene-d8	10.565	98	354860	51.913	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.820%			
62) 4-Bromofluorobenzene	12.847	95	129892	52.160	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 104.320%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	84821	43.654	ug/l	94
3) Chloromethane	2.359	50	94401	42.550	ug/l	96
4) Vinyl Chloride	2.512	62	95237	44.316	ug/l	100
5) Bromomethane	2.942	94	58167	41.035	ug/l	96
6) Chloroethane	3.112	64	62275	40.462	ug/l	98
7) Trichlorofluoromethane	3.495	101	163261	48.786	ug/l	97
8) Diethyl Ether	3.959	74	60533	48.772	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.365	101	95561	49.837	ug/l	99
10) Methyl Iodide	4.589	142	118118	47.834	ug/l	98
11) Tert butyl alcohol	5.518	59	84056	209.300	ug/l	100
12) 1,1-Dichloroethene	4.336	96	91278	49.352	ug/l	98
13) Acrolein	4.183	56	86128	186.809	ug/l	98
14) Allyl chloride	5.024	41	140602	43.772	ug/l	98
15) Acrylonitrile	5.718	53	264703	261.006	ug/l	98
16) Acetone	4.430	43	209499	200.235	ug/l	97
17) Carbon Disulfide	4.712	76	222486	37.994	ug/l	99
18) Methyl Acetate	5.024	43	135697	59.505	ug/l	100
19) Methyl tert-butyl Ether	5.800	73	323965	51.897	ug/l	99
20) Methylene Chloride	5.277	84	108136	50.844	ug/l	93
21) trans-1,2-Dichloroethene	5.783	96	95558	49.315	ug/l	98
22) Diisopropyl ether	6.671	45	344689	51.879	ug/l	99
23) Vinyl Acetate	6.600	43	1175888	238.618	ug/l	100
24) 1,1-Dichloroethane	6.571	63	190033	51.145	ug/l	98
25) 2-Butanone	7.483	43	333066	233.725	ug/l	99
26) 2,2-Dichloropropane	7.494	77	149770	44.706	ug/l	97
27) cis-1,2-Dichloroethene	7.488	96	118609	50.655	ug/l	97
28) Bromochloromethane	7.812	49	78703	47.474	ug/l	97
29) Tetrahydrofuran	7.841	42	221138	251.461	ug/l	97
30) Chloroform	7.965	83	201338	52.167	ug/l	100
31) Cyclohexane	8.259	56	153399	42.521	ug/l	99
32) 1,1,1-Trichloroethane	8.171	97	177818	51.300	ug/l	99
36) 1,1-Dichloropropene	8.371	75	133081	49.312	ug/l	98
37) Ethyl Acetate	7.559	43	133124	48.082	ug/l	99
38) Carbon Tetrachloride	8.359	117	150440	50.818	ug/l	97
39) Methylcyclohexane	9.600	83	138930	46.254	ug/l	99
40) Benzene	8.606	78	428332	50.941	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084448.D
 Acq On : 21 Oct 2024 20:04
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :John Carlone 10/22/2024
 Supervised By :Mahesh Dadoda 10/22/2024

Quant Time: Oct 22 01:42:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.783	41	76939	51.187	ug/l	97
42) 1,2-Dichloroethane	8.671	62	142985	50.160	ug/l	98
43) Isopropyl Acetate	8.688	43	229980	42.408	ug/l	99
44) Trichloroethene	9.353	130	99355	50.592	ug/l	99
45) 1,2-Dichloropropane	9.618	63	106710	53.648	ug/l	97
46) Dibromomethane	9.706	93	74783	55.712	ug/l	99
47) Bromodichloromethane	9.888	83	159633	53.528	ug/l	98
48) Methyl methacrylate	9.682	41	109772	50.850	ug/l	99
49) 1,4-Dioxane	9.694	88	39543	994.968	ug/l	97
51) 4-Methyl-2-Pentanone	10.447	43	708662	268.874	ug/l	100
52) Toluene	10.629	92	268587	52.382	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	155308	51.089	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	170218	52.057	ug/l	98
55) 1,1,2-Trichloroethane	11.018	97	101779	55.371	ug/l	98
56) Ethyl methacrylate	10.871	69	165318	54.658	ug/l	97
57) 1,3-Dichloropropane	11.159	76	175471	53.424	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	390411	278.254	ug/l	98
59) 2-Hexanone	11.194	43	512990	260.410	ug/l	99
60) Dibromochloromethane	11.359	129	122224	56.277	ug/l	100
61) 1,2-Dibromoethane	11.471	107	100984	52.872	ug/l	100
64) Tetrachloroethene	11.100	164	87525	48.943	ug/l	97
65) Chlorobenzene	11.888	112	285234	50.105	ug/l	95
66) 1,1,1,2-Tetrachloroethane	11.959	131	104070	53.075	ug/l	99
67) Ethyl Benzene	11.965	91	518810	51.533	ug/l	100
68) m/p-Xylenes	12.070	106	389131	104.492	ug/l	100
69) o-Xylene	12.400	106	188510	53.503	ug/l	99
70) Styrene	12.412	104	329833	55.268	ug/l	99
71) Bromoform	12.576	173	78482	54.365	ug/l #	97
73) Isopropylbenzene	12.694	105	482626	48.507	ug/l	100
74) N-amyl acetate	12.494	43	205514	45.614	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.935	83	150053	50.155	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	140435m	52.758	ug/l	
77) Bromobenzene	12.976	156	118173	49.239	ug/l	94
78) n-propylbenzene	13.035	91	573481	49.748	ug/l	99
79) 2-Chlorotoluene	13.123	91	356393	48.450	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	411332	50.681	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	55404	49.901	ug/l	89
82) 4-Chlorotoluene	13.223	91	360840	48.730	ug/l	99
83) tert-Butylbenzene	13.435	119	352909	48.923	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	416913	50.992	ug/l	100
85) sec-Butylbenzene	13.617	105	482488	50.124	ug/l	100
86) p-Isopropyltoluene	13.729	119	404005	50.755	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	217666	48.345	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	218499	48.056	ug/l	99
89) n-Butylbenzene	14.053	91	342382	47.437	ug/l	99
90) Hexachloroethane	14.335	117	74567	46.151	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	214956	48.690	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	28592	46.312	ug/l	97
93) 1,2,4-Trichlorobenzene	15.394	180	105156	48.359	ug/l	99
94) Hexachlorobutadiene	15.500	225	47024	43.364	ug/l	98
95) Naphthalene	15.641	128	338368	46.938	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	103615	47.602	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084448.D
Acq On : 21 Oct 2024 20:04
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024

Quant Time: Oct 22 01:42:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084448.D
Acq On : 21 Oct 2024 20:04
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Quant Time: Oct 22 01:42:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

Instrument :

MSVOA_N

ClientSampleId :


VSTDCCC050EC


Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024


Supervised By :Mahesh Dadoda 10/22/2024





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:168 Resp: 16426
Ion Ratio Lower Upper
168 100
99 63.4 54.2 81.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#2
Dichlorodifluoromethane
Concen: 43.654 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04


Tgt Ion: 85 Resp: 84821
Ion Ratio Lower Upper
85 100
87 31.8 17.6 52.9





#3
Chloromethane
Concen: 42.550 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 50 Resp: 9440
Ion Ratio Lower Upper
50 100
52 31.1 26.8 40.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#4
Vinyl Chloride
Concen: 44.316 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04


Tgt Ion: 62 Resp: 95237
Ion Ratio Lower Upper
62 100
64 32.8 26.2 39.4





#5
Bromomethane
Concen: 41.035 ug/l
RT: 2.942 min Scan# 1
Delta R.T. -0.012 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 94 Resp: 5816
Ion Ratio Lower Upper
94 100
96 87.6 72.9 109.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#6
Chloroethane
Concen: 40.462 ug/l
RT: 3.112 min Scan# 197
Delta R.T. -0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04


Tgt Ion: 64 Resp: 62275
Ion Ratio Lower Upper
64 100
66 30.6 23.5 35.3





#7
Trichlorofluoromethane
Concen: 48.786 ug/l
RT: 3.495 min Scan# 262
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:101 Resp: 16326
Ion Ratio Lower Upper
101 100
103 67.3 51.8 77.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#8
Diethyl Ether
Concen: 48.772 ug/l
RT: 3.959 min Scan# 341
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 74 Resp: 60533
Ion Ratio Lower Upper
74 100
45 94.9 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.837 ug/l
RT: 4.365 min Scan# 411
Delta R.T. -0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 9556
Ion Ratio Lower Upper
101 100
85 44.4 36.2 54.4
151 74.8 59.3 88.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#10
Methyl Iodide
Concen: 47.834 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion:142 Resp: 118118
Ion Ratio Lower Upper
142 100
127 45.8 35.3 52.9
141 14.6 11.0 16.6





#11

Tert butyl alcohol

Concen: 209.300 ug/l

RT: 5.518 min Scan# 607

Delta R.T. -0.006 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 59 Resp: 8405

Ion Ratio Lower Upper

59 100





57 10.3 8.2 12.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#12

1,1-Dichloroethene

Concen: 49.352 ug/l

RT: 4.336 min Scan# 405

Delta R.T. -0.006 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04




Tgt Ion: 96 Resp: 91278


Ion Ratio Lower Upper

96 100

61 164.5 128.6 192.8

98 65.0 51.7 77.5





#13

Acrolein

Concen: 186.809 ug/l

RT: 4.183 min Scan# 31

Delta R.T. 0.006 min

Lab File: VN084448.D


Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 56 Resp: 86128

Ion Ratio Lower Upper

56 100



55 73.1 57.2 85.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#14

Allyl chloride

Concen: 43.772 ug/l

RT: 5.024 min Scan# 522

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04




Tgt Ion: 41 Resp: 140602


Ion Ratio Lower Upper

41 100

39 71.5 59.0 88.6


76 38.3 29.8 44.6





#15
Acrylonitrile
Concen: 261.006 ug/l
RT: 5.718 min Scan# 640
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 53 Resp: 26470
Ion Ratio Lower Upper
53 100
52 80.3 66.2 99.4
51 35.5 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#16
Acetone
Concen: 200.235 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 43 Resp: 209499
Ion Ratio Lower Upper
43 100
58 32.1 24.4 36.6





#17

Carbon Disulfide

Concen: 37.994 ug/l

RT: 4.712 min Scan# 469

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 22248

Ion Ratio Lower Upper

76 100





78 9.5 7.3 10.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#18

Methyl Acetate

Concen: 59.505 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN084448.D




Acq: 21 Oct 2024 20:04


Tgt Ion: 43 Resp: 135697

Ion Ratio Lower Upper

43 100

74 24.0 19.2 28.8









#19
Methyl tert-butyl Ether
Concen: 51.897 ug/l
RT: 5.800 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 73 Resp: 32396
Ion Ratio Lower Upper
73 100
57 21.6 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#20
Methylene Chloride
Concen: 50.844 ug/l
RT: 5.277 min Scan# 565
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 84 Resp: 108136
Ion Ratio Lower Upper
84 100
49 120.3 107.2 160.8
51 37.7 31.8 47.8
86 66.1 52.9 79.3





#21

trans-1,2-Dichloroethene

Concen: 49.315 ug/l

RT: 5.783 min Scan# 61

Delta R.T. -0.006 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 96 Resp: 9555

Ion Ratio Lower Upper

96 100

61 138.0 108.4 162.6





98 63.5 49.8 74.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#22

Diisopropyl ether

Concen: 51.879 ug/l

RT: 6.671 min Scan# 802

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Tgt Ion: 45 Resp: 344689




Ion Ratio Lower Upper


45 100

43 52.9 42.3 63.5

87 28.0 21.8 32.6

59 11.4 9.6 14.4









#23
Vinyl Acetate
Concen: 238.618 ug/l
RT: 6.600 min Scan# 790
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 117588
Ion Ratio Lower Upper
43 100
86 10.7 8.6 12.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#24
1,1-Dichloroethane
Concen: 51.145 ug/l
RT: 6.571 min Scan# 785
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 63 Resp: 190033
Ion Ratio Lower Upper
63 100
98 7.3 3.4 10.2
100 4.5 2.0 5.8





#25

2-Butanone

Concen: 233.725 ug/l

RT: 7.483 min Scan# 940

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 333060

Ion Ratio Lower Upper

43 100





72 26.4 21.4 32.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#26

2,2-Dichloropropane

Concen: 44.706 ug/l

RT: 7.494 min Scan# 942

Delta R.T. 0.006 min

Lab File: VN084448.D




Acq: 21 Oct 2024 20:04


Tgt Ion: 77 Resp: 149770

Ion Ratio Lower Upper

77 100

97 23.0 10.8 32.3





#27

cis-1,2-Dichloroethene

Concen: 50.655 ug/l

RT: 7.488 min Scan# 941

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 96 Resp: 118609

Ion Ratio Lower Upper

96 100

61 145.2 0.0 301.8





98 64.1 0.0 128.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#28

Bromochloromethane

Concen: 47.474 ug/l

RT: 7.812 min Scan# 996

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04




Tgt Ion: 49 Resp: 78703


Ion Ratio Lower Upper

49 100

129 1.5 0.0 3.6

130 75.1 58.0 87.0





#29

Tetrahydrofuran

Concen: 251.461 ug/l

RT: 7.841 min Scan# 1001

Delta R.T. 0.006 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 42 Resp: 221138

Ion Ratio Lower Upper

42 100

72 45.3 34.6 51.8





71 41.7 32.2 48.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#30

Chloroform

Concen: 52.167 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. -0.000 min

Lab File: VN084448.D




Acq: 21 Oct 2024 20:04


Tgt Ion: 83 Resp: 201338

Ion Ratio Lower Upper

83 100


85 63.4 51.0 76.4





#31
Cyclohexane
Concen: 42.521 ug/l
RT: 8.259 min Scan# 1072
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 153395
Ion Ratio Lower Upper
56 100
69 29.9 23.8 35.6
84 84.5 68.8 103.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#32
1,1,1-Trichloroethane
Concen: 51.300 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 97 Resp: 177818
Ion Ratio Lower Upper
97 100
99 63.3 51.2 76.8
61 46.9 37.5 56.3





#33

1,2-Dichloroethane-d4

Concen: 47.911 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 116690

Ion Ratio Lower Upper

65 100





67 53.9 0.0 102.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04




Tgt Ion:114 Resp: 281823


Ion Ratio Lower Upper

114 100

63 21.6 0.0 43.8

88 16.3 0.0 31.6





#35

Dibromofluoromethane

Concen: 52.045 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion:113 Resp: 96698

Ion Ratio Lower Upper

113 100

111 102.0 83.3 124.9





192 17.6 13.5 20.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#36

1,1-Dichloropropene

Concen: 49.312 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04




Tgt Ion: 75 Resp: 133081


Ion Ratio Lower Upper

75 100

110 35.1 16.9 50.7

77 31.9 24.7 37.1





#37

Ethyl Acetate

Concen: 48.082 ug/l

RT: 7.559 min Scan# 91

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 133124

Ion Ratio Lower Upper

43 100

61 14.1 10.7 16.1





70 11.1 8.8 13.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#38

Carbon Tetrachloride

Concen: 50.818 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04




Tgt Ion:117 Resp: 150440


Ion Ratio Lower Upper

117 100

119 94.0 78.7 118.1

121 30.4 24.4 36.6





#39

Methylcyclohexane

Concen: 46.254 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 138930

Ion Ratio Lower Upper

83 100

55 79.2 62.8 94.2





98 49.1 38.8 58.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#40

Benzene

Concen: 50.941 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN084448.D




Acq: 21 Oct 2024 20:04


Tgt Ion: 78 Resp: 428332

Ion Ratio Lower Upper

78 100


77 24.4 19.1 28.7





#41
Methacrylonitrile
Concen: 51.187 ug/l
RT: 7.783 min Scan# 991
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 41 Resp: 76939
Ion Ratio Lower Upper
41 100
39 53.7 43.3 64.9
67 74.5 56.1 84.1
52 31.6 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#42
1,2-Dichloroethane
Concen: 50.160 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 62 Resp: 142985
Ion Ratio Lower Upper
62 100
98 9.9 0.0 18.2





#43

Isopropyl Acetate

Concen: 42.408 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:	43	Resp:	229980
Ion	Ratio	Lower	Upper
43	100		
61	26.3	20.4	30.6
87	13.5	10.3	15.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#44

Trichloroethene

Concen: 50.592 ug/l




RT: 9.353 min Scan# 1258


Delta R.T. 0.006 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Tgt Ion:	130	Resp:	99355
Ion	Ratio	Lower	Upper
130	100		
95	100.8	0.0	203.0





#45

1,2-Dichloropropane

Concen: 53.648 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 106710

Ion Ratio Lower Upper

63 100





65 29.3 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#46

Dibromomethane

Concen: 55.712 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04




Tgt Ion: 93 Resp: 74783


Ion Ratio Lower Upper

93 100

95 81.7 65.0 97.4

174 90.5 71.8 107.8





#47

Bromodichloromethane

Concen: 53.528 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 15963

Ion Ratio Lower Upper

83 100

85 66.1 51.5 77.3





127 8.3 5.7 8.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#48

Methyl methacrylate

Concen: 50.850 ug/l

RT: 9.682 min Scan# 1314

Delta R.T. 0.006 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04




Tgt Ion: 41 Resp: 109772


Ion Ratio Lower Upper

41 100

69 86.0 67.1 100.7

39 57.3 45.8 68.6





#49

1,4-Dioxane

Concen: 994.968 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 88 Resp: 3954

Ion Ratio Lower Upper

88 100

43 30.1 25.3 37.9





58 73.5 56.4 84.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#50

Toluene-d8

Concen: 51.913 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN084448.D




Acq: 21 Oct 2024 20:04


Tgt Ion: 98 Resp: 354860

Ion Ratio Lower Upper

98 100

100 65.6 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 268.874 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 708662
Ion Ratio Lower Upper
43 100
58 39.7 31.6 47.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#52
Toluene
Concen: 52.382 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04


Tgt Ion: 92 Resp: 268587
Ion Ratio Lower Upper
92 100
91 171.7 137.3 205.9





#53
t-1,3-Dichloropropene
Concen: 51.089 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 155308
Ion Ratio Lower Upper
75 100
77 32.2 25.4 38.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#54
cis-1,3-Dichloropropene
Concen: 52.057 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04


Tgt Ion: 75 Resp: 170218
Ion Ratio Lower Upper
75 100
77 32.5 24.2 36.2
39 48.0 38.7 58.1





#55
1,1,2-Trichloroethane
Concen: 55.371 ug/l
RT: 11.018 min Scan# 1540
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 97 Resp: 101779
Ion Ratio Lower Upper
97 100
83 88.2 71.0 106.4
85 54.5 41.7 62.5
99 61.1 50.1 75.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024




#56
Ethyl methacrylate
Concen: 54.658 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04



Tgt Ion: 69 Resp: 165318
Ion Ratio Lower Upper
69 100
41 64.1 53.3 79.9
39 34.8 28.8 43.2





#57

1,3-Dichloropropane

Concen: 53.424 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. -0.000 min

Lab File: VN084448.D


Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC



Tgt Ion: 76 Resp: 17547

Ion Ratio Lower Upper

76 100




78 32.5 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#58

2-Chloroethyl Vinyl ether


Concen: 278.254 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04






Tgt Ion: 63 Resp: 390411

Ion Ratio Lower Upper

63 100


106 28.2 21.8 32.6





#59
2-Hexanone
Concen: 260.410 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 512990
Ion Ratio Lower Upper
43 100
58 55.2 27.3 81.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#60
Dibromochloromethane
Concen: 56.277 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04


Tgt Ion:129 Resp: 122224
Ion Ratio Lower Upper
129 100
127 76.7 38.3 114.9





#61
1,2-Dibromoethane
Concen: 52.872 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:107 Resp: 100984
Ion Ratio Lower Upper
107 100
109 94.3 75.4 113.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#62
4-Bromofluorobenzene
Concen: 52.160 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 95 Resp: 129892
Ion Ratio Lower Upper
95 100
174 74.2 0.0 145.2
176 72.0 0.0 140.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion:117 Resp: 256750

Ion Ratio Lower Upper

117 100

82 55.9 47.2 70.8





119 32.6 25.4 38.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#64

Tetrachloroethene

Concen: 48.943 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Tgt Ion:164 Resp: 87525




Ion Ratio Lower Upper


164 100

166 121.2 101.0 151.4

129 99.9 80.8 121.2


131 95.2 78.1 117.1





#65
Chlorobenzene
Concen: 50.105 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:112 Resp: 285234
Ion Ratio Lower Upper
112 100
114 30.9 27.0 40.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 53.075 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion:131 Resp: 104070
Ion Ratio Lower Upper
131 100
133 94.9 47.9 143.7
119 64.4 32.8 98.3









#67
Ethyl Benzene
Concen: 51.533 ug/l
RT: 11.965 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 518810
Ion Ratio Lower Upper
91 100
106 30.6 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#68
m/p-Xylenes
Concen: 104.492 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion:106 Resp: 389131
Ion Ratio Lower Upper
106 100
91 208.9 167.1 250.7









#69
o-Xylene
Concen: 53.503 ug/l
RT: 12.400 min Scan# 1775
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion:106 Resp: 188510
Ion Ratio Lower Upper
106 100
91 221.9 110.2 330.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#70
Styrene
Concen: 55.268 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04


Tgt Ion:104 Resp: 329833
Ion Ratio Lower Upper
104 100
78 51.5 42.7 64.1
103 54.3 43.6 65.4





#71
Bromoform
Concen: 54.365 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:173 Resp: 7848
Ion Ratio Lower Upper
173 100
175 48.9 23.5 70.6
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Tgt Ion:152 Resp: 130386
Ion Ratio Lower Upper
152 100
115 62.2 31.3 93.9
150 175.6 0.0 349.8





#73
Isopropylbenzene
Concen: 48.507 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:105 Resp: 482620
Ion Ratio Lower Upper
105 100
120 25.8 12.9 38.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#74
N-ethyl acetate
Concen: 45.614 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04


Tgt Ion: 43 Resp: 205514
Ion Ratio Lower Upper
43 100
70 44.7 34.4 51.6
55 26.1 19.6 29.4
61 24.5 19.6 29.4





#75
1,1,2,2-Tetrachloroethane
Concen: 50.155 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 15005
Ion Ratio Lower Upper
83 100
131 9.8 4.9 14.7
85 63.2 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#76
1,2,3-Trichloropropane
Concen: 52.758 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 75 Resp: 140435
Ion Ratio Lower Upper
75 100
77 184.0 96.9 290.7





#77

Bromobenzene

Concen: 49.239 ug/l

RT: 12.976 min Scan# 1875

Delta R.T. -0.006 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion:156 Resp: 11817

Ion Ratio Lower Upper

156 100

77 218.7 115.1 345.3





158 94.5 48.9 146.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#78

n-propylbenzene

Concen: 49.748 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. -0.000 min

Lab File: VN084448.D




Acq: 21 Oct 2024 20:04


Tgt Ion: 91 Resp: 573481

Ion Ratio Lower Upper

91 100

120 22.1 10.8 32.4









#79
2-Chlorotoluene
Concen: 48.450 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 35639
Ion Ratio Lower Upper
91 100
126 31.9 15.7 47.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#80
1,3,5-Trimethylbenzene
Concen: 50.681 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04


Tgt Ion:105 Resp: 411332
Ion Ratio Lower Upper
105 100
120 47.5 23.7 71.1





#81
trans-1,4-Dichloro-2-butene
Concen: 49.901 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 55404
Ion Ratio Lower Upper
75 100
53 85.1 78.0 117.0
89 40.5 35.5 53.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#82
4-Chlorotoluene
Concen: 48.730 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04


Tgt Ion: 91 Resp: 360840
Ion Ratio Lower Upper
91 100
126 31.8 15.7 47.1





#83
tert-Butylbenzene
Concen: 48.923 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:119 Resp: 352909
Ion Ratio Lower Upper
119 100
91 68.0 34.1 102.3
134 24.0 12.8 38.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024




#84
1,2,4-Trimethylbenzene
Concen: 50.992 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Tgt Ion:105 Resp: 416913
Ion Ratio Lower Upper
105 100
120 44.3 22.3 66.8





#85
sec-Butylbenzene
Concen: 50.124 ug/l
RT: 13.617 min Scan# 1983
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:105 Resp: 48248
Ion Ratio Lower Upper
105 100
134 19.6 9.8 29.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024




#86
p-Isopropyltoluene
Concen: 50.755 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04



Tgt Ion:119 Resp: 404005
Ion Ratio Lower Upper
119 100
134 25.8 12.6 37.8
91 24.9 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 48.345 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion:146 Resp: 21766

Ion Ratio Lower Upper

146 100

111 42.8 21.8 65.3





148 63.9 31.9 95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#88

1,4-Dichlorobenzene

Concen: 48.056 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04




Tgt Ion:146 Resp: 218499


Ion Ratio Lower Upper

146 100

111 41.1 21.2 63.6

148 63.8 31.9 95.9









#89
n-Butylbenzene
Concen: 47.437 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 34238
Ion Ratio Lower Upper
91 100
92 52.6 26.6 79.7
134 24.3 12.3 36.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#90
Hexachloroethane
Concen: 46.151 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04


Tgt Ion:117 Resp: 74567
Ion Ratio Lower Upper
117 100
201 68.3 34.4 103.2





#91
1,2-Dichlorobenzene
Concen: 48.690 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:146 Resp: 214950
Ion Ratio Lower Upper
146 100
111 44.3 22.0 66.0
148 62.9 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 46.312 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 75 Resp: 28592
Ion Ratio Lower Upper
75 100
155 78.9 38.1 114.5
157 96.6 49.6 149.0





#93

1,2,4-Trichlorobenzene

Concen: 48.359 ug/l

RT: 15.394 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN084448.D


Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC



Tgt Ion:180 Resp: 105150

Ion Ratio Lower Upper

180 100

182 94.7 47.0 141.0




145 31.9 16.4 49.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#94

Hexachlorobutadiene


Concen: 43.364 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04





Tgt Ion:225 Resp: 47024


Ion Ratio Lower Upper

225 100

223 62.1 32.4 97.2


227 63.8 32.4 97.0





#95
Naphthalene
Concen: 46.938 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:128 Resp: 338368
Ion Ratio Lower Upper
128 100
127 12.6 10.5 15.7
129 10.5 8.7 13.1




Manual Integrations
APPROVED

Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#96
1,2,3-Trichlorobenzene
Concen: 47.602 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion:180 Resp: 103615
Ion Ratio Lower Upper
180 100
182 94.8 48.2 144.6
145 35.1 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084448.D
 Acq On : 21 Oct 2024 20:04
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 22 01:42:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.591	0.516	12.7	79	0.00
3 P	Chloromethane	0.675	0.575	14.8	80	0.00
4 C	Vinyl Chloride	0.654	0.580	11.3#	81	0.00
5 T	Bromomethane	0.431	0.354	17.9	77	-0.01
6 T	Chloroethane	0.468	0.379	19.0	85	0.00
7 T	Trichlorofluoromethane	1.019	0.994	2.5	88	0.00
8 T	Diethyl Ether	0.378	0.369	2.4	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.584	0.582	0.3	92	0.00
10 T	Methyl Iodide	0.752	0.719	4.4	84	0.00
11 T	Tert butyl alcohol	0.122	0.102	16.4	78	0.00
12 CM	1,1-Dichloroethene	0.563	0.556	1.2#	89	0.00
13 T	Acrolein	0.140	0.105	25.0#	69	0.00
14 T	Allyl chloride	0.978	0.856	12.5	81	0.00
15 T	Acrylonitrile	0.309	0.322	-4.2	95	0.00
16 T	Acetone	0.318	0.255	19.8	70	0.00
17 T	Carbon Disulfide	1.782	1.354	24.0	74	0.00
18 T	Methyl Acetate	0.694	0.826	-19.0	109	0.00
19 T	Methyl tert-butyl Ether	1.900	1.972	-3.8	91	0.00
20 T	Methylene Chloride	0.647	0.658	-1.7	94	0.00
21 T	trans-1,2-Dichloroethene	0.590	0.582	1.4	88	0.00
22 T	Diisopropyl ether	2.022	2.098	-3.8	91	0.00
23 T	Vinyl Acetate	1.500	1.432	4.5	81	0.00
24 P	1,1-Dichloroethane	1.131	1.157	-2.3	91	0.00
25 T	2-Butanone	0.434	0.406	6.5	84	0.00
26 T	2,2-Dichloropropane	1.020	0.912	10.6	79	0.00
27 T	cis-1,2-Dichloroethene	0.713	0.722	-1.3	91	0.00
28 T	Bromochloromethane	0.505	0.479	5.1	91	0.00
29 T	Tetrahydrofuran	0.268	0.269	-0.4	89	0.00
30 C	Chloroform	1.175	1.226	-4.3#	95	0.00
31 T	Cyclohexane	1.098	0.934	14.9	82	0.00
32 T	1,1,1-Trichloroethane	1.055	1.082	-2.6	92	0.00
33 S	1,2-Dichloroethane-d4	0.741	0.710	4.2	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.330	0.343	-3.9	98	0.00
36 T	1,1-Dichloropropene	0.479	0.472	1.5	86	0.00
37 T	Ethyl Acetate	0.491	0.472	3.9	87	0.00
38 T	Carbon Tetrachloride	0.525	0.534	-1.7	90	0.00
39 T	Methylcyclohexane	0.533	0.493	7.5	78	0.00
40 TM	Benzene	1.492	1.520	-1.9	91	0.00
41 T	Methacrylonitrile	0.267	0.273	-2.2	88	0.00
42 TM	1,2-Dichloroethane	0.506	0.507	-0.2	89	0.00
43 T	Isopropyl Acetate	0.962	0.816	15.2	84	0.00
44 TM	Trichloroethene	0.348	0.353	-1.4	90	0.00
45 C	1,2-Dichloropropane	0.353	0.379	-7.4#	93	0.00
46 T	Dibromomethane	0.238	0.265	-11.3	95	0.00
47 T	Bromodichloromethane	0.529	0.566	-7.0	94	0.00
48 T	Methyl methacrylate	0.383	0.390	-1.8	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084448.D
 Acq On : 21 Oct 2024 20:04
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 22 01:42:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	89	0.00
50 S	Toluene-d8	1.213	1.259	-3.8	94	0.00
51 T	4-Methyl-2-Pentanone	0.468	0.503	-7.5	92	0.00
52 CM	Toluene	0.910	0.953	-4.7#	91	0.00
53 T	t-1,3-Dichloropropene	0.539	0.551	-2.2	87	0.00
54 T	cis-1,3-Dichloropropene	0.580	0.604	-4.1	88	0.00
55 T	1,1,2-Trichloroethane	0.326	0.361	-10.7	96	0.00
56 T	Ethyl methacrylate	0.537	0.587	-9.3	90	0.00
57 T	1,3-Dichloropropane	0.583	0.623	-6.9	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.249	0.277	-11.2	92	0.00
59 T	2-Hexanone	0.349	0.364	-4.3	87	0.00
60 T	Dibromochloromethane	0.385	0.434	-12.7	96	0.00
61 T	1,2-Dibromoethane	0.339	0.358	-5.6	93	0.00
62 S	4-Bromofluorobenzene	0.442	0.461	-4.3	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.348	0.341	2.0	92	0.00
65 PM	Chlorobenzene	1.109	1.111	-0.2	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.382	0.405	-6.0	97	0.00
67 C	Ethyl Benzene	1.961	2.021	-3.1#	92	0.00
68 T	m/p-Xylenes	0.725	0.758	-4.6	91	0.00
69 T	o-Xylene	0.686	0.734	-7.0	92	0.00
70 T	Styrene	1.162	1.285	-10.6	95	0.00
71 P	Bromoform	0.281	0.306	-8.9	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.815	3.702	3.0	91	0.00
74 T	N-amyl acetate	1.728	1.576	8.8	87	0.00
75 P	1,1,2,2-Tetrachloroethane	1.147	1.151	-0.3	99	0.00
76 T	1,2,3-Trichloropropane	1.021	1.077	-5.5	98	0.00
77 T	Bromobenzene	0.920	0.906	1.5	98	0.00
78 T	n-propylbenzene	4.421	4.398	0.5	92	0.00
79 T	2-Chlorotoluene	2.821	2.733	3.1	93	0.00
80 T	1,3,5-Trimethylbenzene	3.112	3.155	-1.4	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.426	0.425	0.2	96	0.00
82 T	4-Chlorotoluene	2.840	2.767	2.6	93	0.00
83 T	tert-Butylbenzene	2.766	2.707	2.1	91	0.00
84 T	1,2,4-Trimethylbenzene	3.135	3.198	-2.0	93	0.00
85 T	sec-Butylbenzene	3.691	3.700	-0.2	91	0.00
86 T	p-Isopropyltoluene	3.052	3.099	-1.5	90	0.00
87 T	1,3-Dichlorobenzene	1.727	1.669	3.4	95	0.00
88 T	1,4-Dichlorobenzene	1.744	1.676	3.9	96	0.00
89 T	n-Butylbenzene	2.768	2.626	5.1	87	0.00
90 T	Hexachloroethane	0.620	0.572	7.7	91	0.00
91 T	1,2-Dichlorobenzene	1.693	1.649	2.6	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.237	0.219	7.6	94	0.00
93 T	1,2,4-Trichlorobenzene	0.834	0.806	3.4	88	0.00
94 T	Hexachlorobutadiene	0.416	0.361	13.2	93	0.00
95 T	Naphthalene	2.764	2.595	6.1	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084448.D
Acq On : 21 Oct 2024 20:04
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 22 01:42:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.835	0.795	4.8	91	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084448.D
 Acq On : 21 Oct 2024 20:04
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 22 01:42:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	50.000	43.654	12.7	79	0.00
3 P	Chloromethane	50.000	42.550	14.9	80	0.00
4 C	Vinyl Chloride	50.000	44.316	11.4#	81	0.00
5 T	Bromomethane	50.000	41.035	17.9	77	-0.01
6 T	Chloroethane	50.000	40.462	19.1	85	0.00
7 T	Trichlorofluoromethane	50.000	48.786	2.4	88	0.00
8 T	Diethyl Ether	50.000	48.772	2.5	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.837	0.3	92	0.00
10 T	Methyl Iodide	50.000	47.834	4.3	84	0.00
11 T	Tert butyl alcohol	250.000	209.300	16.3	78	0.00
12 CM	1,1-Dichloroethene	50.000	49.352	1.3#	89	0.00
13 T	Acrolein	250.000	186.809	25.3#	69	0.00
14 T	Allyl chloride	50.000	43.772	12.5	81	0.00
15 T	Acrylonitrile	250.000	261.006	-4.4	95	0.00
16 T	Acetone	250.000	200.235	19.9	70	0.00
17 T	Carbon Disulfide	50.000	37.994	24.0	74	0.00
18 T	Methyl Acetate	50.000	59.505	-19.0	109	0.00
19 T	Methyl tert-butyl Ether	50.000	51.897	-3.8	91	0.00
20 T	Methylene Chloride	50.000	50.844	-1.7	94	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.315	1.4	88	0.00
22 T	Diisopropyl ether	50.000	51.879	-3.8	91	0.00
23 T	Vinyl Acetate	250.000	238.618	4.6	81	0.00
24 P	1,1-Dichloroethane	50.000	51.145	-2.3	91	0.00
25 T	2-Butanone	250.000	233.725	6.5	84	0.00
26 T	2,2-Dichloropropane	50.000	44.706	10.6	79	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.655	-1.3	91	0.00
28 T	Bromochloromethane	50.000	47.474	5.1	91	0.00
29 T	Tetrahydrofuran	250.000	251.461	-0.6	89	0.00
30 C	Chloroform	50.000	52.167	-4.3#	95	0.00
31 T	Cyclohexane	50.000	42.521	15.0	82	0.00
32 T	1,1,1-Trichloroethane	50.000	51.300	-2.6	92	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.911	4.2	90	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	93	0.00
35 S	Dibromofluoromethane	50.000	52.045	-4.1	98	0.00
36 T	1,1-Dichloropropene	50.000	49.312	1.4	86	0.00
37 T	Ethyl Acetate	50.000	48.082	3.8	87	0.00
38 T	Carbon Tetrachloride	50.000	50.818	-1.6	90	0.00
39 T	Methylcyclohexane	50.000	46.254	7.5	78	0.00
40 TM	Benzene	50.000	50.941	-1.9	91	0.00
41 T	Methacrylonitrile	50.000	51.187	-2.4	88	0.00
42 TM	1,2-Dichloroethane	50.000	50.160	-0.3	89	0.00
43 T	Isopropyl Acetate	50.000	42.408	15.2	84	0.00
44 TM	Trichloroethene	50.000	50.592	-1.2	90	0.00
45 C	1,2-Dichloropropane	50.000	53.648	-7.3#	93	0.00
46 T	Dibromomethane	50.000	55.712	-11.4	95	0.00
47 T	Bromodichloromethane	50.000	53.528	-7.1	94	0.00
48 T	Methyl methacrylate	50.000	50.850	-1.7	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084448.D
 Acq On : 21 Oct 2024 20:04
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 22 01:42:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	994.968	0.5	89	0.00
50 S	Toluene-d8	50.000	51.913	-3.8	94	0.00
51 T	4-Methyl-2-Pentanone	250.000	268.874	-7.5	92	0.00
52 CM	Toluene	50.000	52.382	-4.8#	91	0.00
53 T	t-1,3-Dichloropropene	50.000	51.089	-2.2	87	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.057	-4.1	88	0.00
55 T	1,1,2-Trichloroethane	50.000	55.371	-10.7	96	0.00
56 T	Ethyl methacrylate	50.000	54.658	-9.3	90	0.00
57 T	1,3-Dichloropropane	50.000	53.424	-6.8	94	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	278.254	-11.3	92	0.00
59 T	2-Hexanone	250.000	260.410	-4.2	87	0.00
60 T	Dibromochloromethane	50.000	56.277	-12.6	96	0.00
61 T	1,2-Dibromoethane	50.000	52.872	-5.7	93	0.00
62 S	4-Bromofluorobenzene	50.000	52.160	-4.3	95	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	97	0.00
64 T	Tetrachloroethene	50.000	48.943	2.1	92	0.00
65 PM	Chlorobenzene	50.000	50.105	-0.2	92	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	53.075	-6.2	97	0.00
67 C	Ethyl Benzene	50.000	51.533	-3.1#	92	0.00
68 T	m/p-Xylenes	100.000	104.492	-4.5	91	0.00
69 T	o-Xylene	50.000	53.503	-7.0	92	0.00
70 T	Styrene	50.000	55.268	-10.5	95	0.00
71 P	Bromoform	50.000	54.365	-8.7	94	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	102	0.00
73 T	Isopropylbenzene	50.000	48.507	3.0	91	0.00
74 T	N-ethyl acetate	50.000	45.614	8.8	87	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.155	-0.3	99	0.00
76 T	1,2,3-Trichloropropane	50.000	52.758	-5.5	98	0.00
77 T	Bromobenzene	50.000	49.239	1.5	98	0.00
78 T	n-propylbenzene	50.000	49.748	0.5	92	0.00
79 T	2-Chlorotoluene	50.000	48.450	3.1	93	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.681	-1.4	92	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.901	0.2	96	0.00
82 T	4-Chlorotoluene	50.000	48.730	2.5	93	0.00
83 T	tert-Butylbenzene	50.000	48.923	2.2	91	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.992	-2.0	93	0.00
85 T	sec-Butylbenzene	50.000	50.124	-0.2	91	0.00
86 T	p-Isopropyltoluene	50.000	50.755	-1.5	90	0.00
87 T	1,3-Dichlorobenzene	50.000	48.345	3.3	95	0.00
88 T	1,4-Dichlorobenzene	50.000	48.056	3.9	96	0.00
89 T	n-Butylbenzene	50.000	47.437	5.1	87	0.00
90 T	Hexachloroethane	50.000	46.151	7.7	91	0.00
91 T	1,2-Dichlorobenzene	50.000	48.690	2.6	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	46.312	7.4	94	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.359	3.3	88	0.00
94 T	Hexachlorobutadiene	50.000	43.364	13.3	93	0.00
95 T	Naphthalene	50.000	46.938	6.1	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084448.D
Acq On : 21 Oct 2024 20:04
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 22 01:42:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	47.602	4.8	91	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6





QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084211.D
Acq On : 30 Sep 2024 09:24
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Title : SW846 8260
Last Update : Tue Oct 01 07:11:01 2024



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1843



Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.9	9177	PASS
75	95	30	60	53.9	24795	PASS
95	95	100	100	100.0	46040	PASS
96	95	5	9	6.3	2879	PASS
173	174	0.00	2	0.8	259	PASS
174	95	50	100	71.0	32685	PASS
175	174	5	9	7.8	2541	PASS
176	174	95	101	98.2	32104	PASS
177	176	5	9	7.0	2260	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084427.D
Acq On : 21 Oct 2024 11:14
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Title : SW846 8260
Last Update : Tue Oct 01 07:11:01 2024



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1843

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.0	6183	PASS
75	95	30	60	51.2	18619	PASS
95	95	100	100	100.0	36389	PASS
96	95	5	9	7.6	2782	PASS
173	174	0.00	2	0.9	244	PASS
174	95	50	100	72.7	26461	PASS
175	174	5	9	7.0	1860	PASS
176	174	95	101	96.3	25488	PASS
177	176	5	9	6.6	1686	PASS

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:			
Project:	CTO WE13			Date Received:			
Client Sample ID:	VN1021WBL01			SDG No.:	P4440		
Lab Sample ID:	VN1021WBL01			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	RXI-624	ID :	0.25	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084430.D	1		10/21/24 12:36	VN102124

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	
Project:	CTO WE13	Date Received:	
Client Sample ID:	VN1021WBL01	SDG No.:	P4440
Lab Sample ID:	VN1021WBL01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084430.D	1		10/21/24 12:36	VN102124

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	48.6		81 - 118		97%	SPK: 50
1868-53-7	Dibromofluoromethane	51.1		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	48.8		89 - 112		98%	SPK: 50
460-00-4	4-Bromofluorobenzene	44.7		85 - 114		89%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	167000	8.224				
540-36-3	1,4-Difluorobenzene	295000	9.1				
3114-55-4	Chlorobenzene-d5	254000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	100000	13.788				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084430.D
 Acq On : 21 Oct 2024 12:36
 Operator : JC\MD
 Sample : VN1021WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBL01

Quant Time: Oct 22 01:31:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	167468	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	295479	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	254088	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	99954	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	120667	48.597	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.200%
35) Dibromofluoromethane	8.165	113	99495	51.076	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.160%
50) Toluene-d8	10.565	98	349958	48.830	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.660%
62) 4-Bromofluorobenzene	12.847	95	116683	44.690	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	89.380%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084430.D
Acq On : 21 Oct 2024 12:36
Operator : JC\MD
Sample : VN1021WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBL01

Quant Time: Oct 22 01:31:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084430.D
Acq: 21 Oct 2024 12:36




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBL01


Tgt Ion:168 Resp: 167468
Ion Ratio Lower Upper
168 100
99 65.6 54.2 81.2



#33
1,2-Dichloroethane-d4
Concen: 48.597 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.000 min
Lab File: VN084430.D
Acq: 21 Oct 2024 12:36

Tgt Ion: 65 Resp: 120667
Ion Ratio Lower Upper
65 100
67 52.1 0.0 102.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN084430.D

Acq: 21 Oct 2024 12:36

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBL01





Tgt Ion:114 Resp: 295479

Ion Ratio Lower Upper

114 100

63 21.1 0.0 43.8

88 15.8 0.0 31.6



#35

Dibromofluoromethane

Concen: 51.076 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN084430.D

Acq: 21 Oct 2024 12:36




Tgt Ion:113 Resp: 99495


Ion Ratio Lower Upper

113 100

111 104.6 83.3 124.9

192 18.5 13.5 20.3









#50
Toluene-d8
Concen: 48.830 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.000 min
Lab File: VN084430.D
Acq: 21 Oct 2024 12:36




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBL01


Tgt Ion: 98 Resp: 349958
Ion Ratio Lower Upper
98 100
100 65.2 52.7 79.1



#62
4-Bromofluorobenzene
Concen: 44.690 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN084430.D
Acq: 21 Oct 2024 12:36

Tgt Ion: 95 Resp: 116683
Ion Ratio Lower Upper
95 100
174 75.9 0.0 145.2
176 71.0 0.0 140.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN084430.D
Acq: 21 Oct 2024 12:36




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBL01

Tgt Ion	Ratio	Lower	Upper
117	100		
82	58.2	47.2	70.8
119	32.3	25.4	38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN084430.D
Acq: 21 Oct 2024 12:36

Tgt Ion	Ratio	Lower	Upper
152	100		
115	62.6	31.3	93.9
150	155.5	0.0	349.8



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084430.D
Acq On : 21 Oct 2024 12:36
Operator : JC\MD
Sample : VN1021WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBL01

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Sampling : 1

Start Thrs: 0.2

Stop Thrs : 0

Filtering: 5

Min Area: 3 % of largest Peak

Max Peaks: 100

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M

Title : SW846 8260

Signal : TIC: VN084430.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	7.436	922	932	942	rBV2	4108	10909	1.18%	0.228%
2	8.165	1047	1056	1061	rBV	137448	335323	36.14%	7.019%
3	8.224	1061	1066	1078	rVB	239504	517962	55.83%	10.843%
4	8.577	1118	1126	1137	rBV	143101	329717	35.54%	6.902%
5	9.100	1206	1215	1227	rBV	348480	696007	75.02%	14.570%
6	10.565	1456	1464	1477	rBV	513172	927809	100.00%	19.422%
7	11.865	1677	1685	1700	rVB	450419	783202	84.41%	16.395%
8	12.847	1845	1852	1864	rVB	330020	559221	60.27%	11.706%
9	13.788	2005	2012	2023	rBV	373902	616926	66.49%	12.914%




Sum of corrected areas: 4777076

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084430.D
Acq On : 21 Oct 2024 12:36
Operator : JC\MD
Sample : VN1021WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084430.D
Acq On : 21 Oct 2024 12:36
Operator : JC\MD
Sample : VN1021WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084430.D
Acq On : 21 Oct 2024 12:36
Operator : JC\MD
Sample : VN1021WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--
				#	RT Resp Conc

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	
Project:	CTO WE13			Date Received:	
Client Sample ID:	VN1021WBS01			SDG No.:	P4440
Lab Sample ID:	VN1021WBS01			Matrix:	Water
Analytical Method:	SW8260			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084433.D	1		10/21/24 13:48	VN102124

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	16.1		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	16.7		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	16.5		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	15.0		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	18.1		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.1		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	17.6		0.26	0.75	1.00	ug/L
67-64-1	Acetone	87.5		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	14.2		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	18.1		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	18.6		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	17.4		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.2		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	88.7		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.9		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	18.2		0.25	0.75	1.00	ug/L
67-66-3	Chloroform	19.0		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	18.7		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	16.2		0.19	0.50	1.00	ug/L
71-43-2	Benzene	19.0		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	18.9		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.2		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.6		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	19.5		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	96.8		0.75	2.50	5.00	ug/L
108-88-3	Toluene	19.6		0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	18.5		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	18.9		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.4		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	96.1		1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	
Project:	CTO WE13			Date Received:	
Client Sample ID:	VN1021WBS01			SDG No.:	P4440
Lab Sample ID:	VN1021WBS01			Matrix:	Water
Analytical Method:	SW8260			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084433.D	1		10/21/24 13:48	VN102124

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	20.1		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	18.9		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.2		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	18.5		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	38.5		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	20.0		0.14	0.50	1.00	ug/L
100-42-5	Styrene	19.5		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	19.5		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	18.2		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.4		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	18.6		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	18.8		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	18.3		0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.5		81 - 118		101%	SPK: 50
1868-53-7	Dibromofluoromethane	54.3		80 - 119		109%	SPK: 50
2037-26-5	Toluene-d8	53.9		89 - 112		108%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.5		85 - 114		105%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	190000	8.224				
540-36-3	1,4-Difluorobenzene	323000	9.1				
3114-55-4	Chlorobenzene-d5	284000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	138000	13.788				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084433.D
 Acq On : 21 Oct 2024 13:48
 Operator : JC\MD
 Sample : VN1021WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBS01

Manual Integrations
 APPROVED

Reviewed By :John Carlone 10/22/2024
 Supervised By :Mahesh Dadoda 10/22/2024

Quant Time: Oct 22 01:32:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	189913	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	322919	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	284161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	138325	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	142332	50.547	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 101.100%			
35) Dibromofluoromethane	8.165	113	115508	54.258	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 108.520%			
50) Toluene-d8	10.565	98	422532	53.947	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 107.900%			
62) 4-Bromofluorobenzene	12.847	95	149865	52.521	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 105.040%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.130	85	36261	16.142	ug/l	95
3) Chloromethane	2.359	50	41235	16.076	ug/l	96
4) Vinyl Chloride	2.512	62	41507	16.706	ug/l	100
5) Bromomethane	2.953	94	26963	16.453	ug/l	98
6) Chloroethane	3.124	64	26726	15.020	ug/l	99
7) Trichlorofluoromethane	3.500	101	70146	18.131	ug/l	98
8) Diethyl Ether	3.965	74	24127	16.814	ug/l	93
9) 1,1,2-Trichlorotrifluo...	4.377	101	40161	18.116	ug/l	98
10) Methyl Iodide	4.589	142	48908	17.131	ug/l	96
11) Tert butyl alcohol	5.518	59	36415	78.429	ug/l #	94
12) 1,1-Dichloroethene	4.342	96	37659	17.612	ug/l	97
13) Acrolein	4.183	56	26602	49.907	ug/l	98
14) Allyl chloride	5.024	41	58352	15.713	ug/l	98
15) Acrylonitrile	5.718	53	109279	93.202	ug/l	99
16) Acetone	4.430	43	105837	87.496	ug/l	96
17) Carbon Disulfide	4.712	76	96169	14.205	ug/l	100
18) Methyl Acetate	5.030	43	59251	22.474	ug/l	99
19) Methyl tert-butyl Ether	5.794	73	130984	18.149	ug/l	98
20) Methylene Chloride	5.271	84	45743	18.603	ug/l	91
21) trans-1,2-Dichloroethene	5.789	96	38888	17.359	ug/l	95
22) Diisopropyl ether	6.671	45	141732	18.451	ug/l	98
23) Vinyl Acetate	6.600	43	468614	82.252	ug/l	98
24) 1,1-Dichloroethane	6.571	63	82481	19.201	ug/l	97
25) 2-Butanone	7.483	43	146074	88.663	ug/l	98
26) 2,2-Dichloropropane	7.488	77	71005	18.333	ug/l	98
27) cis-1,2-Dichloroethene	7.494	96	49184	18.169	ug/l	98
28) Bromochloromethane	7.812	49	37550	19.592	ug/l	97
29) Tetrahydrofuran	7.841	42	90421	88.935	ug/l	98
30) Chloroform	7.971	83	84755	18.995	ug/l	100
31) Cyclohexane	8.259	56	66762	16.007	ug/l #	97
32) 1,1,1-Trichloroethane	8.171	97	74940	18.700	ug/l	97
36) 1,1-Dichloropropene	8.371	75	56161	18.162	ug/l	97
37) Ethyl Acetate	7.559	43	57455	18.111	ug/l	99
38) Carbon Tetrachloride	8.365	117	64152	18.912	ug/l	94
39) Methylcyclohexane	9.600	83	55643	16.168	ug/l	96
40) Benzene	8.606	78	183073	19.002	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084433.D
 Acq On : 21 Oct 2024 13:48
 Operator : JC\MD
 Sample : VN1021WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBS01

Manual Integrations
 APPROVED

Reviewed By :John Carlone 10/22/2024
 Supervised By :Mahesh Dadoda 10/22/2024

Quant Time: Oct 22 01:32:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	29860	17.338	ug/l	93
42) 1,2-Dichloroethane	8.671	62	61753	18.907	ug/l	99
43) Isopropyl Acetate	8.688	43	94050	15.136	ug/l	98
44) Trichloroethene	9.347	130	43127	19.166	ug/l	98
45) 1,2-Dichloropropane	9.624	63	44771	19.644	ug/l	98
46) Dibromomethane	9.706	93	31378	20.401	ug/l	96
47) Bromodichloromethane	9.888	83	66599	19.490	ug/l	96
48) Methyl methacrylate	9.682	41	43352	17.526	ug/l	98
49) 1,4-Dioxane	9.694	88	16997	373.245	ug/l	97
51) 4-Methyl-2-Pentanone	10.441	43	292338	96.801	ug/l	100
52) Toluene	10.629	92	114864	19.551	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	64356	18.476	ug/l	98
54) cis-1,3-Dichloropropene	10.312	75	70777	18.891	ug/l	96
55) 1,1,2-Trichloroethane	11.012	97	43053	20.441	ug/l	97
56) Ethyl methacrylate	10.871	69	65255	18.829	ug/l	96
57) 1,3-Dichloropropane	11.165	76	74113	19.693	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	139755	86.930	ug/l	98
59) 2-Hexanone	11.194	43	216944	96.112	ug/l	99
60) Dibromochloromethane	11.359	129	50030	20.104	ug/l	99
61) 1,2-Dibromoethane	11.465	107	41569	18.994	ug/l	98
64) Tetrachloroethene	11.100	164	37454	18.924	ug/l	95
65) Chlorobenzene	11.888	112	121076	19.217	ug/l	94
66) 1,1,1,2-Tetrachloroethane	11.959	131	43109	19.865	ug/l	99
67) Ethyl Benzene	11.965	91	205793	18.470	ug/l	100
68) m/p-Xylenes	12.070	106	158675	38.498	ug/l	100
69) o-Xylene	12.400	106	78180	20.049	ug/l	96
70) Styrene	12.412	104	129069	19.541	ug/l	99
71) Bromoform	12.576	173	31202	19.529	ug/l #	95
73) Isopropylbenzene	12.694	105	191996	18.189	ug/l	98
74) N-amyl acetate	12.494	43	79653	16.664	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.935	83	61726	19.448	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	58038m	20.552	ug/l	
77) Bromobenzene	12.982	156	48284	18.964	ug/l	95
78) n-propylbenzene	13.035	91	231176	18.903	ug/l	99
79) 2-Chlorotoluene	13.123	91	145730	18.674	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	168330	19.550	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	19953	16.940	ug/l	98
82) 4-Chlorotoluene	13.217	91	149766	19.064	ug/l	100
83) tert-Butylbenzene	13.435	119	141261	18.459	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	169548	19.547	ug/l	99
85) sec-Butylbenzene	13.617	105	195385	19.133	ug/l	99
86) p-Isopropyltoluene	13.729	119	161505	19.125	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	88692	18.569	ug/l	100
88) 1,4-Dichlorobenzene	13.812	146	90608	18.784	ug/l	99
89) n-Butylbenzene	14.053	91	132089	17.250	ug/l	99
90) Hexachloroethane	14.329	117	31571	18.419	ug/l	97
91) 1,2-Dichlorobenzene	14.106	146	85799	18.319	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	11177	17.065	ug/l	98
93) 1,2,4-Trichlorobenzene	15.394	180	40526	17.567	ug/l	100
94) Hexachlorobutadiene	15.500	225	19271	16.751	ug/l	98
95) Naphthalene	15.641	128	119595	15.638	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	39826	17.247	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084433.D
Acq On : 21 Oct 2024 13:48
Operator : JC\MD
Sample : VN1021WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024

Quant Time: Oct 22 01:32:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084433.D
Acq On : 21 Oct 2024 13:48
Operator : JC\MD
Sample : VN1021WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01

Manual Integrations

Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024

Quant Time: Oct 22 01:32:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01

Tgt Ion:168 Resp: 18991
Ion Ratio Lower Upper
168 100
99 61.2 54.2 81.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#2
Dichlorodifluoromethane
Concen: 16.142 ug/l
RT: 2.130 min Scan# 30
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 85 Resp: 36261
Ion Ratio Lower Upper
85 100
87 32.5 17.6 52.9









#3
Chloromethane
Concen: 16.076 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01

Tgt Ion: 50 Resp: 41235
Ion Ratio Lower Upper
50 100
52 31.1 26.8 40.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#4
Vinyl Chloride
Concen: 16.706 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48


Tgt Ion: 62 Resp: 41507
Ion Ratio Lower Upper
62 100
64 32.7 26.2 39.4





#5
Bromomethane
Concen: 16.453 ug/l
RT: 2.953 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01



Tgt Ion: 94 Resp: 2696
Ion Ratio Lower Upper
94 100
96 89.6 72.9 109.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#6
Chloroethane
Concen: 15.020 ug/l
RT: 3.124 min Scan# 199
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48


Tgt Ion: 64 Resp: 26726
Ion Ratio Lower Upper
64 100
66 30.2 23.5 35.3





#7
Trichlorofluoromethane
Concen: 18.131 ug/l
RT: 3.500 min Scan# 262
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01



Tgt Ion: 101 Resp: 70140
Ion Ratio Lower Upper
101 100
103 66.6 51.8 77.6




Manual Integrations
APPROVED


Reviewed By : John Carlone 10/22/2024
Supervised By : Mahesh Dadoda 10/22/2024



#8
Diethyl Ether
Concen: 16.814 ug/l
RT: 3.965 min Scan# 342
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 74 Resp: 24127
Ion Ratio Lower Upper
74 100
45 102.5 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.116 ug/l
RT: 4.377 min Scan# 411
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01

Tgt Ion:101 Resp: 4016
Ion Ratio Lower Upper
101 100
85 46.2 36.2 54.4
151 76.3 59.3 88.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#10
Methyl Iodide
Concen: 17.131 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion:142 Resp: 48908
Ion Ratio Lower Upper
142 100
127 47.6 35.3 52.9
141 13.5 11.0 16.6





#11

Tert butyl alcohol

Concen: 78.429 ug/l

RT: 5.518 min Scan# 607

Delta R.T. -0.006 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion: 59 Resp: 3641

Ion Ratio Lower Upper

59 100





57 8.1 8.2 12.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#12

1,1-Dichloroethene

Concen: 17.612 ug/l

RT: 4.342 min Scan# 406

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48




Tgt Ion: 96 Resp: 37659


Ion Ratio Lower Upper

96 100

61 161.0 128.6 192.8

98 57.8 51.7 77.5





#13

Acrolein

Concen: 49.907 ug/l

RT: 4.183 min Scan# 31

Delta R.T. 0.006 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion: 56 Resp: 26602

Ion Ratio Lower Upper

56 100





55 72.9 57.2 85.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#14

Allyl chloride

Concen: 15.713 ug/l

RT: 5.024 min Scan# 522

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48




Tgt Ion: 41 Resp: 58352


Ion Ratio Lower Upper

41 100

39 70.9 59.0 88.6


76 37.4 29.8 44.6





#15
Acrylonitrile
Concen: 93.202 ug/l
RT: 5.718 min Scan# 640
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01



Tgt Ion: 53 Resp: 109279
Ion Ratio Lower Upper
53 100
52 81.7 66.2 99.4
51 35.3 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#16
Acetone
Concen: 87.496 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48


Tgt Ion: 43 Resp: 105837
Ion Ratio Lower Upper
43 100
58 32.9 24.4 36.6





#17
Carbon Disulfide
Concen: 14.205 ug/l
RT: 4.712 min Scan# 469
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01



Tgt Ion: 76 Resp: 96169
Ion Ratio Lower Upper
76 100
78 9.0 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#18
Methyl Acetate
Concen: 22.474 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.012 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 43 Resp: 59251
Ion Ratio Lower Upper
43 100
74 23.6 19.2 28.8





#19

Methyl tert-butyl Ether

Concen: 18.149 ug/l

RT: 5.794 min Scan# 61

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion: 73 Resp: 130984

Ion Ratio Lower Upper

73 100





57 23.1 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#20

Methylene Chloride

Concen: 18.603 ug/l

RT: 5.271 min Scan# 564

Delta R.T. -0.006 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Tgt Ion: 84 Resp: 45743




Ion Ratio Lower Upper


84 100

49 116.0 107.2 160.8

51 39.6 31.8 47.8

86 63.5 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 17.359 ug/l
RT: 5.789 min Scan# 652
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01

Tgt Ion: 96 Resp: 3888
Ion Ratio Lower Upper
96 100
61 130.3 108.4 162.6
98 67.0 49.8 74.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#22
Diisopropyl ether
Concen: 18.451 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 45 Resp: 141732
Ion Ratio Lower Upper
45 100
43 51.9 42.3 63.5
87 28.3 21.8 32.6
59 12.3 9.6 14.4









#23
Vinyl Acetate
Concen: 82.252 ug/l
RT: 6.600 min Scan# 790
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01

Tgt Ion: 43 Resp: 468614
Ion Ratio Lower Upper
43 100
86 11.3 8.6 12.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#24
1,1-Dichloroethane
Concen: 19.201 ug/l
RT: 6.571 min Scan# 785
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 63 Resp: 82481
Ion Ratio Lower Upper
63 100
98 8.3 3.4 10.2
100 4.4 2.0 5.8









#25
2-Butanone
Concen: 88.663 ug/l
RT: 7.483 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01

Tgt Ion: 43 Resp: 146074
Ion Ratio Lower Upper
43 100
72 27.7 21.4 32.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#26
2,2-Dichloropropane
Concen: 18.333 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48


Tgt Ion: 77 Resp: 71005
Ion Ratio Lower Upper
77 100
97 22.6 10.8 32.3





#27
cis-1,2-Dichloroethene
Concen: 18.169 ug/l
RT: 7.494 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01



Tgt Ion: 96 Resp: 49184
Ion Ratio Lower Upper
96 100
61 147.2 0.0 301.8
98 63.3 0.0 128.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#28
Bromochloromethane
Concen: 19.592 ug/l
RT: 7.812 min Scan# 996
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 49 Resp: 37550
Ion Ratio Lower Upper
49 100
129 0.9 0.0 3.6
130 74.6 58.0 87.0





#29

Tetrahydrofuran

Concen: 88.935 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01





Tgt Ion:	42	Resp:	9042
Ion Ratio	Lower	Upper	
42	100		
72	44.5	34.6	51.8
71	41.0	32.2	48.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#30

Chloroform

Concen: 18.995 ug/l




RT: 7.971 min Scan# 1023


Delta R.T. 0.006 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Tgt Ion:	83	Resp:	84755
Ion Ratio	Lower	Upper	
83	100		
85	64.0	51.0	76.4





#31

Cyclohexane

Concen: 16.007 ug/l

RT: 8.259 min Scan# 1072

Delta R.T. 0.006 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion: 56 Resp: 66762

Ion Ratio Lower Upper

56 100

69 35.7 23.8 35.6





84 86.0 68.8 103.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#32

1,1,1-Trichloroethane

Concen: 18.700 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48




Tgt Ion: 97 Resp: 74940


Ion Ratio Lower Upper

97 100

99 62.8 51.2 76.8

61 43.8 37.5 56.3





#33

1,2-Dichloroethane-d4

Concen: 50.547 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion: 65 Resp: 14233

Ion Ratio Lower Upper

65 100





67 52.7 0.0 102.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48




Tgt Ion:114 Resp: 322919


Ion Ratio Lower Upper

114 100

63 20.0 0.0 43.8

88 16.5 0.0 31.6





#35

Dibromofluoromethane

Concen: 54.258 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion:113 Resp: 115508

Ion Ratio Lower Upper

113 100

111 104.0 83.3 124.9





192 18.4 13.5 20.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#36

1,1-Dichloropropene

Concen: 18.162 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48




Tgt Ion: 75 Resp: 56161


Ion Ratio Lower Upper

75 100

110 35.3 16.9 50.7

77 29.3 24.7 37.1





#37

Ethyl Acetate

Concen: 18.111 ug/l

RT: 7.559 min Scan# 91

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion: 43 Resp: 5745

Ion Ratio Lower Upper

43 100

61 12.4 10.7 16.1





70 11.1 8.8 13.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#38

Carbon Tetrachloride

Concen: 18.912 ug/l

RT: 8.365 min Scan# 1090

Delta R.T. 0.006 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48




Tgt Ion:117 Resp: 64152


Ion Ratio Lower Upper

117 100

119 92.6 78.7 118.1

121 27.9 24.4 36.6





#39

Methylcyclohexane

Concen: 16.168 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion: 83 Resp: 5564

Ion Ratio Lower Upper

83 100

55 83.6 62.8 94.2





98 49.3 38.8 58.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#40

Benzene

Concen: 19.002 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN084433.D




Acq: 21 Oct 2024 13:48


Tgt Ion: 78 Resp: 183073

Ion Ratio Lower Upper

78 100


77 23.6 19.1 28.7





#41
Methacrylonitrile
Concen: 17.338 ug/l
RT: 7.777 min Scan# 990
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01



Tgt Ion: 41 Resp: 29860
Ion Ratio Lower Upper
41 100
39 59.2 43.3 64.9
67 75.6 56.1 84.1
52 34.3 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#42
1,2-Dichloroethane
Concen: 18.907 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 62 Resp: 61753
Ion Ratio Lower Upper
62 100
98 9.6 0.0 18.2





#43

Isopropyl Acetate

Concen: 15.136 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01





Tgt Ion:	43	Resp:	94050
Ion	Ratio	Lower	Upper
43	100		
61	27.2	20.4	30.6
87	13.1	10.3	15.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#44

Trichloroethene

Concen: 19.166 ug/l




RT: 9.347 min Scan# 1257


Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Tgt Ion:	130	Resp:	43127
Ion	Ratio	Lower	Upper
130	100		
95	99.7	0.0	203.0





#45

1,2-Dichloropropane

Concen: 19.644 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. 0.006 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion: 63 Resp: 4477

Ion Ratio Lower Upper

63 100





65 32.0 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#46

Dibromomethane

Concen: 20.401 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48




Tgt Ion: 93 Resp: 31378


Ion Ratio Lower Upper

93 100

95 83.6 65.0 97.4

174 94.1 71.8 107.8





#47

Bromodichloromethane

Concen: 19.490 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01





Tgt Ion:	83	Resp:	6659
Ion	Ratio	Lower	Upper
83	100		
85	61.4	51.5	77.3
127	7.9	5.7	8.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#48

Methyl methacrylate

Concen: 17.526 ug/l




RT: 9.682 min Scan# 1314


Delta R.T. 0.006 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Tgt Ion:	41	Resp:	43352
Ion	Ratio	Lower	Upper
41	100		
69	86.0	67.1	100.7
39	56.5	45.8	68.6





#49

1,4-Dioxane

Concen: 373.245 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion: 88 Resp: 1699

Ion Ratio Lower Upper

88 100

43 28.7 25.3 37.9





58 72.3 56.4 84.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#50

Toluene-d8

Concen: 53.947 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN084433.D




Acq: 21 Oct 2024 13:48


Tgt Ion: 98 Resp: 422532

Ion Ratio Lower Upper

98 100

100 65.3 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 96.801 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01

Tgt Ion: 43 Resp: 292338
Ion Ratio Lower Upper
43 100
58 39.7 31.6 47.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#52
Toluene
Concen: 19.551 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48


Tgt Ion: 92 Resp: 114864
Ion Ratio Lower Upper
92 100
91 170.2 137.3 205.9





#53
t-1,3-Dichloropropene
Concen: 18.476 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01



Tgt Ion: 75 Resp: 64350
Ion Ratio Lower Upper
75 100
77 30.8 25.4 38.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#54
cis-1,3-Dichloropropene
Concen: 18.891 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48


Tgt Ion: 75 Resp: 70777
Ion Ratio Lower Upper
75 100
77 33.2 24.2 36.2
39 46.1 38.7 58.1





#55
1,1,2-Trichloroethane
Concen: 20.441 ug/l
RT: 11.012 min Scan# 1540
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01



Tgt Ion: 97 Resp: 4305
Ion Ratio Lower Upper
97 100
83 90.4 71.0 106.4
85 55.6 41.7 62.5
99 60.6 50.1 75.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#56
Ethyl methacrylate
Concen: 18.829 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 69 Resp: 65255
Ion Ratio Lower Upper
69 100
41 63.6 53.3 79.9
39 33.9 28.8 43.2





#57

1,3-Dichloropropane

Concen: 19.693 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.006 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion: 76 Resp: 7411

Ion Ratio Lower Upper

76 100





78 32.5 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#58

2-Chloroethyl Vinyl ether

Concen: 86.930 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN084433.D




Acq: 21 Oct 2024 13:48


Tgt Ion: 63 Resp: 139755

Ion Ratio Lower Upper

63 100


106 28.3 21.8 32.6





#59
2-Hexanone
Concen: 96.112 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01



Tgt Ion: 43 Resp: 216944
Ion Ratio Lower Upper
43 100
58 55.3 27.3 81.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#60
Dibromochloromethane
Concen: 20.104 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48


Tgt Ion:129 Resp: 50030
Ion Ratio Lower Upper
129 100
127 77.4 38.3 114.9





#61
1,2-Dibromoethane
Concen: 18.994 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01



Tgt Ion:107 Resp: 41569
Ion Ratio Lower Upper
107 100
109 96.3 75.4 113.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#62
4-Bromofluorobenzene
Concen: 52.521 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48


Tgt Ion: 95 Resp: 149865
Ion Ratio Lower Upper
95 100
174 74.8 0.0 145.2
176 71.6 0.0 140.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01




Tgt Ion:117 Resp: 28416
Ion Ratio Lower Upper
117 100
82 58.2 47.2 70.8
119 31.9 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024




#64
Tetrachloroethene
Concen: 18.924 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Tgt Ion:164 Resp: 37454
Ion Ratio Lower Upper
164 100
166 129.9 101.0 151.4
129 111.5 80.8 121.2
131 94.8 78.1 117.1





#65
Chlorobenzene
Concen: 19.217 ug/l
RT: 11.888 min Scan# 1689
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01



Tgt Ion:112 Resp: 121070
Ion Ratio Lower Upper
112 100
114 30.2 27.0 40.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 19.865 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion:131 Resp: 43109
Ion Ratio Lower Upper
131 100
133 94.3 47.9 143.7
119 64.6 32.8 98.3









#67
Ethyl Benzene
Concen: 18.470 ug/l
RT: 11.965 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01

Tgt Ion: 91 Resp: 205793
Ion Ratio Lower Upper
91 100
106 30.8 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#68
m/p-Xylenes
Concen: 38.498 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion:106 Resp: 158675
Ion Ratio Lower Upper
106 100
91 208.7 167.1 250.7





#69

o-Xylene

Concen: 20.049 ug/l

RT: 12.400 min Scan# 1775

Delta R.T. 0.006 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion:106 Resp: 78180

Ion Ratio Lower Upper

106 100





91 213.5 110.2 330.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#70

Styrene

Concen: 19.541 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48




Tgt Ion:104 Resp: 129069


Ion Ratio Lower Upper

104 100

78 51.8 42.7 64.1


103 55.0 43.6 65.4





#71
Bromoform
Concen: 19.529 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01




Tgt Ion:173 Resp: 3120
Ion Ratio Lower Upper
173 100
175 50.2 23.5 70.6
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Tgt Ion:152 Resp: 138325
Ion Ratio Lower Upper
152 100
115 62.4 31.3 93.9
150 161.5 0.0 349.8





#73
Isopropylbenzene
Concen: 18.189 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01



Tgt Ion:105 Resp: 191990
Ion Ratio Lower Upper
105 100
120 26.6 12.9 38.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#74
N-ethyl acetate
Concen: 16.664 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 43 Resp: 79653
Ion Ratio Lower Upper
43 100
70 44.0 34.4 51.6
55 28.0 19.6 29.4
61 24.3 19.6 29.4





#75

1,1,2,2-Tetrachloroethane

Concen: 19.448 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion: 83 Resp: 61720

Ion Ratio Lower Upper

83 100

131 10.5 4.9 14.7





85 63.5 31.9 95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#76

1,2,3-Trichloropropane

Concen: 20.552 ug/l m

RT: 12.994 min Scan# 1877

Delta R.T. 0.006 min

Lab File: VN084433.D




Acq: 21 Oct 2024 13:48


Tgt Ion: 75 Resp: 58038

Ion Ratio Lower Upper

75 100


77 182.3 96.9 290.7





#77
Bromobenzene
Concen: 18.964 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01



Tgt Ion:156 Resp: 48284
Ion Ratio Lower Upper
156 100
77 219.2 115.1 345.3
158 96.4 48.9 146.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#78
n-propylbenzene
Concen: 18.903 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 91 Resp: 231176
Ion Ratio Lower Upper
91 100
120 22.3 10.8 32.4





#79

2-Chlorotoluene

Concen: 18.674 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion: 91 Resp: 145730

Ion Ratio Lower Upper

91 100





126 31.7 15.7 47.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#80

1,3,5-Trimethylbenzene

Concen: 19.550 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN084433.D




Acq: 21 Oct 2024 13:48


Tgt Ion:105 Resp: 168330

Ion Ratio Lower Upper

105 100


120 47.3 23.7 71.1





#81
trans-1,4-Dichloro-2-butene
Concen: 16.940 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01



Tgt Ion: 75 Resp: 1995
Ion Ratio Lower Upper
75 100
53 96.1 78.0 117.0
89 46.4 35.5 53.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#82
4-Chlorotoluene
Concen: 19.064 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48


Tgt Ion: 91 Resp: 149766
Ion Ratio Lower Upper
91 100
126 31.4 15.7 47.1





#83
tert-Butylbenzene
Concen: 18.459 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01



Tgt Ion:119 Resp: 14126
Ion Ratio Lower Upper
119 100
91 67.1 34.1 102.3
134 25.8 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#84
1,2,4-Trimethylbenzene
Concen: 19.547 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48


Tgt Ion:105 Resp: 169548
Ion Ratio Lower Upper
105 100
120 43.7 22.3 66.8





#85
 sec-Butylbenzene
 Concen: 19.133 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN084433.D
 Acq: 21 Oct 2024 13:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBS01




Tgt Ion:105 Resp: 19538
 Ion Ratio Lower Upper
 105 100
 134 19.2 9.8 29.4

Manual Integrations
 APPROVED



Reviewed By :John Carlone 10/22/2024
 Supervised By :Mahesh Dadoda 10/22/2024




#86
 p-Isopropyltoluene
 Concen: 19.125 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN084433.D
 Acq: 21 Oct 2024 13:48



Tgt Ion:119 Resp: 161505
 Ion Ratio Lower Upper
 119 100
 134 24.2 12.6 37.8
 91 25.3 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 18.569 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion:146 Resp: 88692

Ion Ratio Lower Upper

146 100

111 43.5 21.8 65.3





148 64.2 31.9 95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#88

1,4-Dichlorobenzene

Concen: 18.784 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48




Tgt Ion:146 Resp: 90608


Ion Ratio Lower Upper

146 100

111 41.4 21.2 63.6

148 63.1 31.9 95.9





#89

n-Butylbenzene

Concen: 17.250 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion: 91 Resp: 132089

Ion Ratio Lower Upper

91 100

92 51.9 26.6 79.7





134 24.5 12.3 36.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#90

Hexachloroethane

Concen: 18.419 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. -0.000 min

Lab File: VN084433.D




Acq: 21 Oct 2024 13:48


Tgt Ion:117 Resp: 31571

Ion Ratio Lower Upper

117 100

201 66.1 34.4 103.2





#91

1,2-Dichlorobenzene

Concen: 18.319 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion:146 Resp: 85799

Ion Ratio Lower Upper

146 100

111 44.9 22.0 66.0





148 64.6 32.1 96.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#92

1,2-Dibromo-3-Chloropropane

Concen: 17.065 ug/l

RT: 14.717 min Scan# 2170

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48




Tgt Ion: 75 Resp: 11177


Ion Ratio Lower Upper

75 100

155 77.7 38.1 114.5

157 96.7 49.6 149.0





#93

1,2,4-Trichlorobenzene

Concen: 17.567 ug/l

RT: 15.394 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion:180 Resp: 40520

Ion Ratio Lower Upper

180 100

182 94.0 47.0 141.0





145 32.8 16.4 49.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#94

Hexachlorobutadiene

Concen: 16.751 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48




Tgt Ion:225 Resp: 19271


Ion Ratio Lower Upper

225 100

223 62.3 32.4 97.2


227 64.9 32.4 97.0





#95
Naphthalene
Concen: 15.638 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01



Tgt Ion:128 Resp: 11959
Ion Ratio Lower Upper
128 100
127 13.1 10.5 15.7
129 10.9 8.7 13.1




Manual Integrations
APPROVED

Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#96
1,2,3-Trichlorobenzene
Concen: 17.247 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion:180 Resp: 39826
Ion Ratio Lower Upper
180 100
182 94.7 48.2 144.6
145 32.6 17.3 51.9



Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	
Project:	CTO WE13			Date Received:	
Client Sample ID:	VN1021WBSD01			SDG No.:	P4440
Lab Sample ID:	VN1021WBSD01			Matrix:	Water
Analytical Method:	SW8260			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084434.D	1		10/21/24 14:22	VN102124

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	16.3		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	17.1		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	17.0		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	16.0		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.0		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.3		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	17.7		0.26	0.75	1.00	ug/L
67-64-1	Acetone	91.6		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	14.5		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.8		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	20.2		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.4		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.6		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	96.8		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.4		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.2		0.25	0.75	1.00	ug/L
67-66-3	Chloroform	20.1		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.8		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	17.0		0.19	0.50	1.00	ug/L
71-43-2	Benzene	19.6		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.6		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.5		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.9		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	20.9		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	100		0.75	2.50	5.00	ug/L
108-88-3	Toluene	20.3		0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	19.7		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.0		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	22.5		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	100		1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	
Project:	CTO WE13			Date Received:	
Client Sample ID:	VN1021WBSD01			SDG No.:	P4440
Lab Sample ID:	VN1021WBSD01			Matrix:	Water
Analytical Method:	SW8260			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN084434.D	1		10/21/24 14:22	VN102124

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	21.8		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	19.3		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.5		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	18.4		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	38.8		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	18.8		0.14	0.50	1.00	ug/L
100-42-5	Styrene	19.8		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	20.7		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	18.4		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	20.6		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.3		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.4		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	18.8		0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.1		81 - 118		102%	SPK: 50
1868-53-7	Dibromofluoromethane	52.8		80 - 119		106%	SPK: 50
2037-26-5	Toluene-d8	51.8		89 - 112		104%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.4		85 - 114		105%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	163000	8.224				
540-36-3	1,4-Difluorobenzene	279000	9.1				
3114-55-4	Chlorobenzene-d5	256000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	124000	13.788				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084434.D
 Acq On : 21 Oct 2024 14:22
 Operator : JC\MD
 Sample : VN1021WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBSD01

Manual Integrations
 APPROVED

Reviewed By :John Carlone 10/22/2024
 Supervised By :Mahesh Dadoda 10/22/2024

Quant Time: Oct 22 01:33:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	163086	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	278516	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	255670	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	124118	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	123457	51.056	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 102.120%			
35) Dibromofluoromethane	8.165	113	97003	52.829	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.660%			
50) Toluene-d8	10.565	98	350013	51.812	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.620%			
62) 4-Bromofluorobenzene	12.847	95	129081	52.449	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 104.900%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	32472	16.833	ug/l	93
3) Chloromethane	2.359	50	35804	16.255	ug/l	98
4) Vinyl Chloride	2.512	62	36441	17.080	ug/l	99
5) Bromomethane	2.959	94	23913	16.992	ug/l	99
6) Chloroethane	3.118	64	24475	16.017	ug/l	96
7) Trichlorofluoromethane	3.501	101	63283	19.047	ug/l	90
8) Diethyl Ether	3.959	74	23522	19.089	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.371	101	36689	19.273	ug/l	98
10) Methyl Iodide	4.589	142	44167	18.016	ug/l	99
11) Tert butyl alcohol	5.518	59	35053	87.914	ug/l	97
12) 1,1-Dichloroethene	4.342	96	32588	17.747	ug/l	96
13) Acrolein	4.189	56	24397	53.300	ug/l	98
14) Allyl chloride	5.024	41	53083	16.646	ug/l	99
15) Acrylonitrile	5.718	53	102471	101.771	ug/l	97
16) Acetone	4.430	43	95163	91.613	ug/l	94
17) Carbon Disulfide	4.712	76	84276	14.496	ug/l	96
18) Methyl Acetate	5.030	43	54347	24.005	ug/l	99
19) Methyl tert-butyl Ether	5.789	73	122872	19.826	ug/l	98
20) Methylene Chloride	5.271	84	42636	20.192	ug/l	84
21) trans-1,2-Dichloroethene	5.783	96	35388	18.395	ug/l	95
22) Diisopropyl ether	6.671	45	130999	19.859	ug/l	99
23) Vinyl Acetate	6.606	43	504977	103.215	ug/l	99
24) 1,1-Dichloroethane	6.571	63	72128	19.553	ug/l	96
25) 2-Butanone	7.483	43	136944	96.795	ug/l	97
26) 2,2-Dichloropropane	7.489	77	64755	19.469	ug/l	100
27) cis-1,2-Dichloroethene	7.489	96	44686	19.222	ug/l	98
28) Bromochloromethane	7.812	49	35788	21.744	ug/l	99
29) Tetrahydrofuran	7.841	42	85661	98.113	ug/l	97
30) Chloroform	7.965	83	76845	20.055	ug/l	97
31) Cyclohexane	8.253	56	58106	16.223	ug/l	97
32) 1,1,1-Trichloroethane	8.165	97	68247	19.832	ug/l	96
36) 1,1-Dichloropropene	8.371	75	49846	18.689	ug/l	99
37) Ethyl Acetate	7.559	43	52425	19.160	ug/l	99
38) Carbon Tetrachloride	8.359	117	56628	19.356	ug/l	96
39) Methylcyclohexane	9.600	83	50490	17.009	ug/l	98
40) Benzene	8.606	78	162968	19.612	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084434.D
 Acq On : 21 Oct 2024 14:22
 Operator : JC\MD
 Sample : VN1021WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBSD01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/22/2024
 Supervised By : Mahesh Dadoda 10/22/2024

Quant Time: Oct 22 01:33:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	29826	20.079	ug/l	97
42) 1,2-Dichloroethane	8.671	62	58026	20.598	ug/l	99
43) Isopropyl Acetate	8.688	43	88049	16.429	ug/l	97
44) Trichloroethene	9.353	130	37926	19.541	ug/l	97
45) 1,2-Dichloropropane	9.624	63	41049	20.882	ug/l	96
46) Dibromomethane	9.712	93	29187	22.002	ug/l	99
47) Bromodichloromethane	9.888	83	61715	20.940	ug/l	98
48) Methyl methacrylate	9.682	41	40695	19.075	ug/l	97
49) 1,4-Dioxane	9.694	88	16333	415.845	ug/l	97
51) 4-Methyl-2-Pentanone	10.447	43	271816	104.354	ug/l	99
52) Toluene	10.629	92	102809	20.289	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	59188	19.701	ug/l	96
54) cis-1,3-Dichloropropene	10.312	75	64640	20.003	ug/l	94
55) 1,1,2-Trichloroethane	11.012	97	40810	22.465	ug/l	96
56) Ethyl methacrylate	10.871	69	59539	19.919	ug/l	98
57) 1,3-Dichloropropane	11.165	76	68931	21.236	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.159	63	141699	102.191	ug/l	97
59) 2-Hexanone	11.194	43	198917	102.176	ug/l	99
60) Dibromochloromethane	11.359	129	46724	21.769	ug/l	98
61) 1,2-Dibromoethane	11.471	107	39150	20.741	ug/l	100
64) Tetrachloroethene	11.106	164	34348	19.288	ug/l	94
65) Chlorobenzene	11.888	112	110738	19.535	ug/l	95
66) 1,1,1,2-Tetrachloroethane	11.959	131	39189	20.071	ug/l	99
67) Ethyl Benzene	11.965	91	184057	18.360	ug/l	99
68) m/p-Xylenes	12.071	106	144045	38.843	ug/l	98
69) o-Xylene	12.400	106	66105	18.841	ug/l	93
70) Styrene	12.412	104	117753	19.814	ug/l	99
71) Bromoform	12.576	173	29825	20.747	ug/l #	99
73) Isopropylbenzene	12.694	105	173820	18.352	ug/l	99
74) N-amyl acetate	12.494	43	73994	17.252	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.935	83	58745	20.627	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	55231m	21.797	ug/l	
77) Bromobenzene	12.976	156	44718	19.574	ug/l	95
78) n-propylbenzene	13.035	91	210981	19.226	ug/l	100
79) 2-Chlorotoluene	13.123	91	129620	18.511	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	148466	19.217	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	18939	17.919	ug/l	99
82) 4-Chlorotoluene	13.218	91	135329	19.198	ug/l	97
83) tert-Butylbenzene	13.435	119	126785	18.464	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	153677	19.745	ug/l	98
85) sec-Butylbenzene	13.618	105	175585	19.162	ug/l	100
86) p-Isopropyltoluene	13.729	119	145633	19.220	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	82553	19.262	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	83788	19.359	ug/l	98
89) n-Butylbenzene	14.059	91	122012	17.758	ug/l	99
90) Hexachloroethane	14.329	117	27657	17.982	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	78881	18.770	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.723	75	10881	18.515	ug/l	97
93) 1,2,4-Trichlorobenzene	15.394	180	38616	18.655	ug/l	99
94) Hexachlorobutadiene	15.506	225	17729	17.175	ug/l	99
95) Naphthalene	15.641	128	115061	16.767	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	37660	18.175	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084434.D
Acq On : 21 Oct 2024 14:22
Operator : JC\MD
Sample : VN1021WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024

Quant Time: Oct 22 01:33:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084434.D
Acq On : 21 Oct 2024 14:22
Operator : JC\MD
Sample : VN1021WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01

Manual Integrations

Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024


Quant Time: Oct 22 01:33:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion:168 Resp: 163086
Ion Ratio Lower Upper
168 100
99 64.3 54.2 81.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#2
Dichlorodifluoromethane
Concen: 16.833 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 85 Resp: 32472
Ion Ratio Lower Upper
85 100
87 31.4 17.6 52.9









#3
Chloromethane
Concen: 16.255 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01

Tgt Ion: 50 Resp: 35804
Ion Ratio Lower Upper
50 100
52 34.5 26.8 40.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#4
Vinyl Chloride
Concen: 17.080 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 62 Resp: 36441
Ion Ratio Lower Upper
62 100
64 33.2 26.2 39.4









#5
Bromomethane
Concen: 16.992 ug/l
RT: 2.959 min Scan# 171
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01

Tgt Ion: 94 Resp: 2391
Ion Ratio Lower Upper
94 100
96 92.4 72.9 109.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#6
Chloroethane
Concen: 16.017 ug/l
RT: 3.118 min Scan# 198
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 64 Resp: 24475
Ion Ratio Lower Upper
64 100
66 31.5 23.5 35.3









#7
Trichlorofluoromethane
Concen: 19.047 ug/l
RT: 3.501 min Scan# 263
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01

Tgt Ion:101 Resp: 6328
Ion Ratio Lower Upper
101 100
103 57.0 51.8 77.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#8
Diethyl Ether
Concen: 19.089 ug/l
RT: 3.959 min Scan# 341
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 74 Resp: 23522
Ion Ratio Lower Upper
74 100
45 92.1 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.273 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01

Tgt Ion:101 Resp: 36689
Ion Ratio Lower Upper
101 100
85 44.0 36.2 54.4
151 75.4 59.3 88.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#10
Methyl Iodide
Concen: 18.016 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion:142 Resp: 44167
Ion Ratio Lower Upper
142 100
127 44.8 35.3 52.9
141 14.8 11.0 16.6





#11

Tert butyl alcohol

Concen: 87.914 ug/l

RT: 5.518 min Scan# 607

Delta R.T. -0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01

Tgt Ion: 59 Resp: 3505

Ion Ratio Lower Upper

59 100





57 11.2 8.2 12.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#12

1,1-Dichloroethene

Concen: 17.747 ug/l

RT: 4.342 min Scan# 406

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22




Tgt Ion: 96 Resp: 32588


Ion Ratio Lower Upper

96 100

61 165.9 128.6 192.8


98 66.6 51.7 77.5





#13
Acrolein
Concen: 53.300 ug/l
RT: 4.189 min Scan# 378
Delta R.T. 0.012 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion: 56 Resp: 2439
Ion Ratio Lower Upper
56 100
55 69.9 57.2 85.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#14
Allyl chloride
Concen: 16.646 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 41 Resp: 53083
Ion Ratio Lower Upper
41 100
39 73.1 59.0 88.6
76 37.9 29.8 44.6





#15

Acrylonitrile

Concen: 101.771 ug/l

RT: 5.718 min Scan# 640

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01

Tgt Ion: 53 Resp: 10247

Ion Ratio Lower Upper

53 100

52 80.5 66.2 99.4





51 34.0 28.8 43.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#16

Acetone

Concen: 91.613 ug/l

RT: 4.430 min Scan# 421

Delta R.T. 0.006 min

Lab File: VN084434.D




Acq: 21 Oct 2024 14:22


Tgt Ion: 43 Resp: 95163

Ion Ratio Lower Upper

43 100


58 33.5 24.4 36.6





#17
Carbon Disulfide
Concen: 14.496 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion: 76 Resp: 84270
Ion Ratio Lower Upper
76 100
78 10.5 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#18
Methyl Acetate
Concen: 24.005 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.012 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 43 Resp: 54347
Ion Ratio Lower Upper
43 100
74 24.7 19.2 28.8









#19
Methyl tert-butyl Ether
Concen: 19.826 ug/l
RT: 5.789 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01

Tgt Ion: 73 Resp: 122872
Ion Ratio Lower Upper
73 100
57 21.3 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#20
Methylene Chloride
Concen: 20.192 ug/l
RT: 5.271 min Scan# 564
Delta R.T. -0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22


Tgt Ion: 84 Resp: 42636
Ion Ratio Lower Upper
84 100
49 108.1 107.2 160.8
51 34.8 31.8 47.8
86 58.8 52.9 79.3





#21
trans-1,2-Dichloroethene
Concen: 18.395 ug/l
RT: 5.783 min Scan# 651
Delta R.T. -0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion: 96 Resp: 3538
Ion Ratio Lower Upper
96 100
61 128.1 108.4 162.6
98 59.8 49.8 74.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#22
Diisopropyl ether
Concen: 19.859 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 45 Resp: 130999
Ion Ratio Lower Upper
45 100
43 51.9 42.3 63.5
87 26.3 21.8 32.6
59 11.8 9.6 14.4









#23
Vinyl Acetate
Concen: 103.215 ug/l
RT: 6.606 min Scan# 790
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01

Tgt Ion: 43 Resp: 50497
Ion Ratio Lower Upper
43 100
86 11.1 8.6 12.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#24
1,1-Dichloroethane
Concen: 19.553 ug/l
RT: 6.571 min Scan# 785
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22


Tgt Ion: 63 Resp: 72128
Ion Ratio Lower Upper
63 100
98 8.1 3.4 10.2
100 5.5 2.0 5.8





#25
2-Butanone
Concen: 96.795 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion: 43 Resp: 136944
Ion Ratio Lower Upper
43 100
72 25.1 21.4 32.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#26
2,2-Dichloropropane
Concen: 19.469 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22


Tgt Ion: 77 Resp: 64755
Ion Ratio Lower Upper
77 100
97 21.7 10.8 32.3





#27
cis-1,2-Dichloroethene
Concen: 19.222 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion: 96 Resp: 44680
Ion Ratio Lower Upper
96 100
61 147.3 0.0 301.8
98 62.9 0.0 128.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#28
Bromochloromethane
Concen: 21.744 ug/l
RT: 7.812 min Scan# 996
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 49 Resp: 35788
Ion Ratio Lower Upper
49 100
129 1.9 0.0 3.6
130 73.3 58.0 87.0





#29

Tetrahydrofuran

Concen: 98.113 ug/l

RT: 7.841 min Scan# 1001

Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01

Tgt Ion: 42 Resp: 8566

Ion Ratio Lower Upper

42 100

72 45.7 34.6 51.8





71 41.7 32.2 48.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#30

Chloroform

Concen: 20.055 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.000 min

Lab File: VN084434.D




Acq: 21 Oct 2024 14:22


Tgt Ion: 83 Resp: 76845

Ion Ratio Lower Upper

83 100


85 65.7 51.0 76.4





#31
Cyclohexane
Concen: 16.223 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion: 56 Resp: 58100
Ion Ratio Lower Upper
56 100
69 32.9 23.8 35.6
84 84.1 68.8 103.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#32
1,1,1-Trichloroethane
Concen: 19.832 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 97 Resp: 68247
Ion Ratio Lower Upper
97 100
99 60.2 51.2 76.8
61 45.1 37.5 56.3





#33

1,2-Dichloroethane-d4

Concen: 51.056 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01

Tgt Ion: 65 Resp: 12345

Ion Ratio Lower Upper

65 100





67 51.6 0.0 102.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22




Tgt Ion:114 Resp: 278516


Ion Ratio Lower Upper

114 100

63 21.6 0.0 43.8

88 17.2 0.0 31.6





#35

Dibromofluoromethane

Concen: 52.829 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01

Tgt Ion:113 Resp: 9700

Ion Ratio Lower Upper

113 100

111 104.0 83.3 124.9





192 18.2 13.5 20.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#36

1,1-Dichloropropene

Concen: 18.689 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22




Tgt Ion: 75 Resp: 49846


Ion Ratio Lower Upper

75 100

110 34.2 16.9 50.7

77 32.0 24.7 37.1





#37

Ethyl Acetate

Concen: 19.160 ug/l

RT: 7.559 min Scan# 91

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01

Tgt Ion: 43 Resp: 5242

Ion Ratio Lower Upper

43 100

61 13.7 10.7 16.1





70 11.4 8.8 13.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#38

Carbon Tetrachloride

Concen: 19.356 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22




Tgt Ion:117 Resp: 56628


Ion Ratio Lower Upper

117 100

119 93.0 78.7 118.1

121 29.8 24.4 36.6





#39

Methylcyclohexane

Concen: 17.009 ug/l

RT: 9.600 min Scan# 11300

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01

Tgt Ion: 83 Resp: 50490

Ion Ratio Lower Upper

83 100

55 81.3 62.8 94.2





98 48.7 38.8 58.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#40

Benzene

Concen: 19.612 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VN084434.D




Acq: 21 Oct 2024 14:22


Tgt Ion: 78 Resp: 162968

Ion Ratio Lower Upper

78 100


77 24.3 19.1 28.7





#41
Methacrylonitrile
Concen: 20.079 ug/l
RT: 7.777 min Scan# 990
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion: 41 Resp: 29820
Ion Ratio Lower Upper
41 100
39 55.1 43.3 64.9
67 74.8 56.1 84.1
52 31.6 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#42
1,2-Dichloroethane
Concen: 20.598 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 62 Resp: 58026
Ion Ratio Lower Upper
62 100
98 9.3 0.0 18.2





#43

Isopropyl Acetate

Concen: 16.429 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01





Tgt Ion: 43	Resp: 88049
Ion Ratio	Lower Upper
43	100
61	27.4 20.4 30.6
87	13.9 10.3 15.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#44

Trichloroethene

Concen: 19.541 ug/l




RT: 9.353 min Scan# 1258


Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Tgt Ion:130	Resp: 37926
Ion Ratio	Lower Upper
130	100
95	98.9 0.0 203.0





#45

1,2-Dichloropropane

Concen: 20.882 ug/l

RT: 9.624 min Scan# 1104

Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01

Tgt Ion: 63 Resp: 41049

Ion Ratio Lower Upper

63 100





65 33.3 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#46

Dibromomethane

Concen: 22.002 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22




Tgt Ion: 93 Resp: 29187


Ion Ratio Lower Upper

93 100

95 82.7 65.0 97.4


174 90.6 71.8 107.8





#47
Bromodichloromethane
Concen: 20.940 ug/l
RT: 9.888 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion: 83 Resp: 61715
Ion Ratio Lower Upper
83 100
85 63.1 51.5 77.3
127 7.7 5.7 8.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#48
Methyl methacrylate
Concen: 19.075 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 41 Resp: 40695
Ion Ratio Lower Upper
41 100
69 87.8 67.1 100.7
39 58.5 45.8 68.6





#49

1,4-Dioxane

Concen: 415.845 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01

Tgt Ion: 88 Resp: 1633

Ion Ratio Lower Upper

88 100

43 29.5 25.3 37.9





58 72.1 56.4 84.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#50

Toluene-d8

Concen: 51.812 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN084434.D




Acq: 21 Oct 2024 14:22


Tgt Ion: 98 Resp: 350013

Ion Ratio Lower Upper

98 100

100 66.0 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 104.354 ug/l
RT: 10.447 min Scan# 1443
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01

Tgt Ion: 43 Resp: 271810
Ion Ratio Lower Upper
43 100
58 40.2 31.6 47.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#52
Toluene
Concen: 20.289 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22


Tgt Ion: 92 Resp: 102809
Ion Ratio Lower Upper
92 100
91 171.1 137.3 205.9





#53
t-1,3-Dichloropropene
Concen: 19.701 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion: 75 Resp: 5918
Ion Ratio Lower Upper
75 100
77 34.0 25.4 38.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#54
cis-1,3-Dichloropropene
Concen: 20.003 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22


Tgt Ion: 75 Resp: 64640
Ion Ratio Lower Upper
75 100
77 33.2 24.2 36.2
39 44.7 38.7 58.1





#55
1,1,2-Trichloroethane
Concen: 22.465 ug/l
RT: 11.012 min Scan# 1540
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion: 97 Resp: 40810
Ion Ratio Lower Upper
97 100
83 90.4 71.0 106.4
85 59.0 41.7 62.5
99 65.5 50.1 75.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#56
Ethyl methacrylate
Concen: 19.919 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 69 Resp: 59539
Ion Ratio Lower Upper
69 100
41 65.6 53.3 79.9
39 34.7 28.8 43.2





#57

1,3-Dichloropropane

Concen: 21.236 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01

Tgt Ion: 76 Resp: 6893

Ion Ratio Lower Upper

76 100





78 33.4 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#58

2-Chloroethyl Vinyl ether

Concen: 102.191 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN084434.D




Acq: 21 Oct 2024 14:22


Tgt Ion: 63 Resp: 141699

Ion Ratio Lower Upper

63 100


106 29.0 21.8 32.6





#59
2-Hexanone
Concen: 102.176 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion: 43 Resp: 198917
Ion Ratio Lower Upper
43 100
58 54.9 27.3 81.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#60
Dibromochloromethane
Concen: 21.769 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22


Tgt Ion:129 Resp: 46724
Ion Ratio Lower Upper
129 100
127 75.3 38.3 114.9





#61
1,2-Dibromoethane
Concen: 20.741 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion:107 Resp: 39150
Ion Ratio Lower Upper
107 100
109 93.8 75.4 113.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#62
4-Bromofluorobenzene
Concen: 52.449 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22


Tgt Ion: 95 Resp: 129081
Ion Ratio Lower Upper
95 100
174 73.0 0.0 145.2
176 70.7 0.0 140.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion:117 Resp: 255670
Ion Ratio Lower Upper
117 100
82 56.8 47.2 70.8
119 32.7 25.4 38.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#64
Tetrachloroethene
Concen: 19.288 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion:164 Resp: 34348
Ion Ratio Lower Upper
164 100
166 121.7 101.0 151.4
129 95.8 80.8 121.2
131 88.2 78.1 117.1









#65
Chlorobenzene
Concen: 19.535 ug/l
RT: 11.888 min Scan# 110738
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01

Tgt Ion:112 Resp: 110738
Ion Ratio Lower Upper
112 100
114 30.9 27.0 40.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 20.071 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion:131 Resp: 39189
Ion Ratio Lower Upper
131 100
133 95.4 47.9 143.7
119 67.4 32.8 98.3









#67
Ethyl Benzene
Concen: 18.360 ug/l
RT: 11.965 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01

Tgt Ion: 91 Resp: 18405
Ion Ratio Lower Upper
91 100
106 29.9 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#68
m/p-Xylenes
Concen: 38.843 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion:106 Resp: 144045
Ion Ratio Lower Upper
106 100
91 206.2 167.1 250.7









#69
o-Xylene
Concen: 18.841 ug/l
RT: 12.400 min Scan# 1775
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01

Tgt Ion:106 Resp: 6610
Ion Ratio Lower Upper
106 100
91 230.9 110.2 330.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#70
Styrene
Concen: 19.814 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22


Tgt Ion:104 Resp: 117753
Ion Ratio Lower Upper
104 100
78 51.9 42.7 64.1
103 54.6 43.6 65.4





#71
Bromoform
Concen: 20.747 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01




Tgt Ion:173 Resp: 2982
Ion Ratio Lower Upper
173 100
175 47.7 23.5 70.6
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Tgt Ion:152 Resp: 124118
Ion Ratio Lower Upper
152 100
115 62.4 31.3 93.9
150 163.2 0.0 349.8





#73
Isopropylbenzene
Concen: 18.352 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion:105 Resp: 173820
Ion Ratio Lower Upper
105 100
120 26.1 12.9 38.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#74
N-ethyl acetate
Concen: 17.252 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22


Tgt Ion: 43 Resp: 73994
Ion Ratio Lower Upper
43 100
70 45.7 34.4 51.6
55 26.6 19.6 29.4
61 24.3 19.6 29.4





#75
1,1,2,2-Tetrachloroethane
Concen: 20.627 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion: 83 Resp: 58749
Ion Ratio Lower Upper
83 100
131 10.1 4.9 14.7
85 64.3 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#76
1,2,3-Trichloropropane
Concen: 21.797 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22


Tgt Ion: 75 Resp: 55231
Ion Ratio Lower Upper
75 100
77 177.9 96.9 290.7





#77
Bromobenzene
Concen: 19.574 ug/l
RT: 12.976 min Scan# 1875
Delta R.T. -0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion:156 Resp: 44718
Ion Ratio Lower Upper
156 100
77 219.8 115.1 345.3
158 95.3 48.9 146.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#78
n-propylbenzene
Concen: 19.226 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 91 Resp: 210981
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4









#79
2-Chlorotoluene
Concen: 18.511 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01

Tgt Ion: 91 Resp: 129620
Ion Ratio Lower Upper
91 100
126 32.2 15.7 47.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#80
1,3,5-Trimethylbenzene
Concen: 19.217 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22


Tgt Ion:105 Resp: 148466
Ion Ratio Lower Upper
105 100
120 47.9 23.7 71.1





#81
trans-1,4-Dichloro-2-butene
Concen: 17.919 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion: 75 Resp: 18939
Ion Ratio Lower Upper
75 100
53 97.4 78.0 117.0
89 47.3 35.5 53.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#82
4-Chlorotoluene
Concen: 19.198 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 91 Resp: 135329
Ion Ratio Lower Upper
91 100
126 32.9 15.7 47.1





#83

tert-Butylbenzene

Concen: 18.464 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01

Tgt Ion:119 Resp: 126785

Ion Ratio Lower Upper

119 100

91 66.9 34.1 102.3





134 23.7 12.8 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#84

1,2,4-Trimethylbenzene

Concen: 19.745 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.000 min

Lab File: VN084434.D




Acq: 21 Oct 2024 14:22


Tgt Ion:105 Resp: 153677

Ion Ratio Lower Upper

105 100


120 43.5 22.3 66.8





#85
 sec-Butylbenzene
 Concen: 19.162 ug/l
 RT: 13.618 min Scan# 1983
 Delta R.T. 0.000 min
 Lab File: VN084434.D
 Acq: 21 Oct 2024 14:22




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBSD01




Tgt Ion:105 Resp: 17558
 Ion Ratio Lower Upper
 105 100
 134 19.4 9.8 29.4

Manual Integrations
 APPROVED



Reviewed By :John Carlone 10/22/2024
 Supervised By :Mahesh Dadoda 10/22/2024




#86
 p-Isopropyltoluene
 Concen: 19.220 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN084434.D
 Acq: 21 Oct 2024 14:22



Tgt Ion:119 Resp: 145633
 Ion Ratio Lower Upper
 119 100
 134 24.5 12.6 37.8
 91 24.9 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 19.262 ug/l

RT: 13.735 min Scan# 2002

Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01

Tgt Ion:146 Resp: 8255

Ion Ratio Lower Upper

146 100

111 42.5 21.8 65.3





148 63.4 31.9 95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 10/22/2024

Supervised By :Mahesh Dadoda 10/22/2024



#88

1,4-Dichlorobenzene

Concen: 19.359 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22




Tgt Ion:146 Resp: 83788


Ion Ratio Lower Upper

146 100

111 41.6 21.2 63.6

148 61.8 31.9 95.9









#89
n-Butylbenzene
Concen: 17.758 ug/l
RT: 14.059 min Scan# 2057
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01

Tgt Ion: 91 Resp: 12201
Ion Ratio Lower Upper
91 100
92 51.8 26.6 79.7
134 24.3 12.3 36.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#90
Hexachloroethane
Concen: 17.982 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22


Tgt Ion:117 Resp: 27657
Ion Ratio Lower Upper
117 100
201 70.8 34.4 103.2





#91
1,2-Dichlorobenzene
Concen: 18.770 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion:146 Resp: 7888
Ion Ratio Lower Upper
146 100
111 43.2 22.0 66.0
148 64.1 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 18.515 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22


Tgt Ion: 75 Resp: 10881
Ion Ratio Lower Upper
75 100
155 73.3 38.1 114.5
157 96.3 49.6 149.0





#93
1,2,4-Trichlorobenzene
Concen: 18.655 ug/l
RT: 15.394 min Scan# 21
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion:180 Resp: 38610
Ion Ratio Lower Upper
180 100
182 92.4 47.0 141.0
145 33.0 16.4 49.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#94
Hexachlorobutadiene
Concen: 17.175 ug/l
RT: 15.506 min Scan# 2304
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22


Tgt Ion:225 Resp: 17729
Ion Ratio Lower Upper
225 100
223 65.9 32.4 97.2
227 65.3 32.4 97.0





#95
Naphthalene
Concen: 16.767 ug/l
RT: 15.641 min Scan# 2326
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01



Tgt Ion:128 Resp: 11506
Ion Ratio Lower Upper
128 100
127 12.4 10.5 15.7
129 10.8 8.7 13.1




Manual Integrations
APPROVED

Reviewed By :John Carlone 10/22/2024
Supervised By :Mahesh Dadoda 10/22/2024



#96
1,2,3-Trichlorobenzene
Concen: 18.175 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion:180 Resp: 37660
Ion Ratio Lower Upper
180 100
182 95.4 48.2 144.6
145 36.1 17.3 51.9



Manual Integration Report

Sequence:	vn093024	Instrument	MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC100	VN084213.D	1,2,3-Trichloropropane	JOHN	10/1/2024 9:36:48 AM	MMDadoda	10/1/2024 5:34:17 PM	Peak Integrated by Software
VSTDICCC050	VN084214.D	1,2,3-Trichloropropane	JOHN	10/1/2024 9:36:52 AM	MMDadoda	10/1/2024 5:34:19 PM	Peak Integrated by Software
VSTDICC020	VN084215.D	1,2,3-Trichloropropane	JOHN	10/1/2024 9:36:57 AM	MMDadoda	10/1/2024 5:34:21 PM	Peak Integrated by Software
VSTDICC010	VN084216.D	1,2,3-Trichloropropane	JOHN	10/1/2024 9:37:06 AM	MMDadoda	10/1/2024 5:34:22 PM	Peak Integrated by Software
VSTDICC005	VN084217.D	1,1,2-Trichlorotrifluoroethane	JOHN	10/1/2024 9:37:11 AM	MMDadoda	10/1/2024 5:34:24 PM	Peak Integrated by Software
VSTDICC005	VN084217.D	1,2,3-Trichloropropane	JOHN	10/1/2024 9:37:11 AM	MMDadoda	10/1/2024 5:34:24 PM	Peak Integrated by Software
VSTDICC005	VN084217.D	Isopropyl Acetate	JOHN	10/1/2024 9:37:11 AM	MMDadoda	10/1/2024 5:34:24 PM	Peak Integrated by Software
VSTDICC005	VN084217.D	trans-1,2-Dichloroethene	JOHN	10/1/2024 9:37:11 AM	MMDadoda	10/1/2024 5:34:24 PM	Peak Integrated by Software
VSTDICC005	VN084217.D	Vinyl Acetate	JOHN	10/1/2024 9:37:11 AM	MMDadoda	10/1/2024 5:34:24 PM	Peak Integrated by Software
VSTDICC001	VN084218.D	1,2,3-Trichloropropane	JOHN	10/1/2024 9:37:57 AM	MMDadoda	10/1/2024 5:34:25 PM	Peak Integrated by Software
VSTDICC001	VN084218.D	1,4-Dichlorobenzene	JOHN	10/1/2024 9:37:57 AM	MMDadoda	10/1/2024 5:34:25 PM	Peak Integrated by Software
VSTDICC001	VN084218.D	Allyl chloride	JOHN	10/1/2024 9:37:57 AM	MMDadoda	10/1/2024 5:34:25 PM	Peak Integrated by Software
VSTDICC001	VN084218.D	Isopropyl Acetate	JOHN	10/1/2024 9:37:57 AM	MMDadoda	10/1/2024 5:34:25 PM	Peak Integrated by Software

Manual Integration Report

Sequence:

vn093024

Instrument

MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDIC001	VN084218.D	Vinyl Acetate	JOHN	10/1/2024 9:37:57 AM	MMDadoda	10/1/2024 5:34:25 PM	Peak Integrated by Software
VSTDICV050	VN084220.D	1,2,3-Trichloropropane	JOHN	10/1/2024 9:38:01 AM	MMDadoda	10/1/2024 5:34:27 PM	Peak Integrated by Software
VSTDCCC050	VN084229.D	1,2,3-Trichloropropane	JOHN	10/1/2024 9:38:17 AM	MMDadoda	10/1/2024 5:34:32 PM	Peak Integrated by Software

Manual Integration Report

Sequence:	vn102124	Instrument	MSVOA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VN084428.D	1,2,3-Trichloropropane	JOHN	10/22/2024 9:41:28 AM	MMDadoda	10/22/2024 3:37:06 PM	Peak Integrated by Software
VN1021WBS01	VN084433.D	1,2,3-Trichloropropane	JOHN	10/22/2024 9:41:37 AM	MMDadoda	10/22/2024 3:37:07 PM	Peak Integrated by Software
VN1021WBSD0 1	VN084434.D	1,2,3-Trichloropropane	JOHN	10/22/2024 9:41:41 AM	MMDadoda	10/22/2024 3:37:09 PM	Peak Integrated by Software
VSTDCCC050	VN084448.D	1,2,3-Trichloropropane	JOHN	10/22/2024 9:41:59 AM	MMDadoda	10/22/2024 3:37:13 PM	Peak Integrated by Software

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN093024

Review By	John Carlone	Review On	10/1/2024 9:40:01 AM
Supervise By	Mahesh Dadoda	Supervise On	10/1/2024 5:34:39 PM
SubDirectory	VN093024	HP Acquire Method	HP Processing Method 82N093024W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP130570		
Initial Calibration Stds	VP130580,VP130582,VP130583,VP130584,VP130585,VP130586		
CCC	VP130571,VP130572		
Internal Standard/PEM			
ICV/I.BLK	VP130587		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN084211.D	30 Sep 2024 09:24	JC\MD	Ok
2	VSTDCCC050	VN084212.D	30 Sep 2024 11:45	JC\MD	Not Ok
3	VSTDICC100	VN084213.D	30 Sep 2024 12:25	JC\MD	Ok,M
4	VSTDICCC050	VN084214.D	30 Sep 2024 12:49	JC\MD	Ok,M
5	VSTDICC020	VN084215.D	30 Sep 2024 13:13	JC\MD	Ok,M
6	VSTDICC010	VN084216.D	30 Sep 2024 13:37	JC\MD	Ok,M
7	VSTDICC005	VN084217.D	30 Sep 2024 14:00	JC\MD	Ok,M
8	VSTDICC001	VN084218.D	30 Sep 2024 14:48	JC\MD	Ok,M
9	IBLK	VN084219.D	30 Sep 2024 15:12	JC\MD	Ok
10	VSTDICV050	VN084220.D	30 Sep 2024 15:36	JC\MD	Ok,M
11	VN0930MBL01	VN084221.D	30 Sep 2024 16:11	JC\MD	Ok
12	VN0930WBL01	VN084222.D	30 Sep 2024 16:35	JC\MD	Ok
13	VN0930WBS01	VN084223.D	30 Sep 2024 17:42	JC\MD	Ok,M
14	VN0930WBSD01	VN084224.D	30 Sep 2024 18:06	JC\MD	Ok,M
15	P4116-24	VN084225.D	30 Sep 2024 18:30	JC\MD	Dilution
16	P4116-25	VN084226.D	30 Sep 2024 18:54	JC\MD	Dilution
17	P4116-26	VN084227.D	30 Sep 2024 19:18	JC\MD	Dilution
18	P4116-27	VN084228.D	30 Sep 2024 19:42	JC\MD	Dilution
19	VSTDCCC050	VN084229.D	30 Sep 2024 20:06	JC\MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN102124

Review By	John Carlone	Review On	10/22/2024 9:50:25 AM
Supervise By	Maresh Dadoda	Supervise On	10/22/2024 3:37:17 PM
SubDirectory	VN102124	HP Acquire Method	HP Processing Method 82N093024W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP130983		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP130986,VP130987		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN084427.D	21 Oct 2024 11:14	JC\MD	Ok
2	VSTDCCC050	VN084428.D	21 Oct 2024 11:38	JC\MD	Ok,M
3	VN1021MBL01	VN084429.D	21 Oct 2024 12:12	JC\MD	Ok
4	VN1021WBL01	VN084430.D	21 Oct 2024 12:36	JC\MD	Ok
5	VN1021MBS01	VN084431.D	21 Oct 2024 13:00	JC\MD	Not Ok
6	P4417-04	VN084432.D	21 Oct 2024 13:24	JC\MD	Ok
7	VN1021WBS01	VN084433.D	21 Oct 2024 13:48	JC\MD	Ok,M
8	VN1021WBSD01	VN084434.D	21 Oct 2024 14:22	JC\MD	Ok,M
9	P4438-03	VN084435.D	21 Oct 2024 14:53	JC\MD	Ok
10	P4438-01	VN084436.D	21 Oct 2024 15:17	JC\MD	Ok
11	P4440-02	VN084437.D	21 Oct 2024 15:41	JC\MD	Ok
12	P4438-02	VN084438.D	21 Oct 2024 16:05	JC\MD	Ok
13	P4438-04	VN084439.D	21 Oct 2024 16:29	JC\MD	Ok
14	P4438-05	VN084440.D	21 Oct 2024 16:53	JC\MD	Ok
15	P4438-06	VN084441.D	21 Oct 2024 17:17	JC\MD	Ok
16	P4438-07	VN084442.D	21 Oct 2024 17:40	JC\MD	Ok
17	P4440-01	VN084443.D	21 Oct 2024 18:04	JC\MD	Ok
18	P4471-03	VN084444.D	21 Oct 2024 18:28	JC\MD	Ok
19	P4460-05	VN084445.D	21 Oct 2024 18:52	JC\MD	Ok
20	P4460-06	VN084446.D	21 Oct 2024 19:16	JC\MD	Ok
21	VN1021MBS02	VN084447.D	21 Oct 2024 19:40	JC\MD	Ok,M

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN102124

Review By	John Carlone	Review On	10/22/2024 9:50:25 AM
Supervise By	Mahesh Dadoda	Supervise On	10/22/2024 3:37:17 PM
SubDirectory	VN102124	HP Acquire Method	HP Processing Method 82N093024W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP130983		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP130986,VP130987		

22	VSTDCCC050	VN084448.D	21 Oct 2024 20:04	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN093024

Review By	John Carlone	Review On	10/1/2024 9:40:01 AM
Supervise By	Mahesh Dadoda	Supervise On	10/1/2024 5:34:39 PM
SubDirectory	VN093024	HP Acquire Method	HP Processing Method 82N093024W.M

STD. NAME	STD REF.#
Tune/Reschk	VP130570
Initial Calibration Stds	VP130580,VP130582,VP130583,VP130584,VP130585,VP130586
CCC	VP130571,VP130572
Internal Standard/PEM	
ICV/I.BLK	VP130587
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN084211.D	30 Sep 2024 09:24		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN084212.D	30 Sep 2024 11:45	Need ICAL	JC\MD	Not Ok
3	VSTDICC100	VSTDICC100	VN084213.D	30 Sep 2024 12:25		JC\MD	Ok,M
4	VSTDICCC050	VSTDICCC050	VN084214.D	30 Sep 2024 12:49		JC\MD	Ok,M
5	VSTDICC020	VSTDICC020	VN084215.D	30 Sep 2024 13:13		JC\MD	Ok,M
6	VSTDICC010	VSTDICC010	VN084216.D	30 Sep 2024 13:37		JC\MD	Ok,M
7	VSTDICC005	VSTDICC005	VN084217.D	30 Sep 2024 14:00		JC\MD	Ok,M
8	VSTDICC001	VSTDICC001	VN084218.D	30 Sep 2024 14:48		JC\MD	Ok,M
9	IBLK	IBLK	VN084219.D	30 Sep 2024 15:12		JC\MD	Ok
10	VSTDICV050	ICVVN093024	VN084220.D	30 Sep 2024 15:36		JC\MD	Ok,M
11	VN0930MBL01	VN0930MBL01	VN084221.D	30 Sep 2024 16:11		JC\MD	Ok
12	VN0930WBL01	VN0930WBL01	VN084222.D	30 Sep 2024 16:35		JC\MD	Ok
13	VN0930WBS01	VN0930WBS01	VN084223.D	30 Sep 2024 17:42		JC\MD	Ok,M
14	VN0930WBSD01	VN0930WBSD01	VN084224.D	30 Sep 2024 18:06		JC\MD	Ok,M
15	P4116-24	RE132D5-20240918	VN084225.D	30 Sep 2024 18:30	vial A pH<2 Need 2x	JC\MD	Dilution
16	P4116-25	RE132D6-20240918	VN084226.D	30 Sep 2024 18:54	vial A pH<2 Need 40X	JC\MD	Dilution
17	P4116-26	RE132D7-20240918	VN084227.D	30 Sep 2024 19:18	vial A pH<2 Need 40X	JC\MD	Dilution
18	P4116-27	DUP05-20240918	VN084228.D	30 Sep 2024 19:42	vial A pH<2 Need 2x	JC\MD	Dilution

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN093024

Review By	John Carlone	Review On	10/1/2024 9:40:01 AM
Supervise By	Mahesh Dadoda	Supervise On	10/1/2024 5:34:39 PM
SubDirectory	VN093024	HP Acquire Method	HP Processing Method 82N093024W.M

STD. NAME	STD REF.#
Tune/Reschk	VP130570
Initial Calibration Stds	VP130580,VP130582,VP130583,VP130584,VP130585,VP130586
CCC	VP130571,VP130572
Internal Standard/PEM	
ICV/I.BLK	VP130587
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	VSTDCCC050	VSTDCCC050EC	VN084229.D	30 Sep 2024 20:06		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN102124

Review By	John Carlone	Review On	10/22/2024 9:50:25 AM
Supervise By	Maresh Dadoda	Supervise On	10/22/2024 3:37:17 PM
SubDirectory	VN102124	HP Acquire Method	HP Processing Method 82N093024W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP130983
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP130986,VP130987

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN084427.D	21 Oct 2024 11:14		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN084428.D	21 Oct 2024 11:38	V13516	JC\MD	Ok,M
3	VN1021MBL01	VN1021MBL01	VN084429.D	21 Oct 2024 12:12		JC\MD	Ok
4	VN1021WBL01	VN1021WBL01	VN084430.D	21 Oct 2024 12:36		JC\MD	Ok
5	VN1021MBS01	VN1021MBS01	VN084431.D	21 Oct 2024 13:00	Recovery failed	JC\MD	Not Ok
6	P4417-04	HOPPER-B-SOLID	VN084432.D	21 Oct 2024 13:24		JC\MD	Ok
7	VN1021WBS01	VN1021WBS01	VN084433.D	21 Oct 2024 13:48		JC\MD	Ok,M
8	VN1021WBSD01	VN1021WBSD01	VN084434.D	21 Oct 2024 14:22		JC\MD	Ok,M
9	P4438-03	VPB190-HYD-2024101	VN084435.D	21 Oct 2024 14:53	vial B pH<2	JC\MD	Ok
10	P4438-01	BP-VPB-190-TB-20241	VN084436.D	21 Oct 2024 15:17	vial B pH<2 TB	JC\MD	Ok
11	P4440-02	BPOW6-11-TB-202410	VN084437.D	21 Oct 2024 15:41	vial B pH<2 TB	JC\MD	Ok
12	P4438-02	BP-VPB-190-EB-20241	VN084438.D	21 Oct 2024 16:05	vial B pH<2 EB	JC\MD	Ok
13	P4438-04	BP-VPB-190-GW-58-60	VN084439.D	21 Oct 2024 16:29	vial B pH<2	JC\MD	Ok
14	P4438-05	BP-VPB-190-GW-98-10	VN084440.D	21 Oct 2024 16:53	vial B pH<2	JC\MD	Ok
15	P4438-06	BP-VPB-190-GW-158-1	VN084441.D	21 Oct 2024 17:17	vial B pH<2	JC\MD	Ok
16	P4438-07	BP-VPB-190-GW-198-2	VN084442.D	21 Oct 2024 17:40	vial B pH<2	JC\MD	Ok
17	P4440-01	BPOW6-11-HYD-20241	VN084443.D	21 Oct 2024 18:04	vial B pH<2	JC\MD	Ok
18	P4471-03	TB100824	VN084444.D	21 Oct 2024 18:28	vial A pH<2 TB	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN102124

Review By	John Carlone	Review On	10/22/2024 9:50:25 AM
Supervise By	Maresh Dadoda	Supervise On	10/22/2024 3:37:17 PM
SubDirectory	VN102124	HP Acquire Method	HP Processing Method 82N093024W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP130983		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP130986,VP130987		

19	P4460-05	TB-10182024	VN084445.D	21 Oct 2024 18:52	vial A pH<2 TB	JC\MD	Ok
20	P4460-06	WB-303-SW	VN084446.D	21 Oct 2024 19:16	vial A pH<2	JC\MD	Ok
21	VN1021MBS02	VN1021MBS02	VN084447.D	21 Oct 2024 19:40		JC\MD	Ok,M
22	VSTDCCC050	VSTDCCC050EC	VN084448.D	21 Oct 2024 20:04		JC\MD	Ok,M

M : Manual Integration

Prep Standard - Chemical Standard Summary

Order ID : P4440

Test : VOCMS Group1

Prepbatch ID :

Sequence ID/Qc Batch ID: vn102124,

Standard ID :

VP128290,VP128298,VP128762,VP128766,VP130314,VP130430,VP130836,VP130983,VP130986,VP130987,

Chemical ID :

V12794,V12798,V13390,V13444,V13462,V13463,V13707,V13804,V13806,V14018,V14019,V14108,V14145,V14147,V14148,V14171,V14172,V14212,V14215,V14217,V14288,V14436,V14443,V14546,V14547,V14548,W 3112,

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
218	BFB, 25PPM	VP128290	06/10/2024	11/23/2024	Semsettin Yesilyurt	None	None	Maresh Dadoda
06/12/2024								

FROM 0.25000ml of V13390 + 24.75000ml of V14148 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
247	8260 Internal Standard, 250PPM	VP128298	06/10/2024	11/23/2024	Semsettin Yesilyurt	None	None	Maresh Dadoda
06/12/2024								

FROM 0.10000ml of V14288 + 9.90000ml of V14148 = Final Quantity: 10.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1810	8260 Working Std(2-CVE)-800ppm	VP128762	07/01/2024	12/11/2024	Semsettin Yesilyurt	None	None	Maresh Dadoda
								07/02/2024

FROM 0.50000ml of V12798 + 1.50000ml of V12794 + 23.00000ml of V14147 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
719	8260 Working STD (BCM)-First source, 400PPM	VP128766	07/01/2024	12/11/2024	Semsettin Yesilyurt	None	None	Maresh Dadoda
								07/02/2024

FROM 1.50000ml of V13462 + 1.50000ml of V13463 + 12.00000ml of V14147 = Final Quantity: 15.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
257	8260 Calibration Working STD Mix-First source, 160PPM	VP130314	09/16/2024	10/26/2024	Semsettin Yesilyurt	None	None	Maresh Dadoda 09/18/2024

FROM 0.40000ml of V13444 + 1.00000ml of V13804 + 1.00000ml of V13806 + 1.00000ml of V14018 + 1.00000ml of V14019 +
1.00000ml of V14171 + 1.00000ml of V14172 + 1.00000ml of V14217 + 1.00000ml of V14436 + 1.00000ml of V14443 +
1.50000ml of V14212 + 1.50000ml of V14215 + 10.60000ml of V14145 + 2.00000ml of V14108 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
617	8260 Surrogate, 400PPM	VP130430	09/20/2024	02/28/2025	Semsettin Yesilyurt	None	None	Maresh Dadoda 09/26/2024

FROM 0.40000ml of V13707 + 24.60000ml of V14145 = Final Quantity: 25.000 ml

VOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
51	8260 Working STD (Acrolein) -first source, 800PPM	VP130836	10/14/2024	11/07/2024	Semsettin Yesilyurt	None	None	Maresh Dadoda
								10/16/2024

FROM 1.00000ml of V14548 + 1.50000ml of V14546 + 1.50000ml of V14547 + 21.00000ml of V14145 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
589	BFB TUNE CHECK	VP130983	10/21/2024	10/22/2024	John Carlone	None	None	Maresh Dadoda
								10/23/2024

FROM 39.98400ml of W3112 + 0.01600ml of VP128290 = Final Quantity: 40.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	VP130986	10/21/2024	10/22/2024	John Carlone	None	None	Mahesh Dadoda 10/23/2024
<u>FROM</u> 39.94450ml of W3112 + 0.00500ml of VP128766 + 0.00500ml of VP130430 + 0.00800ml of VP128298 + 0.01250ml of VP128762 + 0.01250ml of VP130314 + 0.01250ml of VP130836 = Final Quantity: 40.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
620	50 PPB CCC, 8260-Water	VP130987	10/21/2024	10/22/2024	John Carlone	None	None	Mahesh Dadoda 10/23/2024
<u>FROM</u> 39.94450ml of W3112 + 0.00500ml of VP128766 + 0.00500ml of VP130430 + 0.00800ml of VP128298 + 0.01250ml of VP128762 + 0.01250ml of VP130314 + 0.01250ml of VP130836 = Final Quantity: 40.000 ml								

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95318 / 2-Chloroethyl Vinyl Ether (Min = 5)	121321	12/13/2024	06/25/2024 / SAM	03/30/2022 / SAM	V12794

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95318 / 2-Chloroethyl Vinyl Ether (Min = 5)	121321	12/13/2024	06/25/2024 / SAM	03/30/2022 / SAM	V12798

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30067 / BFB tuneing solution	A0191805	12/08/2024	12/08/2023 / SAM	01/13/2023 / SAM	V13390

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30470 / VOA Stock Solution, tert-butanol std, 1mL, P&TM	A0181905	12/14/2024	06/14/2024 / SAM	01/23/2023 / SAM	V13444

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	01/01/2025	07/01/2024 / SAM	01/27/2023 / SAM	V13462

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30225 / VOA Mix, bromochloromethane, 2000ug/mL, P&TM, 1mL/ampul	A0193071	01/01/2025	07/01/2024 / SAM	01/27/2023 / SAM	V13463

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555582 / Custom Mixture, 8260 A/B Surrogate Mix [CS 5179-2]	A0196865	06/10/2025	06/10/2024 / SAM	04/12/2023 / SAM	V13707

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix, 500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0194279	03/16/2025	09/16/2024 / SAM	05/31/2023 / SAM	V13804

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix, 500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0194279	03/16/2025	09/16/2024 / SAM	05/31/2023 / SAM	V13806

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	032922	03/16/2025	09/16/2024 / SAM	11/22/2023 / SAM	V14018

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95319 / Revised Additions Mix (Min = 5)	032922	03/16/2025	09/16/2024 / SAM	11/22/2023 / SAM	V14019

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555408 / Custom Standard, Vinyl Acetate Standard w/ Grav [CS 5066-6] TWO SEPARATE	A0205179	03/16/2025	09/16/2024 / SAM	12/22/2023 / SAM	V14108

LOTS

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	02/28/2025	08/29/2024 / SAM	02/06/2024 / SAM	V14145

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	12/11/2024	06/11/2024 / pedro	02/06/2024 / SAM	V14147

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	11/23/2024	05/23/2024 / pedro	02/06/2024 / SAM	V14148

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	03/16/2025	09/16/2024 / SAM	02/20/2024 / SAM	V14171

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	95317 / Universal VOA Mega Mix (Min order = 5)	021624	03/16/2025	09/16/2024 / SAM	02/20/2024 / SAM	V14172

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A0200785	03/16/2025	09/16/2024 / SAM	02/28/2024 / SAM	V14212

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A0200785	03/16/2025	09/16/2024 / SAM	02/28/2024 / SAM	V14215

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A0200785	03/16/2025	09/16/2024 / SAM	02/28/2024 / SAM	V14217

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555581 / Custom Standard, 8260 Internal Std [CS 5179-1]	A0210184	06/10/2025	06/10/2024 / SAM	04/15/2024 / SAM	V14288

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	09/30/2025	09/16/2024 / SAM	08/15/2024 / SAM	V14436

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30489 / VOA Mix, 8260B Acetates Mix, P&TM, 1mL	A0209618	03/16/2025	09/16/2024 / SAM	08/15/2024 / SAM	V14443

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	100724	11/07/2024	10/14/2024 / SAM	10/09/2024 / SAM	V14546

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	100724	11/07/2024	10/14/2024 / SAM	10/09/2024 / SAM	V14547

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	91980 / Acrolin Std (Min = 5)	100724	11/07/2024	10/14/2024 / SAM	10/09/2024 / SAM	V14548

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112



CERTIFIED WEIGHT REPORT

Part Number: **95317**
Lot Number: **021624**
Description: **Universal VOA Megamix**
69 components
Expiration Date: 021627
Recommended Storage: Freezer (0 °C)
Nominal Concentration (µg/mL): 2000
NIST Test ID#: 8UTB

Solvent(s): **Methanol**
Lot#: **EG359-USQ12**

Weight(s) shown below were combined and diluted to (mL): **100.0** **0.021** Balance Uncertainty: **5E-05** Flask Uncertainty: **0.021**


Formulated By: <i>Prashant Chauhan</i>		021624
Reviewed By: <i>Pedro L. Rentes</i>		021624
		DATE

Compound	(RM#)	Lot	Di.	Initial	Initial	Nominal	Purity	Purity	Uncertainty	Target	Actual	Actual	Expanded	SDS information		
	Part Number	Number	Factor	Vol. (mL)	Conc.(µg/mL)	Conc (µg/mL)	(%)	Uncertainty	Pipette (mL)	Weight(g)	Weight(g)	Conc (µg/mL)	Uncertainty	(Solvent Safety Info. On Attached pg.)		
													(+/-)	CAS#	OSHA PEL (TWA)	LD50
Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2400mg/kg
Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg
Carbon disulfide	(0060)	MKCR8581	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg
cis-1,4-Dichloro-2-butene	(1198)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21058	0.21069	2001.1	8.5	1476-11-5	NA	NA
trans-1,4-Dichloro-2-butene	(0486)	MKBP6041F	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20748	2001.7	8.4	110-57-6	NA	NA
Diethyl ether	(0153)	IK18CA5000K	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	NA	NA
Ethyl methacrylate	(0381)	06126PK	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	NA	or-rat 14800mg/kg
Iodomethane	(0489)	5HBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm (28mg/m3/8H) (skin)	or-rat 70mg/kg
2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 3480mg/kg
Methacrylonitrile	(0442)	03427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	126-98-7	1 ppm (3mg/m3/8H) (skin)	or-rat 1200mg/kg
Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm (35mg/m3/8H) (skin)	or-rat 277mg/kg
Methyl methacrylate	(0404)	MKBW5137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 7872mg/kg
Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H) (skin)	or-rat 700mg/kg
2-Nitropropane	(0461)	14002JX	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	79-46-9	10 ppm (35mg/m3/8H)	or-rat 7200mg/kg
Pentachloroethane	(0450)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	76-01-7	NA	NA
1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7800mg/m3/8H)	or-rat 43g/kg
Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	NA	or-rat 910mg/kg
Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-49-1	NA	or-rat 840mg/kg
cis-1,2-Dichloroethane	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	NA	NA
trans-1,2-Dichloroethane	35171	101623	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	NA	NA
Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 1235mg/kg
1,1-Dichloroethene	32251	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg
Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 930mg/kg
Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg
Chloroform	95321	020724	0.10	10.00	20002.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	67-68-3	50 ppm (240mg/m3) (CL)	or-rat 900mg/kg
Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-85-3	NA	or-rat 105mg/kg
1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	or-rat 725mg/kg
2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	NA	NA
Tetrachloroethane	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.6	20.6	127-18-4	25 ppm (170mg/m3/8H) (final)	or-rat 2629mg/kg
1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10300mg/kg
1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40016.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	96-12-6	0.001 ppm	or-rat 170mg/kg
1,2-Dibromomethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	or-rat 108mg/kg
1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-08-2	50 ppm (8H)	or-rat 870mg/kg
1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2000.2	22.9	78-57-5	75 ppm (350mg/m3/8H)	or-rat 1947mg/kg
1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	NA	un-rat 3600mg/kg
1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	22.9	563-56-6	NA	NA
cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	NA	NA
trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.8	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	NA	NA
Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	22.9	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg
1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	630-20-6	NA	or-rat 670mg/kg
1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H) (skin)	or-rat 800mg/kg
1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (46mg/m3/8H) (skin)	or-rat 830mg/kg
Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	50 ppm (270mg/m3/8H)	or-rat 2400mg/kg
1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (50mg/m3/8H)	or-rat 149.6mg/kg
Benzene	35162	050823	0.05	5.00	40005.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	or-rat 4694mg/kg
Bromobenzene	35162	050823	0.05	5.00	40006.9	2000	NA	NA	0.017	NA	NA	1999.7	22.9	104-51-8	NA	NA
n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg
Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	99-87-6	NA	or-rat 4750mg/kg
p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg
Naphthalene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	or-rat 5000mg/kg
Styrene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	106-98-3	200 ppm	or-rat 5000mg/kg
Toluene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-61-6	NA	or-rat 1390mg/kg
1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 750mg/kg
1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.8	23.0	95-63-6	NA	or-rat 5g/kg
1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.6	2000	NA	NA	0.017	NA	NA	1999.8	22.9	106-57-8	NA	or-rat 5000mg/kg
1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg
m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	98-06-6	NA	NA
tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.6	22.9	135-98-8	NA	or-rat 2240mg/kg
sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2290mg/kg
Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	50 ppm (250mg/m3/8H)	or-rat 3900mg/kg
Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	NA	or-rat 2100mg/kg
Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	541-73-1	50 ppm (300mg/m3) (CL)	or-rat 500mg/kg
2-Dichlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.6	23.0	95-50-1	75 ppm (450mg/m3/8H)	or-rat 900mg/kg
3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	96-52-6	50 ppm (245mg/m3/8H)	or-rat 1400mg/kg
1,4-Dichlorobenzene	35163	101923	0.05	5.00												

Run 16, "P95317 L021624 [2000µg/mL in MeOH]"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments
GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.33
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.00
5	Iodomethane	12.31
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethane	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.24
11	cis-1,2-Dichloroethane	18.00
12	Methacrylonitrile/methyl acrylate/Chloroform	18.46
13	Isobutane/1,1,1-Trichloroethane	19.01
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.49
17	Trichloroethane	23.58
18	1,2-Dichloropropane	24.38
19	Methyl methacrylate	24.52
20	Bromochloromethane	25.13
21	Dibromomethane/2-Hitropropane	25.46
22	cis-1,3-Dichloropropane	27.02
23	Toluene	28.05
24	Ethyl methacrylate/trans-1,3-Dichloropropane	28.73
25	1,1,2-Trichloroethane	29.34
26	Tetrachloroethene/1,3-Dichloropropane	30.24
27	Dibromochloromethane	31.16
28	1,2-Dibromomethane	31.84
29	Chlorobenzene	33.26
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.40
31	m-Xylene/p-Xylene	33.66
32	o-Xylene	35.22
33	Styrene	35.39
34	Isopropylbenzene/Bromoforn	36.18
35	cis-1,4-Dichloro-3-butene	36.48
36	1,1,2,2-Tetrachloroethane	37.23
37	1,2,3-Trichloropropane	37.77
38	n-Propylbenzene	37.93
39	trans-1,4-Dichloro-3-butene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.80
42	2-Chlorotoluene	38.82
43	4-Chlorotoluene	38.77
44	tert-Butylbenzene	39.74
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropyltoluene	41.02
49	1,3-Dichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.53
52	1,2-Dichlorobenzene	43.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.46
55	1,2,4-Trichlorobenzene	49.26
56	Hexachlorobutadiene	49.72
57	Naphthalene	50.26
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2023

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.	
Personal protective equipment	Respiratory protection. Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC. DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT

Part Number: **95317**
Lot Number: **021624**
Description: **Universal VOA Megamix**
69 components
Expiration Date: 021627
Recommended Storage: Freezer (0 °C)
Nominal Concentration (µg/mL): 2000
NIST Test ID#: 8UTB

Solvent(s): **Methanol**
Lot#: **EG359-USQ12**

Weight(s) shown below were combined and diluted to (mL): **100.0** **0.021** Balance Uncertainty: **5E-05** Flask Uncertainty: **0.021**


Formulated By: <i>Prashant Chauhan</i>		021624
Reviewed By: <i>Pedro L. Rentes</i>		021624
DATE		DATE

Compound	(RM#)	Lot	Di.	Initial	Initial	Nominal	Purity	Purity	Uncertainty	Target	Actual	Actual	Expanded	SDS information		
	Part Number	Number	Factor	Vol. (mL)	Conc.(µg/mL)	Conc (µg/mL)	(%)	Uncertainty	Pipette (mL)	Weight(g)	Weight(g)	Conc (µg/mL)	Uncertainty	(Solvent Safety Info. On Attached pg.)		
													(+/-)	CAS#	OSHA PEL (TWA)	LD50
Acetonitrile	(0324)	021644	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20020	2001.3	8.1	75-05-8	40 ppm (70mg/m3/8H)	or-rat 2400mg/kg
Allyl chloride (3-Chloropropene)	(0325)	102396	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	107-05-1	1 ppm (3mg/m3/8H)	or-rat 700mg/kg
Carbon disulfide	(0060)	MKCR8581	NA	NA	NA	2000	99.99	0.2	NA	0.20007	0.20023	2001.6	8.1	75-15-0	4 ppm (12mg/m3) (skin)	or-rat 1200mg/kg
cis-1,4-Dichloro-2-butene	(1198)	14718EF	NA	NA	NA	2000	95	0.2	NA	0.21058	0.21069	2001.1	8.5	1476-11-5	NA	NA
trans-1,4-Dichloro-2-butene	(0486)	MKBP6041F	NA	NA	NA	2000	96.5	0.2	NA	0.20731	0.20748	2001.7	8.4	110-57-6	NA	NA
Diethyl ether	(0153)	IK18CA5000K	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	60-29-7	NA	NA
Ethyl methacrylate	(0381)	06126PK	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20230	2002.3	8.2	97-63-2	NA	or-rat 14800mg/kg
Iodomethane	(0489)	5HBF8718V	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20121	2001.5	8.2	74-88-4	5 ppm (28mg/m3/8H) (skin)	or-rat 70mg/kg
2-Methyl-1-propanol	(0445)	15241EB	NA	NA	NA	2000	99.5	0.2	NA	0.20106	0.20120	2001.4	8.1	78-83-1	50 ppm (150mg/m3/8H)	or-rat 3480mg/kg
Methacrylonitrile	(0442)	00427ET	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20221	2001.4	8.2	126-98-7	1 ppm (3mg/m3/8H) (skin)	or-rat 1200mg/kg
Methyl acrylate	(1075)	SHBK0679	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20040	2001.5	8.1	96-33-3	10 ppm (35mg/m3/8H) (skin)	or-rat 277mg/kg
Methyl methacrylate	(0404)	MKBW5137V	NA	NA	NA	2000	99.9	0.2	NA	0.20025	0.20041	2001.6	8.1	80-62-6	100 ppm (410mg/m3/8H)	or-rat 7872mg/kg
Nitrobenzene	(0228)	01213TV	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20220	2001.3	8.2	98-95-3	1 ppm (5mg/m3/8H) (skin)	or-rat 780mg/kg
2-Nitropropane	(0461)	14002JX	NA	NA	NA	2000	97.3	0.2	NA	0.20560	0.20577	2001.6	8.3	79-46-9	10 ppm (35mg/m3/8H)	or-rat 7200mg/kg
Pentachloroethane	(0450)	HGA01	NA	NA	NA	2000	98	0.2	NA	0.20413	0.20430	2001.6	8.3	76-01-7	NA	NA
1,1,2-Trichlorotrifluoroethane	(0474)	18930	NA	NA	NA	2000	99	0.2	NA	0.20207	0.20225	2001.8	8.2	76-13-1	1000 ppm (7800mg/m3/8H)	or-rat 43g/kg
Bromodichloromethane	35171	101623	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-27-4	NA	or-rat 910mg/kg
Dibromochloromethane	35171	101623	0.05	5.00	40002.1	2000	NA	NA	0.017	NA	NA	1999.6	23.0	124-49-1	NA	or-rat 840mg/kg
cis-1,2-Dichloroethane	35171	101623	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	156-59-2	NA	NA
trans-1,2-Dichloroethane	35171	101623	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.6	23.0	156-60-5	NA	NA
Methylene chloride	35171	101623	0.05	5.00	40002.8	2000	NA	NA	0.017	NA	NA	1999.6	22.9	75-09-2	500 ppm	or-rat 1235mg/kg
1,1-Dichloroethene	32251	102023	0.10	10.00	20001.6	2000	NA	NA	0.042	NA	NA	1999.7	20.4	75-35-4	1 ppm (4mg/m3/8H)	or-rat 200mg/kg
Bromoform	95321	020724	0.10	10.00	20003.2	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-25-2	0.5 ppm (5mg/m3) (skin)	or-rat 930mg/kg
Carbon tetrachloride	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	56-23-5	2 ppm (12.6mg/m3/8H)	or-rat 2350mg/kg
Chloroform	95321	020724	0.10	10.00	20002.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	67-66-3	50 ppm (240mg/m3) (CL)	or-rat 900mg/kg
Dibromomethane	95321	020724	0.10	10.00	20002.9	2000	NA	NA	0.042	NA	NA	1999.8	20.5	74-85-3	NA	or-rat 105mg/kg
1,1-Dichloroethane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.5	75-34-3	100 ppm	or-rat 725mg/kg
2,2-Dichloropropane	95321	020724	0.10	10.00	20003.4	2000	NA	NA	0.042	NA	NA	1999.8	20.4	594-20-7	NA	NA
Tetrachloroethane	95321	020724	0.10	10.00	20201.1	2000	NA	NA	0.042	NA	NA	2019.6	20.6	127-18-4	25 ppm (170mg/m3/8H) (final)	or-rat 2629mg/kg
1,1,1-Trichloroethane	95321	020724	0.10	10.00	20003.0	2000	NA	NA	0.042	NA	NA	1999.8	20.5	71-55-6	350 ppm (1900mg/m3/8H)	or-rat 10300mg/kg
1,2-Dibromo-3-chloropropane	35161	112322	0.05	5.00	40016.5	2000	NA	NA	0.017	NA	NA	2000.3	22.9	96-12-6	0.001 ppm	or-rat 170mg/kg
1,2-Dibromomethane	35161	112322	0.05	5.00	40024.8	2000	NA	NA	0.017	NA	NA	2000.7	22.9	106-93-4	20 ppm (8H)	or-rat 108mg/kg
1,2-Dichloroethane	35161	112322	0.05	5.00	40018.0	2000	NA	NA	0.017	NA	NA	2000.4	22.9	107-08-2	50 ppm (8H)	or-rat 870mg/kg
1,2-Dichloropropane	35161	112322	0.05	5.00	40051.0	2000	NA	NA	0.017	NA	NA	2000.2	22.9	78-57-5	75 ppm (350mg/m3/8H)	or-rat 1947mg/kg
1,3-Dichloropropane	35161	112322	0.05	5.00	40005.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	142-28-9	NA	un-rat 3600mg/kg
1,1-Dichloropropene	35161	112322	0.05	5.00	40012.1	2000	NA	NA	0.017	NA	NA	2000.1	22.9	563-56-6	NA	NA
cis-1,3-Dichloropropene	35161	112322	0.05	5.00	40010.0	2000	NA	NA	0.017	NA	NA	2000.0	23.0	10061-01-5	NA	NA
trans-1,3-Dichloropropene	35161	112322	0.05	5.00	40017.8	2000	NA	NA	0.017	NA	NA	2000.4	23.0	10061-02-6	NA	NA
Hexachloro-1,3-butadiene	35161	112322	0.05	5.00	40021.9	2000	NA	NA	0.017	NA	NA	2000.6	22.9	87-68-3	0.02 ppm (0.24mg/m3/8H)	or-rat 82mg/kg
1,1,1,2-Tetrachloroethane	35161	112322	0.05	5.00	40011.9	2000	NA	NA	0.017	NA	NA	2000.1	22.9	630-20-6	NA	or-rat 670mg/kg
1,1,2,2-Tetrachloroethane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	79-34-5	5 ppm (35mg/m3/8H) (skin)	or-rat 800mg/kg
1,1,2-Trichloroethane	35161	112322	0.05	5.00	40006.6	2000	NA	NA	0.017	NA	NA	1999.8	23.0	79-00-5	10 ppm (46mg/m3/8H) (skin)	or-rat 836mg/kg
Trichloroethene	35161	112322	0.05	5.00	40029.0	2000	NA	NA	0.017	NA	NA	2000.9	22.9	79-01-6	80 ppm (270mg/m3/8H)	or-rat 2400mg/kg
1,2,3-Trichloropropane	35161	112322	0.05	5.00	40007.5	2000	NA	NA	0.017	NA	NA	1999.9	22.9	96-18-4	10 ppm (50mg/m3/8H)	or-rat 149.6mg/kg
Benzene	35162	050823	0.05	5.00	40005.0	2000	NA	NA	0.017	NA	NA	1999.7	22.9	71-43-2	1 ppm	or-rat 4694mg/kg
Bromobenzene	35162	050823	0.05	5.00	40006.9	2000	NA	NA	0.017	NA	NA	1999.8	22.9	104-51-8	NA	NA
n-Butyl benzene	35162	050823	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-41-4	100 ppm (435mg/m3/8H)	or-rat >2000mg/kg
Ethyl benzene	35162	050823	0.05	5.00	40004.8	2000	NA	NA	0.017	NA	NA	1999.7	22.9	99-87-6	NA	or-rat 4750mg/kg
p-Isopropyl toluene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	91-20-3	10 ppm (50mg/m3/8H)	or-rat 490mg/kg
Naphthalene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.7	22.9	100-42-5	100 ppm	or-rat 5000mg/kg
Styrene	35162	050823	0.05	5.00	40006.2	2000	NA	NA	0.017	NA	NA	1999.8	22.9	106-98-3	200 ppm	or-rat 5000mg/kg
Toluene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	87-61-6	NA	or-rat 1390mg/kg
1,2,3-Trichlorobenzene	35162	050823	0.05	5.00	40003.1	2000	NA	NA	0.017	NA	NA	1999.7	22.9	120-82-1	5 ppm (CL) (40mg/m3)	or-rat 750mg/kg
1,2,4-Trichlorobenzene	35162	050823	0.05	5.00	40006.8	2000	NA	NA	0.017	NA	NA	1999.8	23.0	95-63-6	NA	or-rat 5g/kg
1,2,4-Trimethylbenzene	35162	050823	0.05	5.00	40001.6	2000	NA	NA	0.017	NA	NA	1999.8	22.9	106-57-8	NA	or-rat 5000mg/kg
1,3,5-Trimethylbenzene	35162	050823	0.05	5.00	40006.7	2000	NA	NA	0.017	NA	NA	1999.8	22.9	108-38-3	100 ppm (435mg/m3/8H)	or-rat 5g/kg
m-Xylene	35162	050823	0.05	5.00	40005.8	2000	NA	NA	0.017	NA	NA	1999.8	22.9	98-06-6	NA	NA
tert-Butyl benzene	35163	101923	0.05	5.00	40001.2	2000	NA	NA	0.017	NA	NA	1999.6	22.9	135-98-8	NA	or-rat 2240mg/kg
sec-Butyl benzene	35163	101923	0.05	5.00	40002.4	2000	NA	NA	0.017	NA	NA	1999.7	22.9	108-90-7	75 ppm (350mg/m3/8H)	or-rat 2290mg/kg
Chlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.5	22.9	95-49-8	60 ppm (250mg/m3/8H)	or-rat 3900mg/kg
Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	106-43-4	NA	or-rat 2100mg/kg
Chlorotoluene	35163	101923	0.05	5.00	40003.3	2000	NA	NA	0.017	NA	NA	1999.7	22.9	541-73-1	50 ppm (300mg/m3) (CL)	or-rat 500mg/kg
2-Dichlorobenzene	35163	101923	0.05	5.00	40003.8	2000	NA	NA	0.017	NA	NA	1999.6	23.0	96-46-7	78 ppm (450mg/m3/8H)	or-rat 900mg/kg
3-Dichlorobenzene	35163	101923	0.05	5.00	40001.7	2000	NA	NA	0.017	NA	NA	1999.6	23.0	96-52-8	60 ppm (245mg/m3/8H)	or-rat 1400mg/kg
1,4-Dichlorobenzene	35163	101923	0.05	5.00												

Run 16, "P95317 L021624 [2000µg/mL in MeOH]"

Run Length: 60.00 min, 35998 points at 10 points/second.
Created: Sat, Feb 17, 2024 at 8:56:46 AM.
Sampled: Sequence "021624-GC5M1", Method "GC5-M1".
Analyzed using Method "GC5-M1".

Comments
GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydrogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Peak #	Name	FID RT (min.)
1	Ether	9.97
2	1,1,2-Trichloro-1,2,2-trifluoroethane	10.33
3	1,1-Dichloroethane	11.10
4	Acetonitrile	12.00
5	Iodomethane	12.31
6	Allyl chloride	12.56
7	Carbon disulfide/Methylene chloride	13.04
8	trans-1,2-Dichloroethane	14.07
9	1,1-Dichloroethane	15.74
10	2,2-Dichloropropane	17.24
11	cis-1,2-Dichloroethane	18.00
12	Methacrylonitrile/methyl acrylate/Chloroform	18.46
13	Isobutane/1,1,1-Trichloroethane	19.01
14	1,1-Dichloropropane	20.46
15	Carbon tetrachloride	20.79
16	Benzene/1,2-Dichloroethane	21.49
17	Trichloroethane	23.58
18	1,2-Dichloropropane	24.38
19	Methyl methacrylate	24.52
20	Bromochloromethane	25.13
21	Dibromomethane/2-Hydropropane	25.46
22	cis-1,3-Dichloropropane	27.02
23	Toluene	28.05
24	Ethyl methacrylate/trans-1,3-Dichloropropane	28.73
25	1,1,2-Trichloroethane	29.34
26	Tetrachloroethene/1,3-Dichloropropane	30.24
27	Dibromochloromethane	31.16
28	1,2-Dibromomethane	31.84
29	Chlorobenzene	33.26
30	Ethylbenzene/1,1,1,2-Tetrachloroethane	33.40
31	m-Xylene/p-Xylene	33.66
32	o-Xylene	35.22
33	Styrene	35.39
34	Isopropylbenzene/Bromoforn	36.18
35	cis-1,4-Dichloro-3-butene	36.48
36	1,1,2,2-Tetrachloroethane	37.23
37	1,2,3-Trichloropropane	37.77
38	n-Propylbenzene	37.93
39	trans-1,4-Dichloro-3-butene	38.05
40	Bromobenzene	38.14
41	1,3,5-Trimethylbenzene	38.80
42	2-Chlorotoluene	38.82
43	4-Chlorotoluene	38.77
44	tert-Butylbenzene	39.74
45	1,2,4-Trimethylbenzene	39.91
46	Pentachloroethane	40.17
47	sec-Butylbenzene	40.52
48	p-Isopropyltoluene	41.02
49	1,3-Dichlorobenzene	41.42
50	1,4-Dichlorobenzene	41.83
51	n-Butylbenzene	42.53
52	1,2-Dichlorobenzene	43.18
53	1,2-Dibromo-3-chloropropane	46.12
54	Nitrobenzene	46.46
55	1,2,4-Trichlorobenzene	49.26
56	Hexachlorobutadiene	49.72
57	Naphthalene	50.26
58	1,2,3-Trichlorobenzene	51.16

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2023

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))			% (optional)
Methanol	METHYL ALCOHOL	CAS#: 67-56-1	> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability	Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol	67-56-1 TWA 200 ppm
Skin notation	TWA 200 ppm
Potential for skin absorption, ingestion and inhalation.	
Personal protective equipment	Respiratory protection. Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.	

Section IX - Physical/Chemical Characteristics

Boiling Point	65°C	Specific Gravity (H ₂ O = 1)	0.79
Vapor Pressure (mm Hg)	96	Melting Point	-98°C
Vapor Density (AIR = 1)	1.11	Evaporation rate (Butyl Acetate = 1)	4.6
Solubility in Water	COMPLETE		
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	Vapours may form explosive mixture with air.
Conditions to avoid	Heat, flames, sparks, extreme temperature and sunlight.
Materials to avoid	Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids
Hazardous decomposition products formed under fire conditions.	- Carbon oxides

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - rat - 5,628 mg/kg
LC50 Inhalation - rat - 4 h - 64000 ppm
LD50 Dermal - rabbit - 15,800 mg/kg
Toxic if absorbed through skin. Causes skin irritation.
Eye damage/eye irritation
Toxic if inhaled. Causes respiratory tract irritation.
Toxic if swallowed.

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.

LC50 15,400 mg/l - 96 h
EC50 24,500.00 mg/l - 48 h
EC100 10,000.00 mg/l - 24 h

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
UN number: 1230 Class: 3 Packing group: II	UN number: 1230 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION

OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation., Toxic by ingestion, Toxic by skin absorption, Irritant
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC. DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT

Part Number:
Lot Number:
Description:

95319
032922
Revised Additions Mix

Solvent(s):
Methanol
Lot#
EC592-US

Expiration Date:
Recommended Storage:
Nominal Concentration (µg/mL):
NIST Test ID#:

11 components
032925
Refrigerate (4 °C)
Varied
6UTB


Weight(s) shown below were combined and diluted to (mL):

5E-05 Balance Uncertainty
0.012 Flask Uncertainty

Formulated By:	Prashant Chauhan	032922	DATE
Reviewed By:	Pedro L. Rentas	032922	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information	
										(Solvent Safety Info. On Attached pg.)	OSHA PEL (TWA)
1. Acrylonitrile	7	4718CK	10000	99	0.2	1.01015	1.01030	10001.5	40.5	107-13-1	N/A
2. 1-Chlorobutane	1072	MKCM5711	2000	99.99	0.2	0.20003	0.20020	2001.7	8.1	109-69-3	N/A
3. Cyclohexane	1023	28930	2000	99	0.2	0.20203	0.20215	2001.2	8.2	110-82-7	300 ppm (1050mg/m ³ /H)
4. Di-Isopropyl ether (DIPE)	987	00412MX	2000	99	0.2	0.20203	0.20215	2001.2	8.2	109-20-3	500 ppm (2100mg/m ³ /H)
5. 1,4-Dioxane	373	03853KE	40000	99	0.2	4.04060	4.04100	40004.0	161.9	123-91-1	25 ppm (90mg/m ³ /H)(skin)
6. Hexachloroethane	199	12604HBV	2000	99	0.2	0.20203	0.20213	2001.0	8.2	67-72-1	1 ppm (10mg/m ³ /H)(skin)
7. Methylcyclohexane	1627	08046KN	2000	99	0.2	0.20203	0.20215	2001.2	8.2	108-87-2	N/A
8. Methyl tert-butyl ether (MTBE)	209	02197LJ	2000	99.8	0.2	0.20041	0.20055	2001.4	8.1	1634-04-4	N/A
9. Propionitrile	349	1395468	20000	99	0.2	2.02030	2.02045	20001.5	81.0	107-12-0	N/A
10. Tetrahydrofuran	380	SHBH8330	10000	99.9	0.2	1.00105	1.00120	10001.5	40.1	109-99-9	20 ppm (590mg/m ³ /H)
11. 1,2,3,4-Tetramethylbenzene	491	AP01	2000	93	0.2	0.21506	0.21520	2001.3	8.7	488-23-3	N/A

TIC: 95319.D



Name	MSD RT (min.)
Methyl tert-butyl ether (MTBE)	13.56
Acrylonitrile	13.79
Di-isopropyl ether	15.44
Propionitrile	18.53
Tetrahydrofuran	20.17
Cyclohexane	20.58
1-Chlorobutane	20.83
Methylcyclohexane	24.84
1,4-Dioxane	26.84
Hexachloroethane	48.44
1,2,3,4-Tetramethylbenzene	51.62

Safety Data Sheet (SDS)
GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2023

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225	Highly Flammable Liquid and Vapor	H301, 311, 331	Toxic if swallowed, skin contact, inhaled
H370	Cause damage to organs	H351	Suspected of causing cancer
P271	Use in ventilated area	P280	Use gloves, eye protection/face shield
P302,332	If on skin, wash with soap and water	P305,351,338	If in eyes, remove contacts, rinse with water

Signal Word: DANGER

Section III - Composition

Components:	CAS#:	LD50 Oral - Rat	OSHA PEL	% (optional)
Methanol	67-56-1	2,769 mg/kg	200 ppm	> 99

See Certified Weight Report For Other Analytes Present At Trace Quantities.
INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice
If inhaled
If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
Wash with soap and water. Consult a physician.
Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability
Suitable extinguishing media
Protective equipment for fire
Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions
Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions
Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).
Clean up

Section VII. HANDLING AND STORAGE

Precautions for safe handling
Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.
Storage Conditions

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol
67-56-1 TWA 200 ppm
TWA 200 ppm
Skin notation
Potential for skin absorption, ingestion and inhalation.
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Section X. STABILITY AND REACTIVITY			
Boiling Point		65°C	Specific Gravity (H ₂ O = 1)
Vapor Pressure (mm Hg)		96	Melting Point
Vapor Density (AIR = 1)		1.11	Evaporation rate (Butyl Acetate = 1)
Solubility in Water		COMPLETE	
Appearance and Odor		CLEAR, COLORLESS LIQUID WITH CHARACTERISTIC PUNGENT ODOR.	

Section XI. TOXICOLOGICAL INFORMATION	
Chemical stability Stable under recommended storage conditions. Possibility of hazardous reactions Vapours may form explosive mixture with air. Conditions to avoid Heat, flames, sparks, extreme temperature and sunlight. Materials to avoid Acid chlorides, Acid anhydrides, Oxidizing agents, Alkali metals, Reducing agents, Acids Hazardous decomposition products formed under fire conditions. - Carbon oxides	

Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.	
LD50 Oral - rat - 5,628 mg/kg LC50 Inhalation - rat - 4 h - 6400 ppm LD50 Dermal - rabbit - 15,800 mg/kg Toxic if absorbed through skin. Causes skin irritation. Eye damage/eye irritation Toxic if inhaled. Causes respiratory tract irritation. Toxic if swallowed.	

Section XIII. DISPOSAL CONSIDERATIONS	
LC50 15,400 mg/l - 96 h EC50 24,500.00 mg/l - 48 h EC100 10,000.00 mg/l - 24 h	

Section XIV. TRANSPORT INFORMATION	
DOT (US)	IATA
UN number: 1280 Class: 3 Packing group: II	UN number: 1280 Class: 3 Packing group: II
Proper shipping name: Methanol	Proper shipping name: Methanol

Section XV. REGULATORY INFORMATION	
OSHA Hazards Flammable liquid, Target Organ Effect, Toxic by inhalation, Toxic by ingestion, Toxic by skin absorption, Irritant SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.	

Section XVI. Misc. INFORMATION	
The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other chemicals or substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.	



CERTIFIED WEIGHT REPORT

Part Number:
Lot Number:
Description:

95319
032922
Revised Additions Mix

Solvent(s):
Methanol

Lot#
EC592-US

11 components

Expiration Date:

Recommended Storage:

Nominal Concentration (µg/mL):

NIST Test ID#:


Weight(s) shown below were combined and diluted to (mL):

5E-05 Balance Uncertainty
0.012 Flask Uncertainty

Formulated By:	Prashant Chauhan	DATE	032922
Reviewed By:	Pedro L. Rentas	DATE	032922

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information	
										(Solvent Safety Info. On Attached pg.)	OSHA PEL (TWA)
1. Acrylonitrile	7	4718CK	10000	99	0.2	1.01015	1.01030	10001.5	40.5	107-13-1	N/A
2. 1-Chlorobutane	1072	MKCM5711	2000	99.99	0.2	0.20003	0.20020	2001.7	8.1	109-69-3	N/A
3. Cyclohexane	1023	28930	2000	99	0.2	0.20203	0.20215	2001.2	8.2	110-82-7	300 ppm (1050mg/m ³ /H)
4. Di-Isopropyl ether (DIPE)	987	00412MX	2000	99	0.2	0.20203	0.20215	2001.2	8.2	109-20-3	500 ppm (2100mg/m ³ /H)
5. 1,4-Dioxane	373	03853KE	40000	99	0.2	4.04060	4.04100	40004.0	161.9	123-91-1	25 ppm (90mg/m ³ /H)(skin)
6. Hexachloroethane	199	12604HBV	2000	99	0.2	0.20203	0.20213	2001.0	8.2	67-72-1	1 ppm (10mg/m ³ /H)(skin)
7. Methylcyclohexane	1627	08046KN	2000	99	0.2	0.20203	0.20215	2001.2	8.2	108-87-2	N/A
8. Methyl tert-butyl ether (MTBE)	209	02197JJ	2000	99.8	0.2	0.20041	0.20055	2001.4	8.1	1634-04-4	N/A
9. Propionitrile	349	1395468	20000	99	0.2	2.02030	2.02045	20001.5	81.0	107-12-0	N/A
10. Tetrahydrofuran	380	SHBH8330	10000	99.9	0.2	1.00105	1.00120	10001.5	40.1	109-99-9	20 ppm (590mg/m ³ /H)
11. 1,2,3,4-Tetramethylbenzene	491	AP01	2000	93	0.2	0.21506	0.21520	2001.3	8.7	488-23-3	N/A

TIC: 95319.D



Name	MSD RT (min.)
Methyl tert-butyl ether (MTBE)	13.56
Acrylonitrile	13.79
Di-isopropyl ether	15.44
Propionitrile	18.53
Tetrahydrofuran	20.17
Cyclohexane	20.58
1-Chlorobutane	20.83
Methylcyclohexane	24.84
1,4-Dioxane	26.84
Hexachloroethane	48.44
1,2,3,4-Tetramethylbenzene	51.62

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN METHANOL

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr. Hamden CT, 06514	Emergency Telephone International	1-352-323-3500
		Date Prepared/Revised	January 1, 2023

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

H225 H370 P271 P302,332	Highly Flammable Liquid and Vapor Cause damage to organs Use in ventilated area If on skin, wash with soap and water	H301, 311, 331 H351 P280 P305,351,338	Toxic if swallowed, skin contact, inhaled Suspected of causing cancer Use gloves, eye protection/face shield If in eyes, remove contacts, rinse with water
  	Signal Word: DANGER		

Section III - Composition

Components:	CAS#:	LD50 Oral - Rat	OSHA PEL	% (optional)
Methanol	67-56-1	2,769 mg/kg	200 ppm	> 99

See Certified Weight Report For Other Analytes Present At Trace Quantities.
INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice
If inhaled
If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
Wash with soap and water. Consult a physician.
Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Flammability
Suitable extinguishing media
Protective equipment for fire
Flammable in the presence of a source of ignition when the temperature is above the flash point. Keep away from heat/sparks/open flame/hot surface. No smoking.
Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Wear self contained breathing apparatus for fire fighting if necessary.

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions
Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions
Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).
Clean up

Section VII. HANDLING AND STORAGE

Precautions for safe handling
Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge.
Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.
Storage Conditions

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Methanol
67-56-1 TWA 200 ppm
TWA 200 ppm
Skin notation
Potential for skin absorption, ingestion and inhalation.
Personal protective equipment Respiratory protection Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.

Section IX - Physical/Chemical Characteristics

Section X. STABILITY AND REACTIVITY				
Chemical stability				
Possibility of hazardous reactions				
Vapours may form explosive mixture with air:				
Stable under recommended storage conditions.				
Conditions to avoid				
Materials to avoid				
Hazardous decomposition products formed under fire conditions. - Carbon oxides				
Section XI. TOXICOLOGICAL INFORMATION				
LD50 Oral - rat - 5,628 mg/kg				
LC50 Inhalation - rat - 4 h - 6400 ppm				
LD50 Dermal - rabbit - 15,800 mg/kg				
Toxic if absorbed through skin. Causes skin irritation.				
Eye damage/eye irritation				
Toxic if inhaled. Causes respiratory tract irritation.				
Toxic if swallowed.				
Section XII. ECOLOGICAL INFORMATION FOR REPORTABLE QUANTITY OF 5000 lbs.				
LC50				
15,400 mg/l - 96 h				
EC50				
24,500.00 mg/l - 48 h				
EC100				
10,000.00 mg/l - 24 h				
Section XIII. DISPOSAL CONSIDERATIONS				
Dispose with normal Laboratory Solvent Waste.				
Section XIV. TRANSPORT INFORMATION				
DOT (US)				
UN number: 1280 Class: 3 Packing group: II				
Proper shipping name: Methanol				
UN number: 1280 Class: 3 Packing group: II				
Proper shipping name: Methanol				
Section XV. REGULATORY INFORMATION				
OSHA Hazards				
Flammable liquid, Target Organ Effect, Toxic by inhalation, Toxic by ingestion, Toxic by skin absorption, Irritant				
SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.				
Section XVI. Misc. INFORMATION				

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other chemicals or substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC. DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 100724
Description: Acrolein

Solvent(s): Water
Lot# 072324Q


Formulated By: Lawrence Barry	100724 DATE
Reviewed By: Pedro L. Rentas	100724 DATE

Expiration Date: 11/07/24
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB
Weight(s) shown below were combined and diluted to (mL): 20.0
Balance Uncertainty: 5E-05
Flask Uncertainty: 0.002

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (++) (µg/mL)	CAS#	SDS Information	
											(Solvent Safety Info. On Attached pg.)	(OSHA PEL (TWA)) LD50

1. Acrolein	5	103755R02H	5000	97	0.5	0.10307	0.10329	5010.4	51.9	107-02-8	0.1 ppm	01-rat 46mg/kg

Method: GC/MS/MS-1. **Detector:** Mass Selective Detector (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Over Profile:** Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.)
Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas, **NOTE:** Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately.
Long term storage is not recommended. Please contact our technical department if further information is required.



- * The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- * Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- * Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- * All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- * Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN WATER

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271	Use in ventilated area	H315	Causes skin and eye irritation.
P302,332	If on skin, wash with soap and water	P280	Use gloves, eye protection/face shield
		P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#: 7732-18-5	% (optional)
Water		> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water	CAS#: 7732-18-5	TWA: 500 ppm
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H ₂ O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

Vapor Density (AIR = 1)	NA	Evaporation rate (Butyl Acetate = 1)	0°C
Solubility in Water	Completely miscible		NA
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA
Hazardous decomposition products - No data available	

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA
Causes skin irritation.	
Eye irritation	

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
Not dangerous goods	Not dangerous goods
Proper shipping name: Water	Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 100724
Description: Acrolein

Solvent(s): Water
Lot# 072324Q


Formulated By: Lawrence Barry	100724 DATE
Reviewed By: Pedro L. Rentas	100724 DATE

Expiration Date: 11/07/24
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB
Weight(s) shown below were combined and diluted to (mL): 20.0
Balance Uncertainty: 5E-05
Flask Uncertainty: 0.002

SDS Information
(Solvent Safety Info. On Attached pg.)

1. Acrolein 5 103755R02H 5000 97 0.5 0.10307 0.10329 5010.4 51.9 107-02-8 0.1 ppm oil-rat 46mg/kg

Method: GC/MS/MS-1. **Detector:** Mass Selective Detector (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.)
Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas, **NOTE:** Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately.
Long term storage is not recommended. Please contact our technical department if further information is required.



- * The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- * Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- * Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- * All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- * Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS) GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN WATER

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271	Use in ventilated area	H315	Causes skin and eye irritation.
P302,332	If on skin, wash with soap and water	P280	Use gloves, eye protection/face shield
		P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#: 7732-18-5	% (optional)
Water		> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water	CAS#: 7732-18-5	TWA: 500 ppm
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H ₂ O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

Vapor Density (AIR = 1)	NA	Evaporation rate (Butyl Acetate = 1)	0°C
Solubility in Water	NA		NA
Appearance and Odor	Completely miscible CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA
Hazardous decomposition products - No data available	

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA
Causes skin irritation.	
Eye irritation	

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
Not dangerous goods	Not dangerous goods
Proper shipping name: Water	Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT

Part Number: 91980
Lot Number: 100724
Description: Acrolein

Solvent(s): Water
Lot# 072324Q


Formulated By:	Lawrence Barry
Reviewed By:	Pedro L. Rentas
DATE	100724

Expiration Date: 11/07/24
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 5000
NIST Test ID#: 6UTB
Weight(s) shown below were combined and diluted to (mL): 20.0
Balance Uncertainty: 5E-05
Flask Uncertainty: 0.002

SDS Information
(Solvent Safety Info. On Attached pg.)

1. Acrolein 5 103755R02H 5000 97 0.5 0.10307 0.10329 5010.4 51.9 107-02-8 0.1 ppm oil-rat 46mg/kg

Method: GC/MS/MS-1. **Detector:** Mass Selective Detector (Scan mode). **Column:** Vocol (60m X 0.25mm ID X 1.5µm film thickness). **Oven Profile:** Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.)
Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Pedro Rentas, **NOTE:** Due to the instability of acrolein in solution, all solutions of acrolein, and any dilutions thereof, should be used immediately.
Long term storage is not recommended. Please contact our technical department if further information is required.



- * The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- * Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- * Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- * All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- * Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Safety Data Sheet (SDS)

GHS/OSHA Compliant

Section I Product and Company Identification

IDENTITY ANALYTICAL STANDARD DISSOLVED IN WATER

Manufacturer's Name	ABSOLUTE STANDARDS INC	Emergency Telephone USA & CANADA	1-800-535-5053
Address	44 Rossotto Dr.	Emergency Telephone International	1-352-323-3500
	Hamden CT, 06514	Date Prepared/Revised	January 1, 2024

Section II - Hazards Identification

GHS Classification in accordance with 29 CFR 1910 (OSHA HCS)

P271	Use in ventilated area	H315	Causes skin and eye irritation.
P302,332	If on skin, wash with soap and water	P280	Use gloves, eye protection/face shield
		P305,351,338	If in eyes, remove contacts, rinse with water



Signal Word: DANGER

Section III - Composition

Components (Specific Chemical Identity; Common Name(s))	CAS#: 7732-18-5	% (optional)
Water		> 97

See Certified Weight Report For Other Analytes Present At Trace Quantities.

INTENDED USE: REFERENCE MATERIAL

Section IV. FIRST AID MEASURES

General advice	Consult a physician. Show this safety data sheet to the doctor in attendance. Move to safe area.
If inhaled	If inhaled, move person into fresh air. If not breathing, give artificial respiration. Consult a physician.
In case of skin contact	Wash with soap and water. Consult a physician.
In case of eye contact	Rinse thoroughly with plenty of water for at least 15 minutes and consult a physician.
If swallowed	Do NOT induce vomiting. Rinse mouth with water. Consult a physician.

Section V. FIREFIGHTING MEASURES

Suitable extinguishing media	Use water spray, alcohol-resistant foam, dry chemical or carbon dioxide.
Protective equipment for fire	Wear self contained breathing apparatus for fire fighting if necessary.
Hazardous Decomposition products	Carbon oxides

Section VI. ACCIDENTAL RELEASE MEASURES

Personal precautions	Wear respiratory protection. Avoid breathing vapors, mist or gas. Ensure adequate ventilation. Remove all sources of ignition. Vapours accumulate to form explosive concentrations.
Environmental precautions	Prevent further leakage or spillage if safe to do so. Do not let product enter drains.
Clean up	Contain spillage, and then collect and place in container for disposal according to local regulations (see section 13).

Section VII. HANDLING AND STORAGE

Precautions for safe handling	Avoid contact with skin and eyes. Avoid inhalation of vapour or mist.
Storage Conditions	Use ventilation. Keep away from sources of ignition. No smoking. Prevent the build up of electrostatic charge. Keep container tightly closed in a dry and well-ventilated place. Containers which are opened must be carefully resealed and kept upright to prevent leakage.

Section VIII. EXPOSURE CONTROLS/PERSONAL PROTECTION

Water	CAS#: 7732-18-5	TWA: 500 ppm
Personal protective equipment	Respiratory protection	Handle with gloves. Gloves must be inspected prior to use. Eye protection.
Avoid contact with skin, eyes and clothing. Wash hands thoroughly after handling the product.		

Section IX - PHYSICAL/CHEMICAL CHARACTERISTICS

Boiling Point	100°C	Specific Gravity (H ₂ O = 1)	1
Vapor Pressure (mm Hg)		Melting Point	

Vapor Density (AIR = 1)	NA	Evaporation rate (Butyl Acetate = 1)	0°C
Solubility in Water	Completely miscible		NA
Appearance and Odor	CLEAR, COLORLESS LIQUID WITH SLIGHT CHEMICAL ODOR.		

Section X. STABILITY AND REACTIVITY

Chemical stability	Stable under recommended storage conditions.
Possibility of hazardous reactions	NA
Conditions to avoid	NA
Materials to avoid	NA
Hazardous decomposition products - No data available	

Section XI. TOXICOLOGICAL INFORMATION

LD50 Oral - Rat	NA
LC50 Inhalation - Rat	NA
LD50 Dermal - Guinea pig	NA
Causes skin irritation.	
Eye irritation	

Section XII. ECOLOGICAL INFORMATION

LC50	NA
EC50	NA

Section XIII. DISPOSAL CONSIDERATIONS

Dispose with normal Laboratory Solvent Waste.

Section XIV. TRANSPORT INFORMATION

DOT (US)	IATA
Not dangerous goods	Not dangerous goods
Proper shipping name: Water	Proper shipping name: Water

Section XV. REGULATORY INFORMATION

SARA 302: No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

Section XVI. Misc. INFORMATION

The information in this Material Safety Data Sheet meets the requirements of the United States Occupational Safety and Health Act and regulations promulgated thereunder (29 CFR 1910.1200 et. seq.) and Global Harmonized System (GHS). This document is intended only as a guide to the appropriate precautionary handling of the material by trained personnel, or supervised by a person trained in chemical handling. The user is responsible for determining the precautions and dangers of this chemical for his or her particular application. Depending on usage, protective clothing including eye and face guards and respirators must be used to avoid contact with material or breathing chemical vapors/fumes. Exposure to this product may have serious adverse health effects. This chemical may interact with other substances. Since the potential uses are so varied, ABSOLUTE STANDARDS INC. cannot warn of all the potential dangers of use or interaction with other chemicals or substances. ABSOLUTE STANDARDS INC. warrants that the chemical meets the specifications set forth on the label. ABSOLUTE STANDARDS INC DISCLAIMS ANY OTHER WARRANTIES, EXPRESSED OR IMPLIED WITH REGARD TO THE PRODUCT SUPPLIED HEREUNDER, ITS MERCHANTABILITY OR ITS FITNESS FOR A PARTICULAR APPLICATION. The user should recognize that this product can cause severe injury or death, especially if improperly handled or the known dangers of use are not heeded. READ ALL PRECAUTIONARY INFORMATION. As new documented general safety information becomes available, Absolute Standards Inc. will periodically revise this Material Safety Data Sheet. If you have any questions, please call Technical Service at 1-203-281-2917 for assistance.



CERTIFIED WEIGHT REPORT

Part Number: 95318
Lot Number: 121321
Description: 2-Chloroethyl vinyl ether

Solvent(s): Methanol
Lot# EA899-US


Expiration Date: 121324
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 10000
NIST Test ID#: 6UTB
Weight(s) shown below were combined and diluted to (mL): 30.0
5E-05 Balance Uncertainty
0.0003 Flask Uncertainty

Formulated By:	Benson Chan	121321
Reviewed By:	Pedro L. Rentes	121321
		DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (µg/mL)	(Solvent Safety Info. On Attached pg.)	CAS#	OSHA PEL (TWA)	LD50

1. 2-Chloroethyl vinyl ether	74	MKCD0033	10000	99	0.2	0.30320	0.30411	10030.2	40.7	110-75-8	N/A		or rat 250mg/kg

Method: GC/MSD-1.M. **Detector:** MSD. **Column:** (60m X 0.25mm X 1.5 µm). **Oven Profile:** Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min.,
Injector B Temp= 200°C, Detector B Temp. = 220°C. Analyst: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N., and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



CERTIFIED WEIGHT REPORT

Part Number: 95318
Lot Number: 121321
Description: 2-Chloroethyl vinyl ether

Solvent(s): Methanol
Lot# EA899-US


Expiration Date: 121324
Recommended Storage: Refrigerate (4 °C)
Nominal Concentration (µg/mL): 10000
NIST Test ID#: 6UTB
Weight(s) shown below were combined and diluted to (mL): 30.0
5E-05 Balance Uncertainty
0.0003 Flask Uncertainty

Formulated By:	Benson Chan	121321
Reviewed By:	Pedro L. Rentes	121321
DATE		

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (µg/mL)	(Solvent Safety Info. On Attached pg.)	CAS#	OSHA PEL (TWA)	LD50

1. 2-Chloroethyl vinyl ether 74 MKCD0033 10000 99 0.2 0.30320 0.30411 10030.2 40.7 110-75-8 N/A or-rat 250mg/kg

Method: GC/MSD-1.M. **Detector:** MSD. **Column:** (60m X 0.25mm X 1.5 µm). **Oven Profile:** Temp 1 = 35°C (Time 1=10min.), Temp 2 = 200°C (Time 2=8.75 min.), Rate = 4°C/min., **Injector B Temp=200°C, Detector B Temp.=220°C. Analyst:** Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N., and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30470 Lot No.: A0181905
Description : tert-Butanol Standard
tert-Butanol Std 50,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL Pkg Amt: > 1 mL
Expiration Date : February 28, 2025 Storage: 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)												
1	tert-Butanol (TBA) CAS # 75-65-0 Purity 99% (Lot SHBM7694)	50,126.0 µg/mL	<table><tr><td>+/-</td><td>293.4988</td><td>µg/mL</td><td>Gravimetric</td></tr><tr><td>+/-</td><td>1,073.7654</td><td>µg/mL</td><td>Unstressed</td></tr><tr><td>+/-</td><td>1,104.9494</td><td>µg/mL</td><td>Stressed</td></tr></table>	+/-	293.4988	µg/mL	Gravimetric	+/-	1,073.7654	µg/mL	Unstressed	+/-	1,104.9494	µg/mL	Stressed
+/-	293.4988	µg/mL	Gravimetric												
+/-	1,073.7654	µg/mL	Unstressed												
+/-	1,104.9494	µg/mL	Stressed												

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:


200°C

Det. Temp:

250°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 16-Feb-2022

Balance: B442140311

Marlene Cowan - Operations Tech I

Date Passed: 21-Feb-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k \sqrt{U_{gravimetric}^2 + U_{homogeneity}^2 + U_{storage\ stability}^2 + U_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30067 **Lot No.:** A0191805

Description : 4-Bromofluorobenzene Standard

4-Bromofluorobenzene Standard 2,500µg/mL, P&T Methanol,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	2,483.9 µg/mL	+/- 139.5488

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Alicia Leathers - Operation Technician I

Date Mixed: 17-Nov-2022

Balance Serial # B251644995


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 21-Nov-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard
Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30225 **Lot No.:** A0193071

Description : Bromochloromethane Standard

Bromochloromethane 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Bromochloromethane	74-97-5	00008541	99%	2,018.0 µg/mL	+/- 113.3890

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol

CAS # 67-56-1

Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 29-Dec-2022 Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 03-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0194279

Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2029 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00012554	99%	2,001.5 µg/mL	+/- 112.7231
2	Chloromethane (methyl chloride)	74-87-3	SHBK6571	99%	2,001.2 µg/mL	+/- 112.5863
3	Vinyl chloride	75-01-4	00015559	99%	2,001.4 µg/mL	+/- 112.6561
4	Bromomethane (methyl bromide)	74-83-9	101604	99%	2,006.4 µg/mL	+/- 112.8262
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,001.9 µg/mL	+/- 112.5897
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCL8411	99%	2,000.8 µg/mL	+/- 112.6473

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:


Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 03-Feb-2023

Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 07-Feb-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0194279

Description : 502.2 Calibration Mix #1
502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2029 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00012554	99%	2,001.5 µg/mL	+/- 112.7231
2	Chloromethane (methyl chloride)	74-87-3	SHBK6571	99%	2,001.2 µg/mL	+/- 112.5863
3	Vinyl chloride	75-01-4	00015559	99%	2,001.4 µg/mL	+/- 112.6561
4	Bromomethane (methyl bromide)	74-83-9	101604	99%	2,006.4 µg/mL	+/- 112.8262
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,001.9 µg/mL	+/- 112.5897
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCL8411	99%	2,000.8 µg/mL	+/- 112.6473

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:

helium-constant flow 2.0 mL/min.

Temp. Program:

40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


MSD

Split Vent:

Split ratio 10:1

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckar - Mix Technician

Date Mixed: 03-Feb-2023

Balance Serial # B707717271


Christie Mills - Operations Tech II - ARM QC

Date Passed: 07-Feb-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555582 **Lot No.:** A0196865

Description : Custom 8260A/B Surrogate Mix

Custom 8260A/B Surrogate Mix 25,000µg/mL, P&T Methanol,
1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2-Dichloroethane-d4	17060-07-0	PR-32845	99%	25,036.0 µg/mL	+/- 1,417.9179
2	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	25,132.0 µg/mL	+/- 1,423.3549
3	Dibromofluoromethane	1868-53-7	022013	99%	25,040.0 µg/mL	+/- 1,418.1445
4	Toluene-d8	2037-26-5	PR-33397	99%	25,028.0 µg/mL	+/- 1,417.4648

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician I

Date Mixed: 11-Apr-2023

Balance: 1127510105

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0200785

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2026 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBP8774	99%	5,018.5 µg/mL	+/- 173.4162
2	2-Butanone (MEK)	78-93-3	SHBL5543	99%	5,016.0 µg/mL	+/- 173.3298
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP4724	99%	5,010.7 µg/mL	+/- 173.1455
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,015.0 µg/mL	+/- 173.2952

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)

CAS # 67-56-1/7732-18-5

Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 09-Aug-2023

Balance Serial # B707717271


Marlina Cowan - Operations Tech II ARM QC

Date Passed: 16-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0200785

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2026 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBP8774	99%	5,018.5 µg/mL	+/- 173.4162
2	2-Butanone (MEK)	78-93-3	SHBL5543	99%	5,016.0 µg/mL	+/- 173.3298
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP4724	99%	5,010.7 µg/mL	+/- 173.1455
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,015.0 µg/mL	+/- 173.2952

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)

CAS # 67-56-1/7732-18-5

Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 09-Aug-2023

Balance Serial # B707717271


Marlina Cowan - Operations Tech II ARM QC

Date Passed: 16-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0200785

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2026 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBP8774	99%	5,018.5 µg/mL	+/- 173.4162
2	2-Butanone (MEK)	78-93-3	SHBL5543	99%	5,016.0 µg/mL	+/- 173.3298
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP4724	99%	5,010.7 µg/mL	+/- 173.1455
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,015.0 µg/mL	+/- 173.2952

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:


40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 09-Aug-2023

Balance Serial # B707717271


Marlina Cowan - Operations Tech II ARM QC

Date Passed: 16-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

12
CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555408-SL **Lot No.:** A0205179
Description : Custom Vinyl Acetate Standard
Custom Vinyl Acetate Standard 8,000µg/mL, P&T Methanol, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2025 **Storage:** -20°C or colder
Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Vinyl acetate	108-05-4	RP231030CTH	98%	8,075.2 µg/mL	+/- 279.1159

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Daniel Wasson - Operations Tech I

Date Mixed: 06-Dec-2023

Balance Serial # 1127510105


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 11-Dec-2023

APPROVED
By: Jennifer Pollino (Date: 11-Dec-2023)

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024

Balance Serial # B707717271

Dillon Murphy
Dillon Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30489 **Lot No.:** A0209618

Description : 8260B Acetates Mix

8260B Acetates Mix 2,000 µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** -20°C or colder

Handling: This product is photosensitive. **Ship:** On Ice

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Methyl acetate	79-20-9	SHBP3100	99%	2,019.3 µg/mL	+/- 69.7974
2	Vinyl acetate	108-05-4	RP231030CTH	98%	2,016.8 µg/mL	+/- 69.7112
3	Ethyl acetate	141-78-6	SHBQ9682	99%	2,010.7 µg/mL	+/- 69.4979
4	Isopropyl acetate	108-21-4	BCCG7069	99%	2,016.0 µg/mL	+/- 69.6822
5	Propyl acetate	109-60-4	P8XLN	99%	2,008.0 µg/mL	+/- 69.4057
6	Butyl acetate	123-86-4	SHBP6314	99%	2,007.3 µg/mL	+/- 69.3826
7	Amyl acetate	628-63-7	41325/1	97%	2,004.7 µg/mL	+/- 69.2905

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Tech Tips:

Vinyl acetate is a volatile organic ester included in the target lists of several US EPA and other methods. Under acidic conditions, esters react with alcohols to form new esters (transesterification). Methanol-based mixes containing halogenated compounds are slightly acidic, so it is important to minimize exposure of vinyl acetate to mixes of halogenated compounds in methanol. For this

reason, we offer vinyl acetate in individual solution, and suggest that it be introduced into the working level calibration solution immediately before use. This will minimize problems and ensure more consistent results.

Quality Confirmation Test

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:

200°C

Det. Temp:

250°C

Det. Type:


FID

Split Vent:

40 ml/min

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 28-Mar-2024

Balance Serial # B707717271

Dillon Murphy
Dillon Murphy - Operations Technician I

Date Passed: 01-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 555581 **Lot No.:** A0210184

Description: Custom 8260 Internal Standard Mix

Custom 8260 Internal Standard Mix 25,000µg/mL, P&T Methanol, 1mL/ampul

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: April 30, 2027 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	25,212.0 µg/mL	+/- 1,427.8857
2	1,4-Difluorobenzene	540-36-3	MKCS8657	99%	25,220.0 µg/mL	+/- 1,428.3388
3	Chlorobenzene-d5	3114-55-4	PR-31132	99%	25,116.0 µg/mL	+/- 1,422.4487
4	Pentafluorobenzene	363-72-4	MKCR9383	99%	25,180.0 µg/mL	+/- 1,426.0734

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

John Friedline - Operations Technician I

Date Mixed: 11-Apr-2024

Balance: 11275.10105



Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{U_{gravimetric}^2 + U_{homogeneity}^2 + U_{storage\ stability}^2 + U_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (µeq/g)	≤ 0.3	0.2
Titration Base (µeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (µeq/g)	≤ 0.3	0.2
Titration Base (µeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis



Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH ₃ OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (µeq/g)	≤ 0.3	0.2
Titration Base (µeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality



SHIPPING DOCUMENTS

CHEMTECH

CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092

(908) 789-8900 Fax: (908) 78-8922

www.chemtech.net

Chemtech Project Number:

P4440/41

COC Number:

CLIENT INFORMATION

COMPANY: Tetra Tech

ADDRESS: 4433 Corporation Lane Suite 300

CITY: Virginia Beach STATE: VA ZIP: 23462

ATTENTION: Ernie Wu

PHONE: 757-466-4901

FAX: 757-461-4148

PROJECT INFORMATION

PROJECT NAME: NWIRP Bethpage

PROJECT #: 112G08005-WE13 LOCATION: VPB-196

PROJECT MANAGER: Ernie Wu

E-MAIL: ernie.wu@tetratech.com

PHONE: 757-466-4901

FAX: 757-461-4148

BILLING INFORMATION

BILL TO: SEE CONTRACT

PO#

ADDRESS:

CITY:

STATE: ZIP:

ATTENTION:

PHONE:

DATA TURNAROUND INFORMATION

FAX: 2 & 10 DAYS*

HARD COPY: 2 & 10 DAYS*

EDD 2 & 10 DAYS*

* TO BE APPROVED BY CHEMTECH

STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

☐ RESULTS ONLY☐ USEPA CLP☐ RESULTS + QC☐ New York State ASP "B"☐ New Jersey REDUCED☐ New York State ASP "A"☐ New Jersey CLP☐ Other☐ EDD Format

ANALYSIS

VOC(SW846-8260B)

1,4 Dioxane (8270 SIM)

1,4 Dioxane (522)

1

2

3

4

5

6

7

8

9

PRESERVATIVES

COMMENTS

CHEMTECH
SAMPLE
IDPROJECT
SAMPLE IDENTIFICATIONSAMPLE
MATRIXSAMPLE
TYPE

COMP

GRAB

SAMPLE
COLLECTION

DATE

TIME

of Bottles

A

1

2

3

4

5

6

7

8

9

<- Specify Preservatives
A-HCl B-HNO3
C-H2SO4 D-NaOH
E-ICE F-Other

1.

BPOW6-11-HYD-20241016

AQ

X

10/16/24

10:00

5

2

1

2

2.

BPOW6-11-TB-20241016

QA

X

10/16/24

9:00

2

2

3.

4.

5.

6.

7.

8.

9.


10.

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER

DATE/TIME

RECEIVED BY

1. 

10/17/24 1700

1. 

RELINQUISHED BY

DATE/TIME

RECEIVED BY

2.

9:15

2. 

10-18-24

RELINQUISHED BY

DATE/TIME

RECEIVED FOR LAB BY

3.

Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant ☐ Cooler Temp 2.3°

MeOH extraction requires an additional 4oz. Jar for percent solid

☐ Ice in Cooler? 

Comments: Standard TAT

Page 1 of 1

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ OvernightCHEMTECH: ☐ Picked Up ☐ Overnight

Shipment Complete

☐ YES ☐ NO

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT

YELLOW - CHEMTECH COPY

PINK - SAMPLER COPY



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (L-A-B)	L2219
Maine	2024021
Maryland	296
New Hampshire	255423
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488



234 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : P4440	TETR06	Order Date : 10/18/2024 10:22:00 AM	Project Mgr :
Client Name : Tetra Tech NUS, Inc.		Project Name : CTO WE13	Report Type : Level 4
Client Contact : Ernie Wu		Receive DateTime : 10/18/2024 9:15:00 AM	EDD Type : ADAPT
Invoice Name : Tetra Tech NUS, Inc.		Purchase Order :	Hard Copy Date :
Invoice Contact : Ernie Wu			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
P4440-01	BPOW6-11-HYD-20241016	Water	10/16/2024	10:00					
					VOCMS Group1		8260-Low	10 Bus. Days	
P4440-02	BPOW6-11-TB-20241016	Water	10/16/2024	09:00					
					VOCMS Group1		8260-Low	10 Bus. Days	

Relinquished By :

Date / Time :


10/16/24 11:20

Received By :

Date / Time :


10/16/27 11:20 Ref # 4

Storage Area : VOA Refridgerator Room

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084213.D
 Acq On : 30 Sep 2024 12:25
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Quant Time: Oct 01 06:48:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	192402	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	324582	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	288467	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	142880	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	259018	90.797	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 181.600%#	
35) Dibromofluoromethane	8.165	113	200166	93.542	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 187.080%#	
50) Toluene-d8	10.565	98	772146	98.079	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 196.160%#	
62) 4-Bromofluorobenzene	12.847	95	289867	101.065	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 202.140%#	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.124	85	218015	95.796	ug/l	94
3) Chloromethane	2.359	50	238104	91.628	ug/l	98
4) Vinyl Chloride	2.512	62	234942	93.338	ug/l	99
5) Bromomethane	2.936	94	141604	85.288	ug/l	94
6) Chloroethane	3.112	64	144394	80.099	ug/l	92
7) Trichlorofluoromethane	3.495	101	366907	93.608	ug/l	98
8) Diethyl Ether	3.953	74	136436	93.854	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.365	101	208443	92.811	ug/l	99
10) Methyl Iodide	4.589	142	281269	97.248	ug/l	97
11) Tert butyl alcohol	5.518	59	194750	414.018	ug/l	99
12) 1,1-Dichloroethene	4.336	96	206863	95.491	ug/l	95
13) Acrolein	4.177	56	251811	466.304	ug/l	99
14) Allyl chloride	5.018	41	348108	92.526	ug/l	99
15) Acrylonitrile	5.712	53	533270	448.931	ug/l	99
16) Acetone	4.424	43	576139	470.138	ug/l	99
17) Carbon Disulfide	4.712	76	611042	89.090	ug/l	100
18) Methyl Acetate	5.018	43	233708	87.499	ug/l	99
19) Methyl tert-butyl Ether	5.794	73	702366	96.061	ug/l	99
20) Methylene Chloride	5.277	84	228706	91.809	ug/l	91
21) trans-1,2-Dichloroethene	5.783	96	213704	94.160	ug/l	94
22) Diisopropyl ether	6.671	45	745614	95.811	ug/l	98
23) Vinyl Acetate	6.600	43	2895429	498.979	ug/l	99
24) 1,1-Dichloroethane	6.565	63	413523	95.021	ug/l	99
25) 2-Butanone	7.483	43	759354	454.946	ug/l	100
26) 2,2-Dichloropropane	7.488	77	378893	96.560	ug/l	99
27) cis-1,2-Dichloroethene	7.483	96	257696	93.962	ug/l	99
28) Bromochloromethane	7.812	49	181546	93.497	ug/l	100
29) Tetrahydrofuran	7.836	42	453342	440.124	ug/l	99
30) Chloroform	7.965	83	416595	92.156	ug/l	100
31) Cyclohexane	8.253	56	373445	88.380	ug/l	99
32) 1,1,1-Trichloroethane	8.165	97	383566	94.476	ug/l	98
36) 1,1-Dichloropropene	8.371	75	307616	98.969	ug/l	99
37) Ethyl Acetate	7.559	43	293597	92.073	ug/l	100
38) Carbon Tetrachloride	8.359	117	334398	98.077	ug/l	98
39) Methylcyclohexane	9.600	83	368299	106.466	ug/l	99
40) Benzene	8.606	78	931076	96.145	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084213.D
 Acq On : 30 Sep 2024 12:25
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Quant Time: Oct 01 06:48:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	160270	92.581	ug/l	97
42) 1,2-Dichloroethane	8.671	62	311563	94.900	ug/l	100
43) Isopropyl Acetate	8.688	43	521814	83.581	ug/l	100
44) Trichloroethene	9.347	130	216678	95.798	ug/l	97
45) 1,2-Dichloropropane	9.618	63	224724	98.095	ug/l	97
46) Dibromomethane	9.706	93	154739	100.091	ug/l	99
47) Bromodichloromethane	9.888	83	338406	98.526	ug/l	99
48) Methyl methacrylate	9.677	41	244061	98.163	ug/l	99
49) 1,4-Dioxane	9.694	88	82854	1810.110	ug/l	99
51) 4-Methyl-2-Pentanone	10.441	43	1456154	479.699	ug/l	100
52) Toluene	10.629	92	586544	99.324	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	365099	104.280	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	384030	101.975	ug/l	98
55) 1,1,2-Trichloroethane	11.018	97	205711	97.170	ug/l	97
56) Ethyl methacrylate	10.871	69	360916	103.608	ug/l	99
57) 1,3-Dichloropropane	11.165	76	360601	95.326	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	870416	538.640	ug/l	99
59) 2-Hexanone	11.194	43	1105736	487.363	ug/l	99
60) Dibromochloromethane	11.359	129	249435	99.721	ug/l	98
61) 1,2-Dibromoethane	11.465	107	212705	96.694	ug/l	100
64) Tetrachloroethene	11.100	164	186069	92.608	ug/l	99
65) Chlorobenzene	11.888	112	616423	96.376	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.959	131	212695	96.546	ug/l	99
67) Ethyl Benzene	11.959	91	1146786	101.385	ug/l	98
68) m/p-Xylenes	12.071	106	855381	204.437	ug/l	100
69) o-Xylene	12.394	106	412098	104.101	ug/l	99
70) Styrene	12.412	104	711855	106.165	ug/l	99
71) Bromoform	12.576	173	162862	100.412	ug/l #	97
73) Isopropylbenzene	12.694	105	1067798	97.936	ug/l	100
74) N-amyl acetate	12.494	43	467641	94.716	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.935	83	286161	87.285	ug/l	98
76) 1,2,3-Trichloropropane	12.994	75	399963	140.449	ug/l	64
77) Bromobenzene	12.982	156	243075	92.425	ug/l	100
78) n-propylbenzene	13.035	91	1285399	101.754	ug/l	99
79) 2-Chlorotoluene	13.123	91	770341	95.567	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	905999	101.869	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	116289	95.579	ug/l	98
82) 4-Chlorotoluene	13.218	91	790045	97.362	ug/l	100
83) tert-Butylbenzene	13.435	119	785331	99.349	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	913037	101.908	ug/l	99
85) sec-Butylbenzene	13.612	105	1082055	102.580	ug/l	100
86) p-Isopropyltoluene	13.729	119	918730	105.328	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	463998	94.046	ug/l	100
88) 1,4-Dichlorobenzene	13.812	146	465087	93.346	ug/l	99
89) n-Butylbenzene	14.053	91	832034	105.197	ug/l	99
90) Hexachloroethane	14.329	117	168712	95.289	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	444319	91.843	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	59837	88.446	ug/l	98
93) 1,2,4-Trichlorobenzene	15.388	180	245477	103.017	ug/l	99
94) Hexachlorobutadiene	15.500	225	105801	89.034	ug/l	98
95) Naphthalene	15.635	128	784876	99.356	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	233126	97.736	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084213.D
Acq On : 30 Sep 2024 12:25
Operator : JC\MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Quant Time: Oct 01 06:48:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084213.D
 Acq On : 30 Sep 2024 12:25
 Operator : JC\MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC100

Quant Time: Oct 01 06:48:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion:168 Resp: 192402
Ion Ratio Lower Upper
168 100
99 64.4 54.2 81.2



#2
Dichlorodifluoromethane
Concen: 95.796 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 85 Resp: 218015
Ion Ratio Lower Upper
85 100
87 31.6 17.6 52.9









#3
Chloromethane
Concen: 91.628 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 50 Resp: 238104
Ion Ratio Lower Upper
50 100
52 34.7 26.8 40.2



#4
Vinyl Chloride
Concen: 93.338 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.006 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 62 Resp: 234942
Ion Ratio Lower Upper
62 100
64 32.2 26.2 39.4









#5
Bromomethane
Concen: 85.288 ug/l
RT: 2.936 min Scan# 1
Delta R.T. -0.018 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 94 Resp: 141604
Ion Ratio Lower Upper
94 100
96 97.2 72.9 109.3



#6
Chloroethane
Concen: 80.099 ug/l
RT: 3.112 min Scan# 197
Delta R.T. -0.006 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 64 Resp: 144394
Ion Ratio Lower Upper
64 100
66 33.7 23.5 35.3









#7
 Trichlorofluoromethane
 Concen: 93.608 ug/l
 RT: 3.495 min Scan# 20
 Delta R.T. 0.000 min
 Lab File: VN084213.D
 Acq: 30 Sep 2024 12:25




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion:101 Resp: 366907
 Ion Ratio Lower Upper
 101 100
 103 66.3 51.8 77.6



#8
 Diethyl Ether
 Concen: 93.854 ug/l
 RT: 3.953 min Scan# 340
 Delta R.T. -0.006 min
 Lab File: VN084213.D
 Acq: 30 Sep 2024 12:25

Tgt Ion: 74 Resp: 136436
 Ion Ratio Lower Upper
 74 100
 45 96.5 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 92.811 ug/l
RT: 4.365 min Scan# 411
Delta R.T. -0.006 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:101 Resp: 208443
Ion Ratio Lower Upper
101 100
85 46.2 36.2 54.4
151 74.1 59.3 88.9



#10
Methyl Iodide
Concen: 97.248 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.006 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion:142 Resp: 281269
Ion Ratio Lower Upper
142 100
127 46.5 35.3 52.9
141 14.9 11.0 16.6









#11
Tert butyl alcohol
Concen: 414.018 ug/l
RT: 5.518 min Scan# 607
Delta R.T. -0.006 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 59 Resp: 194750
Ion Ratio Lower Upper
59 100
57 9.9 8.2 12.2



#12
1,1-Dichloroethene
Concen: 95.491 ug/l
RT: 4.336 min Scan# 405
Delta R.T. -0.006 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 96 Resp: 206863
Ion Ratio Lower Upper
96 100
61 169.2 128.6 192.8
98 63.7 51.7 77.5





#13

Acrolein

Concen: 466.304 ug/l

RT: 4.177 min Scan# 31

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion: 56 Resp: 251811

Ion Ratio Lower Upper

56 100

55 70.4 57.2 85.8



#14

Allyl chloride

Concen: 92.526 ug/l

RT: 5.018 min Scan# 521

Delta R.T. -0.006 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion: 41 Resp: 348108


Ion Ratio Lower Upper

41 100

39 73.7 59.0 88.6

76 36.4 29.8 44.6









#15
Acrylonitrile
Concen: 448.931 ug/l
RT: 5.712 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 53 Resp: 533270
Ion Ratio Lower Upper
53 100
52 81.8 66.2 99.4
51 35.6 28.8 43.2



#16
Acetone
Concen: 470.138 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 43 Resp: 576139
Ion Ratio Lower Upper
43 100
58 31.3 24.4 36.6





#17

Carbon Disulfide

Concen: 89.090 ug/l

RT: 4.712 min Scan# 4

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC100

Tgt Ion: 76 Resp: 611042

Ion Ratio Lower Upper

76 100

78 9.2 7.3 10.9



#18

Methyl Acetate

Concen: 87.499 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN084213.D




Acq: 30 Sep 2024 12:25


Tgt Ion: 43 Resp: 233708

Ion Ratio Lower Upper

43 100

74 23.4 19.2 28.8









#19
Methyl tert-butyl Ether
Concen: 96.061 ug/l
RT: 5.794 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 73 Resp: 702366
Ion Ratio Lower Upper
73 100
57 21.8 17.7 26.5



#20
Methylene Chloride
Concen: 91.809 ug/l
RT: 5.277 min Scan# 565
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 84 Resp: 228706
Ion Ratio Lower Upper
84 100
49 119.6 107.2 160.8
51 36.6 31.8 47.8
86 62.5 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 94.160 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 96 Resp: 213704
Ion Ratio Lower Upper
96 100
61 145.8 108.4 162.6
98 63.2 49.8 74.8



#22
Diisopropyl ether
Concen: 95.811 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 45 Resp: 745614
Ion Ratio Lower Upper
45 100
43 50.8 42.3 63.5
87 27.4 21.8 32.6
59 11.3 9.6 14.4









#23
 Vinyl Acetate
 Concen: 498.979 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. 0.000 min
 Lab File: VN084213.D
 Acq: 30 Sep 2024 12:25




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion: 43 Resp: 2895429
 Ion Ratio Lower Upper
 43 100
 86 10.4 8.6 12.8



#24
 1,1-Dichloroethane
 Concen: 95.021 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. -0.006 min
 Lab File: VN084213.D
 Acq: 30 Sep 2024 12:25

Tgt Ion: 63 Resp: 413523
 Ion Ratio Lower Upper
 63 100
 98 7.3 3.4 10.2
 100 4.4 2.0 5.8









#25
2-Butanone
Concen: 454.946 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 43 Resp: 759354
Ion Ratio Lower Upper
43 100
72 26.6 21.4 32.2



#26
2,2-Dichloropropane
Concen: 96.560 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 77 Resp: 378893
Ion Ratio Lower Upper
77 100
97 21.7 10.8 32.3









#27
 cis-1,2-Dichloroethene
 Concen: 93.962 ug/l
 RT: 7.483 min Scan# 941
 Delta R.T. -0.006 min
 Lab File: VN084213.D
 Acq: 30 Sep 2024 12:25

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100

Tgt Ion: 96 Resp: 257696




Ion	Ratio	Lower	Upper
96	100		
61	151.7	0.0	301.8
98	63.8	0.0	128.2




#28
 Bromochloromethane
 Concen: 93.497 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.000 min
 Lab File: VN084213.D
 Acq: 30 Sep 2024 12:25

Tgt Ion: 49 Resp: 181546

Ion	Ratio	Lower	Upper
49	100		
129	1.7	0.0	3.6
130	72.1	58.0	87.0





#29

Tetrahydrofuran

Concen: 440.124 ug/l

RT: 7.836 min Scan# 1000

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100





Tgt Ion: 42 Resp: 453342

Ion Ratio Lower Upper

42 100

72 44.0 34.6 51.8

71 41.2 32.2 48.4



#30

Chloroform

Concen: 92.156 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.000 min

Lab File: VN084213.D




Acq: 30 Sep 2024 12:25


Tgt Ion: 83 Resp: 416595

Ion Ratio Lower Upper

83 100

85 63.5 51.0 76.4









#31
Cyclohexane
Concen: 88.380 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 56 Resp: 373445
Ion Ratio Lower Upper
56 100
69 31.2 23.8 35.6
84 85.7 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 94.476 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.006 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 97 Resp: 383566
Ion Ratio Lower Upper
97 100
99 63.3 51.2 76.8
61 45.2 37.5 56.3





#33

1,2-Dichloroethane-d4

Concen: 90.797 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion: 65 Resp: 259018

Ion Ratio Lower Upper

65 100

67 51.9 0.0 102.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion: 114 Resp: 324582


Ion Ratio Lower Upper

114 100

63 22.0 0.0 43.8

88 17.2 0.0 31.6





#35

Dibromofluoromethane

Concen: 93.542 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100





Tgt Ion:113 Resp: 200166

Ion Ratio Lower Upper

113 100

111 101.6 83.3 124.9

192 17.8 13.5 20.3



#36

1,1-Dichloropropene

Concen: 98.969 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion: 75 Resp: 307616


Ion Ratio Lower Upper

75 100

110 33.5 16.9 50.7

77 30.5 24.7 37.1









#37
Ethyl Acetate
Concen: 92.073 ug/l
RT: 7.559 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 43 Resp: 293597
Ion Ratio Lower Upper
43 100
61 13.3 10.7 16.1
70 10.7 8.8 13.2



#38
Carbon Tetrachloride
Concen: 98.077 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 117 Resp: 334398
Ion Ratio Lower Upper
117 100
119 95.6 78.7 118.1
121 29.8 24.4 36.6









#39
Methylcyclohexane
Concen: 106.466 ug/l
RT: 9.600 min Scan# 11300
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 83 Resp: 368299
Ion Ratio Lower Upper
83 100
55 79.2 62.8 94.2
98 48.5 38.8 58.2



#40
Benzene
Concen: 96.145 ug/l
RT: 8.606 min Scan# 1131
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 78 Resp: 931076
Ion Ratio Lower Upper
78 100
77 23.7 19.1 28.7









#41
Methacrylonitrile
Concen: 92.581 ug/l
RT: 7.777 min Scan# 990
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100

Tgt Ion: 41 Resp: 160270




Ion	Ratio	Lower	Upper
41	100		
39	56.6	43.3	64.9
67	72.4	56.1	84.1
52	30.4	24.5	36.7




#42
1,2-Dichloroethane
Concen: 94.900 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 62 Resp: 311563

Ion	Ratio	Lower	Upper
62	100		
98	9.2	0.0	18.2









#43
Isopropyl Acetate
Concen: 83.581 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 43 Resp: 521814
Ion Ratio Lower Upper
43 100
61 25.4 20.4 30.6
87 12.6 10.3 15.5



#44
Trichloroethene
Concen: 95.798 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion:130 Resp: 216678
Ion Ratio Lower Upper
130 100
95 104.9 0.0 203.0





#45

1,2-Dichloropropane

Concen: 98.095 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion: 63 Resp: 224724

Ion Ratio Lower Upper

63 100

65 32.2 24.6 37.0



#46

Dibromomethane

Concen: 100.091 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion: 93 Resp: 154739


Ion Ratio Lower Upper

93 100

95 82.2 65.0 97.4

174 89.5 71.8 107.8









#47
Bromodichloromethane
Concen: 98.526 ug/l
RT: 9.888 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 83 Resp: 338406
Ion Ratio Lower Upper
83 100
85 64.1 51.5 77.3
127 8.4 5.7 8.5



#48
Methyl methacrylate
Concen: 98.163 ug/l
RT: 9.677 min Scan# 1313
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 41 Resp: 244061
Ion Ratio Lower Upper
41 100
69 81.9 67.1 100.7
39 57.1 45.8 68.6





#49

1,4-Dioxane

Concen: 1810.110 ug/l

RT: 9.694 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100





Tgt Ion: 88 Resp: 82854

Ion Ratio Lower Upper

88 100

43 30.8 25.3 37.9

58 70.9 56.4 84.6



#50

Toluene-d8

Concen: 98.079 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN084213.D




Acq: 30 Sep 2024 12:25


Tgt Ion: 98 Resp: 772146

Ion Ratio Lower Upper

98 100

100 65.5 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 479.699 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 43 Resp: 1456154
Ion Ratio Lower Upper
43 100
58 39.3 31.6 47.4



#52
Toluene
Concen: 99.324 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 92 Resp: 586544
Ion Ratio Lower Upper
92 100
91 171.0 137.3 205.9









#53
 t-1,3-Dichloropropene
 Concen: 104.280 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. 0.000 min
 Lab File: VN084213.D
 Acq: 30 Sep 2024 12:25




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion: 75 Resp: 365099
 Ion Ratio Lower Upper
 75 100
 77 32.3 25.4 38.0



#54
 cis-1,3-Dichloropropene
 Concen: 101.975 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. 0.000 min
 Lab File: VN084213.D
 Acq: 30 Sep 2024 12:25

Tgt Ion: 75 Resp: 384030
 Ion Ratio Lower Upper
 75 100
 77 32.5 24.2 36.2
 39 49.2 38.7 58.1





#55

1,1,2-Trichloroethane

Concen: 97.170 ug/l

RT: 11.018 min Scan# 1540

Delta R.T. 0.006 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100

Tgt Ion: 97 Resp: 205711





Ion Ratio Lower Upper

97 100

83 83.8 71.0 106.4

85 54.4 41.7 62.5

99 62.2 50.1 75.1



#56

Ethyl methacrylate

Concen: 103.608 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion: 69 Resp: 360916


Ion Ratio Lower Upper

69 100

41 67.4 53.3 79.9

39 36.1 28.8 43.2









#57
1,3-Dichloropropane
Concen: 95.326 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. 0.006 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 76 Resp: 360601
Ion Ratio Lower Upper
76 100
78 32.6 25.8 38.6



#58
2-Chloroethyl Vinyl ether
Concen: 538.640 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 63 Resp: 870416
Ion Ratio Lower Upper
63 100
106 27.5 21.8 32.6









#59
2-Hexanone
Concen: 487.363 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 43 Resp: 1105736
Ion Ratio Lower Upper
43 100
58 55.0 27.3 81.8



#60
Dibromochloromethane
Concen: 99.721 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion:129 Resp: 249435
Ion Ratio Lower Upper
129 100
127 78.1 38.3 114.9









#61
1,2-Dibromoethane
Concen: 96.694 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:107 Resp: 212705
Ion Ratio Lower Upper
107 100
109 94.0 75.4 113.0



#62
4-Bromofluorobenzene
Concen: 101.065 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 95 Resp: 289867
Ion Ratio Lower Upper
95 100
174 71.4 0.0 145.2
176 68.2 0.0 140.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:117 Resp: 288467
Ion Ratio Lower Upper
117 100
82 59.7 47.2 70.8
119 32.5 25.4 38.0



#64
Tetrachloroethene
Concen: 92.608 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion:164 Resp: 186069
Ion Ratio Lower Upper
164 100
166 127.4 101.0 151.4
129 102.4 80.8 121.2
131 95.7 78.1 117.1





#65

Chlorobenzene

Concen: 96.376 ug/l

RT: 11.888 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC100

Tgt Ion:112 Resp: 616423

Ion Ratio Lower Upper

112 100

114 32.0 27.0 40.6



#66

1,1,1,2-Tetrachloroethane

Concen: 96.546 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion:131 Resp: 212695


Ion Ratio Lower Upper

131 100

133 94.6 47.9 143.7

119 64.8 32.8 98.3









#67
Ethyl Benzene
Concen: 101.385 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 91 Resp: 1146786
Ion Ratio Lower Upper
91 100
106 29.5 24.5 36.7



#68
m/p-Xylenes
Concen: 204.437 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion:106 Resp: 855381
Ion Ratio Lower Upper
106 100
91 208.5 167.1 250.7









#69
o-Xylene
Concen: 104.101 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion:106 Resp: 412098
Ion Ratio Lower Upper
106 100
91 222.0 110.2 330.6



#70
Styrene
Concen: 106.165 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion:104 Resp: 711855
Ion Ratio Lower Upper
104 100
78 52.5 42.7 64.1
103 54.7 43.6 65.4









#71
Bromoform
Concen: 100.412 ug/l
RT: 12.576 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion:173 Resp: 162862
Ion Ratio Lower Upper
173 100
175 49.3 23.5 70.6
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion:152 Resp: 142880
Ion Ratio Lower Upper
152 100
115 63.0 31.3 93.9
150 188.1 0.0 349.8









#73
Isopropylbenzene
Concen: 97.936 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion:105 Resp: 1067798
Ion Ratio Lower Upper
105 100
120 25.9 12.9 38.7



#74
N-ethyl acetate
Concen: 94.716 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 43 Resp: 467641
Ion Ratio Lower Upper
43 100
70 42.8 34.4 51.6
55 24.2 19.6 29.4
61 24.6 19.6 29.4









#75
 1,1,2,2-Tetrachloroethane
 Concen: 87.285 ug/l
 RT: 12.935 min Scan# 1867
 Delta R.T. 0.000 min
 Lab File: VN084213.D
 Acq: 30 Sep 2024 12:25




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion: 83 Resp: 286161
 Ion Ratio Lower Upper
 83 100
 131 10.2 4.9 14.7
 85 65.7 31.9 95.7



#76
 1,2,3-Trichloropropane
 Concen: 140.449 ug/l
 RT: 12.994 min Scan# 1877
 Delta R.T. 0.006 min
 Lab File: VN084213.D
 Acq: 30 Sep 2024 12:25

Tgt Ion: 75 Resp: 399963
 Ion Ratio Lower Upper
 75 100
 77 139.8 96.9 290.7





#77

Bromobenzene

Concen: 92.425 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC100





Tgt Ion:156 Resp: 243075

Ion Ratio Lower Upper

156 100

77 230.0 115.1 345.3

158 98.3 48.9 146.8



#78

n-propylbenzene

Concen: 101.754 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. 0.000 min

Lab File: VN084213.D




Acq: 30 Sep 2024 12:25


Tgt Ion: 91 Resp: 1285399

Ion Ratio Lower Upper

91 100

120 21.9 10.8 32.4









#79
2-Chlorotoluene
Concen: 95.567 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion: 91 Resp: 770341
Ion Ratio Lower Upper
91 100
126 31.4 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 101.869 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 105 Resp: 905999
Ion Ratio Lower Upper
105 100
120 47.3 23.7 71.1









#81
trans-1,4-Dichloro-2-butene
Concen: 95.579 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 75 Resp: 116289
Ion Ratio Lower Upper
75 100
53 95.1 78.0 117.0
89 44.4 35.5 53.3



#82
4-Chlorotoluene
Concen: 97.362 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 91 Resp: 790045
Ion Ratio Lower Upper
91 100
126 31.2 15.7 47.1









#83
tert-Butylbenzene
Concen: 99.349 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC100


Tgt Ion:119 Resp: 785331
Ion Ratio Lower Upper
119 100
91 67.3 34.1 102.3
134 25.2 12.8 38.4



#84
1,2,4-Trimethylbenzene
Concen: 101.908 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion:105 Resp: 913037
Ion Ratio Lower Upper
105 100
120 43.7 22.3 66.8









#85
 sec-Butylbenzene
 Concen: 102.580 ug/l
 RT: 13.612 min Scan# 1983
 Delta R.T. -0.006 min
 Lab File: VN084213.D
 Acq: 30 Sep 2024 12:25




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion:105 Resp: 1082055
 Ion Ratio Lower Upper
 105 100
 134 19.4 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 105.328 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN084213.D
 Acq: 30 Sep 2024 12:25

Tgt Ion:119 Resp: 918730
 Ion Ratio Lower Upper
 119 100
 134 25.2 12.6 37.8
 91 25.1 12.4 37.2









#87
 1,3-Dichlorobenzene
 Concen: 94.046 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN084213.D
 Acq: 30 Sep 2024 12:25




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion:146 Resp: 463998
 Ion Ratio Lower Upper
 146 100
 111 43.1 21.8 65.3
 148 63.9 31.9 95.7



#88
 1,4-Dichlorobenzene
 Concen: 93.346 ug/l
 RT: 13.812 min Scan# 2016
 Delta R.T. 0.000 min
 Lab File: VN084213.D
 Acq: 30 Sep 2024 12:25

Tgt Ion:146 Resp: 465087
 Ion Ratio Lower Upper
 146 100
 111 41.7 21.2 63.6
 148 64.4 31.9 95.9









#89
n-Butylbenzene
Concen: 105.197 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100


Tgt Ion: 91 Resp: 832034
Ion Ratio Lower Upper
91 100
92 53.8 26.6 79.7
134 24.5 12.3 36.9



#90
Hexachloroethane
Concen: 95.289 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion: 117 Resp: 168712
Ion Ratio Lower Upper
117 100
201 67.4 34.4 103.2









#91
 1,2-Dichlorobenzene
 Concen: 91.843 ug/l
 RT: 14.106 min Scan# 2066
 Delta R.T. 0.000 min
 Lab File: VN084213.D
 Acq: 30 Sep 2024 12:25




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC100


Tgt Ion:146 Resp: 444319
 Ion Ratio Lower Upper
 146 100
 111 44.4 22.0 66.0
 148 63.1 32.1 96.3



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 88.446 ug/l
 RT: 14.717 min Scan# 2170
 Delta R.T. 0.000 min
 Lab File: VN084213.D
 Acq: 30 Sep 2024 12:25

Tgt Ion: 75 Resp: 59837
 Ion Ratio Lower Upper
 75 100
 155 77.6 38.1 114.5
 157 96.3 49.6 149.0





#93

1,2,4-Trichlorobenzene

Concen: 103.017 ug/l

RT: 15.388 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC100





Tgt Ion:180 Resp: 245477

Ion Ratio Lower Upper

180 100

182 94.6 47.0 141.0

145 33.9 16.4 49.4



#94

Hexachlorobutadiene

Concen: 89.034 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN084213.D

Acq: 30 Sep 2024 12:25




Tgt Ion:225 Resp: 105801


Ion Ratio Lower Upper

225 100

223 62.0 32.4 97.2

227 64.0 32.4 97.0









#95
Naphthalene
Concen: 99.356 ug/l
RT: 15.635 min Scan# 2326
Delta R.T. 0.000 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC100

Tgt Ion:128 Resp: 784876
Ion Ratio Lower Upper
128 100
127 12.7 10.5 15.7
129 10.7 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 97.736 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN084213.D
Acq: 30 Sep 2024 12:25

Tgt Ion:180 Resp: 233126
Ion Ratio Lower Upper
180 100
182 95.8 48.2 144.6
145 35.9 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084214.D
 Acq On : 30 Sep 2024 12:49
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Quant Time: Oct 01 06:48:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	175379	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	304127	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	264820	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	127701	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	129300	49.725	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.440%
35) Dibromofluoromethane	8.165	113	98531	49.143	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.280%
50) Toluene-d8	10.565	98	377886	51.228	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.460%
62) 4-Bromofluorobenzene	12.847	95	137388	51.124	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.240%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	107530	51.835	ug/l	100
3) Chloromethane	2.359	50	117689	49.686	ug/l	100
4) Vinyl Chloride	2.518	62	117031	51.007	ug/l	100
5) Bromomethane	2.953	94	75848	50.118	ug/l	100
6) Chloroethane	3.118	64	73440	44.693	ug/l	100
7) Trichlorofluoromethane	3.495	101	185982	52.055	ug/l	100
8) Diethyl Ether	3.959	74	68385	51.608	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.371	101	103552	50.583	ug/l	100
10) Methyl Iodide	4.595	142	141310	53.600	ug/l	100
11) Tert butyl alcohol	5.524	59	107986	251.850	ug/l	100
12) 1,1-Dichloroethene	4.342	96	102934	52.128	ug/l	100
13) Acrolein	4.177	56	124940	253.821	ug/l	100
14) Allyl chloride	5.024	41	174518	50.889	ug/l	100
15) Acrylonitrile	5.718	53	277346	256.145	ug/l	100
16) Acetone	4.424	43	299515	268.132	ug/l	100
17) Carbon Disulfide	4.712	76	302183	48.335	ug/l	100
18) Methyl Acetate	5.018	43	124007	50.934	ug/l	100
19) Methyl tert-butyl Ether	5.794	73	356478	53.487	ug/l	100
20) Methylene Chloride	5.277	84	114918	50.609	ug/l	100
21) trans-1,2-Dichloroethene	5.789	96	109048	52.711	ug/l	100
22) Diisopropyl ether	6.671	45	379516	53.501	ug/l	100
23) Vinyl Acetate	6.600	43	1444909	273.176	ug/l	100
24) 1,1-Dichloroethane	6.571	63	209222	52.742	ug/l	100
25) 2-Butanone	7.483	43	396369	260.524	ug/l	100
26) 2,2-Dichloropropane	7.488	77	189745	53.050	ug/l	100
27) cis-1,2-Dichloroethene	7.488	96	129873	51.951	ug/l	100
28) Bromochloromethane	7.812	49	86584	48.919	ug/l	100
29) Tetrahydrofuran	7.835	42	248328	264.488	ug/l	100
30) Chloroform	7.965	83	211178	51.249	ug/l	100
31) Cyclohexane	8.253	56	187282	48.624	ug/l	100
32) 1,1,1-Trichloroethane	8.171	97	193315	52.237	ug/l	100
36) 1,1-Dichloropropene	8.371	75	154855	53.172	ug/l	100
37) Ethyl Acetate	7.559	43	153024	51.216	ug/l	100
38) Carbon Tetrachloride	8.359	117	167086	52.301	ug/l	100
39) Methylcyclohexane	9.600	83	177433	54.741	ug/l	100
40) Benzene	8.606	78	472170	52.037	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084214.D
 Acq On : 30 Sep 2024 12:49
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Quant Time: Oct 01 06:48:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	87365	53.861	ug/l	100
42) 1,2-Dichloroethane	8.671	62	160606	52.210	ug/l	100
43) Isopropyl Acetate	8.688	43	272992	46.667	ug/l	100
44) Trichloroethene	9.347	130	110119	51.961	ug/l	100
45) 1,2-Dichloropropane	9.618	63	114143	53.176	ug/l	100
46) Dibromomethane	9.706	93	79007	54.542	ug/l	100
47) Bromodichloromethane	9.888	83	169263	52.595	ug/l	100
48) Methyl methacrylate	9.677	41	127042	54.534	ug/l	100
49) 1,4-Dioxane	9.694	88	44512	1037.858	ug/l	100
51) 4-Methyl-2-Pentanone	10.441	43	772927	271.750	ug/l	100
52) Toluene	10.629	92	295099	53.332	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	179453	54.703	ug/l	100
54) cis-1,3-Dichloropropene	10.312	75	193952	54.966	ug/l	100
55) 1,1,2-Trichloroethane	11.012	97	106288	53.583	ug/l	100
56) Ethyl methacrylate	10.871	69	183059	56.085	ug/l	100
57) 1,3-Dichloropropane	11.159	76	187372	52.864	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	426364	281.593	ug/l	100
59) 2-Hexanone	11.194	43	588361	276.767	ug/l	100
60) Dibromochloromethane	11.359	129	126992	54.184	ug/l	100
61) 1,2-Dibromoethane	11.465	107	108817	52.794	ug/l	100
64) Tetrachloroethene	11.100	164	94966	51.486	ug/l	100
65) Chlorobenzene	11.888	112	309849	52.770	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	106981	52.897	ug/l	100
67) Ethyl Benzene	11.959	91	561770	54.100	ug/l	100
68) m/p-Xylenes	12.071	106	427131	111.201	ug/l	100
69) o-Xylene	12.394	106	204854	56.370	ug/l	100
70) Styrene	12.412	104	347383	56.434	ug/l	100
71) Bromoform	12.576	173	83487	56.070	ug/l #	100
73) Isopropylbenzene	12.694	105	527625	54.145	ug/l	100
74) N-amyl acetate	12.494	43	237131	53.738	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.935	83	151079	51.560	ug/l	100
76) 1,2,3-Trichloropropane	12.988	75	203575	79.983	ug/l	62
77) Bromobenzene	12.982	156	120918	51.442	ug/l	100
78) n-propylbenzene	13.035	91	626503	55.490	ug/l	100
79) 2-Chlorotoluene	13.123	91	384795	53.411	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	445807	56.084	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	57572	52.943	ug/l	100
82) 4-Chlorotoluene	13.217	91	386854	53.341	ug/l	100
83) tert-Butylbenzene	13.435	119	386532	54.711	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	449720	56.161	ug/l	100
85) sec-Butylbenzene	13.617	105	529076	56.119	ug/l	100
86) p-Isopropyltoluene	13.729	119	448148	57.485	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	228219	51.755	ug/l	100
88) 1,4-Dichlorobenzene	13.812	146	227835	51.163	ug/l	100
89) n-Butylbenzene	14.053	91	394831	55.854	ug/l	100
90) Hexachloroethane	14.329	117	81626	51.582	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	218378	50.506	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	30379	50.241	ug/l	100
93) 1,2,4-Trichlorobenzene	15.388	180	119164	55.953	ug/l	100
94) Hexachlorobutadiene	15.500	225	50680	47.718	ug/l	100
95) Naphthalene	15.635	128	393177	55.688	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	114483	53.701	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084214.D
Acq On : 30 Sep 2024 12:49
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Quant Time: Oct 01 06:48:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 01 06:48:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:168 Resp: 175379
Ion Ratio Lower Upper
168 100
99 67.7 54.2 81.2



#2
Dichlorodifluoromethane
Concen: 51.835 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 85 Resp: 107530
Ion Ratio Lower Upper
85 100
87 35.3 17.6 52.9









#3
Chloromethane
Concen: 49.686 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 50 Resp: 117689
Ion Ratio Lower Upper
50 100
52 33.5 26.8 40.2



#4
Vinyl Chloride
Concen: 51.007 ug/l
RT: 2.518 min Scan# 96
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 62 Resp: 117031
Ion Ratio Lower Upper
62 100
64 32.8 26.2 39.4









#5
Bromomethane
Concen: 50.118 ug/l
RT: 2.953 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 94 Resp: 75848
Ion Ratio Lower Upper
94 100
96 91.1 72.9 109.3



#6
Chloroethane
Concen: 44.693 ug/l
RT: 3.118 min Scan# 198
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 64 Resp: 73440
Ion Ratio Lower Upper
64 100
66 29.4 23.5 35.3









#7
 Trichlorofluoromethane
 Concen: 52.055 ug/l
 RT: 3.495 min Scan# 262
 Delta R.T. 0.000 min
 Lab File: VN084214.D
 Acq: 30 Sep 2024 12:49




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion:101 Resp: 185982
 Ion Ratio Lower Upper
 101 100
 103 64.7 51.8 77.6



#8
 Diethyl Ether
 Concen: 51.608 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. 0.000 min
 Lab File: VN084214.D
 Acq: 30 Sep 2024 12:49

Tgt Ion: 74 Resp: 68385
 Ion Ratio Lower Upper
 74 100
 45 95.4 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.583 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:101 Resp: 103552
Ion Ratio Lower Upper
101 100
85 45.3 36.2 54.4
151 74.1 59.3 88.9



#10
Methyl Iodide
Concen: 53.600 ug/l
RT: 4.595 min Scan# 449
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion:142 Resp: 141310
Ion Ratio Lower Upper
142 100
127 44.1 35.3 52.9
141 13.8 11.0 16.6









#11
Tert butyl alcohol
Concen: 251.850 ug/l
RT: 5.524 min Scan# 607
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 59 Resp: 107986
Ion Ratio Lower Upper
59 100
57 10.2 8.2 12.2



#12
1,1-Dichloroethene
Concen: 52.128 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 96 Resp: 102934
Ion Ratio Lower Upper
96 100
61 160.7 128.6 192.8
98 64.6 51.7 77.5





#13

Acrolein

Concen: 253.821 ug/l

RT: 4.177 min Scan# 31

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion: 56 Resp: 124940

Ion Ratio Lower Upper

56 100

55 71.5 57.2 85.8



#14

Allyl chloride

Concen: 50.889 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49




Tgt Ion: 41 Resp: 174518


Ion Ratio Lower Upper

41 100

39 73.8 59.0 88.6

76 37.2 29.8 44.6









#15
Acrylonitrile
Concen: 256.145 ug/l
RT: 5.718 min Scan# 640
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 53 Resp: 277346
Ion Ratio Lower Upper
53 100
52 82.8 66.2 99.4
51 36.0 28.8 43.2



#16
Acetone
Concen: 268.132 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 43 Resp: 299515
Ion Ratio Lower Upper
43 100
58 30.5 24.4 36.6





#17

Carbon Disulfide

Concen: 48.335 ug/l

RT: 4.712 min Scan# 469

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion: 76 Resp: 302183

Ion Ratio Lower Upper

76 100

78 9.1 7.3 10.9



#18

Methyl Acetate

Concen: 50.934 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion: 43 Resp: 124007

Ion Ratio Lower Upper

43 100

74 24.0 19.2 28.8









#19
Methyl tert-butyl Ether
Concen: 53.487 ug/l
RT: 5.794 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 73 Resp: 356478
Ion Ratio Lower Upper
73 100
57 22.1 17.7 26.5



#20
Methylene Chloride
Concen: 50.609 ug/l
RT: 5.277 min Scan# 565
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 84 Resp: 114918
Ion Ratio Lower Upper
84 100
49 134.0 107.2 160.8
51 39.8 31.8 47.8
86 66.1 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 52.711 ug/l
RT: 5.789 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 96 Resp: 109048
Ion Ratio Lower Upper
96 100
61 135.5 108.4 162.6
98 62.3 49.8 74.8



#22
Diisopropyl ether
Concen: 53.501 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 45 Resp: 379516
Ion Ratio Lower Upper
45 100
43 52.9 42.3 63.5
87 27.2 21.8 32.6
59 12.0 9.6 14.4









#23
 Vinyl Acetate
 Concen: 273.176 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. 0.000 min
 Lab File: VN084214.D
 Acq: 30 Sep 2024 12:49




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 43 Resp: 1444909
 Ion Ratio Lower Upper
 43 100
 86 10.7 8.6 12.8



#24
 1,1-Dichloroethane
 Concen: 52.742 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. 0.000 min
 Lab File: VN084214.D
 Acq: 30 Sep 2024 12:49

Tgt Ion: 63 Resp: 209222
 Ion Ratio Lower Upper
 63 100
 98 6.8 3.4 10.2
 100 3.9 2.0 5.8









#25
2-Butanone
Concen: 260.524 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 396369
Ion Ratio Lower Upper
43 100
72 26.8 21.4 32.2



#26
2,2-Dichloropropane
Concen: 53.050 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 77 Resp: 189745
Ion Ratio Lower Upper
77 100
97 21.5 10.8 32.3









#27
 cis-1,2-Dichloroethene
 Concen: 51.951 ug/l
 RT: 7.488 min Scan# 941
 Delta R.T. 0.000 min
 Lab File: VN084214.D
 Acq: 30 Sep 2024 12:49

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Tgt Ion: 96 Resp: 129873




Ion	Ratio	Lower	Upper
96	100		
61	150.9	0.0	301.8
98	64.1	0.0	128.2




#28
 Bromochloromethane
 Concen: 48.919 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.000 min
 Lab File: VN084214.D
 Acq: 30 Sep 2024 12:49

Tgt Ion: 49 Resp: 86584

Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	3.6
130	72.5	58.0	87.0





#29

Tetrahydrofuran

Concen: 264.488 ug/l

RT: 7.835 min Scan# 1000

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion: 42 Resp: 248328

Ion Ratio Lower Upper

42 100

72 43.2 34.6 51.8

71 40.3 32.2 48.4



#30

Chloroform

Concen: 51.249 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion: 83 Resp: 211178

Ion Ratio Lower Upper

83 100

85 63.7 51.0 76.4









#31
Cyclohexane
Concen: 48.624 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 56 Resp: 187282
Ion Ratio Lower Upper
56 100
69 29.7 23.8 35.6
84 86.0 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 52.237 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 97 Resp: 193315
Ion Ratio Lower Upper
97 100
99 64.0 51.2 76.8
61 46.9 37.5 56.3









#33
 1,2-Dichloroethane-d4
 Concen: 49.725 ug/l
 RT: 8.577 min Scan# 1126
 Delta R.T. 0.000 min
 Lab File: VN084214.D
 Acq: 30 Sep 2024 12:49




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 65 Resp: 129300
 Ion Ratio Lower Upper
 65 100
 67 51.0 0.0 102.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 9.100 min Scan# 1215
 Delta R.T. 0.000 min
 Lab File: VN084214.D
 Acq: 30 Sep 2024 12:49

Tgt Ion: 114 Resp: 304127
 Ion Ratio Lower Upper
 114 100
 63 21.9 0.0 43.8
 88 15.8 0.0 31.6









#35
Dibromofluoromethane
Concen: 49.143 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:113 Resp: 98531
Ion Ratio Lower Upper
113 100
111 104.1 83.3 124.9
192 16.9 13.5 20.3



#36
1,1-Dichloropropene
Concen: 53.172 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 75 Resp: 154855
Ion Ratio Lower Upper
75 100
110 33.8 16.9 50.7
77 30.9 24.7 37.1









#37
Ethyl Acetate
Concen: 51.216 ug/l
RT: 7.559 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 153024
Ion Ratio Lower Upper
43 100
61 13.4 10.7 16.1
70 11.0 8.8 13.2



#38
Carbon Tetrachloride
Concen: 52.301 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 117 Resp: 167086
Ion Ratio Lower Upper
117 100
119 98.4 78.7 118.1
121 30.5 24.4 36.6





#39

Methylcyclohexane

Concen: 54.741 ug/l

RT: 9.600 min Scan# 11300

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion: 83 Resp: 177433

Ion Ratio Lower Upper

83 100

55 78.5 62.8 94.2

98 48.5 38.8 58.2



#40

Benzene

Concen: 52.037 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion: 78 Resp: 472170

Ion Ratio Lower Upper

78 100

77 23.9 19.1 28.7









#41
Methacrylonitrile
Concen: 53.861 ug/l
RT: 7.777 min Scan# 990
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:	41	Resp:	87365
Ion	Ratio	Lower	Upper
41	100		
39	54.1	43.3	64.9
67	70.1	56.1	84.1
52	30.6	24.5	36.7



#42
1,2-Dichloroethane
Concen: 52.210 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion:	62	Resp:	160606
Ion	Ratio	Lower	Upper
62	100		
98	9.1	0.0	18.2









#43
Isopropyl Acetate
Concen: 46.667 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 272992
Ion Ratio Lower Upper
43 100
61 25.5 20.4 30.6
87 12.9 10.3 15.5



#44
Trichloroethene
Concen: 51.961 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 130 Resp: 110119
Ion Ratio Lower Upper
130 100
95 101.5 0.0 203.0









#45
1,2-Dichloropropane
Concen: 53.176 ug/l
RT: 9.618 min Scan# 1303
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 63 Resp: 114143
Ion Ratio Lower Upper
63 100
65 30.8 24.6 37.0



#46
Dibromomethane
Concen: 54.542 ug/l
RT: 9.706 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 93 Resp: 79007
Ion Ratio Lower Upper
93 100
95 81.2 65.0 97.4
174 89.8 71.8 107.8





#47

Bromodichloromethane

Concen: 52.595 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion: 83 Resp: 169263

Ion Ratio Lower Upper

83 100

85 64.4 51.5 77.3

127 7.1 5.7 8.5



#48

Methyl methacrylate

Concen: 54.534 ug/l

RT: 9.677 min Scan# 1313

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49




Tgt Ion: 41 Resp: 127042


Ion Ratio Lower Upper

41 100

69 83.9 67.1 100.7

39 57.2 45.8 68.6





#49

1,4-Dioxane

Concen: 1037.858 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion: 88 Resp: 44512

Ion Ratio Lower Upper

88 100

43 31.6 25.3 37.9

58 70.5 56.4 84.6



#50

Toluene-d8

Concen: 51.228 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion: 98 Resp: 377886

Ion Ratio Lower Upper

98 100

100 65.9 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 271.750 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 772927
Ion Ratio Lower Upper
43 100
58 39.5 31.6 47.4



#52
Toluene
Concen: 53.332 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 92 Resp: 295099
Ion Ratio Lower Upper
92 100
91 171.6 137.3 205.9









#53
 t-1,3-Dichloropropene
 Concen: 54.703 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. 0.000 min
 Lab File: VN084214.D
 Acq: 30 Sep 2024 12:49




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion: 75 Resp: 179453
 Ion Ratio Lower Upper
 75 100
 77 31.7 25.4 38.0



#54
 cis-1,3-Dichloropropene
 Concen: 54.966 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. 0.000 min
 Lab File: VN084214.D
 Acq: 30 Sep 2024 12:49

Tgt Ion: 75 Resp: 193952
 Ion Ratio Lower Upper
 75 100
 77 30.2 24.2 36.2
 39 48.4 38.7 58.1









#55
1,1,2-Trichloroethane
Concen: 53.583 ug/l
RT: 11.012 min Scan# 1540
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 97 Resp: 106288




Ion	Ratio	Lower	Upper
97	100		
83	88.7	71.0	106.4
85	52.1	41.7	62.5
99	62.6	50.1	75.1




#56
Ethyl methacrylate
Concen: 56.085 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 69 Resp: 183059

Ion	Ratio	Lower	Upper
69	100		
41	66.6	53.3	79.9
39	36.0	28.8	43.2









#57
1,3-Dichloropropane
Concen: 52.864 ug/l
RT: 11.159 min Scan# 1565
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 76 Resp: 187372
Ion Ratio Lower Upper
76 100
78 32.2 25.8 38.6



#58
2-Chloroethyl Vinyl ether
Concen: 281.593 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 63 Resp: 426364
Ion Ratio Lower Upper
63 100
106 27.2 21.8 32.6









#59
2-Hexanone
Concen: 276.767 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 43 Resp: 588361
Ion Ratio Lower Upper
43 100
58 54.5 27.3 81.8



#60
Dibromochloromethane
Concen: 54.184 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion:129 Resp: 126992
Ion Ratio Lower Upper
129 100
127 76.6 38.3 114.9









#61
1,2-Dibromoethane
Concen: 52.794 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:107 Resp: 108817
Ion Ratio Lower Upper
107 100
109 94.2 75.4 113.0



#62
4-Bromofluorobenzene
Concen: 51.124 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 95 Resp: 137388
Ion Ratio Lower Upper
95 100
174 72.6 0.0 145.2
176 70.0 0.0 140.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:117 Resp: 264820

Ion Ratio Lower Upper

117 100

82 59.0 47.2 70.8

119 31.7 25.4 38.0



#64

Tetrachloroethene

Concen: 51.486 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Tgt Ion:164 Resp: 94966




Ion Ratio Lower Upper


164 100

166 126.2 101.0 151.4

129 101.0 80.8 121.2

131 97.6 78.1 117.1









#65
Chlorobenzene
Concen: 52.770 ug/l
RT: 11.888 min Scan# 111
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:112 Resp: 309849
Ion Ratio Lower Upper
112 100
114 33.8 27.0 40.6



#66
1,1,1,2-Tetrachloroethane
Concen: 52.897 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion:131 Resp: 106981
Ion Ratio Lower Upper
131 100
133 95.8 47.9 143.7
119 65.5 32.8 98.3





#67

Ethyl Benzene

Concen: 54.100 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tgt Ion: 91 Resp: 561770

Ion Ratio Lower Upper

91 100

106 30.6 24.5 36.7



#68

m/p-Xylenes

Concen: 111.201 ug/l

RT: 12.071 min Scan# 1720

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion:106 Resp: 427131

Ion Ratio Lower Upper

106 100

91 208.9 167.1 250.7









#69
o-Xylene
Concen: 56.370 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:106 Resp: 204854
Ion Ratio Lower Upper
106 100
91 220.4 110.2 330.6



#70
Styrene
Concen: 56.434 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion:104 Resp: 347383
Ion Ratio Lower Upper
104 100
78 53.4 42.7 64.1
103 54.5 43.6 65.4









#71
Bromoform
Concen: 56.070 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:173 Resp: 83487
Ion Ratio Lower Upper
173 100
175 47.1 23.5 70.6
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion:152 Resp: 127701
Ion Ratio Lower Upper
152 100
115 62.6 31.3 93.9
150 174.9 0.0 349.8









#73
Isopropylbenzene
Concen: 54.145 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:105 Resp: 527625
Ion Ratio Lower Upper
105 100
120 25.8 12.9 38.7



#74
N-ethyl acetate
Concen: 53.738 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 43 Resp: 237131
Ion Ratio Lower Upper
43 100
70 43.0 34.4 51.6
55 24.5 19.6 29.4
61 24.5 19.6 29.4





#75

1,1,2,2-Tetrachloroethane

Concen: 51.560 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion: 83 Resp: 151079

Ion Ratio Lower Upper

83 100

131 9.8 4.9 14.7

85 63.8 31.9 95.7



#76

1,2,3-Trichloropropane

Concen: 79.983 ug/l

RT: 12.988 min Scan# 1876

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion: 75 Resp: 203575

Ion Ratio Lower Upper

75 100

77 136.7 96.9 290.7





#77

Bromobenzene

Concen: 51.442 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:156 Resp: 120918

Ion Ratio Lower Upper

156 100

77 230.2 115.1 345.3

158 97.9 48.9 146.8



#78

n-propylbenzene

Concen: 55.490 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion: 91 Resp: 626503

Ion Ratio Lower Upper

91 100

120 21.6 10.8 32.4









#79
2-Chlorotoluene
Concen: 53.411 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 91 Resp: 384795
Ion Ratio Lower Upper
91 100
126 31.4 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 56.084 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 105 Resp: 445807
Ion Ratio Lower Upper
105 100
120 47.4 23.7 71.1









#81
trans-1,4-Dichloro-2-butene
Concen: 52.943 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion: 75 Resp: 57572
Ion Ratio Lower Upper
75 100
53 97.5 78.0 117.0
89 44.4 35.5 53.3



#82
4-Chlorotoluene
Concen: 53.341 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 91 Resp: 386854
Ion Ratio Lower Upper
91 100
126 31.4 15.7 47.1





#83

tert-Butylbenzene

Concen: 54.711 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:119 Resp: 386532

Ion Ratio Lower Upper

119 100

91 68.2 34.1 102.3

134 25.6 12.8 38.4



#84

1,2,4-Trimethylbenzene

Concen: 56.161 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion:105 Resp: 449720

Ion Ratio Lower Upper

105 100

120 44.5 22.3 66.8









#85
 sec-Butylbenzene
 Concen: 56.119 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. 0.000 min
 Lab File: VN084214.D
 Acq: 30 Sep 2024 12:49




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion:105 Resp: 529076
 Ion Ratio Lower Upper
 105 100
 134 19.6 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 57.485 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN084214.D
 Acq: 30 Sep 2024 12:49

Tgt Ion:119 Resp: 448148
 Ion Ratio Lower Upper
 119 100
 134 25.2 12.6 37.8
 91 24.8 12.4 37.2









#87
 1,3-Dichlorobenzene
 Concen: 51.755 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN084214.D
 Acq: 30 Sep 2024 12:49




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tgt Ion:146 Resp: 228219
 Ion Ratio Lower Upper
 146 100
 111 43.5 21.8 65.3
 148 63.8 31.9 95.7



#88
 1,4-Dichlorobenzene
 Concen: 51.163 ug/l
 RT: 13.812 min Scan# 2016
 Delta R.T. 0.000 min
 Lab File: VN084214.D
 Acq: 30 Sep 2024 12:49

Tgt Ion:146 Resp: 227835
 Ion Ratio Lower Upper
 146 100
 111 42.4 21.2 63.6
 148 63.9 31.9 95.9





#89

n-Butylbenzene

Concen: 55.854 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion: 91 Resp: 394831

Ion Ratio Lower Upper

91 100

92 53.1 26.6 79.7

134 24.6 12.3 36.9



#90

Hexachloroethane

Concen: 51.582 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. 0.000 min

Lab File: VN084214.D




Acq: 30 Sep 2024 12:49


Tgt Ion: 117 Resp: 81626

Ion Ratio Lower Upper

117 100

201 68.8 34.4 103.2









#91
1,2-Dichlorobenzene
Concen: 50.506 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tgt Ion:146 Resp: 218378
Ion Ratio Lower Upper
146 100
111 44.0 22.0 66.0
148 64.2 32.1 96.3



#92
1,2-Dibromo-3-Chloropropane
Concen: 50.241 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion: 75 Resp: 30379
Ion Ratio Lower Upper
75 100
155 76.3 38.1 114.5
157 99.3 49.6 149.0





#93

1,2,4-Trichlorobenzene

Concen: 55.953 ug/l

RT: 15.388 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:180 Resp: 119164

Ion Ratio Lower Upper

180 100

182 94.0 47.0 141.0

145 32.9 16.4 49.4



#94

Hexachlorobutadiene

Concen: 47.718 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. 0.000 min

Lab File: VN084214.D

Acq: 30 Sep 2024 12:49




Tgt Ion:225 Resp: 50680


Ion Ratio Lower Upper

225 100

223 64.8 32.4 97.2

227 64.7 32.4 97.0









#95
Naphthalene
Concen: 55.688 ug/l
RT: 15.635 min Scan# 2326
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:128 Resp: 393177
Ion Ratio Lower Upper
128 100
127 13.1 10.5 15.7
129 10.9 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 53.701 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. 0.000 min
Lab File: VN084214.D
Acq: 30 Sep 2024 12:49

Tgt Ion:180 Resp: 114483
Ion Ratio Lower Upper
180 100
182 96.4 48.2 144.6
145 34.6 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084215.D
 Acq On : 30 Sep 2024 13:13
 Operator : JC\MD
 Sample : VSTDIC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020

Quant Time: Oct 01 06:48:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	178087	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	307298	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	270754	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	127763	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	54409	20.606	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	41.220%#
35) Dibromofluoromethane	8.171	113	41903	20.684	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	41.360%#
50) Toluene-d8	10.565	98	152668	20.483	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	40.960%#
62) 4-Bromofluorobenzene	12.847	95	55201	20.329	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	40.660%#

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	42709	20.275	ug/l	94
3) Chloromethane	2.359	50	48218	20.047	ug/l	99
4) Vinyl Chloride	2.518	62	47375	20.334	ug/l	96
5) Bromomethane	2.942	94	30751	20.010	ug/l	97
6) Chloroethane	3.118	64	30650	18.369	ug/l	96
7) Trichlorofluoromethane	3.495	101	74579	20.557	ug/l	95
8) Diethyl Ether	3.959	74	29090	21.619	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.377	101	42958	20.665	ug/l	99
10) Methyl Iodide	4.589	142	54769	20.458	ug/l	97
11) Tert butyl alcohol	5.518	59	45501	104.506	ug/l	98
12) 1,1-Dichloroethene	4.336	96	42450	21.171	ug/l	95
13) Acrolein	4.183	56	48261	96.553	ug/l	98
14) Allyl chloride	5.024	41	77338	22.208	ug/l	92
15) Acrylonitrile	5.718	53	115451	105.004	ug/l	99
16) Acetone	4.430	43	119948	105.747	ug/l	98
17) Carbon Disulfide	4.706	76	124366	19.590	ug/l	99
18) Methyl Acetate	5.030	43	49436	19.996	ug/l	99
19) Methyl tert-butyl Ether	5.794	73	144972	21.421	ug/l	100
20) Methylene Chloride	5.271	84	47091	20.423	ug/l	94
21) trans-1,2-Dichloroethene	5.783	96	44064	20.976	ug/l	92
22) Diisopropyl ether	6.671	45	154575	21.460	ug/l	98
23) Vinyl Acetate	6.600	43	589404	109.739	ug/l	99
24) 1,1-Dichloroethane	6.565	63	82821	20.561	ug/l	98
25) 2-Butanone	7.482	43	165762	107.295	ug/l	100
26) 2,2-Dichloropropane	7.494	77	76824	21.152	ug/l	100
27) cis-1,2-Dichloroethene	7.488	96	52758	20.783	ug/l	99
28) Bromochloromethane	7.812	49	34591	19.246	ug/l	99
29) Tetrahydrofuran	7.835	42	100146	105.041	ug/l	99
30) Chloroform	7.965	83	85848	20.517	ug/l	99
31) Cyclohexane	8.253	56	77216	19.743	ug/l	98
32) 1,1,1-Trichloroethane	8.171	97	77552	20.637	ug/l	97
36) 1,1-Dichloropropene	8.371	75	61556	20.918	ug/l	99
37) Ethyl Acetate	7.559	43	61023	20.213	ug/l	99
38) Carbon Tetrachloride	8.359	117	67913	21.039	ug/l	95
39) Methylcyclohexane	9.600	83	70956	21.665	ug/l	99
40) Benzene	8.606	78	191581	20.896	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084215.D
 Acq On : 30 Sep 2024 13:13
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Quant Time: Oct 01 06:48:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.782	41	34943	21.320	ug/l	99
42) 1,2-Dichloroethane	8.671	62	64453	20.736	ug/l	99
43) Isopropyl Acetate	8.688	43	114278	19.334	ug/l	99
44) Trichloroethene	9.347	130	44412	20.740	ug/l	98
45) 1,2-Dichloropropane	9.618	63	46651	21.509	ug/l	100
46) Dibromomethane	9.706	93	31672	21.639	ug/l	97
47) Bromodichloromethane	9.888	83	68343	21.017	ug/l	96
48) Methyl methacrylate	9.682	41	50498	21.453	ug/l	98
49) 1,4-Dioxane	9.700	88	18411	424.848	ug/l #	93
51) 4-Methyl-2-Pentanone	10.441	43	313636	109.132	ug/l	99
52) Toluene	10.629	92	118412	21.179	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	70403	21.240	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	74778	20.973	ug/l	96
55) 1,1,2-Trichloroethane	11.018	97	42631	21.270	ug/l	98
56) Ethyl methacrylate	10.871	69	72159	21.880	ug/l	99
57) 1,3-Dichloropropane	11.159	76	75915	21.197	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.159	63	142843	93.367	ug/l	99
59) 2-Hexanone	11.194	43	237694	110.658	ug/l	100
60) Dibromochloromethane	11.359	129	50326	21.251	ug/l	97
61) 1,2-Dibromoethane	11.470	107	44554	21.393	ug/l	98
64) Tetrachloroethene	11.100	164	38055	20.179	ug/l	97
65) Chlorobenzene	11.888	112	123787	20.620	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.959	131	44285	21.417	ug/l	96
67) Ethyl Benzene	11.959	91	219635	20.688	ug/l	98
68) m/p-Xylenes	12.070	106	168826	42.989	ug/l	98
69) o-Xylene	12.394	106	79896	21.503	ug/l	98
70) Styrene	12.412	104	134066	21.302	ug/l	98
71) Bromoform	12.576	173	32824	21.561	ug/l #	99
73) Isopropylbenzene	12.694	105	207217	21.254	ug/l	99
74) N-amyl acetate	12.494	43	93392	21.154	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.935	83	60656	20.690	ug/l	98
76) 1,2,3-Trichloropropane	12.994	75	81444	31.983	ug/l	61
77) Bromobenzene	12.976	156	48803	20.752	ug/l	98
78) n-propylbenzene	13.035	91	241164	21.350	ug/l	98
79) 2-Chlorotoluene	13.123	91	147738	20.497	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	171600	21.577	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	22222	20.426	ug/l	92
82) 4-Chlorotoluene	13.217	91	151387	20.864	ug/l	100
83) tert-Butylbenzene	13.435	119	147727	20.899	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	176283	22.004	ug/l	99
85) sec-Butylbenzene	13.611	105	205477	21.784	ug/l	99
86) p-Isopropyltoluene	13.729	119	171766	22.022	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	90970	20.620	ug/l	99
88) 1,4-Dichlorobenzene	13.811	146	88628	19.893	ug/l	96
89) n-Butylbenzene	14.053	91	151334	21.398	ug/l	100
90) Hexachloroethane	14.335	117	32554	20.562	ug/l	95
91) 1,2-Dichlorobenzene	14.106	146	88942	20.560	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	12914	21.347	ug/l	96
93) 1,2,4-Trichlorobenzene	15.388	180	47149	22.128	ug/l	99
94) Hexachlorobutadiene	15.500	225	22742	21.402	ug/l	97
95) Naphthalene	15.641	128	149988	21.233	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	46153	21.639	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084215.D
Acq On : 30 Sep 2024 13:13
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Quant Time: Oct 01 06:48:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084215.D
 Acq On : 30 Sep 2024 13:13
 Operator : JC\MD
 Sample : VSTDIC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020

Quant Time: Oct 01 06:48:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:168 Resp: 178087
Ion Ratio Lower Upper
168 100
99 65.9 54.2 81.2



#2
Dichlorodifluoromethane
Concen: 20.275 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 85 Resp: 42709
Ion Ratio Lower Upper
85 100
87 31.8 17.6 52.9









#3
Chloromethane
Concen: 20.047 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 50 Resp: 48218
Ion Ratio Lower Upper
50 100
52 33.0 26.8 40.2



#4
Vinyl Chloride
Concen: 20.334 ug/l
RT: 2.518 min Scan# 96
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 62 Resp: 47375
Ion Ratio Lower Upper
62 100
64 30.6 26.2 39.4









#5
Bromomethane
Concen: 20.010 ug/l
RT: 2.942 min Scan# 1
Delta R.T. -0.012 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 94 Resp: 30751
Ion Ratio Lower Upper
94 100
96 94.3 72.9 109.3



#6
Chloroethane
Concen: 18.369 ug/l
RT: 3.118 min Scan# 198
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 64 Resp: 30650
Ion Ratio Lower Upper
64 100
66 31.8 23.5 35.3









#7
 Trichlorofluoromethane
 Concen: 20.557 ug/l
 RT: 3.495 min Scan# 262
 Delta R.T. -0.000 min
 Lab File: VN084215.D
 Acq: 30 Sep 2024 13:13




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020


Tgt Ion:101 Resp: 74579
 Ion Ratio Lower Upper
 101 100
 103 61.2 51.8 77.6



#8
 Diethyl Ether
 Concen: 21.619 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. -0.000 min
 Lab File: VN084215.D
 Acq: 30 Sep 2024 13:13

Tgt Ion: 74 Resp: 29090
 Ion Ratio Lower Upper
 74 100
 45 92.8 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.665 ug/l
RT: 4.377 min Scan# 411
Delta R.T. 0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:101 Resp: 42958
Ion Ratio Lower Upper
101 100
85 45.9 36.2 54.4
151 74.9 59.3 88.9



#10
Methyl Iodide
Concen: 20.458 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion:142 Resp: 54769
Ion Ratio Lower Upper
142 100
127 46.7 35.3 52.9
141 13.7 11.0 16.6









#11
Tert butyl alcohol
Concen: 104.506 ug/l
RT: 5.518 min Scan# 607
Delta R.T. -0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 59 Resp: 45501
Ion Ratio Lower Upper
59 100
57 9.4 8.2 12.2



#12
1,1-Dichloroethene
Concen: 21.171 ug/l
RT: 4.336 min Scan# 405
Delta R.T. -0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 96 Resp: 42450
Ion Ratio Lower Upper
96 100
61 166.5 128.6 192.8
98 60.0 51.7 77.5









#13
Acrolein
Concen: 96.553 ug/l
RT: 4.183 min Scan# 31
Delta R.T. 0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 56 Resp: 48261
Ion Ratio Lower Upper
56 100
55 69.7 57.2 85.8



#14
Allyl chloride
Concen: 22.208 ug/l
RT: 5.024 min Scan# 522
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 41 Resp: 77338
Ion Ratio Lower Upper
41 100
39 66.9 59.0 88.6
76 33.5 29.8 44.6









#15
Acrylonitrile
Concen: 105.004 ug/l
RT: 5.718 min Scan# 640
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 53 Resp: 115451




Ion	Ratio	Lower	Upper
53	100		
52	81.7	66.2	99.4
51	35.2	28.8	43.2




#16
Acetone
Concen: 105.747 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 43 Resp: 119948

Ion	Ratio	Lower	Upper
43	100		
58	29.5	24.4	36.6









#17
Carbon Disulfide
Concen: 19.590 ug/l
RT: 4.706 min Scan# 41
Delta R.T. -0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 76 Resp: 124366
Ion Ratio Lower Upper
76 100
78 8.6 7.3 10.9



#18
Methyl Acetate
Concen: 19.996 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.012 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 43 Resp: 49436
Ion Ratio Lower Upper
43 100
74 24.3 19.2 28.8









#19
Methyl tert-butyl Ether
Concen: 21.421 ug/l
RT: 5.794 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 73 Resp: 144972
Ion Ratio Lower Upper
73 100
57 22.1 17.7 26.5



#20
Methylene Chloride
Concen: 20.423 ug/l
RT: 5.271 min Scan# 564
Delta R.T. -0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 84 Resp: 47091
Ion Ratio Lower Upper
84 100
49 122.3 107.2 160.8
51 38.6 31.8 47.8
86 65.8 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 20.976 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 96 Resp: 44064
Ion Ratio Lower Upper
96 100
61 149.0 108.4 162.6
98 63.9 49.8 74.8



#22
Diisopropyl ether
Concen: 21.460 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 45 Resp: 154575
Ion Ratio Lower Upper
45 100
43 51.7 42.3 63.5
87 28.8 21.8 32.6
59 10.5 9.6 14.4









#23
 Vinyl Acetate
 Concen: 109.739 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. -0.000 min
 Lab File: VN084215.D
 Acq: 30 Sep 2024 13:13




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020


Tgt Ion: 43 Resp: 589404
 Ion Ratio Lower Upper
 43 100
 86 10.2 8.6 12.8



#24
 1,1-Dichloroethane
 Concen: 20.561 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. -0.006 min
 Lab File: VN084215.D
 Acq: 30 Sep 2024 13:13

Tgt Ion: 63 Resp: 82821
 Ion Ratio Lower Upper
 63 100
 98 6.8 3.4 10.2
 100 5.3 2.0 5.8









#25
2-Butanone
Concen: 107.295 ug/l
RT: 7.482 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 43 Resp: 165762
Ion Ratio Lower Upper
43 100
72 26.7 21.4 32.2



#26
2,2-Dichloropropane
Concen: 21.152 ug/l
RT: 7.494 min Scan# 942
Delta R.T. 0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 77 Resp: 76824
Ion Ratio Lower Upper
77 100
97 21.3 10.8 32.3









#27
 cis-1,2-Dichloroethene
 Concen: 20.783 ug/l
 RT: 7.488 min Scan# 941
 Delta R.T. -0.000 min
 Lab File: VN084215.D
 Acq: 30 Sep 2024 13:13




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020


Tgt Ion: 96 Resp: 52758
 Ion Ratio Lower Upper
 96 100
 61 149.6 0.0 301.8
 98 63.6 0.0 128.2



#28
 Bromochloromethane
 Concen: 19.246 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. -0.000 min
 Lab File: VN084215.D
 Acq: 30 Sep 2024 13:13

Tgt Ion: 49 Resp: 34591
 Ion Ratio Lower Upper
 49 100
 129 1.6 0.0 3.6
 130 71.9 58.0 87.0





#29

Tetrahydrofuran

Concen: 105.041 ug/l

RT: 7.835 min Scan# 1000

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion: 42 Resp: 100146

Ion Ratio Lower Upper

42 100

72 44.7 34.6 51.8

71 40.3 32.2 48.4



#30

Chloroform

Concen: 20.517 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. -0.000 min

Lab File: VN084215.D




Acq: 30 Sep 2024 13:13


Tgt Ion: 83 Resp: 85848

Ion Ratio Lower Upper

83 100

85 62.7 51.0 76.4









#31
Cyclohexane
Concen: 19.743 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:	56	Resp:	77216
Ion Ratio	Lower	Upper	
56	100		
69	31.9	23.8	35.6
84	87.6	68.8	103.2



#32
1,1,1-Trichloroethane
Concen: 20.637 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion:	97	Resp:	77552
Ion Ratio	Lower	Upper	
97	100		
99	68.3	51.2	76.8
61	46.9	37.5	56.3









#33
 1,2-Dichloroethane-d4
 Concen: 20.606 ug/l
 RT: 8.577 min Scan# 1126
 Delta R.T. -0.000 min
 Lab File: VN084215.D
 Acq: 30 Sep 2024 13:13




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020


Tgt Ion: 65 Resp: 54409
 Ion Ratio Lower Upper
 65 100
 67 49.7 0.0 102.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 9.100 min Scan# 1215
 Delta R.T. -0.000 min
 Lab File: VN084215.D
 Acq: 30 Sep 2024 13:13

Tgt Ion: 114 Resp: 307298
 Ion Ratio Lower Upper
 114 100
 63 21.8 0.0 43.8
 88 16.6 0.0 31.6





#35

Dibromofluoromethane

Concen: 20.684 ug/l

RT: 8.171 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:113 Resp: 41903

Ion Ratio Lower Upper

113 100

111 100.2 83.3 124.9

192 16.8 13.5 20.3



#36

1,1-Dichloropropene

Concen: 20.918 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion: 75 Resp: 61556


Ion Ratio Lower Upper

75 100

110 33.1 16.9 50.7

77 31.1 24.7 37.1









#37
Ethyl Acetate
Concen: 20.213 ug/l
RT: 7.559 min Scan# 91
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 61023
Ion Ratio Lower Upper
43 100
61 14.1 10.7 16.1
70 11.2 8.8 13.2



#38
Carbon Tetrachloride
Concen: 21.039 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 117 Resp: 67913
Ion Ratio Lower Upper
117 100
119 93.0 78.7 118.1
121 29.2 24.4 36.6





#39

Methylcyclohexane

Concen: 21.665 ug/l

RT: 9.600 min Scan# 11300

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion: 83 Resp: 70956

Ion Ratio Lower Upper

83 100

55 77.8 62.8 94.2

98 47.9 38.8 58.2



#40

Benzene

Concen: 20.896 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN084215.D




Acq: 30 Sep 2024 13:13


Tgt Ion: 78 Resp: 191581

Ion Ratio Lower Upper

78 100

77 22.5 19.1 28.7









#41
Methacrylonitrile
Concen: 21.320 ug/l
RT: 7.782 min Scan# 991
Delta R.T. 0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:	41	Resp:	34943
Ion	Ratio	Lower	Upper
41	100		
39	53.8	43.3	64.9
67	71.5	56.1	84.1
52	29.7	24.5	36.7



#42
1,2-Dichloroethane
Concen: 20.736 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion:	62	Resp:	64453
Ion	Ratio	Lower	Upper
62	100		
98	9.5	0.0	18.2









#43
Isopropyl Acetate
Concen: 19.334 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 43 Resp: 114278
Ion Ratio Lower Upper
43 100
61 24.7 20.4 30.6
87 12.7 10.3 15.5



#44
Trichloroethene
Concen: 20.740 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 130 Resp: 44412
Ion Ratio Lower Upper
130 100
95 103.2 0.0 203.0









#45
1,2-Dichloropropane
Concen: 21.509 ug/l
RT: 9.618 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 63 Resp: 46651
Ion Ratio Lower Upper
63 100
65 30.9 24.6 37.0



#46
Dibromomethane
Concen: 21.639 ug/l
RT: 9.706 min Scan# 1318
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 93 Resp: 31672
Ion Ratio Lower Upper
93 100
95 83.8 65.0 97.4
174 92.0 71.8 107.8









#47
Bromodichloromethane
Concen: 21.017 ug/l
RT: 9.888 min Scan# 1349
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 83 Resp: 68343
Ion Ratio Lower Upper
83 100
85 61.4 51.5 77.3
127 8.0 5.7 8.5



#48
Methyl methacrylate
Concen: 21.453 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. 0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 41 Resp: 50498
Ion Ratio Lower Upper
41 100
69 82.3 67.1 100.7
39 59.0 45.8 68.6





#49

1,4-Dioxane

Concen: 424.848 ug/l

RT: 9.700 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion: 88 Resp: 18411

Ion Ratio Lower Upper

88 100

43 22.7 25.3 37.9#

58 72.9 56.4 84.6



#50

Toluene-d8

Concen: 20.483 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN084215.D




Acq: 30 Sep 2024 13:13


Tgt Ion: 98 Resp: 152668

Ion Ratio Lower Upper

98 100

100 67.0 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 109.132 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 43 Resp: 313636
Ion Ratio Lower Upper
43 100
58 38.9 31.6 47.4



#52
Toluene
Concen: 21.179 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 92 Resp: 118412
Ion Ratio Lower Upper
92 100
91 171.0 137.3 205.9









#53
t-1,3-Dichloropropene
Concen: 21.240 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 75 Resp: 70403
Ion Ratio Lower Upper
75 100
77 32.2 25.4 38.0



#54
cis-1,3-Dichloropropene
Concen: 20.973 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 75 Resp: 74778
Ion Ratio Lower Upper
75 100
77 32.6 24.2 36.2
39 51.4 38.7 58.1









#55
1,1,2-Trichloroethane
Concen: 21.270 ug/l
RT: 11.018 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 97	Resp: 42631
Ion Ratio	Lower Upper
97 100	
83 86.8	71.0 106.4
85 53.6	41.7 62.5
99 63.9	50.1 75.1



#56
Ethyl methacrylate
Concen: 21.880 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 69	Resp: 72159
Ion Ratio	Lower Upper
69 100	
41 67.2	53.3 79.9
39 36.6	28.8 43.2





#57

1,3-Dichloropropane

Concen: 21.197 ug/l

RT: 11.159 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC020

Tgt Ion: 76 Resp: 75915

Ion Ratio Lower Upper

76 100

78 31.2 25.8 38.6



#58

2-Chloroethyl Vinyl ether

Concen: 93.367 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN084215.D




Acq: 30 Sep 2024 13:13


Tgt Ion: 63 Resp: 142843

Ion Ratio Lower Upper

63 100

106 27.5 21.8 32.6









#59
2-Hexanone
Concen: 110.658 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 237694
Ion Ratio Lower Upper
43 100
58 54.4 27.3 81.8



#60
Dibromochloromethane
Concen: 21.251 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 129 Resp: 50326
Ion Ratio Lower Upper
129 100
127 78.8 38.3 114.9





#61

1,2-Dibromoethane

Concen: 21.393 ug/l

RT: 11.470 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC020

Tgt Ion:107 Resp: 44554

Ion Ratio Lower Upper

107 100

109 92.5 75.4 113.0



#62

4-Bromofluorobenzene

Concen: 20.329 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion: 95 Resp: 55201


Ion Ratio Lower Upper

95 100

174 73.2 0.0 145.2

176 68.9 0.0 140.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:117 Resp: 270754
Ion Ratio Lower Upper
117 100
82 57.7 47.2 70.8
119 31.2 25.4 38.0



#64
Tetrachloroethene
Concen: 20.179 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion:164 Resp: 38055
Ion Ratio Lower Upper
164 100
166 126.4 101.0 151.4
129 106.4 80.8 121.2
131 94.2 78.1 117.1









#65
Chlorobenzene
Concen: 20.620 ug/l
RT: 11.888 min Scan# 111
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion:112 Resp: 123787
Ion Ratio Lower Upper
112 100
114 33.3 27.0 40.6



#66
1,1,1,2-Tetrachloroethane
Concen: 21.417 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion:131 Resp: 44285
Ion Ratio Lower Upper
131 100
133 92.4 47.9 143.7
119 62.0 32.8 98.3









#67
Ethyl Benzene
Concen: 20.688 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 91 Resp: 219635
Ion Ratio Lower Upper
91 100
106 29.7 24.5 36.7



#68
m/p-Xylenes
Concen: 42.989 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion:106 Resp: 168826
Ion Ratio Lower Upper
106 100
91 206.4 167.1 250.7









#69
o-Xylene
Concen: 21.503 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion:106 Resp: 79896
Ion Ratio Lower Upper
106 100
91 223.3 110.2 330.6



#70
Styrene
Concen: 21.302 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion:104 Resp: 134066
Ion Ratio Lower Upper
104 100
78 51.8 42.7 64.1
103 55.2 43.6 65.4





#71

Bromoform

Concen: 21.561 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:173 Resp: 32824

Ion Ratio Lower Upper

173 100

175 47.8 23.5 70.6

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion:152 Resp: 127763


Ion Ratio Lower Upper

152 100

115 62.3 31.3 93.9

150 164.5 0.0 349.8





#73

Isopropylbenzene

Concen: 21.254 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC020

Tgt Ion:105 Resp: 207217

Ion Ratio Lower Upper

105 100

120 25.0 12.9 38.7



#74

N-ethyl acetate

Concen: 21.154 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Tgt Ion: 43 Resp: 93392




Ion Ratio Lower Upper


43 100

70 42.8 34.4 51.6

55 25.0 19.6 29.4

61 24.6 19.6 29.4









#75
1,1,2,2-Tetrachloroethane
Concen: 20.690 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 83 Resp: 60656
Ion Ratio Lower Upper
83 100
131 10.6 4.9 14.7
85 65.1 31.9 95.7



#76
1,2,3-Trichloropropane
Concen: 31.983 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. 0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 75 Resp: 81444
Ion Ratio Lower Upper
75 100
77 136.1 96.9 290.7









#77
Bromobenzene
Concen: 20.752 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion:156 Resp: 48803
Ion Ratio Lower Upper
156 100
77 227.1 115.1 345.3
158 96.5 48.9 146.8



#78
n-propylbenzene
Concen: 21.350 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 91 Resp: 241164
Ion Ratio Lower Upper
91 100
120 22.4 10.8 32.4









#79
2-Chlorotoluene
Concen: 20.497 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 91 Resp: 147738
Ion Ratio Lower Upper
91 100
126 32.0 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 21.577 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 105 Resp: 171600
Ion Ratio Lower Upper
105 100
120 47.5 23.7 71.1









#81
trans-1,4-Dichloro-2-butene
Concen: 20.426 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 75 Resp: 22222
Ion Ratio Lower Upper
75 100
53 105.8 78.0 117.0
89 48.1 35.5 53.3



#82
4-Chlorotoluene
Concen: 20.864 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 91 Resp: 151387
Ion Ratio Lower Upper
91 100
126 31.6 15.7 47.1









#83
tert-Butylbenzene
Concen: 20.899 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion:119 Resp: 147727
Ion Ratio Lower Upper
119 100
91 68.9 34.1 102.3
134 25.9 12.8 38.4



#84
1,2,4-Trimethylbenzene
Concen: 22.004 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion:105 Resp: 176283
Ion Ratio Lower Upper
105 100
120 43.6 22.3 66.8









#85
 sec-Butylbenzene
 Concen: 21.784 ug/l
 RT: 13.611 min Scan# 1982
 Delta R.T. -0.006 min
 Lab File: VN084215.D
 Acq: 30 Sep 2024 13:13




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020


Tgt Ion:105 Resp: 205477
 Ion Ratio Lower Upper
 105 100
 134 19.4 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 22.022 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN084215.D
 Acq: 30 Sep 2024 13:13

Tgt Ion:119 Resp: 171766
 Ion Ratio Lower Upper
 119 100
 134 25.9 12.6 37.8
 91 24.7 12.4 37.2









#87
1,3-Dichlorobenzene
Concen: 20.620 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion:146 Resp: 90970
Ion Ratio Lower Upper
146 100
111 42.4 21.8 65.3
148 63.4 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 19.893 ug/l
RT: 13.811 min Scan# 2016
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion:146 Resp: 88628
Ion Ratio Lower Upper
146 100
111 45.2 21.2 63.6
148 66.2 31.9 95.9









#89
n-Butylbenzene
Concen: 21.398 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 91 Resp: 151334
Ion Ratio Lower Upper
91 100
92 53.1 26.6 79.7
134 23.9 12.3 36.9



#90
Hexachloroethane
Concen: 20.562 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. 0.006 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 117 Resp: 32554
Ion Ratio Lower Upper
117 100
201 65.1 34.4 103.2









#91
1,2-Dichlorobenzene
Concen: 20.560 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:146 Resp: 88942
Ion Ratio Lower Upper
146 100
111 42.8 22.0 66.0
148 63.6 32.1 96.3



#92
1,2-Dibromo-3-Chloropropane
Concen: 21.347 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN084215.D
Acq: 30 Sep 2024 13:13

Tgt Ion: 75 Resp: 12914
Ion Ratio Lower Upper
75 100
155 76.9 38.1 114.5
157 93.4 49.6 149.0





#93

1,2,4-Trichlorobenzene

Concen: 22.128 ug/l

RT: 15.388 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:180 Resp: 47149

Ion Ratio Lower Upper

180 100

182 94.5 47.0 141.0

145 32.3 16.4 49.4



#94

Hexachlorobutadiene

Concen: 21.402 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion:225 Resp: 22742


Ion Ratio Lower Upper

225 100

223 62.8 32.4 97.2

227 62.7 32.4 97.0





#95

Naphthalene

Concen: 21.233 ug/l

RT: 15.641 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion:128 Resp: 149988

Ion Ratio Lower Upper

128 100

127 12.8 10.5 15.7

129 10.7 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 21.639 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN084215.D

Acq: 30 Sep 2024 13:13




Tgt Ion:180 Resp: 46153

Ion Ratio Lower Upper

180 100

182 90.7 48.2 144.6

145 35.1 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084216.D
 Acq On : 30 Sep 2024 13:37
 Operator : JC\MD
 Sample : VSTDIC010
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC010

Quant Time: Oct 01 06:49:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	176243	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	304981	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	262325	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	119657	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	25083	9.599	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	19.200%#
35) Dibromofluoromethane	8.165	113	19289	9.594	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	19.180%#
50) Toluene-d8	10.565	98	70000	9.463	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	18.920%#
62) 4-Bromofluorobenzene	12.847	95	24756	9.186	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	18.380%#

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	2.130	85	19806	9.501	ug/l	96
3) Chloromethane	2.359	50	22840	9.595	ug/l	97
4) Vinyl Chloride	2.518	62	22584	9.795	ug/l	97
5) Bromomethane	2.965	94	14957	9.835	ug/l	96
6) Chloroethane	3.124	64	15254	9.238	ug/l	89
7) Trichlorofluoromethane	3.494	101	33820	9.420	ug/l	99
8) Diethyl Ether	3.965	74	12422	9.328	ug/l	96
9) 1,1,2-Trichlorotrifluo...	4.377	101	18747	9.113	ug/l	98
10) Methyl Iodide	4.594	142	24268	9.160	ug/l	96
11) Tert butyl alcohol	5.512	59	20490	47.553	ug/l #	86
12) 1,1-Dichloroethene	4.341	96	18798	9.473	ug/l	86
13) Acrolein	4.183	56	22976	46.448	ug/l	98
14) Allyl chloride	5.024	41	31571	9.161	ug/l	98
15) Acrylonitrile	5.718	53	52809	48.533	ug/l	99
16) Acetone	4.424	43	52518	46.785	ug/l	96
17) Carbon Disulfide	4.718	76	58158	9.257	ug/l	99
18) Methyl Acetate	5.024	43	24366	9.959	ug/l	95
19) Methyl tert-butyl Ether	5.794	73	63531	9.486	ug/l	94
20) Methylene Chloride	5.277	84	21801	9.554	ug/l	87
21) trans-1,2-Dichloroethene	5.788	96	20107	9.672	ug/l	85
22) Diisopropyl ether	6.671	45	66959	9.393	ug/l	97
23) Vinyl Acetate	6.606	43	248594	46.769	ug/l	98
24) 1,1-Dichloroethane	6.571	63	38215	9.586	ug/l	99
25) 2-Butanone	7.482	43	73877	48.320	ug/l	99
26) 2,2-Dichloropropane	7.488	77	34655	9.641	ug/l	98
27) cis-1,2-Dichloroethene	7.488	96	23095	9.193	ug/l	98
28) Bromochloromethane	7.812	49	14904	8.379	ug/l	100
29) Tetrahydrofuran	7.835	42	46917	49.725	ug/l	99
30) Chloroform	7.965	83	39355	9.504	ug/l	90
31) Cyclohexane	8.253	56	36916	9.538	ug/l	92
32) 1,1,1-Trichloroethane	8.171	97	35869	9.645	ug/l #	87
36) 1,1-Dichloropropene	8.377	75	28525	9.767	ug/l	99
37) Ethyl Acetate	7.565	43	29754	9.931	ug/l #	81
38) Carbon Tetrachloride	8.365	117	30964	9.665	ug/l	92
39) Methylcyclohexane	9.600	83	30932	9.516	ug/l	95
40) Benzene	8.606	78	86029	9.454	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084216.D
 Acq On : 30 Sep 2024 13:37
 Operator : JC\MD
 Sample : VSTDICC010
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC010

Quant Time: Oct 01 06:49:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	15332	9.426	ug/l	97
42) 1,2-Dichloroethane	8.671	62	30402	9.855	ug/l	97
43) Isopropyl Acetate	8.688	43	52246	8.906	ug/l	99
44) Trichloroethene	9.347	130	20092	9.454	ug/l	99
45) 1,2-Dichloropropane	9.618	63	20703	9.618	ug/l	96
46) Dibromomethane	9.706	93	13770	9.479	ug/l	95
47) Bromodichloromethane	9.888	83	30289	9.385	ug/l #	98
48) Methyl methacrylate	9.682	41	22612	9.679	ug/l	98
49) 1,4-Dioxane	9.694	88	8627	200.587	ug/l #	78
51) 4-Methyl-2-Pentanone	10.441	43	142632	50.007	ug/l	99
52) Toluene	10.629	92	52537	9.468	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	31555	9.592	ug/l	93
54) cis-1,3-Dichloropropene	10.312	75	34684	9.802	ug/l	96
55) 1,1,2-Trichloroethane	11.018	97	19245	9.675	ug/l	97
56) Ethyl methacrylate	10.870	69	31262	9.551	ug/l	97
57) 1,3-Dichloropropane	11.165	76	33613	9.457	ug/l	97
58) 2-Chloroethyl Vinyl ether	10.159	63	71035	46.784	ug/l	99
59) 2-Hexanone	11.194	43	103615	48.604	ug/l	100
60) Dibromochloromethane	11.359	129	22833	9.715	ug/l	99
61) 1,2-Dibromoethane	11.470	107	19220	9.299	ug/l	95
64) Tetrachloroethene	11.100	164	17743	9.711	ug/l	97
65) Chlorobenzene	11.894	112	56099	9.645	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	19227	9.597	ug/l	97
67) Ethyl Benzene	11.964	91	96533	9.385	ug/l	97
68) m/p-Xylenes	12.070	106	72910	19.162	ug/l	100
69) o-Xylene	12.394	106	34965	9.713	ug/l	100
70) Styrene	12.412	104	58367	9.572	ug/l	99
71) Bromoform	12.576	173	13618	9.233	ug/l #	96
73) Isopropylbenzene	12.694	105	88000	9.638	ug/l	99
74) N-amyl acetate	12.494	43	40801	9.868	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	26966	9.822	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	36789	15.426	ug/l	59
77) Bromobenzene	12.982	156	21439	9.734	ug/l	99
78) n-propylbenzene	13.035	91	105334	9.957	ug/l	99
79) 2-Chlorotoluene	13.123	91	65511	9.704	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	73637	9.887	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.729	75	10237	10.047	ug/l	94
82) 4-Chlorotoluene	13.223	91	66449	9.778	ug/l	99
83) tert-Butylbenzene	13.435	119	64152	9.691	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	74245	9.895	ug/l	99
85) sec-Butylbenzene	13.611	105	88913	10.065	ug/l	100
86) p-Isopropyltoluene	13.729	119	71675	9.812	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	39402	9.536	ug/l	98
88) 1,4-Dichlorobenzene	13.811	146	39191	9.392	ug/l	97
89) n-Butylbenzene	14.053	91	61422	9.273	ug/l	99
90) Hexachloroethane	14.335	117	14560	9.820	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	38594	9.526	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	5799	10.235	ug/l	97
93) 1,2,4-Trichlorobenzene	15.388	180	18960	9.501	ug/l	99
94) Hexachlorobutadiene	15.500	225	9592	9.638	ug/l	95
95) Naphthalene	15.635	128	63596	9.613	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	19629	9.826	ug/l	97


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084216.D
Acq On : 30 Sep 2024 13:37
Operator : JC\MD
Sample : VSTDICC010
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010

Quant Time: Oct 01 06:49:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:168 Resp: 176243
Ion Ratio Lower Upper
168 100
99 66.2 54.2 81.2



#2
Dichlorodifluoromethane
Concen: 9.501 ug/l
RT: 2.130 min Scan# 30
Delta R.T. 0.006 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 85 Resp: 19806
Ion Ratio Lower Upper
85 100
87 32.8 17.6 52.9









#3
Chloromethane
Concen: 9.595 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 50 Resp: 22840
Ion Ratio Lower Upper
50 100
52 31.7 26.8 40.2



#4
Vinyl Chloride
Concen: 9.795 ug/l
RT: 2.518 min Scan# 96
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 62 Resp: 22584
Ion Ratio Lower Upper
62 100
64 30.9 26.2 39.4









#5
Bromomethane
Concen: 9.835 ug/l
RT: 2.965 min Scan# 11
Delta R.T. 0.012 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 94 Resp: 14957
Ion Ratio Lower Upper
94 100
96 86.8 72.9 109.3



#6
Chloroethane
Concen: 9.238 ug/l
RT: 3.124 min Scan# 199
Delta R.T. 0.006 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 64 Resp: 15254
Ion Ratio Lower Upper
64 100
66 35.0 23.5 35.3









#7
 Trichlorofluoromethane
 Concen: 9.420 ug/l
 RT: 3.494 min Scan# 210
 Delta R.T. -0.000 min
 Lab File: VN084216.D
 Acq: 30 Sep 2024 13:37




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC010


Tgt Ion:101 Resp: 33820
 Ion Ratio Lower Upper
 101 100
 103 64.2 51.8 77.6



#8
 Diethyl Ether
 Concen: 9.328 ug/l
 RT: 3.965 min Scan# 342
 Delta R.T. 0.006 min
 Lab File: VN084216.D
 Acq: 30 Sep 2024 13:37

Tgt Ion: 74 Resp: 12422
 Ion Ratio Lower Upper
 74 100
 45 99.0 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 9.113 ug/l
RT: 4.377 min Scan# 411
Delta R.T. 0.006 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:101 Resp: 18747
Ion Ratio Lower Upper
101 100
85 47.3 36.2 54.4
151 75.1 59.3 88.9



#10
Methyl Iodide
Concen: 9.160 ug/l
RT: 4.594 min Scan# 449
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion:142 Resp: 24268
Ion Ratio Lower Upper
142 100
127 46.4 35.3 52.9
141 15.9 11.0 16.6









#11
Tert butyl alcohol
Concen: 47.553 ug/l
RT: 5.512 min Scan# 607
Delta R.T. -0.012 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 59 Resp: 20490
Ion Ratio Lower Upper
59 100
57 5.1 8.2 12.2#



#12
1,1-Dichloroethene
Concen: 9.473 ug/l
RT: 4.341 min Scan# 406
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 96 Resp: 18798
Ion Ratio Lower Upper
96 100
61 186.2 128.6 192.8
98 66.5 51.7 77.5





#13

Acrolein

Concen: 46.448 ug/l

RT: 4.183 min Scan# 31

Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC010

Tgt Ion: 56 Resp: 22976

Ion Ratio Lower Upper

56 100

55 72.8 57.2 85.8



#14

Allyl chloride

Concen: 9.161 ug/l

RT: 5.024 min Scan# 522

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion: 41 Resp: 31571


Ion Ratio Lower Upper

41 100

39 71.4 59.0 88.6

76 36.1 29.8 44.6









#15
Acrylonitrile
Concen: 48.533 ug/l
RT: 5.718 min Scan# 640
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 53 Resp: 52809
Ion Ratio Lower Upper
53 100
52 82.2 66.2 99.4
51 35.4 28.8 43.2



#16
Acetone
Concen: 46.785 ug/l
RT: 4.424 min Scan# 420
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 43 Resp: 52518
Ion Ratio Lower Upper
43 100
58 32.8 24.4 36.6









#17
Carbon Disulfide
Concen: 9.257 ug/l
RT: 4.718 min Scan# 41
Delta R.T. 0.006 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 76 Resp: 58158
Ion Ratio Lower Upper
76 100
78 8.6 7.3 10.9



#18
Methyl Acetate
Concen: 9.959 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 43 Resp: 24366
Ion Ratio Lower Upper
43 100
74 21.3 19.2 28.8





#19

Methyl tert-butyl Ether

Concen: 9.486 ug/l

RT: 5.794 min Scan# 61

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC010

Tgt Ion: 73 Resp: 63531

Ion Ratio Lower Upper

73 100

57 19.3 17.7 26.5



#20

Methylene Chloride

Concen: 9.554 ug/l

RT: 5.277 min Scan# 565

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion: 84 Resp: 21801




Ion Ratio Lower Upper


84 100

49 115.7 107.2 160.8

51 36.0 31.8 47.8

86 57.2 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 9.672 ug/l
RT: 5.788 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 96 Resp: 20107
Ion Ratio Lower Upper
96 100
61 153.1 108.4 162.6
98 73.7 49.8 74.8



#22
Diisopropyl ether
Concen: 9.393 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 45 Resp: 66959
Ion Ratio Lower Upper
45 100
43 55.7 42.3 63.5
87 25.9 21.8 32.6
59 11.9 9.6 14.4









#23
 Vinyl Acetate
 Concen: 46.769 ug/l
 RT: 6.606 min Scan# 791
 Delta R.T. 0.006 min
 Lab File: VN084216.D
 Acq: 30 Sep 2024 13:37




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC010


Tgt Ion: 43 Resp: 248594
 Ion Ratio Lower Upper
 43 100
 86 9.8 8.6 12.8



#24
 1,1-Dichloroethane
 Concen: 9.586 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. -0.000 min
 Lab File: VN084216.D
 Acq: 30 Sep 2024 13:37

Tgt Ion: 63 Resp: 38215
 Ion Ratio Lower Upper
 63 100
 98 7.3 3.4 10.2
 100 3.8 2.0 5.8









#25
2-Butanone
Concen: 48.320 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010


Tgt Ion: 43 Resp: 73877
Ion Ratio Lower Upper
43 100
72 27.5 21.4 32.2



#26
2,2-Dichloropropane
Concen: 9.641 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 77 Resp: 34655
Ion Ratio Lower Upper
77 100
97 22.7 10.8 32.3









#27
cis-1,2-Dichloroethene
Concen: 9.193 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010


Tgt Ion	96	Ratio	100	Lower	Upper
61	152.9	0.0	301.8		
98	66.0	0.0	128.2		



#28
Bromochloromethane
Concen: 8.379 ug/l
RT: 7.812 min Scan# 996
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion	49	Ratio	100	Lower	Upper
129	0.4	0.0	3.6		
130	72.3	58.0	87.0		





#29

Tetrahydrofuran

Concen: 49.725 ug/l

RT: 7.835 min Scan# 1000

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion: 42 Resp: 46917

Ion Ratio Lower Upper

42 100

72 42.7 34.6 51.8

71 39.5 32.2 48.4



#30

Chloroform

Concen: 9.504 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. -0.000 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37


Tgt Ion: 83 Resp: 39355

Ion Ratio Lower Upper

83 100

85 71.7 51.0 76.4





#31

Cyclohexane

Concen: 9.538 ug/l

RT: 8.253 min Scan# 1071

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion: 56 Resp: 36916

Ion Ratio Lower Upper

56 100

69 33.8 23.8 35.6

84 79.0 68.8 103.2



#32

1,1,1-Trichloroethane

Concen: 9.645 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion: 97 Resp: 35869


Ion Ratio Lower Upper

97 100

99 46.3 51.2 76.8

61 46.7 37.5 56.3





#33

1,2-Dichloroethane-d4

Concen: 9.599 ug/l

RT: 8.582 min Scan# 1127

Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC010

Tgt Ion: 65 Resp: 25083

Ion Ratio Lower Upper

65 100

67 52.7 0.0 102.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion: 114 Resp: 304981


Ion Ratio Lower Upper

114 100

63 21.5 0.0 43.8

88 16.7 0.0 31.6





#35

Dibromofluoromethane

Concen: 9.594 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion:113 Resp: 19289

Ion Ratio Lower Upper

113 100

111 108.7 83.3 124.9

192 17.7 13.5 20.3



#36

1,1-Dichloropropene

Concen: 9.767 ug/l

RT: 8.377 min Scan# 1092

Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion: 75 Resp: 28525


Ion Ratio Lower Upper

75 100

110 32.3 16.9 50.7

77 31.0 24.7 37.1





#37

Ethyl Acetate

Concen: 9.931 ug/l

RT: 7.565 min Scan# 91

Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010





Tgt Ion: 43 Resp: 29754

Ion Ratio Lower Upper

43 100

61 0.0 10.7 16.1#

70 10.2 8.8 13.2



#38

Carbon Tetrachloride

Concen: 9.665 ug/l

RT: 8.365 min Scan# 1090

Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion: 117 Resp: 30964


Ion Ratio Lower Upper

117 100

119 88.4 78.7 118.1

121 32.4 24.4 36.6





#39

Methylcyclohexane

Concen: 9.516 ug/l

RT: 9.600 min Scan# 11300

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010





Tgt Ion: 83 Resp: 30932

Ion Ratio Lower Upper

83 100

55 84.9 62.8 94.2

98 48.4 38.8 58.2



#40

Benzene

Concen: 9.454 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37


Tgt Ion: 78 Resp: 86029

Ion Ratio Lower Upper

78 100

77 22.8 19.1 28.7









#41
Methacrylonitrile
Concen: 9.426 ug/l
RT: 7.777 min Scan# 990
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:	41	Resp:	15332
Ion	Ratio	Lower	Upper
41	100		
39	55.8	43.3	64.9
67	73.0	56.1	84.1
52	30.4	24.5	36.7



#42
1,2-Dichloroethane
Concen: 9.855 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion:	62	Resp:	30402
Ion	Ratio	Lower	Upper
62	100		
98	8.1	0.0	18.2









#43
Isopropyl Acetate
Concen: 8.906 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:	43	Resp:	52246
Ion	Ratio	Lower	Upper
43	100		
61	25.1	20.4	30.6
87	12.3	10.3	15.5



#44
Trichloroethene
Concen: 9.454 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion:	130	Resp:	20092
Ion	Ratio	Lower	Upper
130	100		
95	102.9	0.0	203.0





#45

1,2-Dichloropropane

Concen: 9.618 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC010

Tgt Ion: 63 Resp: 20703

Ion Ratio Lower Upper

63 100

65 33.0 24.6 37.0



#46

Dibromomethane

Concen: 9.479 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion: 93 Resp: 13770


Ion Ratio Lower Upper

93 100

95 87.2 65.0 97.4

174 92.7 71.8 107.8









#47
Bromodichloromethane
Concen: 9.385 ug/l
RT: 9.888 min Scan# 1349
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 83 Resp: 30289
Ion Ratio Lower Upper
83 100
85 65.7 51.5 77.3
127 9.6 5.7 8.5#



#48
Methyl methacrylate
Concen: 9.679 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. 0.006 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 41 Resp: 22612
Ion Ratio Lower Upper
41 100
69 82.7 67.1 100.7
39 55.6 45.8 68.6









#49
1,4-Dioxane
Concen: 200.587 ug/l
RT: 9.694 min Scan# 1316
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 88 Resp: 8627
Ion Ratio Lower Upper
88 100
43 0.0 25.3 37.9#
58 76.0 56.4 84.6



#50
Toluene-d8
Concen: 9.463 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 98 Resp: 70000
Ion Ratio Lower Upper
98 100
100 66.0 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 50.007 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC010


Tgt Ion: 43 Resp: 142632
Ion Ratio Lower Upper
43 100
58 38.9 31.6 47.4



#52
Toluene
Concen: 9.468 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 92 Resp: 52537
Ion Ratio Lower Upper
92 100
91 170.1 137.3 205.9









#53
t-1,3-Dichloropropene
Concen: 9.592 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 75 Resp: 31555
Ion Ratio Lower Upper
75 100
77 35.8 25.4 38.0



#54
cis-1,3-Dichloropropene
Concen: 9.802 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 75 Resp: 34684
Ion Ratio Lower Upper
75 100
77 33.1 24.2 36.2
39 46.6 38.7 58.1





#55

1,1,2-Trichloroethane

Concen: 9.675 ug/l

RT: 11.018 min Scan# 1540

Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010

Tgt Ion: 97 Resp: 19245





Ion Ratio Lower Upper

97 100

83 85.7 71.0 106.4

85 53.5 41.7 62.5

99 59.9 50.1 75.1



#56

Ethyl methacrylate

Concen: 9.551 ug/l

RT: 10.870 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion: 69 Resp: 31262


Ion Ratio Lower Upper

69 100

41 68.2 53.3 79.9

39 38.1 28.8 43.2





#57

1,3-Dichloropropane

Concen: 9.457 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.006 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC010

Tgt Ion: 76 Resp: 33613

Ion Ratio Lower Upper

76 100

78 33.7 25.8 38.6



#58

2-Chloroethyl Vinyl ether

Concen: 46.784 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37


Tgt Ion: 63 Resp: 71035

Ion Ratio Lower Upper

63 100

106 27.9 21.8 32.6









#59
2-Hexanone
Concen: 48.604 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 43 Resp: 103615
Ion Ratio Lower Upper
43 100
58 54.2 27.3 81.8



#60
Dibromochloromethane
Concen: 9.715 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 129 Resp: 22833
Ion Ratio Lower Upper
129 100
127 75.9 38.3 114.9









#61
1,2-Dibromoethane
Concen: 9.299 ug/l
RT: 11.470 min Scan# 1617
Delta R.T. 0.006 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:107 Resp: 19220
Ion Ratio Lower Upper
107 100
109 98.6 75.4 113.0



#62
4-Bromofluorobenzene
Concen: 9.186 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 95 Resp: 24756
Ion Ratio Lower Upper
95 100
174 72.9 0.0 145.2
176 70.5 0.0 140.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:117 Resp: 262325
Ion Ratio Lower Upper
117 100
82 60.4 47.2 70.8
119 33.9 25.4 38.0



#64
Tetrachloroethene
Concen: 9.711 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion:164 Resp: 17743
Ion Ratio Lower Upper
164 100
166 124.4 101.0 151.4
129 98.0 80.8 121.2
131 94.2 78.1 117.1









#65
Chlorobenzene
Concen: 9.645 ug/l
RT: 11.894 min Scan# 11894
Delta R.T. 0.006 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:112 Resp: 56099
Ion Ratio Lower Upper
112 100
114 32.6 27.0 40.6



#66
1,1,1,2-Tetrachloroethane
Concen: 9.597 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion:131 Resp: 19227
Ion Ratio Lower Upper
131 100
133 100.5 47.9 143.7
119 64.6 32.8 98.3









#67
Ethyl Benzene
Concen: 9.385 ug/l
RT: 11.964 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 91 Resp: 96533
Ion Ratio Lower Upper
91 100
106 29.2 24.5 36.7



#68
m/p-Xylenes
Concen: 19.162 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion:106 Resp: 72910
Ion Ratio Lower Upper
106 100
91 208.7 167.1 250.7









#69
o-Xylene
Concen: 9.713 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:106 Resp: 34965
Ion Ratio Lower Upper
106 100
91 220.8 110.2 330.6



#70
Styrene
Concen: 9.572 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion:104 Resp: 58367
Ion Ratio Lower Upper
104 100
78 52.7 42.7 64.1
103 53.4 43.6 65.4





#71

Bromoform

Concen: 9.233 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion:173 Resp: 13618

Ion Ratio Lower Upper

173 100

175 49.9 23.5 70.6

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion:152 Resp: 119657


Ion Ratio Lower Upper

152 100

115 62.4 31.3 93.9

150 160.7 0.0 349.8





#73

Isopropylbenzene

Concen: 9.638 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC010

Tgt Ion:105 Resp: 88000

Ion Ratio Lower Upper

105 100

120 26.2 12.9 38.7



#74

N-ethyl acetate

Concen: 9.868 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Tgt Ion: 43 Resp: 40801




Ion Ratio Lower Upper


43 100

70 43.4 34.4 51.6

55 24.2 19.6 29.4

61 23.4 19.6 29.4





#75

1,1,2,2-Tetrachloroethane

Concen: 9.822 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion: 83 Resp: 26966

Ion Ratio Lower Upper

83 100

131 9.8 4.9 14.7

85 64.7 31.9 95.7



#76

1,2,3-Trichloropropane

Concen: 15.426 ug/l

RT: 12.994 min Scan# 1877

Delta R.T. 0.006 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37


Tgt Ion: 75 Resp: 36789

Ion Ratio Lower Upper

75 100

77 133.4 96.9 290.7









#77
Bromobenzene
Concen: 9.734 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:156 Resp: 21439
Ion Ratio Lower Upper
156 100
77 229.0 115.1 345.3
158 95.3 48.9 146.8



#78
n-propylbenzene
Concen: 9.957 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 91 Resp: 105334
Ion Ratio Lower Upper
91 100
120 22.1 10.8 32.4





#79

2-Chlorotoluene

Concen: 9.704 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC010

Tgt Ion: 91 Resp: 65511

Ion Ratio Lower Upper

91 100

126 32.0 15.7 47.1



#80

1,3,5-Trimethylbenzene

Concen: 9.887 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37


Tgt Ion: 105 Resp: 73637

Ion Ratio Lower Upper

105 100

120 47.5 23.7 71.1









#81
trans-1,4-Dichloro-2-butene
Concen: 10.047 ug/l
RT: 12.729 min Scan# 1832
Delta R.T. -0.006 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion: 75 Resp: 10237
Ion Ratio Lower Upper
75 100
53 89.4 78.0 117.0
89 44.7 35.5 53.3



#82
4-Chlorotoluene
Concen: 9.778 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.006 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion: 91 Resp: 66449
Ion Ratio Lower Upper
91 100
126 31.0 15.7 47.1





#83

tert-Butylbenzene

Concen: 9.691 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion:119 Resp: 64152

Ion Ratio Lower Upper

119 100

91 67.9 34.1 102.3

134 26.6 12.8 38.4



#84

1,2,4-Trimethylbenzene

Concen: 9.895 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. -0.000 min

Lab File: VN084216.D




Acq: 30 Sep 2024 13:37


Tgt Ion:105 Resp: 74245

Ion Ratio Lower Upper

105 100

120 45.2 22.3 66.8









#85
 sec-Butylbenzene
 Concen: 10.065 ug/l
 RT: 13.611 min Scan# 1982
 Delta R.T. -0.006 min
 Lab File: VN084216.D
 Acq: 30 Sep 2024 13:37




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC010


Tgt Ion:105 Resp: 88913
 Ion Ratio Lower Upper
 105 100
 134 19.4 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 9.812 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN084216.D
 Acq: 30 Sep 2024 13:37

Tgt Ion:119 Resp: 71675
 Ion Ratio Lower Upper
 119 100
 134 24.2 12.6 37.8
 91 24.7 12.4 37.2









#87
 1,3-Dichlorobenzene
 Concen: 9.536 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN084216.D
 Acq: 30 Sep 2024 13:37




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC010


Tgt Ion:146 Resp: 39402
 Ion Ratio Lower Upper
 146 100
 111 43.1 21.8 65.3
 148 65.7 31.9 95.7



#88
 1,4-Dichlorobenzene
 Concen: 9.392 ug/l
 RT: 13.811 min Scan# 2016
 Delta R.T. -0.000 min
 Lab File: VN084216.D
 Acq: 30 Sep 2024 13:37

Tgt Ion:146 Resp: 39191
 Ion Ratio Lower Upper
 146 100
 111 45.4 21.2 63.6
 148 65.5 31.9 95.9









#89
n-Butylbenzene
Concen: 9.273 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010


Tgt Ion:	91	Resp:	61422
Ion	Ratio	Lower	Upper
91	100		
92	53.4	26.6	79.7
134	23.2	12.3	36.9



#90
Hexachloroethane
Concen: 9.820 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. 0.006 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion:	117	Resp:	14560
Ion	Ratio	Lower	Upper
117	100		
201	68.1	34.4	103.2





#91

1,2-Dichlorobenzene

Concen: 9.526 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC010





Tgt Ion:146 Resp: 38594

Ion Ratio Lower Upper

146 100

111 43.1 22.0 66.0

148 63.1 32.1 96.3



#92

1,2-Dibromo-3-Chloropropane

Concen: 10.235 ug/l

RT: 14.717 min Scan# 2170

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion: 75 Resp: 5799


Ion Ratio Lower Upper

75 100

155 73.5 38.1 114.5

157 96.2 49.6 149.0





#93

1,2,4-Trichlorobenzene

Concen: 9.501 ug/l

RT: 15.388 min Scan# 21

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC010





Tgt Ion:180 Resp: 18960

Ion Ratio Lower Upper

180 100

182 93.4 47.0 141.0

145 31.7 16.4 49.4



#94

Hexachlorobutadiene

Concen: 9.638 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN084216.D

Acq: 30 Sep 2024 13:37




Tgt Ion:225 Resp: 9592


Ion Ratio Lower Upper

225 100

223 59.4 32.4 97.2

227 66.8 32.4 97.0









#95
Naphthalene
Concen: 9.613 ug/l
RT: 15.635 min Scan# 2326
Delta R.T. -0.000 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC010

Tgt Ion:128	Resp:	63596
Ion Ratio	Lower	Upper
128	100	
127	12.9	10.5 15.7
129	10.7	8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 9.826 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. 0.006 min
Lab File: VN084216.D
Acq: 30 Sep 2024 13:37

Tgt Ion:180	Resp:	19629
Ion Ratio	Lower	Upper
180	100	
182	93.5	48.2 144.6
145	36.4	17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084217.D
 Acq On : 30 Sep 2024 14:00
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC005

Quant Time: Oct 01 06:49:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	161547	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	288067	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	245710	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	112491	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	13262	5.537	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	11.080%#
35) Dibromofluoromethane	8.171	113	10333	5.441	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	10.880%#
50) Toluene-d8	10.565	98	35782	5.121	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	10.240%#
62) 4-Bromofluorobenzene	12.847	95	13132	5.159	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	10.320%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.130	85	9807	5.132	ug/l	94
3) Chloromethane	2.359	50	12064	5.529	ug/l	92
4) Vinyl Chloride	2.512	62	11701	5.536	ug/l	96
5) Bromomethane	2.947	94	8091	5.804	ug/l	97
6) Chloroethane	3.118	64	8433	5.571	ug/l	94
7) Trichlorofluoromethane	3.495	101	17988	5.466	ug/l	90
8) Diethyl Ether	3.953	74	6430	5.268	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.365	101	5482	2.907	ug/l #	33
10) Methyl Iodide	4.589	142	12344	5.083	ug/l #	97
11) Tert butyl alcohol	5.512	59	11537	29.211	ug/l #	91
12) 1,1-Dichloroethene	4.342	96	10191	5.603	ug/l	95
13) Acrolein	4.177	56	13122	28.940	ug/l	99
14) Allyl chloride	5.030	41	17724	5.611	ug/l	90
15) Acrylonitrile	5.718	53	26537	26.607	ug/l	99
16) Acetone	4.430	43	27153	26.389	ug/l	95
17) Carbon Disulfide	4.712	76	30816	5.351	ug/l	97
18) Methyl Acetate	5.024	43	12354	5.509	ug/l #	72
19) Methyl tert-butyl Ether	5.800	73	33101	5.392	ug/l	97
20) Methylene Chloride	5.265	84	10809	5.168	ug/l	96
21) trans-1,2-Dichloroethene	5.777	96	5005	2.626	ug/l	90
22) Diisopropyl ether	6.671	45	35654	5.457	ug/l	99
23) Vinyl Acetate	6.606	43	144069	29.570	ug/l	99
24) 1,1-Dichloroethane	6.565	63	19800	5.419	ug/l	95
25) 2-Butanone	7.482	43	37739	26.929	ug/l	99
26) 2,2-Dichloropropane	7.488	77	17552	5.327	ug/l	98
27) cis-1,2-Dichloroethene	7.488	96	12388	5.380	ug/l	99
28) Bromochloromethane	7.812	49	9996	6.131	ug/l #	95
29) Tetrahydrofuran	7.841	42	23456	27.121	ug/l	97
30) Chloroform	7.965	83	20341	5.359	ug/l	92
31) Cyclohexane	8.259	56	21337	6.014	ug/l	99
32) 1,1,1-Trichloroethane	8.171	97	18635	5.467	ug/l #	53
36) 1,1-Dichloropropene	8.371	75	14527	5.266	ug/l	99
37) Ethyl Acetate	7.565	43	16138	5.702	ug/l	97
38) Carbon Tetrachloride	8.359	117	15700	5.188	ug/l	95
39) Methylcyclohexane	9.594	83	15377	5.009	ug/l	94
40) Benzene	8.606	78	45343	5.276	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084217.D
 Acq On : 30 Sep 2024 14:00
 Operator : JC\MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC005

Quant Time: Oct 01 06:49:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.782	41	7878	5.128	ug/l	93
42) 1,2-Dichloroethane	8.677	62	15661	5.375	ug/l	99
43) Isopropyl Acetate	8.682	43	29779	5.374	ug/l	95
44) Trichloroethene	9.347	130	10921	5.440	ug/l	87
45) 1,2-Dichloropropane	9.618	63	11181	5.499	ug/l	95
46) Dibromomethane	9.706	93	7207	5.253	ug/l	93
47) Bromodichloromethane	9.888	83	16407	5.382	ug/l #	97
48) Methyl methacrylate	9.676	41	11766	5.332	ug/l	97
49) 1,4-Dioxane	9.688	88	4121	101.444	ug/l #	79
51) 4-Methyl-2-Pentanone	10.441	43	69730	25.883	ug/l	99
52) Toluene	10.629	92	26489	5.054	ug/l	96
53) t-1,3-Dichloropropene	10.835	75	16242	5.227	ug/l	94
54) cis-1,3-Dichloropropene	10.312	75	17448	5.220	ug/l	98
55) 1,1,2-Trichloroethane	11.018	97	9805	5.219	ug/l	89
56) Ethyl methacrylate	10.876	69	15794	5.109	ug/l	96
57) 1,3-Dichloropropane	11.159	76	17954	5.348	ug/l	96
58) 2-Chloroethyl Vinyl ether	10.159	63	38417	26.787	ug/l	99
59) 2-Hexanone	11.194	43	51488	25.570	ug/l	98
60) Dibromochloromethane	11.359	129	11412	5.141	ug/l	96
61) 1,2-Dibromoethane	11.470	107	10394	5.324	ug/l	98
64) Tetrachloroethene	11.106	164	9167	5.356	ug/l	90
65) Chlorobenzene	11.888	112	28818	5.290	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.959	131	10054	5.358	ug/l	97
67) Ethyl Benzene	11.965	91	49891	5.178	ug/l	97
68) m/p-Xylenes	12.065	106	35820	10.051	ug/l	98
69) o-Xylene	12.400	106	18041	5.350	ug/l	96
70) Styrene	12.412	104	28472	4.985	ug/l	99
71) Bromoform	12.582	173	7067	5.115	ug/l #	95
73) Isopropylbenzene	12.694	105	43468	5.064	ug/l	98
74) N-amyl acetate	12.494	43	19775	5.087	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.941	83	14517	5.624	ug/l	94
76) 1,2,3-Trichloropropane	12.994	75	19041	8.493	ug/l	59
77) Bromobenzene	12.976	156	11049	5.336	ug/l	99
78) n-propylbenzene	13.035	91	51120	5.140	ug/l	99
79) 2-Chlorotoluene	13.123	91	33161	5.225	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	36100	5.156	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.735	75	4595	4.797	ug/l #	79
82) 4-Chlorotoluene	13.223	91	33382	5.225	ug/l	99
83) tert-Butylbenzene	13.435	119	31990	5.140	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	36099	5.118	ug/l	100
85) sec-Butylbenzene	13.612	105	40954	4.931	ug/l	98
86) p-Isopropyltoluene	13.729	119	34083	4.963	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	20730	5.337	ug/l	98
88) 1,4-Dichlorobenzene	13.812	146	21249	5.417	ug/l	96
89) n-Butylbenzene	14.059	91	30253	4.858	ug/l	99
90) Hexachloroethane	14.329	117	7628	5.472	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	19905	5.226	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.723	75	3046	5.719	ug/l	92
93) 1,2,4-Trichlorobenzene	15.394	180	9439	5.031	ug/l	95
94) Hexachlorobutadiene	15.500	225	5174	5.530	ug/l	89
95) Naphthalene	15.641	128	30391	4.886	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	9245	4.923	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084217.D
Acq On : 30 Sep 2024 14:00
Operator : JC\MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005

Quant Time: Oct 01 06:49:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:168 Resp: 161547
Ion Ratio Lower Upper
168 100
99 66.3 54.2 81.2



#2
Dichlorodifluoromethane
Concen: 5.132 ug/l
RT: 2.130 min Scan# 30
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 85 Resp: 9807
Ion Ratio Lower Upper
85 100
87 31.8 17.6 52.9









#3
Chloromethane
Concen: 5.529 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 50 Resp: 12064
Ion Ratio Lower Upper
50 100
52 28.9 26.8 40.2



#4
Vinyl Chloride
Concen: 5.536 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 62 Resp: 11701
Ion Ratio Lower Upper
62 100
64 35.0 26.2 39.4









#5
Bromomethane
Concen: 5.804 ug/l
RT: 2.947 min Scan# 1
Delta R.T. -0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion: 94 Resp: 8091
Ion Ratio Lower Upper
94 100
96 88.4 72.9 109.3



#6
Chloroethane
Concen: 5.571 ug/l
RT: 3.118 min Scan# 198
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 64 Resp: 8433
Ion Ratio Lower Upper
64 100
66 32.8 23.5 35.3









#7
 Trichlorofluoromethane
 Concen: 5.466 ug/l
 RT: 3.495 min Scan# 210
 Delta R.T. -0.000 min
 Lab File: VN084217.D
 Acq: 30 Sep 2024 14:00




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion:101 Resp: 17988
 Ion Ratio Lower Upper
 101 100
 103 72.8 51.8 77.6



#8
 Diethyl Ether
 Concen: 5.268 ug/l
 RT: 3.953 min Scan# 340
 Delta R.T. -0.006 min
 Lab File: VN084217.D
 Acq: 30 Sep 2024 14:00

Tgt Ion: 74 Resp: 6430
 Ion Ratio Lower Upper
 74 100
 45 96.9 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 2.907 ug/l
RT: 4.365 min Scan# 411
Delta R.T. -0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:101 Resp: 5482
Ion Ratio Lower Upper
101 100
85 86.0 36.2 54.4#
151 133.2 59.3 88.9#



#10
Methyl Iodide
Concen: 5.083 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion:142 Resp: 12344
Ion Ratio Lower Upper
142 100
127 44.9 35.3 52.9
141 18.0 11.0 16.6#









#11
Tert butyl alcohol
Concen: 29.211 ug/l
RT: 5.512 min Scan# 607
Delta R.T. -0.012 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 59 Resp: 11537
Ion Ratio Lower Upper
59 100
57 7.0 8.2 12.2#



#12
1,1-Dichloroethene
Concen: 5.603 ug/l
RT: 4.342 min Scan# 406
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 96 Resp: 10191
Ion Ratio Lower Upper
96 100
61 164.1 128.6 192.8
98 74.5 51.7 77.5





#13

Acrolein

Concen: 28.940 ug/l

RT: 4.177 min Scan# 31

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC005

Tgt Ion: 56 Resp: 13122

Ion Ratio Lower Upper

56 100

55 70.5 57.2 85.8



#14

Allyl chloride

Concen: 5.611 ug/l

RT: 5.030 min Scan# 523

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00




Tgt Ion: 41 Resp: 17724


Ion Ratio Lower Upper

41 100

39 62.5 59.0 88.6

76 35.6 29.8 44.6









#15
Acrylonitrile
Concen: 26.607 ug/l
RT: 5.718 min Scan# 640
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion:	53	Resp:	26537
Ion	Ratio	Lower	Upper
53	100		
52	82.1	66.2	99.4
51	35.6	28.8	43.2



#16
Acetone
Concen: 26.389 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion:	43	Resp:	27153
Ion	Ratio	Lower	Upper
43	100		
58	27.9	24.4	36.6









#17
Carbon Disulfide
Concen: 5.351 ug/l
RT: 4.712 min Scan# 41
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 76 Resp: 30816
Ion Ratio Lower Upper
76 100
78 7.9 7.3 10.9



#18
Methyl Acetate
Concen: 5.509 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 43 Resp: 12354
Ion Ratio Lower Upper
43 100
74 9.9 19.2 28.8#









#19
Methyl tert-butyl Ether
Concen: 5.392 ug/l
RT: 5.800 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 73 Resp: 33101
Ion Ratio Lower Upper
73 100
57 23.3 17.7 26.5



#20
Methylene Chloride
Concen: 5.168 ug/l
RT: 5.265 min Scan# 563
Delta R.T. -0.012 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 84 Resp: 10809
Ion Ratio Lower Upper
84 100
49 127.8 107.2 160.8
51 38.2 31.8 47.8
86 63.0 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 2.626 ug/l
RT: 5.777 min Scan# 61
Delta R.T. -0.012 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 96 Resp: 5005
Ion Ratio Lower Upper
96 100
61 145.3 108.4 162.6
98 51.4 49.8 74.8



#22
Diisopropyl ether
Concen: 5.457 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 45 Resp: 35654
Ion Ratio Lower Upper
45 100
43 52.7 42.3 63.5
87 28.1 21.8 32.6
59 10.5 9.6 14.4









#23
 Vinyl Acetate
 Concen: 29.570 ug/l
 RT: 6.606 min Scan# 791
 Delta R.T. 0.006 min
 Lab File: VN084217.D
 Acq: 30 Sep 2024 14:00




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion: 43 Resp: 144069
 Ion Ratio Lower Upper
 43 100
 86 11.2 8.6 12.8



#24
 1,1-Dichloroethane
 Concen: 5.419 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. -0.006 min
 Lab File: VN084217.D
 Acq: 30 Sep 2024 14:00

Tgt Ion: 63 Resp: 19800
 Ion Ratio Lower Upper
 63 100
 98 8.9 3.4 10.2
 100 4.5 2.0 5.8









#25
2-Butanone
Concen: 26.929 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 37739
Ion Ratio Lower Upper
43 100
72 26.4 21.4 32.2



#26
2,2-Dichloropropane
Concen: 5.327 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 77 Resp: 17552
Ion Ratio Lower Upper
77 100
97 22.4 10.8 32.3









#27
 cis-1,2-Dichloroethene
 Concen: 5.380 ug/l
 RT: 7.488 min Scan# 941
 Delta R.T. -0.000 min
 Lab File: VN084217.D
 Acq: 30 Sep 2024 14:00

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005

Tgt Ion: 96 Resp: 12388




Ion	Ratio	Lower	Upper
96	100		
61	148.3	0.0	301.8
98	64.1	0.0	128.2




#28
 Bromochloromethane
 Concen: 6.131 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. -0.000 min
 Lab File: VN084217.D
 Acq: 30 Sep 2024 14:00

Tgt Ion: 49 Resp: 9996

Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	3.6
130	68.3	58.0	87.0





#29

Tetrahydrofuran

Concen: 27.121 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion: 42 Resp: 23456

Ion Ratio Lower Upper

42 100

72 40.3 34.6 51.8

71 40.1 32.2 48.4



#30

Chloroform

Concen: 5.359 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. -0.000 min

Lab File: VN084217.D




Acq: 30 Sep 2024 14:00


Tgt Ion: 83 Resp: 20341

Ion Ratio Lower Upper

83 100

85 70.0 51.0 76.4









#31
Cyclohexane
Concen: 6.014 ug/l
RT: 8.259 min Scan# 1071
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 56 Resp: 21337
Ion Ratio Lower Upper
56 100
69 29.5 23.8 35.6
84 85.5 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 5.467 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 97 Resp: 18635
Ion Ratio Lower Upper
97 100
99 0.0 51.2 76.8#
61 47.1 37.5 56.3





#33

1,2-Dichloroethane-d4

Concen: 5.537 ug/l

RT: 8.571 min Scan# 1125

Delta R.T. -0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC005

Tgt Ion: 65 Resp: 13262

Ion Ratio Lower Upper

65 100

67 51.7 0.0 102.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00




Tgt Ion: 114 Resp: 288067


Ion Ratio Lower Upper

114 100

63 21.9 0.0 43.8

88 15.7 0.0 31.6





#35

Dibromofluoromethane

Concen: 5.441 ug/l

RT: 8.171 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005





Tgt Ion:113 Resp: 10333

Ion Ratio Lower Upper

113 100

111 105.9 83.3 124.9

192 16.6 13.5 20.3



#36

1,1-Dichloropropene

Concen: 5.266 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00




Tgt Ion: 75 Resp: 14527


Ion Ratio Lower Upper

75 100

110 34.3 16.9 50.7

77 30.0 24.7 37.1





#37

Ethyl Acetate

Concen: 5.702 ug/l

RT: 7.565 min Scan# 91

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion: 43 Resp: 16138

Ion Ratio Lower Upper

43 100

61 12.7 10.7 16.1

70 9.0 8.8 13.2



#38

Carbon Tetrachloride

Concen: 5.188 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00




Tgt Ion: 117 Resp: 15700


Ion Ratio Lower Upper

117 100

119 93.4 78.7 118.1

121 32.3 24.4 36.6





#39

Methylcyclohexane

Concen: 5.009 ug/l

RT: 9.594 min Scan# 11299

Delta R.T. -0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005





Tgt Ion: 83 Resp: 15377

Ion Ratio Lower Upper

83 100

55 71.9 62.8 94.2

98 46.9 38.8 58.2



#40

Benzene

Concen: 5.276 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN084217.D




Acq: 30 Sep 2024 14:00


Tgt Ion: 78 Resp: 45343

Ion Ratio Lower Upper

78 100

77 23.3 19.1 28.7









#41
Methacrylonitrile
Concen: 5.128 ug/l
RT: 7.782 min Scan# 991
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion:	41	Resp:	7878
Ion	Ratio	Lower	Upper
41	100		
39	63.1	43.3	64.9
67	74.5	56.1	84.1
52	28.7	24.5	36.7



#42
1,2-Dichloroethane
Concen: 5.375 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion:	62	Resp:	15661
Ion	Ratio	Lower	Upper
62	100		
98	9.3	0.0	18.2









#43
Isopropyl Acetate
Concen: 5.374 ug/l
RT: 8.682 min Scan# 1144
Delta R.T. -0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion: 43 Resp: 29779
Ion Ratio Lower Upper
43 100
61 22.6 20.4 30.6
87 11.0 10.3 15.5



#44
Trichloroethene
Concen: 5.440 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 130 Resp: 10921
Ion Ratio Lower Upper
130 100
95 88.8 0.0 203.0









#45
1,2-Dichloropropane
Concen: 5.499 ug/l
RT: 9.618 min Scan# 11181
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 63 Resp: 11181
Ion Ratio Lower Upper
63 100
65 33.3 24.6 37.0



#46
Dibromomethane
Concen: 5.253 ug/l
RT: 9.706 min Scan# 1318
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 93 Resp: 7207
Ion Ratio Lower Upper
93 100
95 88.6 65.0 97.4
174 95.8 71.8 107.8









#47
Bromodichloromethane
Concen: 5.382 ug/l
RT: 9.888 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 83 Resp: 16407
Ion Ratio Lower Upper
83 100
85 66.6 51.5 77.3
127 8.5 5.7 8.5#



#48
Methyl methacrylate
Concen: 5.332 ug/l
RT: 9.676 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 41 Resp: 11766
Ion Ratio Lower Upper
41 100
69 81.4 67.1 100.7
39 60.1 45.8 68.6





#49

1,4-Dioxane

Concen: 101.444 ug/l

RT: 9.688 min Scan# 1

Delta R.T. -0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005





Tgt Ion: 88 Resp: 4121

Ion Ratio Lower Upper

88 100

43 0.0 25.3 37.9#

58 74.7 56.4 84.6



#50

Toluene-d8

Concen: 5.121 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN084217.D




Acq: 30 Sep 2024 14:00


Tgt Ion: 98 Resp: 35782

Ion Ratio Lower Upper

98 100

100 66.5 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 25.883 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 69730
Ion Ratio Lower Upper
43 100
58 39.0 31.6 47.4



#52
Toluene
Concen: 5.054 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 92 Resp: 26489
Ion Ratio Lower Upper
92 100
91 176.6 137.3 205.9









#53
t-1,3-Dichloropropene
Concen: 5.227 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion: 75 Resp: 16242
Ion Ratio Lower Upper
75 100
77 34.9 25.4 38.0



#54
cis-1,3-Dichloropropene
Concen: 5.220 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 75 Resp: 17448
Ion Ratio Lower Upper
75 100
77 32.2 24.2 36.2
39 47.8 38.7 58.1





#55

1,1,2-Trichloroethane

Concen: 5.219 ug/l

RT: 11.018 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005

Tgt Ion: 97 Resp: 9805





Ion Ratio Lower Upper

97 100

83 100.1 71.0 106.4

85 58.2 41.7 62.5

99 70.8 50.1 75.1



#56

Ethyl methacrylate

Concen: 5.109 ug/l

RT: 10.876 min Scan# 1517

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00




Tgt Ion: 69 Resp: 15794


Ion Ratio Lower Upper

69 100

41 65.1 53.3 79.9

39 31.7 28.8 43.2









#57
1,3-Dichloropropane
Concen: 5.348 ug/l
RT: 11.159 min Scan# 111
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion: 76 Resp: 17954
Ion Ratio Lower Upper
76 100
78 29.8 25.8 38.6



#58
2-Chloroethyl Vinyl ether
Concen: 26.787 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 63 Resp: 38417
Ion Ratio Lower Upper
63 100
106 27.8 21.8 32.6









#59
2-Hexanone
Concen: 25.570 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 43 Resp: 51488
Ion Ratio Lower Upper
43 100
58 52.7 27.3 81.8



#60
Dibromochloromethane
Concen: 5.141 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 129 Resp: 11412
Ion Ratio Lower Upper
129 100
127 80.4 38.3 114.9









#61
1,2-Dibromoethane
Concen: 5.324 ug/l
RT: 11.470 min Scan# 1617
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:107 Resp: 10394
Ion Ratio Lower Upper
107 100
109 92.6 75.4 113.0



#62
4-Bromofluorobenzene
Concen: 5.159 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 95 Resp: 13132
Ion Ratio Lower Upper
95 100
174 70.1 0.0 145.2
176 64.1 0.0 140.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:117 Resp: 245710
Ion Ratio Lower Upper
117 100
82 60.4 47.2 70.8
119 32.4 25.4 38.0



#64
Tetrachloroethene
Concen: 5.356 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion:164 Resp: 9167
Ion Ratio Lower Upper
164 100
166 109.1 101.0 151.4
129 90.7 80.8 121.2
131 94.5 78.1 117.1









#65
Chlorobenzene
Concen: 5.290 ug/l
RT: 11.888 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:112 Resp: 28818
Ion Ratio Lower Upper
112 100
114 33.0 27.0 40.6



#66
1,1,1,2-Tetrachloroethane
Concen: 5.358 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion:131 Resp: 10054
Ion Ratio Lower Upper
131 100
133 91.7 47.9 143.7
119 64.8 32.8 98.3









#67
Ethyl Benzene
Concen: 5.178 ug/l
RT: 11.965 min Scan# 1719
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 91 Resp: 49891
Ion Ratio Lower Upper
91 100
106 28.7 24.5 36.7



#68
m/p-Xylenes
Concen: 10.051 ug/l
RT: 12.065 min Scan# 1719
Delta R.T. -0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion:106 Resp: 35820
Ion Ratio Lower Upper
106 100
91 212.5 167.1 250.7









#69
o-Xylene
Concen: 5.350 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC005


Tgt Ion:106 Resp: 18041
Ion Ratio Lower Upper
106 100
91 214.1 110.2 330.6



#70
Styrene
Concen: 4.985 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion:104 Resp: 28472
Ion Ratio Lower Upper
104 100
78 52.7 42.7 64.1
103 54.5 43.6 65.4





#71

Bromoform

Concen: 5.115 ug/l

RT: 12.582 min Scan# 1807

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005





Tgt Ion:173 Resp: 7067

Ion Ratio Lower Upper

173 100

175 50.2 23.5 70.6

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00




Tgt Ion:152 Resp: 112491


Ion Ratio Lower Upper

152 100

115 62.1 31.3 93.9

150 158.0 0.0 349.8





#73

Isopropylbenzene

Concen: 5.064 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC005

Tgt Ion:105 Resp: 43468

Ion Ratio Lower Upper

105 100

120 26.7 12.9 38.7



#74

N-ethyl acetate

Concen: 5.087 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Tgt Ion: 43 Resp: 19775




Ion Ratio Lower Upper


43 100

70 43.5 34.4 51.6

55 26.3 19.6 29.4

61 24.8 19.6 29.4









#75
 1,1,2,2-Tetrachloroethane
 Concen: 5.624 ug/l
 RT: 12.941 min Scan# 1868
 Delta R.T. 0.006 min
 Lab File: VN084217.D
 Acq: 30 Sep 2024 14:00




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion: 83 Resp: 14517
 Ion Ratio Lower Upper
 83 100
 131 9.2 4.9 14.7
 85 59.0 31.9 95.7



#76
 1,2,3-Trichloropropane
 Concen: 8.493 ug/l
 RT: 12.994 min Scan# 1877
 Delta R.T. 0.006 min
 Lab File: VN084217.D
 Acq: 30 Sep 2024 14:00

Tgt Ion: 75 Resp: 19041
 Ion Ratio Lower Upper
 75 100
 77 133.3 96.9 290.7





#77

Bromobenzene

Concen: 5.336 ug/l

RT: 12.976 min Scan# 11

Delta R.T. -0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:156 Resp: 11049

Ion Ratio Lower Upper

156 100

77 229.7 115.1 345.3

158 96.2 48.9 146.8



#78

n-propylbenzene

Concen: 5.140 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. -0.000 min

Lab File: VN084217.D




Acq: 30 Sep 2024 14:00


Tgt Ion: 91 Resp: 51120

Ion Ratio Lower Upper

91 100

120 22.3 10.8 32.4









#79
2-Chlorotoluene
Concen: 5.225 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 91 Resp: 33161
Ion Ratio Lower Upper
91 100
126 31.0 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 5.156 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 105 Resp: 36100
Ion Ratio Lower Upper
105 100
120 49.2 23.7 71.1









#81
trans-1,4-Dichloro-2-butene
Concen: 4.797 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 75 Resp: 4595
Ion Ratio Lower Upper
75 100
53 120.1 78.0 117.0#
89 34.5 35.5 53.3#



#82
4-Chlorotoluene
Concen: 5.225 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 91 Resp: 33382
Ion Ratio Lower Upper
91 100
126 32.1 15.7 47.1





#83

tert-Butylbenzene

Concen: 5.140 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC005





Tgt Ion:119 Resp: 31990

Ion Ratio Lower Upper

119 100

91 68.5 34.1 102.3

134 26.4 12.8 38.4



#84

1,2,4-Trimethylbenzene

Concen: 5.118 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. -0.000 min

Lab File: VN084217.D




Acq: 30 Sep 2024 14:00


Tgt Ion:105 Resp: 36099

Ion Ratio Lower Upper

105 100

120 44.2 22.3 66.8









#85
 sec-Butylbenzene
 Concen: 4.931 ug/l
 RT: 13.612 min Scan# 1982
 Delta R.T. -0.006 min
 Lab File: VN084217.D
 Acq: 30 Sep 2024 14:00




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion:105 Resp: 40954
 Ion Ratio Lower Upper
 105 100
 134 20.6 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 4.963 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN084217.D
 Acq: 30 Sep 2024 14:00

Tgt Ion:119 Resp: 34083
 Ion Ratio Lower Upper
 119 100
 134 25.2 12.6 37.8
 91 26.0 12.4 37.2









#87
 1,3-Dichlorobenzene
 Concen: 5.337 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN084217.D
 Acq: 30 Sep 2024 14:00




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC005


Tgt Ion:146 Resp: 20730
 Ion Ratio Lower Upper
 146 100
 111 45.0 21.8 65.3
 148 64.7 31.9 95.7



#88
 1,4-Dichlorobenzene
 Concen: 5.417 ug/l
 RT: 13.812 min Scan# 2016
 Delta R.T. -0.000 min
 Lab File: VN084217.D
 Acq: 30 Sep 2024 14:00

Tgt Ion:146 Resp: 21249
 Ion Ratio Lower Upper
 146 100
 111 47.9 21.2 63.6
 148 64.5 31.9 95.9









#89
n-Butylbenzene
Concen: 4.858 ug/l
RT: 14.059 min Scan# 2057
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion: 91 Resp: 30253
Ion Ratio Lower Upper
91 100
92 53.8 26.6 79.7
134 24.9 12.3 36.9



#90
Hexachloroethane
Concen: 5.472 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 117 Resp: 7628
Ion Ratio Lower Upper
117 100
201 67.4 34.4 103.2









#91
1,2-Dichlorobenzene
Concen: 5.226 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:146 Resp: 19905
Ion Ratio Lower Upper
146 100
111 44.8 22.0 66.0
148 66.8 32.1 96.3



#92
1,2-Dibromo-3-Chloropropane
Concen: 5.719 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion: 75 Resp: 3046
Ion Ratio Lower Upper
75 100
155 65.1 38.1 114.5
157 95.3 49.6 149.0









#93
1,2,4-Trichlorobenzene
Concen: 5.031 ug/l
RT: 15.394 min Scan# 21
Delta R.T. 0.006 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC005


Tgt Ion:180 Resp: 9439
Ion Ratio Lower Upper
180 100
182 97.8 47.0 141.0
145 37.6 16.4 49.4



#94
Hexachlorobutadiene
Concen: 5.530 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. -0.000 min
Lab File: VN084217.D
Acq: 30 Sep 2024 14:00

Tgt Ion:225 Resp: 5174
Ion Ratio Lower Upper
225 100
223 56.0 32.4 97.2
227 56.4 32.4 97.0





#95

Naphthalene

Concen: 4.886 ug/l

RT: 15.641 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC005





Tgt Ion:128 Resp: 30391

Ion Ratio Lower Upper

128 100

127 13.2 10.5 15.7

129 10.4 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 4.923 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN084217.D

Acq: 30 Sep 2024 14:00




Tgt Ion:180 Resp: 9245

Ion Ratio Lower Upper

180 100

182 94.9 48.2 144.6

145 36.0 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084218.D
 Acq On : 30 Sep 2024 14:48
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Quant Time: Oct 01 06:49:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	160573	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	290146	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	244376	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	101963	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	2443	1.026	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	2.060%#
35) Dibromofluoromethane	8.165	113	2114	1.105	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	2.220%#
50) Toluene-d8	10.565	98	7435	1.056	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	2.120%#
62) 4-Bromofluorobenzene	12.853	95	2451	0.956	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	1.920%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.124	85	1928	1.015	ug/l	83
3) Chloromethane	2.365	50	2217	1.022	ug/l	99
4) Vinyl Chloride	2.512	62	1981	0.943	ug/l	89
5) Bromomethane	2.965	94	1910	1.378	ug/l	93
6) Chloroethane	3.124	64	2029	1.349	ug/l	97
7) Trichlorofluoromethane	3.500	101	3140	0.960	ug/l	94
8) Diethyl Ether	3.965	74	1167	0.962	ug/l	88
9) 1,1,2-Trichlorotrifluo...	4.371	101	1934	1.032	ug/l #	63
10) Methyl Iodide	4.583	142	2446	1.013	ug/l #	80
11) Tert butyl alcohol	5.524	59	1919	4.888	ug/l #	73
12) 1,1-Dichloroethene	4.353	96	1584	0.876	ug/l #	50
13) Acrolein	4.183	56	2550	5.658	ug/l	87
14) Allyl chloride	5.018	41	2333	0.743	ug/l	89
15) Acrylonitrile	5.736	53	4920	4.963	ug/l #	81
16) Acetone	4.436	43	4799	4.692	ug/l	90
17) Carbon Disulfide	4.718	76	6680	1.167	ug/l	100
18) Methyl Acetate	5.030	43	2249	1.009	ug/l #	64
19) Methyl tert-butyl Ether	5.806	73	5319	0.872	ug/l #	83
20) Methylene Chloride	5.277	84	2203	1.060	ug/l #	83
21) trans-1,2-Dichloroethene	5.783	96	1753	0.925	ug/l	86
22) Diisopropyl ether	6.671	45	5639	0.868	ug/l #	96
23) Vinyl Acetate	6.606	43	22441	4.634	ug/l #	92
24) 1,1-Dichloroethane	6.571	63	3358	0.925	ug/l #	83
25) 2-Butanone	7.488	43	6488	4.658	ug/l	91
26) 2,2-Dichloropropane	7.494	77	2902	0.886	ug/l	96
27) cis-1,2-Dichloroethene	7.483	96	2259	0.987	ug/l	92
28) Bromochloromethane	7.824	49	1718	1.060	ug/l #	87
29) Tetrahydrofuran	7.841	42	4006	4.660	ug/l	92
30) Chloroform	7.971	83	3793	1.005	ug/l	99
31) Cyclohexane	8.241	56	6798	1.928	ug/l #	59
32) 1,1,1-Trichloroethane	8.171	97	3120	0.921	ug/l #	53
36) 1,1-Dichloropropene	8.365	75	2420	0.871	ug/l	98
37) Ethyl Acetate	7.559	43	2596	0.911	ug/l #	78
38) Carbon Tetrachloride	8.359	117	2795	0.917	ug/l	90
39) Methylcyclohexane	9.600	83	2486	0.804	ug/l #	88
40) Benzene	8.606	78	8245	0.952	ug/l	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084218.D
 Acq On : 30 Sep 2024 14:48
 Operator : JC\MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC001

Quant Time: Oct 01 06:49:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 06:46:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.782	41	1490	0.963	ug/l #	66
42) 1,2-Dichloroethane	8.671	62	2669	0.909	ug/l	88
43) Isopropyl Acetate	8.688	43	7949	1.424	ug/l #	80
44) Trichloroethene	9.347	130	1885	0.932	ug/l	87
45) 1,2-Dichloropropane	9.624	63	1676	0.818	ug/l #	44
46) Dibromomethane	9.706	93	1144	0.828	ug/l	87
47) Bromodichloromethane	9.894	83	2754	0.897	ug/l #	93
48) Methyl methacrylate	9.676	41	1824	0.821	ug/l	93
49) 1,4-Dioxane	9.694	88	800	19.552	ug/l #	78
51) 4-Methyl-2-Pentanone	10.447	43	11218	4.134	ug/l	97
52) Toluene	10.629	92	4875	0.923	ug/l	91
53) t-1,3-Dichloropropene	10.829	75	2493	0.797	ug/l	98
54) cis-1,3-Dichloropropene	10.306	75	2720	0.808	ug/l	95
55) 1,1,2-Trichloroethane	11.018	97	1669	0.882	ug/l	90
56) Ethyl methacrylate	10.876	69	2402	0.771	ug/l #	77
57) 1,3-Dichloropropane	11.159	76	3092	0.914	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	6179	4.278	ug/l	98
59) 2-Hexanone	11.194	43	8282	4.084	ug/l	96
60) Dibromochloromethane	11.353	129	1916	0.857	ug/l	91
61) 1,2-Dibromoethane	11.465	107	1795	0.913	ug/l	99
64) Tetrachloroethene	11.106	164	1690	0.993	ug/l #	87
65) Chlorobenzene	11.888	112	5025	0.927	ug/l #	86
66) 1,1,1,2-Tetrachloroethane	11.959	131	1632	0.874	ug/l #	66
67) Ethyl Benzene	11.965	91	8582	0.896	ug/l #	88
68) m/p-Xylenes	12.070	106	5869	1.656	ug/l	95
69) o-Xylene	12.400	106	2398	0.715	ug/l	83
70) Styrene	12.412	104	4489	0.790	ug/l	94
71) Bromoform	12.576	173	1168	0.850	ug/l #	94
73) Isopropylbenzene	12.688	105	6991	0.899	ug/l	97
74) N-amyl acetate	12.494	43	3228	0.916	ug/l #	88
75) 1,1,2,2-Tetrachloroethane	12.941	83	2233	0.954	ug/l #	95
76) 1,2,3-Trichloropropane	12.994	75	3395	1.671	ug/l	53
77) Bromobenzene	12.976	156	1818	0.969	ug/l	98
78) n-propylbenzene	13.035	91	7045	0.781	ug/l	95
79) 2-Chlorotoluene	13.129	91	5383	0.936	ug/l	92
80) 1,3,5-Trimethylbenzene	13.170	105	4830	0.761	ug/l	95
81) trans-1,4-Dichloro-2-b...	12.735	75	660	0.760	ug/l #	94
82) 4-Chlorotoluene	13.223	91	5174	0.893	ug/l	100
83) tert-Butylbenzene	13.435	119	4909	0.870	ug/l	88
84) 1,2,4-Trimethylbenzene	13.482	105	4760	0.744	ug/l	100
85) sec-Butylbenzene	13.612	105	5795	0.770	ug/l	99
86) p-Isopropyltoluene	13.723	119	4495	0.722	ug/l	98
87) 1,3-Dichlorobenzene	13.729	146	3424	0.972	ug/l	96
88) 1,4-Dichlorobenzene	13.729	146	3424	0.963	ug/l	96
89) n-Butylbenzene	14.053	91	4866	0.862	ug/l	97
90) Hexachloroethane	14.335	117	1151	0.911	ug/l	82
91) 1,2-Dichlorobenzene	14.100	146	3610	1.046	ug/l	92
92) 1,2-Dibromo-3-Chloropr...	14.711	75	423	0.876	ug/l	65
93) 1,2,4-Trichlorobenzene	15.400	180	1340	0.788	ug/l	89
94) Hexachlorobutadiene	15.494	225	861	1.015	ug/l	90
95) Naphthalene	15.641	128	5031	0.892	ug/l	97
96) 1,2,3-Trichlorobenzene	15.847	180	1531	0.899	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084218.D
Acq On : 30 Sep 2024 14:48
Operator : JC\MD
Sample : VSTDICC001
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Quant Time: Oct 01 06:49:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 01 06:49:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 06:46:32 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion:168 Resp: 160573
Ion Ratio Lower Upper
168 100
99 68.1 54.2 81.2



#2
Dichlorodifluoromethane
Concen: 1.015 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 85 Resp: 1928
Ion Ratio Lower Upper
85 100
87 25.3 17.6 52.9









#3
Chloromethane
Concen: 1.022 ug/l
RT: 2.365 min Scan# 70
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 50 Resp: 2217
Ion Ratio Lower Upper
50 100
52 34.2 26.8 40.2



#4
Vinyl Chloride
Concen: 0.943 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 62 Resp: 1981
Ion Ratio Lower Upper
62 100
64 39.0 26.2 39.4









#5
Bromomethane
Concen: 1.378 ug/l
RT: 2.965 min Scan# 11
Delta R.T. 0.012 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 94 Resp: 1910
Ion Ratio Lower Upper
94 100
96 84.7 72.9 109.3



#6
Chloroethane
Concen: 1.349 ug/l
RT: 3.124 min Scan# 199
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 64 Resp: 2029
Ion Ratio Lower Upper
64 100
66 31.0 23.5 35.3









#7
 Trichlorofluoromethane
 Concen: 0.960 ug/l
 RT: 3.500 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion:101 Resp: 3140
 Ion Ratio Lower Upper
 101 100
 103 60.0 51.8 77.6



#8
 Diethyl Ether
 Concen: 0.962 ug/l
 RT: 3.965 min Scan# 342
 Delta R.T. 0.006 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48

Tgt Ion: 74 Resp: 1167
 Ion Ratio Lower Upper
 74 100
 45 83.9 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 1.032 ug/l
RT: 4.371 min Scan# 411
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:101 Resp: 1934
Ion Ratio Lower Upper
101 100
85 25.4 36.2 54.4#
151 38.8 59.3 88.9#



#10
Methyl Iodide
Concen: 1.013 ug/l
RT: 4.583 min Scan# 447
Delta R.T. -0.012 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion:142 Resp: 2446
Ion Ratio Lower Upper
142 100
127 33.8 35.3 52.9#
141 0.0 11.0 16.6#









#11
Tert butyl alcohol
Concen: 4.888 ug/l
RT: 5.524 min Scan# 607
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 59 Resp: 1919
Ion Ratio Lower Upper
59 100
57 0.0 8.2 12.2#



#12
1,1-Dichloroethene
Concen: 0.876 ug/l
RT: 4.353 min Scan# 408
Delta R.T. 0.012 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 96 Resp: 1584
Ion Ratio Lower Upper
96 100
61 88.5 128.6 192.8#
98 34.3 51.7 77.5#





#13

Acrolein

Concen: 5.658 ug/l

RT: 4.183 min Scan# 31

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC001

Tgt Ion: 56 Resp: 2550

Ion Ratio Lower Upper

56 100

55 60.3 57.2 85.8



#14

Allyl chloride

Concen: 0.743 ug/l

RT: 5.018 min Scan# 521

Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion: 41 Resp: 2333


Ion Ratio Lower Upper

41 100

39 83.1 59.0 88.6

76 29.9 29.8 44.6









#15
Acrylonitrile
Concen: 4.963 ug/l
RT: 5.736 min Scan# 641
Delta R.T. 0.018 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 53 Resp: 4920
Ion Ratio Lower Upper
53 100
52 82.8 66.2 99.4
51 0.0 28.8 43.2#



#16
Acetone
Concen: 4.692 ug/l
RT: 4.436 min Scan# 422
Delta R.T. 0.012 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 43 Resp: 4799
Ion Ratio Lower Upper
43 100
58 24.9 24.4 36.6









#17
Carbon Disulfide
Concen: 1.167 ug/l
RT: 4.718 min Scan# 41
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 76 Resp: 6680
Ion Ratio Lower Upper
76 100
78 9.1 7.3 10.9



#18
Methyl Acetate
Concen: 1.009 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.012 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 43 Resp: 2249
Ion Ratio Lower Upper
43 100
74 6.4 19.2 28.8#





#19

Methyl tert-butyl Ether

Concen: 0.872 ug/l

RT: 5.806 min Scan# 61

Delta R.T. 0.012 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion: 73 Resp: 5319

Ion Ratio Lower Upper

73 100

57 13.9 17.7 26.5#



#20

Methylene Chloride

Concen: 1.060 ug/l

RT: 5.277 min Scan# 565

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion: 84 Resp: 2203




Ion Ratio Lower Upper


84 100

49 100.6 107.2 160.8#

51 43.3 31.8 47.8

86 64.5 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 0.925 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 96 Resp: 1753
Ion Ratio Lower Upper
96 100
61 115.2 108.4 162.6
98 55.5 49.8 74.8



#22
Diisopropyl ether
Concen: 0.868 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 45 Resp: 5639
Ion Ratio Lower Upper
45 100
43 52.8 42.3 63.5
87 22.2 21.8 32.6
59 16.4 9.6 14.4#









#23
 Vinyl Acetate
 Concen: 4.634 ug/l
 RT: 6.606 min Scan# 791
 Delta R.T. 0.006 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion: 43 Resp: 22441
 Ion Ratio Lower Upper
 43 100
 86 13.5 8.6 12.8#



#24
 1,1-Dichloroethane
 Concen: 0.925 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. -0.000 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48

Tgt Ion: 63 Resp: 3358
 Ion Ratio Lower Upper
 63 100
 98 0.0 3.4 10.2#
 100 0.0 2.0 5.8#





#25

2-Butanone

Concen: 4.658 ug/l

RT: 7.488 min Scan# 941

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion: 43 Resp: 6488

Ion Ratio Lower Upper

43 100

72 22.1 21.4 32.2



#26

2,2-Dichloropropane

Concen: 0.886 ug/l

RT: 7.494 min Scan# 942

Delta R.T. 0.006 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion: 77 Resp: 2902

Ion Ratio Lower Upper

77 100

97 19.7 10.8 32.3









#27
 cis-1,2-Dichloroethene
 Concen: 0.987 ug/l
 RT: 7.483 min Scan# 941
 Delta R.T. -0.006 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion: 96 Resp: 2259
 Ion Ratio Lower Upper
 96 100
 61 140.3 0.0 301.8
 98 59.5 0.0 128.2



#28
 Bromochloromethane
 Concen: 1.060 ug/l
 RT: 7.824 min Scan# 998
 Delta R.T. 0.012 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48

Tgt Ion: 49 Resp: 1718
 Ion Ratio Lower Upper
 49 100
 129 0.0 0.0 3.6
 130 61.3 58.0 87.0





#29

Tetrahydrofuran

Concen: 4.660 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion: 42 Resp: 4006

Ion Ratio Lower Upper

42 100

72 40.5 34.6 51.8

71 33.4 32.2 48.4



#30

Chloroform

Concen: 1.005 ug/l

RT: 7.971 min Scan# 1023

Delta R.T. 0.006 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion: 83 Resp: 3793

Ion Ratio Lower Upper

83 100

85 62.8 51.0 76.4









#31
Cyclohexane
Concen: 1.928 ug/l
RT: 8.241 min Scan# 1069
Delta R.T. -0.012 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 56 Resp: 6798
Ion Ratio Lower Upper
56 100
69 87.4 23.8 35.6#
84 103.0 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 0.921 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 97 Resp: 3120
Ion Ratio Lower Upper
97 100
99 0.0 51.2 76.8#
61 46.5 37.5 56.3









#33
 1,2-Dichloroethane-d4
 Concen: 1.026 ug/l
 RT: 8.577 min Scan# 1126
 Delta R.T. -0.000 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion: 65 Resp: 2443
 Ion Ratio Lower Upper
 65 100
 67 54.6 0.0 102.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 9.100 min Scan# 1215
 Delta R.T. -0.000 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48

Tgt Ion: 114 Resp: 290146
 Ion Ratio Lower Upper
 114 100
 63 22.3 0.0 43.8
 88 16.9 0.0 31.6





#35

Dibromofluoromethane

Concen: 1.105 ug/l

RT: 8.165 min Scan# 1090

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:113 Resp: 2114

Ion Ratio Lower Upper

113 100

111 106.9 83.3 124.9

192 7.2 13.5 20.3#



#36

1,1-Dichloropropene

Concen: 0.871 ug/l

RT: 8.365 min Scan# 1090

Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion: 75 Resp: 2420


Ion Ratio Lower Upper

75 100

110 32.6 16.9 50.7

77 31.9 24.7 37.1





#37

Ethyl Acetate

Concen: 0.911 ug/l

RT: 7.559 min Scan# 91

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001





Tgt Ion: 43 Resp: 2596

Ion Ratio Lower Upper

43 100

61 4.8 10.7 16.1#

70 2.4 8.8 13.2#



#38

Carbon Tetrachloride

Concen: 0.917 ug/l

RT: 8.359 min Scan# 1089

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion: 117 Resp: 2795


Ion Ratio Lower Upper

117 100

119 86.1 78.7 118.1

121 31.6 24.4 36.6





#39

Methylcyclohexane

Concen: 0.804 ug/l

RT: 9.600 min Scan# 11300

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion: 83 Resp: 2486

Ion Ratio Lower Upper

83 100

55 87.2 62.8 94.2

98 59.0 38.8 58.2#



#40

Benzene

Concen: 0.952 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion: 78 Resp: 8245

Ion Ratio Lower Upper

78 100

77 21.1 19.1 28.7









#41
Methacrylonitrile
Concen: 0.963 ug/l
RT: 7.782 min Scan# 991
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:	41	Resp:	1490
Ion	Ratio	Lower	Upper
41	100		
39	21.5	43.3	64.9#
67	56.9	56.1	84.1
52	0.0	24.5	36.7#



#42
1,2-Dichloroethane
Concen: 0.909 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion:	62	Resp:	2669
Ion	Ratio	Lower	Upper
62	100		
98	4.9	0.0	18.2





#43

Isopropyl Acetate

Concen: 1.424 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion: 43 Resp: 7949

Ion Ratio Lower Upper

43 100

61 13.7 20.4 30.6#

87 7.0 10.3 15.5#



#44

Trichloroethene

Concen: 0.932 ug/l

RT: 9.347 min Scan# 1257

Delta R.T. -0.000 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion:130 Resp: 1885

Ion Ratio Lower Upper

130 100

95 87.9 0.0 203.0





#45

1,2-Dichloropropane

Concen: 0.818 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :





VSTDICC001

Tgt Ion: 63 Resp: 1676

Ion Ratio Lower Upper

63 100

65 0.0 24.6 37.0#



#46

Dibromomethane

Concen: 0.828 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion: 93 Resp: 1144


Ion Ratio Lower Upper

93 100

95 95.1 65.0 97.4

174 100.3 71.8 107.8









#47
Bromodichloromethane
Concen: 0.897 ug/l
RT: 9.894 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 83 Resp: 2754
Ion Ratio Lower Upper
83 100
85 68.7 51.5 77.3
127 0.0 5.7 8.5#



#48
Methyl methacrylate
Concen: 0.821 ug/l
RT: 9.676 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 41 Resp: 1824
Ion Ratio Lower Upper
41 100
69 86.5 67.1 100.7
39 66.4 45.8 68.6





#49

1,4-Dioxane

Concen: 19.552 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001





Tgt Ion: 88 Resp: 800

Ion Ratio Lower Upper

88 100

43 0.0 25.3 37.9#

58 75.9 56.4 84.6



#50

Toluene-d8

Concen: 1.056 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion: 98 Resp: 7435

Ion Ratio Lower Upper

98 100

100 66.7 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 4.134 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 43 Resp: 11218
Ion Ratio Lower Upper
43 100
58 37.4 31.6 47.4



#52
Toluene
Concen: 0.923 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 92 Resp: 4875
Ion Ratio Lower Upper
92 100
91 159.8 137.3 205.9









#53
t-1,3-Dichloropropene
Concen: 0.797 ug/l
RT: 10.829 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 75 Resp: 2493
Ion Ratio Lower Upper
75 100
77 32.7 25.4 38.0



#54
cis-1,3-Dichloropropene
Concen: 0.808 ug/l
RT: 10.306 min Scan# 1420
Delta R.T. -0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 75 Resp: 2720
Ion Ratio Lower Upper
75 100
77 34.6 24.2 36.2
39 46.3 38.7 58.1





#55

1,1,2-Trichloroethane

Concen: 0.882 ug/l

RT: 11.018 min Scan# 1540

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001

Tgt Ion: 97 Resp: 1669





Ion Ratio Lower Upper

97 100

83 75.7 71.0 106.4

85 57.4 41.7 62.5

99 68.1 50.1 75.1



#56

Ethyl methacrylate

Concen: 0.771 ug/l

RT: 10.876 min Scan# 1517

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion: 69 Resp: 2402


Ion Ratio Lower Upper

69 100

41 64.2 53.3 79.9

39 0.0 28.8 43.2#





#57

1,3-Dichloropropane

Concen: 0.914 ug/l

RT: 11.159 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion: 76 Resp: 3092

Ion Ratio Lower Upper

76 100

78 32.2 25.8 38.6



#58

2-Chloroethyl Vinyl ether

Concen: 4.278 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion: 63 Resp: 6179

Ion Ratio Lower Upper

63 100

106 28.2 21.8 32.6









#59
2-Hexanone
Concen: 4.084 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 43 Resp: 8282
Ion Ratio Lower Upper
43 100
58 51.4 27.3 81.8



#60
Dibromochloromethane
Concen: 0.857 ug/l
RT: 11.353 min Scan# 1598
Delta R.T. -0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 129 Resp: 1916
Ion Ratio Lower Upper
129 100
127 69.2 38.3 114.9









#61
1,2-Dibromoethane
Concen: 0.913 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:107 Resp: 1795
Ion Ratio Lower Upper
107 100
109 94.9 75.4 113.0



#62
4-Bromofluorobenzene
Concen: 0.956 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 95 Resp: 2451
Ion Ratio Lower Upper
95 100
174 71.7 0.0 145.2
176 69.2 0.0 140.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:117 Resp: 244376

Ion Ratio Lower Upper

117 100

82 60.8 47.2 70.8

119 34.0 25.4 38.0



#64

Tetrachloroethene

Concen: 0.993 ug/l

RT: 11.106 min Scan# 1556

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion:164 Resp: 1690




Ion Ratio Lower Upper


164 100

166 92.5 101.0 151.4#

129 103.5 80.8 121.2

131 95.1 78.1 117.1









#65
Chlorobenzene
Concen: 0.927 ug/l
RT: 11.888 min Scan# 111
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:112 Resp: 5025
Ion Ratio Lower Upper
112 100
114 41.6 27.0 40.6#



#66
1,1,1,2-Tetrachloroethane
Concen: 0.874 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion:131 Resp: 1632
Ion Ratio Lower Upper
131 100
133 96.9 47.9 143.7
119 0.0 32.8 98.3#









#67
Ethyl Benzene
Concen: 0.896 ug/l
RT: 11.965 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 91 Resp: 8582
Ion Ratio Lower Upper
91 100
106 24.3 24.5 36.7#



#68
m/p-Xylenes
Concen: 1.656 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 106 Resp: 5869
Ion Ratio Lower Upper
106 100
91 201.7 167.1 250.7









#69
o-Xylene
Concen: 0.715 ug/l
RT: 12.400 min Scan# 1775
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:106 Resp: 2398
Ion Ratio Lower Upper
106 100
91 248.5 110.2 330.6



#70
Styrene
Concen: 0.790 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion:104 Resp: 4489
Ion Ratio Lower Upper
104 100
78 47.1 42.7 64.1
103 52.2 43.6 65.4









#71
Bromoform
Concen: 0.850 ug/l
RT: 12.576 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:173 Resp: 1168
Ion Ratio Lower Upper
173 100
175 43.1 23.5 70.6
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion:152 Resp: 101963
Ion Ratio Lower Upper
152 100
115 62.6 31.3 93.9
150 159.0 0.0 349.8





#73

Isopropylbenzene

Concen: 0.899 ug/l

RT: 12.688 min Scan# 1825

Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion:105 Resp: 6991

Ion Ratio Lower Upper

105 100

120 24.1 12.9 38.7



#74

N-ethyl acetate

Concen: 0.916 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. -0.000 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Tgt Ion: 43 Resp: 3228




Ion Ratio Lower Upper


43 100

70 35.2 34.4 51.6

55 21.8 19.6 29.4

61 14.8 19.6 29.4#









#75
1,1,2,2-Tetrachloroethane
Concen: 0.954 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion: 83 Resp: 2233
Ion Ratio Lower Upper
83 100
131 0.0 4.9 14.7#
85 65.3 31.9 95.7



#76
1,2,3-Trichloropropane
Concen: 1.671 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 75 Resp: 3395
Ion Ratio Lower Upper
75 100
77 124.4 96.9 290.7









#77
Bromobenzene
Concen: 0.969 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:156 Resp: 1818
Ion Ratio Lower Upper
156 100
77 232.3 115.1 345.3
158 92.7 48.9 146.8



#78
n-propylbenzene
Concen: 0.781 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 91 Resp: 7045
Ion Ratio Lower Upper
91 100
120 24.2 10.8 32.4





#79

2-Chlorotoluene

Concen: 0.936 ug/l

RT: 13.129 min Scan# 1899

Delta R.T. 0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :





VSTDIC001

Tgt Ion: 91 Resp: 5383

Ion Ratio Lower Upper

91 100

126 27.2 15.7 47.1



#80

1,3,5-Trimethylbenzene

Concen: 0.761 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VN084218.D




Acq: 30 Sep 2024 14:48


Tgt Ion: 105 Resp: 4830

Ion Ratio Lower Upper

105 100

120 50.6 23.7 71.1









#81
trans-1,4-Dichloro-2-butene
Concen: 0.760 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC001


Tgt Ion: 75 Resp: 660
Ion Ratio Lower Upper
75 100
53 99.1 78.0 117.0
89 55.5 35.5 53.3#



#82
4-Chlorotoluene
Concen: 0.893 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion: 91 Resp: 5174
Ion Ratio Lower Upper
91 100
126 31.3 15.7 47.1









#83
tert-Butylbenzene
Concen: 0.870 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:119 Resp: 4909
Ion Ratio Lower Upper
119 100
91 59.0 34.1 102.3
134 18.3 12.8 38.4



#84
1,2,4-Trimethylbenzene
Concen: 0.744 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion:105 Resp: 4760
Ion Ratio Lower Upper
105 100
120 44.7 22.3 66.8









#85
 sec-Butylbenzene
 Concen: 0.770 ug/l
 RT: 13.612 min Scan# 1982
 Delta R.T. -0.006 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion:105 Resp: 5795
 Ion Ratio Lower Upper
 105 100
 134 20.1 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 0.722 ug/l
 RT: 13.723 min Scan# 2001
 Delta R.T. -0.006 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48

Tgt Ion:119 Resp: 4495
 Ion Ratio Lower Upper
 119 100
 134 24.2 12.6 37.8
 91 23.4 12.4 37.2









#87
 1,3-Dichlorobenzene
 Concen: 0.972 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC001


Tgt Ion:146 Resp: 3424
 Ion Ratio Lower Upper
 146 100
 111 45.5 21.8 65.3
 148 67.0 31.9 95.7



#88
 1,4-Dichlorobenzene
 Concen: 0.963 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.082 min
 Lab File: VN084218.D
 Acq: 30 Sep 2024 14:48

Tgt Ion:146 Resp: 3424
 Ion Ratio Lower Upper
 146 100
 111 45.5 21.2 63.6
 148 67.0 31.9 95.9









#89
n-Butylbenzene
Concen: 0.862 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001


Tgt Ion:	91	Resp:	4866
Ion	Ratio	Lower	Upper
91	100		
92	52.1	26.6	79.7
134	21.4	12.3	36.9



#90
Hexachloroethane
Concen: 0.911 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion:	117	Resp:	1151
Ion	Ratio	Lower	Upper
117	100		
201	54.3	34.4	103.2





#91

1,2-Dichlorobenzene

Concen: 1.046 ug/l

RT: 14.100 min Scan# 2066

Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC001





Tgt Ion:146 Resp: 3610

Ion Ratio Lower Upper

146 100

111 39.9 22.0 66.0

148 57.1 32.1 96.3



#92

1,2-Dibromo-3-Chloropropane

Concen: 0.876 ug/l

RT: 14.711 min Scan# 2169

Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion: 75 Resp: 423


Ion Ratio Lower Upper

75 100

155 51.3 38.1 114.5

157 59.1 49.6 149.0





#93

1,2,4-Trichlorobenzene

Concen: 0.788 ug/l

RT: 15.400 min Scan# 21

Delta R.T. 0.012 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC001





Tgt Ion:180 Resp: 1340

Ion Ratio Lower Upper

180 100

182 103.0 47.0 141.0

145 41.3 16.4 49.4



#94

Hexachlorobutadiene

Concen: 1.015 ug/l

RT: 15.494 min Scan# 2302

Delta R.T. -0.006 min

Lab File: VN084218.D

Acq: 30 Sep 2024 14:48




Tgt Ion:225 Resp: 861


Ion Ratio Lower Upper

225 100

223 52.7 32.4 97.2

227 60.7 32.4 97.0









#95
Naphthalene
Concen: 0.892 ug/l
RT: 15.641 min Scan# 21
Delta R.T. 0.006 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC001

Tgt Ion:128	Resp:	5031
Ion Ratio	Lower	Upper
128	100	
127	11.5	10.5 15.7
129	11.3	8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 0.899 ug/l
RT: 15.847 min Scan# 2362
Delta R.T. 0.012 min
Lab File: VN084218.D
Acq: 30 Sep 2024 14:48

Tgt Ion:180	Resp:	1531
Ion Ratio	Lower	Upper
180	100	
182	101.0	48.2 144.6
145	37.5	17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084220.D
 Acq On : 30 Sep 2024 15:36
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN093024

Quant Time: Oct 01 07:14:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	174241	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	300916	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	263364	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	130851	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	130630	50.564	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.120%	
35) Dibromofluoromethane	8.165	113	100200	50.508	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.020%	
50) Toluene-d8	10.565	98	387775	53.129	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.260%	
62) 4-Bromofluorobenzene	12.847	95	140850	52.971	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 105.940%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	108641	52.712	ug/l	95
3) Chloromethane	2.359	50	121140	51.477	ug/l	97
4) Vinyl Chloride	2.512	62	116792	51.235	ug/l	99
5) Bromomethane	2.947	94	71654	47.656	ug/l	99
6) Chloroethane	3.118	64	73658	45.119	ug/l	94
7) Trichlorofluoromethane	3.495	101	188482	53.099	ug/l	98
8) Diethyl Ether	3.959	74	70360	53.445	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.371	101	104895	51.573	ug/l	99
10) Methyl Iodide	4.589	142	141430	53.996	ug/l	92
11) Tert butyl alcohol	5.524	59	109545	257.154	ug/l	99
12) 1,1-Dichloroethene	4.342	96	104103	53.064	ug/l	95
13) Acrolein	4.177	56	125971	257.587	ug/l	97
14) Allyl chloride	5.024	41	174077	51.092	ug/l	100
15) Acrylonitrile	5.718	53	293962	273.264	ug/l	99
16) Acetone	4.430	43	308020	277.547	ug/l	97
17) Carbon Disulfide	4.712	76	302450	48.693	ug/l	100
18) Methyl Acetate	5.024	43	130494	53.948	ug/l	98
19) Methyl tert-butyl Ether	5.800	73	360019	54.371	ug/l	96
20) Methylene Chloride	5.277	84	117572	52.116	ug/l	89
21) trans-1,2-Dichloroethene	5.783	96	108894	52.980	ug/l	97
22) Diisopropyl ether	6.671	45	384647	54.579	ug/l	99
23) Vinyl Acetate	6.600	43	1492704	285.569	ug/l	99
24) 1,1-Dichloroethane	6.565	63	211544	53.676	ug/l	98
25) 2-Butanone	7.482	43	426860	282.397	ug/l	98
26) 2,2-Dichloropropane	7.488	77	189858	53.428	ug/l	99
27) cis-1,2-Dichloroethene	7.488	96	128843	51.876	ug/l	98
28) Bromochloromethane	7.818	49	84632	48.129	ug/l	97
29) Tetrahydrofuran	7.835	42	262996	281.940	ug/l	99
30) Chloroform	7.965	83	213634	52.184	ug/l	95
31) Cyclohexane	8.259	56	190602	49.809	ug/l	100
32) 1,1,1-Trichloroethane	8.171	97	196319	53.395	ug/l	99
36) 1,1-Dichloropropene	8.371	75	155365	53.917	ug/l	99
37) Ethyl Acetate	7.559	43	159764	54.043	ug/l	99
38) Carbon Tetrachloride	8.359	117	169403	53.593	ug/l	94
39) Methylcyclohexane	9.600	83	181676	56.648	ug/l	98
40) Benzene	8.606	78	478538	53.301	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
 Data File : VN084220.D
 Acq On : 30 Sep 2024 15:36
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN093024

Quant Time: Oct 01 07:14:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	89407	55.708	ug/l	98
42) 1,2-Dichloroethane	8.671	62	164758	54.131	ug/l	99
43) Isopropyl Acetate	8.688	43	278668	48.126	ug/l	100
44) Trichloroethene	9.353	130	109810	52.368	ug/l	100
45) 1,2-Dichloropropane	9.618	63	116161	54.694	ug/l	99
46) Dibromomethane	9.706	93	80284	56.015	ug/l	100
47) Bromodichloromethane	9.888	83	173015	54.335	ug/l	98
48) Methyl methacrylate	9.682	41	132287	57.391	ug/l	98
49) 1,4-Dioxane	9.694	88	47169	1111.545	ug/l	99
51) 4-Methyl-2-Pentanone	10.441	43	808630	287.336	ug/l	100
52) Toluene	10.629	92	296651	54.185	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	183153	56.426	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	193475	55.416	ug/l	98
55) 1,1,2-Trichloroethane	11.018	97	108550	55.307	ug/l	98
56) Ethyl methacrylate	10.876	69	189121	58.561	ug/l	100
57) 1,3-Dichloropropane	11.165	76	191966	54.738	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	415454	277.315	ug/l	98
59) 2-Hexanone	11.194	43	621000	295.238	ug/l	100
60) Dibromochloromethane	11.359	129	127786	55.105	ug/l	99
61) 1,2-Dibromoethane	11.470	107	111244	54.548	ug/l	98
64) Tetrachloroethene	11.100	164	96460	52.585	ug/l	99
65) Chlorobenzene	11.888	112	312728	53.555	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.959	131	111388	55.381	ug/l	97
67) Ethyl Benzene	11.965	91	573790	55.563	ug/l	99
68) m/p-Xylenes	12.070	106	434515	113.748	ug/l	99
69) o-Xylene	12.400	106	209737	58.032	ug/l	99
70) Styrene	12.412	104	355892	58.136	ug/l	99
71) Bromoform	12.576	173	84787	57.258	ug/l #	99
73) Isopropylbenzene	12.694	105	535240	53.604	ug/l	100
74) N-amyl acetate	12.494	43	249223	55.118	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.935	83	157322	52.398	ug/l	100
76) 1,2,3-Trichloropropane	12.994	75	213162	79.795	ug/l	58
77) Bromobenzene	12.976	156	124463	51.676	ug/l	98
78) n-propylbenzene	13.035	91	638464	55.188	ug/l	99
79) 2-Chlorotoluene	13.123	91	390502	52.898	ug/l	100
80) 1,3,5-Trimethylbenzene	13.170	105	456943	56.101	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	58987	52.939	ug/l	99
82) 4-Chlorotoluene	13.217	91	397246	53.455	ug/l	100
83) tert-Butylbenzene	13.435	119	390221	53.903	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	461123	56.199	ug/l	98
85) sec-Butylbenzene	13.617	105	536685	55.556	ug/l	99
86) p-Isopropyltoluene	13.729	119	452245	56.614	ug/l	100
87) 1,3-Dichlorobenzene	13.729	146	234190	51.830	ug/l	99
88) 1,4-Dichlorobenzene	13.811	146	234070	51.298	ug/l	99
89) n-Butylbenzene	14.053	91	396001	54.671	ug/l	100
90) Hexachloroethane	14.329	117	82363	50.795	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	226320	51.082	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	32296	52.125	ug/l	96
93) 1,2,4-Trichlorobenzene	15.394	180	114157	52.311	ug/l	99
94) Hexachlorobutadiene	15.500	225	50461	46.368	ug/l	97
95) Naphthalene	15.641	128	380549	52.602	ug/l	99
96) 1,2,3-Trichlorobenzene	15.835	180	111086	50.853	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084220.D
Acq On : 30 Sep 2024 15:36
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024

Quant Time: Oct 01 07:14:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN093024\
Data File : VN084220.D
Acq On : 30 Sep 2024 15:36
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Quant Time: Oct 01 07:14:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024







Tgt Ion:168 Resp: 174241
Ion Ratio Lower Upper
168 100
99 66.5 54.2 81.2



#2
Dichlorodifluoromethane
Concen: 52.712 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 85 Resp: 108641
Ion Ratio Lower Upper
85 100
87 32.5 17.6 52.9









#3
Chloromethane
Concen: 51.477 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 50 Resp: 121140
Ion Ratio Lower Upper
50 100
52 31.9 26.8 40.2



#4
Vinyl Chloride
Concen: 51.235 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.006 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 62 Resp: 116792
Ion Ratio Lower Upper
62 100
64 33.3 26.2 39.4









#5
Bromomethane
Concen: 47.656 ug/l
RT: 2.947 min Scan# 1
Delta R.T. -0.006 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 94 Resp: 71654
Ion Ratio Lower Upper
94 100
96 91.7 72.9 109.3



#6
Chloroethane
Concen: 45.119 ug/l
RT: 3.118 min Scan# 198
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 64 Resp: 73658
Ion Ratio Lower Upper
64 100
66 32.7 23.5 35.3









#7
 Trichlorofluoromethane
 Concen: 53.099 ug/l
 RT: 3.495 min Scan# 262
 Delta R.T. -0.000 min
 Lab File: VN084220.D
 Acq: 30 Sep 2024 15:36




Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN093024


Tgt Ion:101 Resp: 188482
 Ion Ratio Lower Upper
 101 100
 103 63.4 51.8 77.6



#8
 Diethyl Ether
 Concen: 53.445 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. -0.000 min
 Lab File: VN084220.D
 Acq: 30 Sep 2024 15:36

Tgt Ion: 74 Resp: 70360
 Ion Ratio Lower Upper
 74 100
 45 97.0 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 51.573 ug/l
RT: 4.371 min Scan# 411
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion:101 Resp: 104895
Ion Ratio Lower Upper
101 100
85 46.0 36.2 54.4
151 74.0 59.3 88.9



#10
Methyl Iodide
Concen: 53.996 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.006 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion:142 Resp: 141430
Ion Ratio Lower Upper
142 100
127 50.3 35.3 52.9
141 14.7 11.0 16.6









#11
Tert butyl alcohol
Concen: 257.154 ug/l
RT: 5.524 min Scan# 607
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 59 Resp: 109545
Ion Ratio Lower Upper
59 100
57 10.6 8.2 12.2



#12
1,1-Dichloroethene
Concen: 53.064 ug/l
RT: 4.342 min Scan# 406
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 96 Resp: 104103
Ion Ratio Lower Upper
96 100
61 166.3 128.6 192.8
98 60.4 51.7 77.5









#13
Acrolein
Concen: 257.587 ug/l
RT: 4.177 min Scan# 31
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 56 Resp: 125971
Ion Ratio Lower Upper
56 100
55 69.4 57.2 85.8



#14
Allyl chloride
Concen: 51.092 ug/l
RT: 5.024 min Scan# 522
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 41 Resp: 174077
Ion Ratio Lower Upper
41 100
39 73.7 59.0 88.6
76 37.1 29.8 44.6









#15
Acrylonitrile
Concen: 273.264 ug/l
RT: 5.718 min Scan# 640
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 53 Resp: 293962
Ion Ratio Lower Upper
53 100
52 81.3 66.2 99.4
51 35.7 28.8 43.2



#16
Acetone
Concen: 277.547 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.006 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 43 Resp: 308020
Ion Ratio Lower Upper
43 100
58 32.2 24.4 36.6









#17
Carbon Disulfide
Concen: 48.693 ug/l
RT: 4.712 min Scan# 41
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 76 Resp: 302450
Ion Ratio Lower Upper
76 100
78 9.1 7.3 10.9



#18
Methyl Acetate
Concen: 53.948 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 43 Resp: 130494
Ion Ratio Lower Upper
43 100
74 23.0 19.2 28.8









#19
Methyl tert-butyl Ether
Concen: 54.371 ug/l
RT: 5.800 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICV093024


Tgt Ion: 73 Resp: 360019
Ion Ratio Lower Upper
73 100
57 20.3 17.7 26.5



#20
Methylene Chloride
Concen: 52.116 ug/l
RT: 5.277 min Scan# 565
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 84 Resp: 117572
Ion Ratio Lower Upper
84 100
49 120.2 107.2 160.8
51 36.9 31.8 47.8
86 56.5 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 52.980 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 96 Resp: 108894
Ion Ratio Lower Upper
96 100
61 139.8 108.4 162.6
98 62.9 49.8 74.8



#22
Diisopropyl ether
Concen: 54.579 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 45 Resp: 384647
Ion Ratio Lower Upper
45 100
43 53.5 42.3 63.5
87 27.0 21.8 32.6
59 11.3 9.6 14.4









#23
 Vinyl Acetate
 Concen: 285.569 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. -0.000 min
 Lab File: VN084220.D
 Acq: 30 Sep 2024 15:36




Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN093024


Tgt Ion: 43 Resp: 1492704
 Ion Ratio Lower Upper
 43 100
 86 10.5 8.6 12.8



#24
 1,1-Dichloroethane
 Concen: 53.676 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. -0.006 min
 Lab File: VN084220.D
 Acq: 30 Sep 2024 15:36

Tgt Ion: 63 Resp: 211544
 Ion Ratio Lower Upper
 63 100
 98 6.4 3.4 10.2
 100 4.9 2.0 5.8









#25
2-Butanone
Concen: 282.397 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 43 Resp: 426860
Ion Ratio Lower Upper
43 100
72 25.9 21.4 32.2



#26
2,2-Dichloropropane
Concen: 53.428 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 77 Resp: 189858
Ion Ratio Lower Upper
77 100
97 22.1 10.8 32.3









#27
 cis-1,2-Dichloroethene
 Concen: 51.876 ug/l
 RT: 7.488 min Scan# 941
 Delta R.T. -0.000 min
 Lab File: VN084220.D
 Acq: 30 Sep 2024 15:36

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN093024

Tgt Ion: 96 Resp: 128843




Ion	Ratio	Lower	Upper
96	100		
61	154.8	0.0	301.8
98	64.2	0.0	128.2




#28
 Bromochloromethane
 Concen: 48.129 ug/l
 RT: 7.818 min Scan# 997
 Delta R.T. 0.006 min
 Lab File: VN084220.D
 Acq: 30 Sep 2024 15:36

Tgt Ion: 49 Resp: 84632

Ion	Ratio	Lower	Upper
49	100		
129	1.6	0.0	3.6
130	70.0	58.0	87.0









#29
Tetrahydrofuran
Concen: 281.940 ug/l
RT: 7.835 min Scan# 1000
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 42 Resp: 262996
Ion Ratio Lower Upper
42 100
72 42.8 34.6 51.8
71 40.0 32.2 48.4



#30
Chloroform
Concen: 52.184 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36


Tgt Ion: 83 Resp: 213634
Ion Ratio Lower Upper
83 100
85 67.5 51.0 76.4








#31
Cyclohexane
Concen: 49.809 ug/l
RT: 8.259 min Scan# 1072
Delta R.T. 0.006 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36


Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024





Tgt Ion: 56 Resp: 190602
Ion Ratio Lower Upper
56 100
69 30.1 23.8 35.6
84 85.8 68.8 103.2




#32
1,1,1-Trichloroethane
Concen: 53.395 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36



Tgt Ion: 97 Resp: 196319
Ion Ratio Lower Upper
97 100
99 64.5 51.2 76.8
61 47.6 37.5 56.3





#33

1,2-Dichloroethane-d4

Concen: 50.564 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :





ICVNV093024

Tgt Ion: 65 Resp: 130630

Ion Ratio Lower Upper

65 100

67 53.0 0.0 102.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion: 114 Resp: 300916


Ion Ratio Lower Upper

114 100

63 20.6 0.0 43.8

88 17.0 0.0 31.6





#35

Dibromofluoromethane

Concen: 50.508 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024





Tgt Ion:113 Resp: 100200

Ion Ratio Lower Upper

113 100

111 102.6 83.3 124.9

192 17.3 13.5 20.3



#36

1,1-Dichloropropene

Concen: 53.917 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion: 75 Resp: 155365


Ion Ratio Lower Upper

75 100

110 33.2 16.9 50.7

77 30.7 24.7 37.1









#37
Ethyl Acetate
Concen: 54.043 ug/l
RT: 7.559 min Scan# 91
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 43 Resp: 159764
Ion Ratio Lower Upper
43 100
61 13.5 10.7 16.1
70 10.0 8.8 13.2



#38
Carbon Tetrachloride
Concen: 53.593 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 117 Resp: 169403
Ion Ratio Lower Upper
117 100
119 90.5 78.7 118.1
121 29.9 24.4 36.6





#39

Methylcyclohexane

Concen: 56.648 ug/l

RT: 9.600 min Scan# 11300

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024





Tgt Ion: 83 Resp: 181676

Ion Ratio Lower Upper

83 100

55 76.1 62.8 94.2

98 48.6 38.8 58.2



#40

Benzene

Concen: 53.301 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN084220.D




Acq: 30 Sep 2024 15:36


Tgt Ion: 78 Resp: 478538

Ion Ratio Lower Upper

78 100

77 23.3 19.1 28.7









#41
Methacrylonitrile
Concen: 55.708 ug/l
RT: 7.777 min Scan# 990
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion:	41	Resp:	89407
Ion	Ratio	Lower	Upper
41	100		
39	56.2	43.3	64.9
67	70.8	56.1	84.1
52	30.2	24.5	36.7



#42
1,2-Dichloroethane
Concen: 54.131 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion:	62	Resp:	164758
Ion	Ratio	Lower	Upper
62	100		
98	8.6	0.0	18.2









#43
Isopropyl Acetate
Concen: 48.126 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 43 Resp: 278668
Ion Ratio Lower Upper
43 100
61 25.8 20.4 30.6
87 12.9 10.3 15.5



#44
Trichloroethene
Concen: 52.368 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.006 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion:130 Resp: 109810
Ion Ratio Lower Upper
130 100
95 101.8 0.0 203.0





#45

1,2-Dichloropropane

Concen: 54.694 ug/l

RT: 9.618 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :





ICV093024

Tgt Ion: 63 Resp: 116161

Ion Ratio Lower Upper

63 100

65 30.4 24.6 37.0



#46

Dibromomethane

Concen: 56.015 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion: 93 Resp: 80284


Ion Ratio Lower Upper

93 100

95 81.5 65.0 97.4

174 89.6 71.8 107.8









#47
Bromodichloromethane
Concen: 54.335 ug/l
RT: 9.888 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 83 Resp: 173015
Ion Ratio Lower Upper
83 100
85 62.5 51.5 77.3
127 7.9 5.7 8.5



#48
Methyl methacrylate
Concen: 57.391 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. 0.006 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 41 Resp: 132287
Ion Ratio Lower Upper
41 100
69 81.0 67.1 100.7
39 56.9 45.8 68.6





#49

1,4-Dioxane

Concen: 1111.545 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024





Tgt Ion: 88 Resp: 47169

Ion Ratio Lower Upper

88 100

43 31.2 25.3 37.9

58 71.7 56.4 84.6



#50

Toluene-d8

Concen: 53.129 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN084220.D




Acq: 30 Sep 2024 15:36


Tgt Ion: 98 Resp: 387775

Ion Ratio Lower Upper

98 100

100 65.1 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 287.336 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 43 Resp: 808630
Ion Ratio Lower Upper
43 100
58 39.5 31.6 47.4



#52
Toluene
Concen: 54.185 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 92 Resp: 296651
Ion Ratio Lower Upper
92 100
91 173.0 137.3 205.9









#53
 t-1,3-Dichloropropene
 Concen: 56.426 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. -0.000 min
 Lab File: VN084220.D
 Acq: 30 Sep 2024 15:36




Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN093024


Tgt Ion: 75 Resp: 183153
 Ion Ratio Lower Upper
 75 100
 77 32.1 25.4 38.0



#54
 cis-1,3-Dichloropropene
 Concen: 55.416 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. -0.000 min
 Lab File: VN084220.D
 Acq: 30 Sep 2024 15:36

Tgt Ion: 75 Resp: 193475
 Ion Ratio Lower Upper
 75 100
 77 31.6 24.2 36.2
 39 49.4 38.7 58.1





#55

1,1,2-Trichloroethane

Concen: 55.307 ug/l

RT: 11.018 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024

Tgt Ion: 97 Resp: 108550





Ion Ratio Lower Upper

97 100

83 85.0 71.0 106.4

85 53.6 41.7 62.5

99 62.4 50.1 75.1



#56

Ethyl methacrylate

Concen: 58.561 ug/l

RT: 10.876 min Scan# 1517

Delta R.T. 0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion: 69 Resp: 189121


Ion Ratio Lower Upper

69 100

41 66.4 53.3 79.9

39 35.9 28.8 43.2





#57

1,3-Dichloropropane

Concen: 54.738 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :





ICVVN093024

Tgt Ion: 76 Resp: 191966

Ion Ratio Lower Upper

76 100

78 32.0 25.8 38.6



#58

2-Chloroethyl Vinyl ether

Concen: 277.315 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN084220.D




Acq: 30 Sep 2024 15:36


Tgt Ion: 63 Resp: 415454

Ion Ratio Lower Upper

63 100

106 28.0 21.8 32.6









#59
2-Hexanone
Concen: 295.238 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 43 Resp: 621000
Ion Ratio Lower Upper
43 100
58 54.4 27.3 81.8



#60
Dibromochloromethane
Concen: 55.105 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 129 Resp: 127786
Ion Ratio Lower Upper
129 100
127 77.9 38.3 114.9





#61

1,2-Dibromoethane

Concen: 54.548 ug/l

RT: 11.470 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :





ICVVN093024

Tgt Ion:107 Resp: 111244

Ion Ratio Lower Upper

107 100

109 92.5 75.4 113.0



#62

4-Bromofluorobenzene

Concen: 52.971 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion: 95 Resp: 140850


Ion Ratio Lower Upper

95 100

174 70.8 0.0 145.2

176 69.2 0.0 140.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion:117 Resp: 263364
Ion Ratio Lower Upper
117 100
82 61.2 47.2 70.8
119 32.9 25.4 38.0



#64
Tetrachloroethene
Concen: 52.585 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion:164 Resp: 96460
Ion Ratio Lower Upper
164 100
166 126.7 101.0 151.4
129 102.6 80.8 121.2
131 98.1 78.1 117.1









#65
Chlorobenzene
Concen: 53.555 ug/l
RT: 11.888 min Scan# 112
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion:112 Resp: 312728
Ion Ratio Lower Upper
112 100
114 32.0 27.0 40.6



#66
1,1,1,2-Tetrachloroethane
Concen: 55.381 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion:131 Resp: 111388
Ion Ratio Lower Upper
131 100
133 93.4 47.9 143.7
119 62.9 32.8 98.3









#67
Ethyl Benzene
Concen: 55.563 ug/l
RT: 11.965 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 91 Resp: 573790
Ion Ratio Lower Upper
91 100
106 30.3 24.5 36.7



#68
m/p-Xylenes
Concen: 113.748 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 106 Resp: 434515
Ion Ratio Lower Upper
106 100
91 208.1 167.1 250.7









#69
o-Xylene
Concen: 58.032 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. 0.006 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion:106 Resp: 209737
Ion Ratio Lower Upper
106 100
91 219.4 110.2 330.6



#70
Styrene
Concen: 58.136 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion:104 Resp: 355892
Ion Ratio Lower Upper
104 100
78 52.7 42.7 64.1
103 55.1 43.6 65.4









#71
Bromoform
Concen: 57.258 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion:173 Resp: 84787
Ion Ratio Lower Upper
173 100
175 48.0 23.5 70.6
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion:152 Resp: 130851
Ion Ratio Lower Upper
152 100
115 62.2 31.3 93.9
150 174.1 0.0 349.8









#73
Isopropylbenzene
Concen: 53.604 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion:105 Resp: 535240
Ion Ratio Lower Upper
105 100
120 25.9 12.9 38.7



#74
N-ethyl acetate
Concen: 55.118 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 43 Resp: 249223
Ion Ratio Lower Upper
43 100
70 42.5 34.4 51.6
55 24.2 19.6 29.4
61 23.8 19.6 29.4









#75
1,1,2,2-Tetrachloroethane
Concen: 52.398 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 83 Resp: 157322
Ion Ratio Lower Upper
83 100
131 10.0 4.9 14.7
85 63.6 31.9 95.7



#76
1,2,3-Trichloropropane
Concen: 79.795 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. 0.006 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 75 Resp: 213162
Ion Ratio Lower Upper
75 100
77 131.8 96.9 290.7





#77

Bromobenzene

Concen: 51.676 ug/l

RT: 12.976 min Scan# 1874

Delta R.T. -0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024





Tgt Ion:156 Resp: 124463

Ion Ratio Lower Upper

156 100

77 225.7 115.1 345.3

158 96.8 48.9 146.8



#78

n-propylbenzene

Concen: 55.188 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. -0.000 min

Lab File: VN084220.D




Acq: 30 Sep 2024 15:36


Tgt Ion: 91 Resp: 638464

Ion Ratio Lower Upper

91 100

120 21.9 10.8 32.4









#79
2-Chlorotoluene
Concen: 52.898 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 91 Resp: 390502
Ion Ratio Lower Upper
91 100
126 31.2 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 56.101 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 105 Resp: 456943
Ion Ratio Lower Upper
105 100
120 47.0 23.7 71.1









#81
trans-1,4-Dichloro-2-butene
Concen: 52.939 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 75 Resp: 58987
Ion Ratio Lower Upper
75 100
53 98.1 78.0 117.0
89 45.7 35.5 53.3



#82
4-Chlorotoluene
Concen: 53.455 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 91 Resp: 397246
Ion Ratio Lower Upper
91 100
126 31.3 15.7 47.1









#83
tert-Butylbenzene
Concen: 53.903 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion:119 Resp: 390221
Ion Ratio Lower Upper
119 100
91 68.6 34.1 102.3
134 23.9 12.8 38.4



#84
1,2,4-Trimethylbenzene
Concen: 56.199 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion:105 Resp: 461123
Ion Ratio Lower Upper
105 100
120 43.4 22.3 66.8









#85
 sec-Butylbenzene
 Concen: 55.556 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN084220.D
 Acq: 30 Sep 2024 15:36




Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN093024


Tgt Ion:105 Resp: 536685
 Ion Ratio Lower Upper
 105 100
 134 19.2 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 56.614 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN084220.D
 Acq: 30 Sep 2024 15:36

Tgt Ion:119 Resp: 452245
 Ion Ratio Lower Upper
 119 100
 134 25.3 12.6 37.8
 91 24.9 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 51.830 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024





Tgt Ion:146 Resp: 234190

Ion Ratio Lower Upper

146 100

111 42.7 21.8 65.3

148 63.9 31.9 95.7



#88

1,4-Dichlorobenzene

Concen: 51.298 ug/l

RT: 13.811 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion:146 Resp: 234070


Ion Ratio Lower Upper

146 100

111 42.2 21.2 63.6

148 63.2 31.9 95.9









#89
n-Butylbenzene
Concen: 54.671 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion: 91 Resp: 396001
Ion Ratio Lower Upper
91 100
92 53.1 26.6 79.7
134 24.7 12.3 36.9



#90
Hexachloroethane
Concen: 50.795 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 117 Resp: 82363
Ion Ratio Lower Upper
117 100
201 68.4 34.4 103.2









#91
1,2-Dichlorobenzene
Concen: 51.082 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36




Instrument :
MSVOA_N
ClientSampleId :
ICVVN093024


Tgt Ion:146 Resp: 226320
Ion Ratio Lower Upper
146 100
111 43.9 22.0 66.0
148 64.7 32.1 96.3



#92
1,2-Dibromo-3-Chloropropane
Concen: 52.125 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN084220.D
Acq: 30 Sep 2024 15:36

Tgt Ion: 75 Resp: 32296
Ion Ratio Lower Upper
75 100
155 74.4 38.1 114.5
157 94.7 49.6 149.0





#93

1,2,4-Trichlorobenzene

Concen: 52.311 ug/l

RT: 15.394 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024





Tgt Ion:180 Resp: 114157

Ion Ratio Lower Upper

180 100

182 95.1 47.0 141.0

145 33.8 16.4 49.4



#94

Hexachlorobutadiene

Concen: 46.368 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion:225 Resp: 50461


Ion Ratio Lower Upper

225 100

223 61.5 32.4 97.2

227 62.7 32.4 97.0





#95

Naphthalene

Concen: 52.602 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. 0.006 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36

Instrument :

MSVOA_N

ClientSampleId :

ICVVN093024





Tgt Ion:128 Resp: 380549

Ion Ratio Lower Upper

128 100

127 12.6 10.5 15.7

129 11.0 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 50.853 ug/l

RT: 15.835 min Scan# 2360

Delta R.T. -0.000 min

Lab File: VN084220.D

Acq: 30 Sep 2024 15:36




Tgt Ion:180 Resp: 111086

Ion Ratio Lower Upper

180 100

182 95.5 48.2 144.6

145 34.9 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084428.D
 Acq On : 21 Oct 2024 11:38
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Quant Time: Oct 22 01:29:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	157355	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	262800	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	244635	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	127638	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	124506	53.365	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.740%	
35) Dibromofluoromethane	8.165	113	101751	58.729	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 117.460%	
50) Toluene-d8	10.565	98	376165	59.014	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 118.020%#	
62) 4-Bromofluorobenzene	12.847	95	137246	59.102	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 118.200%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	83308	44.759	ug/l	98
3) Chloromethane	2.359	50	94679	44.550	ug/l	98
4) Vinyl Chloride	2.512	62	93996	45.660	ug/l	97
5) Bromomethane	2.936	94	58206	42.866	ug/l	93
6) Chloroethane	3.106	64	63113	42.808	ug/l	94
7) Trichlorofluoromethane	3.489	101	164207	51.224	ug/l	99
8) Diethyl Ether	3.959	74	59329	49.902	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.365	101	94908	51.670	ug/l	99
10) Methyl Iodide	4.583	142	116312	49.171	ug/l	96
11) Tert butyl alcohol	5.530	59	94486	245.606	ug/l	100
12) 1,1-Dichloroethene	4.342	96	89116	50.300	ug/l	98
13) Acrolein	4.177	56	85169	192.843	ug/l	100
14) Allyl chloride	5.018	41	134288	43.643	ug/l	98
15) Acrylonitrile	5.718	53	275307	283.386	ug/l	99
16) Acetone	4.430	43	249463	248.905	ug/l	100
17) Carbon Disulfide	4.706	76	224021	39.937	ug/l	99
18) Methyl Acetate	5.024	43	143956	65.900	ug/l	99
19) Methyl tert-butyl Ether	5.794	73	319663	53.457	ug/l	98
20) Methylene Chloride	5.271	84	105509	51.787	ug/l	88
21) trans-1,2-Dichloroethene	5.783	96	93324	50.278	ug/l	94
22) Diisopropyl ether	6.671	45	336994	52.949	ug/l	99
23) Vinyl Acetate	6.606	43	1190800	252.258	ug/l	100
24) 1,1-Dichloroethane	6.571	63	187107	52.570	ug/l	98
25) 2-Butanone	7.483	43	367222	269.013	ug/l	99
26) 2,2-Dichloropropane	7.488	77	159617	49.738	ug/l	97
27) cis-1,2-Dichloroethene	7.488	96	118267	52.728	ug/l	96
28) Bromochloromethane	7.812	49	83905	52.836	ug/l	93
29) Tetrahydrofuran	7.841	42	238619	283.258	ug/l	98
30) Chloroform	7.965	83	201405	54.476	ug/l	98
31) Cyclohexane	8.259	56	150876	43.659	ug/l	98
32) 1,1,1-Trichloroethane	8.165	97	175417	52.830	ug/l	98
36) 1,1-Dichloropropene	8.371	75	130914	52.021	ug/l	98
37) Ethyl Acetate	7.559	43	137280	53.172	ug/l	98
38) Carbon Tetrachloride	8.359	117	149628	54.202	ug/l	97
39) Methylcyclohexane	9.600	83	137117	48.955	ug/l	97
40) Benzene	8.606	78	429647	54.796	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084428.D
 Acq On : 21 Oct 2024 11:38
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050

Quant Time: Oct 22 01:29:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	79086	56.424	ug/l	96
42) 1,2-Dichloroethane	8.671	62	144178	54.240	ug/l	99
43) Isopropyl Acetate	8.688	43	233616	46.197	ug/l	98
44) Trichloroethene	9.353	130	98494	53.784	ug/l	100
45) 1,2-Dichloropropane	9.618	63	105490	56.873	ug/l	96
46) Dibromomethane	9.706	93	74735	59.706	ug/l	98
47) Bromodichloromethane	9.888	83	159166	57.235	ug/l #	96
48) Methyl methacrylate	9.682	41	112649	55.960	ug/l	99
49) 1,4-Dioxane	9.694	88	46681	1259.594	ug/l	97
51) 4-Methyl-2-Pentanone	10.441	43	757152	308.066	ug/l	99
52) Toluene	10.629	92	268859	56.231	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	157928	55.712	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	171021	56.089	ug/l	96
55) 1,1,2-Trichloroethane	11.018	97	104404	60.910	ug/l	97
56) Ethyl methacrylate	10.871	69	164383	58.283	ug/l	96
57) 1,3-Dichloropropane	11.165	76	176631	57.670	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	397738	303.996	ug/l	97
59) 2-Hexanone	11.194	43	560104	304.908	ug/l	97
60) Dibromochloromethane	11.359	129	122440	60.458	ug/l	99
61) 1,2-Dibromoethane	11.471	107	102008	57.274	ug/l	100
64) Tetrachloroethene	11.100	164	89495	52.523	ug/l	97
65) Chlorobenzene	11.888	112	289014	53.283	ug/l	93
66) 1,1,1,2-Tetrachloroethane	11.959	131	103631	55.468	ug/l	98
67) Ethyl Benzene	11.965	91	506469	52.799	ug/l	99
68) m/p-Xylenes	12.071	106	394932	111.301	ug/l	99
69) o-Xylene	12.394	106	187215	55.767	ug/l	100
70) Styrene	12.412	104	331204	58.246	ug/l	98
71) Bromoform	12.576	173	81253	59.072	ug/l #	100
73) Isopropylbenzene	12.694	105	484656	49.760	ug/l	100
74) N-amyl acetate	12.494	43	204351	46.332	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.935	83	152956	52.226	ug/l	98
76) 1,2,3-Trichloropropane	12.994	75	200192	76.826	ug/l	58
77) Bromobenzene	12.976	156	118952	50.631	ug/l	96
78) n-propylbenzene	13.035	91	576662	51.101	ug/l	98
79) 2-Chlorotoluene	13.123	91	354683	49.256	ug/l	98
80) 1,3,5-Trimethylbenzene	13.170	105	416797	52.460	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	58513	53.835	ug/l	90
82) 4-Chlorotoluene	13.218	91	364438	50.275	ug/l	99
83) tert-Butylbenzene	13.435	119	352099	49.862	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	421969	52.722	ug/l	99
85) sec-Butylbenzene	13.617	105	492135	52.226	ug/l	99
86) p-Isopropyltoluene	13.729	119	412249	52.906	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	226954	51.493	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	223977	50.322	ug/l	99
89) n-Butylbenzene	14.053	91	349709	49.495	ug/l	99
90) Hexachloroethane	14.329	117	76242	48.204	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	220370	50.991	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.717	75	30011	49.657	ug/l	97
93) 1,2,4-Trichlorobenzene	15.388	180	109356	51.373	ug/l	99
94) Hexachlorobutadiene	15.500	225	50783	47.838	ug/l	96
95) Naphthalene	15.641	128	361412	51.214	ug/l	100
96) 1,2,3-Trichlorobenzene	15.841	180	110846	52.021	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084428.D
Acq On : 21 Oct 2024 11:38
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Quant Time: Oct 22 01:29:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:168 Resp: 157355
Ion Ratio Lower Upper
168 100
99 61.7 54.2 81.2



#2
Dichlorodifluoromethane
Concen: 44.759 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 85 Resp: 83308
Ion Ratio Lower Upper
85 100
87 34.4 17.6 52.9









#3
Chloromethane
Concen: 44.550 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 50 Resp: 94679
Ion Ratio Lower Upper
50 100
52 32.3 26.8 40.2



#4
Vinyl Chloride
Concen: 45.660 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 62 Resp: 93996
Ion Ratio Lower Upper
62 100
64 30.9 26.2 39.4









#5
Bromomethane
Concen: 42.866 ug/l
RT: 2.936 min Scan# 11
Delta R.T. -0.018 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 94 Resp: 58206
Ion Ratio Lower Upper
94 100
96 98.2 72.9 109.3



#6
Chloroethane
Concen: 42.808 ug/l
RT: 3.106 min Scan# 196
Delta R.T. -0.012 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 64 Resp: 63113
Ion Ratio Lower Upper
64 100
66 32.4 23.5 35.3









#7
 Trichlorofluoromethane
 Concen: 51.224 ug/l
 RT: 3.489 min Scan# 210
 Delta R.T. -0.006 min
 Lab File: VN084428.D
 Acq: 21 Oct 2024 11:38




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion:101 Resp: 164207
 Ion Ratio Lower Upper
 101 100
 103 64.0 51.8 77.6



#8
 Diethyl Ether
 Concen: 49.902 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. 0.000 min
 Lab File: VN084428.D
 Acq: 21 Oct 2024 11:38

Tgt Ion: 74 Resp: 59329
 Ion Ratio Lower Upper
 74 100
 45 95.6 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 51.670 ug/l
RT: 4.365 min Scan# 411
Delta R.T. -0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 94908
Ion Ratio Lower Upper
101 100
85 44.2 36.2 54.4
151 75.0 59.3 88.9



#10
Methyl Iodide
Concen: 49.171 ug/l
RT: 4.583 min Scan# 447
Delta R.T. -0.012 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion:142 Resp: 116312
Ion Ratio Lower Upper
142 100
127 47.5 35.3 52.9
141 14.4 11.0 16.6





#11

Tert butyl alcohol

Concen: 245.606 ug/l

RT: 5.530 min Scan# 607

Delta R.T. 0.006 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 59 Resp: 94486

Ion Ratio Lower Upper

59 100

57 10.2 8.2 12.2



#12

1,1-Dichloroethene

Concen: 50.300 ug/l

RT: 4.342 min Scan# 406

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38




Tgt Ion: 96 Resp: 89116


Ion Ratio Lower Upper

96 100

61 162.3 128.6 192.8

98 60.6 51.7 77.5









#13
Acrolein
Concen: 192.843 ug/l
RT: 4.177 min Scan# 31
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 85169
Ion Ratio Lower Upper
56 100
55 71.1 57.2 85.8



#14
Allyl chloride
Concen: 43.643 ug/l
RT: 5.018 min Scan# 521
Delta R.T. -0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 41 Resp: 134288
Ion Ratio Lower Upper
41 100
39 73.1 59.0 88.6
76 39.4 29.8 44.6









#15
Acrylonitrile
Concen: 283.386 ug/l
RT: 5.718 min Scan# 640
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 53 Resp: 275307
Ion Ratio Lower Upper
53 100
52 82.1 66.2 99.4
51 35.2 28.8 43.2



#16
Acetone
Concen: 248.905 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 43 Resp: 249463
Ion Ratio Lower Upper
43 100
58 30.4 24.4 36.6









#17
Carbon Disulfide
Concen: 39.937 ug/l
RT: 4.706 min Scan# 41
Delta R.T. -0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 224021
Ion Ratio Lower Upper
76 100
78 9.4 7.3 10.9



#18
Methyl Acetate
Concen: 65.900 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 43 Resp: 143956
Ion Ratio Lower Upper
43 100
74 24.6 19.2 28.8









#19
Methyl tert-butyl Ether
Concen: 53.457 ug/l
RT: 5.794 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 73 Resp: 319663
Ion Ratio Lower Upper
73 100
57 23.3 17.7 26.5



#20
Methylene Chloride
Concen: 51.787 ug/l
RT: 5.271 min Scan# 564
Delta R.T. -0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 84 Resp: 105509
Ion Ratio Lower Upper
84 100
49 112.4 107.2 160.8
51 37.8 31.8 47.8
86 61.9 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 50.278 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 93324
Ion Ratio Lower Upper
96 100
61 126.8 108.4 162.6
98 59.2 49.8 74.8



#22
Diisopropyl ether
Concen: 52.949 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 45 Resp: 336994
Ion Ratio Lower Upper
45 100
43 52.6 42.3 63.5
87 26.8 21.8 32.6
59 12.8 9.6 14.4









#23
 Vinyl Acetate
 Concen: 252.258 ug/l
 RT: 6.606 min Scan# 791
 Delta R.T. 0.006 min
 Lab File: VN084428.D
 Acq: 21 Oct 2024 11:38




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 43 Resp: 1190800
 Ion Ratio Lower Upper
 43 100
 86 10.9 8.6 12.8



#24
 1,1-Dichloroethane
 Concen: 52.570 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. 0.000 min
 Lab File: VN084428.D
 Acq: 21 Oct 2024 11:38

Tgt Ion: 63 Resp: 187107
 Ion Ratio Lower Upper
 63 100
 98 6.2 3.4 10.2
 100 4.9 2.0 5.8









#25
2-Butanone
Concen: 269.013 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 367222
Ion Ratio Lower Upper
43 100
72 27.4 21.4 32.2



#26
2,2-Dichloropropane
Concen: 49.738 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 77 Resp: 159617
Ion Ratio Lower Upper
77 100
97 23.0 10.8 32.3





#27

cis-1,2-Dichloroethene

Concen: 52.728 ug/l

RT: 7.488 min Scan# 941

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion: 96 Resp: 118267

Ion Ratio Lower Upper

96 100

61 144.9 0.0 301.8

98 65.0 0.0 128.2



#28

Bromochloromethane

Concen: 52.836 ug/l

RT: 7.812 min Scan# 996

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38




Tgt Ion: 49 Resp: 83905


Ion Ratio Lower Upper

49 100

129 1.7 0.0 3.6

130 78.3 58.0 87.0





#29

Tetrahydrofuran

Concen: 283.258 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion: 42 Resp: 238619

Ion Ratio Lower Upper

42 100

72 45.2 34.6 51.8

71 41.3 32.2 48.4



#30

Chloroform

Concen: 54.476 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.000 min

Lab File: VN084428.D




Acq: 21 Oct 2024 11:38


Tgt Ion: 83 Resp: 201405

Ion Ratio Lower Upper

83 100


85 65.0 51.0 76.4








#31
Cyclohexane
Concen: 43.659 ug/l
RT: 8.259 min Scan# 1072
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050







Tgt Ion: 56 Resp: 150876
Ion Ratio Lower Upper
56 100
69 27.6 23.8 35.6
84 84.5 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 52.830 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 97 Resp: 175417
Ion Ratio Lower Upper
97 100
99 65.5 51.2 76.8
61 46.3 37.5 56.3





#33

1,2-Dichloroethane-d4

Concen: 53.365 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 65 Resp: 124506

Ion Ratio Lower Upper

65 100

67 53.5 0.0 102.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38




Tgt Ion: 114 Resp: 262800


Ion Ratio Lower Upper

114 100

63 22.4 0.0 43.8

88 16.4 0.0 31.6





#35

Dibromofluoromethane

Concen: 58.729 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion:113 Resp: 101751

Ion Ratio Lower Upper

113 100

111 102.8 83.3 124.9

192 17.8 13.5 20.3



#36

1,1-Dichloropropene

Concen: 52.021 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38




Tgt Ion: 75 Resp: 130914


Ion Ratio Lower Upper

75 100

110 34.9 16.9 50.7

77 31.6 24.7 37.1









#37
Ethyl Acetate
Concen: 53.172 ug/l
RT: 7.559 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 137280
Ion Ratio Lower Upper
43 100
61 14.4 10.7 16.1
70 11.2 8.8 13.2



#38
Carbon Tetrachloride
Concen: 54.202 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 117 Resp: 149628
Ion Ratio Lower Upper
117 100
119 95.2 78.7 118.1
121 28.3 24.4 36.6





#39

Methylcyclohexane

Concen: 48.955 ug/l

RT: 9.600 min Scan# 11300

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion: 83 Resp: 137117

Ion Ratio Lower Upper

83 100

55 74.6 62.8 94.2

98 49.4 38.8 58.2



#40

Benzene

Concen: 54.796 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VN084428.D




Acq: 21 Oct 2024 11:38


Tgt Ion: 78 Resp: 429647

Ion Ratio Lower Upper

78 100

77 23.3 19.1 28.7









#41
Methacrylonitrile
Concen: 56.424 ug/l
RT: 7.777 min Scan# 990
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:	41	Resp:	79086
Ion	Ratio	Lower	Upper
41	100		
39	56.4	43.3	64.9
67	75.1	56.1	84.1
52	30.9	24.5	36.7



#42
1,2-Dichloroethane
Concen: 54.240 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion:	62	Resp:	144178
Ion	Ratio	Lower	Upper
62	100		
98	9.6	0.0	18.2









#43
Isopropyl Acetate
Concen: 46.197 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 233616
Ion Ratio Lower Upper
43 100
61 26.8 20.4 30.6
87 13.4 10.3 15.5



#44
Trichloroethene
Concen: 53.784 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion:130 Resp: 98494
Ion Ratio Lower Upper
130 100
95 102.0 0.0 203.0





#45

1,2-Dichloropropane

Concen: 56.873 ug/l

RT: 9.618 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 63 Resp: 105490

Ion Ratio Lower Upper

63 100

65 32.8 24.6 37.0



#46

Dibromomethane

Concen: 59.706 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38




Tgt Ion: 93 Resp: 74735


Ion Ratio Lower Upper

93 100

95 80.7 65.0 97.4

174 93.5 71.8 107.8









#47
Bromodichloromethane
Concen: 57.235 ug/l
RT: 9.888 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 159166
Ion Ratio Lower Upper
83 100
85 61.4 51.5 77.3
127 8.9 5.7 8.5#



#48
Methyl methacrylate
Concen: 55.960 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 41 Resp: 112649
Ion Ratio Lower Upper
41 100
69 84.7 67.1 100.7
39 55.9 45.8 68.6





#49

1,4-Dioxane

Concen: 1259.594 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion: 88 Resp: 46681

Ion Ratio Lower Upper

88 100

43 28.2 25.3 37.9

58 71.6 56.4 84.6



#50

Toluene-d8

Concen: 59.014 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN084428.D




Acq: 21 Oct 2024 11:38


Tgt Ion: 98 Resp: 376165

Ion Ratio Lower Upper

98 100

100 64.8 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 308.066 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 757152
Ion Ratio Lower Upper
43 100
58 39.9 31.6 47.4



#52
Toluene
Concen: 56.231 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 92 Resp: 268859
Ion Ratio Lower Upper
92 100
91 170.0 137.3 205.9









#53
 t-1,3-Dichloropropene
 Concen: 55.712 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. 0.000 min
 Lab File: VN084428.D
 Acq: 21 Oct 2024 11:38




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 75 Resp: 157928
 Ion Ratio Lower Upper
 75 100
 77 32.5 25.4 38.0



#54
 cis-1,3-Dichloropropene
 Concen: 56.089 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. 0.000 min
 Lab File: VN084428.D
 Acq: 21 Oct 2024 11:38

Tgt Ion: 75 Resp: 171021
 Ion Ratio Lower Upper
 75 100
 77 30.9 24.2 36.2
 39 44.8 38.7 58.1









#55
1,1,2-Trichloroethane
Concen: 60.910 ug/l
RT: 11.018 min Scan# 1540
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050

Tgt Ion: 97 Resp: 104404




Ion	Ratio	Lower	Upper
97	100		
83	83.3	71.0	106.4
85	52.3	41.7	62.5
99	60.4	50.1	75.1




#56
Ethyl methacrylate
Concen: 58.283 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 69 Resp: 164383

Ion	Ratio	Lower	Upper
69	100		
41	63.2	53.3	79.9
39	34.8	28.8	43.2





#57

1,3-Dichloropropane

Concen: 57.670 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.006 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 76 Resp: 176631

Ion Ratio Lower Upper

76 100

78 32.5 25.8 38.6



#58

2-Chloroethyl Vinyl ether

Concen: 303.996 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN084428.D




Acq: 21 Oct 2024 11:38


Tgt Ion: 63 Resp: 397738

Ion Ratio Lower Upper

63 100

106 28.9 21.8 32.6









#59
2-Hexanone
Concen: 304.908 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 560104
Ion Ratio Lower Upper
43 100
58 56.4 27.3 81.8



#60
Dibromochloromethane
Concen: 60.458 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 129 Resp: 122440
Ion Ratio Lower Upper
129 100
127 75.8 38.3 114.9





#61

1,2-Dibromoethane

Concen: 57.274 ug/l

RT: 11.471 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion:107 Resp: 102008

Ion Ratio Lower Upper

107 100

109 94.4 75.4 113.0



#62

4-Bromofluorobenzene

Concen: 59.102 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38




Tgt Ion: 95 Resp: 137246


Ion Ratio Lower Upper

95 100

174 76.1 0.0 145.2

176 71.4 0.0 140.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion:117 Resp: 244635

Ion Ratio Lower Upper

117 100

82 56.4 47.2 70.8

119 30.2 25.4 38.0



#64

Tetrachloroethene

Concen: 52.523 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Tgt Ion:164 Resp: 89495




Ion Ratio Lower Upper


164 100

166 126.0 101.0 151.4

129 95.5 80.8 121.2

131 94.8 78.1 117.1









#65
Chlorobenzene
Concen: 53.283 ug/l
RT: 11.888 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:112 Resp: 289014
Ion Ratio Lower Upper
112 100
114 29.8 27.0 40.6



#66
1,1,1,2-Tetrachloroethane
Concen: 55.468 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion:131 Resp: 103631
Ion Ratio Lower Upper
131 100
133 93.7 47.9 143.7
119 65.0 32.8 98.3





#67

Ethyl Benzene

Concen: 52.799 ug/l

RT: 11.965 min Scan# 1701

Delta R.T. 0.006 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050

Tgt Ion: 91 Resp: 506469

Ion Ratio Lower Upper

91 100

106 31.0 24.5 36.7



#68

m/p-Xylenes

Concen: 111.301 ug/l

RT: 12.071 min Scan# 1720

Delta R.T. 0.000 min

Lab File: VN084428.D




Acq: 21 Oct 2024 11:38


Tgt Ion:106 Resp: 394932

Ion Ratio Lower Upper

106 100

91 206.8 167.1 250.7









#69
o-Xylene
Concen: 55.767 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:106 Resp: 187215
Ion Ratio Lower Upper
106 100
91 219.7 110.2 330.6



#70
Styrene
Concen: 58.246 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion:104 Resp: 331204
Ion Ratio Lower Upper
104 100
78 50.7 42.7 64.1
103 54.3 43.6 65.4





#71

Bromoform

Concen: 59.072 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion:173 Resp: 81253

Ion Ratio Lower Upper

173 100

175 47.1 23.5 70.6

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38




Tgt Ion:152 Resp: 127638


Ion Ratio Lower Upper

152 100

115 61.0 31.3 93.9

150 173.0 0.0 349.8









#73
Isopropylbenzene
Concen: 49.760 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:105 Resp: 484656
Ion Ratio Lower Upper
105 100
120 25.6 12.9 38.7



#74
N-ethyl acetate
Concen: 46.332 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 43 Resp: 204351
Ion Ratio Lower Upper
43 100
70 44.9 34.4 51.6
55 26.9 19.6 29.4
61 25.1 19.6 29.4









#75
1,1,2,2-Tetrachloroethane
Concen: 52.226 ug/l
RT: 12.935 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 152956
Ion Ratio Lower Upper
83 100
131 10.5 4.9 14.7
85 65.3 31.9 95.7



#76
1,2,3-Trichloropropane
Concen: 76.826 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. 0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 75 Resp: 200192
Ion Ratio Lower Upper
75 100
77 131.0 96.9 290.7









#77
Bromobenzene
Concen: 50.631 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.006 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:156 Resp: 118952
Ion Ratio Lower Upper
156 100
77 220.5 115.1 345.3
158 97.0 48.9 146.8



#78
n-propylbenzene
Concen: 51.101 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 91 Resp: 576662
Ion Ratio Lower Upper
91 100
120 22.6 10.8 32.4









#79
2-Chlorotoluene
Concen: 49.256 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 354683
Ion Ratio Lower Upper
91 100
126 32.7 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 52.460 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion:105 Resp: 416797
Ion Ratio Lower Upper
105 100
120 47.7 23.7 71.1









#81
trans-1,4-Dichloro-2-butene
Concen: 53.835 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 58513
Ion Ratio Lower Upper
75 100
53 86.4 78.0 117.0
89 40.2 35.5 53.3



#82
4-Chlorotoluene
Concen: 50.275 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion: 91 Resp: 364438
Ion Ratio Lower Upper
91 100
126 31.9 15.7 47.1









#83
 tert-Butylbenzene
 Concen: 49.862 ug/l
 RT: 13.435 min Scan# 1952
 Delta R.T. 0.000 min
 Lab File: VN084428.D
 Acq: 21 Oct 2024 11:38




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion:119 Resp: 352099
 Ion Ratio Lower Upper
 119 100
 91 68.1 34.1 102.3
 134 26.4 12.8 38.4



#84
 1,2,4-Trimethylbenzene
 Concen: 52.722 ug/l
 RT: 13.482 min Scan# 1960
 Delta R.T. 0.000 min
 Lab File: VN084428.D
 Acq: 21 Oct 2024 11:38

Tgt Ion:105 Resp: 421969
 Ion Ratio Lower Upper
 105 100
 120 44.9 22.3 66.8









#85
 sec-Butylbenzene
 Concen: 52.226 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. 0.000 min
 Lab File: VN084428.D
 Acq: 21 Oct 2024 11:38




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion:105 Resp: 492135
 Ion Ratio Lower Upper
 105 100
 134 19.2 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 52.906 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN084428.D
 Acq: 21 Oct 2024 11:38

Tgt Ion:119 Resp: 412249
 Ion Ratio Lower Upper
 119 100
 134 26.0 12.6 37.8
 91 24.7 12.4 37.2









#87
 1,3-Dichlorobenzene
 Concen: 51.493 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN084428.D
 Acq: 21 Oct 2024 11:38




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050


Tgt Ion:146 Resp: 226954
 Ion Ratio Lower Upper
 146 100
 111 42.6 21.8 65.3
 148 63.9 31.9 95.7



#88
 1,4-Dichlorobenzene
 Concen: 50.322 ug/l
 RT: 13.812 min Scan# 2016
 Delta R.T. 0.000 min
 Lab File: VN084428.D
 Acq: 21 Oct 2024 11:38

Tgt Ion:146 Resp: 223977
 Ion Ratio Lower Upper
 146 100
 111 41.7 21.2 63.6
 148 63.1 31.9 95.9





#89

n-Butylbenzene

Concen: 49.495 ug/l

RT: 14.053 min Scan# 2057

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion: 91 Resp: 349709

Ion Ratio Lower Upper

91 100

92 52.7 26.6 79.7

134 24.9 12.3 36.9



#90

Hexachloroethane

Concen: 48.204 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. 0.000 min

Lab File: VN084428.D




Acq: 21 Oct 2024 11:38


Tgt Ion: 117 Resp: 76242

Ion Ratio Lower Upper

117 100

201 69.3 34.4 103.2





#91

1,2-Dichlorobenzene

Concen: 50.991 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion:146 Resp: 220370

Ion Ratio Lower Upper

146 100

111 44.0 22.0 66.0

148 64.4 32.1 96.3



#92

1,2-Dibromo-3-Chloropropane

Concen: 49.657 ug/l

RT: 14.717 min Scan# 2170

Delta R.T. 0.000 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38




Tgt Ion: 75 Resp: 30011


Ion Ratio Lower Upper

75 100

155 79.3 38.1 114.5

157 102.7 49.6 149.0









#93
1,2,4-Trichlorobenzene
Concen: 51.373 ug/l
RT: 15.388 min Scan# 21
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050


Tgt Ion:180 Resp: 109356
Ion Ratio Lower Upper
180 100
182 94.6 47.0 141.0
145 32.7 16.4 49.4



#94
Hexachlorobutadiene
Concen: 47.838 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. 0.000 min
Lab File: VN084428.D
Acq: 21 Oct 2024 11:38

Tgt Ion:225 Resp: 50783
Ion Ratio Lower Upper
225 100
223 60.9 32.4 97.2
227 62.5 32.4 97.0





#95

Naphthalene

Concen: 51.214 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. 0.006 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050





Tgt Ion:128 Resp: 361412

Ion Ratio Lower Upper

128 100

127 13.1 10.5 15.7

129 11.2 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 52.021 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN084428.D

Acq: 21 Oct 2024 11:38




Tgt Ion:180 Resp: 110846

Ion Ratio Lower Upper

180 100

182 95.2 48.2 144.6

145 34.9 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084433.D
 Acq On : 21 Oct 2024 13:48
 Operator : JC\MD
 Sample : VN1021WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBS01

Quant Time: Oct 22 01:32:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	189913	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	322919	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	284161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	138325	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	142332	50.547	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.100%
35) Dibromofluoromethane	8.165	113	115508	54.258	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.520%
50) Toluene-d8	10.565	98	422532	53.947	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.900%
62) 4-Bromofluorobenzene	12.847	95	149865	52.521	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.040%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.130	85	36261	16.142	ug/l	95
3) Chloromethane	2.359	50	41235	16.076	ug/l	96
4) Vinyl Chloride	2.512	62	41507	16.706	ug/l	100
5) Bromomethane	2.953	94	26963	16.453	ug/l	98
6) Chloroethane	3.124	64	26726	15.020	ug/l	99
7) Trichlorofluoromethane	3.500	101	70146	18.131	ug/l	98
8) Diethyl Ether	3.965	74	24127	16.814	ug/l	93
9) 1,1,2-Trichlorotrifluo...	4.377	101	40161	18.116	ug/l	98
10) Methyl Iodide	4.589	142	48908	17.131	ug/l	96
11) Tert butyl alcohol	5.518	59	36415	78.429	ug/l #	94
12) 1,1-Dichloroethene	4.342	96	37659	17.612	ug/l	97
13) Acrolein	4.183	56	26602	49.907	ug/l	98
14) Allyl chloride	5.024	41	58352	15.713	ug/l	98
15) Acrylonitrile	5.718	53	109279	93.202	ug/l	99
16) Acetone	4.430	43	105837	87.496	ug/l	96
17) Carbon Disulfide	4.712	76	96169	14.205	ug/l	100
18) Methyl Acetate	5.030	43	59251	22.474	ug/l	99
19) Methyl tert-butyl Ether	5.794	73	130984	18.149	ug/l	98
20) Methylene Chloride	5.271	84	45743	18.603	ug/l	91
21) trans-1,2-Dichloroethene	5.789	96	38888	17.359	ug/l	95
22) Diisopropyl ether	6.671	45	141732	18.451	ug/l	98
23) Vinyl Acetate	6.600	43	468614	82.252	ug/l	98
24) 1,1-Dichloroethane	6.571	63	82481	19.201	ug/l	97
25) 2-Butanone	7.483	43	146074	88.663	ug/l	98
26) 2,2-Dichloropropane	7.488	77	71005	18.333	ug/l	98
27) cis-1,2-Dichloroethene	7.494	96	49184	18.169	ug/l	98
28) Bromochloromethane	7.812	49	37550	19.592	ug/l	97
29) Tetrahydrofuran	7.841	42	90421	88.935	ug/l	98
30) Chloroform	7.971	83	84755	18.995	ug/l	100
31) Cyclohexane	8.259	56	66762	16.007	ug/l #	97
32) 1,1,1-Trichloroethane	8.171	97	74940	18.700	ug/l	97
36) 1,1-Dichloropropene	8.371	75	56161	18.162	ug/l	97
37) Ethyl Acetate	7.559	43	57455	18.111	ug/l	99
38) Carbon Tetrachloride	8.365	117	64152	18.912	ug/l	94
39) Methylcyclohexane	9.600	83	55643	16.168	ug/l	96
40) Benzene	8.606	78	183073	19.002	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084433.D
 Acq On : 21 Oct 2024 13:48
 Operator : JC\MD
 Sample : VN1021WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBS01

Quant Time: Oct 22 01:32:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	29860	17.338	ug/l	93
42) 1,2-Dichloroethane	8.671	62	61753	18.907	ug/l	99
43) Isopropyl Acetate	8.688	43	94050	15.136	ug/l	98
44) Trichloroethene	9.347	130	43127	19.166	ug/l	98
45) 1,2-Dichloropropane	9.624	63	44771	19.644	ug/l	98
46) Dibromomethane	9.706	93	31378	20.401	ug/l	96
47) Bromodichloromethane	9.888	83	66599	19.490	ug/l	96
48) Methyl methacrylate	9.682	41	43352	17.526	ug/l	98
49) 1,4-Dioxane	9.694	88	16997	373.245	ug/l	97
51) 4-Methyl-2-Pentanone	10.441	43	292338	96.801	ug/l	100
52) Toluene	10.629	92	114864	19.551	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	64356	18.476	ug/l	98
54) cis-1,3-Dichloropropene	10.312	75	70777	18.891	ug/l	96
55) 1,1,2-Trichloroethane	11.012	97	43053	20.441	ug/l	97
56) Ethyl methacrylate	10.871	69	65255	18.829	ug/l	96
57) 1,3-Dichloropropane	11.165	76	74113	19.693	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	139755	86.930	ug/l	98
59) 2-Hexanone	11.194	43	216944	96.112	ug/l	99
60) Dibromochloromethane	11.359	129	50030	20.104	ug/l	99
61) 1,2-Dibromoethane	11.465	107	41569	18.994	ug/l	98
64) Tetrachloroethene	11.100	164	37454	18.924	ug/l	95
65) Chlorobenzene	11.888	112	121076	19.217	ug/l	94
66) 1,1,1,2-Tetrachloroethane	11.959	131	43109	19.865	ug/l	99
67) Ethyl Benzene	11.965	91	205793	18.470	ug/l	100
68) m/p-Xylenes	12.070	106	158675	38.498	ug/l	100
69) o-Xylene	12.400	106	78180	20.049	ug/l	96
70) Styrene	12.412	104	129069	19.541	ug/l	99
71) Bromoform	12.576	173	31202	19.529	ug/l #	95
73) Isopropylbenzene	12.694	105	191996	18.189	ug/l	98
74) N-amyl acetate	12.494	43	79653	16.664	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.935	83	61726	19.448	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	79646	28.204	ug/l	59
77) Bromobenzene	12.982	156	48284	18.964	ug/l	95
78) n-propylbenzene	13.035	91	231176	18.903	ug/l	99
79) 2-Chlorotoluene	13.123	91	145730	18.674	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	168330	19.550	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	19953	16.940	ug/l	98
82) 4-Chlorotoluene	13.217	91	149766	19.064	ug/l	100
83) tert-Butylbenzene	13.435	119	141261	18.459	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	169548	19.547	ug/l	99
85) sec-Butylbenzene	13.617	105	195385	19.133	ug/l	99
86) p-Isopropyltoluene	13.729	119	161505	19.125	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	88692	18.569	ug/l	100
88) 1,4-Dichlorobenzene	13.812	146	90608	18.784	ug/l	99
89) n-Butylbenzene	14.053	91	132089	17.250	ug/l	99
90) Hexachloroethane	14.329	117	31571	18.419	ug/l	97
91) 1,2-Dichlorobenzene	14.106	146	85799	18.319	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	11177	17.065	ug/l	98
93) 1,2,4-Trichlorobenzene	15.394	180	40526	17.567	ug/l	100
94) Hexachlorobutadiene	15.500	225	19271	16.751	ug/l	98
95) Naphthalene	15.641	128	119595	15.638	ug/l	100
96) 1,2,3-Trichlorobenzene	15.835	180	39826	17.247	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084433.D
Acq On : 21 Oct 2024 13:48
Operator : JC\MD
Sample : VN1021WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01

Quant Time: Oct 22 01:32:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084433.D
 Acq On : 21 Oct 2024 13:48
 Operator : JC\MD
 Sample : VN1021WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBS01

Quant Time: Oct 22 01:32:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion:168 Resp: 189913
Ion Ratio Lower Upper
168 100
99 61.2 54.2 81.2



#2
Dichlorodifluoromethane
Concen: 16.142 ug/l
RT: 2.130 min Scan# 30
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 85 Resp: 36261
Ion Ratio Lower Upper
85 100
87 32.5 17.6 52.9









#3
Chloromethane
Concen: 16.076 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 50 Resp: 41235
Ion Ratio Lower Upper
50 100
52 31.1 26.8 40.2



#4
Vinyl Chloride
Concen: 16.706 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 62 Resp: 41507
Ion Ratio Lower Upper
62 100
64 32.7 26.2 39.4









#5
Bromomethane
Concen: 16.453 ug/l
RT: 2.953 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 94 Resp: 26963
Ion Ratio Lower Upper
94 100
96 89.6 72.9 109.3



#6
Chloroethane
Concen: 15.020 ug/l
RT: 3.124 min Scan# 199
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 64 Resp: 26726
Ion Ratio Lower Upper
64 100
66 30.2 23.5 35.3









#7
Trichlorofluoromethane
Concen: 18.131 ug/l
RT: 3.500 min Scan# 21
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion:101 Resp: 70146
Ion Ratio Lower Upper
101 100
103 66.6 51.8 77.6



#8
Diethyl Ether
Concen: 16.814 ug/l
RT: 3.965 min Scan# 342
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 74 Resp: 24127
Ion Ratio Lower Upper
74 100
45 102.5 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.116 ug/l
RT: 4.377 min Scan# 411
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion:101 Resp: 40161
Ion Ratio Lower Upper
101 100
85 46.2 36.2 54.4
151 76.3 59.3 88.9



#10
Methyl Iodide
Concen: 17.131 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion:142 Resp: 48908
Ion Ratio Lower Upper
142 100
127 47.6 35.3 52.9
141 13.5 11.0 16.6









#11
Tert butyl alcohol
Concen: 78.429 ug/l
RT: 5.518 min Scan# 607
Delta R.T. -0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 59 Resp: 36415
Ion Ratio Lower Upper
59 100
57 8.1 8.2 12.2#



#12
1,1-Dichloroethene
Concen: 17.612 ug/l
RT: 4.342 min Scan# 406
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 96 Resp: 37659
Ion Ratio Lower Upper
96 100
61 161.0 128.6 192.8
98 57.8 51.7 77.5









#13
Acrolein
Concen: 49.907 ug/l
RT: 4.183 min Scan# 31
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 56 Resp: 26602
Ion Ratio Lower Upper
56 100
55 72.9 57.2 85.8



#14
Allyl chloride
Concen: 15.713 ug/l
RT: 5.024 min Scan# 522
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 41 Resp: 58352
Ion Ratio Lower Upper
41 100
39 70.9 59.0 88.6
76 37.4 29.8 44.6









#15
Acrylonitrile
Concen: 93.202 ug/l
RT: 5.718 min Scan# 640
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01

Tgt Ion: 53 Resp: 109279




Ion	Ratio	Lower	Upper
53	100		
52	81.7	66.2	99.4
51	35.3	28.8	43.2




#16
Acetone
Concen: 87.496 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 43 Resp: 105837

Ion	Ratio	Lower	Upper
43	100		
58	32.9	24.4	36.6









#17
Carbon Disulfide
Concen: 14.205 ug/l
RT: 4.712 min Scan# 41
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 76 Resp: 96169
Ion Ratio Lower Upper
76 100
78 9.0 7.3 10.9



#18
Methyl Acetate
Concen: 22.474 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.012 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 43 Resp: 59251
Ion Ratio Lower Upper
43 100
74 23.6 19.2 28.8









#19
Methyl tert-butyl Ether
Concen: 18.149 ug/l
RT: 5.794 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 73 Resp: 130984
Ion Ratio Lower Upper
73 100
57 23.1 17.7 26.5



#20
Methylene Chloride
Concen: 18.603 ug/l
RT: 5.271 min Scan# 564
Delta R.T. -0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 84 Resp: 45743
Ion Ratio Lower Upper
84 100
49 116.0 107.2 160.8
51 39.6 31.8 47.8
86 63.5 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 17.359 ug/l
RT: 5.789 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 96 Resp: 38888
Ion Ratio Lower Upper
96 100
61 130.3 108.4 162.6
98 67.0 49.8 74.8



#22
Diisopropyl ether
Concen: 18.451 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 45 Resp: 141732
Ion Ratio Lower Upper
45 100
43 51.9 42.3 63.5
87 28.3 21.8 32.6
59 12.3 9.6 14.4









#23
 Vinyl Acetate
 Concen: 82.252 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. -0.000 min
 Lab File: VN084433.D
 Acq: 21 Oct 2024 13:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBS01


Tgt Ion: 43 Resp: 468614
 Ion Ratio Lower Upper
 43 100
 86 11.3 8.6 12.8



#24
 1,1-Dichloroethane
 Concen: 19.201 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. -0.000 min
 Lab File: VN084433.D
 Acq: 21 Oct 2024 13:48

Tgt Ion: 63 Resp: 82481
 Ion Ratio Lower Upper
 63 100
 98 8.3 3.4 10.2
 100 4.4 2.0 5.8









#25
2-Butanone
Concen: 88.663 ug/l
RT: 7.483 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 43 Resp: 146074
Ion Ratio Lower Upper
43 100
72 27.7 21.4 32.2



#26
2,2-Dichloropropane
Concen: 18.333 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 77 Resp: 71005
Ion Ratio Lower Upper
77 100
97 22.6 10.8 32.3









#27
 cis-1,2-Dichloroethene
 Concen: 18.169 ug/l
 RT: 7.494 min Scan# 941
 Delta R.T. 0.006 min
 Lab File: VN084433.D
 Acq: 21 Oct 2024 13:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBS01


Tgt Ion: 96 Resp: 49184
 Ion Ratio Lower Upper
 96 100
 61 147.2 0.0 301.8
 98 63.3 0.0 128.2



#28
 Bromochloromethane
 Concen: 19.592 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. -0.000 min
 Lab File: VN084433.D
 Acq: 21 Oct 2024 13:48

Tgt Ion: 49 Resp: 37550
 Ion Ratio Lower Upper
 49 100
 129 0.9 0.0 3.6
 130 74.6 58.0 87.0





#29

Tetrahydrofuran

Concen: 88.935 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01





Tgt Ion: 42 Resp: 90421

Ion Ratio Lower Upper

42 100

72 44.5 34.6 51.8

71 41.0 32.2 48.4



#30

Chloroform

Concen: 18.995 ug/l

RT: 7.971 min Scan# 1023

Delta R.T. 0.006 min

Lab File: VN084433.D




Acq: 21 Oct 2024 13:48


Tgt Ion: 83 Resp: 84755

Ion Ratio Lower Upper

83 100

85 64.0 51.0 76.4









#31
Cyclohexane
Concen: 16.007 ug/l
RT: 8.259 min Scan# 1071
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 56 Resp: 66762
Ion Ratio Lower Upper
56 100
69 35.7 23.8 35.6#
84 86.0 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 18.700 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 97 Resp: 74940
Ion Ratio Lower Upper
97 100
99 62.8 51.2 76.8
61 43.8 37.5 56.3





#33

1,2-Dichloroethane-d4

Concen: 50.547 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :





VN1021WBS01

Tgt Ion: 65 Resp: 142332

Ion Ratio Lower Upper

65 100

67 52.7 0.0 102.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48




Tgt Ion: 114 Resp: 322919


Ion Ratio Lower Upper

114 100

63 20.0 0.0 43.8

88 16.5 0.0 31.6





#35

Dibromofluoromethane

Concen: 54.258 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01





Tgt Ion:113 Resp: 115508

Ion Ratio Lower Upper

113 100

111 104.0 83.3 124.9

192 18.4 13.5 20.3



#36

1,1-Dichloropropene

Concen: 18.162 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48




Tgt Ion: 75 Resp: 56161


Ion Ratio Lower Upper

75 100

110 35.3 16.9 50.7

77 29.3 24.7 37.1









#37
Ethyl Acetate
Concen: 18.111 ug/l
RT: 7.559 min Scan# 91
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 43 Resp: 57455
Ion Ratio Lower Upper
43 100
61 12.4 10.7 16.1
70 11.1 8.8 13.2



#38
Carbon Tetrachloride
Concen: 18.912 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 117 Resp: 64152
Ion Ratio Lower Upper
117 100
119 92.6 78.7 118.1
121 27.9 24.4 36.6





#39

Methylcyclohexane

Concen: 16.168 ug/l

RT: 9.600 min Scan# 11300

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01





Tgt Ion: 83 Resp: 55643

Ion Ratio Lower Upper

83 100

55 83.6 62.8 94.2

98 49.3 38.8 58.2



#40

Benzene

Concen: 19.002 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN084433.D




Acq: 21 Oct 2024 13:48


Tgt Ion: 78 Resp: 183073

Ion Ratio Lower Upper

78 100

77 23.6 19.1 28.7









#41
Methacrylonitrile
Concen: 17.338 ug/l
RT: 7.777 min Scan# 990
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion:	41	Resp:	29860
Ion	Ratio	Lower	Upper
41	100		
39	59.2	43.3	64.9
67	75.6	56.1	84.1
52	34.3	24.5	36.7



#42
1,2-Dichloroethane
Concen: 18.907 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion:	62	Resp:	61753
Ion	Ratio	Lower	Upper
62	100		
98	9.6	0.0	18.2









#43
Isopropyl Acetate
Concen: 15.136 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 43	Resp:	94050
Ion Ratio	Lower	Upper
43	100	
61	27.2	20.4 30.6
87	13.1	10.3 15.5



#44
Trichloroethene
Concen: 19.166 ug/l
RT: 9.347 min Scan# 1257
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 130	Resp:	43127
Ion Ratio	Lower	Upper
130	100	
95	99.7	0.0 203.0





#45

1,2-Dichloropropane

Concen: 19.644 ug/l

RT: 9.624 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :





VN1021WBS01

Tgt Ion: 63 Resp: 44771

Ion Ratio Lower Upper

63 100

65 32.0 24.6 37.0



#46

Dibromomethane

Concen: 20.401 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48




Tgt Ion: 93 Resp: 31378


Ion Ratio Lower Upper

93 100

95 83.6 65.0 97.4

174 94.1 71.8 107.8









#47
Bromodichloromethane
Concen: 19.490 ug/l
RT: 9.888 min Scan# 1349
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 83 Resp: 66599
Ion Ratio Lower Upper
83 100
85 61.4 51.5 77.3
127 7.9 5.7 8.5



#48
Methyl methacrylate
Concen: 17.526 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 41 Resp: 43352
Ion Ratio Lower Upper
41 100
69 86.0 67.1 100.7
39 56.5 45.8 68.6





#49

1,4-Dioxane

Concen: 373.245 ug/l

RT: 9.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01





Tgt Ion: 88 Resp: 16997

Ion Ratio Lower Upper

88 100

43 28.7 25.3 37.9

58 72.3 56.4 84.6



#50

Toluene-d8

Concen: 53.947 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN084433.D




Acq: 21 Oct 2024 13:48


Tgt Ion: 98 Resp: 422532

Ion Ratio Lower Upper

98 100

100 65.3 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 96.801 ug/l
RT: 10.441 min Scan# 1443
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 43 Resp: 292338
Ion Ratio Lower Upper
43 100
58 39.7 31.6 47.4



#52
Toluene
Concen: 19.551 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 92 Resp: 114864
Ion Ratio Lower Upper
92 100
91 170.2 137.3 205.9









#53
 t-1,3-Dichloropropene
 Concen: 18.476 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. -0.000 min
 Lab File: VN084433.D
 Acq: 21 Oct 2024 13:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBS01


Tgt Ion: 75 Resp: 64356
 Ion Ratio Lower Upper
 75 100
 77 30.8 25.4 38.0



#54
 cis-1,3-Dichloropropene
 Concen: 18.891 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. -0.000 min
 Lab File: VN084433.D
 Acq: 21 Oct 2024 13:48

Tgt Ion: 75 Resp: 70777
 Ion Ratio Lower Upper
 75 100
 77 33.2 24.2 36.2
 39 46.1 38.7 58.1





#55

1,1,2-Trichloroethane

Concen: 20.441 ug/l

RT: 11.012 min Scan# 1540

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01

Tgt Ion: 97 Resp: 43053





Ion Ratio Lower Upper

97 100

83 90.4 71.0 106.4

85 55.6 41.7 62.5

99 60.6 50.1 75.1



#56

Ethyl methacrylate

Concen: 18.829 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48




Tgt Ion: 69 Resp: 65255


Ion Ratio Lower Upper

69 100

41 63.6 53.3 79.9

39 33.9 28.8 43.2









#57
1,3-Dichloropropane
Concen: 19.693 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 76 Resp: 74113
Ion Ratio Lower Upper
76 100
78 32.5 25.8 38.6



#58
2-Chloroethyl Vinyl ether
Concen: 86.930 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 63 Resp: 139755
Ion Ratio Lower Upper
63 100
106 28.3 21.8 32.6









#59
2-Hexanone
Concen: 96.112 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 43 Resp: 216944
Ion Ratio Lower Upper
43 100
58 55.3 27.3 81.8



#60
Dibromochloromethane
Concen: 20.104 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 129 Resp: 50030
Ion Ratio Lower Upper
129 100
127 77.4 38.3 114.9









#61
1,2-Dibromoethane
Concen: 18.994 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion:107 Resp: 41569
Ion Ratio Lower Upper
107 100
109 96.3 75.4 113.0



#62
4-Bromofluorobenzene
Concen: 52.521 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 95 Resp: 149865
Ion Ratio Lower Upper
95 100
174 74.8 0.0 145.2
176 71.6 0.0 140.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion:117 Resp: 284161
Ion Ratio Lower Upper
117 100
82 58.2 47.2 70.8
119 31.9 25.4 38.0



#64
Tetrachloroethene
Concen: 18.924 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion:164 Resp: 37454
Ion Ratio Lower Upper
164 100
166 129.9 101.0 151.4
129 111.5 80.8 121.2
131 94.8 78.1 117.1









#65
Chlorobenzene
Concen: 19.217 ug/l
RT: 11.888 min Scan# 112
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion:112 Resp: 121076
Ion Ratio Lower Upper
112 100
114 30.2 27.0 40.6



#66
1,1,1,2-Tetrachloroethane
Concen: 19.865 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion:131 Resp: 43109
Ion Ratio Lower Upper
131 100
133 94.3 47.9 143.7
119 64.6 32.8 98.3









#67
Ethyl Benzene
Concen: 18.470 ug/l
RT: 11.965 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 91 Resp: 205793
Ion Ratio Lower Upper
91 100
106 30.8 24.5 36.7



#68
m/p-Xylenes
Concen: 38.498 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion:106 Resp: 158675
Ion Ratio Lower Upper
106 100
91 208.7 167.1 250.7









#69
o-Xylene
Concen: 20.049 ug/l
RT: 12.400 min Scan# 1775
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion:106 Resp: 78180
Ion Ratio Lower Upper
106 100
91 213.5 110.2 330.6



#70
Styrene
Concen: 19.541 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion:104 Resp: 129069
Ion Ratio Lower Upper
104 100
78 51.8 42.7 64.1
103 55.0 43.6 65.4









#71
Bromoform
Concen: 19.529 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion:173 Resp: 31202
Ion Ratio Lower Upper
173 100
175 50.2 23.5 70.6
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion:152 Resp: 138325
Ion Ratio Lower Upper
152 100
115 62.4 31.3 93.9
150 161.5 0.0 349.8









#73
Isopropylbenzene
Concen: 18.189 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion:105 Resp: 191996
Ion Ratio Lower Upper
105 100
120 26.6 12.9 38.7



#74
N-ethyl acetate
Concen: 16.664 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 43 Resp: 79653
Ion Ratio Lower Upper
43 100
70 44.0 34.4 51.6
55 28.0 19.6 29.4
61 24.3 19.6 29.4





#75

1,1,2,2-Tetrachloroethane

Concen: 19.448 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01





Tgt Ion: 83 Resp: 61726

Ion Ratio Lower Upper

83 100

131 10.5 4.9 14.7

85 63.5 31.9 95.7



#76

1,2,3-Trichloropropane

Concen: 28.204 ug/l

RT: 12.994 min Scan# 1877

Delta R.T. 0.006 min

Lab File: VN084433.D




Acq: 21 Oct 2024 13:48


Tgt Ion: 75 Resp: 79646

Ion Ratio Lower Upper

75 100

77 132.9 96.9 290.7





#77

Bromobenzene

Concen: 18.964 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01





Tgt Ion:156 Resp: 48284

Ion Ratio Lower Upper

156 100

77 219.2 115.1 345.3

158 96.4 48.9 146.8



#78

n-propylbenzene

Concen: 18.903 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. -0.000 min

Lab File: VN084433.D




Acq: 21 Oct 2024 13:48


Tgt Ion: 91 Resp: 231176

Ion Ratio Lower Upper

91 100

120 22.3 10.8 32.4









#79
2-Chlorotoluene
Concen: 18.674 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 91 Resp: 145730
Ion Ratio Lower Upper
91 100
126 31.7 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 19.550 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion:105 Resp: 168330
Ion Ratio Lower Upper
105 100
120 47.3 23.7 71.1









#81
trans-1,4-Dichloro-2-butene
Concen: 16.940 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 75 Resp: 19953
Ion Ratio Lower Upper
75 100
53 96.1 78.0 117.0
89 46.4 35.5 53.3



#82
4-Chlorotoluene
Concen: 19.064 ug/l
RT: 13.217 min Scan# 1915
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 91 Resp: 149766
Ion Ratio Lower Upper
91 100
126 31.4 15.7 47.1









#83
tert-Butylbenzene
Concen: 18.459 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion:119 Resp: 141261
Ion Ratio Lower Upper
119 100
91 67.1 34.1 102.3
134 25.8 12.8 38.4



#84
1,2,4-Trimethylbenzene
Concen: 19.547 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion:105 Resp: 169548
Ion Ratio Lower Upper
105 100
120 43.7 22.3 66.8









#85
 sec-Butylbenzene
 Concen: 19.133 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN084433.D
 Acq: 21 Oct 2024 13:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBS01


Tgt Ion:105 Resp: 195385
 Ion Ratio Lower Upper
 105 100
 134 19.2 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 19.125 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN084433.D
 Acq: 21 Oct 2024 13:48

Tgt Ion:119 Resp: 161505
 Ion Ratio Lower Upper
 119 100
 134 24.2 12.6 37.8
 91 25.3 12.4 37.2









#87
 1,3-Dichlorobenzene
 Concen: 18.569 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN084433.D
 Acq: 21 Oct 2024 13:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBS01


Tgt Ion:146 Resp: 88692
 Ion Ratio Lower Upper
 146 100
 111 43.5 21.8 65.3
 148 64.2 31.9 95.7



#88
 1,4-Dichlorobenzene
 Concen: 18.784 ug/l
 RT: 13.812 min Scan# 2016
 Delta R.T. -0.000 min
 Lab File: VN084433.D
 Acq: 21 Oct 2024 13:48

Tgt Ion:146 Resp: 90608
 Ion Ratio Lower Upper
 146 100
 111 41.4 21.2 63.6
 148 63.1 31.9 95.9









#89
n-Butylbenzene
Concen: 17.250 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01


Tgt Ion: 91 Resp: 132089
Ion Ratio Lower Upper
91 100
92 51.9 26.6 79.7
134 24.5 12.3 36.9



#90
Hexachloroethane
Concen: 18.419 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion: 117 Resp: 31571
Ion Ratio Lower Upper
117 100
201 66.1 34.4 103.2





#91

1,2-Dichlorobenzene

Concen: 18.319 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBS01





Tgt Ion:146 Resp: 85799

Ion Ratio Lower Upper

146 100

111 44.9 22.0 66.0

148 64.6 32.1 96.3



#92

1,2-Dibromo-3-Chloropropane

Concen: 17.065 ug/l

RT: 14.717 min Scan# 2170

Delta R.T. -0.000 min

Lab File: VN084433.D

Acq: 21 Oct 2024 13:48




Tgt Ion: 75 Resp: 11177


Ion Ratio Lower Upper

75 100

155 77.7 38.1 114.5

157 96.7 49.6 149.0









#93
 1,2,4-Trichlorobenzene
 Concen: 17.567 ug/l
 RT: 15.394 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: VN084433.D
 Acq: 21 Oct 2024 13:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBS01


Tgt Ion:180 Resp: 40526
 Ion Ratio Lower Upper
 180 100
 182 94.0 47.0 141.0
 145 32.8 16.4 49.4



#94
 Hexachlorobutadiene
 Concen: 16.751 ug/l
 RT: 15.500 min Scan# 2303
 Delta R.T. -0.000 min
 Lab File: VN084433.D
 Acq: 21 Oct 2024 13:48

Tgt Ion:225 Resp: 19271
 Ion Ratio Lower Upper
 225 100
 223 62.3 32.4 97.2
 227 64.9 32.4 97.0









#95
Naphthalene
Concen: 15.638 ug/l
RT: 15.641 min Scan# 21
Delta R.T. 0.006 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBS01

Tgt Ion:128 Resp: 119595
Ion Ratio Lower Upper
128 100
127 13.1 10.5 15.7
129 10.9 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 17.247 ug/l
RT: 15.835 min Scan# 2360
Delta R.T. -0.000 min
Lab File: VN084433.D
Acq: 21 Oct 2024 13:48

Tgt Ion:180 Resp: 39826
Ion Ratio Lower Upper
180 100
182 94.7 48.2 144.6
145 32.6 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084434.D
 Acq On : 21 Oct 2024 14:22
 Operator : JC\MD
 Sample : VN1021WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBSD01

Quant Time: Oct 22 01:33:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	163086	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	278516	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	255670	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	124118	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	123457	51.056	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.120%	
35) Dibromofluoromethane	8.165	113	97003	52.829	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.660%	
50) Toluene-d8	10.565	98	350013	51.812	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.620%	
62) 4-Bromofluorobenzene	12.847	95	129081	52.449	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.900%	

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	2.124	85	32472	16.833	ug/l	93
3) Chloromethane	2.359	50	35804	16.255	ug/l	98
4) Vinyl Chloride	2.512	62	36441	17.080	ug/l	99
5) Bromomethane	2.959	94	23913	16.992	ug/l	99
6) Chloroethane	3.118	64	24475	16.017	ug/l	96
7) Trichlorofluoromethane	3.501	101	63283	19.047	ug/l	90
8) Diethyl Ether	3.959	74	23522	19.089	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.371	101	36689	19.273	ug/l	98
10) Methyl Iodide	4.589	142	44167	18.016	ug/l	99
11) Tert butyl alcohol	5.518	59	35053	87.914	ug/l	97
12) 1,1-Dichloroethene	4.342	96	32588	17.747	ug/l	96
13) Acrolein	4.189	56	24397	53.300	ug/l	98
14) Allyl chloride	5.024	41	53083	16.646	ug/l	99
15) Acrylonitrile	5.718	53	102471	101.771	ug/l	97
16) Acetone	4.430	43	95163	91.613	ug/l	94
17) Carbon Disulfide	4.712	76	84276	14.496	ug/l	96
18) Methyl Acetate	5.030	43	54347	24.005	ug/l	99
19) Methyl tert-butyl Ether	5.789	73	122872	19.826	ug/l	98
20) Methylene Chloride	5.271	84	42636	20.192	ug/l	84
21) trans-1,2-Dichloroethene	5.783	96	35388	18.395	ug/l	95
22) Diisopropyl ether	6.671	45	130999	19.859	ug/l	99
23) Vinyl Acetate	6.606	43	504977	103.215	ug/l	99
24) 1,1-Dichloroethane	6.571	63	72128	19.553	ug/l	96
25) 2-Butanone	7.483	43	136944	96.795	ug/l	97
26) 2,2-Dichloropropane	7.489	77	64755	19.469	ug/l	100
27) cis-1,2-Dichloroethene	7.489	96	44686	19.222	ug/l	98
28) Bromochloromethane	7.812	49	35788	21.744	ug/l	99
29) Tetrahydrofuran	7.841	42	85661	98.113	ug/l	97
30) Chloroform	7.965	83	76845	20.055	ug/l	97
31) Cyclohexane	8.253	56	58106	16.223	ug/l	97
32) 1,1,1-Trichloroethane	8.165	97	68247	19.832	ug/l	96
36) 1,1-Dichloropropene	8.371	75	49846	18.689	ug/l	99
37) Ethyl Acetate	7.559	43	52425	19.160	ug/l	99
38) Carbon Tetrachloride	8.359	117	56628	19.356	ug/l	96
39) Methylcyclohexane	9.600	83	50490	17.009	ug/l	98
40) Benzene	8.606	78	162968	19.612	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084434.D
 Acq On : 21 Oct 2024 14:22
 Operator : JC\MD
 Sample : VN1021WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBSD01

Quant Time: Oct 22 01:33:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	29826	20.079	ug/l	97
42) 1,2-Dichloroethane	8.671	62	58026	20.598	ug/l	99
43) Isopropyl Acetate	8.688	43	88049	16.429	ug/l	97
44) Trichloroethene	9.353	130	37926	19.541	ug/l	97
45) 1,2-Dichloropropane	9.624	63	41049	20.882	ug/l	96
46) Dibromomethane	9.712	93	29187	22.002	ug/l	99
47) Bromodichloromethane	9.888	83	61715	20.940	ug/l	98
48) Methyl methacrylate	9.682	41	40695	19.075	ug/l	97
49) 1,4-Dioxane	9.694	88	16333	415.845	ug/l	97
51) 4-Methyl-2-Pentanone	10.447	43	271816	104.354	ug/l	99
52) Toluene	10.629	92	102809	20.289	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	59188	19.701	ug/l	96
54) cis-1,3-Dichloropropene	10.312	75	64640	20.003	ug/l	94
55) 1,1,2-Trichloroethane	11.012	97	40810	22.465	ug/l	96
56) Ethyl methacrylate	10.871	69	59539	19.919	ug/l	98
57) 1,3-Dichloropropane	11.165	76	68931	21.236	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.159	63	141699	102.191	ug/l	97
59) 2-Hexanone	11.194	43	198917	102.176	ug/l	99
60) Dibromochloromethane	11.359	129	46724	21.769	ug/l	98
61) 1,2-Dibromoethane	11.471	107	39150	20.741	ug/l	100
64) Tetrachloroethene	11.106	164	34348	19.288	ug/l	94
65) Chlorobenzene	11.888	112	110738	19.535	ug/l	95
66) 1,1,1,2-Tetrachloroethane	11.959	131	39189	20.071	ug/l	99
67) Ethyl Benzene	11.965	91	184057	18.360	ug/l	99
68) m/p-Xylenes	12.071	106	144045	38.843	ug/l	98
69) o-Xylene	12.400	106	66105	18.841	ug/l	93
70) Styrene	12.412	104	117753	19.814	ug/l	99
71) Bromoform	12.576	173	29825	20.747	ug/l #	99
73) Isopropylbenzene	12.694	105	173820	18.352	ug/l	99
74) N-amyl acetate	12.494	43	73994	17.252	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.935	83	58745	20.627	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	74985	29.593	ug/l	58
77) Bromobenzene	12.976	156	44718	19.574	ug/l	95
78) n-propylbenzene	13.035	91	210981	19.226	ug/l	100
79) 2-Chlorotoluene	13.123	91	129620	18.511	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	148466	19.217	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	18939	17.919	ug/l	99
82) 4-Chlorotoluene	13.218	91	135329	19.198	ug/l	97
83) tert-Butylbenzene	13.435	119	126785	18.464	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	153677	19.745	ug/l	98
85) sec-Butylbenzene	13.618	105	175585	19.162	ug/l	100
86) p-Isopropyltoluene	13.729	119	145633	19.220	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	82553	19.262	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	83788	19.359	ug/l	98
89) n-Butylbenzene	14.059	91	122012	17.758	ug/l	99
90) Hexachloroethane	14.329	117	27657	17.982	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	78881	18.770	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.723	75	10881	18.515	ug/l	97
93) 1,2,4-Trichlorobenzene	15.394	180	38616	18.655	ug/l	99
94) Hexachlorobutadiene	15.506	225	17729	17.175	ug/l	99
95) Naphthalene	15.641	128	115061	16.767	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	37660	18.175	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084434.D
Acq On : 21 Oct 2024 14:22
Operator : JC\MD
Sample : VN1021WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01

Quant Time: Oct 22 01:33:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084434.D
 Acq On : 21 Oct 2024 14:22
 Operator : JC\MD
 Sample : VN1021WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBSD01

Quant Time: Oct 22 01:33:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion:168 Resp: 163086
Ion Ratio Lower Upper
168 100
99 64.3 54.2 81.2



#2
Dichlorodifluoromethane
Concen: 16.833 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 85 Resp: 32472
Ion Ratio Lower Upper
85 100
87 31.4 17.6 52.9









#3
Chloromethane
Concen: 16.255 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion: 50 Resp: 35804
Ion Ratio Lower Upper
50 100
52 34.5 26.8 40.2



#4
Vinyl Chloride
Concen: 17.080 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 62 Resp: 36441
Ion Ratio Lower Upper
62 100
64 33.2 26.2 39.4









#5
Bromomethane
Concen: 16.992 ug/l
RT: 2.959 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion: 94 Resp: 23913
Ion Ratio Lower Upper
94 100
96 92.4 72.9 109.3



#6
Chloroethane
Concen: 16.017 ug/l
RT: 3.118 min Scan# 198
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 64 Resp: 24475
Ion Ratio Lower Upper
64 100
66 31.5 23.5 35.3









#7
 Trichlorofluoromethane
 Concen: 19.047 ug/l
 RT: 3.501 min Scan# 263
 Delta R.T. 0.006 min
 Lab File: VN084434.D
 Acq: 21 Oct 2024 14:22




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBSD01


Tgt Ion:101 Resp: 63283
 Ion Ratio Lower Upper
 101 100
 103 57.0 51.8 77.6



#8
 Diethyl Ether
 Concen: 19.089 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. 0.000 min
 Lab File: VN084434.D
 Acq: 21 Oct 2024 14:22

Tgt Ion: 74 Resp: 23522
 Ion Ratio Lower Upper
 74 100
 45 92.1 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.273 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion:101 Resp: 36689
Ion Ratio Lower Upper
101 100
85 44.0 36.2 54.4
151 75.4 59.3 88.9



#10
Methyl Iodide
Concen: 18.016 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion:142 Resp: 44167
Ion Ratio Lower Upper
142 100
127 44.8 35.3 52.9
141 14.8 11.0 16.6









#11
Tert butyl alcohol
Concen: 87.914 ug/l
RT: 5.518 min Scan# 607
Delta R.T. -0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion: 59 Resp: 35053
Ion Ratio Lower Upper
59 100
57 11.2 8.2 12.2



#12
1,1-Dichloroethene
Concen: 17.747 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 96 Resp: 32588
Ion Ratio Lower Upper
96 100
61 165.9 128.6 192.8
98 66.6 51.7 77.5









#13
Acrolein
Concen: 53.300 ug/l
RT: 4.189 min Scan# 378
Delta R.T. 0.012 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion: 56 Resp: 24397
Ion Ratio Lower Upper
56 100
55 69.9 57.2 85.8



#14
Allyl chloride
Concen: 16.646 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 41 Resp: 53083
Ion Ratio Lower Upper
41 100
39 73.1 59.0 88.6
76 37.9 29.8 44.6









#15
Acrylonitrile
Concen: 101.771 ug/l
RT: 5.718 min Scan# 640
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01

Tgt Ion: 53 Resp: 102471




Ion	Ratio	Lower	Upper
53	100		
52	80.5	66.2	99.4
51	34.0	28.8	43.2




#16
Acetone
Concen: 91.613 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 43 Resp: 95163

Ion	Ratio	Lower	Upper
43	100		
58	33.5	24.4	36.6









#17
Carbon Disulfide
Concen: 14.496 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion: 76 Resp: 84276
Ion Ratio Lower Upper
76 100
78 10.5 7.3 10.9



#18
Methyl Acetate
Concen: 24.005 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.012 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 43 Resp: 54347
Ion Ratio Lower Upper
43 100
74 24.7 19.2 28.8









#19
Methyl tert-butyl Ether
Concen: 19.826 ug/l
RT: 5.789 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion: 73 Resp: 122872
Ion Ratio Lower Upper
73 100
57 21.3 17.7 26.5



#20
Methylene Chloride
Concen: 20.192 ug/l
RT: 5.271 min Scan# 564
Delta R.T. -0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 84 Resp: 42636
Ion Ratio Lower Upper
84 100
49 108.1 107.2 160.8
51 34.8 31.8 47.8
86 58.8 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 18.395 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion: 96 Resp: 35388
Ion Ratio Lower Upper
96 100
61 128.1 108.4 162.6
98 59.8 49.8 74.8



#22
Diisopropyl ether
Concen: 19.859 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 45 Resp: 130999
Ion Ratio Lower Upper
45 100
43 51.9 42.3 63.5
87 26.3 21.8 32.6
59 11.8 9.6 14.4









#23
 Vinyl Acetate
 Concen: 103.215 ug/l
 RT: 6.606 min Scan# 791
 Delta R.T. 0.006 min
 Lab File: VN084434.D
 Acq: 21 Oct 2024 14:22




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBSD01


Tgt Ion: 43 Resp: 504977
 Ion Ratio Lower Upper
 43 100
 86 11.1 8.6 12.8



#24
 1,1-Dichloroethane
 Concen: 19.553 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. 0.000 min
 Lab File: VN084434.D
 Acq: 21 Oct 2024 14:22

Tgt Ion: 63 Resp: 72128
 Ion Ratio Lower Upper
 63 100
 98 8.1 3.4 10.2
 100 5.5 2.0 5.8









#25
2-Butanone
Concen: 96.795 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion: 43 Resp: 136944
Ion Ratio Lower Upper
43 100
72 25.1 21.4 32.2



#26
2,2-Dichloropropane
Concen: 19.469 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 77 Resp: 64755
Ion Ratio Lower Upper
77 100
97 21.7 10.8 32.3









#27
 cis-1,2-Dichloroethene
 Concen: 19.222 ug/l
 RT: 7.489 min Scan# 941
 Delta R.T. 0.000 min
 Lab File: VN084434.D
 Acq: 21 Oct 2024 14:22

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBSD01

Tgt Ion: 96 Resp: 44686




Ion	Ratio	Lower	Upper
96	100		
61	147.3	0.0	301.8
98	62.9	0.0	128.2




#28
 Bromochloromethane
 Concen: 21.744 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. 0.000 min
 Lab File: VN084434.D
 Acq: 21 Oct 2024 14:22

Tgt Ion: 49 Resp: 35788

Ion	Ratio	Lower	Upper
49	100		
129	1.9	0.0	3.6
130	73.3	58.0	87.0





#29

Tetrahydrofuran

Concen: 98.113 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01





Tgt Ion: 42 Resp: 85661

Ion Ratio Lower Upper

42 100

72 45.7 34.6 51.8

71 41.7 32.2 48.4



#30

Chloroform

Concen: 20.055 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. 0.000 min

Lab File: VN084434.D




Acq: 21 Oct 2024 14:22


Tgt Ion: 83 Resp: 76845

Ion Ratio Lower Upper

83 100

85 65.7 51.0 76.4









#31
Cyclohexane
Concen: 16.223 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion: 56 Resp: 58106
Ion Ratio Lower Upper
56 100
69 32.9 23.8 35.6
84 84.1 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 19.832 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 97 Resp: 68247
Ion Ratio Lower Upper
97 100
99 60.2 51.2 76.8
61 45.1 37.5 56.3









#33
 1,2-Dichloroethane-d4
 Concen: 51.056 ug/l
 RT: 8.577 min Scan# 1126
 Delta R.T. 0.000 min
 Lab File: VN084434.D
 Acq: 21 Oct 2024 14:22




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBSD01


Tgt Ion: 65 Resp: 123457
 Ion Ratio Lower Upper
 65 100
 67 51.6 0.0 102.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 9.100 min Scan# 1215
 Delta R.T. 0.000 min
 Lab File: VN084434.D
 Acq: 21 Oct 2024 14:22

Tgt Ion: 114 Resp: 278516
 Ion Ratio Lower Upper
 114 100
 63 21.6 0.0 43.8
 88 17.2 0.0 31.6





#35

Dibromofluoromethane

Concen: 52.829 ug/l

RT: 8.165 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01





Tgt Ion:113 Resp: 97003

Ion Ratio Lower Upper

113 100

111 104.0 83.3 124.9

192 18.2 13.5 20.3



#36

1,1-Dichloropropene

Concen: 18.689 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22




Tgt Ion: 75 Resp: 49846


Ion Ratio Lower Upper

75 100

110 34.2 16.9 50.7

77 32.0 24.7 37.1









#37
Ethyl Acetate
Concen: 19.160 ug/l
RT: 7.559 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion:	43	Resp:	52425
Ion	Ratio	Lower	Upper
43	100		
61	13.7	10.7	16.1
70	11.4	8.8	13.2



#38
Carbon Tetrachloride
Concen: 19.356 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion:	117	Resp:	56628
Ion	Ratio	Lower	Upper
117	100		
119	93.0	78.7	118.1
121	29.8	24.4	36.6









#39
Methylcyclohexane
Concen: 17.009 ug/l
RT: 9.600 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion: 83 Resp: 50490
Ion Ratio Lower Upper
83 100
55 81.3 62.8 94.2
98 48.7 38.8 58.2



#40
Benzene
Concen: 19.612 ug/l
RT: 8.606 min Scan# 1131
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 78 Resp: 162968
Ion Ratio Lower Upper
78 100
77 24.3 19.1 28.7









#41
Methacrylonitrile
Concen: 20.079 ug/l
RT: 7.777 min Scan# 990
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion:	41	Resp:	29826
Ion	Ratio	Lower	Upper
41	100		
39	55.1	43.3	64.9
67	74.8	56.1	84.1
52	31.6	24.5	36.7



#42
1,2-Dichloroethane
Concen: 20.598 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion:	62	Resp:	58026
Ion	Ratio	Lower	Upper
62	100		
98	9.3	0.0	18.2









#43
Isopropyl Acetate
Concen: 16.429 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion: 43	Resp:	88049
Ion Ratio	Lower	Upper
43	100	
61	27.4	20.4 30.6
87	13.9	10.3 15.5



#44
Trichloroethene
Concen: 19.541 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 130	Resp:	37926
Ion Ratio	Lower	Upper
130	100	
95	98.9	0.0 203.0





#45

1,2-Dichloropropane

Concen: 20.882 ug/l

RT: 9.624 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :





VN1021WBSD01

Tgt Ion: 63 Resp: 41049

Ion Ratio Lower Upper

63 100

65 33.3 24.6 37.0



#46

Dibromomethane

Concen: 22.002 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22




Tgt Ion: 93 Resp: 29187


Ion Ratio Lower Upper

93 100

95 82.7 65.0 97.4

174 90.6 71.8 107.8









#47
Bromodichloromethane
Concen: 20.940 ug/l
RT: 9.888 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion: 83 Resp: 61715
Ion Ratio Lower Upper
83 100
85 63.1 51.5 77.3
127 7.7 5.7 8.5



#48
Methyl methacrylate
Concen: 19.075 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 41 Resp: 40695
Ion Ratio Lower Upper
41 100
69 87.8 67.1 100.7
39 58.5 45.8 68.6





#49

1,4-Dioxane

Concen: 415.845 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01





Tgt Ion: 88 Resp: 16333

Ion Ratio Lower Upper

88 100

43 29.5 25.3 37.9

58 72.1 56.4 84.6



#50

Toluene-d8

Concen: 51.812 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.000 min

Lab File: VN084434.D




Acq: 21 Oct 2024 14:22


Tgt Ion: 98 Resp: 350013

Ion Ratio Lower Upper

98 100

100 66.0 52.7 79.1









#51
4-Methyl-2-Pentanone
Concen: 104.354 ug/l
RT: 10.447 min Scan# 1443
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion: 43 Resp: 271816
Ion Ratio Lower Upper
43 100
58 40.2 31.6 47.4



#52
Toluene
Concen: 20.289 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 92 Resp: 102809
Ion Ratio Lower Upper
92 100
91 171.1 137.3 205.9









#53
 t-1,3-Dichloropropene
 Concen: 19.701 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. 0.000 min
 Lab File: VN084434.D
 Acq: 21 Oct 2024 14:22




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBSD01


Tgt Ion: 75 Resp: 59188
 Ion Ratio Lower Upper
 75 100
 77 34.0 25.4 38.0



#54
 cis-1,3-Dichloropropene
 Concen: 20.003 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. 0.000 min
 Lab File: VN084434.D
 Acq: 21 Oct 2024 14:22

Tgt Ion: 75 Resp: 64640
 Ion Ratio Lower Upper
 75 100
 77 33.2 24.2 36.2
 39 44.7 38.7 58.1





#55

1,1,2-Trichloroethane

Concen: 22.465 ug/l

RT: 11.012 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01

Tgt Ion: 97 Resp: 40810





Ion Ratio Lower Upper

97 100

83 90.4 71.0 106.4

85 59.0 41.7 62.5

99 65.5 50.1 75.1



#56

Ethyl methacrylate

Concen: 19.919 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22




Tgt Ion: 69 Resp: 59539


Ion Ratio Lower Upper

69 100

41 65.6 53.3 79.9

39 34.7 28.8 43.2





#57

1,3-Dichloropropane

Concen: 21.236 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :





VN1021WBSD01

Tgt Ion: 76 Resp: 68931

Ion Ratio Lower Upper

76 100

78 33.4 25.8 38.6



#58

2-Chloroethyl Vinyl ether

Concen: 102.191 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. 0.000 min

Lab File: VN084434.D




Acq: 21 Oct 2024 14:22


Tgt Ion: 63 Resp: 141699

Ion Ratio Lower Upper

63 100

106 29.0 21.8 32.6









#59
2-Hexanone
Concen: 102.176 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion: 43 Resp: 198917
Ion Ratio Lower Upper
43 100
58 54.9 27.3 81.8



#60
Dibromochloromethane
Concen: 21.769 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 129 Resp: 46724
Ion Ratio Lower Upper
129 100
127 75.3 38.3 114.9





#61

1,2-Dibromoethane

Concen: 20.741 ug/l

RT: 11.471 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :





VN1021WBSD01

Tgt Ion:107 Resp: 39150

Ion Ratio Lower Upper

107 100

109 93.8 75.4 113.0



#62

4-Bromofluorobenzene

Concen: 52.449 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22




Tgt Ion: 95 Resp: 129081


Ion Ratio Lower Upper

95 100

174 73.0 0.0 145.2

176 70.7 0.0 140.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01





Tgt Ion:117 Resp: 255670

Ion Ratio Lower Upper

117 100

82 56.8 47.2 70.8

119 32.7 25.4 38.0



#64

Tetrachloroethene

Concen: 19.288 ug/l

RT: 11.106 min Scan# 1556

Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Tgt Ion:164 Resp: 34348




Ion Ratio Lower Upper


164 100

166 121.7 101.0 151.4

129 95.8 80.8 121.2

131 88.2 78.1 117.1









#65
Chlorobenzene
Concen: 19.535 ug/l
RT: 11.888 min Scan# 1107
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion:112 Resp: 110738
Ion Ratio Lower Upper
112 100
114 30.9 27.0 40.6



#66
1,1,1,2-Tetrachloroethane
Concen: 20.071 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion:131 Resp: 39189
Ion Ratio Lower Upper
131 100
133 95.4 47.9 143.7
119 67.4 32.8 98.3









#67
Ethyl Benzene
Concen: 18.360 ug/l
RT: 11.965 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion: 91 Resp: 184057
Ion Ratio Lower Upper
91 100
106 29.9 24.5 36.7



#68
m/p-Xylenes
Concen: 38.843 ug/l
RT: 12.071 min Scan# 1720
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion:106 Resp: 144045
Ion Ratio Lower Upper
106 100
91 206.2 167.1 250.7





#69

o-Xylene

Concen: 18.841 ug/l

RT: 12.400 min Scan# 1776

Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :





VN1021WBSD01

Tgt Ion:106 Resp: 66105

Ion Ratio Lower Upper

106 100

91 230.9 110.2 330.6



#70

Styrene

Concen: 19.814 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22




Tgt Ion:104 Resp: 117753


Ion Ratio Lower Upper

104 100

78 51.9 42.7 64.1

103 54.6 43.6 65.4





#71

Bromoform

Concen: 20.747 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01





Tgt Ion:173 Resp: 29825

Ion Ratio Lower Upper

173 100

175 47.7 23.5 70.6

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22




Tgt Ion:152 Resp: 124118


Ion Ratio Lower Upper

152 100

115 62.4 31.3 93.9

150 163.2 0.0 349.8





#73

Isopropylbenzene

Concen: 18.352 ug/l

RT: 12.694 min Scan# 1826

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :





VN1021WBSD01

Tgt Ion:105 Resp: 173820

Ion Ratio Lower Upper

105 100

120 26.1 12.9 38.7



#74

N-ethyl acetate

Concen: 17.252 ug/l

RT: 12.494 min Scan# 1792

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Tgt Ion: 43 Resp: 73994




Ion Ratio Lower Upper


43 100

70 45.7 34.4 51.6

55 26.6 19.6 29.4

61 24.3 19.6 29.4





#75

1,1,2,2-Tetrachloroethane

Concen: 20.627 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01





Tgt Ion: 83 Resp: 58745

Ion Ratio Lower Upper

83 100

131 10.1 4.9 14.7

85 64.3 31.9 95.7



#76

1,2,3-Trichloropropane

Concen: 29.593 ug/l

RT: 12.994 min Scan# 1877

Delta R.T. 0.006 min

Lab File: VN084434.D




Acq: 21 Oct 2024 14:22


Tgt Ion: 75 Resp: 74985

Ion Ratio Lower Upper

75 100

77 131.1 96.9 290.7









#77
Bromobenzene
Concen: 19.574 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion:156 Resp: 44718
Ion Ratio Lower Upper
156 100
77 219.8 115.1 345.3
158 95.3 48.9 146.8



#78
n-propylbenzene
Concen: 19.226 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 91 Resp: 210981
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4





#79

2-Chlorotoluene

Concen: 18.511 ug/l

RT: 13.123 min Scan# 1899

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :





VN1021WBSD01

Tgt Ion: 91 Resp: 129620

Ion Ratio Lower Upper

91 100

126 32.2 15.7 47.1



#80

1,3,5-Trimethylbenzene

Concen: 19.217 ug/l

RT: 13.170 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VN084434.D




Acq: 21 Oct 2024 14:22


Tgt Ion:105 Resp: 148466

Ion Ratio Lower Upper

105 100

120 47.9 23.7 71.1









#81
trans-1,4-Dichloro-2-butene
Concen: 17.919 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion: 75 Resp: 18939
Ion Ratio Lower Upper
75 100
53 97.4 78.0 117.0
89 47.3 35.5 53.3



#82
4-Chlorotoluene
Concen: 19.198 ug/l
RT: 13.218 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion: 91 Resp: 135329
Ion Ratio Lower Upper
91 100
126 32.9 15.7 47.1





#83

tert-Butylbenzene

Concen: 18.464 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01





Tgt Ion:119 Resp: 126785

Ion Ratio Lower Upper

119 100

91 66.9 34.1 102.3

134 23.7 12.8 38.4



#84

1,2,4-Trimethylbenzene

Concen: 19.745 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. 0.000 min

Lab File: VN084434.D




Acq: 21 Oct 2024 14:22


Tgt Ion:105 Resp: 153677

Ion Ratio Lower Upper

105 100

120 43.5 22.3 66.8









#85
 sec-Butylbenzene
 Concen: 19.162 ug/l
 RT: 13.618 min Scan# 1983
 Delta R.T. 0.000 min
 Lab File: VN084434.D
 Acq: 21 Oct 2024 14:22




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1021WBSD01


Tgt Ion:105 Resp: 175585
 Ion Ratio Lower Upper
 105 100
 134 19.4 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 19.220 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN084434.D
 Acq: 21 Oct 2024 14:22

Tgt Ion:119 Resp: 145633
 Ion Ratio Lower Upper
 119 100
 134 24.5 12.6 37.8
 91 24.9 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 19.262 ug/l

RT: 13.735 min Scan# 2002

Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01





Tgt Ion:146 Resp: 82553

Ion Ratio Lower Upper

146 100

111 42.5 21.8 65.3

148 63.4 31.9 95.7



#88

1,4-Dichlorobenzene

Concen: 19.359 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22




Tgt Ion:146 Resp: 83788


Ion Ratio Lower Upper

146 100

111 41.6 21.2 63.6

148 61.8 31.9 95.9





#89

n-Butylbenzene

Concen: 17.758 ug/l

RT: 14.059 min Scan# 2058

Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01





Tgt Ion: 91 Resp: 122012

Ion Ratio Lower Upper

91 100

92 51.8 26.6 79.7

134 24.3 12.3 36.9



#90

Hexachloroethane

Concen: 17.982 ug/l

RT: 14.329 min Scan# 2104

Delta R.T. 0.000 min

Lab File: VN084434.D




Acq: 21 Oct 2024 14:22


Tgt Ion: 117 Resp: 27657

Ion Ratio Lower Upper

117 100

201 70.8 34.4 103.2





#91

1,2-Dichlorobenzene

Concen: 18.770 ug/l

RT: 14.106 min Scan# 2066

Delta R.T. 0.000 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01





Tgt Ion:146 Resp: 78881

Ion Ratio Lower Upper

146 100

111 43.2 22.0 66.0

148 64.1 32.1 96.3



#92

1,2-Dibromo-3-Chloropropane

Concen: 18.515 ug/l

RT: 14.723 min Scan# 2171

Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22




Tgt Ion: 75 Resp: 10881


Ion Ratio Lower Upper

75 100

155 73.3 38.1 114.5

157 96.3 49.6 149.0









#93
1,2,4-Trichlorobenzene
Concen: 18.655 ug/l
RT: 15.394 min Scan# 21
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22




Instrument :
MSVOA_N
ClientSampleId :
VN1021WBSD01


Tgt Ion:180 Resp: 38616
Ion Ratio Lower Upper
180 100
182 92.4 47.0 141.0
145 33.0 16.4 49.4



#94
Hexachlorobutadiene
Concen: 17.175 ug/l
RT: 15.506 min Scan# 2304
Delta R.T. 0.006 min
Lab File: VN084434.D
Acq: 21 Oct 2024 14:22

Tgt Ion:225 Resp: 17729
Ion Ratio Lower Upper
225 100
223 65.9 32.4 97.2
227 65.3 32.4 97.0





#95

Naphthalene

Concen: 16.767 ug/l

RT: 15.641 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22

Instrument :

MSVOA_N

ClientSampleId :

VN1021WBSD01





Tgt Ion:128 Resp: 115061

Ion Ratio Lower Upper

128 100

127 12.4 10.5 15.7

129 10.8 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 18.175 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN084434.D

Acq: 21 Oct 2024 14:22




Tgt Ion:180 Resp: 37660

Ion Ratio Lower Upper

180 100

182 95.4 48.2 144.6

145 36.1 17.3 51.9



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084448.D
 Acq On : 21 Oct 2024 20:04
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Oct 22 01:42:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	164267	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	281823	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	256750	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	130386	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	116690	47.911	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.820%
35) Dibromofluoromethane	8.165	113	96698	52.045	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.100%
50) Toluene-d8	10.565	98	354860	51.913	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.820%
62) 4-Bromofluorobenzene	12.847	95	129892	52.160	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.320%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	84821	43.654	ug/l	94
3) Chloromethane	2.359	50	94401	42.550	ug/l	96
4) Vinyl Chloride	2.512	62	95237	44.316	ug/l	100
5) Bromomethane	2.942	94	58167	41.035	ug/l	96
6) Chloroethane	3.112	64	62275	40.462	ug/l	98
7) Trichlorofluoromethane	3.495	101	163261	48.786	ug/l	97
8) Diethyl Ether	3.959	74	60533	48.772	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.365	101	95561	49.837	ug/l	99
10) Methyl Iodide	4.589	142	118118	47.834	ug/l	98
11) Tert butyl alcohol	5.518	59	84056	209.300	ug/l	100
12) 1,1-Dichloroethene	4.336	96	91278	49.352	ug/l	98
13) Acrolein	4.183	56	86128	186.809	ug/l	98
14) Allyl chloride	5.024	41	140602	43.772	ug/l	98
15) Acrylonitrile	5.718	53	264703	261.006	ug/l	98
16) Acetone	4.430	43	209499	200.235	ug/l	97
17) Carbon Disulfide	4.712	76	222486	37.994	ug/l	99
18) Methyl Acetate	5.024	43	135697	59.505	ug/l	100
19) Methyl tert-butyl Ether	5.800	73	323965	51.897	ug/l	99
20) Methylene Chloride	5.277	84	108136	50.844	ug/l	93
21) trans-1,2-Dichloroethene	5.783	96	95558	49.315	ug/l	98
22) Diisopropyl ether	6.671	45	344689	51.879	ug/l	99
23) Vinyl Acetate	6.600	43	1175888	238.618	ug/l	100
24) 1,1-Dichloroethane	6.571	63	190033	51.145	ug/l	98
25) 2-Butanone	7.483	43	333066	233.725	ug/l	99
26) 2,2-Dichloropropane	7.494	77	149770	44.706	ug/l	97
27) cis-1,2-Dichloroethene	7.488	96	118609	50.655	ug/l	97
28) Bromochloromethane	7.812	49	78703	47.474	ug/l	97
29) Tetrahydrofuran	7.841	42	221138	251.461	ug/l	97
30) Chloroform	7.965	83	201338	52.167	ug/l	100
31) Cyclohexane	8.259	56	153399	42.521	ug/l	99
32) 1,1,1-Trichloroethane	8.171	97	177818	51.300	ug/l	99
36) 1,1-Dichloropropene	8.371	75	133081	49.312	ug/l	98
37) Ethyl Acetate	7.559	43	133124	48.082	ug/l	99
38) Carbon Tetrachloride	8.359	117	150440	50.818	ug/l	97
39) Methylcyclohexane	9.600	83	138930	46.254	ug/l	99
40) Benzene	8.606	78	428332	50.941	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
 Data File : VN084448.D
 Acq On : 21 Oct 2024 20:04
 Operator : JC\MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDC0050EC

Quant Time: Oct 22 01:42:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 07:11:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.783	41	76939	51.187	ug/l	97
42) 1,2-Dichloroethane	8.671	62	142985	50.160	ug/l	98
43) Isopropyl Acetate	8.688	43	229980	42.408	ug/l	99
44) Trichloroethene	9.353	130	99355	50.592	ug/l	99
45) 1,2-Dichloropropane	9.618	63	106710	53.648	ug/l	97
46) Dibromomethane	9.706	93	74783	55.712	ug/l	99
47) Bromodichloromethane	9.888	83	159633	53.528	ug/l	98
48) Methyl methacrylate	9.682	41	109772	50.850	ug/l	99
49) 1,4-Dioxane	9.694	88	39543	994.968	ug/l	97
51) 4-Methyl-2-Pentanone	10.447	43	708662	268.874	ug/l	100
52) Toluene	10.629	92	268587	52.382	ug/l	100
53) t-1,3-Dichloropropene	10.835	75	155308	51.089	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	170218	52.057	ug/l	98
55) 1,1,2-Trichloroethane	11.018	97	101779	55.371	ug/l	98
56) Ethyl methacrylate	10.871	69	165318	54.658	ug/l	97
57) 1,3-Dichloropropane	11.159	76	175471	53.424	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	390411	278.254	ug/l	98
59) 2-Hexanone	11.194	43	512990	260.410	ug/l	99
60) Dibromochloromethane	11.359	129	122224	56.277	ug/l	100
61) 1,2-Dibromoethane	11.471	107	100984	52.872	ug/l	100
64) Tetrachloroethene	11.100	164	87525	48.943	ug/l	97
65) Chlorobenzene	11.888	112	285234	50.105	ug/l	95
66) 1,1,1,2-Tetrachloroethane	11.959	131	104070	53.075	ug/l	99
67) Ethyl Benzene	11.965	91	518810	51.533	ug/l	100
68) m/p-Xylenes	12.070	106	389131	104.492	ug/l	100
69) o-Xylene	12.400	106	188510	53.503	ug/l	99
70) Styrene	12.412	104	329833	55.268	ug/l	99
71) Bromoform	12.576	173	78482	54.365	ug/l #	97
73) Isopropylbenzene	12.694	105	482626	48.507	ug/l	100
74) N-amyl acetate	12.494	43	205514	45.614	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.935	83	150053	50.155	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	189767	71.291	ug/l	61
77) Bromobenzene	12.976	156	118173	49.239	ug/l	94
78) n-propylbenzene	13.035	91	573481	49.748	ug/l	99
79) 2-Chlorotoluene	13.123	91	356393	48.450	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	411332	50.681	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	55404	49.901	ug/l	89
82) 4-Chlorotoluene	13.223	91	360840	48.730	ug/l	99
83) tert-Butylbenzene	13.435	119	352909	48.923	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	416913	50.992	ug/l	100
85) sec-Butylbenzene	13.617	105	482488	50.124	ug/l	100
86) p-Isopropyltoluene	13.729	119	404005	50.755	ug/l	99
87) 1,3-Dichlorobenzene	13.729	146	217666	48.345	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	218499	48.056	ug/l	99
89) n-Butylbenzene	14.053	91	342382	47.437	ug/l	99
90) Hexachloroethane	14.335	117	74567	46.151	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	214956	48.690	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	28592	46.312	ug/l	97
93) 1,2,4-Trichlorobenzene	15.394	180	105156	48.359	ug/l	99
94) Hexachlorobutadiene	15.500	225	47024	43.364	ug/l	98
95) Naphthalene	15.641	128	338368	46.938	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	103615	47.602	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084448.D
Acq On : 21 Oct 2024 20:04
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Quant Time: Oct 22 01:42:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102124\
Data File : VN084448.D
Acq On : 21 Oct 2024 20:04
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Quant Time: Oct 22 01:42:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N093024W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 07:11:01 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:168 Resp: 164267
Ion Ratio Lower Upper
168 100
99 63.4 54.2 81.2



#2
Dichlorodifluoromethane
Concen: 43.654 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 85 Resp: 84821
Ion Ratio Lower Upper
85 100
87 31.8 17.6 52.9









#3
Chloromethane
Concen: 42.550 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 50 Resp: 94401
Ion Ratio Lower Upper
50 100
52 31.1 26.8 40.2



#4
Vinyl Chloride
Concen: 44.316 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 62 Resp: 95237
Ion Ratio Lower Upper
62 100
64 32.8 26.2 39.4









#5
Bromomethane
Concen: 41.035 ug/l
RT: 2.942 min Scan# 11
Delta R.T. -0.012 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 94 Resp: 58167
Ion Ratio Lower Upper
94 100
96 87.6 72.9 109.3



#6
Chloroethane
Concen: 40.462 ug/l
RT: 3.112 min Scan# 197
Delta R.T. -0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 64 Resp: 62275
Ion Ratio Lower Upper
64 100
66 30.6 23.5 35.3









#7
Trichlorofluoromethane
Concen: 48.786 ug/l
RT: 3.495 min Scan# 262
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 163261
Ion Ratio Lower Upper
101 100
103 67.3 51.8 77.6



#8
Diethyl Ether
Concen: 48.772 ug/l
RT: 3.959 min Scan# 341
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 74 Resp: 60533
Ion Ratio Lower Upper
74 100
45 94.9 47.7 143.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.837 ug/l
RT: 4.365 min Scan# 411
Delta R.T. -0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 95561
Ion Ratio Lower Upper
101 100
85 44.4 36.2 54.4
151 74.8 59.3 88.9



#10
Methyl Iodide
Concen: 47.834 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion:142 Resp: 118118
Ion Ratio Lower Upper
142 100
127 45.8 35.3 52.9
141 14.6 11.0 16.6









#11
Tert butyl alcohol
Concen: 209.300 ug/l
RT: 5.518 min Scan# 607
Delta R.T. -0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 59 Resp: 84056
Ion Ratio Lower Upper
59 100
57 10.3 8.2 12.2



#12
1,1-Dichloroethene
Concen: 49.352 ug/l
RT: 4.336 min Scan# 405
Delta R.T. -0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 96 Resp: 91278
Ion Ratio Lower Upper
96 100
61 164.5 128.6 192.8
98 65.0 51.7 77.5









#13
Acrolein
Concen: 186.809 ug/l
RT: 4.183 min Scan# 31
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 86128
Ion Ratio Lower Upper
56 100
55 73.1 57.2 85.8



#14
Allyl chloride
Concen: 43.772 ug/l
RT: 5.024 min Scan# 522
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 41 Resp: 140602
Ion Ratio Lower Upper
41 100
39 71.5 59.0 88.6
76 38.3 29.8 44.6









#15
Acrylonitrile
Concen: 261.006 ug/l
RT: 5.718 min Scan# 640
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 53 Resp: 264703
Ion Ratio Lower Upper
53 100
52 80.3 66.2 99.4
51 35.5 28.8 43.2



#16
Acetone
Concen: 200.235 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 43 Resp: 209499
Ion Ratio Lower Upper
43 100
58 32.1 24.4 36.6









#17
Carbon Disulfide
Concen: 37.994 ug/l
RT: 4.712 min Scan# 41
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 222486
Ion Ratio Lower Upper
76 100
78 9.5 7.3 10.9



#18
Methyl Acetate
Concen: 59.505 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 43 Resp: 135697
Ion Ratio Lower Upper
43 100
74 24.0 19.2 28.8









#19
Methyl tert-butyl Ether
Concen: 51.897 ug/l
RT: 5.800 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 73 Resp: 323965
Ion Ratio Lower Upper
73 100
57 21.6 17.7 26.5



#20
Methylene Chloride
Concen: 50.844 ug/l
RT: 5.277 min Scan# 565
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 84 Resp: 108136
Ion Ratio Lower Upper
84 100
49 120.3 107.2 160.8
51 37.7 31.8 47.8
86 66.1 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 49.315 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 96 Resp: 95558
Ion Ratio Lower Upper
96 100
61 138.0 108.4 162.6
98 63.5 49.8 74.8



#22
Diisopropyl ether
Concen: 51.879 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 45 Resp: 344689
Ion Ratio Lower Upper
45 100
43 52.9 42.3 63.5
87 28.0 21.8 32.6
59 11.4 9.6 14.4









#23
 Vinyl Acetate
 Concen: 238.618 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. -0.000 min
 Lab File: VN084448.D
 Acq: 21 Oct 2024 20:04




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 43 Resp: 1175888
 Ion Ratio Lower Upper
 43 100
 86 10.7 8.6 12.8



#24
 1,1-Dichloroethane
 Concen: 51.145 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. -0.000 min
 Lab File: VN084448.D
 Acq: 21 Oct 2024 20:04

Tgt Ion: 63 Resp: 190033
 Ion Ratio Lower Upper
 63 100
 98 7.3 3.4 10.2
 100 4.5 2.0 5.8









#25
2-Butanone
Concen: 233.725 ug/l
RT: 7.483 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 333066
Ion Ratio Lower Upper
43 100
72 26.4 21.4 32.2



#26
2,2-Dichloropropane
Concen: 44.706 ug/l
RT: 7.494 min Scan# 942
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 77 Resp: 149770
Ion Ratio Lower Upper
77 100
97 23.0 10.8 32.3









#27
cis-1,2-Dichloroethene
Concen: 50.655 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 96 Resp: 118609
Ion Ratio Lower Upper
96 100
61 145.2 0.0 301.8
98 64.1 0.0 128.2



#28
Bromochloromethane
Concen: 47.474 ug/l
RT: 7.812 min Scan# 996
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 49 Resp: 78703
Ion Ratio Lower Upper
49 100
129 1.5 0.0 3.6
130 75.1 58.0 87.0





#29

Tetrahydrofuran

Concen: 251.461 ug/l

RT: 7.841 min Scan# 1000

Delta R.T. 0.006 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 42 Resp: 221138

Ion Ratio Lower Upper

42 100

72 45.3 34.6 51.8

71 41.7 32.2 48.4



#30

Chloroform

Concen: 52.167 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. -0.000 min

Lab File: VN084448.D




Acq: 21 Oct 2024 20:04


Tgt Ion: 83 Resp: 201338

Ion Ratio Lower Upper

83 100

85 63.4 51.0 76.4









#31
Cyclohexane
Concen: 42.521 ug/l
RT: 8.259 min Scan# 1072
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 153399
Ion Ratio Lower Upper
56 100
69 29.9 23.8 35.6
84 84.5 68.8 103.2



#32
1,1,1-Trichloroethane
Concen: 51.300 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 97 Resp: 177818
Ion Ratio Lower Upper
97 100
99 63.3 51.2 76.8
61 46.9 37.5 56.3









#33
 1,2-Dichloroethane-d4
 Concen: 47.911 ug/l
 RT: 8.577 min Scan# 1126
 Delta R.T. -0.000 min
 Lab File: VN084448.D
 Acq: 21 Oct 2024 20:04




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 65 Resp: 116690
 Ion Ratio Lower Upper
 65 100
 67 53.9 0.0 102.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 9.100 min Scan# 1215
 Delta R.T. -0.000 min
 Lab File: VN084448.D
 Acq: 21 Oct 2024 20:04

Tgt Ion: 114 Resp: 281823
 Ion Ratio Lower Upper
 114 100
 63 21.6 0.0 43.8
 88 16.3 0.0 31.6









#35
Dibromofluoromethane
Concen: 52.045 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:113 Resp: 96698
Ion Ratio Lower Upper
113 100
111 102.0 83.3 124.9
192 17.6 13.5 20.3



#36
1,1-Dichloropropene
Concen: 49.312 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 75 Resp: 133081
Ion Ratio Lower Upper
75 100
110 35.1 16.9 50.7
77 31.9 24.7 37.1









#37
Ethyl Acetate
Concen: 48.082 ug/l
RT: 7.559 min Scan# 91
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 133124
Ion Ratio Lower Upper
43 100
61 14.1 10.7 16.1
70 11.1 8.8 13.2



#38
Carbon Tetrachloride
Concen: 50.818 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 117 Resp: 150440
Ion Ratio Lower Upper
117 100
119 94.0 78.7 118.1
121 30.4 24.4 36.6





#39

Methylcyclohexane

Concen: 46.254 ug/l

RT: 9.600 min Scan# 11300

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83 Resp: 138930

Ion Ratio Lower Upper

83 100

55 79.2 62.8 94.2

98 49.1 38.8 58.2



#40

Benzene

Concen: 50.941 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN084448.D




Acq: 21 Oct 2024 20:04


Tgt Ion: 78 Resp: 428332

Ion Ratio Lower Upper

78 100

77 24.4 19.1 28.7









#41
Methacrylonitrile
Concen: 51.187 ug/l
RT: 7.783 min Scan# 991
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	41	Resp:	76939
Ion	Ratio	Lower	Upper
41	100		
39	53.7	43.3	64.9
67	74.5	56.1	84.1
52	31.6	24.5	36.7



#42
1,2-Dichloroethane
Concen: 50.160 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion:	62	Resp:	142985
Ion	Ratio	Lower	Upper
62	100		
98	9.9	0.0	18.2









#43
Isopropyl Acetate
Concen: 42.408 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 229980
Ion Ratio Lower Upper
43 100
61 26.3 20.4 30.6
87 13.5 10.3 15.5



#44
Trichloroethene
Concen: 50.592 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion:130 Resp: 99355
Ion Ratio Lower Upper
130 100
95 100.8 0.0 203.0





#45

1,2-Dichloropropane

Concen: 53.648 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 63 Resp: 106710

Ion Ratio Lower Upper

63 100

65 29.3 24.6 37.0



#46

Dibromomethane

Concen: 55.712 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04




Tgt Ion: 93 Resp: 74783


Ion Ratio Lower Upper

93 100

95 81.7 65.0 97.4

174 90.5 71.8 107.8









#47
Bromodichloromethane
Concen: 53.528 ug/l
RT: 9.888 min Scan# 1349
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 159633
Ion Ratio Lower Upper
83 100
85 66.1 51.5 77.3
127 8.3 5.7 8.5



#48
Methyl methacrylate
Concen: 50.850 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 41 Resp: 109772
Ion Ratio Lower Upper
41 100
69 86.0 67.1 100.7
39 57.3 45.8 68.6





#49

1,4-Dioxane

Concen: 994.968 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 88 Resp: 39543

Ion Ratio Lower Upper

88 100

43 30.1 25.3 37.9

58 73.5 56.4 84.6



#50

Toluene-d8

Concen: 51.913 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN084448.D




Acq: 21 Oct 2024 20:04


Tgt Ion: 98 Resp: 354860

Ion Ratio Lower Upper

98 100

100 65.6 52.7 79.1





#51

4-Methyl-2-Pentanone

Concen: 268.874 ug/l

RT: 10.447 min Scan# 1443

Delta R.T. 0.006 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 43 Resp: 708662

Ion Ratio Lower Upper

43 100

58 39.7 31.6 47.4



#52

Toluene

Concen: 52.382 ug/l

RT: 10.629 min Scan# 1475

Delta R.T. -0.000 min

Lab File: VN084448.D




Acq: 21 Oct 2024 20:04


Tgt Ion: 92 Resp: 268587

Ion Ratio Lower Upper

92 100

91 171.7 137.3 205.9









#53
t-1,3-Dichloropropene
Concen: 51.089 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 155308
Ion Ratio Lower Upper
75 100
77 32.2 25.4 38.0



#54
cis-1,3-Dichloropropene
Concen: 52.057 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 75 Resp: 170218
Ion Ratio Lower Upper
75 100
77 32.5 24.2 36.2
39 48.0 38.7 58.1





#55

1,1,2-Trichloroethane

Concen: 55.371 ug/l

RT: 11.018 min Scan# 1540

Delta R.T. 0.006 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 97 Resp: 101779





Ion Ratio Lower Upper

97 100

83 88.2 71.0 106.4

85 54.5 41.7 62.5

99 61.1 50.1 75.1



#56

Ethyl methacrylate

Concen: 54.658 ug/l

RT: 10.871 min Scan# 1516

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04




Tgt Ion: 69 Resp: 165318


Ion Ratio Lower Upper

69 100

41 64.1 53.3 79.9

39 34.8 28.8 43.2





#57

1,3-Dichloropropane

Concen: 53.424 ug/l

RT: 11.159 min Scan# 1565

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 76 Resp: 175471

Ion Ratio Lower Upper

76 100

78 32.5 25.8 38.6



#58

2-Chloroethyl Vinyl ether

Concen: 278.254 ug/l

RT: 10.159 min Scan# 1395

Delta R.T. -0.000 min

Lab File: VN084448.D




Acq: 21 Oct 2024 20:04


Tgt Ion: 63 Resp: 390411

Ion Ratio Lower Upper

63 100

106 28.2 21.8 32.6









#59
2-Hexanone
Concen: 260.410 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 512990
Ion Ratio Lower Upper
43 100
58 55.2 27.3 81.8



#60
Dibromochloromethane
Concen: 56.277 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion:129 Resp: 122224
Ion Ratio Lower Upper
129 100
127 76.7 38.3 114.9









#61
1,2-Dibromoethane
Concen: 52.872 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:107 Resp: 100984
Ion Ratio Lower Upper
107 100
109 94.3 75.4 113.0



#62
4-Bromofluorobenzene
Concen: 52.160 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 95 Resp: 129892
Ion Ratio Lower Upper
95 100
174 74.2 0.0 145.2
176 72.0 0.0 140.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:117 Resp: 256750

Ion Ratio Lower Upper

117 100

82 55.9 47.2 70.8

119 32.6 25.4 38.0



#64

Tetrachloroethene

Concen: 48.943 ug/l

RT: 11.100 min Scan# 1555

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Tgt Ion:164 Resp: 87525




Ion Ratio Lower Upper


164 100

166 121.2 101.0 151.4

129 99.9 80.8 121.2

131 95.2 78.1 117.1





#65

Chlorobenzene

Concen: 50.105 ug/l

RT: 11.888 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion:112 Resp: 285234

Ion Ratio Lower Upper

112 100

114 30.9 27.0 40.6



#66

1,1,1,2-Tetrachloroethane

Concen: 53.075 ug/l

RT: 11.959 min Scan# 1701

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04




Tgt Ion:131 Resp: 104070


Ion Ratio Lower Upper

131 100

133 94.9 47.9 143.7

119 64.4 32.8 98.3





#67

Ethyl Benzene

Concen: 51.533 ug/l

RT: 11.965 min Scan# 1701

Delta R.T. 0.006 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 91 Resp: 518810

Ion Ratio Lower Upper

91 100

106 30.6 24.5 36.7



#68

m/p-Xylenes

Concen: 104.492 ug/l

RT: 12.070 min Scan# 1720

Delta R.T. -0.000 min

Lab File: VN084448.D




Acq: 21 Oct 2024 20:04


Tgt Ion:106 Resp: 389131

Ion Ratio Lower Upper

106 100

91 208.9 167.1 250.7









#69
o-Xylene
Concen: 53.503 ug/l
RT: 12.400 min Scan# 1775
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:106 Resp: 188510
Ion Ratio Lower Upper
106 100
91 221.9 110.2 330.6



#70
Styrene
Concen: 55.268 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion:104 Resp: 329833
Ion Ratio Lower Upper
104 100
78 51.5 42.7 64.1
103 54.3 43.6 65.4





#71

Bromoform

Concen: 54.365 ug/l

RT: 12.576 min Scan# 1806

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:173 Resp: 78482

Ion Ratio Lower Upper

173 100

175 48.9 23.5 70.6

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04




Tgt Ion:152 Resp: 130386


Ion Ratio Lower Upper

152 100

115 62.2 31.3 93.9

150 175.6 0.0 349.8









#73
Isopropylbenzene
Concen: 48.507 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:105 Resp: 482626
Ion Ratio Lower Upper
105 100
120 25.8 12.9 38.7



#74
N-ethyl acetate
Concen: 45.614 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 43 Resp: 205514
Ion Ratio Lower Upper
43 100
70 44.7 34.4 51.6
55 26.1 19.6 29.4
61 24.5 19.6 29.4





#75

1,1,2,2-Tetrachloroethane

Concen: 50.155 ug/l

RT: 12.935 min Scan# 1867

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83 Resp: 150053

Ion Ratio Lower Upper

83 100

131 9.8 4.9 14.7

85 63.2 31.9 95.7



#76

1,2,3-Trichloropropane

Concen: 71.291 ug/l

RT: 12.994 min Scan# 1877

Delta R.T. 0.006 min

Lab File: VN084448.D




Acq: 21 Oct 2024 20:04


Tgt Ion: 75 Resp: 189767

Ion Ratio Lower Upper

75 100

77 136.2 96.9 290.7









#77
Bromobenzene
Concen: 49.239 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:156 Resp: 118173
Ion Ratio Lower Upper
156 100
77 218.7 115.1 345.3
158 94.5 48.9 146.8



#78
n-propylbenzene
Concen: 49.748 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 91 Resp: 573481
Ion Ratio Lower Upper
91 100
120 22.1 10.8 32.4









#79
2-Chlorotoluene
Concen: 48.450 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 356393
Ion Ratio Lower Upper
91 100
126 31.9 15.7 47.1



#80
1,3,5-Trimethylbenzene
Concen: 50.681 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 105 Resp: 411332
Ion Ratio Lower Upper
105 100
120 47.5 23.7 71.1









#81
trans-1,4-Dichloro-2-butene
Concen: 49.901 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 55404
Ion Ratio Lower Upper
75 100
53 85.1 78.0 117.0
89 40.5 35.5 53.3



#82
4-Chlorotoluene
Concen: 48.730 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 91 Resp: 360840
Ion Ratio Lower Upper
91 100
126 31.8 15.7 47.1









#83
tert-Butylbenzene
Concen: 48.923 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:119 Resp: 352909
Ion Ratio Lower Upper
119 100
91 68.0 34.1 102.3
134 24.0 12.8 38.4



#84
1,2,4-Trimethylbenzene
Concen: 50.992 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion:105 Resp: 416913
Ion Ratio Lower Upper
105 100
120 44.3 22.3 66.8









#85
 sec-Butylbenzene
 Concen: 50.124 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN084448.D
 Acq: 21 Oct 2024 20:04




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:105 Resp: 482488
 Ion Ratio Lower Upper
 105 100
 134 19.6 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 50.755 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN084448.D
 Acq: 21 Oct 2024 20:04

Tgt Ion:119 Resp: 404005
 Ion Ratio Lower Upper
 119 100
 134 25.8 12.6 37.8
 91 24.9 12.4 37.2









#87
1,3-Dichlorobenzene
Concen: 48.345 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:146 Resp: 217666
Ion Ratio Lower Upper
146 100
111 42.8 21.8 65.3
148 63.9 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 48.056 ug/l
RT: 13.812 min Scan# 2016
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion:146 Resp: 218499
Ion Ratio Lower Upper
146 100
111 41.1 21.2 63.6
148 63.8 31.9 95.9









#89
n-Butylbenzene
Concen: 47.437 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 342382
Ion Ratio Lower Upper
91 100
92 52.6 26.6 79.7
134 24.3 12.3 36.9



#90
Hexachloroethane
Concen: 46.151 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. 0.006 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 117 Resp: 74567
Ion Ratio Lower Upper
117 100
201 68.3 34.4 103.2









#91
1,2-Dichlorobenzene
Concen: 48.690 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:146 Resp: 214956
Ion Ratio Lower Upper
146 100
111 44.3 22.0 66.0
148 62.9 32.1 96.3



#92
1,2-Dibromo-3-Chloropropane
Concen: 46.312 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN084448.D
Acq: 21 Oct 2024 20:04

Tgt Ion: 75 Resp: 28592
Ion Ratio Lower Upper
75 100
155 78.9 38.1 114.5
157 96.6 49.6 149.0





#93

1,2,4-Trichlorobenzene

Concen: 48.359 ug/l

RT: 15.394 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:180 Resp: 105156

Ion Ratio Lower Upper

180 100

182 94.7 47.0 141.0

145 31.9 16.4 49.4



#94

Hexachlorobutadiene

Concen: 43.364 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04




Tgt Ion:225 Resp: 47024


Ion Ratio Lower Upper

225 100

223 62.1 32.4 97.2

227 63.8 32.4 97.0





#95

Naphthalene

Concen: 46.938 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. 0.006 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:128 Resp: 338368

Ion Ratio Lower Upper

128 100

127 12.6 10.5 15.7

129 10.5 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 47.602 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.006 min

Lab File: VN084448.D

Acq: 21 Oct 2024 20:04

Tgt Ion:180 Resp: 103615

Ion Ratio Lower Upper

180 100

182 94.8 48.2 144.6

145 35.1 17.3 51.9

