

Surrogate Summary

SDG No.: P4440

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4440-01	BPOW6-11-HYD-20241016	1,2-Dichloroethane-d4	50	51.3	103	81	118
		Dibromofluoromethane	50	51.5	103	80	119
		Toluene-d8	50	49.7	99	89	112
		4-Bromofluorobenzene	50	46.9	94	85	114
P4440-02	BPOW6-11-TB-20241016	1,2-Dichloroethane-d4	50	47.5	95	81	118
		Dibromofluoromethane	50	50.9	102	80	119
		Toluene-d8	50	49.1	98	89	112
		4-Bromofluorobenzene	50	45.6	91	85	114
VN1021WBL01	VN1021WBL01	1,2-Dichloroethane-d4	50	48.6	97	81	118
		Dibromofluoromethane	50	51.1	102	80	119
		Toluene-d8	50	48.8	98	89	112
		4-Bromofluorobenzene	50	44.7	89	85	114
VN1021WBS01	VN1021WBS01	1,2-Dichloroethane-d4	50	50.5	101	81	118
		Dibromofluoromethane	50	54.3	109	80	119
		Toluene-d8	50	54.0	108	89	112
		4-Bromofluorobenzene	50	52.5	105	85	114
VN1021WBSD0	VN1021WBSD01	1,2-Dichloroethane-d4	50	51.1	102	81	118
		Dibromofluoromethane	50	52.8	106	80	119
		Toluene-d8	50	51.8	104	89	112
		4-Bromofluorobenzene	50	52.5	105	85	114