

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG NO.: P4440
 Lab File ID: VN084428.D Date Analyzed: 10/21/2024
 Instrument ID: MSVOA_N Time Analyzed: 11:38
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	127638	13.788				
UPPER LIMIT	255276	14.288				
LOWER LIMIT	63819	13.288				
EPA SAMPLE NO.						
BPOW6-11-HYD-20241016	101220	13.79				
BPOW6-11-TB-20241016	103232	13.79				
VN1021WBL01	99954	13.79				
VN1021WBS01	138325	13.79				
VN1021WBSD01	124118	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.