

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P4440 SAS No.: P4440 SDG NO.: P4440
Lab File ID: VN084211.D BFB Injection Date: 09/30/2024
Instrument ID: MSVOA_N BFB Injection Time: 09:24
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.9
75	30.0 - 60.0% of mass 95	53.9
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.6 (0.8) 1
174	50.0 - 100.0% of mass 95	71
175	5.0 - 9.0% of mass 174	5.5 (7.8) 1
176	95.0 - 101.0% of mass 174	69.7 (98.2) 1
177	5.0 - 9.0% of mass 176	4.9 (7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC100	VSTDICC100	VN084213.D	09/30/2024	12:25
VSTDICCC050	VSTDICCC050	VN084214.D	09/30/2024	12:49
VSTDICC020	VSTDICC020	VN084215.D	09/30/2024	13:13
VSTDICC010	VSTDICC010	VN084216.D	09/30/2024	13:37
VSTDICC005	VSTDICC005	VN084217.D	09/30/2024	14:00
VSTDICC001	VSTDICC001	VN084218.D	09/30/2024	14:48