

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4442 SAS No.: P4442 SDG NO.: P4442

Continuing Calib Date: 10/23/2024 Initial Calibration Date(s): 10/07/2024 10/08/2024

Continuing Calib Time: 13:47 Initial Calibration Time(s): 17:21 02:48

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.42	4.42	4.35	4.49	0.01
Aroclor-1242-2	(2)	4.43	4.44	4.37	4.51	0.01
Aroclor-1242-3	(3)	4.49	4.50	4.43	4.57	0.01
Aroclor-1242-4	(4)	4.58	4.59	4.52	4.66	0.01
Aroclor-1242-5	(5)	5.29	5.30	5.23	5.37	0.01
TCX		3.33	3.33	3.28	3.38	0.00
DCB		8.45	8.46	8.36	8.56	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4442 SAS No.: P4442 SDG NO.: P4442

Continuing Calib Date: 10/23/2024 Initial Calibration Date(s): 10/07/2024 10/08/2024

Continuing Calib Time: 13:47 Initial Calibration Time(s): 17:21 02:48

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	3.73	3.74	3.67	3.81	0.01
Aroclor-1242-2	(2)	3.75	3.75	3.68	3.82	0.01
Aroclor-1242-3	(3)	3.90	3.91	3.84	3.98	0.01
Aroclor-1242-4	(4)	3.99	4.00	3.93	4.07	0.01
Aroclor-1242-5	(5)	4.48	4.49	4.42	4.56	0.01
TCX		2.74	2.74	2.69	2.79	0.00
DCB		7.48	7.48	7.38	7.58	0.01