

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4442 SAS No.: P4442 SDG NO.: P4442

Continuing Calib Date: 10/23/2024 Initial Calibration Date(s): 10/07/2024 10/08/2024

Continuing Calib Time: 14:01 Initial Calibration Time(s): 17:21 02:48

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.42	4.42	4.35	4.49	0.00
Aroclor-1248-2	(2)	4.68	4.68	4.61	4.75	0.00
Aroclor-1248-3	(3)	4.87	4.87	4.80	4.94	0.00
Aroclor-1248-4	(4)	5.26	5.26	5.19	5.33	0.00
Aroclor-1248-5	(5)	5.29	5.30	5.23	5.37	0.01
TCX		3.33	3.33	3.28	3.38	0.00
DCB		8.45	8.46	8.36	8.56	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4442 SAS No.: P4442 SDG NO.: P4442

Continuing Calib Date: 10/23/2024 Initial Calibration Date(s): 10/07/2024 10/08/2024

Continuing Calib Time: 14:01 Initial Calibration Time(s): 17:21 02:48

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	3.73	3.74	3.67	3.81	0.01
Aroclor-1248-2	(2)	3.95	3.96	3.89	4.03	0.01
Aroclor-1248-3	(3)	3.99	4.00	3.93	4.07	0.01
Aroclor-1248-4	(4)	4.15	4.15	4.08	4.22	0.00
Aroclor-1248-5	(5)	4.52	4.53	4.46	4.60	0.01
TCX		2.74	2.74	2.69	2.79	0.00
DCB		7.48	7.48	7.38	7.58	0.01