

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4442 SAS No.: P4442 SDG NO.: P4442

Continuing Calib Date: 10/23/2024 Initial Calibration Date(s): 10/07/2024 10/08/2024

Continuing Calib Time: 14:16 Initial Calibration Time(s): 17:21 02:48

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.23	5.23	5.16	5.30	0.00
Aroclor-1254-2	(2)	5.45	5.45	5.38	5.52	0.00
Aroclor-1254-3	(3)	5.80	5.80	5.73	5.87	0.00
Aroclor-1254-4	(4)	6.08	6.08	6.01	6.15	0.00
Aroclor-1254-5	(5)	6.49	6.50	6.43	6.57	0.01
TCX		3.33	3.33	3.28	3.38	0.00
DCB		8.45	8.46	8.36	8.56	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P4442 SAS No.: P4442 SDG NO.: P4442

Continuing Calib Date: 10/23/2024 Initial Calibration Date(s): 10/07/2024 10/08/2024

Continuing Calib Time: 14:16 Initial Calibration Time(s): 17:21 02:48

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	4.48	4.49	4.42	4.56	0.01
Aroclor-1254-2	(2)	4.63	4.63	4.56	4.70	0.00
Aroclor-1254-3	(3)	5.01	5.01	4.94	5.08	0.00
Aroclor-1254-4	(4)	5.23	5.24	5.17	5.31	0.01
Aroclor-1254-5	(5)	5.63	5.64	5.57	5.71	0.01
TCX		2.74	2.74	2.69	2.79	0.00
DCB		7.47	7.48	7.38	7.58	0.01