



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Portal Partners Tri-Venture

Project Name: Amtrak Sawtooth Bridges 2024

Project # N/A

Chemtech Project # P4460

Test Name: SVOC-TCL BNA -20

A. Number of Samples and Date of Receipt:

3 Solid samples were received on 10/18/2024.

2 Water samples were received on 10/18/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Corrosivity, EPH, EPH, Hexavalent Chromium, Ignitability, Mercury, Metals ICP-TAL, METALS-TAL, PCB, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, Trivalent Chromium, VOC-TCLVOA-10 and VOC-TCLVOA-10. This data package contains results for SVOC-TCL BNA -20.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df. The samples were analyzed on instrument BNA_M using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GG. The analysis of SVOC-TCL BNA -20 was based on method 8270E and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS {P4460-03MS} with File ID: BF140006.D recoveries met the requirements for all compounds except for 4,6-Dinitro-2-methylphenol[148%], 4-Chloroaniline[34%] and Hexachlorocyclopentadiene[164%], these compounds did not meet the NJDKQP criteria but met the in-house criteria.

The MSD {P4460-03MSD} with File ID: BF140007.D recoveries met the acceptable requirements except for 4,6-Dinitro-2-methylphenol[157%], 4-Chloroaniline[36%], Di-n-octyl phthalate[138%], these compounds did not meet the NJDKQP criteria but met the in-house criteria, while Hexachlorocyclopentadiene[179%], Atrazine[133%], Dibenz(a,h)anthracene[138%], Indeno(1,2,3-cd)pyrene[138%], these compounds did not

meet the NJDKQP criteria and in-house criteria, due to matrix interference no corrective action is required.

The RPD for {P4460-03MSD} with File ID: BF140007.D met criteria except for Benzo(g,h,i)perylene[38%], Dibenz(a,h)anthracene[37%] and Indeno(1,2,3-cd)pyrene [37%], these compounds did not meet the NJDKQP criteria and in-house criteria due to matrix interference, no corrective action is required.

The Blank Spike for {PB164286BS} with File ID: BF139963.D met requirements for all samples except for 4-Chloroaniline[65%], Hexachlorocyclopentadiene[161%], these compounds did not meet the NJDKQP criteria but met the in-house criteria.

The Blank Spike for {PB164369BS} with File ID: BM048237.D met requirements for all samples except for 3-Nitroaniline[60%], 4-Chloroaniline[40%], these compounds did not meet the NJDKQP criteria but met the in-house criteria. While Hexachlorocyclopentadiene [210%], this compound did not meet the NJDKQP criteria and in-house criteria, but associate samples have no positive hit for these compounds therefore no corrective action was taken.

The Blank Spike Duplicate for {PB164369BSD} with File ID: BM048233.D met requirements for all samples except for 3,3-Dichlorobenzidine[69%], 3-Nitroaniline [58%], 4-Chloroaniline[41%], these compounds did not meet the NJDKQP criteria but met the in-house criteria, while Hexachlorocyclopentadiene[210%], this compound did not meet the NJDKQP criteria and in-house criteria, but associate samples have no positive hit for these compounds therefore no corrective action was taken .

The Blank analysis did not indicate the presence of lab contamination.

The % RSD is greater than 20% in the Initial Calibration (8270-BF101824.M) for 2,4-Dinitrophenol, this compound is passing on Linear Regression

The Continuous Calibration File ID BF139952.D met the requirements except for 2,4-Dinitrophenol and 4,6-Dinitro-2-methylphenol, The associate samples have no positive hit for these compounds therefore no corrective action was taken. .

The Continuous Calibration File ID BF140001.D met the requirements except for 2,4-Dinitrophenol and 4,6-Dinitro-2-methylphenol, The associate samples have no positive hit for these compounds therefore no corrective action was taken.

The Continuous Calibration File ID BF140050.D met the requirements except for 2,4-Dinitrophenol and 4,6-Dinitro-2-methylphenol, The associate samples have no positive hit for these compounds therefore no corrective action was taken.
The Tuning criteria met requirements.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

E. Additional Comments:

The Sample WB-303-TOP, have the concentration of target compound below method detection limits; therefore it is not reported as Hit in Form1.

Sample # WB-303-SW, was received with limited volume.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____