

CASE NARRATIVE

Portal Partners Tri-Venture

Project Name: Amtrak Sawtooth Bridges 2024

Project # N/A

Chemtech Project # P4460

Test Name: TCLP Herbicide

A. Number of Samples and Date of Receipt:

3 Solid samples were received on 10/18/2024.

2 Water samples were received on 10/18/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Corrosivity, EPH, EPH, Hexavalent Chromium, Ignitability, Mercury, Metals ICP-TAL, METALS-TAL, PCB, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, Trivalent Chromium, VOC-TCLVOA-10 and VOC-TCLVOA-10. This data package contains results for TCLP Herbicide.

C. Analytical Techniques:

The analysis was performed on instrument ECD_S. The front column is RTX-CLPesticides which is 30 meters, 0.32 mm ID, 0.5 um df, Catalog # 11139. The rear column is RTX-CLPesticides2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 11324. The analysis of TCLP Herbicides was based on method 8151A and extraction was done based on method 3510 and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for WB-301-BOTMS [2,4-DCAA(1) - 28%, 2,4-DCAA(2) - 23%], WB-301-BOTMSD [2,4-DCAA(1) - 28%, 2,4-DCAA(2) - 22%], these compounds did not meet the NJDKQP criteria and in-house criteria and surrogate failure for MS-MSD confirms with Original sample while, PB164261TB[2,4-DCAA(2) - 57%], this compound did not meet the NJDKQP criteria but met the in-house criteria.

The Retention Times were acceptable for all samples.

The MS {P4397-06MS} with File ID: PS028042.D recoveries met the requirements for all compounds except for 2,4,5-TP(Silvex)[212%], this compound did not meet the NJDKQP criteria and in-house criteria due to matrix interference.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

The MSD {P4397-06MSD} with File ID: PS028043.D recoveries met the acceptable requirements except for 2,4,5-TP(Silvex)[226%], this compound did not meet the NJDKQP criteria and in-house criteria due to matrix interference.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

E. Additional Comments:

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____