

Matrix Spike Summary

Client:	Portal Partners Tri-Venture	SDG No.:	P4460
Project:	Amtrak Sawtooth Bridges 2024	Sample ID:	P4460-03
Client ID:	WB-303-BOTMS	Percent Solids for Spike Sample:	80.1

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	75-125	1450		0.098	U	1600	40	91		10/22/2024

Matrix Spike Summary

Client:	Portal Partners Tri-Venture	SDG No.:	P4460
Project:	Amtrak Sawtooth Bridges 2024	Sample ID:	P4460-03
Client ID:	WB-303-BOTMS	Percent Solids for Spike Sample:	80.1

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	85-115	46.7		0.098	U	49.9	2	94		10/22/2024

Matrix Spike Summary

Client:	Portal Partners Tri-Venture	SDG No.:	P4460
Project:	Amtrak Sawtooth Bridges 2024	Sample ID:	P4460-03
Client ID:	WB-303-BOTMS	Percent Solids for Spike Sample:	80.1

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	75-125	38.6		0.098	U	49.9	2	77		10/22/2024

Matrix Spike Summary

Client:	Portal Partners Tri-Venture	SDG No.:	P4460
Project:	Amtrak Sawtooth Bridges 2024	Sample ID:	P4460-06
Client ID:	WB-303-SWMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/L	90-111	1.00		0.0030	U	1.0	2	100		10/18/2024

Matrix Spike Summary

Client:	Portal Partners Tri-Venture	SDG No.:	P4460
Project:	Amtrak Sawtooth Bridges 2024	Sample ID:	P4460-06
Client ID:	WB-303-SWMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/L	90-111	1.00		0.0030	U	1.0	2	100		10/18/2024