


LB133043

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.009	-0.009	-51	0.000	-51					1
0.2	0.200	0.188	-0.012	803	0.0 %	803					6
2.5	2.500	2.536	0.036	10942	0.0 %	10942					1
5.0	5.000	5.018	0.018	21666	0.0 %	21666					0
7.5	7.500	7.443	-0.057	32140	0.0 %	32140					0
10.0	10.000	10.024	0.024	43290	0.0 %	43290					0

LB133043 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Oct 2024 11:27:59

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	-51	-	0.00007471B	PPB	22 Oct 2024 11:34:54	S	Std
0.2 - 1	50.2	803	-	0.20007471B	PPB	22 Oct 2024 11:37:10	S	Std
2.5 - 1	52.5	10942	-	2.50007471B	PPB	22 Oct 2024 11:39:27	S	Std
5.0 - 1	55	21666	-	5.00007471B	PPB	22 Oct 2024 11:41:44	S	Std
7.5 - 1	57.5	32140	-	7.50007471B	PPB	22 Oct 2024 11:44:02	S	Std
10.0 - 1	510	43290	-	10.00007471B	PPB	22 Oct 2024 11:51:53	S	Std
ICV47 - 1	ICV47	18397	4.2614	-7471B	PPB	22 Oct 2024 11:54:48	U	SMPL
ICB47 - 1	ICB47	-332	-0.0745	-7471B	PPB	22 Oct 2024 11:57:04	U	SMPL
CCV52 - 1	CCV52	21695	5.0249	-7471B	PPB	22 Oct 2024 11:59:22	U	SMPL
CCB52 - 1	CCB52	-319	-0.0715	-7471B	PPB	22 Oct 2024 12:01:37	U	SMPL
CRA - 1	CRA	905	0.2119	-7471B	PPB	22 Oct 2024 12:03:55	U	SMPL
HighStd - 1	HighStd	43748	10.1304	-7471B	PPB	22 Oct 2024 12:06:10	U	SMPL
ChkStd - 1	ChkStd	29070	6.7323	-7471B	PPB	22 Oct 2024 12:08:25	U	SMPL
PB164322BL - 1	PBS	-70	-0.0138	-7471B	PPB	22 Oct 2024 12:13:27	U	SMPL
PB164322BS - 1	LCSS	18397	4.2614	-7471B	PPB	22 Oct 2024 12:15:43	U	SMPL
P3390-01 - 1	LOD-MDL-SOIL-01-QT3-2024	196	0.0477	-7471B	PPB	22 Oct 2024 12:20:39	U	SMPL
P3390-02 - 1	LOQ-SOIL-02-QT3-2024	920	0.2153	-7471B	PPB	22 Oct 2024 12:22:54	U	SMPL
P4443-01 - 1	OG-315-HR-502-COMP-29	9510	2.2040	-7471B	PPB	22 Oct 2024 12:30:05	U	SMPL
P4443-01DUP - 1	OG-315-HR-502-COMP-29DUP	9110	2.1114	-7471B	PPB	22 Oct 2024 12:32:24	U	SMPL
P4443-01MS - 1	OG-315-HR-502-COMP-29MS	29896	6.9235	-7471B	PPB	22 Oct 2024 12:34:41	U	SMPL
CCV53 - 1	CCV53	22325	5.1708	-7471B	PPB	22 Oct 2024 12:36:58	U	SMPL
CCB53 - 1	CCB53	-327	-0.0733	-7471B	PPB	22 Oct 2024 12:39:18	U	SMPL
P4443-01MSD - 1	OG-315-HR-502-COMP-29MSD	29032	6.7235	-7471B	PPB	22 Oct 2024 12:41:36	U	SMPL
P4443-06 - 1	OG-315-HR-502-COMP-30	29287	6.7826	-7471B	PPB	22 Oct 2024 12:43:51	U	SMPL
P4452-01 - 1	ETGI-285	-36	-0.0060	-7471B	PPB	22 Oct 2024 12:46:11	U	SMPL
P4455-01 - 1	SU-4-101824	1626	0.3788	-7471B	PPB	22 Oct 2024 12:48:31	U	SMPL
P4458-01 - 1	280517	2010	0.4677	-7471B	PPB	22 Oct 2024 12:50:46	U	SMPL
P4460-02 - 1	WB-303-TOP	215348	49.8574	-7471B	PPB	22 Oct 2024 12:53:02	U	SMPL
P4460-03 - 1	WB-303-BOT	-278	-0.0620	-7471B	PPB	22 Oct 2024 12:55:20	U	SMPL
P4467-01 - 1	TP-1	327	0.0781	-7471B	PPB	22 Oct 2024 12:57:58	U	SMPL
P4468-01 - 1	ETGI-331	237	0.0572	-7471B	PPB	22 Oct 2024 13:00:15	U	SMPL
P4468-03 - 1	ETGI-329	2568	0.5969	-7471B	PPB	22 Oct 2024 13:02:30	U	SMPL
CCV54 - 1	CCV54	22313	5.1680	-7471B	PPB	22 Oct 2024 13:04:45	U	SMPL
CCB54 - 1	CCB54	-306	-0.0685	-7471B	PPB	22 Oct 2024 13:07:02	U	SMPL
P4468-05 - 1	ETGI-345	1307	0.3049	-7471B	PPB	22 Oct 2024 13:09:21	U	SMPL
P4470-01 - 1	CL-01-102124	1090	0.2547	-7471B	PPB	22 Oct 2024 13:11:36	U	SMPL
P4471-01 - 1	B-180-SB01	28392	6.5754	-7471B	PPB	22 Oct 2024 13:13:52	U	SMPL
P4471-02 - 1	B-180-SB02	3346	0.7770	-7471B	PPB	22 Oct 2024 13:16:07	U	SMPL
P4472-01 - 1	BP-F-28	258	0.0621	-7471B	PPB	22 Oct 2024 13:18:28	U	SMPL
P4472-05 - 1	BP-F-6	1138	0.2658	-7471B	PPB	22 Oct 2024 13:20:45	U	SMPL
P4443-01LX5 - 1		1377	0.3211	-7471B	PPB	22 Oct 2024 13:23:02	U	SMPL
P4443-01A - 1		25513	5.9088	-7471B	PPB	22 Oct 2024 13:25:19	U	SMPL
P4460-02DLX10 - 1		24153	5.5940	-7471B	PPB	22 Oct 2024 13:27:36	U	SMPL
CCV55 - 1	CCV55	22954	5.3164	-7471B	PPB	22 Oct 2024 13:29:54	U	SMPL
CCB55 - 1	CCB55	-273	-0.0608	-7471B	PPB	22 Oct 2024 13:32:12	U	SMPL