

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4460

Instrument: CV1

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB164322BL		SOLID		Batch Number:	PB164322		Prep Date:	10/21/2024	
	Mercury	0.012	<0.012	U	0.012	CV	10/22/2024	12:13	LB133043
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB164407BL		WATER		Batch Number:	PB164407		Prep Date:	10/24/2024	
	Mercury	0.20	<0.20	U	0.20	CV	10/25/2024	12:42	LB133129

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4460

Instrument: P4

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB164289BL	SOLID			Batch Number:	PB164289		Prep Date:	10/21/2024	
	Aluminum	4.39	<4.39	U	4.39	P	10/24/2024	23:37	LB133110
	Antimony	2.19	<2.19	U	2.19	P	10/24/2024	23:37	LB133110
	Arsenic	0.88	<0.88	U	0.88	P	10/24/2024	23:37	LB133110
	Barium	4.39	<4.39	U	4.39	P	10/24/2024	23:37	LB133110
	Beryllium	0.26	<0.26	U	0.26	P	10/24/2024	23:37	LB133110
	Cadmium	0.26	<0.26	U	0.26	P	10/24/2024	23:37	LB133110
	Calcium	87.7	<87.7	U	87.7	P	10/24/2024	23:37	LB133110
	Chromium	0.44	<0.44	U	0.44	P	10/24/2024	23:37	LB133110
	Cobalt	1.32	<1.32	U	1.32	P	10/24/2024	23:37	LB133110
	Copper	0.88	<0.88	U	0.88	P	10/24/2024	23:37	LB133110
	Iron	4.39	<4.39	U	4.39	P	10/24/2024	23:37	LB133110
	Lead	0.53	<0.53	U	0.53	P	10/24/2024	23:37	LB133110
	Magnesium	87.7	<87.7	U	87.7	P	10/24/2024	23:37	LB133110
	Manganese	0.88	<0.88	U	0.88	P	10/24/2024	23:37	LB133110
	Nickel	1.75	<1.75	U	1.75	P	10/24/2024	23:37	LB133110
	Potassium	87.7	<87.7	U	87.7	P	10/24/2024	23:37	LB133110
	Selenium	0.88	<0.88	U	0.88	P	10/24/2024	23:37	LB133110
	Silver	0.44	<0.44	U	0.44	P	10/24/2024	23:37	LB133110
	Sodium	87.7	<87.7	U	87.7	P	10/24/2024	23:37	LB133110
	Thallium	1.75	<1.75	U	1.75	P	10/24/2024	23:37	LB133110
	Vanadium	1.75	<1.75	U	1.75	P	10/24/2024	23:37	LB133110
	Zinc	1.75	<1.75	U	1.75	P	10/24/2024	23:37	LB133110
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB164347BL	WATER			Batch Number:	PB164347		Prep Date:	10/23/2024	
	Aluminum	50.0	<50.0	U	50.0	P	10/24/2024	19:49	LB133110
	Antimony	25.0	<25.0	U	25.0	P	10/24/2024	19:49	LB133110
	Arsenic	10.0	<10.0	U	10.0	P	10/24/2024	19:49	LB133110
	Barium	50.0	<50.0	U	50.0	P	10/24/2024	19:49	LB133110
	Beryllium	3.00	<3.00	U	3.00	P	10/24/2024	19:49	LB133110
	Cadmium	3.00	<3.00	U	3.00	P	10/24/2024	19:49	LB133110
	Calcium	1000	<1000	U	1000	P	10/24/2024	19:49	LB133110
	Chromium	5.00	<5.00	U	5.00	P	10/24/2024	19:49	LB133110
	Cobalt	15.0	<15.0	U	15.0	P	10/24/2024	19:49	LB133110
	Copper	10.0	<10.0	U	10.0	P	10/24/2024	19:49	LB133110
	Iron	50.0	<50.0	U	50.0	P	10/24/2024	19:49	LB133110
	Lead	6.00	<6.00	U	6.00	P	10/24/2024	19:49	LB133110
	Magnesium	1000	<1000	U	1000	P	10/24/2024	19:49	LB133110
	Manganese	10.0	<10.0	U	10.0	P	10/24/2024	19:49	LB133110
	Nickel	20.0	<20.0	U	20.0	P	10/24/2024	19:49	LB133110

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4460

Instrument: P4

Potassium	1000	<1000	U	1000	P	10/24/2024	19:49	LB133110
Selenium	10.0	<10.0	U	10.0	P	10/24/2024	19:49	LB133110
Silver	5.00	<5.00	U	5.00	P	10/24/2024	19:49	LB133110
Sodium	332	<1000	J	1000	P	10/24/2024	19:49	LB133110
Thallium	20.0	<20.0	U	20.0	P	10/24/2024	19:49	LB133110
Vanadium	20.0	<20.0	U	20.0	P	10/24/2024	19:49	LB133110
Zinc	20.0	<20.0	U	20.0	P	10/24/2024	19:49	LB133110