

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P4460

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4460 **SAS No.:** P4460

Matrix: Solid **Level:** LOW **Client ID:** WB-303-BOTA

Sample ID: P4460-03 **Spiked ID:** P4460-03A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Antimony	mg/Kg	75 - 125	31.2		3.03	U	48.5	64		P
Barium	mg/Kg	75 - 125	29.1		26.9		12.1	18		P
Beryllium	mg/Kg	75 - 125	9.18		0.68		12.1	70		P
Chromium	mg/Kg	75 - 125	31.6		14.6		24.2	70		P
Cobalt	mg/Kg	75 - 125	20.1		10.5		12.1	79		P
Copper	mg/Kg	75 - 125	27.1		16.5		18.2	59		P
Nickel	mg/Kg	75 - 125	44.7		19.8		30.3	82		P
Potassium	mg/Kg	75 - 125	1510		1320		610	31		P
Selenium	mg/Kg	75 - 125	83.3		1.21	U	120	69		P
Vanadium	mg/Kg	75 - 125	28.5		18.2		18.2	57		P
Zinc	mg/Kg	75 - 125	42.8		40.5		12.1	20		P

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P4460

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4460 **SAS No.:** P4460

Matrix: Water **Level:** LOW **Client ID:** WB-303-SWA

Sample ID: P4460-06 **Spiked ID:** P4460-06A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	ug/L	75 - 125	1530		426		10000	11		P
Antimony	ug/L	75 - 125	540		3.58	J	400	134		P
Arsenic	ug/L	75 - 125	496		10.0	U	400	124		P
Selenium	ug/L	75 - 125	1270		10.0	U	1000	127		P
Silver	ug/L	75 - 125	63.1		5.00	U	37.5	168		P
Zinc	ug/L	75 - 125	161		19.8	J	100	141		P

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P4460
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4460 **SAS No.:** P4460
Matrix: Water **Level:** LOW **Client ID:** 34839-40A
Sample ID: P4546-03 **Spiked ID:** P4546-03A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	82 - 119	4.38		0.28		4.00	103		CV