



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Portal Partners Tri-Venture				SDG No.:	P4460				
Contract:	PORT06	Lab Code:	CHEM		Case No.:	P4460	SAS No.:	P4460		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB48	Mercury	0.20	+/-0.20	U	0.20	CV	10/23/2024	09:46	LB133065	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4460

Contract: PORT06

Lab Code: CHEM

Case No.: P4460

SAS No.: P4460

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB56	Mercury	0.20	+/-0.20	U	0.20	CV	10/23/2024	09:51	LB133065
CCB57	Mercury	0.20	+/-0.20	U	0.20	CV	10/23/2024	10:24	LB133065
CCB58	Mercury	0.20	+/-0.20	U	0.20	CV	10/23/2024	10:51	LB133065
CCB59	Mercury	0.20	+/-0.20	U	0.20	CV	10/23/2024	11:18	LB133065

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4460

Contract: PORT06

Lab Code: CHEM

Case No.: P4460

SAS No.: P4460

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	10/23/2024	15:16	LB133086
	Barium	100	+/-100	U	100	P	10/23/2024	15:16	LB133086
	Cadmium	6.00	+/-6.00	U	6.00	P	10/23/2024	15:16	LB133086
	Chromium	10.0	+/-10.0	U	10.0	P	10/23/2024	15:16	LB133086
	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	15:16	LB133086
	Selenium	20.0	+/-20.0	U	20.0	P	10/23/2024	15:16	LB133086
	Silver	10.0	+/-10.0	U	10.0	P	10/23/2024	15:16	LB133086

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4460							
Contract: PORT06	Lab Code: CHEM	Case No.: P4460	SAS No.: P4460						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	10/23/2024	16:02	LB133086
	Barium	100	+/-100	U	100	P	10/23/2024	16:02	LB133086
	Cadmium	1.78	+/-6.00	J	6.00	P	10/23/2024	16:02	LB133086
	Chromium	10.0	+/-10.0	U	10.0	P	10/23/2024	16:02	LB133086
	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	16:02	LB133086
	Selenium	20.0	+/-20.0	U	20.0	P	10/23/2024	16:02	LB133086
	Silver	1.20	+/-10.0	J	10.0	P	10/23/2024	16:02	LB133086
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	10/23/2024	16:52	LB133086
	Barium	100	+/-100	U	100	P	10/23/2024	16:52	LB133086
	Cadmium	6.00	+/-6.00	U	6.00	P	10/23/2024	16:52	LB133086
	Chromium	10.0	+/-10.0	U	10.0	P	10/23/2024	16:52	LB133086
	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	16:52	LB133086
	Selenium	20.0	+/-20.0	U	20.0	P	10/23/2024	16:52	LB133086
	Silver	10.0	+/-10.0	U	10.0	P	10/23/2024	16:52	LB133086
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	10/23/2024	17:46	LB133086
	Barium	100	+/-100	U	100	P	10/23/2024	17:46	LB133086
	Cadmium	6.00	+/-6.00	U	6.00	P	10/23/2024	17:46	LB133086
	Chromium	10.0	+/-10.0	U	10.0	P	10/23/2024	17:46	LB133086
	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	17:46	LB133086
	Selenium	20.0	+/-20.0	U	20.0	P	10/23/2024	17:46	LB133086
	Silver	10.0	+/-10.0	U	10.0	P	10/23/2024	17:46	LB133086
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	10/23/2024	18:36	LB133086
	Barium	100	+/-100	U	100	P	10/23/2024	18:36	LB133086
	Cadmium	6.00	+/-6.00	U	6.00	P	10/23/2024	18:36	LB133086
	Chromium	10.0	+/-10.0	U	10.0	P	10/23/2024	18:36	LB133086
	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	18:36	LB133086
	Selenium	20.0	+/-20.0	U	20.0	P	10/23/2024	18:36	LB133086
	Silver	10.0	+/-10.0	U	10.0	P	10/23/2024	18:36	LB133086
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	10/23/2024	19:28	LB133086
	Barium	100	+/-100	U	100	P	10/23/2024	19:28	LB133086
	Cadmium	6.00	+/-6.00	U	6.00	P	10/23/2024	19:28	LB133086
	Chromium	10.0	+/-10.0	U	10.0	P	10/23/2024	19:28	LB133086
	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	19:28	LB133086
	Selenium	20.0	+/-20.0	U	20.0	P	10/23/2024	19:28	LB133086
	Silver	10.0	+/-10.0	U	10.0	P	10/23/2024	19:28	LB133086
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	10/23/2024	20:21	LB133086
	Barium	100	+/-100	U	100	P	10/23/2024	20:21	LB133086
	Cadmium	6.00	+/-6.00	U	6.00	P	10/23/2024	20:21	LB133086
	Chromium	10.0	+/-10.0	U	10.0	P	10/23/2024	20:21	LB133086

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture		SDG No.: P4460							
Contract: PORT06	Lab Code: CHEM	Case No.: P4460	SAS No.: P4460						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	20:21	LB133086
	Selenium	20.0	+/-20.0	U	20.0	P	10/23/2024	20:21	LB133086
	Silver	10.0	+/-10.0	U	10.0	P	10/23/2024	20:21	LB133086
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	10/23/2024	21:34	LB133086
	Barium	100	+/-100	U	100	P	10/23/2024	21:34	LB133086
	Cadmium	6.00	+/-6.00	U	6.00	P	10/23/2024	21:34	LB133086
	Chromium	10.0	+/-10.0	U	10.0	P	10/23/2024	21:34	LB133086
	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	21:34	LB133086
	Selenium	20.0	+/-20.0	U	20.0	P	10/23/2024	21:34	LB133086
	Silver	10.0	+/-10.0	U	10.0	P	10/23/2024	21:34	LB133086
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	10/23/2024	22:25	LB133086
	Barium	100	+/-100	U	100	P	10/23/2024	22:25	LB133086
	Cadmium	6.00	+/-6.00	U	6.00	P	10/23/2024	22:25	LB133086
	Chromium	10.0	+/-10.0	U	10.0	P	10/23/2024	22:25	LB133086
	Lead	12.0	+/-12.0	U	12.0	P	10/23/2024	22:25	LB133086
	Selenium	20.0	+/-20.0	U	20.0	P	10/23/2024	22:25	LB133086
	Silver	10.0	+/-10.0	U	10.0	P	10/23/2024	22:25	LB133086

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4460

Contract: PORT06

Lab Code: CHEM

Case No.: P4460

SAS No.: P4460

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------