

Continuing Calibration Report for SequenceID : FC102224AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067484.D

Aliphatic C9-C12	7061530.000	60.000	3.175	6.455	117692.167	128354.488	8.307
Aliphatic C12-C16	4662702.000	40.000	6.456	9.845	116567.550	130957.677	10.988
Aliphatic C16-C21	6773563.000	60.000	9.846	13.203	112892.717	131476.102	14.134
Aliphatic C21-C28	9096898.000	80.000	13.204	16.857	113711.225	126823.664	10.339
Aliphatic C28-C40	11940513.000	120.000	16.858	21.701	99504.275	105215.848	5.428
Aliphatic EPH	39535206.000	360.000	3.175	21.701	109820.017	121110.937	9.323