

Continuing Calibration Report for SequenceID : FC102224AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067497.D

Aliphatic C9-C12	7120345.000	60.000	3.175	6.455	118672.417	128354.488	7.543
Aliphatic C12-C16	4707454.000	40.000	6.456	9.845	117686.350	130957.677	10.134
Aliphatic C16-C21	6832576.000	60.000	9.846	13.203	113876.267	131476.102	13.386
Aliphatic C21-C28	9183247.000	80.000	13.204	16.857	114790.588	126823.664	9.488
Aliphatic C28-C40	12317337.000	120.000	16.858	21.701	102644.475	105215.848	2.444
Aliphatic EPH	40160959.000	360.000	3.175	21.701	111558.219	121110.937	7.888