

Continuing Calibration Report for SequenceID : FC102224AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067503.D

Aliphatic C9-C12	7278914.000	60.000	3.175	6.455	121315.233	128354.488	5.484
Aliphatic C12-C16	4776060.000	40.000	6.456	9.845	119401.500	130957.677	8.824
Aliphatic C16-C21	6956163.000	60.000	9.846	13.203	115936.050	131476.102	11.820
Aliphatic C21-C28	9397788.000	80.000	13.204	16.857	117472.350	126823.664	7.373
Aliphatic C28-C40	12533926.000	120.000	16.858	21.701	104449.383	105215.848	0.728
Aliphatic EPH	40942851.000	360.000	3.175	21.701	113730.142	121110.937	6.094