

Continuing Calibration Report for SequenceID : FC102424AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067516.D

Aliphatic C9-C12	7891359.000	60.000	3.174	6.454	131522.650	135469.641	2.914
Aliphatic C12-C16	5116825.000	40.000	6.455	9.844	127920.625	131681.730	2.856
Aliphatic C16-C21	7443152.000	60.000	9.845	13.202	124052.533	127235.092	2.501
Aliphatic C21-C28	9423790.000	80.000	13.203	16.857	117797.375	120090.398	1.909
Aliphatic C28-C40	12056218.000	120.000	16.858	21.703	100468.483	101093.195	0.618
Aliphatic EPH	41931344.000	360.000	3.174	21.703	116475.956	118799.912	1.956