

Continuing Calibration Report for SequenceID : FC102424AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067522.D

Aliphatic C9-C12	8032920.000	60.000	3.174	6.454	133882.000	135469.641	1.172
Aliphatic C12-C16	5187653.000	40.000	6.455	9.844	129691.325	131681.730	1.512
Aliphatic C16-C21	7523928.000	60.000	9.845	13.202	125398.800	127235.092	1.443
Aliphatic C21-C28	9556657.000	80.000	13.203	16.857	119458.213	120090.398	0.526
Aliphatic C28-C40	12426370.000	120.000	16.858	21.703	103553.083	101093.195	-2.433
Aliphatic EPH	42727528.000	360.000	3.174	21.703	118687.578	118799.912	0.095