

Continuing Calibration Report for SequenceID : FD102424AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048578.D

Aromatic C10-C12	7282027.000	40.000	4.095	5.813	182050.675	176202.694	-3.319
Aromatic C12-C16	10947020.000	60.000	5.814	8.419	182450.333	177273.493	-2.920
Aromatic C16-C21	13483086.000	80.000	8.420	12.681	168538.575	163832.550	-2.872
Aromatic C21-C36	28447129.000	180.000	12.682	18.087	158039.606	152713.153	-3.488
Aromatic EPH	60159262.000	360.000	4.095	18.087	167109.061	161887.469	-3.225