

Continuing Calibration Report for SequenceID : FD102424AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048584.D

Aromatic C10-C12	6996046.000	40.000	4.095	5.813	174901.150	176202.694	0.739
Aromatic C12-C16	10520876.000	60.000	5.814	8.419	175347.933	177273.493	1.086
Aromatic C16-C21	12885484.000	80.000	8.420	12.681	161068.550	163832.550	1.687
Aromatic C21-C36	27316669.000	180.000	12.682	18.087	151759.272	152713.153	0.625
Aromatic EPH	57719075.000	360.000	4.095	18.087	160330.764	161887.469	0.962