

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** P4460 **SAS No :** P4460 **SDG No:** P4460
Sample No : PB164309BS **Datafile:** FC067486.D
Client ID : PB164309BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C9-C12	10	5.99	60		(40-140)
Aliphatic C12-C16	6.7	5.14	77		(40-140)
Aliphatic C16-C21	10	8.64	86		(40-140)
Aliphatic C21-C28	13.3	11.7	88		(40-140)
Aliphatic C28-C40	20.0	21.2	106		(40-140)