

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** P4460 **SAS No :** P4460 **SDG No:** P4460
Sample No : PB164368BS **Datafile:** FC067518.D
Client ID : PB164368BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aliphatic C9-C12	300	176	59		(40-140)
Aliphatic C12-C16	200	159	80		(40-140)
Aliphatic C16-C21	300	281	94		(40-140)
Aliphatic C21-C28	400	391	98		(40-140)
Aliphatic C28-C40	600	692	115		(40-140)