

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** P4460 **SAS No :** P4460 **SDG No:** P4460
Sample No : PB164368BSD **Datafile:** FC067519.D
Client ID : PB164368BSD

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C9-C12	300	174	58		1.2 (40-140)	50
Aliphatic C12-C16	200	157	78		1.3 (40-140)	50
Aliphatic C16-C21	300	277	92		1.5 (40-140)	50
Aliphatic C21-C28	400	387	97		1.1 (40-140)	50
Aliphatic C28-C40	600	686	114		0.581 (40-140)	50