

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** P4460 **SAS No :** P4460 **SDG No:** P4460
Sample No : PB164368BSD **Datafile:** FD048581.D
Client ID : PB164368BSD

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aromatic C10-C12	200	174	87		0 (40-140)	50
Aromatic C12-C16	300	291	97		0.309 (40-140)	50
Aromatic C16-C21	500	610	122		0 (40-140)	50
Aromatic C21-C36	800	923	115		0.32 (40-140)	50