

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 10/21/2024

Lab Sample No.: AR1660CCC500 Data File : PO107293.D Time Analyzed: 11:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.521	5.423	5.623	506.980	500.000	1.4
Aroclor-1016-2	5.544	5.445	5.645	491.800	500.000	-1.6
Aroclor-1016-3	5.605	5.507	5.707	459.940	500.000	-8.0
Aroclor-1016-4	5.702	5.604	5.804	484.940	500.000	-3.0
Aroclor-1016-5	5.997	5.899	6.099	482.770	500.000	-3.4
Aroclor-1260-1	7.124	7.026	7.226	492.220	500.000	-1.6
Aroclor-1260-2	7.379	7.282	7.482	524.180	500.000	4.8
Aroclor-1260-3	7.742	7.643	7.843	510.390	500.000	2.1
Aroclor-1260-4	7.966	7.868	8.068	510.260	500.000	2.1
Aroclor-1260-5	8.280	8.181	8.381	544.830	500.000	9.0
Decachlorobiphenyl	10.057	9.958	10.158	50.010	50.000	0.0
Tetrachloro-m-xylene	4.373	4.274	4.474	51.260	50.000	2.5

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 10/21/2024

Lab Sample No.: AR1660CCC500 Data File : PO107293.D Time Analyzed: 11:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.626	4.826	514.630	500.000	2.9
Aroclor-1016-2	4.745	4.645	4.845	527.080	500.000	5.4
Aroclor-1016-3	4.921	4.821	5.021	492.510	500.000	-1.5
Aroclor-1016-4	4.961	4.862	5.062	451.280	500.000	-9.7
Aroclor-1016-5	5.175	5.076	5.276	491.380	500.000	-1.7
Aroclor-1260-1	6.206	6.107	6.307	518.020	500.000	3.6
Aroclor-1260-2	6.393	6.294	6.494	537.770	500.000	7.6
Aroclor-1260-3	6.546	6.448	6.648	527.570	500.000	5.5
Aroclor-1260-4	7.016	6.918	7.118	517.840	500.000	3.6
Aroclor-1260-5	7.257	7.159	7.359	546.300	500.000	9.3
Decachlorobiphenyl	8.636	8.539	8.739	51.800	50.000	3.6
Tetrachloro-m-xylene	3.645	3.545	3.745	52.920	50.000	5.8