

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 10/21/2024

Lab Sample No.: AR1660CCC500 Data File : PO107308.D Time Analyzed: 18:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.521	5.423	5.623	483.570	500.000	-3.3
Aroclor-1016-2	5.543	5.445	5.645	471.960	500.000	-5.6
Aroclor-1016-3	5.605	5.507	5.707	442.600	500.000	-11.5
Aroclor-1016-4	5.702	5.604	5.804	476.570	500.000	-4.7
Aroclor-1016-5	5.996	5.899	6.099	497.990	500.000	-0.4
Aroclor-1260-1	7.123	7.026	7.226	499.660	500.000	-0.1
Aroclor-1260-2	7.378	7.282	7.482	507.460	500.000	1.5
Aroclor-1260-3	7.740	7.643	7.843	513.820	500.000	2.8
Aroclor-1260-4	7.965	7.868	8.068	504.850	500.000	1.0
Aroclor-1260-5	8.278	8.181	8.381	514.530	500.000	2.9
Decachlorobiphenyl	10.053	9.958	10.158	48.130	50.000	-3.7
Tetrachloro-m-xylene	4.372	4.274	4.474	50.560	50.000	1.1

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 10/21/2024

Lab Sample No.: AR1660CCC500 Data File : PO107308.D Time Analyzed: 18:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.626	4.826	512.880	500.000	2.6
Aroclor-1016-2	4.745	4.645	4.845	533.630	500.000	6.7
Aroclor-1016-3	4.920	4.821	5.021	489.740	500.000	-2.1
Aroclor-1016-4	4.961	4.862	5.062	457.260	500.000	-8.5
Aroclor-1016-5	5.175	5.076	5.276	513.220	500.000	2.6
Aroclor-1260-1	6.206	6.107	6.307	486.800	500.000	-2.6
Aroclor-1260-2	6.393	6.294	6.494	524.290	500.000	4.9
Aroclor-1260-3	6.547	6.448	6.648	495.030	500.000	-1.0
Aroclor-1260-4	7.017	6.918	7.118	478.400	500.000	-4.3
Aroclor-1260-5	7.257	7.159	7.359	487.470	500.000	-2.5
Decachlorobiphenyl	8.637	8.539	8.739	44.100	50.000	-11.8
Tetrachloro-m-xylene	3.645	3.545	3.745	53.390	50.000	6.8