

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL03 Date Analyzed: 10/21/2024

Lab Sample No.: AR1660CCC500 Data File : PO107323.D Time Analyzed: 23:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.521	5.423	5.623	519.640	500.000	3.9
Aroclor-1016-2	5.543	5.445	5.645	506.910	500.000	1.4
Aroclor-1016-3	5.606	5.507	5.707	479.170	500.000	-4.2
Aroclor-1016-4	5.701	5.604	5.804	506.100	500.000	1.2
Aroclor-1016-5	5.997	5.899	6.099	508.720	500.000	1.7
Aroclor-1260-1	7.122	7.026	7.226	535.060	500.000	7.0
Aroclor-1260-2	7.379	7.282	7.482	547.990	500.000	9.6
Aroclor-1260-3	7.740	7.643	7.843	510.820	500.000	2.2
Aroclor-1260-4	7.964	7.868	8.068	540.140	500.000	8.0
Aroclor-1260-5	8.278	8.181	8.381	580.330	500.000	16.1
Decachlorobiphenyl	10.055	9.958	10.158	49.390	50.000	-1.2
Tetrachloro-m-xylene	4.372	4.274	4.474	52.850	50.000	5.7

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL03 Date Analyzed: 10/21/2024

Lab Sample No.: AR1660CCC500 Data File : PO107323.D Time Analyzed: 23:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.626	4.826	522.590	500.000	4.5
Aroclor-1016-2	4.745	4.645	4.845	548.030	500.000	9.6
Aroclor-1016-3	4.921	4.821	5.021	508.560	500.000	1.7
Aroclor-1016-4	4.962	4.862	5.062	466.520	500.000	-6.7
Aroclor-1016-5	5.175	5.076	5.276	555.540	500.000	11.1
Aroclor-1260-1	6.206	6.107	6.307	543.800	500.000	8.8
Aroclor-1260-2	6.393	6.294	6.494	572.690	500.000	14.5
Aroclor-1260-3	6.547	6.448	6.648	543.390	500.000	8.7
Aroclor-1260-4	7.018	6.918	7.118	544.220	500.000	8.8
Aroclor-1260-5	7.257	7.159	7.359	556.140	500.000	11.2
Decachlorobiphenyl	8.637	8.539	8.739	49.000	50.000	-2.0
Tetrachloro-m-xylene	3.645	3.545	3.745	53.780	50.000	7.6