

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL04 Date Analyzed: 10/23/2024

Lab Sample No.: AR1660CCC500 Data File : PP068166.D Time Analyzed: 15:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.921	5.819	6.019	491.140	500.000	-1.8
Aroclor-1016-2	5.944	5.842	6.042	505.430	500.000	1.1
Aroclor-1016-3	6.007	5.905	6.105	491.780	500.000	-1.6
Aroclor-1016-4	6.105	6.003	6.203	485.430	500.000	-2.9
Aroclor-1016-5	6.399	6.298	6.498	476.670	500.000	-4.7
Aroclor-1260-1	7.523	7.422	7.622	449.530	500.000	-10.1
Aroclor-1260-2	7.776	7.675	7.875	448.260	500.000	-10.3
Aroclor-1260-3	8.138	8.037	8.237	457.380	500.000	-8.5
Aroclor-1260-4	8.376	8.275	8.475	456.620	500.000	-8.7
Aroclor-1260-5	8.713	8.612	8.812	462.630	500.000	-7.5
Decachlorobiphenyl	10.670	10.569	10.769	48.800	50.000	-2.4
Tetrachloro-m-xylene	4.757	4.655	4.855	51.470	50.000	2.9

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL04 Date Analyzed: 10/23/2024

Lab Sample No.: AR1660CCC500 Data File : PP068166.D Time Analyzed: 15:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.157	5.061	5.261	465.090	500.000	-7.0
Aroclor-1016-2	5.177	5.081	5.281	473.530	500.000	-5.3
Aroclor-1016-3	5.357	5.262	5.462	476.580	500.000	-4.7
Aroclor-1016-4	5.397	5.301	5.501	471.310	500.000	-5.7
Aroclor-1016-5	5.616	5.520	5.720	467.820	500.000	-6.4
Aroclor-1260-1	6.659	6.564	6.764	451.860	500.000	-9.6
Aroclor-1260-2	6.846	6.750	6.950	457.130	500.000	-8.6
Aroclor-1260-3	7.004	6.908	7.108	455.780	500.000	-8.8
Aroclor-1260-4	7.478	7.383	7.583	458.570	500.000	-8.3
Aroclor-1260-5	7.716	7.622	7.822	455.920	500.000	-8.8
Decachlorobiphenyl	9.209	9.119	9.319	46.300	50.000	-7.4
Tetrachloro-m-xylene	4.050	3.952	4.152	48.480	50.000	-3.0