

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL05 Date Analyzed: 10/23/2024

Lab Sample No.: AR1660CCC500 Data File : PP068180.D Time Analyzed: 20:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.923	5.819	6.019	497.330	500.000	-0.5
Aroclor-1016-2	5.945	5.842	6.042	503.300	500.000	0.7
Aroclor-1016-3	6.008	5.905	6.105	489.880	500.000	-2.0
Aroclor-1016-4	6.106	6.003	6.203	480.170	500.000	-4.0
Aroclor-1016-5	6.400	6.298	6.498	480.560	500.000	-3.9
Aroclor-1260-1	7.525	7.422	7.622	461.200	500.000	-7.8
Aroclor-1260-2	7.778	7.675	7.875	466.470	500.000	-6.7
Aroclor-1260-3	8.139	8.037	8.237	469.470	500.000	-6.1
Aroclor-1260-4	8.378	8.275	8.475	471.690	500.000	-5.7
Aroclor-1260-5	8.714	8.612	8.812	474.540	500.000	-5.1
Decachlorobiphenyl	10.672	10.569	10.769	49.760	50.000	-0.5
Tetrachloro-m-xylene	4.758	4.655	4.855	51.720	50.000	3.4

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL05 Date Analyzed: 10/23/2024

Lab Sample No.: AR1660CCC500 Data File : PP068180.D Time Analyzed: 20:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.158	5.061	5.261	466.650	500.000	-6.7
Aroclor-1016-2	5.178	5.081	5.281	471.050	500.000	-5.8
Aroclor-1016-3	5.358	5.262	5.462	476.420	500.000	-4.7
Aroclor-1016-4	5.397	5.301	5.501	471.550	500.000	-5.7
Aroclor-1016-5	5.617	5.520	5.720	475.810	500.000	-4.8
Aroclor-1260-1	6.660	6.564	6.764	465.690	500.000	-6.9
Aroclor-1260-2	6.847	6.750	6.950	475.110	500.000	-5.0
Aroclor-1260-3	7.005	6.908	7.108	465.600	500.000	-6.9
Aroclor-1260-4	7.478	7.383	7.583	470.220	500.000	-6.0
Aroclor-1260-5	7.717	7.622	7.822	470.390	500.000	-5.9
Decachlorobiphenyl	9.212	9.119	9.319	48.140	50.000	-3.7
Tetrachloro-m-xylene	4.051	3.952	4.152	47.960	50.000	-4.1