

Surrogate Summary

SDG No.: **P4460**

Client: **Portal Partners Tri-Venture**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO107183.D	PIBLK-PO107183.D	Tetrachloro-m-xylene	1	20	22.5	112		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	23.4	117		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	22.1	110		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	22.8	114		70 (60)	130 (140)
I.BLK-PO107297.D	PIBLK-PO107297.D	Tetrachloro-m-xylene	1	20	24.1	121		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	23.0	115		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	24.1	120		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	22.6	113		70 (60)	130 (140)
PB164287BL	PB164287BL	Tetrachloro-m-xylene	1	20	21.4	107		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	21.7	108		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.4	107		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	21.9	110		30 (32)	150 (175)
PB164287BS	PB164287BS	Tetrachloro-m-xylene	1	20	21.8	109		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	21.5	107		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	21.3	106		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	22.9	114		30 (32)	150 (175)
P4455-01MS	SU-4-101824MS	Tetrachloro-m-xylene	1	20	16.1	81		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	16.3	82		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	23.2	116		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	14.3	71		30 (32)	150 (175)
P4455-01MSD	SU-4-101824MSD	Tetrachloro-m-xylene	1	20	18.2	91		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	15.0	75		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	23.4	117		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	13.3	67		30 (32)	150 (175)
I.BLK-PO107312.D	PIBLK-PO107312.D	Tetrachloro-m-xylene	1	20	24.2	121		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	29.0	145	*	70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	24.2	121		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	27.2	136	*	70 (60)	130 (140)
P4460-02	WB-303-TOP	Tetrachloro-m-xylene	1	20	11.7	59		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	10.4	52		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	17.0	85		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	11.3	57		30 (32)	150 (175)
P4460-03	WB-303-BOT	Tetrachloro-m-xylene	1	20	16.4	82		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	14.9	74		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	18.0	90		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	13.2	66		30 (32)	150 (175)
I.BLK-PO107327.D	PIBLK-PO107327.D	Tetrachloro-m-xylene	1	20	24.4	122		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	21.9	110		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	24.4	122		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	21.4	107		70 (60)	130 (140)
I.BLK-PP067586.D	PIBLK-PP067586.D	Tetrachloro-m-xylene	1	20	21.1	105		70 (60)	130 (140)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: **P4460**

Client: **Portal Partners Tri-Venture**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP067586.D	PIBLK-PP067586.D	Decachlorobiphenyl	1	20	23.6	118		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	22.1	110		70 (60)	130 (140)
I.BLK-PP068170.D	PIBLK-PP068170.D	Decachlorobiphenyl	2	20	23.5	117		70 (60)	130 (140)
		Tetrachloro-m-xylene	1	20	20.8	104		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	21.9	110		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	19.5	98		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.9	104		70 (60)	130 (140)
PB164342BL	PB164342BL	Tetrachloro-m-xylene	1	20	20.1	101		30 (10)	150 (157)
		Decachlorobiphenyl	1	20	21.6	108		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	19.0	95		30 (10)	150 (157)
		Decachlorobiphenyl	2	20	21.1	105		30 (10)	150 (173)
PB164342BS	PB164342BS	Tetrachloro-m-xylene	1	20	20.8	104		30 (10)	150 (157)
		Decachlorobiphenyl	1	20	21.4	107		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	18.0	90		30 (10)	150 (157)
		Decachlorobiphenyl	2	20	20.9	104		30 (10)	150 (173)
PB164342BSD	PB164342BSD	Tetrachloro-m-xylene	1	20	20.6	103		30 (10)	150 (157)
		Decachlorobiphenyl	1	20	22.2	111		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	18.4	92		30 (10)	150 (157)
		Decachlorobiphenyl	2	20	21.6	108		30 (10)	150 (173)
P4460-06	WB-303-SW	Tetrachloro-m-xylene	1	20	18.5	92		30 (10)	150 (157)
		Decachlorobiphenyl	1	20	20.7	104		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	17.9	89		30 (10)	150 (157)
		Decachlorobiphenyl	2	20	19.9	100		30 (10)	150 (173)
		Tetrachloro-m-xylene	1	20	20.9	105		70 (60)	130 (140)
I.BLK-PP068184.D	PIBLK-PP068184.D	Decachlorobiphenyl	1	20	22.5	112		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	19.6	98		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	21.6	108		70 (60)	130 (140)
		Tetrachloro-m-xylene	1	20	20.9	105		70 (60)	130 (140)