

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

Instrument ID: ECD_S Calibration Date(s): 10/23/2024 10/23/2024
Calibration Times: 11:28 13:04

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS028008.D</u>	CF 500 =	<u>PS028009.D</u>		
CF 750 =		<u>PS028010.D</u>	CF 1000 =	<u>PS028011.D</u>	CF 1500 =	<u>PS028012.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	19584000000	16293000000	15358800000	15287500000	14230600000	16150800000	13
2,4-D	4140980000	3388880000	3183890000	3192570000	3036110000	3388490000	13
2,4-DCAA	3343560000	2711910000	2549070000	2475610000	2404680000	2696970000	14

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

Instrument ID: ECD_S Calibration Date(s): 10/23/2024 10/23/2024

Calibration Times: 11:28 13:04

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS028008.D</u>	CF 500 =	<u>PS028009.D</u>		
CF 750 =		<u>PS028010.D</u>	CF 1000 =	<u>PS028011.D</u>	CF 1500 =	<u>PS028012.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	4575720000	4585960000	4376840000	4595510000	4689140000	4564630000	3
2,4-D	1125240000	990045000	959343000	966564000	944531000	997145000	7
2,4-DCAA	1149150000	929743000	912345000	889645000	862510000	948678000	12