

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 10/23/2024 10/23/2024

Client Sample No.: CCAL01 Date Analyzed: 10/24/2024

Lab Sample No.: HSTDCCC750 Data File : PS028036.D Time Analyzed: 11:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.028	8.929	9.129	684.860	712.500	-3.9
2,4-D	8.176	8.077	8.277	665.580	705.000	-5.6
2,4-DCAA	7.091	6.991	7.191	709.960	750.000	-5.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/23/2024 10/23/2024

Client Sample No.: CCAL01 Date Analyzed: 10/24/2024

Lab Sample No.: HSTDCCC750 Data File : PS028036.D Time Analyzed: 11:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.725	9.625	9.825	736.810	712.500	3.4
2,4-D	8.835	8.737	8.937	709.890	705.000	0.7
2,4-DCAA	7.613	7.514	7.714	733.180	750.000	-2.2