

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 10/23/2024 10/23/2024

Client Sample No.: CCAL03 Date Analyzed: 10/25/2024

Lab Sample No.: HSTDCCC750 Data File : PS028050.D Time Analyzed: 12:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.029	8.929	9.129	702.550	712.500	-1.4
2,4-D	8.177	8.077	8.277	696.610	705.000	-1.2
2,4-DCAA	7.091	6.991	7.191	734.600	750.000	-2.1

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/23/2024 10/23/2024

Client Sample No.: CCAL03 Date Analyzed: 10/25/2024

Lab Sample No.: HSTDCCC750 Data File : PS028050.D Time Analyzed: 12:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.722	9.625	9.825	670.830	712.500	-5.8
2,4-D	8.835	8.737	8.937	632.170	705.000	-10.3
2,4-DCAA	7.610	7.514	7.714	719.220	750.000	-4.1