

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 10/23/2024 10/23/2024

Client Sample No.: CCAL04 Date Analyzed: 10/25/2024

Lab Sample No.: HSTDCCC750 Data File : PS028056.D Time Analyzed: 18:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.029	8.929	9.129	713.180	712.500	0.1
2,4-D	8.176	8.077	8.277	698.470	705.000	-0.9
2,4-DCAA	7.090	6.991	7.191	745.250	750.000	-0.6

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/23/2024 10/23/2024

Client Sample No.: CCAL04 Date Analyzed: 10/25/2024

Lab Sample No.: HSTDCCC750 Data File : PS028056.D Time Analyzed: 18:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.722	9.625	9.825	812.130	712.500	14.0
2,4-D	8.833	8.737	8.937	721.570	705.000	2.4
2,4-DCAA	7.611	7.514	7.714	745.670	750.000	-0.6