

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM **Case No.:** P4460 **SAS No.:** P4460 **SDG NO.:** P4460

GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/23/2024 10/23/2024

Client Sample No.: CCAL05 **Date Analyzed:** 10/28/2024

Lab Sample No.: HSTDCCC750 **Data File :** PS028072.D **Time Analyzed:** 10:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.025	8.929	9.129	699.070	712.500	-1.9
2,4-D	8.173	8.077	8.277	675.860	705.000	-4.1
2,4-DCAA	7.088	6.991	7.191	727.270	750.000	-3.0

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/23/2024 10/23/2024

Client Sample No.: CCAL05 Date Analyzed: 10/28/2024

Lab Sample No.: HSTDCCC750 Data File : PS028072.D Time Analyzed: 10:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.719	9.625	9.825	764.520	712.500	7.3
2,4-D	8.832	8.737	8.937	690.400	705.000	-2.1
2,4-DCAA	7.609	7.514	7.714	754.150	750.000	0.6