

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM **Case No.:** P4460 **SAS No.:** P4460 **SDG NO.:** P4460

GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/23/2024 10/23/2024

Client Sample No.: CCAL06 **Date Analyzed:** 10/28/2024

Lab Sample No.: HSTDCCC750 **Data File :** PS028076.D **Time Analyzed:** 17:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.029	8.929	9.129	715.650	712.500	0.4
2,4-D	8.176	8.077	8.277	699.590	705.000	-0.8
2,4-DCAA	7.090	6.991	7.191	733.900	750.000	-2.1

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/23/2024 10/23/2024

Client Sample No.: CCAL06 Date Analyzed: 10/28/2024

Lab Sample No.: HSTDCCC750 Data File : PS028076.D Time Analyzed: 17:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.719	9.625	9.825	743.740	712.500	4.4
2,4-D	8.829	8.737	8.937	729.960	705.000	3.5
2,4-DCAA	7.606	7.514	7.714	729.190	750.000	-2.8