

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS102324

Review By	Abdul	Review On	10/24/2024 2:22:53 PM
Supervise By	Ankita	Supervise On	10/24/2024 2:25:32 PM
SubDirectory	PS102324	HP Acquire Method	HP Processing Method ps102324 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC Internal Standard/PEM	PP23462		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23469		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS028006.D	23 Oct 2024 10:40	AR\AJ	Ok
2	I.BLK	PS028007.D	23 Oct 2024 11:04	AR\AJ	Ok
3	HSTDICC200	PS028008.D	23 Oct 2024 11:28	AR\AJ	Ok,M
4	HSTDICC500	PS028009.D	23 Oct 2024 11:52	AR\AJ	Ok
5	HSTDICC750	PS028010.D	23 Oct 2024 12:16	AR\AJ	Ok
6	HSTDICC1000	PS028011.D	23 Oct 2024 12:40	AR\AJ	Ok,M
7	HSTDICC1500	PS028012.D	23 Oct 2024 13:04	AR\AJ	Ok
8	HSTDICV750	PS028013.D	23 Oct 2024 13:28	AR\AJ	Ok
9	I.BLK	PS028014.D	23 Oct 2024 14:14	AR\AJ	Ok
10	HSTDCCC750	PS028015.D	23 Oct 2024 14:38	AR\AJ	Ok,M
11	P4443-01	PS028016.D	23 Oct 2024 15:02	AR\AJ	Ok,M
12	P4443-06	PS028017.D	23 Oct 2024 15:26	AR\AJ	Ok,M
13	P4458-01	PS028018.D	23 Oct 2024 15:49	AR\AJ	Ok,M
14	P4458-01MS	PS028019.D	23 Oct 2024 16:13	AR\AJ	Ok,M
15	P4458-01MSD	PS028020.D	23 Oct 2024 16:37	AR\AJ	Ok,M
16	PB174307BL	PS028021.D	23 Oct 2024 17:01	AR\AJ	Ok,M
17	PB174307BS	PS028022.D	23 Oct 2024 17:25	AR\AJ	Ok
18	I.BLK	PS028023.D	23 Oct 2024 18:14	AR\AJ	Ok
19	HSTDCCC750	PS028024.D	23 Oct 2024 18:38	AR\AJ	Ok,M
20	P4468-03	PS028025.D	23 Oct 2024 19:02	AR\AJ	Ok,M
21	P4468-05	PS028026.D	23 Oct 2024 19:26	AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS102324

Review By	Abdul	Review On	10/24/2024 2:22:53 PM
Supervise By	Ankita	Supervise On	10/24/2024 2:25:32 PM
SubDirectory	PS102324	HP Acquire Method	HP Processing Method ps102324 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC Internal Standard/PEM	PP23462		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23469		

22	P4467-01	PS028027.D	23 Oct 2024 19:50	AR\AJ	Ok
23	I.BLK	PS028028.D	24 Oct 2024 01:27	AR\AJ	Ok
24	HSTDCCC750	PS028029.D	24 Oct 2024 01:51	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS102424

Review By	Abdul	Review On	10/25/2024 3:28:01 PM
Supervise By	Ankita	Supervise On	10/28/2024 9:12:45 AM
SubDirectory	PS102424	HP Acquire Method	HP Processing Method ps102324 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC Internal Standard/PEM	PP23462		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23469		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS028030.D	24 Oct 2024 09:02	AR\AJ	Ok
2	I.BLK	PS028031.D	24 Oct 2024 09:26	AR\AJ	Ok,M
3	HSTDCCC750	PS028032.D	24 Oct 2024 09:50	AR\AJ	Ok
4	P4472-01	PS028033.D	24 Oct 2024 10:13	AR\AJ	Ok
5	P4472-05	PS028034.D	24 Oct 2024 10:37	AR\AJ	Ok
6	I.BLK	PS028035.D	24 Oct 2024 11:01	AR\AJ	Ok,M
7	HSTDCCC750	PS028036.D	24 Oct 2024 11:25	AR\AJ	Ok
8	PB164378BL	PS028037.D	24 Oct 2024 17:09	AR\AJ	Ok,M
9	PB164378BS	PS028038.D	24 Oct 2024 17:33	AR\AJ	Ok
10	PB164261TB	PS028039.D	24 Oct 2024 17:57	AR\AJ	Not Ok
11	PB164336TB	PS028040.D	24 Oct 2024 18:21	AR\AJ	Not Ok
12	P4397-06	PS028041.D	24 Oct 2024 18:45	AR\AJ	Ok,M
13	P4397-06MS	PS028042.D	24 Oct 2024 19:09	AR\AJ	Ok,M
14	P4397-06MSD	PS028043.D	24 Oct 2024 19:32	AR\AJ	Ok,M
15	P4460-04	PS028044.D	24 Oct 2024 19:56	AR\AJ	Not Ok
16	P4462-02	PS028045.D	24 Oct 2024 20:20	AR\AJ	Ok
17	I.BLK	PS028046.D	24 Oct 2024 20:44	AR\AJ	Ok,M
18	HSTDCCC750	PS028047.D	24 Oct 2024 21:08	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS102524

Review By	Abdul	Review On	10/28/2024 9:31:35 AM
Supervise By	Ankita	Supervise On	10/28/2024 10:48:49 AM
SubDirectory	PS102524	HP Acquire Method	HP Processing Method ps102324 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC Internal Standard/PEM	PP23462		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23469		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS028048.D	25 Oct 2024 11:18	AR\AJ	Ok
2	I.BLK	PS028049.D	25 Oct 2024 11:42	AR\AJ	Ok
3	HSTDCCC750	PS028050.D	25 Oct 2024 12:32	AR\AJ	Ok
4	P4485-01	PS028051.D	25 Oct 2024 16:25	AR\AJ	Ok,M
5	P4460-04	PS028052.D	25 Oct 2024 16:49	AR\AJ	Ok,M
6	PB164411BL	PS028053.D	25 Oct 2024 17:13	AR\AJ	Ok,M
7	PB164411BS	PS028054.D	25 Oct 2024 17:41	AR\AJ	Ok
8	I.BLK	PS028055.D	25 Oct 2024 18:05	AR\AJ	Ok
9	HSTDCCC750	PS028056.D	25 Oct 2024 18:28	AR\AJ	Ok
10	P4487-01	PS028057.D	25 Oct 2024 19:16	AR\AJ	Ok,M
11	P4487-01MS	PS028058.D	25 Oct 2024 19:40	AR\AJ	Ok,M
12	P4487-01MSD	PS028059.D	25 Oct 2024 20:04	AR\AJ	Ok,M
13	P4487-05	PS028060.D	25 Oct 2024 20:28	AR\AJ	Ok,M
14	P4508-01	PS028061.D	25 Oct 2024 20:52	AR\AJ	Ok
15	P4508-05	PS028062.D	25 Oct 2024 21:16	AR\AJ	Ok,M
16	P4508-09	PS028063.D	25 Oct 2024 21:39	AR\AJ	Ok
17	P4509-01	PS028064.D	25 Oct 2024 22:03	AR\AJ	Ok,M
18	P4545-01	PS028065.D	25 Oct 2024 22:27	AR\AJ	Ok
19	P4547-01	PS028066.D	25 Oct 2024 22:51	AR\AJ	Ok,M
20	P4547-05	PS028067.D	25 Oct 2024 23:15	AR\AJ	Ok
21	I.BLK	PS028068.D	25 Oct 2024 23:39	AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QCBatch ID # PS102524

Review By	Abdul	Review On	10/28/2024 9:31:35 AM
Supervise By	Ankita	Supervise On	10/28/2024 10:48:49 AM
SubDirectory	PS102524	HP Acquire Method	HP Processing Method ps102324 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC	PP23462		
Internal Standard/PEM			
ICV/I.BLK	PP23469		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	HSTDCCC750	PS028069.D	26 Oct 2024 00:03	AR\AJ	Ok,M
----	------------	------------	-------------------	-------	------

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS102824

Review By	Abdul	Review On	10/29/2024 8:52:20 AM
Supervise By	Ankita	Supervise On	10/29/2024 8:52:48 AM
SubDirectory	PS102824	HP Acquire Method	HP Processing Method ps102324 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	P23457,PP23458,PP23459,PP23460,PP23461,PP23462,PP23467,PP23468,PP23469		
CCC Internal Standard/PEM	PP23462		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23469		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS028070.D	28 Oct 2024 09:44	AR\AJ	Ok
2	I.BLK	PS028071.D	28 Oct 2024 10:08	AR\AJ	Ok,M
3	HSTDCCC750	PS028072.D	28 Oct 2024 10:32	AR\AJ	Ok,M
4	PB164261TB	PS028073.D	28 Oct 2024 13:11	AR\AJ	Ok,M
5	PB164336TB	PS028074.D	28 Oct 2024 13:35	AR\AJ	Ok,M
6	I.BLK	PS028075.D	28 Oct 2024 13:59	AR\AJ	Ok
7	HSTDCCC750	PS028076.D	28 Oct 2024 17:17	AR\AJ	Ok

M : Manual Integration