

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/21/2024 10/21/2024

Client Sample No.: CCAL01 Date Analyzed: 10/24/2024

Lab Sample No.: PSTDCCC050 Data File : PL092591.D Time Analyzed: 10:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.057	8.955	9.155	51.250	50.000	2.5
Endrin	6.576	6.474	6.674	44.700	50.000	-10.6
gamma-BHC (Lindane)	4.330	4.227	4.427	48.930	50.000	-2.1
Heptachlor	4.918	4.815	5.015	47.520	50.000	-5.0
Heptachlor epoxide	5.684	5.583	5.783	46.660	50.000	-6.7
Methoxychlor	7.501	7.400	7.600	51.960	50.000	3.9
Tetrachloro-m-xylene	3.541	3.438	3.638	51.820	50.000	3.6

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4460 SAS No.: P4460 SDG NO.: P4460

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/21/2024 10/21/2024

Client Sample No.: CCAL01 Date Analyzed: 10/24/2024

Lab Sample No.: PSTDCCC050 Data File : PL092591.D Time Analyzed: 10:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.918	7.817	8.017	58.240	50.000	16.5
Endrin	5.643	5.541	5.741	53.080	50.000	6.2
gamma-BHC (Lindane)	3.612	3.510	3.710	56.090	50.000	12.2
Heptachlor	3.951	3.849	4.049	56.020	50.000	12.0
Heptachlor epoxide	4.733	4.631	4.831	55.710	50.000	11.4
Methoxychlor	6.616	6.515	6.715	58.450	50.000	16.9
Tetrachloro-m-xylene	2.779	2.677	2.877	56.670	50.000	13.3